

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Tully Environmental, Inc **SDG No.:** Q2349
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q2349 **SAS No.:** Q2349
ICS Source: EPA **Instrument ID:** P5

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Copper	-11.5	2.0	575	-18	22	06/19/2025	15:10	LB136210
	Iron	104000	100000	104	85600	116500	06/19/2025	15:10	LB136210
	Lead	6.98			-12	12	06/19/2025	15:10	LB136210
ICSAB01	Copper	491	510	96	434	588	06/19/2025	15:16	LB136210
	Iron	103000	99000	104	84400	114500	06/19/2025	15:16	LB136210
	Lead	51.1	49.0	104	37	61	06/19/2025	15:16	LB136210
ICSA	Copper	-18.5	2.0	925	-18	22	06/19/2025	15:23	LB136210
	Iron	104000	100000	104	85600	116500	06/19/2025	15:23	LB136210
	Lead	31.5			-12	12	06/19/2025	15:23	LB136210
ICSAB	Copper	464	510	91	434	588	06/19/2025	15:28	LB136210
	Iron	103000	99000	104	84400	114500	06/19/2025	15:28	LB136210
	Lead	48.9	49.0	100	37	61	06/19/2025	15:28	LB136210