

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>Tully Environmental, Inc</u>	level: <u>low</u>	sdg no.: <u>Q2349</u>
contract: <u>TULL01</u>	lab code: <u>CHEM</u>	case no.: <u>Q2349</u> sas no.: <u>Q2349</u>
matrix: <u>Water</u>	sample id: <u>Q2348-02</u>	client id: <u>002-35TH-AVE(MAY)MS</u>
Percent Solids for Sample: NA	Spiked ID: Q2348-02MS	Percent Solids for Spike Sample: NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	213		64.4		150	99		P
Iron	ug/L	75 - 125	16100		14100		1500	133		P
Lead	ug/L	75 - 125	539		66.9		500	94		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>Tully Environmental, Inc</u>	level: <u>low</u>	sdg no.: <u>Q2349</u>
contract: <u>TULL01</u>	lab code: <u>CHEM</u>	case no.: <u>Q2349</u> sas no.: <u>Q2349</u>
matrix: <u>Water</u>	sample id: <u>Q2348-02</u>	client id: <u>002-35TH-AVE(MAY)MSD</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q2348-02MSD</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	233		64.4		150	112		P
Iron	ug/L	75 - 125	17500		14100		1500	225		P
Lead	ug/L	75 - 125	574		66.9		500	102		P