

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tully Environmental, Inc

SDG No.: Q2349

Contract: TULL01

Lab Code: CHEM

Case No.: Q2349

SAS No.: Q2349

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168539BS							
Copper	ug/L	150	165		110	85 - 115	P
Iron	ug/L	1500	1610		107	85 - 115	P
Lead	ug/L	500	477		95	85 - 115	P