

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TULL01

Lab Code: CHEM Case No.: Q2349 SAS No.: Q2349 SDG No.: Q2349

Instrument ID: MSVOA_N Calibration Date/Time: 06/25/2025 09:20

Lab File ID: VN087163.D Init. Calib. Date(s): 06/25/2025 06/25/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 08:59 10:24

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF020	MIN RRF	%D	MAX%D
Benzene	1.481	1.309		-11.61	
Toluene	1.694	1.550		-8.5	
Ethyl Benzene	1.855	1.649		-11.1	
m/p-Xylenes	0.711	0.636		-10.55	
o-Xylene	0.678	0.604		-10.91	
1,2-Dichloroethane-d4	2.260	2.309		2.17	
Toluene-d8	1.347	1.399		3.86	
4-Bromofluorobenzene	0.463	0.448		-3.24	

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.