

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TULL01
 Lab Code: CHEM Case No.: Q2349 SAS No.: Q2349 SDG NO.: Q2349
 Lab File ID: VN087163.D Date Analyzed: 06/25/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:20
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	35021	7.82	196036	9.11	185687	11.87
UPPER LIMIT	70042	8.324	392072	9.606	371374	12.365
LOWER LIMIT	17510.5	7.324	98018	8.606	92843.5	11.365
EPA SAMPLE NO.						
001-WILLETS-PT-BLVD (JUNE)	35391	7.82	192737	9.11	179463	11.87
001-WILLETS-PT-BLVD (JUNE) DL	36259	7.82	194067	9.11	178157	11.87
002-35TH-AVE (JUNE)	37564	7.82	195128	9.11	182621	11.87
002-35TH-AVE (JUNE) DL	36435	7.82	187691	9.11	177573	11.87
VN0625WBL01	33553	7.82	188572	9.11	167117	11.87
VN0625WBS01	40080	7.82	221679	9.11	199254	11.87
VN0625WBSD01	37318	7.82	209203	9.11	180451	11.87

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.