

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TULL01
 Lab Code: CHEM Case No.: Q2349 SAS No.: Q2349 SDG NO.: Q2349
 Lab File ID: VN087163.D Date Analyzed: 06/25/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:20
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	0	0				
UPPER LIMIT	0					
LOWER LIMIT	0					
EPA SAMPLE NO.						
001-WILLETS-PT-BLVD (JUNE)	0	0.00				
001-WILLETS-PT-BLVD (JUNE) DL	0	0.00				
002-35TH-AVE (JUNE)	0	0.00				
002-35TH-AVE (JUNE) DL	0	0.00				
VN0625WBL01	0	0.00				
VN0625WBS01	0	0.00				
VN0625WBSD01	0	0.00				

IS4 =

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.