

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: WATE03
Lab Code: CHEM Case No.: Q2359 SAS No.: Q2359 SDG No.: Q2359
Instrument ID: MSVOA_L Calibration Date(s): 05/29/2025 05/29/2025
Heated Purge: (Y/N) N Calibration Time(s): 11:09 14:24
GC Column: RTX-1 ID: 0.32 (mm)

LAB FILE ID:	RRF010 = VL042579.D			RRF002 = VL042580.D		RRF001 = VL042581.D		RRF0.5 = VL042582.D		RRF0.1 = VL042583.D		RRF.03 = VL042584.D	
COMPOUND	RRF010	RRF002	RRF001	RRF0.5	RRF0.1	RRF.03	RRF	% RSD					
Dichlorodifluoromethane	1.135	1.300	1.353	1.257			1.229	8.8					
Chloromethane	0.461	0.507	0.529	0.502			0.487	7.9					
Vinyl Chloride	0.437	0.468	0.487	0.529	0.465	0.675	0.496	17.7					
Bromomethane	0.216	0.230	0.251	0.310			0.241	17.7					
Chloroethane	0.175	0.198	0.212	0.214			0.193	11.6					
Tetrahydrofuran	0.388	0.426	0.446	0.445			0.417	7.3					
Trichlorofluoromethane	1.071	1.202	1.209	1.193			1.144	7					
1,1,2-Trichlorotrifluoroethane	0.838	0.951	0.980	0.903			0.898	7.8					
Dichlorotetrafluoroethane	0.948	1.075	1.098	1.099			1.024	9.1					
tert-Butyl alcohol	1.237	1.279	1.334	1.337			1.255	8.1					
Heptane	1.777	1.981	2.082	2.079			1.921	9.4					
1,1-Dichloroethene	0.377	0.437	0.451	0.451			0.416	10.1					
Acetone	0.928	1.299	1.484	1.404			1.205	22.4					
Carbon Disulfide	1.069	1.198	1.267	1.174			1.150	8.2					
Methyl tert-Butyl Ether	0.550	0.621	0.693	0.664			0.608	12.6					
Methylene Chloride	0.326	0.418	0.457	0.473			0.398	18.6					
trans-1,2-Dichloroethene	0.427	0.466	0.517	0.496			0.465	9.2					
1,1-Dichloroethane	0.903	1.004	1.046	0.996			0.957	9					
Cyclohexane	1.146	1.288	1.363	1.318			1.242	9.4					
2-Butanone	0.655	0.746	0.754	0.760			0.713	7.7					
Carbon Tetrachloride	0.603	0.670	0.684	0.665	0.706	0.681	0.660	5.8					
cis-1,2-Dichloroethene	1.271	1.465	1.461	1.419			1.370	8					
Chloroform	1.755	2.034	2.120	2.099			1.948	9.7					
1,1,1-Trichloroethane	1.794	2.029	2.043	2.004	2.083	2.207	1.990	7.9					
2,2,4-Trimethylpentane	1.522	1.765	1.830	1.811			1.684	9.7					
Benzene	0.900	1.027	1.052	1.051			0.985	8.2					
1,2-Dichloroethane	0.448	0.513	0.509	0.518			0.489	6.7					
Trichloroethene	0.376	0.438	0.455	0.468	0.432	0.389	0.419	9.3					
1,2-Dichloropropane	0.329	0.373	0.374	0.415			0.364	9.9					
Bromodichloromethane	0.656	0.721	0.737	0.702			0.695	5.2					

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name:	<u>CHEMTECH</u>	Contract:	<u>WATE03</u>
Lab Code:	<u>CHEM</u>	Case No.:	<u>Q2359</u>
Instrument ID:	<u>MSVOA_L</u>	SAS No.:	<u>Q2359</u>
Heated Purge: (Y/N)	<u>N</u>	SDG No.:	<u>Q2359</u>
GC Column:	<u>RTX-1</u>	Calibration Date(s):	<u>05/29/2025</u>
ID:	<u>0.32</u> (mm)	Calibration Time(s):	<u>11:09</u>
			<u>14:24</u>

LAB FILE ID:		RRF010 = VL042579.D		RRF002 = VL042580.D		RRF001 = VL042581.D		
		RRF0.5 = VL042582.D		RRF0.1 = VL042583.D		RRF.03 = VL042584.D		
COMPOUND	RRF010	RRF002	RRF001	RRF0.5	RRF0.1	RRF.03	RRF	% RSD
4-Methyl-2-Pentanone	0.958	1.002	1.030	1.036			0.989	5
Toluene	1.054	1.197	1.190	1.210			1.138	7.4
t-1,3-Dichloropropene	0.424	0.456	0.455	0.433			0.437	4
cis-1,3-Dichloropropene	0.527	0.555	0.568	0.567			0.548	4
1,1,2-Trichloroethane	0.343	0.394	0.401	0.406			0.377	8.7
Dibromochloromethane	0.570	0.635	0.642	0.600			0.603	5.6
1,2-Dibromoethane	0.516	0.576	0.590	0.578	0.575		0.558	6.1
Tetrachloroethene	0.333	0.390	0.383	0.391	0.407	0.351	0.368	8.9
Chlorobenzene	0.868	1.014	1.028	1.081			0.971	10.2
Ethyl Benzene	1.547	1.833	1.786	1.770			1.696	8.2
m/p-Xylene	1.212	1.431	1.412	1.411			1.335	8.5
o-Xylene	1.187	1.441	1.410	1.430			1.331	9.9
Styrene	0.590	0.678	0.658	0.626			0.629	6.1
Bromoform	0.508	0.553	0.519	0.524			0.521	3.8
1,1,2,2-Tetrachloroethane	0.723	0.855	0.861	0.879	0.828	1.054	0.847	13.2
2-Chlorotoluene	1.343	1.596	1.620	1.603			1.500	9.7
1,3,5-Trimethylbenzene	1.235	1.475	1.482	1.441			1.374	9.3
1,2,4-Trimethylbenzene	1.335	1.666	1.675	1.640			1.524	12.3
1,3-Dichlorobenzene	0.839	1.040	1.034	1.031			0.957	11.1
1,4-Dichlorobenzene	0.849	1.037	1.040	0.998			0.955	10.2
1,2-Dichlorobenzene	0.798	1.027	1.008	1.019			0.929	13.1
1,2,4-Trichlorobenzene	0.735	0.944	0.889	0.716			0.802	13.3
Hexachloro-1,3-Butadiene	0.607	0.880	0.893	0.872			0.769	20.1
1,3-Butadiene	0.491	0.597	0.636	0.649			0.571	14.1
Naphthalene	1.384	1.785	1.634	1.366	1.088		1.437	16.9
4-Ethyltoluene	1.481	1.823	1.797	1.704			1.660	9.9
1-Bromo-4-Fluorobenzene	0.744	0.741	0.751	0.755	0.751	0.752	0.750	0.7
Hexane	1.350	1.526	1.638	1.690			1.498	11.8
Allyl Chloride	0.789	0.866	0.921	0.924			0.850	9.2
1,4-Dioxane	151.601	168.112	178.377	171.890			162.953	8.7

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: WATE03
 Lab Code: CHEM Case No.: Q2359 SAS No.: Q2359 SDG No.: Q2359
 Instrument ID: MSVOA_L Calibration Date(s): 05/29/2025 05/29/2025
 Heated Purge: (Y/N) N Calibration Time(s): 11:09 14:24
 GC Column: RTX-1 ID: 0.32 (mm)

LAB FILE ID:								
RRF010 = VL042579.D			RRF002 = VL042580.D			RRF001 = VL042581.D		
RRF0.5 = VL042582.D			RRF0.1 = VL042583.D			RRF.03 = VL042584.D		
COMPOUND	RRF010	RRF002	RRF001	RRF0.5	RRF0.1	RRF.03	RRF	% RSD
Methyl Methacrylate	0.366	0.406	0.384	0.398			0.382	5.4

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.