

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: WATE03
Lab Code: CHEM Case No.: Q2359 SAS No.: Q2359 SDG NO.: Q2359
Lab File ID: VL042646.D BFB Injection Date: 06/19/2025
Instrument ID: MSVOA_L BFB Injection Time: 08:10
GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	8.0 - 40.0% of mass 95	22.8
75	30.0 - 66.0% of mass 95	55.3
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.4
173	Less than 2.0% of mass 174	0.8 (1.3) 1
174	50.0 - 120.0% of mass 95	59.4
175	4.0 - 9.0% of mass 174	4 (6.7) 1
176	93.0 - 101.0% of mass 174	57.6 (97) 1
177	5.0 - 9.0% of mass 176	3.7 (6.3) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC010	VSTDCCC010	VL042647.D	06/19/2025	08:56
VL0619ABL01	VL0619ABL01	VL042648.D	06/19/2025	09:42
VL0619ABS01	VL0619ABS01	VL042649.D	06/19/2025	10:26
SS-1DL	Q2359-01DL	VL042650.D	06/19/2025	11:09
SS-1	Q2359-01	VL042651.D	06/19/2025	11:58
SS-1DUP	Q2359-01DUP	VL042652.D	06/19/2025	12:34