

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2361 SAS No.: Q2361 SDG NO.: Q2361

EPA Sample No.: SSTDICCC0.4 Date Analyzed: 06/26/2025

Lab File ID: BN037388.D Time Analyzed: 11:53

Instrument ID: BNA\_N GC Column: ZB-GR ID: 0.25 (mm)

|                     | IS1 (DCB)<br>AREA # | RT #  | IS2 (NPT)<br>AREA # | RT #  | IS3 (ANT)<br>AREA # | RT #   |
|---------------------|---------------------|-------|---------------------|-------|---------------------|--------|
| 12 HOUR STD         | 2117                | 7.568 | 4565                | 10.34 | 3048                | 14.21  |
| UPPER LIMIT         | 4234                | 8.068 | 9130                | 10.84 | 6096                | 14.713 |
| LOWER LIMIT         | 1058.5              | 7.068 | 2282.5              | 9.84  | 1524                | 13.713 |
| EPA SAMPLE NO.      |                     |       |                     |       |                     |        |
| 01 TT205S1-20250617 | 1481                | 7.57  | 3305                | 10.34 | 2183                | 14.21  |

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.

\* Values outside of QC limits.