

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2361 SAS No.: Q2361 SDG NO.: Q2361

EPA Sample No.: SSTDICCC0.4 Date Analyzed: 06/26/2025

Lab File ID: BN037388.D Time Analyzed: 11:53

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	6284	16.959	5612	21.162	5950	23.345
UPPER LIMIT	12568	17.459	11224	21.662	11900	23.845
LOWER LIMIT	3142	16.459	2806	20.662	2975	22.845
EPA SAMPLE NO.						
01 TT205S1-20250617	4358	16.97	4233	21.16	3947	23.35

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.