

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q2361 SAS No.: Q2361 SDG No.: Q2361

Instrument ID: MSVOA_X Calibration Date/Time: 06/19/2025 09:43

Lab File ID: VX046758.D Init. Calib. Date(s): 06/17/2025 06/17/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:19 17:18

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.534	0.565		5.8	20
Chloromethane	0.576	0.595	0.1	3.3	20
Vinyl Chloride	0.614	0.641		4.4	20
Bromomethane	0.346	0.384		10.98	20
Chloroethane	0.372	0.387		4.03	20
Trichlorofluoromethane	0.933	0.957		2.57	20
1,1,2-Trichlorotrifluoroethane	0.572	0.592		3.5	20
1,1-Dichloroethene	0.552	0.572		3.62	20
Acetone	0.231	0.185		-19.91	20
Carbon Disulfide	1.668	1.595		-4.38	20
Methyl tert-butyl Ether	1.720	1.569		-8.78	20
Methyl Acetate	0.586	0.480		-18.09	20
Methylene Chloride	0.641	0.618		-3.59	20
trans-1,2-Dichloroethene	0.591	0.587		-0.68	20
1,1-Dichloroethane	1.092	1.081	0.1	-1.01	20
Cyclohexane	1.036	1.018		-1.74	20
2-Butanone	0.326	0.263		-19.33	20
Carbon Tetrachloride	0.518	0.515		-0.58	20
cis-1,2-Dichloroethene	0.694	0.686		-1.15	20
Bromochloromethane	0.495	0.476		-3.84	20
Chloroform	1.092	1.093		0.09	20
1,1,1-Trichloroethane	0.958	0.939		-1.98	20
Methylcyclohexane	0.628	0.627		-0.32	20
Benzene	1.413	1.415		0.14	20
1,2-Dichloroethane	0.497	0.481		-3.22	20
Trichloroethene	0.360	0.357		-0.83	20
1,2-Dichloropropane	0.350	0.353		0.86	20
Bromodichloromethane	0.514	0.521		1.36	20
4-Methyl-2-Pentanone	0.411	0.339		-17.52	20
Toluene	0.876	0.877		0.11	20
t-1,3-Dichloropropene	0.477	0.481		0.84	20
cis-1,3-Dichloropropene	0.541	0.545		0.74	20
1,1,2-Trichloroethane	0.327	0.311		-4.89	20
2-Hexanone	0.284	0.229		-19.37	20
Dibromochloromethane	0.385	0.377		-2.08	20
1,2-Dibromoethane	0.332	0.317		-4.52	20
Tetrachloroethene	0.346	0.344		-0.58	20
Chlorobenzene	1.104	1.089	0.3	-1.36	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q2361 SAS No.: Q2361 SDG No.: Q2361

Instrument ID: MSVOA_X Calibration Date/Time: 06/19/2025 09:43

Lab File ID: VX046758.D Init. Calib. Date(s): 06/17/2025 06/17/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:19 17:18

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.929	1.918		-0.57	20
m/p-Xylenes	0.719	0.724		0.69	20
o-Xylene	0.673	0.687		2.08	20
Styrene	1.179	1.192		1.1	20
Bromoform	0.282	0.263	0.1	-6.74	20
Isopropylbenzene	3.682	3.747		1.76	20
1,1,2,2-Tetrachloroethane	1.028	0.933	0.3	-9.24	20
1,3-Dichlorobenzene	1.728	1.670		-3.36	20
1,4-Dichlorobenzene	1.775	1.693		-4.62	20
1,2-Dichlorobenzene	1.615	1.598		-1.05	20
1,2-Dibromo-3-Chloropropane	0.203	0.164		-19.21	20
1,2,4-Trichlorobenzene	1.148	1.111		-3.22	20
1,2,3-Trichlorobenzene	1.098	1.064		-3.1	20
1,2-Dichloroethane-d4	0.692	0.639		-7.66	20
Dibromofluoromethane	0.331	0.327		-1.21	20
Toluene-d8	1.189	1.166		-1.93	20
4-Bromofluorobenzene	0.443	0.433		-2.26	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.