

LAB CHRONICLE

OrderID:	Q2361	OrderDate:	6/18/2025 3:09:00 PM
Client:	AECOM Technical Services, Inc.	Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258
Contact:	Eleanor Vivadou	Location:	D51,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2361-01	TT205S1-20250617	Water	VOCMS Group1	8260-Low	06/17/25		06/19/25	06/18/25



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q2361
Client: AECOM Technical Services, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID:	TT205S1-20250617								
Q2361-01	TT205S1-20250617	Water	Trichloroethene	7.00		0.090	0.75	1.00	ug/L
			Total Voc :	7.00					
			Total Concentration:	7.00					



SAMPLE DATA

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	06/17/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	06/18/25	
Client Sample ID:	TT205S1-20250617		SDG No.:	Q2361	
Lab Sample ID:	Q2361-01		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046764.D	1		06/19/25 12:12	VX061925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	7.00		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	06/17/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	06/18/25	
Client Sample ID:	TT205S1-20250617		SDG No.:	Q2361	
Lab Sample ID:	Q2361-01		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046764.D	1		06/19/25 12:12	VX061925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	47.5		81 - 118		95%	SPK: 50
1868-53-7	Dibromofluoromethane	49.1		80 - 119		98%	SPK: 50
2037-26-5	Toluene-d8	50.5		89 - 112		101%	SPK: 50
460-00-4	4-Bromofluorobenzene	48.7		85 - 114		97%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	120000	5.562				
540-36-3	1,4-Difluorobenzene	201000	6.769				
3114-55-4	Chlorobenzene-d5	178000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	84200	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	06/17/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	06/18/25	
Client Sample ID:	TT205S1-20250617		SDG No.:	Q2361	
Lab Sample ID:	Q2361-01		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046764.D	1		06/19/25 12:12	VX061925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SDG No.: Q2361

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2361-01	TT205S1-20250617	1,2-Dichloroethane-d4	50	47.5	95	81	118
		Dibromofluoromethane	50	49.1	98	80	119
		Toluene-d8	50	50.5	101	89	112
		4-Bromofluorobenzene	50	48.7	97	85	114
VX0619WBL01	VX0619WBL01	1,2-Dichloroethane-d4	50	48.3	97	81	118
		Dibromofluoromethane	50	49.0	98	80	119
		Toluene-d8	50	50.0	100	89	112
		4-Bromofluorobenzene	50	49.4	99	85	114
VX0619WBS01	VX0619WBS01	1,2-Dichloroethane-d4	50	49.8	100	81	118
		Dibromofluoromethane	50	52.0	104	80	119
		Toluene-d8	50	52.3	105	89	112
		4-Bromofluorobenzene	50	54.1	108	85	114
VX0619WBSD0	VX0619WBSD01	1,2-Dichloroethane-d4	50	51.5	103	81	118
		Dibromofluoromethane	50	52.8	106	80	119
		Toluene-d8	50	52.7	105	89	112
		4-Bromofluorobenzene	50	54.9	110	85	114

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2361

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Datafile : VX046762.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0619WBS01	Dichlorodifluoromethane	20	19.1	ug/L	96			32	152	
	Chloromethane	20	18.7	ug/L	94			50	139	
	Vinyl chloride	20	18.9	ug/L	95			58	137	
	Bromomethane	20	21.2	ug/L	106			53	141	
	Chloroethane	20	19.9	ug/L	100			60	138	
	Trichlorofluoromethane	20	18.4	ug/L	92			65	141	
	1,1,2-Trichlorotrifluoroethane	20	18.5	ug/L	93			70	136	
	1,1-Dichloroethene	20	18.6	ug/L	93			71	131	
	Acetone	100	74.1	ug/L	74			39	160	
	Carbon disulfide	20	17.0	ug/L	85			64	133	
	Methyl tert-butyl Ether	20	17.0	ug/L	85			71	124	
	Methyl Acetate	20	16.0	ug/L	80			56	136	
	Methylene Chloride	20	18.1	ug/L	91			74	124	
	trans-1,2-Dichloroethene	20	18.3	ug/L	92			75	124	
	1,1-Dichloroethane	20	18.5	ug/L	93			77	125	
	Cyclohexane	20	18.8	ug/L	94			71	130	
	2-Butanone	100	78.9	ug/L	79			56	143	
	Carbon Tetrachloride	20	18.0	ug/L	90			72	136	
	cis-1,2-Dichloroethene	20	18.6	ug/L	93			78	123	
	Bromochloromethane	20	20.8	ug/L	104			78	123	
	Chloroform	20	19.0	ug/L	95			79	124	
	1,1,1-Trichloroethane	20	17.7	ug/L	89			74	131	
	Methylcyclohexane	20	18.5	ug/L	93			72	132	
	Benzene	20	18.9	ug/L	95			79	120	
	1,2-Dichloroethane	20	18.5	ug/L	93			73	128	
	Trichloroethene	20	18.4	ug/L	92			79	123	
	1,2-Dichloropropane	20	18.4	ug/L	92			78	122	
	Bromodichloromethane	20	18.9	ug/L	95			79	125	
	4-Methyl-2-Pentanone	100	81.6	ug/L	82			67	130	
	Toluene	20	19.0	ug/L	95			80	121	
	t-1,3-Dichloropropene	20	18.0	ug/L	90			73	127	
	cis-1,3-Dichloropropene	20	18.2	ug/L	91			75	124	
	1,1,2-Trichloroethane	20	18.5	ug/L	93			80	119	
	2-Hexanone	100	78.9	ug/L	79			57	139	
	Dibromochloromethane	20	18.4	ug/L	92			74	126	
	1,2-Dibromoethane	20	18.0	ug/L	90			77	121	
	Tetrachloroethene	20	17.7	ug/L	89			74	129	
	Chlorobenzene	20	18.0	ug/L	90			82	118	
	Ethyl Benzene	20	18.0	ug/L	90			79	121	
	m/p-Xylenes	40	37.2	ug/L	93			80	121	
	o-Xylene	20	18.6	ug/L	93			78	122	
	Styrene	20	18.3	ug/L	92			78	123	
	Bromoform	20	16.7	ug/L	84			66	130	
	Isopropylbenzene	20	18.1	ug/L	91			72	131	
	1,1,2,2-Tetrachloroethane	20	17.3	ug/L	86			71	121	
	1,3-Dichlorobenzene	20	17.8	ug/L	89			80	119	
	1,4-Dichlorobenzene	20	17.5	ug/L	88			79	118	
	1,2-Dichlorobenzene	20	18.1	ug/L	91			80	119	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2361

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Datafile : VX046762.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VX0619WBS01	1,2-Dibromo-3-Chloropropane	20	15.1	ug/L	76			62	128	
	1,2,4-Trichlorobenzene	20	17.1	ug/L	86			69	130	
	1,2,3-Trichlorobenzene	20	17.6	ug/L	88			69	129	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2361

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Datafile : VX046763.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0619WBSD01	Dichlorodifluoromethane	20	19.8	ug/L	99	3		32	152	20
	Chloromethane	20	19.5	ug/L	98	4		50	139	20
	Vinyl chloride	20	19.7	ug/L	99	4		58	137	20
	Bromomethane	20	22.2	ug/L	111	5		53	141	20
	Chloroethane	20	20.6	ug/L	103	3		60	138	20
	Trichlorofluoromethane	20	19.3	ug/L	97	5		65	141	20
	1,1,2-Trichlorotrifluoroethane	20	19.7	ug/L	99	6		70	136	20
	1,1-Dichloroethene	20	19.5	ug/L	98	5		71	131	20
	Acetone	100	81.5	ug/L	82	10		39	160	20
	Carbon disulfide	20	18.0	ug/L	90	6		64	133	20
	Methyl tert-butyl Ether	20	18.7	ug/L	94	10		71	124	20
	Methyl Acetate	20	17.8	ug/L	89	11		56	136	20
	Methylene Chloride	20	19.3	ug/L	97	6		74	124	20
	trans-1,2-Dichloroethene	20	19.2	ug/L	96	4		75	124	20
	1,1-Dichloroethane	20	19.4	ug/L	97	4		77	125	20
	Cyclohexane	20	19.7	ug/L	99	5		71	130	20
	2-Butanone	100	87.0	ug/L	87	10		56	143	20
	Carbon Tetrachloride	20	19.1	ug/L	96	6		72	136	20
	cis-1,2-Dichloroethene	20	19.4	ug/L	97	4		78	123	20
	Bromochloromethane	20	23.1	ug/L	116	11		78	123	20
	Chloroform	20	20.5	ug/L	103	8		79	124	20
	1,1,1-Trichloroethane	20	19.3	ug/L	97	9		74	131	20
	Methylcyclohexane	20	19.0	ug/L	95	2		72	132	20
	Benzene	20	20.0	ug/L	100	5		79	120	20
	1,2-Dichloroethane	20	20.1	ug/L	101	8		73	128	20
	Trichloroethene	20	19.3	ug/L	97	5		79	123	20
	1,2-Dichloropropane	20	20.3	ug/L	102	10		78	122	20
	Bromodichloromethane	20	20.1	ug/L	101	6		79	125	20
	4-Methyl-2-Pentanone	100	91.8	ug/L	92	11		67	130	20
	Toluene	20	20.3	ug/L	102	7		80	121	20
	t-1,3-Dichloropropene	20	19.5	ug/L	98	9		73	127	20
	cis-1,3-Dichloropropene	20	19.9	ug/L	100	9		75	124	20
	1,1,2-Trichloroethane	20	20.4	ug/L	102	9		80	119	20
	2-Hexanone	100	88.8	ug/L	89	12		57	139	20
	Dibromochloromethane	20	19.9	ug/L	100	8		74	126	20
	1,2-Dibromoethane	20	20.1	ug/L	101	12		77	121	20
	Tetrachloroethene	20	18.8	ug/L	94	5		74	129	20
	Chlorobenzene	20	19.4	ug/L	97	7		82	118	20
	Ethyl Benzene	20	19.1	ug/L	96	6		79	121	20
	m/p-Xylenes	40	38.8	ug/L	97	4		80	121	20
	o-Xylene	20	19.8	ug/L	99	6		78	122	20
	Styrene	20	19.7	ug/L	99	7		78	123	20
	Bromoform	20	18.4	ug/L	92	9		66	130	20
	Isopropylbenzene	20	18.6	ug/L	93	2		72	131	20
	1,1,2,2-Tetrachloroethane	20	18.4	ug/L	92	7		71	121	20
	1,3-Dichlorobenzene	20	18.6	ug/L	93	4		80	119	20
	1,4-Dichlorobenzene	20	17.8	ug/L	89	1		79	118	20
	1,2-Dichlorobenzene	20	19.0	ug/L	95	4		80	119	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2361

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Datafile : VX046763.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0619WBSD01	1,2-Dibromo-3-Chloropropane	20	16.1	ug/L	81	6		62	128	20
	1,2,4-Trichlorobenzene	20	18.3	ug/L	92	7		69	130	20
	1,2,3-Trichlorobenzene	20	18.2	ug/L	91	3		69	129	20

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VX0619WBL01

Lab Name: CHEMTECH

Contract: AECO15

Lab Code: CHEM Case No.: Q2361

SAS No.: Q2361 SDG NO.: Q2361

Lab File ID: VX046760.D

Lab Sample ID: VX0619WBL01

Date Analyzed: 06/19/2025

Time Analyzed: 10:33

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_X

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VX0619WBS01	VX0619WBS01	VX046762.D	06/19/2025
VX0619WBSD01	VX0619WBSD01	VX046763.D	06/19/2025
TT205S1-20250617	Q2361-01	VX046764.D	06/19/2025

COMMENTS: _____

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: Q2361 SAS No.: Q2361 SDG NO.: Q2361
Lab File ID: VX046715.D BFB Injection Date: 06/17/2025
Instrument ID: MSVOA_X BFB Injection Time: 08:46
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	18.7
75	30.0 - 60.0% of mass 95	50.7
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.5
173	Less than 2.0% of mass 174	0.5 (0.7) 1
174	50.0 - 100.0% of mass 95	74.8
175	5.0 - 9.0% of mass 174	5.5 (7.4) 1
176	95.0 - 101.0% of mass 174	72 (96.2) 1
177	5.0 - 9.0% of mass 176	4.3 (6) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC005	VSTDIC005	VX046718.D	06/17/2025	11:19
VSTDIC020	VSTDIC020	VX046719.D	06/17/2025	13:59
VSTDICCC050	VSTDICCC050	VX046720.D	06/17/2025	14:20
VSTDIC100	VSTDIC100	VX046721.D	06/17/2025	14:41
VSTDIC150	VSTDIC150	VX046722.D	06/17/2025	15:02
VSTDIC001	VSTDIC001	VX046725.D	06/17/2025	17:18

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: Q2361 SAS No.: Q2361 SDG NO.: Q2361
Lab File ID: VX046757.D BFB Injection Date: 06/19/2025
Instrument ID: MSVOA_X BFB Injection Time: 08:42
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	18.8
75	30.0 - 60.0% of mass 95	50.7
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.3
173	Less than 2.0% of mass 174	0.7 (1) 1
174	50.0 - 100.0% of mass 95	72.6
175	5.0 - 9.0% of mass 174	5.7 (7.8) 1
176	95.0 - 101.0% of mass 174	71.1 (97.9) 1
177	5.0 - 9.0% of mass 176	4.5 (6.3) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX046758.D	06/19/2025	09:43
VX0619WBL01	VX0619WBL01	VX046760.D	06/19/2025	10:33
VX0619WBS01	VX0619WBS01	VX046762.D	06/19/2025	11:24
VX0619WBSD01	VX0619WBSD01	VX046763.D	06/19/2025	11:51
TT205S1-20250617	Q2361-01	VX046764.D	06/19/2025	12:12
VSTDCCC050EC	VSTDCCC050	VX046775.D	06/19/2025	17:10

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: Q2361 SAS No.: Q2361 SDG NO.: Q2361
Lab File ID: VX046758.D Date Analyzed: 06/19/2025
Instrument ID: MSVOA_X Time Analyzed: 09:43
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	137182	5.56	224265	6.76	197740	10.05
UPPER LIMIT	274364	6.056	448530	7.263	395480	10.549
LOWER LIMIT	68591	5.056	112133	6.263	98870	9.549
EPA SAMPLE NO.						
TT205S1-20250617	120112	5.56	200995	6.77	177542	10.06
VX0619WBL01	124937	5.56	212303	6.77	191555	10.06
VX0619WBS01	137936	5.56	227140	6.77	207463	10.06
VX0619WBSD01	120371	5.56	199496	6.76	183807	10.06

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: Q2361 SAS No.: Q2361 SDG NO.: Q2361
Lab File ID: VX046758.D Date Analyzed: 06/19/2025
Instrument ID: MSVOA_X Time Analyzed: 09:43
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	96867	12.018				
UPPER LIMIT	193734	12.518				
LOWER LIMIT	48433.5	11.518				
EPA SAMPLE NO.						
TT205S1-20250617	84203	12.02				
VX0619WBL01	93040	12.02				
VX0619WBS01	104652	12.02				
VX0619WBSD01	94851	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0619WBL01		SDG No.:	Q2361
Lab Sample ID:	VX0619WBL01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046760.D	1		06/19/25 10:33	VX061925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0619WBL01		SDG No.:	Q2361
Lab Sample ID:	VX0619WBL01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046760.D	1		06/19/25 10:33	VX061925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	48.3		81 - 118		97%	SPK: 50
1868-53-7	Dibromofluoromethane	49.0		80 - 119		98%	SPK: 50
2037-26-5	Toluene-d8	50.0		89 - 112		100%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.4		85 - 114		99%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	125000	5.562				
540-36-3	1,4-Difluorobenzene	212000	6.769				
3114-55-4	Chlorobenzene-d5	192000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	93000	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0619WBL01		SDG No.:	Q2361
Lab Sample ID:	VX0619WBL01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046760.D	1		06/19/25 10:33	VX061925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0619WBS01		SDG No.:	Q2361
Lab Sample ID:	VX0619WBS01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046762.D	1		06/19/25 11:24	VX061925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	19.1		0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	18.7		0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	18.9		0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	21.2		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	19.9		0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	18.4		0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	18.5		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	18.6		0.23	0.75	1.00	ug/L
67-64-1	Acetone	74.1		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	17.0		0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	17.0		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	16.0		0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	18.1		0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	18.3		0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	18.5		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	18.8		1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	78.9		0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	18.0		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	18.6		0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	20.8		0.22	0.50	1.00	ug/L
67-66-3	Chloroform	19.0		0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	17.7		0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	18.5		0.16	0.50	1.00	ug/L
71-43-2	Benzene	18.9		0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	18.5		0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	18.4		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	18.4		0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	18.9		0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	81.6		0.68	2.50	5.00	ug/L
108-88-3	Toluene	19.0		0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0619WBS01		SDG No.:	Q2361
Lab Sample ID:	VX0619WBS01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046762.D	1		06/19/25 11:24	VX061925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	18.0		0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	18.2		0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	18.5		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	78.9		0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	18.4		0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	18.0		0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	17.7		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	18.0		0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	18.0		0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	37.2		0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	18.6		0.12	0.50	1.00	ug/L
100-42-5	Styrene	18.3		0.15	0.50	1.00	ug/L
75-25-2	Bromoform	16.7		0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	18.1		0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	17.3		0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	17.8		0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	17.5		0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	18.1		0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	15.1		0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	17.1		0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	17.6		0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	49.8		81 - 118		100%	SPK: 50
1868-53-7	Dibromofluoromethane	52.0		80 - 119		104%	SPK: 50
2037-26-5	Toluene-d8	52.3		89 - 112		105%	SPK: 50
460-00-4	4-Bromofluorobenzene	54.1		85 - 114		108%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	138000	5.562				
540-36-3	1,4-Difluorobenzene	227000	6.769				
3114-55-4	Chlorobenzene-d5	207000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	105000	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0619WBS01		SDG No.:	Q2361
Lab Sample ID:	VX0619WBS01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046762.D	1		06/19/25 11:24	VX061925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0619WBSD01		SDG No.:	Q2361
Lab Sample ID:	VX0619WBSD01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046763.D	1		06/19/25 11:51	VX061925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	19.8		0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	19.5		0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	19.7		0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	22.2		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	20.6		0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	19.3		0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	19.7		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	19.5		0.23	0.75	1.00	ug/L
67-64-1	Acetone	81.5		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	18.0		0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	18.7		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	17.8		0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	19.3		0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	19.2		0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	19.4		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	19.7		1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	87.0		0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	19.1		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	19.4		0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	23.1		0.22	0.50	1.00	ug/L
67-66-3	Chloroform	20.5		0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	19.3		0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	19.0		0.16	0.50	1.00	ug/L
71-43-2	Benzene	20.0		0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	20.1		0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	19.3		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	20.3		0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	20.1		0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	91.8		0.68	2.50	5.00	ug/L
108-88-3	Toluene	20.3		0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0619WBSD01		SDG No.:	Q2361
Lab Sample ID:	VX0619WBSD01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046763.D	1		06/19/25 11:51	VX061925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	19.5		0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	19.9		0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	20.4		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	88.8		0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	19.9		0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	20.1		0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	18.8		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	19.4		0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	19.1		0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	38.8		0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	19.8		0.12	0.50	1.00	ug/L
100-42-5	Styrene	19.7		0.15	0.50	1.00	ug/L
75-25-2	Bromoform	18.4		0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	18.6		0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	18.4		0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	18.6		0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	17.8		0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	19.0		0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	16.1		0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	18.3		0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	18.2		0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	51.5		81 - 118		103%	SPK: 50
1868-53-7	Dibromofluoromethane	52.8		80 - 119		106%	SPK: 50
2037-26-5	Toluene-d8	52.7		89 - 112		105%	SPK: 50
460-00-4	4-Bromofluorobenzene	54.9		85 - 114		110%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	120000	5.562				
540-36-3	1,4-Difluorobenzene	199000	6.763				
3114-55-4	Chlorobenzene-d5	184000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	94900	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0619WBSD01		SDG No.:	Q2361
Lab Sample ID:	VX0619WBSD01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046763.D	1		06/19/25 11:51	VX061925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q2361 SAS No.: Q2361 SDG No.: Q2361

Instrument ID: MSVOA_X Calibration Date(s): 06/17/2025 06/17/2025

Heated Purge: (Y/N) N Calibration Time(s): 11:19 17:18

GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:		RRF005 = VX046718.D		RRF020 = VX046719.D		RRF050 = VX046720.D		RRF100 = VX046721.D		RRF150 = VX046722.D		RRF001 = VX046725.D	
COMPOUND		RRF005	RRF020	RRF050	RRF100	RRF150	RRF001	RRF	% RSD				
Dichlorodifluoromethane		0.464	0.507	0.583	0.627	0.598	0.425	0.534	15.1				
Chloromethane		0.541	0.570	0.593	0.641	0.627	0.485	0.576	10				
Vinyl Chloride		0.569	0.632	0.634	0.687	0.644	0.521	0.614	9.7				
Bromomethane		0.371	0.367	0.369	0.341	0.280		0.346	11.1				
Chloroethane		0.350	0.388	0.377	0.405	0.381	0.332	0.372	7.1				
Trichlorofluoromethane		0.897	0.974	0.967	1.030	0.972	0.757	0.933	10.3				
1,1,2-Trichlorotrifluoroethane		0.563	0.608	0.576	0.623	0.594	0.470	0.572	9.6				
1,1-Dichloroethene		0.527	0.574	0.569	0.609	0.583	0.451	0.552	10.2				
Acetone		0.219	0.222	0.220	0.238	0.234	0.251	0.231	5.6				
Carbon Disulfide		1.503	1.644	1.599	1.735	1.656	1.869	1.668	7.5				
Methyl tert-butyl Ether		1.628	1.803	1.752	1.898	1.808	1.427	1.720	9.8				
Methyl Acetate		0.577	0.590	0.581	0.650	0.647	0.470	0.586	11.2				
Methylene Chloride		0.655	0.655	0.610	0.666	0.624	0.637	0.641	3.3				
trans-1,2-Dichloroethene		0.569	0.627	0.589	0.637	0.595	0.529	0.591	6.7				
1,1-Dichloroethane		1.054	1.153	1.088	1.170	1.114	0.972	1.092	6.6				
Cyclohexane		0.983	1.086	1.013	1.074	1.021		1.036	4.2				
2-Butanone		0.319	0.325	0.338	0.371	0.353	0.251	0.326	12.7				
Carbon Tetrachloride		0.509	0.537	0.515	0.548	0.531	0.470	0.518	5.4				
cis-1,2-Dichloroethene		0.681	0.731	0.686	0.741	0.704	0.623	0.694	6.1				
Bromochloromethane		0.527	0.459	0.485	0.519	0.500	0.480	0.495	5.2				
Chloroform		1.102	1.190	1.114	1.181	1.109	0.857	1.092	11.1				
1,1,1-Trichloroethane		0.931	1.012	0.956	1.046	0.990	0.812	0.958	8.6				
Methylcyclohexane		0.631	0.642	0.629	0.672	0.647	0.550	0.628	6.6				
Benzene		1.431	1.477	1.417	1.509	1.427	1.218	1.413	7.2				
1,2-Dichloroethane		0.508	0.523	0.495	0.528	0.497	0.429	0.497	7.2				
Trichloroethene		0.355	0.374	0.356	0.389	0.366	0.320	0.360	6.5				
1,2-Dichloropropane		0.347	0.372	0.346	0.374	0.357	0.305	0.350	7.2				
Bromodichloromethane		0.510	0.545	0.522	0.562	0.540	0.404	0.514	11.1				
4-Methyl-2-Pentanone		0.408	0.414	0.426	0.460	0.439	0.322	0.411	11.6				
Toluene		0.884	0.927	0.888	0.927	0.881	0.750	0.876	7.4				

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q2361 SAS No.: Q2361 SDG No.: Q2361

Instrument ID: MSVOA_X Calibration Date(s): 06/17/2025 06/17/2025

Heated Purge: (Y/N) N Calibration Time(s): 11:19 17:18

GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:		RRF005 = VX046718.D		RRF020 = VX046719.D		RRF050 = VX046720.D			
		RRF100 = VX046721.D		RRF150 = VX046722.D		RRF001 = VX046725.D			
COMPOUND	RRF005	RRF020	RRF050	RRF100	RRF150	RRF001	RRF	% RSD	
t-1,3-Dichloropropene	0.438	0.484	0.488	0.555	0.540	0.355	0.477	15.3	
cis-1,3-Dichloropropene	0.516	0.555	0.548	0.603	0.589	0.436	0.541	11.1	
1,1,2-Trichloroethane	0.330	0.341	0.326	0.347	0.329	0.287	0.327	6.4	
2-Hexanone	0.285	0.285	0.297	0.320	0.305	0.214	0.284	13	
Dibromochloromethane	0.380	0.402	0.388	0.419	0.402	0.318	0.385	9.3	
1,2-Dibromoethane	0.333	0.348	0.335	0.363	0.346	0.267	0.332	10.1	
Tetrachloroethene	0.345	0.353	0.340	0.360	0.339	0.341	0.346	2.4	
Chlorobenzene	1.114	1.148	1.091	1.165	1.100	1.005	1.104	5.1	
Ethyl Benzene	1.905	2.030	1.933	2.062	1.945	1.696	1.929	6.7	
m/p-Xylenes	0.705	0.764	0.724	0.765	0.718	0.635	0.719	6.7	
o-Xylene	0.686	0.729	0.692	0.739	0.698	0.498	0.673	13.1	
Styrene	1.144	1.256	1.208	1.267	1.203	0.993	1.179	8.6	
Bromoform	0.268	0.295	0.287	0.315	0.304	0.226	0.282	11.3	
Isopropylbenzene	3.593	3.914	3.723	4.004	3.814	3.048	3.682	9.3	
1,1,2,2-Tetrachloroethane	1.037	1.075	1.044	1.124	1.074	0.816	1.028	10.6	
1,3-Dichlorobenzene	1.703	1.787	1.678	1.786	1.719	1.694	1.728	2.7	
1,4-Dichlorobenzene	1.743	1.830	1.653	1.789	1.702	1.932	1.775	5.6	
1,2-Dichlorobenzene	1.604	1.701	1.592	1.703	1.628	1.460	1.615	5.5	
1,2-Dibromo-3-Chloropropane	0.196	0.205	0.211	0.235	0.237	0.134	0.203	18.6	
1,2,4-Trichlorobenzene	1.040	1.138	1.127	1.253	1.160	1.170	1.148	6	
1,2,3-Trichlorobenzene	1.062	1.129	1.082	1.207	1.153	0.957	1.098	7.9	
1,2-Dichloroethane-d4	0.801	0.567	0.649	0.726	0.714		0.692	12.7	
Dibromofluoromethane	0.362	0.267	0.320	0.354	0.351		0.331	11.9	
Toluene-d8	1.342	0.973	1.144	1.250	1.234		1.189	11.7	
4-Bromofluorobenzene	0.513	0.354	0.426	0.466	0.456		0.443	13.3	

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q2361 SAS No.: Q2361 SDG No.: Q2361

Instrument ID: MSVOA_X Calibration Date/Time: 06/19/2025 09:43

Lab File ID: VX046758.D Init. Calib. Date(s): 06/17/2025 06/17/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:19 17:18

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.534	0.565		5.8	20
Chloromethane	0.576	0.595	0.1	3.3	20
Vinyl Chloride	0.614	0.641		4.4	20
Bromomethane	0.346	0.384		10.98	20
Chloroethane	0.372	0.387		4.03	20
Trichlorofluoromethane	0.933	0.957		2.57	20
1,1,2-Trichlorotrifluoroethane	0.572	0.592		3.5	20
1,1-Dichloroethene	0.552	0.572		3.62	20
Acetone	0.231	0.185		-19.91	20
Carbon Disulfide	1.668	1.595		-4.38	20
Methyl tert-butyl Ether	1.720	1.569		-8.78	20
Methyl Acetate	0.586	0.480		-18.09	20
Methylene Chloride	0.641	0.618		-3.59	20
trans-1,2-Dichloroethene	0.591	0.587		-0.68	20
1,1-Dichloroethane	1.092	1.081	0.1	-1.01	20
Cyclohexane	1.036	1.018		-1.74	20
2-Butanone	0.326	0.263		-19.33	20
Carbon Tetrachloride	0.518	0.515		-0.58	20
cis-1,2-Dichloroethene	0.694	0.686		-1.15	20
Bromochloromethane	0.495	0.476		-3.84	20
Chloroform	1.092	1.093		0.09	20
1,1,1-Trichloroethane	0.958	0.939		-1.98	20
Methylcyclohexane	0.628	0.627		-0.32	20
Benzene	1.413	1.415		0.14	20
1,2-Dichloroethane	0.497	0.481		-3.22	20
Trichloroethene	0.360	0.357		-0.83	20
1,2-Dichloropropane	0.350	0.353		0.86	20
Bromodichloromethane	0.514	0.521		1.36	20
4-Methyl-2-Pentanone	0.411	0.339		-17.52	20
Toluene	0.876	0.877		0.11	20
t-1,3-Dichloropropene	0.477	0.481		0.84	20
cis-1,3-Dichloropropene	0.541	0.545		0.74	20
1,1,2-Trichloroethane	0.327	0.311		-4.89	20
2-Hexanone	0.284	0.229		-19.37	20
Dibromochloromethane	0.385	0.377		-2.08	20
1,2-Dibromoethane	0.332	0.317		-4.52	20
Tetrachloroethene	0.346	0.344		-0.58	20
Chlorobenzene	1.104	1.089	0.3	-1.36	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q2361 SAS No.: Q2361 SDG No.: Q2361

Instrument ID: MSVOA_X Calibration Date/Time: 06/19/2025 09:43

Lab File ID: VX046758.D Init. Calib. Date(s): 06/17/2025 06/17/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:19 17:18

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.929	1.918		-0.57	20
m/p-Xylenes	0.719	0.724		0.69	20
o-Xylene	0.673	0.687		2.08	20
Styrene	1.179	1.192		1.1	20
Bromoform	0.282	0.263	0.1	-6.74	20
Isopropylbenzene	3.682	3.747		1.76	20
1,1,2,2-Tetrachloroethane	1.028	0.933	0.3	-9.24	20
1,3-Dichlorobenzene	1.728	1.670		-3.36	20
1,4-Dichlorobenzene	1.775	1.693		-4.62	20
1,2-Dichlorobenzene	1.615	1.598		-1.05	20
1,2-Dibromo-3-Chloropropane	0.203	0.164		-19.21	20
1,2,4-Trichlorobenzene	1.148	1.111		-3.22	20
1,2,3-Trichlorobenzene	1.098	1.064		-3.1	20
1,2-Dichloroethane-d4	0.692	0.639		-7.66	20
Dibromofluoromethane	0.331	0.327		-1.21	20
Toluene-d8	1.189	1.166		-1.93	20
4-Bromofluorobenzene	0.443	0.433		-2.26	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q2361 SAS No.: Q2361 SDG No.: Q2361

Instrument ID: MSVOA_X Calibration Date/Time: 06/19/2025 17:10

Lab File ID: VX046775.D Init. Calib. Date(s): 06/17/2025 06/17/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:19 17:18

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.534	0.573		7.3	50
Chloromethane	0.576	0.601	0.1	4.34	50
Vinyl Chloride	0.614	0.661		7.66	50
Bromomethane	0.346	0.382		10.4	50
Chloroethane	0.372	0.413		11.02	50
Trichlorofluoromethane	0.933	0.983		5.36	50
1,1,2-Trichlorotrifluoroethane	0.572	0.614		7.34	50
1,1-Dichloroethene	0.552	0.592		7.25	50
Acetone	0.231	0.223		-3.46	50
Carbon Disulfide	1.668	1.588		-4.8	50
Methyl tert-butyl Ether	1.720	1.740		1.16	50
Methyl Acetate	0.586	0.608		3.75	50
Methylene Chloride	0.641	0.645		0.62	50
trans-1,2-Dichloroethene	0.591	0.612		3.55	50
1,1-Dichloroethane	1.092	1.144	0.1	4.76	50
Cyclohexane	1.036	1.061		2.41	50
2-Butanone	0.326	0.334		2.45	50
Carbon Tetrachloride	0.518	0.513		-0.96	50
cis-1,2-Dichloroethene	0.694	0.729		5.04	50
Bromochloromethane	0.495	0.513		3.64	50
Chloroform	1.092	1.155		5.77	50
1,1,1-Trichloroethane	0.958	0.973		1.57	50
Methylcyclohexane	0.628	0.645		2.71	50
Benzene	1.413	1.472		4.18	50
1,2-Dichloroethane	0.497	0.505		1.61	50
Trichloroethene	0.360	0.369		2.5	50
1,2-Dichloropropane	0.350	0.363		3.71	50
Bromodichloromethane	0.514	0.529		2.92	50
4-Methyl-2-Pentanone	0.411	0.419		1.95	50
Toluene	0.876	0.902		2.97	50
t-1,3-Dichloropropene	0.477	0.490		2.72	50
cis-1,3-Dichloropropene	0.541	0.565		4.44	50
1,1,2-Trichloroethane	0.327	0.339		3.67	50
2-Hexanone	0.284	0.287		1.06	50
Dibromochloromethane	0.385	0.396		2.86	50
1,2-Dibromoethane	0.332	0.339		2.11	50
Tetrachloroethene	0.346	0.348		0.58	50
Chlorobenzene	1.104	1.134	0.3	2.72	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q2361 SAS No.: Q2361 SDG No.: Q2361

Instrument ID: MSVOA_X Calibration Date/Time: 06/19/2025 17:10

Lab File ID: VX046775.D Init. Calib. Date(s): 06/17/2025 06/17/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:19 17:18

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.929	1.995		3.42	50
m/p-Xylenes	0.719	0.745		3.62	50
o-Xylene	0.673	0.716		6.39	50
Styrene	1.179	1.253		6.28	50
Bromoform	0.282	0.283	0.1	0.35	50
Isopropylbenzene	3.682	3.814		3.59	50
1,1,2,2-Tetrachloroethane	1.028	1.047	0.3	1.85	50
1,3-Dichlorobenzene	1.728	1.702		-1.5	50
1,4-Dichlorobenzene	1.775	1.686		-5.01	50
1,2-Dichlorobenzene	1.615	1.619		0.25	50
1,2-Dibromo-3-Chloropropane	0.203	0.194		-4.43	50
1,2,4-Trichlorobenzene	1.148	1.143		-0.44	50
1,2,3-Trichlorobenzene	1.098	1.103		0.46	50
1,2-Dichloroethane-d4	0.692	0.648		-6.36	50
Dibromofluoromethane	0.331	0.320		-3.32	50
Toluene-d8	1.189	1.138		-4.29	50
4-Bromofluorobenzene	0.443	0.431		-2.71	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.