

Cover Page

Order ID : Q2361

Project ID : NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258

Client : AECOM Technical Services, Inc.

Lab Sample Number

Q2361-01

Client Sample Number

TT205S1-20250617

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 6/24/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

AECOM Technical Services, Inc.

Project Name: NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258

Project # N/A

Order ID # Q2361

Test Name: VOCMS Group1

A. Number of Samples and Date of Receipt:

1 Water sample was received on 06/18/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: SVOC-SIMGroup1 and VOCMS Group1. This data package contains results for VOCMS Group1.

C. Analytical Techniques:

The analysis performed on instrument MSVOA_X were done using GC column DB-624UI 20m 0.18mm 1.0 um. Cat#121-1324UI. The analysis of VOCMS Group1 was based on method 8260D.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The RPD met criteria .

The Blank Spike met requirements for all samples .

The Blank Spike Duplicate met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Continuous Calibration met the requirements .

The Tuning criteria met requirements.

E. Additional Comments:

Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.

The not QT review data is reported in the Miscellaneous.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012; NEW YORK LAB ID#: 11376

GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY

ORDER ID: Q2361

MATRIX: Water

METHOD: 8260D

		NA	NO	YES
1.	Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)			✓
2.	GC/MS Tuning Specifications BFB Meet Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)			✓
3.	GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 Series.			✓
4.	GC/MS Calibration - Initial Calibration performed before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series.			✓
5.	GC/MS Calibration Requirements. The Initial Calibration met the requirements . The Continuous Calibration met the requirements .			✓
6.	Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
7.	Surrogate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable ranges.			✓
8.	Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range. The Blank Spike met requirements for all samples . The Blank Spike Duplicate met requirements for all samples .			✓
9.	Internal Standard Area/Retention Time Shift Meet Criteria Comments:			✓
10.	Analysis Holding Time Met If not met, list number of days exceeded for each sample:			✓

ADDITIONAL COMMENTS:

Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.

The not QT review data is reported in the Miscellaneous.

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012; NEW YORK LAB ID#: 11376

GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY (CONTINUED)

NA NO YES

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

QA REVIEW

Date

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q2361

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: PRATHA PARGHI

Date: 06/24/2025



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

LAB CHRONICLE

OrderID:	Q2361	OrderDate:	6/18/2025 3:09:00 PM
Client:	AECOM Technical Services, Inc.	Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258
Contact:	Eleanor Vivadou	Location:	D51,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2361-01	TT205S1-20250617	Water	VOCMS Group1	8260-Low	06/17/25		06/19/25	06/18/25



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q2361
Client: AECOM Technical Services, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID:	TT205S1-20250617								
Q2361-01	TT205S1-20250617	Water	Trichloroethene	7.00		0.090	0.75	1.00	ug/L
			Total Voc :	7.00					
			Total Concentration:	7.00					



QC SUMMARY

Surrogate Summary

SDG No.: Q2361

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2361-01	TT205S1-20250617	1,2-Dichloroethane-d4	50	47.5	95	81	118
		Dibromofluoromethane	50	49.1	98	80	119
		Toluene-d8	50	50.5	101	89	112
		4-Bromofluorobenzene	50	48.7	97	85	114
VX0619WBL01	VX0619WBL01	1,2-Dichloroethane-d4	50	48.3	97	81	118
		Dibromofluoromethane	50	49.0	98	80	119
		Toluene-d8	50	50.0	100	89	112
		4-Bromofluorobenzene	50	49.4	99	85	114
VX0619WBS01	VX0619WBS01	1,2-Dichloroethane-d4	50	49.8	100	81	118
		Dibromofluoromethane	50	52.0	104	80	119
		Toluene-d8	50	52.3	105	89	112
		4-Bromofluorobenzene	50	54.1	108	85	114
VX0619WBSD0	VX0619WBSD01	1,2-Dichloroethane-d4	50	51.5	103	81	118
		Dibromofluoromethane	50	52.8	106	80	119
		Toluene-d8	50	52.7	105	89	112
		4-Bromofluorobenzene	50	54.9	110	85	114

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2361

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Datafile : VX046762.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0619WBS01	Dichlorodifluoromethane	20	19.1	ug/L	96			32	152	
	Chloromethane	20	18.7	ug/L	94			50	139	
	Vinyl chloride	20	18.9	ug/L	95			58	137	
	Bromomethane	20	21.2	ug/L	106			53	141	
	Chloroethane	20	19.9	ug/L	100			60	138	
	Trichlorofluoromethane	20	18.4	ug/L	92			65	141	
	1,1,2-Trichlorotrifluoroethane	20	18.5	ug/L	93			70	136	
	1,1-Dichloroethene	20	18.6	ug/L	93			71	131	
	Acetone	100	74.1	ug/L	74			39	160	
	Carbon disulfide	20	17.0	ug/L	85			64	133	
	Methyl tert-butyl Ether	20	17.0	ug/L	85			71	124	
	Methyl Acetate	20	16.0	ug/L	80			56	136	
	Methylene Chloride	20	18.1	ug/L	91			74	124	
	trans-1,2-Dichloroethene	20	18.3	ug/L	92			75	124	
	1,1-Dichloroethane	20	18.5	ug/L	93			77	125	
	Cyclohexane	20	18.8	ug/L	94			71	130	
	2-Butanone	100	78.9	ug/L	79			56	143	
	Carbon Tetrachloride	20	18.0	ug/L	90			72	136	
	cis-1,2-Dichloroethene	20	18.6	ug/L	93			78	123	
	Bromochloromethane	20	20.8	ug/L	104			78	123	
	Chloroform	20	19.0	ug/L	95			79	124	
	1,1,1-Trichloroethane	20	17.7	ug/L	89			74	131	
	Methylcyclohexane	20	18.5	ug/L	93			72	132	
	Benzene	20	18.9	ug/L	95			79	120	
	1,2-Dichloroethane	20	18.5	ug/L	93			73	128	
	Trichloroethene	20	18.4	ug/L	92			79	123	
	1,2-Dichloropropane	20	18.4	ug/L	92			78	122	
	Bromodichloromethane	20	18.9	ug/L	95			79	125	
	4-Methyl-2-Pentanone	100	81.6	ug/L	82			67	130	
	Toluene	20	19.0	ug/L	95			80	121	
	t-1,3-Dichloropropene	20	18.0	ug/L	90			73	127	
	cis-1,3-Dichloropropene	20	18.2	ug/L	91			75	124	
	1,1,2-Trichloroethane	20	18.5	ug/L	93			80	119	
	2-Hexanone	100	78.9	ug/L	79			57	139	
	Dibromochloromethane	20	18.4	ug/L	92			74	126	
	1,2-Dibromoethane	20	18.0	ug/L	90			77	121	
	Tetrachloroethene	20	17.7	ug/L	89			74	129	
	Chlorobenzene	20	18.0	ug/L	90			82	118	
	Ethyl Benzene	20	18.0	ug/L	90			79	121	
	m/p-Xylenes	40	37.2	ug/L	93			80	121	
	o-Xylene	20	18.6	ug/L	93			78	122	
	Styrene	20	18.3	ug/L	92			78	123	
	Bromoform	20	16.7	ug/L	84			66	130	
	Isopropylbenzene	20	18.1	ug/L	91			72	131	
	1,1,2,2-Tetrachloroethane	20	17.3	ug/L	86			71	121	
	1,3-Dichlorobenzene	20	17.8	ug/L	89			80	119	
	1,4-Dichlorobenzene	20	17.5	ug/L	88			79	118	
	1,2-Dichlorobenzene	20	18.1	ug/L	91			80	119	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2361

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Datafile : VX046762.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VX0619WBS01	1,2-Dibromo-3-Chloropropane	20	15.1	ug/L	76			62	128	
	1,2,4-Trichlorobenzene	20	17.1	ug/L	86			69	130	
	1,2,3-Trichlorobenzene	20	17.6	ug/L	88			69	129	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2361

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Datafile : VX046763.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0619WBSD01	Dichlorodifluoromethane	20	19.8	ug/L	99	3		32	152	20
	Chloromethane	20	19.5	ug/L	98	4		50	139	20
	Vinyl chloride	20	19.7	ug/L	99	4		58	137	20
	Bromomethane	20	22.2	ug/L	111	5		53	141	20
	Chloroethane	20	20.6	ug/L	103	3		60	138	20
	Trichlorofluoromethane	20	19.3	ug/L	97	5		65	141	20
	1,1,2-Trichlorotrifluoroethane	20	19.7	ug/L	99	6		70	136	20
	1,1-Dichloroethene	20	19.5	ug/L	98	5		71	131	20
	Acetone	100	81.5	ug/L	82	10		39	160	20
	Carbon disulfide	20	18.0	ug/L	90	6		64	133	20
	Methyl tert-butyl Ether	20	18.7	ug/L	94	10		71	124	20
	Methyl Acetate	20	17.8	ug/L	89	11		56	136	20
	Methylene Chloride	20	19.3	ug/L	97	6		74	124	20
	trans-1,2-Dichloroethene	20	19.2	ug/L	96	4		75	124	20
	1,1-Dichloroethane	20	19.4	ug/L	97	4		77	125	20
	Cyclohexane	20	19.7	ug/L	99	5		71	130	20
	2-Butanone	100	87.0	ug/L	87	10		56	143	20
	Carbon Tetrachloride	20	19.1	ug/L	96	6		72	136	20
	cis-1,2-Dichloroethene	20	19.4	ug/L	97	4		78	123	20
	Bromochloromethane	20	23.1	ug/L	116	11		78	123	20
	Chloroform	20	20.5	ug/L	103	8		79	124	20
	1,1,1-Trichloroethane	20	19.3	ug/L	97	9		74	131	20
	Methylcyclohexane	20	19.0	ug/L	95	2		72	132	20
	Benzene	20	20.0	ug/L	100	5		79	120	20
	1,2-Dichloroethane	20	20.1	ug/L	101	8		73	128	20
	Trichloroethene	20	19.3	ug/L	97	5		79	123	20
	1,2-Dichloropropane	20	20.3	ug/L	102	10		78	122	20
	Bromodichloromethane	20	20.1	ug/L	101	6		79	125	20
	4-Methyl-2-Pentanone	100	91.8	ug/L	92	11		67	130	20
	Toluene	20	20.3	ug/L	102	7		80	121	20
	t-1,3-Dichloropropene	20	19.5	ug/L	98	9		73	127	20
	cis-1,3-Dichloropropene	20	19.9	ug/L	100	9		75	124	20
	1,1,2-Trichloroethane	20	20.4	ug/L	102	9		80	119	20
	2-Hexanone	100	88.8	ug/L	89	12		57	139	20
	Dibromochloromethane	20	19.9	ug/L	100	8		74	126	20
	1,2-Dibromoethane	20	20.1	ug/L	101	12		77	121	20
	Tetrachloroethene	20	18.8	ug/L	94	5		74	129	20
	Chlorobenzene	20	19.4	ug/L	97	7		82	118	20
	Ethyl Benzene	20	19.1	ug/L	96	6		79	121	20
	m/p-Xylenes	40	38.8	ug/L	97	4		80	121	20
	o-Xylene	20	19.8	ug/L	99	6		78	122	20
	Styrene	20	19.7	ug/L	99	7		78	123	20
	Bromoform	20	18.4	ug/L	92	9		66	130	20
	Isopropylbenzene	20	18.6	ug/L	93	2		72	131	20
	1,1,2,2-Tetrachloroethane	20	18.4	ug/L	92	7		71	121	20
	1,3-Dichlorobenzene	20	18.6	ug/L	93	4		80	119	20
	1,4-Dichlorobenzene	20	17.8	ug/L	89	1		79	118	20
	1,2-Dichlorobenzene	20	19.0	ug/L	95	4		80	119	20



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2361

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Datafile : VX046763.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0619WBSD01	1,2-Dibromo-3-Chloropropane	20	16.1	ug/L	81	6		62	128	20
	1,2,4-Trichlorobenzene	20	18.3	ug/L	92	7		69	130	20
	1,2,3-Trichlorobenzene	20	18.2	ug/L	91	3		69	129	20

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VX0619WBL01

Lab Name: CHEMTECH

Contract: AECO15

Lab Code: CHEM Case No.: Q2361

SAS No.: Q2361 SDG NO.: Q2361

Lab File ID: VX046760.D

Lab Sample ID: VX0619WBL01

Date Analyzed: 06/19/2025

Time Analyzed: 10:33

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_X

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VX0619WBS01	VX0619WBS01	VX046762.D	06/19/2025
VX0619WBSD01	VX0619WBSD01	VX046763.D	06/19/2025
TT205S1-20250617	Q2361-01	VX046764.D	06/19/2025

COMMENTS: _____



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: Q2361 SAS No.: Q2361 SDG NO.: Q2361
Lab File ID: VX046715.D BFB Injection Date: 06/17/2025
Instrument ID: MSVOA_X BFB Injection Time: 08:46
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	18.7
75	30.0 - 60.0% of mass 95	50.7
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.5
173	Less than 2.0% of mass 174	0.5 (0.7) 1
174	50.0 - 100.0% of mass 95	74.8
175	5.0 - 9.0% of mass 174	5.5 (7.4) 1
176	95.0 - 101.0% of mass 174	72 (96.2) 1
177	5.0 - 9.0% of mass 176	4.3 (6) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC005	VSTDIC005	VX046718.D	06/17/2025	11:19
VSTDIC020	VSTDIC020	VX046719.D	06/17/2025	13:59
VSTDIC050	VSTDIC050	VX046720.D	06/17/2025	14:20
VSTDIC100	VSTDIC100	VX046721.D	06/17/2025	14:41
VSTDIC150	VSTDIC150	VX046722.D	06/17/2025	15:02
VSTDIC001	VSTDIC001	VX046725.D	06/17/2025	17:18



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: Q2361 SAS No.: Q2361 SDG NO.: Q2361
Lab File ID: VX046757.D BFB Injection Date: 06/19/2025
Instrument ID: MSVOA_X BFB Injection Time: 08:42
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	18.8
75	30.0 - 60.0% of mass 95	50.7
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.3
173	Less than 2.0% of mass 174	0.7 (1) 1
174	50.0 - 100.0% of mass 95	72.6
175	5.0 - 9.0% of mass 174	5.7 (7.8) 1
176	95.0 - 101.0% of mass 174	71.1 (97.9) 1
177	5.0 - 9.0% of mass 176	4.5 (6.3) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX046758.D	06/19/2025	09:43
VX0619WBL01	VX0619WBL01	VX046760.D	06/19/2025	10:33
VX0619WBS01	VX0619WBS01	VX046762.D	06/19/2025	11:24
VX0619WBSD01	VX0619WBSD01	VX046763.D	06/19/2025	11:51
TT205S1-20250617	Q2361-01	VX046764.D	06/19/2025	12:12
VSTDCCC050EC	VSTDCCC050	VX046775.D	06/19/2025	17:10

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: Q2361 SAS No.: Q2361 SDG NO.: Q2361
 Lab File ID: VX046758.D Date Analyzed: 06/19/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:43
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	137182	5.56	224265	6.76	197740	10.05
UPPER LIMIT	274364	6.056	448530	7.263	395480	10.549
LOWER LIMIT	68591	5.056	112133	6.263	98870	9.549
EPA SAMPLE NO.						
TT205S1-20250617	120112	5.56	200995	6.77	177542	10.06
VX0619WBL01	124937	5.56	212303	6.77	191555	10.06
VX0619WBS01	137936	5.56	227140	6.77	207463	10.06
VX0619WBSD01	120371	5.56	199496	6.76	183807	10.06

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: Q2361 SAS No.: Q2361 SDG NO.: Q2361
 Lab File ID: VX046758.D Date Analyzed: 06/19/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:43
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	96867	12.018				
UPPER LIMIT	193734	12.518				
LOWER LIMIT	48433.5	11.518				
EPA SAMPLE NO.						
TT205S1-20250617	84203	12.02				
VX0619WBL01	93040	12.02				
VX0619WBS01	104652	12.02				
VX0619WBSD01	94851	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.



SAMPLE DATA

Report of Analysis

Client:	AECOM Technical Services, Inc.			Date Collected:	06/17/25		
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258			Date Received:	06/18/25		
Client Sample ID:	TT205S1-20250617			SDG No.:	Q2361		
Lab Sample ID:	Q2361-01			Matrix:	Water		
Analytical Method:	8260D			% Solid:	0		
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL	
Soil Aliquot Vol:			uL	Test:	VOCMS Group1		
GC Column:	DB-624UI	ID :	0.18	Level :	LOW		
Prep Method :							

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046764.D	1		06/19/25 12:12	VX061925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	7.00		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	06/17/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	06/18/25	
Client Sample ID:	TT205S1-20250617		SDG No.:	Q2361	
Lab Sample ID:	Q2361-01		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046764.D	1		06/19/25 12:12	VX061925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	47.5		81 - 118		95%	SPK: 50
1868-53-7	Dibromofluoromethane	49.1		80 - 119		98%	SPK: 50
2037-26-5	Toluene-d8	50.5		89 - 112		101%	SPK: 50
460-00-4	4-Bromofluorobenzene	48.7		85 - 114		97%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	120000	5.562				
540-36-3	1,4-Difluorobenzene	201000	6.769				
3114-55-4	Chlorobenzene-d5	178000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	84200	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.	Date Collected:	06/17/25
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258	Date Received:	06/18/25
Client Sample ID:	TT205S1-20250617	SDG No.:	Q2361
Lab Sample ID:	Q2361-01	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046764.D	1		06/19/25 12:12	VX061925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
 Data File : VX046764.D
 Acq On : 19 Jun 2025 12:12
 Operator : JC/MD
 Sample : Q2361-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT205S1-20250617

Quant Time: Jun 20 05:18:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	5.562	168	120112	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	200995	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	177542	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	84203	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	78866	47.471	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.940%
35) Dibromofluoromethane	5.403	113	65287	49.094	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.180%
50) Toluene-d8	8.653	98	241409	50.527	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.060%
62) 4-Bromofluorobenzene	11.079	95	86656	48.656	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	97.320%
Target Compounds						
					Qvalue	
44) Trichloroethene	7.135	130	10170	7.024	ug/l	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
Data File : VX046764.D
Acq On : 19 Jun 2025 12:12
Operator : JC/MD
Sample : Q2361-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT205S1-20250617

Quant Time: Jun 20 05:18:31 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 03:09:16 2025
Response via : Initial Calibration



#1
Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.562 min Scan# 71

Delta R.T. -0.000 min

Lab File: VX046764.D

Acq: 19 Jun 2025 12:12

Instrument :

MSVOA_X

ClientSampleId :





TT205S1-20250617

Tgt Ion:168 Resp: 120112

Ion Ratio Lower Upper

168 100

99 60.6 48.5 72.7



#33

1,2-Dichloroethane-d4

Concen: 47.471 ug/l

RT: 5.964 min Scan# 800

Delta R.T. -0.000 min

Lab File: VX046764.D




Acq: 19 Jun 2025 12:12


Tgt Ion: 65 Resp: 78866

Ion Ratio Lower Upper

65 100

67 52.4 0.0 105.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.769 min Scan# 91

Delta R.T. -0.000 min

Lab File: VX046764.D

Acq: 19 Jun 2025 12:12

Instrument :

MSVOA_X

ClientSampleId :

TT205S1-20250617





Tgt Ion:114 Resp: 200995

Ion Ratio Lower Upper

114 100

63 21.0 0.0 44.2

88 15.2 0.0 33.2



#35

Dibromofluoromethane

Concen: 49.094 ug/l

RT: 5.403 min Scan# 708

Delta R.T. 0.006 min

Lab File: VX046764.D

Acq: 19 Jun 2025 12:12




Tgt Ion:113 Resp: 65287


Ion Ratio Lower Upper

113 100

111 103.6 82.0 123.0

192 18.7 15.3 22.9









#44
Trichloroethene
Concen: 7.024 ug/l
RT: 7.135 min Scan# 991
Delta R.T. 0.006 min
Lab File: VX046764.D
Acq: 19 Jun 2025 12:12




Instrument :
MSVOA_X
ClientSampleId :
TT205S1-20250617


Tgt Ion:130 Resp: 10170
Ion Ratio Lower Upper
130 100
95 103.9 0.0 202.6



#50
Toluene-d8
Concen: 50.527 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. 0.006 min
Lab File: VX046764.D
Acq: 19 Jun 2025 12:12

Tgt Ion: 98 Resp: 241409
Ion Ratio Lower Upper
98 100
100 65.9 53.0 79.4









#62
4-Bromofluorobenzene
Concen: 48.656 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046764.D
Acq: 19 Jun 2025 12:12




Instrument :
MSVOA_X
ClientSampleId :
TT205S1-20250617


Tgt Ion: 95 Resp: 86656
Ion Ratio Lower Upper
95 100
174 76.0 0.0 150.4
176 73.1 0.0 145.0



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX046764.D
Acq: 19 Jun 2025 12:12

Tgt Ion: 117 Resp: 177542
Ion Ratio Lower Upper
117 100
82 55.1 44.6 66.8
119 32.5 25.8 38.8





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046764.D


Acq: 19 Jun 2025 12:12

Instrument :

MSVOA_X

ClientSampleId :

TT205S1-20250617





Tgt Ion:152 Resp: 84203

Ion Ratio Lower Upper

152 100

115 62.1 43.2 129.6

150 161.1 0.0 346.8





CALIBRATION SUMMARY



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: Q2361 SAS No.: Q2361 SDG No.: Q2361
 Instrument ID: MSVOA_X Calibration Date(s): 06/17/2025 06/17/2025
 Heated Purge: (Y/N) N Calibration Time(s): 11:19 17:18
 GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:	RRF005 = VX046718.D	RRF020 = VX046719.D	RRF050 = VX046720.D					
	RRF100 = VX046721.D	RRF150 = VX046722.D	RRF001 = VX046725.D					
COMPOUND	RRF005	RRF020	RRF050	RRF100	RRF150	RRF001	RRF	% RSD
Dichlorodifluoromethane	0.464	0.507	0.583	0.627	0.598	0.425	0.534	15.1
Chloromethane	0.541	0.570	0.593	0.641	0.627	0.485	0.576	10
Vinyl Chloride	0.569	0.632	0.634	0.687	0.644	0.521	0.614	9.7
Bromomethane	0.371	0.367	0.369	0.341	0.280		0.346	11.1
Chloroethane	0.350	0.388	0.377	0.405	0.381	0.332	0.372	7.1
Trichlorofluoromethane	0.897	0.974	0.967	1.030	0.972	0.757	0.933	10.3
1,1,2-Trichlorotrifluoroethane	0.563	0.608	0.576	0.623	0.594	0.470	0.572	9.6
1,1-Dichloroethene	0.527	0.574	0.569	0.609	0.583	0.451	0.552	10.2
Acetone	0.219	0.222	0.220	0.238	0.234	0.251	0.231	5.6
Carbon Disulfide	1.503	1.644	1.599	1.735	1.656	1.869	1.668	7.5
Methyl tert-butyl Ether	1.628	1.803	1.752	1.898	1.808	1.427	1.720	9.8
Methyl Acetate	0.577	0.590	0.581	0.650	0.647	0.470	0.586	11.2
Methylene Chloride	0.655	0.655	0.610	0.666	0.624	0.637	0.641	3.3
trans-1,2-Dichloroethene	0.569	0.627	0.589	0.637	0.595	0.529	0.591	6.7
1,1-Dichloroethane	1.054	1.153	1.088	1.170	1.114	0.972	1.092	6.6
Cyclohexane	0.983	1.086	1.013	1.074	1.021		1.036	4.2
2-Butanone	0.319	0.325	0.338	0.371	0.353	0.251	0.326	12.7
Carbon Tetrachloride	0.509	0.537	0.515	0.548	0.531	0.470	0.518	5.4
cis-1,2-Dichloroethene	0.681	0.731	0.686	0.741	0.704	0.623	0.694	6.1
Bromochloromethane	0.527	0.459	0.485	0.519	0.500	0.480	0.495	5.2
Chloroform	1.102	1.190	1.114	1.181	1.109	0.857	1.092	11.1
1,1,1-Trichloroethane	0.931	1.012	0.956	1.046	0.990	0.812	0.958	8.6
Methylcyclohexane	0.631	0.642	0.629	0.672	0.647	0.550	0.628	6.6
Benzene	1.431	1.477	1.417	1.509	1.427	1.218	1.413	7.2
1,2-Dichloroethane	0.508	0.523	0.495	0.528	0.497	0.429	0.497	7.2
Trichloroethene	0.355	0.374	0.356	0.389	0.366	0.320	0.360	6.5
1,2-Dichloropropane	0.347	0.372	0.346	0.374	0.357	0.305	0.350	7.2
Bromodichloromethane	0.510	0.545	0.522	0.562	0.540	0.404	0.514	11.1
4-Methyl-2-Pentanone	0.408	0.414	0.426	0.460	0.439	0.322	0.411	11.6
Toluene	0.884	0.927	0.888	0.927	0.881	0.750	0.876	7.4

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: Q2361 SAS No.: Q2361 SDG No.: Q2361
 Instrument ID: MSVOA_X Calibration Date(s): 06/17/2025 06/17/2025
 Heated Purge: (Y/N) N Calibration Time(s): 11:19 17:18
 GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:	RRF005 = VX046718.D			RRF020 = VX046719.D		RRF050 = VX046720.D		
	RRF100 = VX046721.D			RRF150 = VX046722.D		RRF001 = VX046725.D		
COMPOUND	RRF005	RRF020	RRF050	RRF100	RRF150	RRF001	RRF	% RSD
t-1,3-Dichloropropene	0.438	0.484	0.488	0.555	0.540	0.355	0.477	15.3
cis-1,3-Dichloropropene	0.516	0.555	0.548	0.603	0.589	0.436	0.541	11.1
1,1,2-Trichloroethane	0.330	0.341	0.326	0.347	0.329	0.287	0.327	6.4
2-Hexanone	0.285	0.285	0.297	0.320	0.305	0.214	0.284	13
Dibromochloromethane	0.380	0.402	0.388	0.419	0.402	0.318	0.385	9.3
1,2-Dibromoethane	0.333	0.348	0.335	0.363	0.346	0.267	0.332	10.1
Tetrachloroethene	0.345	0.353	0.340	0.360	0.339	0.341	0.346	2.4
Chlorobenzene	1.114	1.148	1.091	1.165	1.100	1.005	1.104	5.1
Ethyl Benzene	1.905	2.030	1.933	2.062	1.945	1.696	1.929	6.7
m/p-Xylenes	0.705	0.764	0.724	0.765	0.718	0.635	0.719	6.7
o-Xylene	0.686	0.729	0.692	0.739	0.698	0.498	0.673	13.1
Styrene	1.144	1.256	1.208	1.267	1.203	0.993	1.179	8.6
Bromoform	0.268	0.295	0.287	0.315	0.304	0.226	0.282	11.3
Isopropylbenzene	3.593	3.914	3.723	4.004	3.814	3.048	3.682	9.3
1,1,2,2-Tetrachloroethane	1.037	1.075	1.044	1.124	1.074	0.816	1.028	10.6
1,3-Dichlorobenzene	1.703	1.787	1.678	1.786	1.719	1.694	1.728	2.7
1,4-Dichlorobenzene	1.743	1.830	1.653	1.789	1.702	1.932	1.775	5.6
1,2-Dichlorobenzene	1.604	1.701	1.592	1.703	1.628	1.460	1.615	5.5
1,2-Dibromo-3-Chloropropane	0.196	0.205	0.211	0.235	0.237	0.134	0.203	18.6
1,2,4-Trichlorobenzene	1.040	1.138	1.127	1.253	1.160	1.170	1.148	6
1,2,3-Trichlorobenzene	1.062	1.129	1.082	1.207	1.153	0.957	1.098	7.9
1,2-Dichloroethane-d4	0.801	0.567	0.649	0.726	0.714		0.692	12.7
Dibromofluoromethane	0.362	0.267	0.320	0.354	0.351		0.331	11.9
Toluene-d8	1.342	0.973	1.144	1.250	1.234		1.189	11.7
4-Bromofluorobenzene	0.513	0.354	0.426	0.466	0.456		0.443	13.3

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.

Method Path : Z:\voasrv\HPCHEM1\MSVOA_X\Method\
 Method File : 82X061725W.M
 Title : SW846 8260
 Last Update : Wed Jun 18 03:09:16 2025
 Response Via : Initial Calibration

Calibration Files

1 =VX046725.D 5 =VX046718.D 20 =VX046719.D 50 =VX046720.D 100 =VX046721.D 150 =VX046722.D

Compound		1	5	20	50	100	150	Avg	%RSD

1) I	Pentafluorobenzene	-----ISTD-----							
2) T	Dichlorodifluo...	0.425	0.464	0.507	0.583	0.627	0.598	0.534	15.13
3) P	Chloromethane	0.485	0.541	0.570	0.593	0.641	0.627	0.576	9.97
4) C	Vinyl Chloride	0.521	0.569	0.632	0.634	0.687	0.644	0.614	9.68#
5) T	Bromomethane	0.371	0.367	0.369	0.341	0.280	0.346		11.13
6) T	Chloroethane	0.332	0.350	0.388	0.377	0.405	0.381	0.372	7.12
7) T	Trichlorofluor...	0.757	0.897	0.974	0.967	1.030	0.972	0.933	10.30
8) T	Diethyl Ether	0.265	0.309	0.337	0.322	0.352	0.336	0.320	9.58
9) T	1,1,2-Trichlor...	0.470	0.563	0.608	0.576	0.623	0.594	0.572	9.55
10) T	Methyl Iodide	0.447	0.566	0.654	0.788	0.752	0.642		21.71
11) T	Tert butyl alc...	0.048	0.061	0.062	0.071	0.069	0.062		14.45
12) CM	1,1-Dichloroet...	0.451	0.527	0.574	0.569	0.609	0.583	0.552	10.21#
13) T	Acrolein	0.095	0.074	0.087	0.098	0.097	0.090		11.04
14) T	Allyl chloride	0.876	0.934	1.042	0.999	1.102	1.051	1.001	8.30
15) T	Acrylonitrile	0.226	0.270	0.287	0.285	0.312	0.295	0.279	10.61
16) T	Acetone	0.251	0.219	0.222	0.220	0.238	0.234	0.231	5.60
17) T	Carbon Disulfide	1.869	1.503	1.644	1.599	1.735	1.656	1.668	7.46
18) T	Methyl Acetate	0.470	0.577	0.590	0.581	0.650	0.647	0.586	11.17
19) T	Methyl tert-bu...	1.427	1.628	1.803	1.752	1.898	1.808	1.720	9.79
20) T	Methylene Chlo...	0.637	0.655	0.655	0.610	0.666	0.624	0.641	3.32
21) T	trans-1,2-Dich...	0.529	0.569	0.627	0.589	0.637	0.595	0.591	6.68
22) T	Diisopropyl ether	1.531	1.883	2.073	1.967	2.113	1.973	1.923	10.87
23) T	Vinyl Acetate	1.138	1.423	1.608	1.596	1.747	1.664	1.529	14.33
24) P	1,1-Dichloroet...	0.972	1.054	1.153	1.088	1.170	1.114	1.092	6.63
25) T	2-Butanone	0.251	0.319	0.325	0.338	0.371	0.353	0.326	12.66
26) T	2,2-Dichloropr...	0.688	0.768	0.798	0.761	0.878	0.886	0.796	9.49
27) T	cis-1,2-Dichlo...	0.623	0.681	0.731	0.686	0.741	0.704	0.694	6.06
28) T	Bromochloromet...	0.480	0.527	0.459	0.485	0.519	0.500	0.495	5.18
29) T	Tetrahydrofuran	0.162	0.205	0.211	0.216	0.240	0.229	0.211	12.74
30) C	Chloroform	0.857	1.102	1.190	1.114	1.181	1.109	1.092	11.08#
31) T	Cyclohexane	0.983	1.086	1.013	1.074	1.021	1.036		4.20
32) T	1,1,1-Trichlor...	0.812	0.931	1.012	0.956	1.046	0.990	0.958	8.58
33) S	1,2-Dichloroet...	0.801	0.567	0.649	0.726	0.714	0.692		12.70
34) I	1,4-Difluorobenzene	-----ISTD-----							
35) S	Dibromofluorom...	0.362	0.267	0.320	0.354	0.351	0.331		11.89
36) T	1,1-Dichloropr...	0.446	0.477	0.486	0.477	0.500	0.480	0.478	3.73
37) T	Ethyl Acetate	0.443	0.404	0.440	0.431	0.468	0.459	0.441	5.11
38) T	Carbon Tetrach...	0.470	0.509	0.537	0.515	0.548	0.531	0.518	5.37
39) T	Methylcyclohexane	0.550	0.631	0.642	0.629	0.672	0.647	0.628	6.61
40) TM	Benzene	1.218	1.431	1.477	1.417	1.509	1.427	1.413	7.22
41) T	Methacrylonitrile	0.190	0.207	0.242	0.239	0.267	0.266	0.235	13.31
42) TM	1,2-Dichloroet...	0.429	0.508	0.523	0.495	0.528	0.497	0.497	7.23
43) T	Isopropyl Acetate	0.491	0.678	0.689	0.716	0.795	0.770	0.690	15.63
44) TM	Trichloroethene	0.320	0.355	0.374	0.356	0.389	0.366	0.360	6.47
45) C	1,2-Dichloropr...	0.305	0.347	0.372	0.346	0.374	0.357	0.350	7.18#
46) T	Dibromomethane	0.228	0.243	0.260	0.251	0.270	0.259	0.252	5.92
47) T	Bromodichlorom...	0.404	0.510	0.545	0.522	0.562	0.540	0.514	11.07
48) T	Methyl methacr...	0.232	0.316	0.352	0.370	0.409	0.396	0.346	18.71
49) T	1,4-Dioxane	0.001	0.004	0.005	0.005	0.005	0.005	0.004	42.21
50) S	Toluene-d8	1.342	0.973	1.144	1.250	1.234	1.189		11.72
51) T	4-Methyl-2-Pen...	0.322	0.408	0.414	0.426	0.460	0.439	0.411	11.62
52) CM	Toluene	0.750	0.884	0.927	0.888	0.927	0.881	0.876	7.43#
53) T	t-1,3-Dichloro...	0.355	0.438	0.484	0.488	0.555	0.540	0.477	15.29
54) T	cis-1,3-Dichlo...	0.436	0.516	0.555	0.548	0.603	0.589	0.541	11.13
55) T	1,1,2-Trichlor...	0.287	0.330	0.341	0.326	0.347	0.329	0.327	6.43
56) T	Ethyl methacry...	0.327	0.465	0.493	0.507	0.560	0.540	0.482	17.24

Method Path : Z:\voasrv\HPCHEM1\MSVOA_X\Method\
 Method File : 82X061725W.M

57)	T	1,3-Dichloropr...	0.497	0.575	0.576	0.562	0.600	0.563	0.562	6.16
58)	T	2-Chloroethyl ...	0.167	0.241	0.235	0.281	0.296	0.282	0.250	18.96
59)	T	2-Hexanone	0.214	0.285	0.285	0.297	0.320	0.305	0.284	13.04
60)	T	Dibromochlorom...	0.318	0.380	0.402	0.388	0.419	0.402	0.385	9.25
61)	T	1,2-Dibromoethane	0.267	0.333	0.348	0.335	0.363	0.346	0.332	10.10
62)	S	4-Bromofluorob...	0.513	0.354	0.426	0.466	0.456	0.443		13.28
63)	I	Chlorobenzene-d5	-----ISTD-----							
64)	T	Tetrachloroethene	0.341	0.345	0.353	0.340	0.360	0.339	0.346	2.43
65)	PM	Chlorobenzene	1.005	1.114	1.148	1.091	1.165	1.100	1.104	5.09
66)	T	1,1,1,2-Tetrac...	0.302	0.358	0.398	0.373	0.408	0.386	0.371	10.22
67)	C	Ethyl Benzene	1.696	1.905	2.030	1.933	2.062	1.945	1.929	6.68#
68)	T	m/p-Xylenes	0.635	0.705	0.764	0.724	0.765	0.718	0.719	6.68
69)	T	o-Xylene	0.498	0.686	0.729	0.692	0.739	0.698	0.673	13.15
70)	T	Styrene	0.993	1.144	1.256	1.208	1.267	1.203	1.179	8.57
71)	P	Bromoform	0.226	0.268	0.295	0.287	0.315	0.304	0.282	11.35
72)	I	1,4-Dichlorobenzen...	-----ISTD-----							
73)	T	Isopropylbenzene	3.048	3.593	3.914	3.723	4.004	3.814	3.682	9.31
74)	T	N-amyl acetate	0.927	1.237	1.455	1.467	1.660	1.625	1.395	19.64
75)	P	1,1,2,2-Tetrac...	0.816	1.037	1.075	1.044	1.124	1.074	1.028	10.56
76)	T	1,2,3-Trichlor...	0.650	0.959	1.066	0.990	0.938	0.892	0.916	15.58
77)	T	Bromobenzene	0.744	0.861	0.953	0.892	0.956	0.911	0.886	8.87
78)	T	n-propylbenzene	3.642	4.243	4.682	4.430	4.737	4.509	4.374	9.16
79)	T	2-Chlorotoluene	2.363	2.596	2.736	2.586	2.770	2.618	2.611	5.50
80)	T	1,3,5-Trimethy...	2.497	2.888	3.297	3.072	3.294	3.108	3.026	9.94
81)	T	trans-1,4-Dich...	0.264	0.297	0.311	0.367	0.381	0.324		15.12
82)	T	4-Chlorotoluene	2.734	2.980	3.255	3.024	3.227	3.066	3.048	6.21
83)	T	tert-Butylbenzene	2.658	3.043	3.316	3.122	3.395	3.234	3.128	8.41
84)	T	1,2,4-Trimethy...	2.455	2.992	3.270	3.061	3.302	3.162	3.040	10.20
85)	T	sec-Butylbenzene	3.301	3.920	4.198	3.999	4.272	4.030	3.953	8.73
86)	T	p-Isopropyltol...	2.905	3.275	3.526	3.340	3.574	3.394	3.336	7.17
87)	T	1,3-Dichlorobe...	1.694	1.703	1.787	1.678	1.786	1.719	1.728	2.74
88)	T	1,4-Dichlorobe...	1.932	1.743	1.830	1.653	1.789	1.702	1.775	5.59
89)	T	n-Butylbenzene	2.986	2.970	3.284	3.154	3.415	3.194	3.167	5.43
90)	T	Hexachloroethane	0.522	0.523	0.601	0.584	0.647	0.627	0.584	8.97
91)	T	1,2-Dichlorobe...	1.460	1.604	1.701	1.592	1.703	1.627	1.615	5.55
92)	T	1,2-Dibromo-3-...	0.134	0.196	0.205	0.211	0.235	0.237	0.203	18.57
93)	T	1,2,4-Trichlor...	1.170	1.040	1.138	1.127	1.253	1.160	1.148	5.99
94)	T	Hexachlorobuta...	0.495	0.490	0.501	0.479	0.512	0.420	0.483	6.78
95)	T	Naphthalene	2.636	2.887	3.301	3.335	3.744	3.720	3.270	13.54
96)	T	1,2,3-Trichlor...	0.957	1.062	1.129	1.082	1.206	1.153	1.098	7.87

(#) = Out of Range

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
 Data File : VX046718.D
 Acq On : 17 Jun 2025 11:19
 Operator : JC/MD
 Sample : VSTDICC005
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC005

Quant Time: Jun 18 02:37:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 02:34:13 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
 Supervised By :Mahesh Dadoda 06/18/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	130530	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	217019	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	195782	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	100990	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	10456	5.791	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	11.580%#		
35) Dibromofluoromethane	5.403	113	7865	5.478	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	10.960%#		
50) Toluene-d8	8.653	98	29120	5.645	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	11.280%#		
62) 4-Bromofluorobenzene	11.079	95	11137	5.791	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	11.580%#		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.179	85	6051	4.342	ug/l	99
3) Chloromethane	1.307	50	7066	4.698	ug/l	90
4) Vinyl Chloride	1.386	62	7425	4.629	ug/l	99
5) Bromomethane	1.617	94	4837	5.360	ug/l	95
6) Chloroethane	1.703	64	4575	4.706	ug/l	92
7) Trichlorofluoromethane	1.904	101	11714	4.810	ug/l	92
8) Diethyl Ether	2.148	74	4038	4.831	ug/l	98
9) 1,1,2-Trichlorotrifluo...	2.349	101	7353	4.921	ug/l	96
10) Methyl Iodide	2.465	142	5830	7.181	ug/l	99
11) Tert butyl alcohol	2.959	59	3134	19.340	ug/l	99
12) 1,1-Dichloroethene	2.337	96	6884	4.775	ug/l	91
13) Acrolein	2.251	56	6208	26.301	ug/l	100
14) Allyl chloride	2.678	41	12195	4.668	ug/l	99
15) Acrylonitrile	3.081	53	17611	24.169	ug/l	99
16) Acetone	2.392	43	14275	23.710	ug/l	98
17) Carbon Disulfide	2.526	76	19624	4.507	ug/l	100
18) Methyl Acetate	2.721	43	7530	4.924	ug/l	100
19) Methyl tert-butyl Ether	3.123	73	21254	4.734	ug/l	97
20) Methylene Chloride	2.800	84	8556	5.110	ug/l	97
21) trans-1,2-Dichloroethene	3.111	96	7425	4.813	ug/l	97
22) Diisopropyl ether	3.769	45	24580	4.895	ug/l #	67
23) Vinyl Acetate	3.739	43	92889	23.264	ug/l	98
24) 1,1-Dichloroethane	3.623	63	13761	4.827	ug/l	99
25) 2-Butanone	4.580	43	20821	24.452	ug/l	97
26) 2,2-Dichloropropane	4.483	77	10020	4.819	ug/l	96
27) cis-1,2-Dichloroethene	4.507	96	8891	4.906	ug/l	99
28) Bromochloromethane	4.910	49	6884	5.326	ug/l #	96
29) Tetrahydrofuran	5.025	42	13402	24.367	ug/l	99
30) Chloroform	5.099	83	14385	5.046	ug/l	86
31) Cyclohexane	5.483	56	12835	4.748	ug/l	94
32) 1,1,1-Trichloroethane	5.397	97	12150	4.859	ug/l	97
36) 1,1-Dichloropropene	5.702	75	10359	4.998	ug/l	99
37) Ethyl Acetate	4.733	43	8763	4.580	ug/l	96
38) Carbon Tetrachloride	5.690	117	11043	4.909	ug/l	98
39) Methylcyclohexane	7.391	83	13704	5.024	ug/l #	87
40) Benzene	6.050	78	31061	5.064	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
 Data File : VX046718.D
 Acq On : 17 Jun 2025 11:19
 Operator : JC/MD
 Sample : VSTDICC005
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC005

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
 Supervised By :Mahesh Dadoda 06/18/2025

Quant Time: Jun 18 02:37:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 02:34:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.940	41	4488	4.400	ug/l #	92
42) 1,2-Dichloroethane	6.104	62	11029	5.116	ug/l	98
43) Isopropyl Acetate	6.348	43	14711	4.914	ug/l	97
44) Trichloroethene	7.135	130	7703	4.927	ug/l	93
45) 1,2-Dichloropropane	7.433	63	7529	4.952	ug/l	97
46) Dibromomethane	7.592	93	5267	4.821	ug/l	95
47) Bromodichloromethane	7.824	83	11072	4.966	ug/l	96
48) Methyl methacrylate	7.702	41	6864	4.571	ug/l	99
49) 1,4-Dioxane	7.671	88	1949	113.091	ug/l	95
51) 4-Methyl-2-Pentanone	8.574	43	44282	24.800	ug/l	97
52) Toluene	8.720	92	19190	5.046	ug/l	99
53) t-1,3-Dichloropropene	8.982	75	9497	4.591	ug/l	99
54) cis-1,3-Dichloropropene	8.372	75	11188	4.764	ug/l	95
55) 1,1,2-Trichloroethane	9.153	97	7151	5.046	ug/l	95
56) Ethyl methacrylate	9.122	69	10101	4.828	ug/l	99
57) 1,3-Dichloropropane	9.311	76	12483	5.115	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.244	63	26158	24.071	ug/l	100
59) 2-Hexanone	9.433	43	30938	25.062	ug/l	99
60) Dibromochloromethane	9.518	129	8249	4.939	ug/l	99
61) 1,2-Dibromoethane	9.610	107	7221	5.009	ug/l	98
64) Tetrachloroethene	9.275	164	6755	4.980	ug/l	97
65) Chlorobenzene	10.079	112	21807	5.046	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.165	131	7006	4.823	ug/l	98
67) Ethyl Benzene	10.195	91	37295	4.939	ug/l	98
68) m/p-Xylenes	10.299	106	27609	9.812	ug/l	96
69) o-Xylene	10.640	106	13421	5.090	ug/l	97
70) Styrene	10.659	104	22403	4.855	ug/l	100
71) Bromoform	10.799	173	5248	4.744	ug/l #	100
73) Isopropylbenzene	10.963	105	36281	4.878	ug/l	99
74) N-amyl acetate	10.841	43	12496	4.434	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.213	83	10468	5.040	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	9687m	4.722	ug/l	
77) Bromobenzene	11.201	156	8694	4.857	ug/l	95
78) n-propylbenzene	11.305	91	42848	4.850	ug/l	100
79) 2-Chlorotoluene	11.366	91	26221	4.971	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	29167	4.772	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	2663	4.070	ug/l	94
82) 4-Chlorotoluene	11.457	91	30097	4.889	ug/l	99
83) tert-Butylbenzene	11.713	119	30728	4.864	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	30213	4.920	ug/l	98
85) sec-Butylbenzene	11.890	105	39588	4.958	ug/l	100
86) p-Isopropyltoluene	12.006	119	33076	4.909	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	17198	4.928	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	17600	4.910	ug/l	95
89) n-Butylbenzene	12.329	91	29990	4.688	ug/l	98
90) Hexachloroethane	12.536	117	5285	4.480	ug/l	97
91) 1,2-Dichlorobenzene	12.335	146	16203	4.968	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	1982	4.834	ug/l	94
93) 1,2,4-Trichlorobenzene	13.585	180	10507	4.532	ug/l	99
94) Hexachlorobutadiene	13.725	225	4953	5.078	ug/l	95
95) Naphthalene	13.774	128	29158	4.414	ug/l	99
96) 1,2,3-Trichlorobenzene	13.963	180	10725	4.835	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
Data File : VX046718.D
Acq On : 17 Jun 2025 11:19
Operator : JC/MD
Sample : VSTDICC005
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025

Quant Time: Jun 18 02:37:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 02:34:13 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
Data File : VX046718.D
Acq On : 17 Jun 2025 11:19
Operator : JC/MD
Sample : VSTDIC005
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1


Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025


Quant Time: Jun 18 02:37:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 02:34:13 2025
Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.562 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005



Tgt Ion:168 Resp: 130530
Ion Ratio Lower Upper
168 100
99 60.0 48.5 72.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#2
Dichlorodifluoromethane
Concen: 4.342 ug/l
RT: 1.179 min Scan# 15
Delta R.T. -0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Tgt Ion: 85 Resp: 6051
Ion Ratio Lower Upper
85 100
87 32.7 16.6 49.8









#3
Chloromethane
Concen: 4.698 ug/l
RT: 1.307 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Tgt Ion: 50 Resp: 7060
Ion Ratio Lower Upper
50 100
52 36.4 24.9 37.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#4
Vinyl Chloride
Concen: 4.629 ug/l
RT: 1.386 min Scan# 49
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Tgt Ion: 62 Resp: 7425
Ion Ratio Lower Upper
62 100
64 29.3 23.9 35.9









#5
Bromomethane
Concen: 5.360 ug/l
RT: 1.617 min Scan# 81
Delta R.T. -0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005

Tgt Ion: 94 Resp: 483
Ion Ratio Lower Upper
94 100
96 95.3 72.3 108.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#6
Chloroethane
Concen: 4.706 ug/l
RT: 1.703 min Scan# 101
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19


Tgt Ion: 64 Resp: 4575
Ion Ratio Lower Upper
64 100
66 29.4 27.1 40.7





#7
Trichlorofluoromethane
Concen: 4.810 ug/l
RT: 1.904 min Scan# 11714
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005



Tgt Ion:101 Resp: 11714
Ion Ratio Lower Upper
101 100
103 70.9 51.8 77.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#8
Diethyl Ether
Concen: 4.831 ug/l
RT: 2.148 min Scan# 174
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Tgt Ion: 74 Resp: 4038
Ion Ratio Lower Upper
74 100
45 106.5 52.3 157.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 4.921 ug/l
RT: 2.349 min Scan# 206
Delta R.T. 0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005

Tgt Ion:101 Resp: 735
Ion Ratio Lower Upper
101 100
85 43.2 35.4 53.0
151 71.5 61.2 91.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#10
Methyl Iodide
Concen: 7.181 ug/l
RT: 2.465 min Scan# 226
Delta R.T. -0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19


Tgt Ion:142 Resp: 5830
Ion Ratio Lower Upper
142 100
127 45.5 37.0 55.6
141 14.6 11.5 17.3





#11
Tert butyl alcohol
Concen: 19.340 ug/l
RT: 2.959 min Scan# 307
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005




Tgt Ion: 59 Resp: 3134
Ion Ratio Lower Upper
59 100
57 10.9 8.3 12.5

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025




#12
1,1-Dichloroethene
Concen: 4.775 ug/l
RT: 2.337 min Scan# 205
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19



Tgt Ion: 96 Resp: 6884
Ion Ratio Lower Upper
96 100
61 171.4 126.6 189.8
98 68.2 50.7 76.1





#13

Acrolein

Concen: 26.301 ug/l

RT: 2.251 min Scan# 190

Delta R.T. 0.006 min

Lab File: VX046718.D

Acq: 17 Jun 2025 11:19

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC005

Tgt Ion: 56 Resp: 6208

Ion Ratio Lower Upper

56 100





55 68.8 55.2 82.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#14

Allyl chloride

Concen: 4.668 ug/l

RT: 2.678 min Scan# 261

Delta R.T. -0.000 min

Lab File: VX046718.D

Acq: 17 Jun 2025 11:19




Tgt Ion: 41 Resp: 12195


Ion Ratio Lower Upper

41 100

39 68.1 53.6 80.4


76 33.8 26.8 40.2





#15
Acrylonitrile
Concen: 24.169 ug/l
RT: 3.081 min Scan# 31
Delta R.T. 0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005




Tgt Ion: 53 Resp: 1761
Ion Ratio Lower Upper
53 100
52 81.8 65.1 97.7
51 35.6 28.9 43.3

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025




#16
Acetone
Concen: 23.710 ug/l
RT: 2.392 min Scan# 214
Delta R.T. 0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19




Tgt Ion: 43 Resp: 14275
Ion Ratio Lower Upper
43 100
58 31.7 24.5 36.7





#17
Carbon Disulfide
Concen: 4.507 ug/l
RT: 2.526 min Scan# 21
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005




Tgt Ion: 76 Resp: 19624
Ion Ratio Lower Upper
76 100
78 9.3 7.5 11.3

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025




#18
Methyl Acetate
Concen: 4.924 ug/l
RT: 2.721 min Scan# 268
Delta R.T. 0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19



Tgt Ion: 43 Resp: 7530
Ion Ratio Lower Upper
43 100
74 22.8 18.4 27.6









#19
Methyl tert-butyl Ether
Concen: 4.734 ug/l
RT: 3.123 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005

Tgt Ion: 73 Resp: 21254
Ion Ratio Lower Upper
73 100
57 24.0 18.2 27.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#20
Methylene Chloride
Concen: 5.110 ug/l
RT: 2.800 min Scan# 281
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Tgt Ion: 84 Resp: 8556
Ion Ratio Lower Upper
84 100
49 130.6 101.3 151.9
51 37.6 31.0 46.6
86 65.2 51.1 76.7









#21
trans-1,2-Dichloroethene
Concen: 4.813 ug/l
RT: 3.111 min Scan# 311
Delta R.T. 0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Tgt Ion: 96 Resp: 7425
Ion Ratio Lower Upper
96 100
61 144.2 112.5 168.7
98 68.5 51.9 77.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#22
Diisopropyl ether
Concen: 4.895 ug/l
RT: 3.769 min Scan# 440
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Tgt Ion: 45 Resp: 24580
Ion Ratio Lower Upper
45 100
43 160.9 91.9 137.9#
87 25.1 22.0 33.0
59 11.3 9.1 13.7









#23
 Vinyl Acetate
 Concen: 23.264 ug/l
 RT: 3.739 min Scan# 41
 Delta R.T. 0.006 min
 Lab File: VX046718.D
 Acq: 17 Jun 2025 11:19

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC005

Tgt Ion: 43 Resp: 92889
 Ion Ratio Lower Upper
 43 100
 86 9.5 8.2 12.2




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
 Supervised By :Mahesh Dadoda 06/18/2025



#24
 1,1-Dichloroethane
 Concen: 4.827 ug/l
 RT: 3.623 min Scan# 416
 Delta R.T. -0.000 min
 Lab File: VX046718.D
 Acq: 17 Jun 2025 11:19

Tgt Ion: 63 Resp: 13761
 Ion Ratio Lower Upper
 63 100
 98 8.5 4.0 11.9
 100 4.2 2.2 6.6









#25
2-Butanone
Concen: 24.452 ug/l
RT: 4.580 min Scan# 51
Delta R.T. 0.018 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005

Tgt Ion: 43 Resp: 2082
Ion Ratio Lower Upper
43 100
72 26.8 20.2 30.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#26
2,2-Dichloropropane
Concen: 4.819 ug/l
RT: 4.483 min Scan# 557
Delta R.T. -0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19


Tgt Ion: 77 Resp: 10020
Ion Ratio Lower Upper
77 100
97 21.7 11.8 35.3





#27
cis-1,2-Dichloroethene
Concen: 4.906 ug/l
RT: 4.507 min Scan# 560
Delta R.T. 0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005



Tgt Ion: 96 Resp: 889
Ion Ratio Lower Upper
96 100
61 146.2 0.0 291.0
98 64.6 0.0 126.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#28
Bromochloromethane
Concen: 5.326 ug/l
RT: 4.910 min Scan# 627
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19


Tgt Ion: 49 Resp: 6884
Ion Ratio Lower Upper
49 100
129 0.0 0.0 3.8
130 73.3 61.4 92.0





#29
Tetrahydrofuran
Concen: 24.367 ug/l
RT: 5.025 min Scan# 644
Delta R.T. 0.012 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005



Tgt Ion: 42 Resp: 1340
Ion Ratio Lower Upper
42 100
72 44.7 34.9 52.3
71 40.5 32.2 48.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#30
Chloroform
Concen: 5.046 ug/l
RT: 5.099 min Scan# 658
Delta R.T. -0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Tgt Ion: 83 Resp: 14385
Ion Ratio Lower Upper
83 100
85 56.2 53.7 80.5









#31
Cyclohexane
Concen: 4.748 ug/l
RT: 5.483 min Scan# 721
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005

Tgt Ion: 56 Resp: 1283
Ion Ratio Lower Upper
56 100
69 28.2 24.8 37.2
84 79.6 68.2 102.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#32
1,1,1-Trichloroethane
Concen: 4.859 ug/l
RT: 5.397 min Scan# 707
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Tgt Ion: 97 Resp: 12150
Ion Ratio Lower Upper
97 100
99 65.5 51.3 76.9
61 47.8 36.4 54.6









#33
 1,2-Dichloroethane-d4
 Concen: 5.791 ug/l
 RT: 5.964 min Scan# 800
 Delta R.T. -0.000 min
 Lab File: VX046718.D
 Acq: 17 Jun 2025 11:19

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC005

Tgt Ion: 65 Resp: 10450
 Ion Ratio Lower Upper
 65 100
 67 50.9 0.0 105.4




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
 Supervised By :Mahesh Dadoda 06/18/2025



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.769 min Scan# 932
 Delta R.T. -0.000 min
 Lab File: VX046718.D
 Acq: 17 Jun 2025 11:19


Tgt Ion:114 Resp: 217019
 Ion Ratio Lower Upper
 114 100
 63 21.8 0.0 44.2
 88 16.4 0.0 33.2





#35
Dibromofluoromethane
Concen: 5.478 ug/l
RT: 5.403 min Scan# 707
Delta R.T. 0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005



Tgt Ion: 113 Resp: 786
Ion Ratio Lower Upper
113 100
111 104.7 82.0 123.0
192 16.4 15.3 22.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#36
1,1-Dichloropropene
Concen: 4.998 ug/l
RT: 5.702 min Scan# 757
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Tgt Ion: 75 Resp: 10359
Ion Ratio Lower Upper
75 100
110 34.8 17.2 51.5
77 30.4 24.2 36.4





#37

Ethyl Acetate

Concen: 4.580 ug/l

RT: 4.733 min Scan# 598

Delta R.T. 0.012 min

Lab File: VX046718.D

Acq: 17 Jun 2025 11:19

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC005





Tgt Ion: 43 Resp: 876

Ion	Ratio	Lower	Upper
43	100		
61	10.1	9.7	14.5
70	8.9	8.1	12.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#38

Carbon Tetrachloride

Concen: 4.909 ug/l

RT: 5.690 min Scan# 755




Delta R.T. -0.000 min


Lab File: VX046718.D

Acq: 17 Jun 2025 11:19

Tgt Ion:117 Resp: 11043

Ion	Ratio	Lower	Upper
117	100		
119	95.6	75.2	112.8
121	31.6	24.1	36.1









#39
Methylcyclohexane
Concen: 5.024 ug/l
RT: 7.391 min Scan# 1034
Delta R.T. 0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Tgt Ion: 83 Resp: 13704
Ion Ratio Lower Upper
83 100
55 69.0 63.0 94.4
98 38.4 38.8 58.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#40
Benzene
Concen: 5.064 ug/l
RT: 6.050 min Scan# 814
Delta R.T. 0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19


Tgt Ion: 78 Resp: 31061
Ion Ratio Lower Upper
78 100
77 25.1 18.8 28.2





#41
Methacrylonitrile
Concen: 4.400 ug/l
RT: 4.940 min Scan# 61
Delta R.T. 0.012 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005



Tgt Ion: 41 Resp: 448
Ion Ratio Lower Upper
41 100
39 61.2 45.9 68.9
67 74.8 61.2 91.8
52 17.5 25.5 38.3#




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#42
1,2-Dichloroethane
Concen: 5.116 ug/l
RT: 6.104 min Scan# 823
Delta R.T. 0.012 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19


Tgt Ion: 62 Resp: 11029
Ion Ratio Lower Upper
62 100
98 8.1 0.0 17.8





#43
Isopropyl Acetate
Concen: 4.914 ug/l
RT: 6.348 min Scan# 863
Delta R.T. 0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005



Tgt Ion: 43 Resp: 1471
Ion Ratio Lower Upper
43 100
61 17.3 15.3 22.9
87 11.6 10.0 15.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#44
Trichloroethene
Concen: 4.927 ug/l
RT: 7.135 min Scan# 992
Delta R.T. 0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Tgt Ion:130 Resp: 7703
Ion Ratio Lower Upper
130 100
95 94.6 0.0 202.6





#45

1,2-Dichloropropane

Concen: 4.952 ug/l

RT: 7.433 min Scan# 1041

Delta R.T. -0.000 min

Lab File: VX046718.D

Acq: 17 Jun 2025 11:19

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 63 Resp: 7529

Ion Ratio Lower Upper

63 100





65 30.2 22.9 34.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#46

Dibromomethane

Concen: 4.821 ug/l

RT: 7.592 min Scan# 1067

Delta R.T. 0.006 min

Lab File: VX046718.D

Acq: 17 Jun 2025 11:19




Tgt Ion: 93 Resp: 5267


Ion Ratio Lower Upper

93 100

95 86.9 66.2 99.4


174 102.0 77.0 115.6





#47
Bromodichloromethane
Concen: 4.966 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005




Tgt Ion: 83 Resp: 11072
Ion Ratio Lower Upper
83 100
85 61.1 51.4 77.2
127 7.9 6.9 10.3

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025




#48
Methyl methacrylate
Concen: 4.571 ug/l
RT: 7.702 min Scan# 1085
Delta R.T. 0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19




Tgt Ion: 41 Resp: 6864
Ion Ratio Lower Upper
41 100
69 79.8 63.4 95.2
39 58.5 45.4 68.0





#49
1,4-Dioxane
Concen: 113.091 ug/l
RT: 7.671 min Scan# 1080
Delta R.T. 0.012 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005



Tgt Ion: 88 Resp: 1949
Ion Ratio Lower Upper
88 100
43 44.9 31.2 46.8
58 70.8 54.7 82.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#50
Toluene-d8
Concen: 5.645 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. 0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Tgt Ion: 98 Resp: 29120
Ion Ratio Lower Upper
98 100
100 68.4 53.0 79.4









#51
4-Methyl-2-Pentanone
Concen: 24.800 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005

Tgt Ion: 43 Resp: 4428
Ion Ratio Lower Upper
43 100
58 36.8 30.8 46.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#52
Toluene
Concen: 5.046 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Tgt Ion: 92 Resp: 19190
Ion Ratio Lower Upper
92 100
91 170.2 134.8 202.2





#53

t-1,3-Dichloropropene

Concen: 4.591 ug/l

RT: 8.982 min Scan# 1195

Delta R.T. 0.006 min

Lab File: VX046718.D

Acq: 17 Jun 2025 11:19

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 75 Resp: 949

Ion Ratio Lower Upper

75 100





77 32.7 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#54

cis-1,3-Dichloropropene

Concen: 4.764 ug/l

RT: 8.372 min Scan# 1195

Delta R.T. 0.006 min

Lab File: VX046718.D

Acq: 17 Jun 2025 11:19




Tgt Ion: 75 Resp: 11188


Ion Ratio Lower Upper

75 100

77 31.7 25.3 37.9

39 54.9 39.2 58.8





#55

1,1,2-Trichloroethane

Concen: 5.046 ug/l

RT: 9.153 min Scan# 1323

Delta R.T. -0.000 min

Lab File: VX046718.D

Acq: 17 Jun 2025 11:19

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC005

Tgt Ion:	97	Resp:	715
Ion	Ratio	Lower	Upper
97	100		
83	90.8	67.2	100.8
85	55.8	44.1	66.1
99	58.5	50.3	75.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#56

Ethyl methacrylate

Concen: 4.828 ug/l




RT: 9.122 min Scan# 1318


Delta R.T. 0.006 min

Lab File: VX046718.D

Acq: 17 Jun 2025 11:19

Tgt Ion:	69	Resp:	10101
Ion	Ratio	Lower	Upper
69	100		
41	68.6	55.4	83.2
39	36.5	29.7	44.5





#57

1,3-Dichloropropane

Concen: 5.115 ug/l

RT: 9.311 min Scan# 1134

Delta R.T. 0.006 min

Lab File: VX046718.D

Acq: 17 Jun 2025 11:19

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 76 Resp: 1248

Ion Ratio Lower Upper

76 100





78 31.7 25.7 38.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#58

2-Chloroethyl Vinyl ether

Concen: 24.071 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. -0.000 min

Lab File: VX046718.D




Acq: 17 Jun 2025 11:19


Tgt Ion: 63 Resp: 26158

Ion Ratio Lower Upper

63 100

106 28.2 22.6 34.0









#59
2-Hexanone
Concen: 25.062 ug/l
RT: 9.433 min Scan# 1369
Delta R.T. 0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Tgt Ion: 43 Resp: 30938
Ion Ratio Lower Upper
43 100
58 52.6 26.8 80.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#60
Dibromochloromethane
Concen: 4.939 ug/l
RT: 9.518 min Scan# 1383
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19


Tgt Ion:129 Resp: 8249
Ion Ratio Lower Upper
129 100
127 76.2 38.4 115.2





#61
1,2-Dibromoethane
Concen: 5.009 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005



Tgt Ion:107 Resp: 722
Ion Ratio Lower Upper
107 100
109 92.8 75.8 113.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#62
4-Bromofluorobenzene
Concen: 5.791 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19


Tgt Ion: 95 Resp: 11137
Ion Ratio Lower Upper
95 100
174 77.4 0.0 150.4
176 73.9 0.0 145.0





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005



Tgt Ion:117 Resp: 19578
Ion Ratio Lower Upper
117 100
82 55.2 44.6 66.8
119 32.2 25.8 38.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#64
Tetrachloroethene
Concen: 4.980 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Tgt Ion:164 Resp: 6755
Ion Ratio Lower Upper
164 100
166 125.4 104.6 157.0
129 94.1 76.3 114.5
131 89.4 74.1 111.1









#65
Chlorobenzene
Concen: 5.046 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005

Tgt Ion:112 Resp: 2180
Ion Ratio Lower Upper
112 100
114 31.6 25.5 38.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 4.823 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. 0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Tgt Ion:131 Resp: 7006
Ion Ratio Lower Upper
131 100
133 94.8 47.3 141.8
119 68.5 32.1 96.3









#67
Ethyl Benzene
Concen: 4.939 ug/l
RT: 10.195 min Scan# 1493
Delta R.T. 0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005

Tgt Ion: 91 Resp: 37295
Ion Ratio Lower Upper
91 100
106 29.0 24.2 36.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#68
m/p-Xylenes
Concen: 9.812 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Tgt Ion:106 Resp: 27609
Ion Ratio Lower Upper
106 100
91 212.5 164.8 247.2









#69
o-Xylene
Concen: 5.090 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005

Tgt Ion:106 Resp: 1342
Ion Ratio Lower Upper
106 100
91 218.1 111.3 333.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#70
Styrene
Concen: 4.855 ug/l
RT: 10.659 min Scan# 1570
Delta R.T. 0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Tgt Ion:104 Resp: 22403
Ion Ratio Lower Upper
104 100
78 52.1 41.4 62.2
103 54.0 43.1 64.7





#71

Bromoform

Concen: 4.744 ug/l

RT: 10.799 min Scan# 1593

Delta R.T. -0.000 min

Lab File: VX046718.D

Acq: 17 Jun 2025 11:19

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC005

Tgt Ion:173	Resp:	524
Ion Ratio	Lower	Upper
173	100	
175	48.3	24.1 72.4
254	0.0	0.1 0.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 12.018 min Scan# 1793


Delta R.T. -0.000 min

Lab File: VX046718.D

Acq: 17 Jun 2025 11:19

Tgt Ion:152	Resp:	100990
Ion Ratio	Lower	Upper
152	100	
115	64.0	43.2 129.6
150	156.3	0.0 346.8





#73

Isopropylbenzene

Concen: 4.878 ug/l

RT: 10.963 min Scan# 1620

Delta R.T. 0.006 min

Lab File: VX046718.D

Acq: 17 Jun 2025 11:19

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC005

Tgt Ion:105 Resp: 3628
 Ion Ratio Lower Upper
 105 100
 120 26.3 13.0 39.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
 Supervised By :Mahesh Dadoda 06/18/2025



#74

N-ethyl acetate

Concen: 4.434 ug/l




RT: 10.841 min Scan# 1600


Delta R.T. -0.000 min

Lab File: VX046718.D

Acq: 17 Jun 2025 11:19

Tgt Ion: 43 Resp: 12496
 Ion Ratio Lower Upper
 43 100
 70 41.0 33.4 50.0
 55 23.9 19.1 28.7
 61 22.9 19.0 28.4









#75
1,1,2,2-Tetrachloroethane
Concen: 5.040 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. 0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005

Tgt Ion: 83 Resp: 10468
Ion Ratio Lower Upper
83 100
131 9.0 4.9 14.6
85 64.4 32.0 96.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#76
1,2,3-Trichloropropane
Concen: 4.722 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Tgt Ion: 75 Resp: 9687
Ion Ratio Lower Upper
75 100
77 34.2 18.5 55.5





#77

Bromobenzene

Concen: 4.857 ug/l

RT: 11.201 min Scan# 1659

Delta R.T. 0.006 min

Lab File: VX046718.D

Acq: 17 Jun 2025 11:19

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC005

Tgt Ion:	156	Resp:	8694
Ion Ratio	Lower	Upper	
156	100		
77	173.3	82.8	248.6
158	99.4	47.6	142.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#78

n-propylbenzene

Concen: 4.850 ug/l




RT: 11.305 min Scan# 1676


Delta R.T. 0.006 min

Lab File: VX046718.D

Acq: 17 Jun 2025 11:19

Tgt Ion:	91	Resp:	42848
Ion Ratio	Lower	Upper	
91	100		
120	22.7	11.4	34.1









#79
2-Chlorotoluene
Concen: 4.971 ug/l
RT: 11.366 min Scan# 1686
Delta R.T. 0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005

Tgt Ion: 91 Resp: 2622
Ion Ratio Lower Upper
91 100
126 33.1 16.6 49.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#80
1,3,5-Trimethylbenzene
Concen: 4.772 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Tgt Ion:105 Resp: 29167
Ion Ratio Lower Upper
105 100
120 47.7 24.1 72.2









#81
trans-1,4-Dichloro-2-butene
Concen: 4.070 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005

Tgt Ion: 75 Resp: 266
Ion Ratio Lower Upper
75 100
53 114.8 86.8 130.2
89 42.7 36.9 55.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#82
4-Chlorotoluene
Concen: 4.889 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19


Tgt Ion: 91 Resp: 30097
Ion Ratio Lower Upper
91 100
126 28.4 14.4 43.2





#83
tert-Butylbenzene
Concen: 4.864 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005




Tgt Ion:119 Resp: 30728
Ion Ratio Lower Upper
119 100
91 60.3 29.5 88.6
134 23.0 11.7 35.0

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025




#84
1,2,4-Trimethylbenzene
Concen: 4.920 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19



Tgt Ion:105 Resp: 30213
Ion Ratio Lower Upper
105 100
120 45.2 22.1 66.3





#85
 sec-Butylbenzene
 Concen: 4.958 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. -0.000 min
 Lab File: VX046718.D
 Acq: 17 Jun 2025 11:19

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC005

Tgt Ion:105 Resp: 3958
 Ion Ratio Lower Upper
 105 100
 134 19.6 9.8 29.4

Manual Integrations




APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
 Supervised By :Mahesh Dadoda 06/18/2025



#86
 p-Isopropyltoluene
 Concen: 4.909 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. -0.000 min
 Lab File: VX046718.D
 Acq: 17 Jun 2025 11:19

Tgt Ion:119 Resp: 33076
 Ion Ratio Lower Upper
 119 100
 134 26.2 12.9 38.7
 91 23.9 12.3 36.9





#87

1,3-Dichlorobenzene

Concen: 4.928 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX046718.D

Acq: 17 Jun 2025 11:19

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion:146 Resp: 17198





Ion	Ratio	Lower	Upper
146	100		
111	42.0	20.6	61.8
148	63.3	31.6	94.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#88

1,4-Dichlorobenzene

Concen: 4.910 ug/l

RT: 12.036 min Scan# 1796




Delta R.T. -0.000 min


Lab File: VX046718.D

Acq: 17 Jun 2025 11:19

Tgt Ion:146 Resp: 17600

Ion	Ratio	Lower	Upper
146	100		
111	46.6	20.3	61.0
148	66.0	32.1	96.5









#89
n-Butylbenzene
Concen: 4.688 ug/l
RT: 12.329 min Scan# 1844
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005

Tgt Ion: 91 Resp: 29990
Ion Ratio Lower Upper
91 100
92 55.6 27.1 81.3
134 26.3 12.9 38.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#90
Hexachloroethane
Concen: 4.480 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19


Tgt Ion:117 Resp: 5285
Ion Ratio Lower Upper
117 100
201 71.1 34.4 103.3





#91
1,2-Dichlorobenzene
Concen: 4.968 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005




Tgt Ion:146 Resp: 1620
Ion Ratio Lower Upper
146 100
111 44.1 21.7 65.1
148 64.0 32.0 96.2

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 4.834 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19



Tgt Ion: 75 Resp: 1982
Ion Ratio Lower Upper
75 100
155 72.2 41.9 125.7
157 102.4 51.4 154.3









#93
1,2,4-Trichlorobenzene
Concen: 4.532 ug/l
RT: 13.585 min Scan# 2050
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005

Tgt Ion:180 Resp: 1050
Ion Ratio Lower Upper
180 100
182 94.5 46.7 140.1
145 31.8 15.8 47.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#94
Hexachlorobutadiene
Concen: 5.078 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19


Tgt Ion:225 Resp: 4953
Ion Ratio Lower Upper
225 100
223 61.5 32.5 97.5
227 60.3 32.2 96.6





#95
Naphthalene
Concen: 4.414 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005




Tgt Ion:128 Resp: 2915
Ion Ratio Lower Upper
128 100
127 12.6 10.1 15.1
129 10.3 8.7 13.1

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#96
1,2,3-Trichlorobenzene
Concen: 4.835 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. 0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19



Tgt Ion:180 Resp: 10725
Ion Ratio Lower Upper
180 100
182 92.3 47.8 143.4
145 32.0 17.3 51.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
 Data File : VX046719.D
 Acq On : 17 Jun 2025 13:59
 Operator : JC/MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC020

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
 Supervised By :Mahesh Dadoda 06/18/2025

Quant Time: Jun 18 02:38:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 02:34:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	134732	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	226567	50.000	ug/l	-0.01
63) Chlorobenzene-d5	10.049	117	197947	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	98285	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	30582	16.410	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	32.820%#		
35) Dibromofluoromethane	5.385	113	24156	16.114	ug/l	-0.01
Spiked Amount 50.000	Range 75 - 124		Recovery =	32.220%#		
50) Toluene-d8	8.647	98	88197	16.376	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	32.760%#		
62) 4-Bromofluorobenzene	11.079	95	32067	15.973	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	31.940%#		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.185	85	27297	18.977	ug/l	97
3) Chloromethane	1.307	50	30731	19.793	ug/l	98
4) Vinyl Chloride	1.386	62	34049	20.565	ug/l	96
5) Bromomethane	1.618	94	19805	21.261	ug/l	91
6) Chloroethane	1.697	64	20911	20.838	ug/l	97
7) Trichlorofluoromethane	1.904	101	52480	20.875	ug/l	99
8) Diethyl Ether	2.142	74	18139	21.024	ug/l	100
9) 1,1,2-Trichlorotrifluo...	2.343	101	32778	21.253	ug/l	98
10) Methyl Iodide	2.465	142	30525	18.775	ug/l	99
11) Tert butyl alcohol	2.953	59	16360	97.808	ug/l	96
12) 1,1-Dichloroethene	2.331	96	30924	20.780	ug/l	92
13) Acrolein	2.246	56	20036	82.237	ug/l	98
14) Allyl chloride	2.672	41	56147	20.823	ug/l	99
15) Acrylonitrile	3.069	53	77278	102.746	ug/l	99
16) Acetone	2.386	43	59747	96.140	ug/l	97
17) Carbon Disulfide	2.526	76	88625	19.721	ug/l	99
18) Methyl Acetate	2.709	43	31812	20.154	ug/l	99
19) Methyl tert-butyl Ether	3.117	73	97187	20.974	ug/l	95
20) Methylene Chloride	2.800	84	35302	20.426	ug/l	99
21) trans-1,2-Dichloroethene	3.105	96	33796	21.224	ug/l	95
22) Diisopropyl ether	3.764	45	111712	21.555	ug/l	95
23) Vinyl Acetate	3.727	43	433313	105.140	ug/l	100
24) 1,1-Dichloroethane	3.617	63	62164	21.126	ug/l	98
25) 2-Butanone	4.556	43	87671	99.748	ug/l	99
26) 2,2-Dichloropropane	4.483	77	42992	20.032	ug/l	99
27) cis-1,2-Dichloroethene	4.495	96	39374	21.047	ug/l	99
28) Bromochloromethane	4.904	49	24738	18.544	ug/l	100
29) Tetrahydrofuran	5.007	42	56765	99.990	ug/l	98
30) Chloroform	5.093	83	64109	21.785	ug/l	100
31) Cyclohexane	5.477	56	58552	20.982	ug/l	98
32) 1,1,1-Trichloroethane	5.385	97	54542	21.133	ug/l	99
36) 1,1-Dichloropropene	5.696	75	44023	20.344	ug/l	99
37) Ethyl Acetate	4.715	43	39916	19.984	ug/l	99
38) Carbon Tetrachloride	5.684	117	48694	20.735	ug/l	97
39) Methylcyclohexane	7.379	83	58170	20.428	ug/l	97
40) Benzene	6.038	78	133872	20.904	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
 Data File : VX046719.D
 Acq On : 17 Jun 2025 13:59
 Operator : JC/MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC020

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
 Supervised By :Mahesh Dadoda 06/18/2025

Quant Time: Jun 18 02:38:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 02:34:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	21970	20.632	ug/l	97
42) 1,2-Dichloroethane	6.086	62	47430	21.076	ug/l	99
43) Isopropyl Acetate	6.342	43	62432	19.975	ug/l	100
44) Trichloroethene	7.129	130	33927	20.788	ug/l	100
45) 1,2-Dichloropropane	7.428	63	33688	21.226	ug/l	94
46) Dibromomethane	7.580	93	23569	20.663	ug/l	99
47) Bromodichloromethane	7.818	83	49411	21.227	ug/l	98
48) Methyl methacrylate	7.690	41	31856	20.322	ug/l	98
49) 1,4-Dioxane	7.659	88	8340	382.035	ug/l	93
51) 4-Methyl-2-Pentanone	8.568	43	187509	100.586	ug/l	99
52) Toluene	8.714	92	84030	21.166	ug/l	100
53) t-1,3-Dichloropropene	8.976	75	43852	20.304	ug/l	99
54) cis-1,3-Dichloropropene	8.366	75	50326	20.526	ug/l	98
55) 1,1,2-Trichloroethane	9.147	97	30865	20.862	ug/l	94
56) Ethyl methacrylate	9.116	69	44702	20.468	ug/l	99
57) 1,3-Dichloropropane	9.305	76	52181	20.480	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.238	63	106505	93.878	ug/l	99
59) 2-Hexanone	9.427	43	129218	100.265	ug/l	99
60) Dibromochloromethane	9.519	129	36392	20.872	ug/l	99
61) 1,2-Dibromoethane	9.604	107	31507	20.936	ug/l	100
64) Tetrachloroethene	9.269	164	27988	20.408	ug/l	97
65) Chlorobenzene	10.073	112	90868	20.796	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.159	131	31508	21.455	ug/l	100
67) Ethyl Benzene	10.189	91	160761	21.056	ug/l	99
68) m/p-Xylenes	10.299	106	121014	42.539	ug/l	99
69) o-Xylene	10.640	106	57689	21.638	ug/l	100
70) Styrene	10.653	104	99462	21.318	ug/l	99
71) Bromoform	10.799	173	23372	20.898	ug/l #	100
73) Isopropylbenzene	10.957	105	153861	21.255	ug/l	99
74) N-amyl acetate	10.842	43	57185	20.851	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	42272	20.914	ug/l	100
76) 1,2,3-Trichloropropane	11.238	75	41901m	20.989	ug/l	
77) Bromobenzene	11.195	156	37464	21.507	ug/l	99
78) n-propylbenzene	11.299	91	184078	21.410	ug/l	100
79) 2-Chlorotoluene	11.360	91	107552	20.952	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	129635	21.793	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	11694	18.362	ug/l	98
82) 4-Chlorotoluene	11.451	91	127954	21.358	ug/l	100
83) tert-Butylbenzene	11.713	119	130359	21.201	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	128566	21.512	ug/l	99
85) sec-Butylbenzene	11.890	105	165047	21.239	ug/l	99
86) p-Isopropyltoluene	12.006	119	138623	21.142	ug/l	100
87) 1,3-Dichlorobenzene	11.969	146	70243	20.681	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	71953	20.625	ug/l	99
89) n-Butylbenzene	12.329	91	129101	20.738	ug/l	100
90) Hexachloroethane	12.536	117	23645	20.594	ug/l	97
91) 1,2-Dichlorobenzene	12.329	146	66883	21.073	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	8071	20.226	ug/l	98
93) 1,2,4-Trichlorobenzene	13.585	180	44726	19.821	ug/l	97
94) Hexachlorobutadiene	13.725	225	19698	20.752	ug/l	99
95) Naphthalene	13.774	128	129768	20.186	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	44380	20.558	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
Data File : VX046719.D
Acq On : 17 Jun 2025 13:59
Operator : JC/MD
Sample : VSTDICC020
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025

Quant Time: Jun 18 02:38:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 02:34:13 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Tgt Ion:168 Resp: 134732
Ion Ratio Lower Upper
168 100
99 58.4 48.5 72.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#2
Dichlorodifluoromethane
Concen: 18.977 ug/l
RT: 1.185 min Scan# 16
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Tgt Ion: 85 Resp: 27297
Ion Ratio Lower Upper
85 100
87 31.6 16.6 49.8









#3
Chloromethane
Concen: 19.793 ug/l
RT: 1.307 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Tgt Ion: 50 Resp: 3073
Ion Ratio Lower Upper
50 100
52 32.1 24.9 37.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#4
Vinyl Chloride
Concen: 20.565 ug/l
RT: 1.386 min Scan# 49
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Tgt Ion: 62 Resp: 34049
Ion Ratio Lower Upper
62 100
64 32.0 23.9 35.9









#5
Bromomethane
Concen: 21.261 ug/l
RT: 1.618 min Scan# 81
Delta R.T. -0.006 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Tgt Ion: 94 Resp: 1980
Ion Ratio Lower Upper
94 100
96 99.2 72.3 108.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#6
Chloroethane
Concen: 20.838 ug/l
RT: 1.697 min Scan# 100
Delta R.T. -0.006 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59


Tgt Ion: 64 Resp: 20911
Ion Ratio Lower Upper
64 100
66 32.4 27.1 40.7





#7
Trichlorofluoromethane
Concen: 20.875 ug/l
RT: 1.904 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020



Tgt Ion:101 Resp: 52480
Ion Ratio Lower Upper
101 100
103 65.7 51.8 77.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#8
Diethyl Ether
Concen: 21.024 ug/l
RT: 2.142 min Scan# 173
Delta R.T. -0.006 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Tgt Ion: 74 Resp: 18139
Ion Ratio Lower Upper
74 100
45 104.8 52.3 157.1





#9

1,1,2-Trichlorotrifluoroethane

Concen: 21.253 ug/l

RT: 2.343 min Scan# 206

Delta R.T. -0.000 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC020

Tgt Ion:	101	Resp:	32778
Ion	Ratio	Lower	Upper
101	100		
85	43.4	35.4	53.0
151	74.1	61.2	91.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#10

Methyl Iodide

Concen: 18.775 ug/l




RT: 2.465 min Scan# 226


Delta R.T. -0.006 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Tgt Ion:	142	Resp:	30525
Ion	Ratio	Lower	Upper
142	100		
127	46.7	37.0	55.6
141	13.7	11.5	17.3









#11
Tert butyl alcohol
Concen: 97.808 ug/l
RT: 2.953 min Scan# 307
Delta R.T. -0.006 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Tgt Ion: 59 Resp: 16360
Ion Ratio Lower Upper
59 100
57 8.9 8.3 12.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#12
1,1-Dichloroethene
Concen: 20.780 ug/l
RT: 2.331 min Scan# 204
Delta R.T. -0.006 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Tgt Ion: 96 Resp: 30924
Ion Ratio Lower Upper
96 100
61 171.1 126.6 189.8
98 64.9 50.7 76.1





#13

Acrolein

Concen: 82.237 ug/l

RT: 2.246 min Scan# 190

Delta R.T. -0.000 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 56 Resp: 20030

Ion Ratio Lower Upper

56 100





55 70.4 55.2 82.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#14

Allyl chloride

Concen: 20.823 ug/l

RT: 2.672 min Scan# 260

Delta R.T. -0.006 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59




Tgt Ion: 41 Resp: 56147


Ion Ratio Lower Upper

41 100

39 66.0 53.6 80.4

76 33.0 26.8 40.2





#15

Acrylonitrile

Concen: 102.746 ug/l

RT: 3.069 min Scan# 31

Delta R.T. -0.006 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC020

Tgt Ion: 53	Resp: 77278
Ion Ratio	Lower Upper
53 100	
52 82.5	65.1 97.7
51 35.1	28.9 43.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#16

Acetone

Concen: 96.140 ug/l




RT: 2.386 min Scan# 213


Delta R.T. -0.000 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Tgt Ion: 43	Resp: 59747
Ion Ratio	Lower Upper
43 100	
58 28.7	24.5 36.7





#17

Carbon Disulfide

Concen: 19.721 ug/l

RT: 2.526 min Scan# 21

Delta R.T. -0.000 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 76 Resp: 8862

Ion Ratio Lower Upper

76 100





78 8.9 7.5 11.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#18

Methyl Acetate

Concen: 20.154 ug/l

RT: 2.709 min Scan# 266

Delta R.T. -0.006 min

Lab File: VX046719.D




Acq: 17 Jun 2025 13:59


Tgt Ion: 43 Resp: 31812

Ion Ratio Lower Upper

43 100

74 22.7 18.4 27.6









#19
Methyl tert-butyl Ether
Concen: 20.974 ug/l
RT: 3.117 min Scan# 311
Delta R.T. -0.006 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Tgt Ion: 73 Resp: 9718
Ion Ratio Lower Upper
73 100
57 20.2 18.2 27.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#20
Methylene Chloride
Concen: 20.426 ug/l
RT: 2.800 min Scan# 281
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Tgt Ion: 84 Resp: 35302
Ion Ratio Lower Upper
84 100
49 127.8 101.3 151.9
51 39.4 31.0 46.6
86 66.0 51.1 76.7









#21
trans-1,2-Dichloroethene
Concen: 21.224 ug/l
RT: 3.105 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Tgt Ion: 96 Resp: 33790
Ion Ratio Lower Upper
96 100
61 133.3 112.5 168.7
98 62.1 51.9 77.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#22
Diisopropyl ether
Concen: 21.555 ug/l
RT: 3.764 min Scan# 439
Delta R.T. -0.006 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Tgt Ion: 45 Resp: 111712
Ion Ratio Lower Upper
45 100
43 120.9 91.9 137.9
87 25.2 22.0 33.0
59 11.1 9.1 13.7









#23
Vinyl Acetate
Concen: 105.140 ug/l
RT: 3.727 min Scan# 41
Delta R.T. -0.006 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Tgt Ion: 43 Resp: 43331
Ion Ratio Lower Upper
43 100
86 10.2 8.2 12.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#24
1,1-Dichloroethane
Concen: 21.126 ug/l
RT: 3.617 min Scan# 415
Delta R.T. -0.006 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Tgt Ion: 63 Resp: 62164
Ion Ratio Lower Upper
63 100
98 7.1 4.0 11.9
100 4.5 2.2 6.6





#25

2-Butanone

Concen: 99.748 ug/l

RT: 4.556 min Scan# 50

Delta R.T. -0.006 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 43 Resp: 8767

Ion Ratio Lower Upper

43 100





72 24.9 20.2 30.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#26

2,2-Dichloropropane

Concen: 20.032 ug/l

RT: 4.483 min Scan# 557

Delta R.T. -0.006 min

Lab File: VX046719.D




Acq: 17 Jun 2025 13:59


Tgt Ion: 77 Resp: 42992

Ion Ratio Lower Upper

77 100

97 22.9 11.8 35.3





#27

cis-1,2-Dichloroethene

Concen: 21.047 ug/l

RT: 4.495 min Scan# 51

Delta R.T. -0.006 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC020

Tgt Ion:	96	Resp:	39374
Ion	Ratio	Lower	Upper
96	100		
61	144.0	0.0	291.0
98	62.5	0.0	126.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#28

Bromochloromethane

Concen: 18.544 ug/l




RT: 4.904 min Scan# 626


Delta R.T. -0.006 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Tgt Ion:	49	Resp:	24738
Ion	Ratio	Lower	Upper
49	100		
129	1.7	0.0	3.8
130	76.8	61.4	92.0





#29

Tetrahydrofuran

Concen: 99.990 ug/l

RT: 5.007 min Scan# 64

Delta R.T. -0.006 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC020

Tgt Ion:	42	Resp:	5676
Ion	Ratio	Lower	Upper
42	100		
72	42.0	34.9	52.3
71	40.8	32.2	48.2

Manual Integrations

APPROVED

Reviewed By :	Semsettin Yesilyurt	06/18/2025
Supervised By :	Mahesh Dadoda	06/18/2025



#30

Chloroform

Concen: 21.785 ug/l




RT: 5.093 min Scan# 657


Delta R.T. -0.012 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59


Tgt Ion:	83	Resp:	64109
Ion	Ratio	Lower	Upper
83	100		
85	67.0	53.7	80.5





#31
Cyclohexane
Concen: 20.982 ug/l
RT: 5.477 min Scan# 705
Delta R.T. -0.006 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020



Tgt Ion: 56 Resp: 58552
Ion Ratio Lower Upper
56 100
69 28.3 24.8 37.2
84 84.5 68.2 102.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#32
1,1,1-Trichloroethane
Concen: 21.133 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.012 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Tgt Ion: 97 Resp: 54542
Ion Ratio Lower Upper
97 100
99 63.4 51.3 76.9
61 46.4 36.4 54.6





#33

1,2-Dichloroethane-d4

Concen: 16.410 ug/l

RT: 5.964 min Scan# 800

Delta R.T. -0.000 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 65 Resp: 3058

Ion Ratio Lower Upper

65 100





67 52.5 0.0 105.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. -0.012 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59




Tgt Ion:114 Resp: 226567


Ion Ratio Lower Upper

114 100

63 21.2 0.0 44.2

88 15.5 0.0 33.2





#35

Dibromofluoromethane

Concen: 16.114 ug/l

RT: 5.385 min Scan# 705

Delta R.T. -0.012 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC020

Tgt Ion:	113	Resp:	2415
Ion	Ratio	Lower	Upper
113	100		
111	104.6	82.0	123.0
192	19.1	15.3	22.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#36

1,1-Dichloropropene

Concen: 20.344 ug/l




RT: 5.696 min Scan# 756


Delta R.T. -0.006 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Tgt Ion:	75	Resp:	44023
Ion	Ratio	Lower	Upper
75	100		
110	34.9	17.2	51.5
77	30.6	24.2	36.4





#37

Ethyl Acetate

Concen: 19.984 ug/l

RT: 4.715 min Scan# 591

Delta R.T. -0.006 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC020

Tgt Ion:	43	Resp:	39910
Ion	Ratio	Lower	Upper
43	100		
61	11.7	9.7	14.5
70	9.7	8.1	12.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#38

Carbon Tetrachloride

Concen: 20.735 ug/l




RT: 5.684 min Scan# 754


Delta R.T. -0.006 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Tgt Ion:	117	Resp:	48694
Ion	Ratio	Lower	Upper
117	100		
119	90.9	75.2	112.8
121	31.8	24.1	36.1





#39

Methylcyclohexane

Concen: 20.428 ug/l

RT: 7.379 min Scan# 1032

Delta R.T. -0.006 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC020

Tgt Ion:	83	Resp:	58170
Ion	Ratio	Lower	Upper
83	100		
55	76.1	63.0	94.4
98	50.7	38.8	58.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#40

Benzene

Concen: 20.904 ug/l




RT: 6.038 min Scan# 812


Delta R.T. -0.006 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59


Tgt Ion:	78	Resp:	133872
Ion	Ratio	Lower	Upper
78	100		
77	23.3	18.8	28.2





#41
Methacrylonitrile
Concen: 20.632 ug/l
RT: 4.922 min Scan# 61
Delta R.T. -0.006 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020



Tgt Ion: 41 Resp: 21970
Ion Ratio Lower Upper
41 100
39 54.8 45.9 68.9
67 73.2 61.2 91.8
52 30.8 25.5 38.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#42
1,2-Dichloroethane
Concen: 21.076 ug/l
RT: 6.086 min Scan# 820
Delta R.T. -0.006 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Tgt Ion: 62 Resp: 47430
Ion Ratio Lower Upper
62 100
98 8.7 0.0 17.8





#43

Isopropyl Acetate

Concen: 19.975 ug/l

RT: 6.342 min Scan# 80

Delta R.T. -0.000 min

Lab File: VX046719.D


Acq: 17 Jun 2025 13:59

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020






Tgt Ion:	43	Resp:	6243
Ion	Ratio	Lower	Upper
43	100		
61	19.2	15.3	22.9
87	12.5	10.0	15.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#44

Trichloroethene


Concen: 20.788 ug/l

RT: 7.129 min Scan# 991



Delta R.T. -0.000 min


Lab File: VX046719.D

Acq: 17 Jun 2025 13:59



Tgt Ion:	130	Resp:	33927
Ion	Ratio	Lower	Upper
130	100		
95	101.1	0.0	202.6





#45

1,2-Dichloropropane

Concen: 21.226 ug/l

RT: 7.428 min Scan# 1040

Delta R.T. -0.006 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 63 Resp: 3368

Ion Ratio Lower Upper

63 100





65 31.6 22.9 34.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#46

Dibromomethane

Concen: 20.663 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. -0.006 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59




Tgt Ion: 93 Resp: 23569


Ion Ratio Lower Upper

93 100

95 84.4 66.2 99.4

174 97.1 77.0 115.6





#47

Bromodichloromethane

Concen: 21.227 ug/l

RT: 7.818 min Scan# 1104

Delta R.T. -0.006 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC020

Tgt Ion:	83	Resp:	4941
Ion	Ratio	Lower	Upper
83	100		
85	65.7	51.4	77.2
127	8.5	6.9	10.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#48

Methyl methacrylate

Concen: 20.322 ug/l




RT: 7.690 min Scan# 1083


Delta R.T. -0.006 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Tgt Ion:	41	Resp:	31856
Ion	Ratio	Lower	Upper
41	100		
69	82.0	63.4	95.2
39	57.2	45.4	68.0





#49

1,4-Dioxane

Concen: 382.035 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. 0.000 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020





Tgt Ion: 88 Resp: 8340

Ion	Ratio	Lower	Upper
88	100		
43	37.1	31.2	46.8
58	75.7	54.7	82.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#50

Toluene-d8

Concen: 16.376 ug/l

RT: 8.647 min Scan# 1240




Delta R.T. -0.000 min


Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Tgt Ion: 98 Resp: 88197

Ion	Ratio	Lower	Upper
98	100		
100	67.7	53.0	79.4





#51

4-Methyl-2-Pentanone

Concen: 100.586 ug/l

RT: 8.568 min Scan# 1227

Delta R.T. -0.006 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 43 Resp: 187509

Ion Ratio Lower Upper

43 100





58 37.8 30.8 46.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#52

Toluene

Concen: 21.166 ug/l

RT: 8.714 min Scan# 1251

Delta R.T. -0.006 min

Lab File: VX046719.D




Acq: 17 Jun 2025 13:59


Tgt Ion: 92 Resp: 84030

Ion Ratio Lower Upper

92 100

91 169.0 134.8 202.2





#53

t-1,3-Dichloropropene

Concen: 20.304 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020

Tgt Ion: 75 Resp: 43852

Ion Ratio Lower Upper

75 100





77 31.4 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#54

cis-1,3-Dichloropropene

Concen: 20.526 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59




Tgt Ion: 75 Resp: 50326


Ion Ratio Lower Upper

75 100

77 30.9 25.3 37.9

39 47.9 39.2 58.8





#55

1,1,2-Trichloroethane

Concen: 20.862 ug/l

RT: 9.147 min Scan# 1323

Delta R.T. -0.006 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 97 Resp: 3086

Ion Ratio Lower Upper

97 100

83 91.5 67.2 100.8

85 53.2 44.1 66.1





99 68.3 50.3 75.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#56

Ethyl methacrylate

Concen: 20.468 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. -0.000 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59




Tgt Ion: 69 Resp: 44702


Ion Ratio Lower Upper

69 100

41 70.7 55.4 83.2

39 37.3 29.7 44.5





#57

1,3-Dichloropropane

Concen: 20.480 ug/l

RT: 9.305 min Scan# 1173

Delta R.T. -0.000 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020

Tgt Ion: 76 Resp: 5218

Ion Ratio Lower Upper

76 100





78 32.6 25.7 38.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#58

2-Chloroethyl Vinyl ether

Concen: 93.878 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. -0.006 min

Lab File: VX046719.D




Acq: 17 Jun 2025 13:59


Tgt Ion: 63 Resp: 106505

Ion Ratio Lower Upper

63 100

106 27.9 22.6 34.0





#59

2-Hexanone

Concen: 100.265 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. -0.000 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 43 Resp: 129218

Ion Ratio Lower Upper

43 100





58 52.8 26.8 80.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#60

Dibromochloromethane

Concen: 20.872 ug/l

RT: 9.519 min Scan# 1383

Delta R.T. -0.000 min

Lab File: VX046719.D




Acq: 17 Jun 2025 13:59


Tgt Ion:129 Resp: 36392

Ion Ratio Lower Upper

129 100

127 78.0 38.4 115.2





#61

1,2-Dibromoethane

Concen: 20.936 ug/l

RT: 9.604 min Scan# 11

Delta R.T. -0.006 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion:107 Resp: 3150

Ion Ratio Lower Upper

107 100





109 95.0 75.8 113.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#62

4-Bromofluorobenzene

Concen: 15.973 ug/l

RT: 11.079 min Scan# 1639

Delta R.T. -0.000 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59




Tgt Ion: 95 Resp: 32067


Ion Ratio Lower Upper

95 100

174 78.1 0.0 150.4

176 75.2 0.0 145.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. -0.006 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020

Tgt Ion:117 Resp: 19794
Ion Ratio Lower Upper
117 100
82 58.9 44.6 66.8
119 33.0 25.8 38.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#64
Tetrachloroethene
Concen: 20.408 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. -0.006 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Tgt Ion:164 Resp: 27988
Ion Ratio Lower Upper
164 100
166 133.4 104.6 157.0
129 102.4 76.3 114.5
131 93.3 74.1 111.1









#65
Chlorobenzene
Concen: 20.796 ug/l
RT: 10.073 min Scan# 1475
Delta R.T. -0.006 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020

Tgt Ion:112 Resp: 9086
Ion Ratio Lower Upper
112 100
114 31.9 25.5 38.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 21.455 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Tgt Ion:131 Resp: 31508
Ion Ratio Lower Upper
131 100
133 94.3 47.3 141.8
119 63.4 32.1 96.3





#67

Ethyl Benzene

Concen: 21.056 ug/l

RT: 10.189 min Scan# 1493

Delta R.T. -0.000 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 91 Resp: 16076

Ion Ratio Lower Upper

91 100





106 30.5 24.2 36.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#68

m/p-Xylenes

Concen: 42.539 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. -0.000 min

Lab File: VX046719.D




Acq: 17 Jun 2025 13:59


Tgt Ion:106 Resp: 121014

Ion Ratio Lower Upper

106 100

91 207.7 164.8 247.2





#69

o-Xylene

Concen: 21.638 ug/l

RT: 10.640 min Scan# 1567

Delta R.T. -0.000 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion:106 Resp: 57689

Ion Ratio Lower Upper

106 100





91 222.2 111.3 333.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#70

Styrene

Concen: 21.318 ug/l

RT: 10.653 min Scan# 1569

Delta R.T. -0.000 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59




Tgt Ion:104 Resp: 99462


Ion Ratio Lower Upper

104 100

78 52.6 41.4 62.2


103 54.4 43.1 64.7





#71
Bromoform
Concen: 20.898 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020



Tgt Ion:173 Resp: 23372
Ion Ratio Lower Upper
173 100
175 48.0 24.1 72.4
254 0.0 0.1 0.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Tgt Ion:152 Resp: 98285
Ion Ratio Lower Upper
152 100
115 71.7 43.2 129.6
150 166.5 0.0 346.8









#73
Isopropylbenzene
Concen: 21.255 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020

Tgt Ion:105 Resp: 15386
Ion Ratio Lower Upper
105 100
120 25.6 13.0 39.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#74
N-ethyl acetate
Concen: 20.851 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Tgt Ion: 43 Resp: 57185
Ion Ratio Lower Upper
43 100
70 40.3 33.4 50.0
55 23.9 19.1 28.7
61 23.3 19.0 28.4





#75

1,1,2,2-Tetrachloroethane

Concen: 20.914 ug/l

RT: 11.207 min Scan# 1660

Delta R.T. -0.000 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC020

Tgt Ion:	83	Resp:	4227
Ion	Ratio	Lower	Upper
83	100		
131	10.0	4.9	14.6
85	64.0	32.0	96.0

Manual Integrations

APPROVED

Reviewed By : Semsettin Yesilyurt	06/18/2025
Supervised By : Mahesh Dadoda	06/18/2025



#76

1,2,3-Trichloropropane

Concen: 20.989 ug/l m




RT: 11.238 min Scan# 1665


Delta R.T. -0.000 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59


Tgt Ion:	75	Resp:	41901
Ion	Ratio	Lower	Upper
75	100		
77	34.6	18.5	55.5





#77
Bromobenzene
Concen: 21.507 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020




Tgt Ion:156 Resp: 37464
Ion Ratio Lower Upper
156 100
77 163.6 82.8 248.6
158 96.6 47.6 142.9

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025




#78
n-propylbenzene
Concen: 21.410 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59



Tgt Ion: 91 Resp: 184078
Ion Ratio Lower Upper
91 100
120 22.5 11.4 34.1





#79

2-Chlorotoluene

Concen: 20.952 ug/l

RT: 11.360 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 91 Resp: 107552

Ion Ratio Lower Upper

91 100





126 33.4 16.6 49.7

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#80

1,3,5-Trimethylbenzene

Concen: 21.793 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. -0.000 min

Lab File: VX046719.D




Acq: 17 Jun 2025 13:59


Tgt Ion:105 Resp: 129635

Ion Ratio Lower Upper

105 100

120 47.6 24.1 72.2





#81

trans-1,4-Dichloro-2-butene

Concen: 18.362 ug/l

RT: 11.018 min Scan# 1629

Delta R.T. -0.000 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC020

Tgt Ion:	75	Resp:	11694
Ion Ratio	Lower	Upper	
75	100		
53	111.5	86.8	130.2
89	46.7	36.9	55.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#82

4-Chlorotoluene

Concen: 21.358 ug/l




RT: 11.451 min Scan# 1700


Delta R.T. -0.000 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59


Tgt Ion:	91	Resp:	127954
Ion Ratio	Lower	Upper	
91	100		
126	28.6	14.4	43.2





#83
tert-Butylbenzene
Concen: 21.201 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020



Tgt Ion:119 Resp: 130359
Ion Ratio Lower Upper
119 100
91 58.4 29.5 88.6
134 23.0 11.7 35.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#84
1,2,4-Trimethylbenzene
Concen: 21.512 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59


Tgt Ion:105 Resp: 128566
Ion Ratio Lower Upper
105 100
120 43.5 22.1 66.3





#85
sec-Butylbenzene
Concen: 21.239 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020




Tgt Ion:105 Resp: 16504
Ion Ratio Lower Upper
105 100
134 20.0 9.8 29.4

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025




#86
p-Isopropyltoluene
Concen: 21.142 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59




Tgt Ion:119 Resp: 138623
Ion Ratio Lower Upper
119 100
134 25.9 12.9 38.7
91 24.8 12.3 36.9





#87
1,3-Dichlorobenzene
Concen: 20.681 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020




Tgt Ion:146 Resp: 7024
Ion Ratio Lower Upper
146 100
111 40.5 20.6 61.8
148 62.4 31.6 94.8

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025




#88
1,4-Dichlorobenzene
Concen: 20.625 ug/l
RT: 12.036 min Scan# 1796
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59



Tgt Ion:146 Resp: 71953
Ion Ratio Lower Upper
146 100
111 40.7 20.3 61.0
148 62.6 32.1 96.5





#89

n-Butylbenzene

Concen: 20.738 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. -0.000 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 91 Resp: 12910

Ion Ratio Lower Upper

91 100

92 54.2 27.1 81.3





134 25.8 12.9 38.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#90

Hexachloroethane

Concen: 20.594 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX046719.D




Acq: 17 Jun 2025 13:59


Tgt Ion:117 Resp: 23645

Ion Ratio Lower Upper

117 100


201 71.6 34.4 103.3





#91
1,2-Dichlorobenzene
Concen: 21.073 ug/l
RT: 12.329 min Scan# 1844
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020



Tgt Ion:146 Resp: 6688
Ion Ratio Lower Upper
146 100
111 43.0 21.7 65.1
148 63.5 32.0 96.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 20.226 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59


Tgt Ion: 75 Resp: 8071
Ion Ratio Lower Upper
75 100
155 79.6 41.9 125.7
157 102.7 51.4 154.3





#93
1,2,4-Trichlorobenzene
Concen: 19.821 ug/l
RT: 13.585 min Scan# 2050
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020



Tgt Ion:180 Resp: 44720
Ion Ratio Lower Upper
180 100
182 96.1 46.7 140.1
145 32.8 15.8 47.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#94
Hexachlorobutadiene
Concen: 20.752 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.006 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59


Tgt Ion:225 Resp: 19698
Ion Ratio Lower Upper
225 100
223 66.0 32.5 97.5
227 65.3 32.2 96.6





#95
Naphthalene
Concen: 20.186 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020



Tgt Ion:128 Resp: 129768
Ion Ratio Lower Upper
128 100
127 12.7 10.1 15.1
129 10.8 8.7 13.1




Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#96
1,2,3-Trichlorobenzene
Concen: 20.558 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Tgt Ion:180 Resp: 44380
Ion Ratio Lower Upper
180 100
182 95.4 47.8 143.4
145 34.4 17.3 51.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
 Data File : VX046720.D
 Acq On : 17 Jun 2025 14:20
 Operator : JC/MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
 Supervised By :Mahesh Dadoda 06/18/2025

Quant Time: Jun 18 02:39:23 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 02:34:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	123412	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	204400	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	179407	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	89016	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	80133	46.944	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	93.880%		
35) Dibromofluoromethane	5.397	113	65447	48.394	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.780%		
50) Toluene-d8	8.647	98	233825	48.124	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	96.240%		
62) 4-Bromofluorobenzene	11.079	95	87131	48.107	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	96.220%		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.185	85	71910	54.577	ug/l	100
3) Chloromethane	1.307	50	73175	51.454	ug/l	100
4) Vinyl Chloride	1.386	62	78242	51.592	ug/l	100
5) Bromomethane	1.624	94	45535	53.367	ug/l	100
6) Chloroethane	1.703	64	46570	50.665	ug/l	100
7) Trichlorofluoromethane	1.904	101	119335	51.823	ug/l	100
8) Diethyl Ether	2.148	74	39712	50.251	ug/l	100
9) 1,1,2-Trichlorotrifluo...	2.343	101	71037	50.285	ug/l	100
10) Methyl Iodide	2.471	142	80710	46.019	ug/l	100
11) Tert butyl alcohol	2.959	59	38256	249.692	ug/l	100
12) 1,1-Dichloroethene	2.337	96	70202	51.500	ug/l	100
13) Acrolein	2.246	56	53791	241.036	ug/l	100
14) Allyl chloride	2.678	41	123265	49.909	ug/l	100
15) Acrylonitrile	3.075	53	175846	255.243	ug/l	100
16) Acetone	2.386	43	135683	238.356	ug/l	100
17) Carbon Disulfide	2.526	76	197344	47.941	ug/l	100
18) Methyl Acetate	2.715	43	71685	49.582	ug/l	100
19) Methyl tert-butyl Ether	3.123	73	216266	50.953	ug/l	100
20) Methylene Chloride	2.800	84	75315	47.576	ug/l	100
21) trans-1,2-Dichloroethene	3.105	96	72648	49.807	ug/l	100
22) Diisopropyl ether	3.770	45	242789	51.143	ug/l	100
23) Vinyl Acetate	3.733	43	985040	260.935	ug/l	100
24) 1,1-Dichloroethane	3.623	63	134324	49.835	ug/l	100
25) 2-Butanone	4.562	43	208346	258.788	ug/l	100
26) 2,2-Dichloropropane	4.489	77	93956	47.794	ug/l	100
27) cis-1,2-Dichloroethene	4.501	96	84638	49.392	ug/l	100
28) Bromochloromethane	4.910	49	59857	48.984	ug/l	100
29) Tetrahydrofuran	5.013	42	133555	256.831	ug/l	100
30) Chloroform	5.105	83	137459	50.996	ug/l	100
31) Cyclohexane	5.483	56	125014	48.909	ug/l	100
32) 1,1,1-Trichloroethane	5.397	97	117954	49.895	ug/l	100
36) 1,1-Dichloropropene	5.702	75	97413	49.899	ug/l	100
37) Ethyl Acetate	4.721	43	88076	48.877	ug/l	100
38) Carbon Tetrachloride	5.690	117	105202	49.656	ug/l	100
39) Methylcyclohexane	7.385	83	128553	50.041	ug/l	100
40) Benzene	6.044	78	289724	50.146	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
 Data File : VX046720.D
 Acq On : 17 Jun 2025 14:20
 Operator : JC/MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 06/18/2025
 Supervised By : Mahesh Dadoda 06/18/2025

Quant Time: Jun 18 02:39:23 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 02:34:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.928	41	48765	50.762	ug/l	100
42) 1,2-Dichloroethane	6.092	62	101218	49.855	ug/l	100
43) Isopropyl Acetate	6.342	43	146377	51.911	ug/l	100
44) Trichloroethene	7.129	130	72796	49.441	ug/l	100
45) 1,2-Dichloropropane	7.434	63	70702	49.378	ug/l	100
46) Dibromomethane	7.586	93	51288	49.840	ug/l	100
47) Bromodichloromethane	7.824	83	106651	50.787	ug/l	100
48) Methyl methacrylate	7.696	41	75708	53.535	ug/l	100
49) 1,4-Dioxane	7.659	88	19802	962.534	ug/l	100
51) 4-Methyl-2-Pentanone	8.574	43	435000	258.656	ug/l	100
52) Toluene	8.720	92	181414	50.650	ug/l	100
53) t-1,3-Dichloropropene	8.976	75	99727	51.183	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	111917	50.598	ug/l	100
55) 1,1,2-Trichloroethane	9.153	97	66688	49.963	ug/l	100
56) Ethyl methacrylate	9.116	69	103697	52.629	ug/l	100
57) 1,3-Dichloropropane	9.305	76	114824	49.955	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.245	63	287531	280.927	ug/l	100
59) 2-Hexanone	9.427	43	303322	260.883	ug/l	100
60) Dibromochloromethane	9.519	129	79227	50.368	ug/l	100
61) 1,2-Dibromoethane	9.610	107	68515	50.464	ug/l	100
64) Tetrachloroethene	9.275	164	61048	49.115	ug/l	100
65) Chlorobenzene	10.080	112	195752	49.429	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.159	131	67004	50.341	ug/l	100
67) Ethyl Benzene	10.189	91	346812	50.119	ug/l	100
68) m/p-Xylenes	10.299	106	259899	100.800	ug/l	100
69) o-Xylene	10.640	106	124120	51.366	ug/l	100
70) Styrene	10.653	104	216738	51.254	ug/l	100
71) Bromoform	10.799	173	51485	50.792	ug/l #	100
73) Isopropylbenzene	10.957	105	331391	50.548	ug/l	100
74) N-amyl acetate	10.842	43	130606	52.581	ug/l	100
75) 1,1,2,2-Tetrachloroethane	11.207	83	92893	50.745	ug/l	100
76) 1,2,3-Trichloropropane	11.238	75	88091m	48.721	ug/l	
77) Bromobenzene	11.195	156	79404	50.331	ug/l	100
78) n-propylbenzene	11.299	91	394375	50.646	ug/l	100
79) 2-Chlorotoluene	11.360	91	230178	49.509	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	273477	50.761	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	27677	47.984	ug/l	100
82) 4-Chlorotoluene	11.451	91	269198	49.613	ug/l	100
83) tert-Butylbenzene	11.713	119	277915	49.905	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	272510	50.345	ug/l	100
85) sec-Butylbenzene	11.890	105	355947	50.574	ug/l	100
86) p-Isopropyltoluene	12.006	119	297314	50.067	ug/l	100
87) 1,3-Dichlorobenzene	11.969	146	149402	48.568	ug/l	100
88) 1,4-Dichlorobenzene	12.037	146	147146	46.571	ug/l	100
89) n-Butylbenzene	12.329	91	280794	49.801	ug/l	100
90) Hexachloroethane	12.536	117	51961	49.968	ug/l	100
91) 1,2-Dichlorobenzene	12.329	146	141743	49.309	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	12.939	75	18747	51.873	ug/l	100
93) 1,2,4-Trichlorobenzene	13.585	180	100346	49.101	ug/l	100
94) Hexachlorobutadiene	13.719	225	42659	49.621	ug/l	100
95) Naphthalene	13.774	128	296912	50.994	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	96318	49.263	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
Data File : VX046720.D
Acq On : 17 Jun 2025 14:20
Operator : JC/MD
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025

Quant Time: Jun 18 02:39:23 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 02:34:13 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


```
Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\  
Data File : VX046720.D  
Acq On    : 17 Jun 2025  14:20  
Operator  : JC/MD  
Sample    : VSTDICCC050  
Misc      : 5.0mL/MSVOA_X/WATER  
ALS Vial  : 6      Sample Multiplier: 1
```


Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025

Quant Time: Jun 18 02:39:23 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 02:34:13 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.562 min Scan# 734
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion:168 Resp: 123412
Ion Ratio Lower Upper
168 100
99 60.6 48.5 72.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#2
Dichlorodifluoromethane
Concen: 54.577 ug/l
RT: 1.185 min Scan# 16
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Tgt Ion: 85 Resp: 71910
Ion Ratio Lower Upper
85 100
87 33.2 16.6 49.8









#3
Chloromethane
Concen: 51.454 ug/l
RT: 1.307 min Scan# 30
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 50 Resp: 73179
Ion Ratio Lower Upper
50 100
52 31.1 24.9 37.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#4
Vinyl Chloride
Concen: 51.592 ug/l
RT: 1.386 min Scan# 49
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Tgt Ion: 62 Resp: 78242
Ion Ratio Lower Upper
62 100
64 29.9 23.9 35.9









#5
Bromomethane
Concen: 53.367 ug/l
RT: 1.624 min Scan# 88
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 94 Resp: 4553
Ion Ratio Lower Upper
94 100
96 90.4 72.3 108.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#6
Chloroethane
Concen: 50.665 ug/l
RT: 1.703 min Scan# 101
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20


Tgt Ion: 64 Resp: 46570
Ion Ratio Lower Upper
64 100
66 33.9 27.1 40.7





#7
Trichlorofluoromethane
Concen: 51.823 ug/l
RT: 1.904 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion:101 Resp: 11933
Ion Ratio Lower Upper
101 100
103 64.8 51.8 77.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#8
Diethyl Ether
Concen: 50.251 ug/l
RT: 2.148 min Scan# 174
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Tgt Ion: 74 Resp: 39712
Ion Ratio Lower Upper
74 100
45 104.7 52.3 157.1





#9

1,1,2-Trichlorotrifluoroethane

Concen: 50.285 ug/l

RT: 2.343 min Scan# 206

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion:	101	Resp:	7103
Ion	Ratio	Lower	Upper
101	100		
85	44.2	35.4	53.0
151	76.5	61.2	91.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#10

Methyl Iodide

Concen: 46.019 ug/l




RT: 2.471 min Scan# 227


Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20

Tgt Ion:	142	Resp:	80710
Ion	Ratio	Lower	Upper
142	100		
127	46.3	37.0	55.6
141	14.4	11.5	17.3





#11

Tert butyl alcohol

Concen: 249.692 ug/l

RT: 2.959 min Scan# 307

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 59 Resp: 38250

Ion Ratio Lower Upper

59 100





57 10.4 8.3 12.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#12

1,1-Dichloroethene

Concen: 51.500 ug/l

RT: 2.337 min Scan# 205

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20




Tgt Ion: 96 Resp: 70202


Ion Ratio Lower Upper

96 100

61 158.2 126.6 189.8

98 63.4 50.7 76.1





#13

Acrolein

Concen: 241.036 ug/l

RT: 2.246 min Scan# 190

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 56 Resp: 5379

Ion Ratio Lower Upper

56 100





55 69.0 55.2 82.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#14

Allyl chloride

Concen: 49.909 ug/l

RT: 2.678 min Scan# 261

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20




Tgt Ion: 41 Resp: 123265


Ion Ratio Lower Upper

41 100

39 67.0 53.6 80.4

76 33.5 26.8 40.2





#15

Acrylonitrile

Concen: 255.243 ug/l

RT: 3.075 min Scan# 31

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 53 Resp: 175840

Ion Ratio Lower Upper

53 100

52 81.4 65.1 97.7





51 36.1 28.9 43.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#16

Acetone

Concen: 238.356 ug/l

RT: 2.386 min Scan# 213

Delta R.T. 0.000 min

Lab File: VX046720.D




Acq: 17 Jun 2025 14:20


Tgt Ion: 43 Resp: 135683

Ion Ratio Lower Upper

43 100


58 30.6 24.5 36.7





#17
Carbon Disulfide
Concen: 47.941 ug/l
RT: 2.526 min Scan# 21
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion: 76 Resp: 197344
Ion Ratio Lower Upper
76 100
78 9.4 7.5 11.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#18
Methyl Acetate
Concen: 49.582 ug/l
RT: 2.715 min Scan# 267
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Tgt Ion: 43 Resp: 71685
Ion Ratio Lower Upper
43 100
74 23.0 18.4 27.6









#19
Methyl tert-butyl Ether
Concen: 50.953 ug/l
RT: 3.123 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 73 Resp: 21626
Ion Ratio Lower Upper
73 100
57 22.7 18.2 27.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#20
Methylene Chloride
Concen: 47.576 ug/l
RT: 2.800 min Scan# 281
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Tgt Ion: 84 Resp: 75315
Ion Ratio Lower Upper
84 100
49 126.6 101.3 151.9
51 38.8 31.0 46.6
86 63.9 51.1 76.7









#21
trans-1,2-Dichloroethene
Concen: 49.807 ug/l
RT: 3.105 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 96 Resp: 72648
Ion Ratio Lower Upper
96 100
61 140.6 112.5 168.7
98 64.9 51.9 77.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#22
Diisopropyl ether
Concen: 51.143 ug/l
RT: 3.770 min Scan# 440
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Tgt Ion: 45 Resp: 242789
Ion Ratio Lower Upper
45 100
43 114.9 91.9 137.9
87 27.5 22.0 33.0
59 11.4 9.1 13.7









#23
Vinyl Acetate
Concen: 260.935 ug/l
RT: 3.733 min Scan# 41
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 43 Resp: 985040
Ion Ratio Lower Upper
43 100
86 10.2 8.2 12.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#24
1,1-Dichloroethane
Concen: 49.835 ug/l
RT: 3.623 min Scan# 416
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Tgt Ion: 63 Resp: 134324
Ion Ratio Lower Upper
63 100
98 7.9 4.0 11.9
100 4.4 2.2 6.6









#25
2-Butanone
Concen: 258.788 ug/l
RT: 4.562 min Scan# 570
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 43 Resp: 208340
Ion Ratio Lower Upper
43 100
72 25.3 20.2 30.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#26
2,2-Dichloropropane
Concen: 47.794 ug/l
RT: 4.489 min Scan# 558
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Tgt Ion: 77 Resp: 93956
Ion Ratio Lower Upper
77 100
97 23.5 11.8 35.3





#27

cis-1,2-Dichloroethene

Concen: 49.392 ug/l

RT: 4.501 min Scan# 560

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 96 Resp: 84638

Ion Ratio Lower Upper

96 100

61 145.5 0.0 291.0





98 63.4 0.0 126.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#28

Bromochloromethane

Concen: 48.984 ug/l

RT: 4.910 min Scan# 627

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20




Tgt Ion: 49 Resp: 59857


Ion Ratio Lower Upper

49 100

129 1.9 0.0 3.8

130 76.7 61.4 92.0





#29

Tetrahydrofuran

Concen: 256.831 ug/l

RT: 5.013 min Scan# 644

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 42 Resp: 13355

Ion Ratio Lower Upper

42 100

72 43.6 34.9 52.3





71 40.2 32.2 48.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#30

Chloroform

Concen: 50.996 ug/l

RT: 5.105 min Scan# 659

Delta R.T. 0.000 min

Lab File: VX046720.D




Acq: 17 Jun 2025 14:20


Tgt Ion: 83 Resp: 137459

Ion Ratio Lower Upper

83 100


85 67.1 53.7 80.5





#31
Cyclohexane
Concen: 48.909 ug/l
RT: 5.483 min Scan# 721
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion: 56 Resp: 125014
Ion Ratio Lower Upper
56 100
69 31.0 24.8 37.2
84 85.2 68.2 102.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#32
1,1,1-Trichloroethane
Concen: 49.895 ug/l
RT: 5.397 min Scan# 707
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Tgt Ion: 97 Resp: 117954
Ion Ratio Lower Upper
97 100
99 64.1 51.3 76.9
61 45.5 36.4 54.6





#33

1,2-Dichloroethane-d4

Concen: 46.944 ug/l

RT: 5.964 min Scan# 800

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 65 Resp: 8013

Ion Ratio Lower Upper

65 100





67 52.7 0.0 105.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.769 min Scan# 932

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20




Tgt Ion:114 Resp: 204400


Ion Ratio Lower Upper

114 100

63 22.1 0.0 44.2

88 16.6 0.0 33.2





#35

Dibromofluoromethane

Concen: 48.394 ug/l

RT: 5.397 min Scan# 707

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion:	113	Resp:	6544
Ion	Ratio	Lower	Upper
113	100		
111	102.5	82.0	123.0
192	19.1	15.3	22.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#36

1,1-Dichloropropene

Concen: 49.899 ug/l




RT: 5.702 min Scan# 757


Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20

Tgt Ion:	75	Resp:	97413
Ion	Ratio	Lower	Upper
75	100		
110	34.3	17.2	51.5
77	30.3	24.2	36.4





#37

Ethyl Acetate

Concen: 48.877 ug/l

RT: 4.721 min Scan# 596

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion:	43	Resp:	88070
Ion	Ratio	Lower	Upper
43	100		
61	12.1	9.7	14.5
70	10.1	8.1	12.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#38

Carbon Tetrachloride

Concen: 49.656 ug/l




RT: 5.690 min Scan# 755


Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20

Tgt Ion:	117	Resp:	105202
Ion	Ratio	Lower	Upper
117	100		
119	94.0	75.2	112.8
121	30.1	24.1	36.1





#39

Methylcyclohexane

Concen: 50.041 ug/l

RT: 7.385 min Scan# 1033

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 83 Resp: 12855

Ion Ratio Lower Upper

83 100

55 78.7 63.0 94.4





98 48.5 38.8 58.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#40

Benzene

Concen: 50.146 ug/l

RT: 6.044 min Scan# 813

Delta R.T. 0.000 min

Lab File: VX046720.D




Acq: 17 Jun 2025 14:20


Tgt Ion: 78 Resp: 289724

Ion Ratio Lower Upper

78 100


77 23.5 18.8 28.2





#41
Methacrylonitrile
Concen: 50.762 ug/l
RT: 4.928 min Scan# 61
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion: 41 Resp: 4876
Ion Ratio Lower Upper
41 100
39 57.4 45.9 68.9
67 76.5 61.2 91.8
52 31.9 25.5 38.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#42
1,2-Dichloroethane
Concen: 49.855 ug/l
RT: 6.092 min Scan# 821
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Tgt Ion: 62 Resp: 101218
Ion Ratio Lower Upper
62 100
98 8.9 0.0 17.8





#43

Isopropyl Acetate

Concen: 51.911 ug/l

RT: 6.342 min Scan# 862

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20

Instrument :

MSVOA_X

Client Sample Id :

VSTDICCC050

Tgt Ion: 43 Resp: 14637

Ion Ratio Lower Upper

43 100

61 19.1 15.3 22.9





87 12.5 10.0 15.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#44

Trichloroethene

Concen: 49.441 ug/l

RT: 7.129 min Scan# 991

Delta R.T. 0.000 min

Lab File: VX046720.D




Acq: 17 Jun 2025 14:20


Tgt Ion:130 Resp: 72796

Ion Ratio Lower Upper

130 100

95 101.3 0.0 202.6





#45

1,2-Dichloropropane

Concen: 49.378 ug/l

RT: 7.434 min Scan# 1041

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 63 Resp: 7070

Ion Ratio Lower Upper

63 100





65 28.6 22.9 34.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#46

Dibromomethane

Concen: 49.840 ug/l

RT: 7.586 min Scan# 1066

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20




Tgt Ion: 93 Resp: 51288


Ion Ratio Lower Upper

93 100

95 82.8 66.2 99.4


174 96.3 77.0 115.6





#47
Bromodichloromethane
Concen: 50.787 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion: 83 Resp: 10665
Ion Ratio Lower Upper
83 100
85 64.3 51.4 77.2
127 8.6 6.9 10.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#48
Methyl methacrylate
Concen: 53.535 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Tgt Ion: 41 Resp: 75708
Ion Ratio Lower Upper
41 100
69 79.3 63.4 95.2
39 56.7 45.4 68.0





#49

1,4-Dioxane

Concen: 962.534 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 88 Resp: 1980

Ion Ratio Lower Upper

88 100

43 39.0 31.2 46.8





58 68.4 54.7 82.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#50

Toluene-d8

Concen: 48.124 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. 0.000 min

Lab File: VX046720.D




Acq: 17 Jun 2025 14:20


Tgt Ion: 98 Resp: 233825

Ion Ratio Lower Upper

98 100

100 66.2 53.0 79.4









#51
4-Methyl-2-Pentanone
Concen: 258.656 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 43 Resp: 435000
Ion Ratio Lower Upper
43 100
58 38.5 30.8 46.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#52
Toluene
Concen: 50.650 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Tgt Ion: 92 Resp: 181414
Ion Ratio Lower Upper
92 100
91 168.5 134.8 202.2





#53

t-1,3-Dichloropropene

Concen: 51.183 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 75 Resp: 9972

Ion Ratio Lower Upper

75 100





77 32.0 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#54

cis-1,3-Dichloropropene

Concen: 50.598 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20




Tgt Ion: 75 Resp: 111917


Ion Ratio Lower Upper

75 100

77 31.6 25.3 37.9

39 49.0 39.2 58.8





#55

1,1,2-Trichloroethane

Concen: 49.963 ug/l

RT: 9.153 min Scan# 1323

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 97 Resp: 6668

Ion Ratio Lower Upper

97 100

83 84.0 67.2 100.8

85 55.1 44.1 66.1





99 62.9 50.3 75.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#56

Ethyl methacrylate

Concen: 52.629 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20




Tgt Ion: 69 Resp: 103697


Ion Ratio Lower Upper

69 100

41 69.3 55.4 83.2

39 37.1 29.7 44.5





#57

1,3-Dichloropropane

Concen: 49.955 ug/l

RT: 9.305 min Scan# 114824

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 76 Resp: 114824





Ion Ratio Lower Upper

76 100

78 32.1 25.7 38.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025

#58

2-Chloroethyl Vinyl ether

Concen: 280.927 ug/l

RT: 8.245 min Scan# 1174

Delta R.T. 0.000 min

Lab File: VX046720.D




Acq: 17 Jun 2025 14:20


Tgt Ion: 63 Resp: 287531

Ion Ratio Lower Upper

63 100

106 28.3 22.6 34.0





#59

2-Hexanone

Concen: 260.883 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 43 Resp: 30332

Ion Ratio Lower Upper

43 100





58 53.6 26.8 80.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#60

Dibromochloromethane

Concen: 50.368 ug/l

RT: 9.519 min Scan# 1383

Delta R.T. 0.000 min

Lab File: VX046720.D




Acq: 17 Jun 2025 14:20


Tgt Ion:129 Resp: 79227

Ion Ratio Lower Upper

129 100


127 76.8 38.4 115.2





#61
1,2-Dibromoethane
Concen: 50.464 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion:107 Resp: 6851
Ion Ratio Lower Upper
107 100
109 94.7 75.8 113.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#62
4-Bromofluorobenzene
Concen: 48.107 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Tgt Ion: 95 Resp: 87131
Ion Ratio Lower Upper
95 100
174 75.2 0.0 150.4
176 72.5 0.0 145.0





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.055 min Scan# 1471

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion:117	Resp: 17940
Ion Ratio	Lower Upper
117	100
82	55.7 44.6 66.8
119	32.3 25.8 38.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#64

Tetrachloroethene

Concen: 49.115 ug/l




RT: 9.275 min Scan# 1343


Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20

Tgt Ion:164	Resp: 61048
Ion Ratio	Lower Upper
164	100
166	130.8 104.6 157.0
129	95.4 76.3 114.5
131	92.6 74.1 111.1









#65
Chlorobenzene
Concen: 49.429 ug/l
RT: 10.080 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion:112 Resp: 195752
Ion Ratio Lower Upper
112 100
114 31.9 25.5 38.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 50.341 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20


Tgt Ion:131 Resp: 67004
Ion Ratio Lower Upper
131 100
133 94.5 47.3 141.8
119 64.2 32.1 96.3





#67
Ethyl Benzene
Concen: 50.119 ug/l
RT: 10.189 min Scan# 1493
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion: 91 Resp: 346812
Ion Ratio Lower Upper
91 100
106 30.2 24.2 36.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#68
m/p-Xylenes
Concen: 100.800 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Tgt Ion:106 Resp: 259899
Ion Ratio Lower Upper
106 100
91 206.0 164.8 247.2









#69
o-Xylene
Concen: 51.366 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion:106 Resp: 124120
Ion Ratio Lower Upper
106 100
91 222.6 111.3 333.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#70
Styrene
Concen: 51.254 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20


Tgt Ion:104 Resp: 216738
Ion Ratio Lower Upper
104 100
78 51.8 41.4 62.2
103 53.9 43.1 64.7





#71
Bromoform
Concen: 50.792 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion:173 Resp: 5148
Ion Ratio Lower Upper
173 100
175 48.3 24.1 72.4
254 0.1 0.1 0.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Tgt Ion:152 Resp: 89016
Ion Ratio Lower Upper
152 100
115 86.4 43.2 129.6
150 173.4 0.0 346.8









#73
Isopropylbenzene
Concen: 50.548 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion:105 Resp: 33139
Ion Ratio Lower Upper
105 100
120 26.0 13.0 39.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#74
N-ethyl acetate
Concen: 52.581 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Tgt Ion: 43 Resp: 130606
Ion Ratio Lower Upper
43 100
70 41.7 33.4 50.0
55 23.9 19.1 28.7
61 23.7 19.0 28.4









#75
1,1,2,2-Tetrachloroethane
Concen: 50.745 ug/l
RT: 11.207 min Scan# 1660
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 83 Resp: 9289
Ion Ratio Lower Upper
83 100
131 9.7 4.9 14.6
85 64.0 32.0 96.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#76
1,2,3-Trichloropropane
Concen: 48.721 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Tgt Ion: 75 Resp: 88091
Ion Ratio Lower Upper
75 100
77 37.0 18.5 55.5









#77
Bromobenzene
Concen: 50.331 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion:156 Resp: 79404
Ion Ratio Lower Upper
156 100
77 165.7 82.8 248.6
158 95.3 47.6 142.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#78
n-propylbenzene
Concen: 50.646 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Tgt Ion: 91 Resp: 394375
Ion Ratio Lower Upper
91 100
120 22.7 11.4 34.1









#79
2-Chlorotoluene
Concen: 49.509 ug/l
RT: 11.360 min Scan# 1685
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 91 Resp: 230178
Ion Ratio Lower Upper
91 100
126 33.1 16.6 49.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#80
1,3,5-Trimethylbenzene
Concen: 50.761 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Tgt Ion:105 Resp: 273477
Ion Ratio Lower Upper
105 100
120 48.1 24.1 72.2









#81
trans-1,4-Dichloro-2-butene
Concen: 47.984 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion:	75	Resp:	2767
Ion	Ratio	Lower	Upper
75	100		
53	108.5	86.8	130.2
89	46.1	36.9	55.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#82
4-Chlorotoluene
Concen: 49.613 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20


Tgt Ion:	91	Resp:	269198
Ion	Ratio	Lower	Upper
91	100		
126	28.8	14.4	43.2





#83
tert-Butylbenzene
Concen: 49.905 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion:119 Resp: 277915
Ion Ratio Lower Upper
119 100
91 59.1 29.5 88.6
134 23.3 11.7 35.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#84
1,2,4-Trimethylbenzene
Concen: 50.345 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Tgt Ion:105 Resp: 272510
Ion Ratio Lower Upper
105 100
120 44.2 22.1 66.3









#85
sec-Butylbenzene
Concen: 50.574 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion:105 Resp: 35594
Ion Ratio Lower Upper
105 100
134 19.6 9.8 29.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#86
p-Isopropyltoluene
Concen: 50.067 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Tgt Ion:119 Resp: 297314
Ion Ratio Lower Upper
119 100
134 25.8 12.9 38.7
91 24.6 12.3 36.9





#87

1,3-Dichlorobenzene

Concen: 48.568 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion:146 Resp: 14940

Ion Ratio Lower Upper

146 100

111 41.2 20.6 61.8





148 63.2 31.6 94.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#88

1,4-Dichlorobenzene

Concen: 46.571 ug/l

RT: 12.037 min Scan# 1796

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20




Tgt Ion:146 Resp: 147146


Ion Ratio Lower Upper

146 100

111 40.7 20.3 61.0

148 64.3 32.1 96.5





#89

n-Butylbenzene

Concen: 49.801 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion:	91	Resp:	280794
Ion	Ratio	Lower	Upper
91	100		
92	54.2	27.1	81.3
134	25.7	12.9	38.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#90

Hexachloroethane

Concen: 49.968 ug/l




RT: 12.536 min Scan# 1878


Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20


Tgt Ion:	117	Resp:	51961
Ion	Ratio	Lower	Upper
117	100		
201	68.9	34.4	103.3





#91
1,2-Dichlorobenzene
Concen: 49.309 ug/l
RT: 12.329 min Scan# 1844
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion:146 Resp: 14174
Ion Ratio Lower Upper
146 100
111 43.4 21.7 65.1
148 64.1 32.0 96.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 51.873 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Tgt Ion: 75 Resp: 18747
Ion Ratio Lower Upper
75 100
155 83.8 41.9 125.7
157 102.9 51.4 154.3





#93

1,2,4-Trichlorobenzene

Concen: 49.101 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion:	180	Resp:	100340
Ion Ratio	Lower	Upper	
180	100		
182	93.4	46.7	140.1
145	31.6	15.8	47.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#94

Hexachlorobutadiene

Concen: 49.621 ug/l




RT: 13.719 min Scan# 2072


Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20


Tgt Ion:	225	Resp:	42659
Ion Ratio	Lower	Upper	
225	100		
223	65.0	32.5	97.5
227	64.4	32.2	96.6





#95
Naphthalene
Concen: 50.994 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion:128 Resp: 296912
Ion Ratio Lower Upper
128 100
127 12.6 10.1 15.1
129 10.9 8.7 13.1




Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#96
1,2,3-Trichlorobenzene
Concen: 49.263 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Tgt Ion:180 Resp: 96318
Ion Ratio Lower Upper
180 100
182 95.6 47.8 143.4
145 34.6 17.3 51.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
 Data File : VX046721.D
 Acq On : 17 Jun 2025 14:41
 Operator : JC/MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC100

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
 Supervised By :Mahesh Dadoda 06/18/2025

Quant Time: Jun 18 02:40:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 02:34:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	110482	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	183300	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	160869	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	78522	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	160514	105.037	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 210.080%#			
35) Dibromofluoromethane	5.391	113	129932	107.136	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 214.280%#			
50) Toluene-d8	8.647	98	458174	105.153	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 210.300%#			
62) 4-Bromofluorobenzene	11.079	95	170795	105.156	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 210.320%#			
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.179	85	138589	117.494	ug/l	99
3) Chloromethane	1.307	50	141568	111.195	ug/l	99
4) Vinyl Chloride	1.386	62	151879	111.868	ug/l	97
5) Bromomethane	1.611	94	75361	98.660	ug/l	94
6) Chloroethane	1.697	64	89532	108.805	ug/l	96
7) Trichlorofluoromethane	1.898	101	227688	110.449	ug/l	98
8) Diethyl Ether	2.148	74	77791	109.956	ug/l	96
9) 1,1,2-Trichlorotrifluo...	2.343	101	137759	108.929	ug/l	99
10) Methyl Iodide	2.465	142	174182	104.824	ug/l	99
11) Tert butyl alcohol	2.965	59	78366	571.345	ug/l	99
12) 1,1-Dichloroethene	2.331	96	134660	110.347	ug/l	98
13) Acrolein	2.246	56	108796	544.566	ug/l	99
14) Allyl chloride	2.678	41	243477	110.119	ug/l	100
15) Acrylonitrile	3.075	53	344981	559.348	ug/l	100
16) Acetone	2.386	43	263024	516.133	ug/l	99
17) Carbon Disulfide	2.526	76	383348	104.026	ug/l	98
18) Methyl Acetate	2.709	43	143551	110.909	ug/l	100
19) Methyl tert-butyl Ether	3.123	73	419445	110.387	ug/l	98
20) Methylene Chloride	2.800	84	147144	103.828	ug/l	98
21) trans-1,2-Dichloroethene	3.105	96	140792	107.824	ug/l	97
22) Diisopropyl ether	3.770	45	466831	109.846	ug/l	98
23) Vinyl Acetate	3.733	43	1930136	571.126	ug/l	99
24) 1,1-Dichloroethane	3.623	63	258546	107.149	ug/l	98
25) 2-Butanone	4.562	43	409960	568.810	ug/l	99
26) 2,2-Dichloropropane	4.489	77	194006	110.239	ug/l	99
27) cis-1,2-Dichloroethene	4.501	96	163672	106.691	ug/l	99
28) Bromochloromethane	4.910	49	114720	104.869	ug/l	99
29) Tetrahydrofuran	5.013	42	265233	569.746	ug/l	100
30) Chloroform	5.105	83	260888	108.113	ug/l	96
31) Cyclohexane	5.477	56	237424	103.757	ug/l	96
32) 1,1,1-Trichloroethane	5.391	97	231101	109.198	ug/l	99
36) 1,1-Dichloropropene	5.702	75	183355	104.733	ug/l	98
37) Ethyl Acetate	4.721	43	171444	106.094	ug/l	97
38) Carbon Tetrachloride	5.690	117	200914	105.749	ug/l	98
39) Methylcyclohexane	7.385	83	246277	106.902	ug/l	99
40) Benzene	6.044	78	553336	106.798	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
 Data File : VX046721.D
 Acq On : 17 Jun 2025 14:41
 Operator : JC/MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC100

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 06/18/2025
 Supervised By : Mahesh Dadoda 06/18/2025

Quant Time: Jun 18 02:40:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 02:34:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.928	41	97782	113.502	ug/l	97
42) 1,2-Dichloroethane	6.092	62	193498	106.278	ug/l	99
43) Isopropyl Acetate	6.342	43	291467	115.264	ug/l	100
44) Trichloroethene	7.129	130	142659	108.043	ug/l	96
45) 1,2-Dichloropropane	7.434	63	137291	106.920	ug/l	96
46) Dibromomethane	7.586	93	98964	107.240	ug/l	99
47) Bromodichloromethane	7.824	83	205870	109.319	ug/l	100
48) Methyl methacrylate	7.696	41	150002	118.279	ug/l	99
49) 1,4-Dioxane	7.659	88	38203	2040.447	ug/l	95
51) 4-Methyl-2-Pentanone	8.574	43	843944	559.584	ug/l	100
52) Toluene	8.720	92	339735	105.772	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	203470	116.448	ug/l	99
54) cis-1,3-Dichloropropene	8.366	75	221241	111.537	ug/l	98
55) 1,1,2-Trichloroethane	9.153	97	127195	106.264	ug/l	99
56) Ethyl methacrylate	9.116	69	205153	116.106	ug/l	100
57) 1,3-Dichloropropane	9.305	76	220021	106.740	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.244	63	541734	590.219	ug/l	100
59) 2-Hexanone	9.427	43	587221	563.200	ug/l	100
60) Dibromochloromethane	9.519	129	153731	108.983	ug/l	99
61) 1,2-Dibromoethane	9.610	107	133253	109.444	ug/l	99
64) Tetrachloroethene	9.275	164	115774	103.876	ug/l	98
65) Chlorobenzene	10.079	112	374913	105.577	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.159	131	131134	109.876	ug/l	99
67) Ethyl Benzene	10.189	91	663369	106.914	ug/l	100
68) m/p-Xylenes	10.299	106	492448	213.003	ug/l	100
69) o-Xylene	10.640	106	237631	109.675	ug/l	99
70) Styrene	10.653	104	407566	107.488	ug/l	99
71) Bromoform	10.799	173	101329	111.485	ug/l #	99
73) Isopropylbenzene	10.957	105	628803	108.731	ug/l	100
74) N-amyl acetate	10.842	43	260620	118.945	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	176523	109.317	ug/l	100
76) 1,2,3-Trichloropropane	11.238	75	147349m	92.387	ug/l	
77) Bromobenzene	11.195	156	150110	107.864	ug/l	99
78) n-propylbenzene	11.299	91	743979	108.310	ug/l	100
79) 2-Chlorotoluene	11.360	91	434989	106.065	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	517286	108.848	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	57608	113.225	ug/l	96
82) 4-Chlorotoluene	11.451	91	506852	105.896	ug/l	99
83) tert-Butylbenzene	11.713	119	533235	108.548	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	518524	108.598	ug/l	100
85) sec-Butylbenzene	11.890	105	670896	108.062	ug/l	99
86) p-Isopropyltoluene	12.006	119	561258	107.145	ug/l	100
87) 1,3-Dichlorobenzene	11.969	146	280540	103.388	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	280884	100.779	ug/l	99
89) n-Butylbenzene	12.329	91	536268	107.822	ug/l	100
90) Hexachloroethane	12.536	117	101647	110.811	ug/l	96
91) 1,2-Dichlorobenzene	12.335	146	267432	105.466	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	36858	115.616	ug/l	95
93) 1,2,4-Trichlorobenzene	13.585	180	196715	109.119	ug/l	99
94) Hexachlorobutadiene	13.725	225	80407	106.029	ug/l	98
95) Naphthalene	13.774	128	587914	114.468	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	189473	109.861	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
Data File : VX046721.D
Acq On : 17 Jun 2025 14:41
Operator : JC/MD
Sample : VSTDICC100
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025

Quant Time: Jun 18 02:40:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 02:34:13 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
 Data File : VX046721.D
 Acq On : 17 Jun 2025 14:41
 Operator : JC/MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
 Supervised By :Mahesh Dadoda 06/18/2025


Quant Time: Jun 18 02:40:16 2025


Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M

Quant Title : SW846 8260

QLast Update : Wed Jun 18 02:34:13 2025

Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Tgt Ion:168 Resp: 11048
Ion Ratio Lower Upper
168 100
99 61.4 48.5 72.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#2
Dichlorodifluoromethane
Concen: 117.494 ug/l
RT: 1.179 min Scan# 15
Delta R.T. -0.006 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Tgt Ion: 85 Resp: 138589
Ion Ratio Lower Upper
85 100
87 32.6 16.6 49.8









#3
Chloromethane
Concen: 111.195 ug/l
RT: 1.307 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Tgt Ion: 50 Resp: 141568
Ion Ratio Lower Upper
50 100
52 31.8 24.9 37.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#4
Vinyl Chloride
Concen: 111.868 ug/l
RT: 1.386 min Scan# 49
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Tgt Ion: 62 Resp: 151879
Ion Ratio Lower Upper
62 100
64 31.7 23.9 35.9









#5
Bromomethane
Concen: 98.660 ug/l
RT: 1.611 min Scan# 80
Delta R.T. -0.012 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Tgt Ion: 94 Resp: 7536
Ion Ratio Lower Upper
94 100
96 95.8 72.3 108.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#6
Chloroethane
Concen: 108.805 ug/l
RT: 1.697 min Scan# 100
Delta R.T. -0.006 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41


Tgt Ion: 64 Resp: 89532
Ion Ratio Lower Upper
64 100
66 31.8 27.1 40.7





#7
Trichlorofluoromethane
Concen: 110.449 ug/l
RT: 1.898 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion:101 Resp: 22768
Ion Ratio Lower Upper
101 100
103 63.5 51.8 77.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#8
Diethyl Ether
Concen: 109.956 ug/l
RT: 2.148 min Scan# 174
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Tgt Ion: 74 Resp: 77791
Ion Ratio Lower Upper
74 100
45 100.7 52.3 157.1





#9

1,1,2-Trichlorotrifluoroethane

Concen: 108.929 ug/l

RT: 2.343 min Scan# 206

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion:101 Resp: 137759

Ion Ratio Lower Upper

101 100

85 43.8 35.4 53.0





151 76.0 61.2 91.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#10

Methyl Iodide

Concen: 104.824 ug/l

RT: 2.465 min Scan# 226

Delta R.T. -0.006 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41




Tgt Ion:142 Resp: 174182


Ion Ratio Lower Upper

142 100

127 45.7 37.0 55.6

141 14.0 11.5 17.3





#11

Tert butyl alcohol

Concen: 571.345 ug/l

RT: 2.965 min Scan# 307

Delta R.T. 0.006 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 59 Resp: 78360

Ion Ratio Lower Upper

59 100





57 10.0 8.3 12.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#12

1,1-Dichloroethene

Concen: 110.347 ug/l

RT: 2.331 min Scan# 204

Delta R.T. -0.006 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41




Tgt Ion: 96 Resp: 134660


Ion Ratio Lower Upper

96 100

61 161.9 126.6 189.8

98 63.0 50.7 76.1





#13

Acrolein

Concen: 544.566 ug/l

RT: 2.246 min Scan# 190

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion: 56 Resp: 108790

Ion Ratio Lower Upper

56 100





55 69.7 55.2 82.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#14

Allyl chloride

Concen: 110.119 ug/l

RT: 2.678 min Scan# 261

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41




Tgt Ion: 41 Resp: 243477


Ion Ratio Lower Upper

41 100

39 66.5 53.6 80.4

76 33.6 26.8 40.2





#15

Acrylonitrile

Concen: 559.348 ug/l

RT: 3.075 min Scan# 31

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion: 53 Resp: 34498

Ion Ratio Lower Upper

53 100

52 81.9 65.1 97.7





51 36.0 28.9 43.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#16

Acetone

Concen: 516.133 ug/l

RT: 2.386 min Scan# 213

Delta R.T. -0.000 min

Lab File: VX046721.D




Acq: 17 Jun 2025 14:41


Tgt Ion: 43 Resp: 263024

Ion Ratio Lower Upper

43 100

58 30.2 24.5 36.7





#17

Carbon Disulfide

Concen: 104.026 ug/l

RT: 2.526 min Scan# 21

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 76 Resp: 383348

Ion Ratio Lower Upper

76 100





78 8.8 7.5 11.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#18

Methyl Acetate

Concen: 110.909 ug/l

RT: 2.709 min Scan# 266

Delta R.T. -0.006 min

Lab File: VX046721.D




Acq: 17 Jun 2025 14:41


Tgt Ion: 43 Resp: 143551

Ion Ratio Lower Upper

43 100

74 23.2 18.4 27.6









#19
Methyl tert-butyl Ether
Concen: 110.387 ug/l
RT: 3.123 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Tgt Ion: 73 Resp: 41944
Ion Ratio Lower Upper
73 100
57 21.8 18.2 27.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#20
Methylene Chloride
Concen: 103.828 ug/l
RT: 2.800 min Scan# 281
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41


Tgt Ion: 84 Resp: 147144
Ion Ratio Lower Upper
84 100
49 122.3 101.3 151.9
51 38.8 31.0 46.6
86 63.5 51.1 76.7





#21
trans-1,2-Dichloroethene
Concen: 107.824 ug/l
RT: 3.105 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion: 96 Resp: 140792
Ion Ratio Lower Upper
96 100
61 136.9 112.5 168.7
98 63.1 51.9 77.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#22
Diisopropyl ether
Concen: 109.846 ug/l
RT: 3.770 min Scan# 440
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Tgt Ion: 45 Resp: 466831
Ion Ratio Lower Upper
45 100
43 112.3 91.9 137.9
87 26.8 22.0 33.0
59 11.4 9.1 13.7









#23
Vinyl Acetate
Concen: 571.126 ug/l
RT: 3.733 min Scan# 41
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Tgt Ion: 43 Resp: 1930130
Ion Ratio Lower Upper
43 100
86 10.4 8.2 12.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#24
1,1-Dichloroethane
Concen: 107.149 ug/l
RT: 3.623 min Scan# 416
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Tgt Ion: 63 Resp: 258546
Ion Ratio Lower Upper
63 100
98 7.0 4.0 11.9
100 4.6 2.2 6.6





#25

2-Butanone

Concen: 568.810 ug/l

RT: 4.562 min Scan# 51

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 43 Resp: 409960

Ion Ratio Lower Upper

43 100





72 24.6 20.2 30.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#26

2,2-Dichloropropane

Concen: 110.239 ug/l

RT: 4.489 min Scan# 558

Delta R.T. -0.000 min

Lab File: VX046721.D




Acq: 17 Jun 2025 14:41


Tgt Ion: 77 Resp: 194006

Ion Ratio Lower Upper

77 100

97 22.8 11.8 35.3





#27

cis-1,2-Dichloroethene

Concen: 106.691 ug/l

RT: 4.501 min Scan# 560

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 96 Resp: 163672

Ion Ratio Lower Upper

96 100

61 144.8 0.0 291.0





98 64.6 0.0 126.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#28

Bromochloromethane

Concen: 104.869 ug/l

RT: 4.910 min Scan# 627

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41




Tgt Ion: 49 Resp: 114720


Ion Ratio Lower Upper

49 100

129 1.9 0.0 3.8

130 77.6 61.4 92.0





#29

Tetrahydrofuran

Concen: 569.746 ug/l

RT: 5.013 min Scan# 644

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion: 42 Resp: 26523

Ion Ratio Lower Upper

42 100

72 43.5 34.9 52.3





71 40.2 32.2 48.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#30

Chloroform

Concen: 108.113 ug/l

RT: 5.105 min Scan# 659

Delta R.T. -0.000 min

Lab File: VX046721.D




Acq: 17 Jun 2025 14:41


Tgt Ion: 83 Resp: 260888

Ion Ratio Lower Upper

83 100


85 63.6 53.7 80.5





#31
Cyclohexane
Concen: 103.757 ug/l
RT: 5.477 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion: 56 Resp: 237424
Ion Ratio Lower Upper
56 100
69 28.9 24.8 37.2
84 81.9 68.2 102.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#32
1,1,1-Trichloroethane
Concen: 109.198 ug/l
RT: 5.391 min Scan# 706
Delta R.T. -0.006 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Tgt Ion: 97 Resp: 231101
Ion Ratio Lower Upper
97 100
99 63.5 51.3 76.9
61 44.1 36.4 54.6





#33

1,2-Dichloroethane-d4

Concen: 105.037 ug/l

RT: 5.958 min Scan# 799

Delta R.T. -0.006 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 65 Resp: 160514

Ion Ratio Lower Upper

65 100





67 52.1 0.0 105.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. -0.006 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41




Tgt Ion:114 Resp: 183300


Ion Ratio Lower Upper

114 100

63 21.0 0.0 44.2

88 16.4 0.0 33.2





#35

Dibromofluoromethane

Concen: 107.136 ug/l

RT: 5.391 min Scan# 707

Delta R.T. -0.006 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion:113 Resp: 129932

Ion Ratio Lower Upper

113 100

111 102.2 82.0 123.0





192 18.7 15.3 22.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#36

1,1-Dichloropropene

Concen: 104.733 ug/l

RT: 5.702 min Scan# 757

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41




Tgt Ion: 75 Resp: 183355


Ion Ratio Lower Upper

75 100

110 35.3 17.2 51.5

77 31.2 24.2 36.4





#37

Ethyl Acetate

Concen: 106.094 ug/l

RT: 4.721 min Scan# 596

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 43 Resp: 17144

Ion Ratio Lower Upper

43 100

61 13.8 9.7 14.5





70 10.4 8.1 12.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#38

Carbon Tetrachloride

Concen: 105.749 ug/l

RT: 5.690 min Scan# 755

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41




Tgt Ion:117 Resp: 200914


Ion Ratio Lower Upper

117 100

119 96.0 75.2 112.8

121 30.8 24.1 36.1





#39

Methylcyclohexane

Concen: 106.902 ug/l

RT: 7.385 min Scan# 1033

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion: 83 Resp: 24627

Ion Ratio Lower Upper

83 100

55 77.9 63.0 94.4





98 48.8 38.8 58.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#40

Benzene

Concen: 106.798 ug/l

RT: 6.044 min Scan# 813

Delta R.T. -0.000 min

Lab File: VX046721.D




Acq: 17 Jun 2025 14:41


Tgt Ion: 78 Resp: 553336

Ion Ratio Lower Upper

78 100

77 23.8 18.8 28.2









#41
Methacrylonitrile
Concen: 113.502 ug/l
RT: 4.928 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Tgt Ion: 41 Resp: 9778
Ion Ratio Lower Upper
41 100
39 54.8 45.9 68.9
67 73.6 61.2 91.8
52 31.6 25.5 38.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#42
1,2-Dichloroethane
Concen: 106.278 ug/l
RT: 6.092 min Scan# 821
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Tgt Ion: 62 Resp: 193498
Ion Ratio Lower Upper
62 100
98 9.1 0.0 17.8





#43

Isopropyl Acetate

Concen: 115.264 ug/l

RT: 6.342 min Scan# 862

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion: 43 Resp: 29146

Ion Ratio Lower Upper

43 100

61 19.4 15.3 22.9





87 12.5 10.0 15.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#44

Trichloroethene

Concen: 108.043 ug/l

RT: 7.129 min Scan# 991

Delta R.T. -0.000 min

Lab File: VX046721.D




Acq: 17 Jun 2025 14:41


Tgt Ion:130 Resp: 142659

Ion Ratio Lower Upper

130 100

95 97.1 0.0 202.6





#45

1,2-Dichloropropane

Concen: 106.920 ug/l

RT: 7.434 min Scan# 1041

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion: 63 Resp: 13729

Ion Ratio Lower Upper

63 100





65 30.8 22.9 34.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#46

Dibromomethane

Concen: 107.240 ug/l

RT: 7.586 min Scan# 1066

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41




Tgt Ion: 93 Resp: 98964


Ion Ratio Lower Upper

93 100

95 82.5 66.2 99.4

174 97.4 77.0 115.6





#47

Bromodichloromethane

Concen: 109.319 ug/l

RT: 7.824 min Scan# 1105

Delta R.T. -0.000 min

Lab File: VX046721.D


Acq: 17 Jun 2025 14:41

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100






Tgt Ion:	83	Resp:	205870
Ion	Ratio	Lower	Upper
83	100		
85	64.4	51.4	77.2
127	8.8	6.9	10.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#48

Methyl methacrylate

Concen: 118.279 ug/l




RT: 7.696 min Scan# 1084


Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41

Tgt Ion:	41	Resp:	150002
Ion	Ratio	Lower	Upper
41	100		
69	80.4	63.4	95.2
39	56.2	45.4	68.0





#49

1,4-Dioxane

Concen: 2040.447 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. 0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 88 Resp: 3820

Ion Ratio Lower Upper

88 100

43 37.8 31.2 46.8





58 73.7 54.7 82.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#50

Toluene-d8

Concen: 105.153 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. -0.000 min

Lab File: VX046721.D




Acq: 17 Jun 2025 14:41


Tgt Ion: 98 Resp: 458174

Ion Ratio Lower Upper

98 100

100 66.7 53.0 79.4





#51

4-Methyl-2-Pentanone

Concen: 559.584 ug/l

RT: 8.574 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 43 Resp: 843944

Ion Ratio Lower Upper





43 100

58 38.5 30.8 46.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#52

Toluene

Concen: 105.772 ug/l

RT: 8.720 min Scan# 1252

Delta R.T. -0.000 min

Lab File: VX046721.D




Acq: 17 Jun 2025 14:41


Tgt Ion: 92 Resp: 339735

Ion Ratio Lower Upper

92 100

91 170.4 134.8 202.2









#53
t-1,3-Dichloropropene
Concen: 116.448 ug/l
RT: 8.976 min Scan# 1294
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Tgt Ion: 75 Resp: 203470
Ion Ratio Lower Upper
75 100
77 31.6 25.6 38.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#54
cis-1,3-Dichloropropene
Concen: 111.537 ug/l
RT: 8.366 min Scan# 1194
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Tgt Ion: 75 Resp: 221241
Ion Ratio Lower Upper
75 100
77 32.0 25.3 37.9
39 51.0 39.2 58.8





#55

1,1,2-Trichloroethane

Concen: 106.264 ug/l

RT: 9.153 min Scan# 1323

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 97 Resp: 12719

Ion Ratio Lower Upper

97 100

83 84.7 67.2 100.8

85 55.4 44.1 66.1





99 62.7 50.3 75.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#56

Ethyl methacrylate

Concen: 116.106 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41




Tgt Ion: 69 Resp: 205153


Ion Ratio Lower Upper

69 100

41 69.6 55.4 83.2

39 36.9 29.7 44.5





#57

1,3-Dichloropropane

Concen: 106.740 ug/l

RT: 9.305 min Scan# 11348

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion: 76 Resp: 22002

Ion Ratio Lower Upper

76 100





78 32.1 25.7 38.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#58

2-Chloroethyl Vinyl ether

Concen: 590.219 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. -0.000 min

Lab File: VX046721.D




Acq: 17 Jun 2025 14:41


Tgt Ion: 63 Resp: 541734

Ion Ratio Lower Upper

63 100

106 28.2 22.6 34.0





#59

2-Hexanone

Concen: 563.200 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion: 43 Resp: 58722

Ion Ratio Lower Upper

43 100





58 54.0 26.8 80.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#60

Dibromochloromethane

Concen: 108.983 ug/l

RT: 9.519 min Scan# 1383

Delta R.T. -0.000 min

Lab File: VX046721.D




Acq: 17 Jun 2025 14:41


Tgt Ion:129 Resp: 153731

Ion Ratio Lower Upper

129 100


127 77.6 38.4 115.2





#61
1,2-Dibromoethane
Concen: 109.444 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion:107 Resp: 13325
Ion Ratio Lower Upper
107 100
109 93.8 75.8 113.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#62
4-Bromofluorobenzene
Concen: 105.156 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41


Tgt Ion: 95 Resp: 170795
Ion Ratio Lower Upper
95 100
174 76.4 0.0 150.4
176 72.7 0.0 145.0





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100



Tgt Ion:117 Resp: 160869
Ion Ratio Lower Upper
117 100
82 54.4 44.6 66.8
119 32.1 25.8 38.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#64
Tetrachloroethene
Concen: 103.876 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Tgt Ion:164 Resp: 115774
Ion Ratio Lower Upper
164 100
166 128.0 104.6 157.0
129 96.2 76.3 114.5
131 91.4 74.1 111.1









#65
Chlorobenzene
Concen: 105.577 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Tgt Ion:112 Resp: 374911
Ion Ratio Lower Upper
112 100
114 31.7 25.5 38.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 109.876 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Tgt Ion:131 Resp: 131134
Ion Ratio Lower Upper
131 100
133 93.8 47.3 141.8
119 63.0 32.1 96.3





#67

Ethyl Benzene

Concen: 106.914 ug/l

RT: 10.189 min Scan# 1493

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion: 91 Resp: 663369

Ion Ratio Lower Upper

91 100





106 30.4 24.2 36.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#68

m/p-Xylenes

Concen: 213.003 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. -0.000 min

Lab File: VX046721.D




Acq: 17 Jun 2025 14:41


Tgt Ion:106 Resp: 492448

Ion Ratio Lower Upper

106 100

91 206.0 164.8 247.2









#69
o-Xylene
Concen: 109.675 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Tgt Ion:106 Resp: 23763
Ion Ratio Lower Upper
106 100
91 220.5 111.3 333.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#70
Styrene
Concen: 107.488 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41


Tgt Ion:104 Resp: 407566
Ion Ratio Lower Upper
104 100
78 52.3 41.4 62.2
103 54.3 43.1 64.7





#71
Bromoform
Concen: 111.485 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion:173 Resp: 101329
Ion Ratio Lower Upper
173 100
175 48.9 24.1 72.4
254 0.1 0.1 0.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Tgt Ion:152 Resp: 78522
Ion Ratio Lower Upper
152 100
115 115.4 43.2 129.6
150 191.6 0.0 346.8









#73
Isopropylbenzene
Concen: 108.731 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Tgt Ion:105 Resp: 62880
Ion Ratio Lower Upper
105 100
120 26.1 13.0 39.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#74
N-ethyl acetate
Concen: 118.945 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Tgt Ion: 43 Resp: 260620
Ion Ratio Lower Upper
43 100
70 40.8 33.4 50.0
55 23.2 19.1 28.7
61 24.1 19.0 28.4









#75
1,1,2,2-Tetrachloroethane
Concen: 109.317 ug/l
RT: 11.207 min Scan# 1660
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Tgt Ion: 83 Resp: 17652
Ion Ratio Lower Upper
83 100
131 10.3 4.9 14.6
85 63.8 32.0 96.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#76
1,2,3-Trichloropropane
Concen: 92.387 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Tgt Ion: 75 Resp: 147349
Ion Ratio Lower Upper
75 100
77 42.6 18.5 55.5





#77

Bromobenzene

Concen: 107.864 ug/l

RT: 11.195 min Scan# 1658

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion:156 Resp: 150110

Ion Ratio Lower Upper

156 100

77 166.5 82.8 248.6





158 97.7 47.6 142.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#78

n-propylbenzene

Concen: 108.310 ug/l

RT: 11.299 min Scan# 1675

Delta R.T. -0.000 min

Lab File: VX046721.D




Acq: 17 Jun 2025 14:41


Tgt Ion: 91 Resp: 743979

Ion Ratio Lower Upper

91 100

120 22.6 11.4 34.1





#79

2-Chlorotoluene

Concen: 106.065 ug/l

RT: 11.360 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 91 Resp: 434989

Ion Ratio Lower Upper

91 100





126 33.2 16.6 49.7

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#80

1,3,5-Trimethylbenzene

Concen: 108.848 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. -0.000 min

Lab File: VX046721.D




Acq: 17 Jun 2025 14:41


Tgt Ion:105 Resp: 517286

Ion Ratio Lower Upper

105 100

120 48.0 24.1 72.2









#81
trans-1,4-Dichloro-2-butene
Concen: 113.225 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Tgt Ion: 75 Resp: 57608
Ion Ratio Lower Upper
75 100
53 103.4 86.8 130.2
89 46.9 36.9 55.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#82
4-Chlorotoluene
Concen: 105.896 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41


Tgt Ion: 91 Resp: 506852
Ion Ratio Lower Upper
91 100
126 29.1 14.4 43.2





#83
tert-Butylbenzene
Concen: 108.548 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion:119 Resp: 533235
Ion Ratio Lower Upper
119 100
91 59.1 29.5 88.6
134 22.7 11.7 35.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#84
1,2,4-Trimethylbenzene
Concen: 108.598 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41


Tgt Ion:105 Resp: 518524
Ion Ratio Lower Upper
105 100
120 44.0 22.1 66.3





#85
sec-Butylbenzene
Concen: 108.062 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion:105 Resp: 670890
Ion Ratio Lower Upper
105 100
134 20.0 9.8 29.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#86
p-Isopropyltoluene
Concen: 107.145 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41


Tgt Ion:119 Resp: 561258
Ion Ratio Lower Upper
119 100
134 25.7 12.9 38.7
91 24.7 12.3 36.9





#87
1,3-Dichlorobenzene
Concen: 103.388 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100




Tgt Ion:146 Resp: 280540
Ion Ratio Lower Upper
146 100
111 41.7 20.6 61.8
148 63.9 31.6 94.8

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025




#88
1,4-Dichlorobenzene
Concen: 100.779 ug/l
RT: 12.036 min Scan# 1796
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41



Tgt Ion:146 Resp: 280884
Ion Ratio Lower Upper
146 100
111 39.5 20.3 61.0
148 63.9 32.1 96.5





#89

n-Butylbenzene

Concen: 107.822 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 91 Resp: 53626

Ion Ratio Lower Upper

91 100

92 54.5 27.1 81.3





134 25.6 12.9 38.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#90

Hexachloroethane

Concen: 110.811 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX046721.D




Acq: 17 Jun 2025 14:41


Tgt Ion:117 Resp: 101647

Ion Ratio Lower Upper

117 100


201 72.5 34.4 103.3





#91
1,2-Dichlorobenzene
Concen: 105.466 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.006 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion:146 Resp: 26743
Ion Ratio Lower Upper
146 100
111 43.2 21.7 65.1
148 63.5 32.0 96.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 115.616 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Tgt Ion: 75 Resp: 36858
Ion Ratio Lower Upper
75 100
155 85.8 41.9 125.7
157 110.6 51.4 154.3





#93

1,2,4-Trichlorobenzene

Concen: 109.119 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion:180 Resp: 19671

Ion Ratio Lower Upper

180 100

182 94.3 46.7 140.1





145 32.0 15.8 47.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#94

Hexachlorobutadiene

Concen: 106.029 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. 0.006 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41




Tgt Ion:225 Resp: 80407


Ion Ratio Lower Upper

225 100

223 62.8 32.5 97.5


227 63.0 32.2 96.6





#95
Naphthalene
Concen: 114.468 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion:128 Resp: 587914
Ion Ratio Lower Upper
128 100
127 12.7 10.1 15.1
129 10.9 8.7 13.1




Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#96
1,2,3-Trichlorobenzene
Concen: 109.861 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Tgt Ion:180 Resp: 189473
Ion Ratio Lower Upper
180 100
182 96.0 47.8 143.4
145 34.4 17.3 51.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
 Data File : VX046722.D
 Acq On : 17 Jun 2025 15:02
 Operator : JC/MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC150

Quant Time: Jun 18 02:41:13 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 02:34:13 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
 Supervised By :Mahesh Dadoda 06/18/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	107848	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	178346	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	156970	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	75829	50.000	ug/l	# 0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	5.964	65	230915	154.797	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 309.600%#	
35) Dibromofluoromethane	5.397	113	187536	158.930	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 317.860%#	
50) Toluene-d8	8.653	98	660221	155.733	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 311.460%#	
62) 4-Bromofluorobenzene	11.079	95	244007	154.404	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 308.800%#	

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.179	85	193321	167.898	ug/l	99
3) Chloromethane	1.307	50	202755	163.144	ug/l	97
4) Vinyl Chloride	1.386	62	208285	157.161	ug/l	96
5) Bromomethane	1.612	94	90708	121.652	ug/l	97
6) Chloroethane	1.691	64	123323	153.530	ug/l	96
7) Trichlorofluoromethane	1.898	101	314552	156.313	ug/l	98
8) Diethyl Ether	2.142	74	108801	157.544	ug/l	97
9) 1,1,2-Trichlorotrifluo...	2.343	101	192071	155.583	ug/l	99
10) Methyl Iodide	2.465	142	243423	148.201	ug/l	99
11) Tert butyl alcohol	2.971	59	111150	830.155	ug/l	99
12) 1,1-Dichloroethene	2.331	96	188770	158.466	ug/l	97
13) Acrolein	2.246	56	156844	804.238	ug/l	99
14) Allyl chloride	2.678	41	340117	157.583	ug/l	99
15) Acrylonitrile	3.075	53	477449	793.036	ug/l	99
16) Acetone	2.386	43	378554	760.980	ug/l	99
17) Carbon Disulfide	2.526	76	535657	148.907	ug/l	99
18) Methyl Acetate	2.715	43	209349	165.695	ug/l	99
19) Methyl tert-butyl Ether	3.123	73	585091	157.742	ug/l	99
20) Methylene Chloride	2.800	84	201983	146.004	ug/l	99
21) trans-1,2-Dichloroethene	3.105	96	192490	151.016	ug/l	99
22) Diisopropyl ether	3.770	45	638477	153.904	ug/l	99
23) Vinyl Acetate	3.733	43	2691267	815.794	ug/l	99
24) 1,1-Dichloroethane	3.623	63	360395	153.006	ug/l	98
25) 2-Butanone	4.562	43	570672	811.132	ug/l	99
26) 2,2-Dichloropropane	4.483	77	286713	166.896	ug/l	99
27) cis-1,2-Dichloroethene	4.495	96	227788	152.112	ug/l	99
28) Bromochloromethane	4.910	49	161801	151.520	ug/l	100
29) Tetrahydrofuran	5.013	42	370760	815.879	ug/l	100
30) Chloroform	5.105	83	358782	152.312	ug/l	99
31) Cyclohexane	5.483	56	330235	147.842	ug/l	98
32) 1,1,1-Trichloroethane	5.397	97	320382	155.082	ug/l	99
36) 1,1-Dichloropropene	5.702	75	256619	150.653	ug/l	99
37) Ethyl Acetate	4.721	43	245724	156.284	ug/l	99
38) Carbon Tetrachloride	5.684	117	284066	153.668	ug/l	99
39) Methylcyclohexane	7.385	83	346118	154.413	ug/l	100
40) Benzene	6.044	78	763341	151.423	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
 Data File : VX046722.D
 Acq On : 17 Jun 2025 15:02
 Operator : JC/MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC150

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
 Supervised By :Mahesh Dadoda 06/18/2025

Quant Time: Jun 18 02:41:13 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 02:34:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.928	41	142283	169.745	ug/l	94
42) 1,2-Dichloroethane	6.092	62	265740	150.011	ug/l	99
43) Isopropyl Acetate	6.342	43	412067	167.483	ug/l	99
44) Trichloroethene	7.129	130	195943	152.520	ug/l	99
45) 1,2-Dichloropropane	7.434	63	191181	153.025	ug/l	95
46) Dibromomethane	7.580	93	138541	154.297	ug/l	99
47) Bromodichloromethane	7.824	83	288826	157.630	ug/l	100
48) Methyl methacrylate	7.696	41	211908	171.734	ug/l	99
49) 1,4-Dioxane	7.659	88	54646	2987.386	ug/l	96
51) 4-Methyl-2-Pentanone	8.574	43	1173666	799.825	ug/l	100
52) Toluene	8.720	92	471273	150.800	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	288893	169.929	ug/l	99
54) cis-1,3-Dichloropropene	8.366	75	314965	163.198	ug/l	98
55) 1,1,2-Trichloroethane	9.153	97	175973	151.099	ug/l	99
56) Ethyl methacrylate	9.116	69	288720	167.940	ug/l	99
57) 1,3-Dichloropropane	9.305	76	301418	150.290	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.244	63	754730	845.118	ug/l	100
59) 2-Hexanone	9.433	43	816655	805.005	ug/l	99
60) Dibromochloromethane	9.519	129	215347	156.904	ug/l	99
61) 1,2-Dibromoethane	9.610	107	185268	156.392	ug/l	99
64) Tetrachloroethene	9.275	164	159752	146.895	ug/l	98
65) Chlorobenzene	10.080	112	517859	149.453	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.165	131	181981	156.268	ug/l	99
67) Ethyl Benzene	10.195	91	915724	151.251	ug/l	98
68) m/p-Xylenes	10.299	106	676085	299.697	ug/l	99
69) o-Xylene	10.640	106	328782	155.514	ug/l	98
70) Styrene	10.653	104	566606	153.143	ug/l	100
71) Bromoform	10.799	173	143229	161.498	ug/l	100
73) Isopropylbenzene	10.964	105	867728	155.374	ug/l	100
74) N-amyl acetate	10.842	43	369757	174.748	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	244406	156.730	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	202926m	131.753	ug/l	
77) Bromobenzene	11.195	156	207348	154.285	ug/l	98
78) n-propylbenzene	11.305	91	1025742	154.633	ug/l	100
79) 2-Chlorotoluene	11.366	91	595555	150.373	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	707005	154.053	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	86677	176.408	ug/l	93
82) 4-Chlorotoluene	11.451	91	697556	150.915	ug/l	99
83) tert-Butylbenzene	11.713	119	735750	155.093	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	719283	155.994	ug/l	100
85) sec-Butylbenzene	11.890	105	916769	152.910	ug/l	99
86) p-Isopropyltoluene	12.006	119	771998	152.610	ug/l	100
87) 1,3-Dichlorobenzene	11.969	146	391049	149.232	ug/l	100
88) 1,4-Dichlorobenzene	12.036	146	387139	143.836	ug/l	99
89) n-Butylbenzene	12.329	91	726563	151.271	ug/l	100
90) Hexachloroethane	12.536	117	142663	161.049	ug/l	98
91) 1,2-Dichlorobenzene	12.329	146	370234	151.192	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	12.939	75	54014	175.448	ug/l	97
93) 1,2,4-Trichlorobenzene	13.585	180	263803	151.531	ug/l	97
94) Hexachlorobutadiene	13.725	225	95520	130.431	ug/l	100
95) Naphthalene	13.774	128	846206	170.609	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	262348	157.517	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
Data File : VX046722.D
Acq On : 17 Jun 2025 15:02
Operator : JC/MD
Sample : VSTDICC150
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025

Quant Time: Jun 18 02:41:13 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 02:34:13 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
 Data File : VX046722.D
 Acq On : 17 Jun 2025 15:02
 Operator : JC/MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1


Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC150

Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
 Supervised By :Mahesh Dadoda 06/18/2025

Quant Time: Jun 18 02:41:13 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 02:34:13 2025
 Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.562 min Scan# 71
Delta R.T. 0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Tgt Ion:168 Resp: 107848
Ion Ratio Lower Upper
168 100
99 59.3 48.5 72.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#2
Dichlorodifluoromethane
Concen: 167.898 ug/l
RT: 1.179 min Scan# 15
Delta R.T. -0.006 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02


Tgt Ion: 85 Resp: 193321
Ion Ratio Lower Upper
85 100
87 32.6 16.6 49.8





#3
Chloromethane
Concen: 163.144 ug/l
RT: 1.307 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150



Tgt Ion: 50 Resp: 20275
Ion Ratio Lower Upper
50 100
52 32.6 24.9 37.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#4
Vinyl Chloride
Concen: 157.161 ug/l
RT: 1.386 min Scan# 49
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02


Tgt Ion: 62 Resp: 208285
Ion Ratio Lower Upper
62 100
64 31.9 23.9 35.9





#5
Bromomethane
Concen: 121.652 ug/l
RT: 1.612 min Scan# 80
Delta R.T. -0.012 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150



Tgt Ion: 94 Resp: 90708
Ion Ratio Lower Upper
94 100
96 92.9 72.3 108.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#6
Chloroethane
Concen: 153.530 ug/l
RT: 1.691 min Scan# 99
Delta R.T. -0.012 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02


Tgt Ion: 64 Resp: 123323
Ion Ratio Lower Upper
64 100
66 31.7 27.1 40.7





#7
Trichlorofluoromethane
Concen: 156.313 ug/l
RT: 1.898 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150



Tgt Ion:101 Resp: 314552
Ion Ratio Lower Upper
101 100
103 63.4 51.8 77.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#8
Diethyl Ether
Concen: 157.544 ug/l
RT: 2.142 min Scan# 173
Delta R.T. -0.006 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion: 74 Resp: 108801
Ion Ratio Lower Upper
74 100
45 101.4 52.3 157.1





#9

1,1,2-Trichlorotrifluoroethane

Concen: 155.583 ug/l

RT: 2.343 min Scan# 206

Delta R.T. -0.000 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC150

Tgt Ion:101 Resp: 19207

Ion Ratio Lower Upper

101 100

85 43.7 35.4 53.0





151 75.4 61.2 91.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#10

Methyl Iodide

Concen: 148.201 ug/l

RT: 2.465 min Scan# 226

Delta R.T. -0.006 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02




Tgt Ion:142 Resp: 243423


Ion Ratio Lower Upper

142 100

127 45.5 37.0 55.6


141 13.6 11.5 17.3





#11
Tert butyl alcohol
Concen: 830.155 ug/l
RT: 2.971 min Scan# 307
Delta R.T. 0.012 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150



Tgt Ion: 59 Resp: 111150
Ion Ratio Lower Upper
59 100
57 10.7 8.3 12.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#12
1,1-Dichloroethene
Concen: 158.466 ug/l
RT: 2.331 min Scan# 204
Delta R.T. -0.006 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion: 96 Resp: 188770
Ion Ratio Lower Upper
96 100
61 163.4 126.6 189.8
98 63.6 50.7 76.1





#13

Acrolein

Concen: 804.238 ug/l

RT: 2.246 min Scan# 190

Delta R.T. -0.000 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 56 Resp: 156844

Ion Ratio Lower Upper





56 100

55 69.8 55.2 82.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#14

Allyl chloride

Concen: 157.583 ug/l

RT: 2.678 min Scan# 261

Delta R.T. -0.000 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02




Tgt Ion: 41 Resp: 340117


Ion Ratio Lower Upper

41 100

39 65.9 53.6 80.4


76 33.6 26.8 40.2





#15
Acrylonitrile
Concen: 793.036 ug/l
RT: 3.075 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150



Tgt Ion: 53 Resp: 477449
Ion Ratio Lower Upper
53 100
52 82.8 65.1 97.7
51 36.3 28.9 43.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#16
Acetone
Concen: 760.980 ug/l
RT: 2.386 min Scan# 213
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion: 43 Resp: 378554
Ion Ratio Lower Upper
43 100
58 30.2 24.5 36.7





#17

Carbon Disulfide

Concen: 148.907 ug/l

RT: 2.526 min Scan# 21

Delta R.T. -0.000 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 76 Resp: 53565

Ion Ratio Lower Upper

76 100





78 9.0 7.5 11.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#18

Methyl Acetate

Concen: 165.695 ug/l

RT: 2.715 min Scan# 267

Delta R.T. -0.000 min

Lab File: VX046722.D




Acq: 17 Jun 2025 15:02


Tgt Ion: 43 Resp: 209349

Ion Ratio Lower Upper

43 100

74 22.4 18.4 27.6









#19
Methyl tert-butyl Ether
Concen: 157.742 ug/l
RT: 3.123 min Scan# 311
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150

Tgt Ion: 73 Resp: 58509
Ion Ratio Lower Upper
73 100
57 22.1 18.2 27.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#20
Methylene Chloride
Concen: 146.004 ug/l
RT: 2.800 min Scan# 281
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02


Tgt Ion: 84 Resp: 201983
Ion Ratio Lower Upper
84 100
49 126.6 101.3 151.9
51 39.0 31.0 46.6
86 66.5 51.1 76.7





#21
trans-1,2-Dichloroethene
Concen: 151.016 ug/l
RT: 3.105 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150



Tgt Ion: 96 Resp: 192490
Ion Ratio Lower Upper
96 100
61 139.6 112.5 168.7
98 62.5 51.9 77.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#22
Diisopropyl ether
Concen: 153.904 ug/l
RT: 3.770 min Scan# 440
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion: 45 Resp: 638477
Ion Ratio Lower Upper
45 100
43 115.7 91.9 137.9
87 26.6 22.0 33.0
59 11.1 9.1 13.7









#23
Vinyl Acetate
Concen: 815.794 ug/l
RT: 3.733 min Scan# 41
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Tgt Ion: 43 Resp: 269126
Ion Ratio Lower Upper
43 100
86 10.4 8.2 12.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#24
1,1-Dichloroethane
Concen: 153.006 ug/l
RT: 3.623 min Scan# 416
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion: 63 Resp: 360395
Ion Ratio Lower Upper
63 100
98 7.1 4.0 11.9
100 4.3 2.2 6.6





#25

2-Butanone

Concen: 811.132 ug/l

RT: 4.562 min Scan# 51

Delta R.T. -0.000 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 43 Resp: 570672

Ion Ratio Lower Upper

43 100





72 24.7 20.2 30.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#26

2,2-Dichloropropane

Concen: 166.896 ug/l

RT: 4.483 min Scan# 557

Delta R.T. -0.006 min

Lab File: VX046722.D




Acq: 17 Jun 2025 15:02


Tgt Ion: 77 Resp: 286713

Ion Ratio Lower Upper

77 100

97 22.9 11.8 35.3





#27

cis-1,2-Dichloroethene

Concen: 152.112 ug/l

RT: 4.495 min Scan# 51

Delta R.T. -0.006 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 96 Resp: 22778

Ion Ratio Lower Upper

96 100

61 145.0 0.0 291.0





98 64.7 0.0 126.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#28

Bromochloromethane

Concen: 151.520 ug/l

RT: 4.910 min Scan# 627

Delta R.T. -0.000 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02




Tgt Ion: 49 Resp: 161801


Ion Ratio Lower Upper

49 100

129 2.0 0.0 3.8

130 76.9 61.4 92.0





#29

Tetrahydrofuran

Concen: 815.879 ug/l

RT: 5.013 min Scan# 644

Delta R.T. -0.000 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 42 Resp: 370760

Ion Ratio Lower Upper

42 100

72 43.8 34.9 52.3





71 40.1 32.2 48.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#30

Chloroform

Concen: 152.312 ug/l

RT: 5.105 min Scan# 659

Delta R.T. -0.000 min

Lab File: VX046722.D




Acq: 17 Jun 2025 15:02


Tgt Ion: 83 Resp: 358782

Ion Ratio Lower Upper

83 100


85 66.2 53.7 80.5





#31
Cyclohexane
Concen: 147.842 ug/l
RT: 5.483 min Scan# 721
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150



Tgt Ion: 56 Resp: 330235
Ion Ratio Lower Upper
56 100
69 29.4 24.8 37.2
84 83.6 68.2 102.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#32
1,1,1-Trichloroethane
Concen: 155.082 ug/l
RT: 5.397 min Scan# 707
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion: 97 Resp: 320382
Ion Ratio Lower Upper
97 100
99 64.4 51.3 76.9
61 44.3 36.4 54.6





#33

1,2-Dichloroethane-d4

Concen: 154.797 ug/l

RT: 5.964 min Scan# 800

Delta R.T. -0.000 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 65 Resp: 230915

Ion Ratio Lower Upper

65 100





67 52.1 0.0 105.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.769 min Scan# 932

Delta R.T. -0.000 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02




Tgt Ion:114 Resp: 178346


Ion Ratio Lower Upper

114 100

63 21.5 0.0 44.2

88 16.1 0.0 33.2





#35

Dibromofluoromethane

Concen: 158.930 ug/l

RT: 5.397 min Scan# 707

Delta R.T. -0.000 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion:113 Resp: 187530

Ion Ratio Lower Upper

113 100

111 101.9 82.0 123.0





192 19.2 15.3 22.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#36

1,1-Dichloropropene

Concen: 150.653 ug/l

RT: 5.702 min Scan# 757

Delta R.T. -0.000 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02




Tgt Ion: 75 Resp: 256619


Ion Ratio Lower Upper

75 100

110 35.0 17.2 51.5

77 31.0 24.2 36.4





#37

Ethyl Acetate

Concen: 156.284 ug/l

RT: 4.721 min Scan# 596

Delta R.T. -0.000 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 43 Resp: 245724

Ion Ratio Lower Upper

43 100

61 12.3 9.7 14.5





70 10.4 8.1 12.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#38

Carbon Tetrachloride

Concen: 153.668 ug/l

RT: 5.684 min Scan# 754

Delta R.T. -0.006 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02




Tgt Ion:117 Resp: 284066


Ion Ratio Lower Upper

117 100

119 94.6 75.2 112.8

121 30.9 24.1 36.1





#39

Methylcyclohexane

Concen: 154.413 ug/l

RT: 7.385 min Scan# 1033

Delta R.T. -0.000 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 83 Resp: 346118

Ion Ratio Lower Upper

83 100

55 78.6 63.0 94.4





98 48.5 38.8 58.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#40

Benzene

Concen: 151.423 ug/l

RT: 6.044 min Scan# 813

Delta R.T. -0.000 min

Lab File: VX046722.D




Acq: 17 Jun 2025 15:02


Tgt Ion: 78 Resp: 763341

Ion Ratio Lower Upper

78 100

77 23.5 18.8 28.2









#41
Methacrylonitrile
Concen: 169.745 ug/l
RT: 4.928 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150

Tgt Ion: 41 Resp: 14228
Ion Ratio Lower Upper
41 100
39 52.0 45.9 68.9
67 71.3 61.2 91.8
52 29.4 25.5 38.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#42
1,2-Dichloroethane
Concen: 150.011 ug/l
RT: 6.092 min Scan# 821
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion: 62 Resp: 265740
Ion Ratio Lower Upper
62 100
98 9.3 0.0 17.8





#43

Isopropyl Acetate

Concen: 167.483 ug/l

RT: 6.342 min Scan# 862

Delta R.T. -0.000 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC150

Tgt Ion:	43	Resp:	41206
Ion	Ratio	Lower	Upper
43	100		
61	19.6	15.3	22.9
87	12.8	10.0	15.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#44

Trichloroethene

Concen: 152.520 ug/l




RT: 7.129 min Scan# 991


Delta R.T. -0.000 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02

Tgt Ion:	130	Resp:	195943
Ion	Ratio	Lower	Upper
130	100		
95	100.8	0.0	202.6





#45

1,2-Dichloropropane

Concen: 153.025 ug/l

RT: 7.434 min Scan# 1041

Delta R.T. -0.000 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC150

Tgt Ion: 63 Resp: 19118

Ion Ratio Lower Upper

63 100





65 31.4 22.9 34.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#46

Dibromomethane

Concen: 154.297 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. -0.006 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02




Tgt Ion: 93 Resp: 138541


Ion Ratio Lower Upper

93 100

95 82.2 66.2 99.4


174 97.0 77.0 115.6





#47
Bromodichloromethane
Concen: 157.630 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150



Tgt Ion: 83 Resp: 288820
Ion Ratio Lower Upper
83 100
85 64.0 51.4 77.2
127 8.8 6.9 10.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#48
Methyl methacrylate
Concen: 171.734 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion: 41 Resp: 211908
Ion Ratio Lower Upper
41 100
69 80.7 63.4 95.2
39 56.3 45.4 68.0





#49

1,4-Dioxane

Concen: 2987.386 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. 0.000 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC150





Tgt Ion: 88 Resp: 54640

Ion	Ratio	Lower	Upper
88	100		
43	37.6	31.2	46.8
58	71.8	54.7	82.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#50

Toluene-d8

Concen: 155.733 ug/l

RT: 8.653 min Scan# 1241




Delta R.T. 0.006 min


Lab File: VX046722.D

Acq: 17 Jun 2025 15:02

Tgt Ion: 98 Resp: 660221

Ion	Ratio	Lower	Upper
98	100		
100	66.6	53.0	79.4









#51
4-Methyl-2-Pentanone
Concen: 799.825 ug/l
RT: 8.574 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150

Tgt Ion: 43 Resp: 117366
Ion Ratio Lower Upper
43 100
58 38.8 30.8 46.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#52
Toluene
Concen: 150.800 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02


Tgt Ion: 92 Resp: 471273
Ion Ratio Lower Upper
92 100
91 169.7 134.8 202.2





#53
t-1,3-Dichloropropene
Concen: 169.929 ug/l
RT: 8.976 min Scan# 1294
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150



Tgt Ion: 75 Resp: 28889
Ion Ratio Lower Upper
75 100
77 31.5 25.6 38.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#54
cis-1,3-Dichloropropene
Concen: 163.198 ug/l
RT: 8.366 min Scan# 1194
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion: 75 Resp: 314965
Ion Ratio Lower Upper
75 100
77 32.0 25.3 37.9
39 51.4 39.2 58.8





#55

1,1,2-Trichloroethane

Concen: 151.099 ug/l

RT: 9.153 min Scan# 1323

Delta R.T. -0.000 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC150

Tgt Ion: 97 Resp: 17597

Ion Ratio Lower Upper

97 100

83 85.2 67.2 100.8

85 55.5 44.1 66.1





99 62.6 50.3 75.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#56

Ethyl methacrylate

Concen: 167.940 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. -0.000 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02




Tgt Ion: 69 Resp: 288720


Ion Ratio Lower Upper

69 100

41 68.8 55.4 83.2

39 37.3 29.7 44.5





#57

1,3-Dichloropropane

Concen: 150.290 ug/l

RT: 9.305 min Scan# 11348

Delta R.T. -0.000 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 76 Resp: 301418





Ion Ratio Lower Upper

76 100

78 32.4 25.7 38.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025

#58

2-Chloroethyl Vinyl ether

Concen: 845.118 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. -0.000 min

Lab File: VX046722.D




Acq: 17 Jun 2025 15:02


Tgt Ion: 63 Resp: 754730

Ion Ratio Lower Upper

63 100

106 28.4 22.6 34.0





#59

2-Hexanone

Concen: 805.005 ug/l

RT: 9.433 min Scan# 1369

Delta R.T. 0.006 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 43 Resp: 81665

Ion Ratio Lower Upper

43

100

58

54.1

26.8





80.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#60

Dibromochloromethane

Concen: 156.904 ug/l

RT: 9.519 min Scan# 1383

Delta R.T. -0.000 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02

Tgt Ion:129 Resp: 215347

Ion Ratio Lower Upper

129




100


127

77.3

38.4


115.2





#61
1,2-Dibromoethane
Concen: 156.392 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150



Tgt Ion:107 Resp: 185268
Ion Ratio Lower Upper
107 100
109 93.6 75.8 113.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#62
4-Bromofluorobenzene
Concen: 154.404 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02


Tgt Ion: 95 Resp: 244007
Ion Ratio Lower Upper
95 100
174 75.6 0.0 150.4
176 72.7 0.0 145.0





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150



Tgt Ion:117 Resp: 156970
Ion Ratio Lower Upper
117 100
82 56.4 44.6 66.8
119 33.1 25.8 38.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#64
Tetrachloroethene
Concen: 146.895 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion:164 Resp: 159752
Ion Ratio Lower Upper
164 100
166 125.9 104.6 157.0
129 97.1 76.3 114.5
131 91.9 74.1 111.1









#65
Chlorobenzene
Concen: 149.453 ug/l
RT: 10.080 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Tgt Ion:112 Resp: 51785
Ion Ratio Lower Upper
112 100
114 31.7 25.5 38.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 156.268 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. 0.006 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion:131 Resp: 181981
Ion Ratio Lower Upper
131 100
133 95.5 47.3 141.8
119 64.7 32.1 96.3





#67

Ethyl Benzene

Concen: 151.251 ug/l

RT: 10.195 min Scan# 1493

Delta R.T. 0.006 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 91 Resp: 915724

Ion Ratio Lower Upper

91 100





106 31.0 24.2 36.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#68

m/p-Xylenes

Concen: 299.697 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. -0.000 min

Lab File: VX046722.D




Acq: 17 Jun 2025 15:02


Tgt Ion:106 Resp: 676085

Ion Ratio Lower Upper

106 100

91 208.3 164.8 247.2





#69

o-Xylene

Concen: 155.514 ug/l

RT: 10.640 min Scan# 1567

Delta R.T. -0.000 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion:106 Resp: 32878

Ion Ratio Lower Upper

106 100





91 220.1 111.3 333.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#70

Styrene

Concen: 153.143 ug/l

RT: 10.653 min Scan# 1569

Delta R.T. -0.000 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02




Tgt Ion:104 Resp: 566606


Ion Ratio Lower Upper

104 100

78 52.2 41.4 62.2


103 54.1 43.1 64.7





#71
Bromoform
Concen: 161.498 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150



Tgt Ion:173 Resp: 143229
Ion Ratio Lower Upper
173 100
175 48.6 24.1 72.4
254 0.1 0.1 0.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion:152 Resp: 75829
Ion Ratio Lower Upper
152 100
115 139.0 43.2 129.6#
150 208.8 0.0 346.8









#73
Isopropylbenzene
Concen: 155.374 ug/l
RT: 10.964 min Scan# 1620
Delta R.T. 0.006 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150

Tgt Ion:105 Resp: 867728
Ion Ratio Lower Upper
105 100
120 26.0 13.0 39.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#74
N-ethyl acetate
Concen: 174.748 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02


Tgt Ion: 43 Resp: 369757
Ion Ratio Lower Upper
43 100
70 41.3 33.4 50.0
55 23.3 19.1 28.7
61 23.9 19.0 28.4





#75
1,1,2,2-Tetrachloroethane
Concen: 156.730 ug/l
RT: 11.207 min Scan# 1660
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150



Tgt Ion: 83 Resp: 244400
Ion Ratio Lower Upper
83 100
131 10.1 4.9 14.6
85 64.4 32.0 96.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#76
1,2,3-Trichloropropane
Concen: 131.753 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02


Tgt Ion: 75 Resp: 202926
Ion Ratio Lower Upper
75 100
77 43.5 18.5 55.5





#77
Bromobenzene
Concen: 154.285 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150



Tgt Ion:156 Resp: 207348
Ion Ratio Lower Upper
156 100
77 167.9 82.8 248.6
158 96.7 47.6 142.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#78
n-propylbenzene
Concen: 154.633 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. 0.006 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion: 91 Resp: 1025742
Ion Ratio Lower Upper
91 100
120 22.6 11.4 34.1





#79

2-Chlorotoluene

Concen: 150.373 ug/l

RT: 11.366 min Scan# 1

Delta R.T. 0.006 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 91 Resp: 59555

Ion Ratio Lower Upper

91 100





126 33.4 16.6 49.7

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#80

1,3,5-Trimethylbenzene

Concen: 154.053 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. -0.000 min

Lab File: VX046722.D




Acq: 17 Jun 2025 15:02


Tgt Ion:105 Resp: 707005

Ion Ratio Lower Upper

105 100

120 48.5 24.1 72.2





#81

trans-1,4-Dichloro-2-butene

Concen: 176.408 ug/l

RT: 11.018 min Scan# 1629

Delta R.T. -0.000 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC150

Tgt Ion:	75	Resp:	8667
Ion Ratio	Lower	Upper	
75	100		
53	98.9	86.8	130.2
89	45.8	36.9	55.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#82

4-Chlorotoluene

Concen: 150.915 ug/l




RT: 11.451 min Scan# 1700


Delta R.T. -0.000 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02

Tgt Ion:	91	Resp:	697556
Ion Ratio	Lower	Upper	
91	100		
126	29.1	14.4	43.2





#83

tert-Butylbenzene

Concen: 155.093 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. -0.000 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion:119 Resp: 735750

Ion Ratio Lower Upper

119 100

91 59.0 29.5 88.6





134 22.8 11.7 35.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#84

1,2,4-Trimethylbenzene

Concen: 155.994 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. -0.000 min

Lab File: VX046722.D




Acq: 17 Jun 2025 15:02


Tgt Ion:105 Resp: 719283

Ion Ratio Lower Upper

105 100

120 44.4 22.1 66.3









#85
sec-Butylbenzene
Concen: 152.910 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Tgt Ion:105 Resp: 916769
Ion Ratio Lower Upper
105 100
134 19.9 9.8 29.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#86
p-Isopropyltoluene
Concen: 152.610 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion:119 Resp: 771998
Ion Ratio Lower Upper
119 100
134 25.9 12.9 38.7
91 24.4 12.3 36.9





#87

1,3-Dichlorobenzene

Concen: 149.232 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion:146 Resp: 391049

Ion Ratio Lower Upper

146 100

111 41.3 20.6 61.8





148 63.2 31.6 94.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#88

1,4-Dichlorobenzene

Concen: 143.836 ug/l

RT: 12.036 min Scan# 1796

Delta R.T. -0.000 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02




Tgt Ion:146 Resp: 387139


Ion Ratio Lower Upper

146 100

111 40.3 20.3 61.0

148 63.5 32.1 96.5





#89

n-Butylbenzene

Concen: 151.271 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. -0.000 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 91 Resp: 72656

Ion Ratio Lower Upper

91 100

92 54.7 27.1 81.3





134 25.6 12.9 38.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#90

Hexachloroethane

Concen: 161.049 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX046722.D




Acq: 17 Jun 2025 15:02


Tgt Ion:117 Resp: 142663

Ion Ratio Lower Upper

117 100


201 70.7 34.4 103.3





#91
1,2-Dichlorobenzene
Concen: 151.192 ug/l
RT: 12.329 min Scan# 1844
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150



Tgt Ion:146 Resp: 370234
Ion Ratio Lower Upper
146 100
111 43.3 21.7 65.1
148 63.8 32.0 96.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 175.448 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion: 75 Resp: 54014
Ion Ratio Lower Upper
75 100
155 84.2 41.9 125.7
157 108.5 51.4 154.3





#93

1,2,4-Trichlorobenzene

Concen: 151.531 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. -0.000 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion:180 Resp: 26380

Ion Ratio Lower Upper

180 100

182 96.9 46.7 140.1





145 32.2 15.8 47.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#94

Hexachlorobutadiene

Concen: 130.431 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. 0.006 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02




Tgt Ion:225 Resp: 95520


Ion Ratio Lower Upper

225 100

223 64.5 32.5 97.5


227 64.2 32.2 96.6





#95
Naphthalene
Concen: 170.609 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150



Tgt Ion:128 Resp: 846200
Ion Ratio Lower Upper
128 100
127 12.7 10.1 15.1
129 10.9 8.7 13.1




Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#96
1,2,3-Trichlorobenzene
Concen: 157.517 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion:180 Resp: 262348
Ion Ratio Lower Upper
180 100
182 93.7 47.8 143.4
145 33.8 17.3 51.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
 Data File : VX046725.D
 Acq On : 17 Jun 2025 17:18
 Operator : JC/MD
 Sample : VSTDICC001
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC001

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
 Supervised By :Mahesh Dadoda 06/18/2025

Quant Time: Jun 18 02:42:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 02:34:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	136219	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	226094	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	205370	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	103593	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	0.000	65	0d	0.000	ug/l	
Spiked Amount	50.000	Range	74 - 125	Recovery	=	0.000%#
35) Dibromofluoromethane	0.000	113	0d	0.000	ug/l	
Spiked Amount	50.000	Range	75 - 124	Recovery	=	0.000%#
50) Toluene-d8	0.000	98	0d	0.000	ug/l	
Spiked Amount	50.000	Range	86 - 113	Recovery	=	0.000%#
62) 4-Bromofluorobenzene	0.000	95	0d	0.000	ug/l	
Spiked Amount	50.000	Range	77 - 121	Recovery	=	0.000%#

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.179	85	1159	0.797	ug/l	83
3) Chloromethane	1.307	50	1322	0.842	ug/l	93
4) Vinyl Chloride	1.386	62	1419	0.848	ug/l	89
6) Chloroethane	1.703	64	905	0.892	ug/l	90
7) Trichlorofluoromethane	1.904	101	2062	0.811	ug/l	97
8) Diethyl Ether	2.148	74	722	0.828	ug/l	87
9) 1,1,2-Trichlorotrifluo...	2.343	101	1280	0.821	ug/l	99
12) 1,1-Dichloroethene	2.337	96	1228	0.816	ug/l #	76
14) Allyl chloride	2.672	41	2386	0.875	ug/l	96
15) Acrylonitrile	3.081	53	3075	4.044	ug/l	96
16) Acetone	2.386	43	3424	5.449	ug/l #	80
17) Carbon Disulfide	2.526	76	5092	1.121	ug/l	96
18) Methyl Acetate	2.721	43	1280	0.802	ug/l #	87
19) Methyl tert-butyl Ether	3.129	73	3888	0.830	ug/l #	89
20) Methylene Chloride	2.800	84	1736	0.994	ug/l	91
21) trans-1,2-Dichloroethene	3.111	96	1441	0.895	ug/l #	81
22) Diisopropyl ether	3.770	45	4170	0.796	ug/l #	71
23) Vinyl Acetate	3.745	43	15507	3.722	ug/l	95
24) 1,1-Dichloroethane	3.617	63	2648	0.890	ug/l #	79
25) 2-Butanone	4.593	43	3422	3.851	ug/l	93
26) 2,2-Dichloropropane	4.489	77	1874m	0.864	ug/l	
27) cis-1,2-Dichloroethene	4.507	96	1698	0.898	ug/l	97
28) Bromochloromethane	4.922	49	1307	0.969	ug/l #	91
29) Tetrahydrofuran	5.032	42	2212	3.854	ug/l #	54
30) Chloroform	5.105	83	2336m	0.785	ug/l	
32) 1,1,1-Trichloroethane	5.391	97	2212	0.848	ug/l #	50
36) 1,1-Dichloropropene	5.708	75	2016	0.934	ug/l	90
37) Ethyl Acetate	4.739	43	2002m	1.004	ug/l	
38) Carbon Tetrachloride	5.684	117	2124	0.906	ug/l	92
39) Methylcyclohexane	7.379	83	2485	0.875	ug/l	87
40) Benzene	6.044	78	5507	0.862	ug/l	97
41) Methacrylonitrile	4.965	41	857m	0.806	ug/l	
42) 1,2-Dichloroethane	6.098	62	1938	0.863	ug/l #	75
43) Isopropyl Acetate	6.367	43	2218	0.711	ug/l #	83
44) Trichloroethene	7.135	130	1448	0.889	ug/l	87
45) 1,2-Dichloropropane	7.434	63	1380	0.871	ug/l #	89

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
 Data File : VX046725.D
 Acq On : 17 Jun 2025 17:18
 Operator : JC/MD
 Sample : VSTDICC001
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC001

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
 Supervised By :Mahesh Dadoda 06/18/2025

Quant Time: Jun 18 02:42:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 02:34:13 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) Dibromomethane	7.592	93	1030	0.905	ug/l	89
47) Bromodichloromethane	7.824	83	1825	0.786	ug/l #	82
48) Methyl methacrylate	7.720	41	1050	0.671	ug/l #	80
49) 1,4-Dioxane	7.671	88	56	28.694	ug/l #	77
51) 4-Methyl-2-Pentanone	8.580	43	7273	3.910	ug/l	97
52) Toluene	8.726	92	3393	0.856	ug/l	96
53) t-1,3-Dichloropropene	8.994	75	1607	0.746	ug/l #	87
54) cis-1,3-Dichloropropene	8.372	75	1971	0.806	ug/l #	57
55) 1,1,2-Trichloroethane	9.153	97	1297	0.878	ug/l #	83
56) Ethyl methacrylate	9.122	69	1477	0.678	ug/l	93
57) 1,3-Dichloropropane	9.311	76	2249	0.885	ug/l	95
58) 2-Chloroethyl Vinyl ether	8.251	63	3778	3.337	ug/l	92
59) 2-Hexanone	9.439	43	4833	3.758	ug/l	94
60) Dibromochloromethane	9.525	129	1436	0.825	ug/l	99
61) 1,2-Dibromoethane	9.610	107	1209	0.805	ug/l	90
64) Tetrachloroethene	9.275	164	1399	0.983	ug/l	95
65) Chlorobenzene	10.073	112	4127	0.910	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	1242	0.815	ug/l #	64
67) Ethyl Benzene	10.195	91	6967	0.880	ug/l	98
68) m/p-Xylenes	10.305	106	5214	1.767	ug/l	98
69) o-Xylene	10.646	106	2045	0.739	ug/l	80
70) Styrene	10.659	104	4077	0.842	ug/l	98
71) Bromoform	10.799	173	927	0.799	ug/l #	96
73) Isopropylbenzene	10.963	105	6314	0.828	ug/l	97
74) N-amyl acetate	10.848	43	1921	0.665	ug/l #	85
75) 1,1,2,2-Tetrachloroethane	11.213	83	1690	0.793	ug/l	93
76) 1,2,3-Trichloropropane	11.238	75	1346m	0.640	ug/l	
77) Bromobenzene	11.201	156	1541	0.839	ug/l	95
78) n-propylbenzene	11.305	91	7545	0.833	ug/l	98
79) 2-Chlorotoluene	11.366	91	4896	0.905	ug/l	93
80) 1,3,5-Trimethylbenzene	11.451	105	5174	0.825	ug/l	94
82) 4-Chlorotoluene	11.457	91	5664	0.897	ug/l	99
83) tert-Butylbenzene	11.713	119	5507	0.850	ug/l	95
84) 1,2,4-Trimethylbenzene	11.750	105	5087	0.808	ug/l	99
85) sec-Butylbenzene	11.890	105	6839	0.835	ug/l	98
86) p-Isopropyltoluene	12.006	119	6018	0.871	ug/l	92
87) 1,3-Dichlorobenzene	11.969	146	3509	0.980	ug/l	96
88) 1,4-Dichlorobenzene	12.036	146	4003m	1.089	ug/l	
89) n-Butylbenzene	12.335	91	6186	0.943	ug/l	96
90) Hexachloroethane	12.536	117	1081	0.893	ug/l	90
91) 1,2-Dichlorobenzene	12.335	146	3024	0.904	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.951	75	277	0.659	ug/l	95
93) 1,2,4-Trichlorobenzene	13.591	180	2424	1.019	ug/l	99
94) Hexachlorobutadiene	13.725	225	1025	1.025	ug/l	97
95) Naphthalene	13.780	128	5461	0.806	ug/l	96
96) 1,2,3-Trichlorobenzene	13.963	180	1982	0.871	ug/l	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Tgt Ion:168 Resp: 136219
Ion Ratio Lower Upper
168 100
99 59.2 48.5 72.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#2
Dichlorodifluoromethane
Concen: 0.797 ug/l
RT: 1.179 min Scan# 15
Delta R.T. -0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Tgt Ion: 85 Resp: 1159
Ion Ratio Lower Upper
85 100
87 23.5 16.6 49.8









#3
Chloromethane
Concen: 0.842 ug/l
RT: 1.307 min Scan# 36
Delta R.T. -0.000 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Tgt Ion: 50 Resp: 132
Ion Ratio Lower Upper
50 100
52 35.2 24.9 37.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#4
Vinyl Chloride
Concen: 0.848 ug/l
RT: 1.386 min Scan# 49
Delta R.T. -0.000 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18


Tgt Ion: 62 Resp: 1419
Ion Ratio Lower Upper
62 100
64 35.7 23.9 35.9





#6
Chloroethane
Concen: 0.892 ug/l
RT: 1.703 min Scan# 101
Delta R.T. -0.000 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001



Tgt Ion: 64 Resp: 909
Ion Ratio Lower Upper
64 100
66 28.0 27.1 40.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#7
Trichlorofluoromethane
Concen: 0.811 ug/l
RT: 1.904 min Scan# 134
Delta R.T. -0.000 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Tgt Ion:101 Resp: 2062
Ion Ratio Lower Upper
101 100
103 67.4 51.8 77.8





#8

Diethyl Ether

Concen: 0.828 ug/l

RT: 2.148 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion: 74 Resp: 72

Ion Ratio Lower Upper

74 100





45 118.6 52.3 157.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#9

1,1,2-Trichlorotrifluoroethane

Concen: 0.821 ug/l

RT: 2.343 min Scan# 206

Delta R.T. -0.000 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18




Tgt Ion:101 Resp: 1280


Ion Ratio Lower Upper

101 100

85 45.9 35.4 53.0

151 76.5 61.2 91.8





#12

1,1-Dichloroethene

Concen: 0.816 ug/l

RT: 2.337 min Scan# 205

Delta R.T. -0.000 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion: 96 Resp: 1228

Ion Ratio Lower Upper

96 100

61 200.9 126.6 189.8





98 61.6 50.7 76.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#14

Allyl chloride

Concen: 0.875 ug/l

RT: 2.672 min Scan# 260

Delta R.T. -0.006 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18




Tgt Ion: 41 Resp: 2386


Ion Ratio Lower Upper

41 100

39 64.2 53.6 80.4

76 31.4 26.8 40.2









#15
Acrylonitrile
Concen: 4.044 ug/l
RT: 3.081 min Scan# 31
Delta R.T. 0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Tgt Ion: 53 Resp: 3075
Ion Ratio Lower Upper
53 100
52 79.2 65.1 97.7
51 31.7 28.9 43.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#16
Acetone
Concen: 5.449 ug/l
RT: 2.386 min Scan# 213
Delta R.T. -0.000 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Tgt Ion: 43 Resp: 3424
Ion Ratio Lower Upper
43 100
58 19.6 24.5 36.7#





#17

Carbon Disulfide

Concen: 1.121 ug/l

RT: 2.526 min Scan# 21

Delta R.T. -0.000 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion: 76 Resp: 5092

Ion Ratio Lower Upper

76 100





78 8.1 7.5 11.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#18

Methyl Acetate

Concen: 0.802 ug/l

RT: 2.721 min Scan# 268

Delta R.T. 0.006 min

Lab File: VX046725.D




Acq: 17 Jun 2025 17:18


Tgt Ion: 43 Resp: 1280

Ion Ratio Lower Upper

43 100

74 16.9 18.4 27.6#









#19
Methyl tert-butyl Ether
Concen: 0.830 ug/l
RT: 3.129 min Scan# 31
Delta R.T. 0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001

Tgt Ion: 73 Resp: 388
Ion Ratio Lower Upper
73 100
57 17.3 18.2 27.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#20
Methylene Chloride
Concen: 0.994 ug/l
RT: 2.800 min Scan# 281
Delta R.T. -0.000 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Tgt Ion: 84 Resp: 1736
Ion Ratio Lower Upper
84 100
49 142.6 101.3 151.9
51 38.8 31.0 46.6
86 66.8 51.1 76.7









#21
trans-1,2-Dichloroethene
Concen: 0.895 ug/l
RT: 3.111 min Scan# 311
Delta R.T. 0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Tgt Ion: 96 Resp: 144
Ion Ratio Lower Upper
96 100
61 118.7 112.5 168.7
98 47.7 51.9 77.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#22
Diisopropyl ether
Concen: 0.796 ug/l
RT: 3.770 min Scan# 440
Delta R.T. -0.000 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Tgt Ion: 45 Resp: 4170
Ion Ratio Lower Upper
45 100
43 155.2 91.9 137.9#
87 24.5 22.0 33.0
59 12.4 9.1 13.7





#23

Vinyl Acetate

Concen: 3.722 ug/l

RT: 3.745 min Scan# 41

Delta R.T. 0.012 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC001

Tgt Ion: 43 Resp: 1550

Ion Ratio Lower Upper

43 100





86 12.0 8.2 12.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#24

1,1-Dichloroethane

Concen: 0.890 ug/l

RT: 3.617 min Scan# 415

Delta R.T. -0.006 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18




Tgt Ion: 63 Resp: 2648


Ion Ratio Lower Upper

63 100

98 0.0 4.0 11.9#

100 10.4 2.2 6.6#





#25

2-Butanone

Concen: 3.851 ug/l

RT: 4.593 min Scan# 51

Delta R.T. 0.030 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC001

Tgt Ion:	43	Resp:	342
Ion	Ratio	Lower	Upper
43	100		
72	22.0	20.2	30.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt	06/18/2025
Supervised By :Mahesh Dadoda	06/18/2025



#26

2,2-Dichloropropane

Concen: 0.864 ug/l m




RT: 4.489 min Scan# 558


Delta R.T. -0.000 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Tgt Ion:	77	Resp:	1874
Ion	Ratio	Lower	Upper
77	100		
97	9.8	11.8	35.3#





#27

cis-1,2-Dichloroethene

Concen: 0.898 ug/l

RT: 4.507 min Scan# 560

Delta R.T. 0.006 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion: 96 Resp: 1698

Ion Ratio Lower Upper

96 100

61 143.8 0.0 291.0





98 69.4 0.0 126.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#28

Bromochloromethane

Concen: 0.969 ug/l

RT: 4.922 min Scan# 629

Delta R.T. 0.012 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18




Tgt Ion: 49 Resp: 1307


Ion Ratio Lower Upper

49 100

129 0.0 0.0 3.8

130 68.6 61.4 92.0





#29

Tetrahydrofuran

Concen: 3.854 ug/l

RT: 5.032 min Scan# 644

Delta R.T. 0.018 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC001

Tgt Ion:	42	Resp:	221
Ion	Ratio	Lower	Upper
42	100		
72	0.0	34.9	52.3
71	26.1	32.2	48.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#30

Chloroform

Concen: 0.785 ug/l m




RT: 5.105 min Scan# 659


Delta R.T. -0.000 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Tgt Ion:	83	Resp:	2336
Ion	Ratio	Lower	Upper
83	100		
85	52.8	53.7	80.5





#32

1,1,1-Trichloroethane

Concen: 0.848 ug/l

RT: 5.391 min Scan# 70

Delta R.T. -0.006 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC001

Tgt Ion: 97 Resp: 221

Ion Ratio Lower Upper

97 100

99 0.0 51.3 76.9





61 41.9 36.4 54.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#33

1,2-Dichloroethane-d4

Concen: 0.000 ug/l

Expected RT: 5.96 min

Lab File: VX046725.D


Acq: 17 Jun 2025 17:18

Tgt Ion: 65


Sig Exp Ratio

65 100

67 52.7



Ion 66.90 (66.60 to 67.60): VX046725.D\data.ms



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 91

Delta R.T. -0.006 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC001

Tgt Ion:114 Resp: 226094

Ion Ratio Lower Upper

114 100

63 21.2 0.0 44.2





88 16.3 0.0 33.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#35

Dibromofluoromethane

Concen: 0.000 ug/l

Expected RT: 5.40 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18


Tgt Ion: 113


Sig Exp Ratio

113 100

111 102.5

192 19.1





#36

1,1-Dichloropropene

Concen: 0.934 ug/l

RT: 5.708 min Scan# 758

Delta R.T. 0.006 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC001

Tgt Ion:	75	Resp:	2010
Ion	Ratio	Lower	Upper
75	100		
110	27.6	17.2	51.5
77	25.5	24.2	36.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#37

Ethyl Acetate

Concen: 1.004 ug/l m




RT: 4.739 min Scan# 599


Delta R.T. 0.018 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Tgt Ion:	43	Resp:	2002
Ion	Ratio	Lower	Upper
43	100		
61	0.0	9.7	14.5#
70	0.0	8.1	12.1#





#38

Carbon Tetrachloride

Concen: 0.906 ug/l

RT: 5.684 min Scan# 754

Delta R.T. -0.006 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion:117 Resp: 2124

Ion Ratio Lower Upper

117 100

119 102.3 75.2 112.8





121 32.4 24.1 36.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#39

Methylcyclohexane

Concen: 0.875 ug/l

RT: 7.379 min Scan# 1032

Delta R.T. -0.006 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18




Tgt Ion: 83 Resp: 2485


Ion Ratio Lower Upper

83 100

55 91.5 63.0 94.4

98 55.1 38.8 58.2





#40

Benzene

Concen: 0.862 ug/l

RT: 6.044 min Scan# 813

Delta R.T. -0.000 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC001

Tgt Ion: 78 Resp: 550

Ion Ratio Lower Upper

78 100





77 22.0 18.8 28.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#41

Methacrylonitrile

Concen: 0.806 ug/l m

RT: 4.965 min Scan# 636

Delta R.T. 0.036 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Tgt Ion: 41 Resp: 857




Ion Ratio Lower Upper


41 100

39 0.0 45.9 68.9#

67 20.4 61.2 91.8#

52 0.0 25.5 38.3#





#42

1,2-Dichloroethane

Concen: 0.863 ug/l

RT: 6.098 min Scan# 81

Delta R.T. 0.006 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion: 62 Resp: 1938

Ion Ratio Lower Upper

62 100





98 0.0 0.0 17.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#43

Isopropyl Acetate

Concen: 0.711 ug/l

RT: 6.367 min Scan# 866

Delta R.T. 0.024 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18




Tgt Ion: 43 Resp: 2218


Ion Ratio Lower Upper

43 100

61 9.3 15.3 22.9#


87 8.6 10.0 15.0#





#44
Trichloroethene
Concen: 0.889 ug/l
RT: 7.135 min Scan# 991
Delta R.T. 0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001



Tgt Ion:130 Resp: 1448
Ion Ratio Lower Upper
130 100
95 114.0 0.0 202.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#45
1,2-Dichloropropane
Concen: 0.871 ug/l
RT: 7.434 min Scan# 1041
Delta R.T. -0.000 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Tgt Ion: 63 Resp: 1380
Ion Ratio Lower Upper
63 100
65 34.6 22.9 34.3#





#46

Dibromomethane

Concen: 0.905 ug/l

RT: 7.592 min Scan# 1066

Delta R.T. 0.006 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion: 93 Resp: 1030

Ion Ratio Lower Upper

93 100

95 71.1 66.2 99.4





174 86.7 77.0 115.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#47

Bromodichloromethane

Concen: 0.786 ug/l

RT: 7.824 min Scan# 1105

Delta R.T. -0.000 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18




Tgt Ion: 83 Resp: 1825


Ion Ratio Lower Upper

83 100

85 80.3 51.4 77.2#

127 7.7 6.9 10.3





#48

Methyl methacrylate

Concen: 0.671 ug/l

RT: 7.720 min Scan# 1088

Delta R.T. 0.024 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC001

Tgt Ion: 41	Resp: 1050
Ion Ratio	Lower Upper
41 100	
69 71.8	63.4 95.2
39 30.3	45.4 68.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#49

1,4-Dioxane

Concen: 28.694 ug/l




RT: 7.671 min Scan# 1080


Delta R.T. 0.012 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Tgt Ion: 88	Resp: 56
Ion Ratio	Lower Upper
88 100	
43 75.0	31.2 46.8
58 66.1	54.7 82.1





#50
Toluene-d8
Concen: 0.000 ug/l
Expected RT: 8.65 min



Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Tgt Ion: 98
Sig Exp Ratio
98 100
100 66.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#51
4-Methyl-2-Pentanone
Concen: 3.910 ug/l
RT: 8.580 min Scan# 1229
Delta R.T. 0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Tgt Ion: 43 Resp: 7273
Ion Ratio Lower Upper
43 100
58 36.9 30.8 46.2





#52

Toluene

Concen: 0.856 ug/l

RT: 8.726 min Scan# 1252

Delta R.T. 0.006 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion: 92 Resp: 339

Ion Ratio Lower Upper

92 100





91 163.1 134.8 202.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#53

t-1,3-Dichloropropene

Concen: 0.746 ug/l

RT: 8.994 min Scan# 1297

Delta R.T. 0.018 min

Lab File: VX046725.D




Acq: 17 Jun 2025 17:18


Tgt Ion: 75 Resp: 1607

Ion Ratio Lower Upper

75 100

77 39.3 25.6 38.4#





#54

cis-1,3-Dichloropropene

Concen: 0.806 ug/l

RT: 8.372 min Scan# 1194

Delta R.T. 0.006 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC001

Tgt Ion:	75	Resp:	197
Ion	Ratio	Lower	Upper
75	100		
77	30.3	25.3	37.9
39	96.1	39.2	58.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#55

1,1,2-Trichloroethane

Concen: 0.878 ug/l




RT: 9.153 min Scan# 1323


Delta R.T. -0.000 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Tgt Ion:	97	Resp:	1297
Ion	Ratio	Lower	Upper
97	100		
83	66.9	67.2	100.8#
85	42.3	44.1	66.1#
99	53.1	50.3	75.5









#56
Ethyl methacrylate
Concen: 0.678 ug/l
RT: 9.122 min Scan# 1317
Delta R.T. 0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Tgt Ion: 69 Resp: 147
Ion Ratio Lower Upper
69 100
41 63.2 55.4 83.2
39 41.0 29.7 44.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#57
1,3-Dichloropropane
Concen: 0.885 ug/l
RT: 9.311 min Scan# 1349
Delta R.T. 0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Tgt Ion: 76 Resp: 2249
Ion Ratio Lower Upper
76 100
78 29.5 25.7 38.5





#58

2-Chloroethyl Vinyl ether

Concen: 3.337 ug/l

RT: 8.251 min Scan# 1175

Delta R.T. 0.006 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion: 63 Resp: 3778

Ion Ratio Lower Upper

63 100





106 24.2 22.6 34.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#59

2-Hexanone

Concen: 3.758 ug/l

RT: 9.439 min Scan# 1370

Delta R.T. 0.012 min

Lab File: VX046725.D




Acq: 17 Jun 2025 17:18


Tgt Ion: 43 Resp: 4833

Ion Ratio Lower Upper

43 100


58 49.0 26.8 80.4





#60
Dibromochloromethane
Concen: 0.825 ug/l
RT: 9.525 min Scan# 1384
Delta R.T. 0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001



Tgt Ion:129 Resp: 1430
Ion Ratio Lower Upper
129 100
127 75.8 38.4 115.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#61
1,2-Dibromoethane
Concen: 0.805 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. -0.000 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Tgt Ion:107 Resp: 1209
Ion Ratio Lower Upper
107 100
109 104.6 75.8 113.6





#62
4-Bromofluorobenzene
Concen: 0.000 ug/l
Expected RT: 11.08 min



Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Tgt Ion: 95
Sig Exp Ratio
95 100
174 75.2
176 72.5

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. -0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18


Tgt Ion:117 Resp: 205370
Ion Ratio Lower Upper
117 100
82 58.0 44.6 66.8
119 32.1 25.8 38.8





#64
Tetrachloroethene
Concen: 0.983 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001



Tgt Ion:164 Resp: 1399
Ion Ratio Lower Upper
164 100
166 119.7 104.6 157.0
129 96.4 76.3 114.5
131 95.9 74.1 111.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#65
Chlorobenzene
Concen: 0.910 ug/l
RT: 10.073 min Scan# 1474
Delta R.T. -0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18


Tgt Ion:112 Resp: 4127
Ion Ratio Lower Upper
112 100
114 31.3 25.5 38.3





#66
1,1,1,2-Tetrachloroethane
Concen: 0.815 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001



Tgt Ion:131 Resp: 124
Ion Ratio Lower Upper
131 100
133 89.0 47.3 141.8
119 0.0 32.1 96.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#67
Ethyl Benzene
Concen: 0.880 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Tgt Ion: 91 Resp: 6967
Ion Ratio Lower Upper
91 100
106 29.3 24.2 36.2





#68

m/p-Xylenes

Concen: 1.767 ug/l

RT: 10.305 min Scan# 1511

Delta R.T. 0.006 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion:106 Resp: 5214

Ion Ratio Lower Upper

106 100





91 202.7 164.8 247.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#69

o-Xylene

Concen: 0.739 ug/l

RT: 10.646 min Scan# 1568

Delta R.T. 0.006 min

Lab File: VX046725.D




Acq: 17 Jun 2025 17:18


Tgt Ion:106 Resp: 2045

Ion Ratio Lower Upper

106 100


91 255.0 111.3 333.9





#70
Styrene
Concen: 0.842 ug/l
RT: 10.659 min Scan# 1569
Delta R.T. 0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001



Tgt Ion:104 Resp: 407
Ion Ratio Lower Upper
104 100
78 50.7 41.4 62.2
103 52.8 43.1 64.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#71
Bromoform
Concen: 0.799 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18


Tgt Ion:173 Resp: 927
Ion Ratio Lower Upper
173 100
175 45.4 24.1 72.4
254 0.0 0.1 0.1#





#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001




Tgt Ion:152 Resp: 10359
Ion Ratio Lower Upper
152 100
115 61.3 43.2 129.6
150 159.2 0.0 346.8

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025




#73
Isopropylbenzene
Concen: 0.828 ug/l
RT: 10.963 min Scan# 1620
Delta R.T. 0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18




Tgt Ion:105 Resp: 6314
Ion Ratio Lower Upper
105 100
120 27.6 13.0 39.0





#74
N-amy1 acetate
Concen: 0.665 ug/l
RT: 10.848 min Scan# 1601
Delta R.T. 0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001



Tgt Ion: 43 Resp: 192.0
Ion Ratio Lower Upper
43 100
70 38.4 33.4 50.0
55 20.1 19.1 28.7
61 5.0 19.0 28.4#




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#75
1,1,2,2-Tetrachloroethane
Concen: 0.793 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. 0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Tgt Ion: 83 Resp: 1690
Ion Ratio Lower Upper
83 100
131 10.8 4.9 14.6
85 70.3 32.0 96.0









#76
1,2,3-Trichloropropane
Concen: 0.640 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001

Tgt Ion: 75 Resp: 1340
Ion Ratio Lower Upper
75 100
77 30.3 18.5 55.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#77
Bromobenzene
Concen: 0.839 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Tgt Ion:156 Resp: 1541
Ion Ratio Lower Upper
156 100
77 169.3 82.8 248.6
158 103.2 47.6 142.9









#78
n-propylbenzene
Concen: 0.833 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. 0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Tgt Ion: 91 Resp: 7549
Ion Ratio Lower Upper
91 100
120 23.6 11.4 34.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#79
2-Chlorotoluene
Concen: 0.905 ug/l
RT: 11.366 min Scan# 1686
Delta R.T. 0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Tgt Ion: 91 Resp: 4896
Ion Ratio Lower Upper
91 100
126 28.9 16.6 49.7





#80

1,3,5-Trimethylbenzene

Concen: 0.825 ug/l

RT: 11.451 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion:105 Resp: 5174

Ion Ratio Lower Upper

105 100





120 44.2 24.1 72.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#82

4-Chlorotoluene

Concen: 0.897 ug/l

RT: 11.457 min Scan# 1701

Delta R.T. 0.006 min

Lab File: VX046725.D




Acq: 17 Jun 2025 17:18


Tgt Ion: 91 Resp: 5664

Ion Ratio Lower Upper

91 100

126 29.2 14.4 43.2





#83

tert-Butylbenzene

Concen: 0.850 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. -0.000 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC001

Tgt Ion:	119	Resp:	550
Ion	Ratio	Lower	Upper
119	100		
91	55.1	29.5	88.6
134	21.6	11.7	35.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#84

1,2,4-Trimethylbenzene

Concen: 0.808 ug/l




RT: 11.750 min Scan# 1749


Delta R.T. -0.000 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Tgt Ion:	105	Resp:	5087
Ion	Ratio	Lower	Upper
105	100		
120	43.6	22.1	66.3





#85

sec-Butylbenzene

Concen: 0.835 ug/l

RT: 11.890 min Scan# 1772

Delta R.T. -0.000 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC001

Tgt Ion:105 Resp: 6839

Ion Ratio Lower Upper

105 100





134 20.3 9.8 29.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#86

p-Isopropyltoluene

Concen: 0.871 ug/l

RT: 12.006 min Scan# 1791

Delta R.T. -0.000 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18




Tgt Ion:119 Resp: 6018


Ion Ratio Lower Upper

119 100

134 24.5 12.9 38.7


91 31.1 12.3 36.9





#87
1,3-Dichlorobenzene
Concen: 0.980 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. -0.000 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001




Tgt Ion:146 Resp: 3509
Ion Ratio Lower Upper
146 100
111 40.8 20.6 61.8
148 58.8 31.6 94.8

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025




#88
1,4-Dichlorobenzene
Concen: 1.089 ug/l m
RT: 12.036 min Scan# 1796
Delta R.T. -0.000 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18



Tgt Ion:146 Resp: 4003
Ion Ratio Lower Upper
146 100
111 35.8 20.3 61.0
148 51.6 32.1 96.5





#89

n-Butylbenzene

Concen: 0.943 ug/l

RT: 12.335 min Scan# 1844

Delta R.T. 0.006 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion: 91 Resp: 6180

Ion Ratio Lower Upper

91 100

92 51.3 27.1 81.3





134 24.5 12.9 38.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#90

Hexachloroethane

Concen: 0.893 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX046725.D




Acq: 17 Jun 2025 17:18


Tgt Ion:117 Resp: 1081

Ion Ratio Lower Upper

117 100


201 60.5 34.4 103.3





#91
1,2-Dichlorobenzene
Concen: 0.904 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001



Tgt Ion:146 Resp: 3024
Ion Ratio Lower Upper
146 100
111 44.1 21.7 65.1
148 63.1 32.0 96.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 0.659 ug/l
RT: 12.951 min Scan# 1946
Delta R.T. 0.012 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Tgt Ion: 75 Resp: 277
Ion Ratio Lower Upper
75 100
155 73.6 41.9 125.7
157 102.5 51.4 154.3





#93

1,2,4-Trichlorobenzene

Concen: 1.019 ug/l

RT: 13.591 min Scan# 2050

Delta R.T. 0.006 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC001

Tgt Ion:	180	Resp:	242
Ion Ratio	Lower	Upper	
180	100		
182	93.2	46.7	140.1
145	30.0	15.8	47.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#94

Hexachlorobutadiene

Concen: 1.025 ug/l




RT: 13.725 min Scan# 2073


Delta R.T. 0.006 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18


Tgt Ion:	225	Resp:	1025
Ion Ratio	Lower	Upper	
225	100		
223	62.1	32.5	97.5
227	62.4	32.2	96.6





#95
Naphthalene
Concen: 0.806 ug/l
RT: 13.780 min Scan# 2081
Delta R.T. 0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001



Tgt Ion:128 Resp: 546
Ion Ratio Lower Upper
128 100
127 11.8 10.1 15.1
129 8.9 8.7 13.1




Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#96
1,2,3-Trichlorobenzene
Concen: 0.871 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. 0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Tgt Ion:180 Resp: 1982
Ion Ratio Lower Upper
180 100
182 99.9 47.8 143.4
145 34.1 17.3 51.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
 Data File : VX046726.D
 Acq On : 17 Jun 2025 18:00
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX061725

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 06/18/2025
 Supervised By : Mahesh Dadoda 06/18/2025

Quant Time: Jun 18 07:50:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	129144	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	209541	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	184761	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	91992	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	81629	45.697	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	91.400%		
35) Dibromofluoromethane	5.397	113	66647	48.072	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.140%		
50) Toluene-d8	8.647	98	236781	47.537	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	95.080%		
62) 4-Bromofluorobenzene	11.079	95	88588	47.712	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	95.420%		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.179	85	75554	54.798	ug/l	100
3) Chloromethane	1.307	50	76483	51.393	ug/l	99
4) Vinyl Chloride	1.386	62	82156	51.768	ug/l	95
5) Bromomethane	1.618	94	48957	54.831	ug/l	96
6) Chloroethane	1.703	64	49931	51.911	ug/l	99
7) Trichlorofluoromethane	1.904	101	123768	51.363	ug/l	98
8) Diethyl Ether	2.148	74	42437	51.316	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.343	101	75525	51.089	ug/l	99
10) Methyl Iodide	2.465	142	84139	45.861	ug/l	99
11) Tert butyl alcohol	2.959	59	40385	251.888	ug/l	100
12) 1,1-Dichloroethene	2.337	96	74097	51.945	ug/l	99
13) Acrolein	2.246	56	59544	254.972	ug/l	99
14) Allyl chloride	2.678	41	131610	50.922	ug/l	99
15) Acrylonitrile	3.075	53	187601	260.219	ug/l	98
16) Acetone	2.386	43	146745	246.347	ug/l	98
17) Carbon Disulfide	2.526	76	205516	47.710	ug/l	98
18) Methyl Acetate	2.715	43	80049	52.909	ug/l	98
19) Methyl tert-butyl Ether	3.123	73	226557	51.008	ug/l	99
20) Methylene Chloride	2.800	84	80668	48.696	ug/l	99
21) trans-1,2-Dichloroethene	3.105	96	76889	50.375	ug/l	98
22) Diisopropyl ether	3.770	45	257775	51.890	ug/l	99
23) Vinyl Acetate	3.733	43	1044679	264.451	ug/l	100
24) 1,1-Dichloroethane	3.623	63	142442	50.502	ug/l	99
25) 2-Butanone	4.562	43	222870	264.542	ug/l	98
26) 2,2-Dichloropropane	4.483	77	104151	50.629	ug/l	99
27) cis-1,2-Dichloroethene	4.501	96	89665	50.003	ug/l	99
28) Bromochloromethane	4.904	49	64330	50.308	ug/l	99
29) Tetrahydrofuran	5.013	42	143570	263.836	ug/l	99
30) Chloroform	5.105	83	144108	51.089	ug/l	98
31) Cyclohexane	5.483	56	132301	49.462	ug/l	99
32) 1,1,1-Trichloroethane	5.391	97	123796	50.042	ug/l	99
36) 1,1-Dichloropropene	5.702	75	100241	50.088	ug/l	99
37) Ethyl Acetate	4.721	43	94953	51.401	ug/l	99
38) Carbon Tetrachloride	5.690	117	109471	50.403	ug/l	98
39) Methylcyclohexane	7.385	83	132086	50.155	ug/l	97
40) Benzene	6.044	78	304346	51.385	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
 Data File : VX046726.D
 Acq On : 17 Jun 2025 18:00
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICSVX061725

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 06/18/2025
 Supervised By : Mahesh Dadoda 06/18/2025

Quant Time: Jun 18 07:50:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.928	41	54501	55.341	ug/l	97
42) 1,2-Dichloroethane	6.099	62	106085	50.970	ug/l	99
43) Isopropyl Acetate	6.342	43	155773	53.888	ug/l	99
44) Trichloroethene	7.129	130	76981	51.000	ug/l	97
45) 1,2-Dichloropropane	7.434	63	74773	50.940	ug/l	94
46) Dibromomethane	7.586	93	54149	51.329	ug/l	100
47) Bromodichloromethane	7.824	83	110823	51.479	ug/l	100
48) Methyl methacrylate	7.696	41	79868	55.091	ug/l	99
49) 1,4-Dioxane	7.659	88	20314	964.253	ug/l	97
51) 4-Methyl-2-Pentanone	8.574	43	470527	272.916	ug/l	99
52) Toluene	8.720	92	188131	51.237	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	105435	52.785	ug/l	99
54) cis-1,3-Dichloropropene	8.366	75	117487	51.813	ug/l	97
55) 1,1,2-Trichloroethane	9.153	97	70739	51.698	ug/l	99
56) Ethyl methacrylate	9.116	69	109612	54.266	ug/l	100
57) 1,3-Dichloropropane	9.305	76	121649	51.625	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.244	63	293709	279.922	ug/l	100
59) 2-Hexanone	9.427	43	326564	273.982	ug/l	99
60) Dibromochloromethane	9.519	129	82234	50.997	ug/l	99
61) 1,2-Dibromoethane	9.610	107	72749	52.268	ug/l	98
64) Tetrachloroethene	9.275	164	62943	49.172	ug/l	98
65) Chlorobenzene	10.079	112	203947	50.006	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	69709	50.856	ug/l	99
67) Ethyl Benzene	10.195	91	359779	50.487	ug/l	98
68) m/p-Xylenes	10.299	106	272935	102.789	ug/l	99
69) o-Xylene	10.640	106	130980	52.635	ug/l	97
70) Styrene	10.653	104	227060	52.139	ug/l	100
71) Bromoform	10.799	173	53660	51.404	ug/l #	100
73) Isopropylbenzene	10.957	105	346098	51.083	ug/l	100
74) N-amyl acetate	10.842	43	138890	54.107	ug/l	100
75) 1,1,2,2-Tetrachloroethane	11.207	83	98960	52.310	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	81355m	48.286	ug/l	
77) Bromobenzene	11.195	156	82515	50.611	ug/l	99
78) n-propylbenzene	11.305	91	412345	51.240	ug/l	100
79) 2-Chlorotoluene	11.360	91	239904	49.931	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	288104	51.746	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	29929	50.210	ug/l	99
82) 4-Chlorotoluene	11.451	91	280848	50.085	ug/l	98
83) tert-Butylbenzene	11.713	119	287513	49.958	ug/l	98
84) 1,2,4-Trimethylbenzene	11.750	105	286365	51.193	ug/l	99
85) sec-Butylbenzene	11.890	105	371523	51.079	ug/l	100
86) p-Isopropyltoluene	12.006	119	311698	50.791	ug/l	100
87) 1,3-Dichlorobenzene	11.969	146	154181	48.501	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	153158	46.906	ug/l	100
89) n-Butylbenzene	12.329	91	296174	50.829	ug/l	100
90) Hexachloroethane	12.536	117	52831	49.161	ug/l	98
91) 1,2-Dichlorobenzene	12.335	146	147561	49.672	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	12.939	75	19851	53.151	ug/l	97
93) 1,2,4-Trichlorobenzene	13.585	180	104267	49.369	ug/l	99
94) Hexachlorobutadiene	13.719	225	45958	51.729	ug/l	98
95) Naphthalene	13.774	128	315669	52.462	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	103451	51.200	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
Data File : VX046726.D
Acq On : 17 Jun 2025 18:00
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025

Quant Time: Jun 18 07:50:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 03:09:16 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


```
Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
Data File : VX046726.D
Acq On    : 17 Jun 2025  18:00
Operator  : JC/MD
Sample    : VSTDICV050
Misc      : 5.0mL/MSVOA_X/WATER
ALS Vial  : 14   Sample Multiplier: 1
```


Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725

Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025

Quant Time: Jun 18 07:50:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 03:09:16 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.562 min Scan# 71
Delta R.T. 0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725

Tgt Ion:168 Resp: 129144
Ion Ratio Lower Upper
168 100
99 61.8 48.5 72.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#2
Dichlorodifluoromethane
Concen: 54.798 ug/l
RT: 1.179 min Scan# 15
Delta R.T. -0.006 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Tgt Ion: 85 Resp: 75554
Ion Ratio Lower Upper
85 100
87 33.0 16.6 49.8









#3
Chloromethane
Concen: 51.393 ug/l
RT: 1.307 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725

Tgt Ion: 50 Resp: 7648
Ion Ratio Lower Upper
50 100
52 31.8 24.9 37.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#4
Vinyl Chloride
Concen: 51.768 ug/l
RT: 1.386 min Scan# 49
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Tgt Ion: 62 Resp: 82156
Ion Ratio Lower Upper
62 100
64 32.7 23.9 35.9









#5
Bromomethane
Concen: 54.831 ug/l
RT: 1.618 min Scan# 81
Delta R.T. -0.006 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725

Tgt Ion: 94 Resp: 4895
Ion Ratio Lower Upper
94 100
96 94.4 72.3 108.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#6
Chloroethane
Concen: 51.911 ug/l
RT: 1.703 min Scan# 101
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00


Tgt Ion: 64 Resp: 49931
Ion Ratio Lower Upper
64 100
66 33.5 27.1 40.7





#7
Trichlorofluoromethane
Concen: 51.363 ug/l
RT: 1.904 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00




Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725



Tgt Ion:101 Resp: 123768
Ion Ratio Lower Upper
101 100
103 63.4 51.8 77.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#8
Diethyl Ether
Concen: 51.316 ug/l
RT: 2.148 min Scan# 174
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Tgt Ion: 74 Resp: 42437
Ion Ratio Lower Upper
74 100
45 103.7 52.3 157.1





#9

1,1,2-Trichlorotrifluoroethane

Concen: 51.089 ug/l

RT: 2.343 min Scan# 206

Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00

Instrument :

MSVOA_X

ClientSampleId :





ICVX061725

Tgt Ion:	101	Resp:	7552
Ion	Ratio	Lower	Upper
101	100		
85	43.6	35.4	53.0
151	75.9	61.2	91.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#10

Methyl Iodide

Concen: 45.861 ug/l




RT: 2.465 min Scan# 226


Delta R.T. -0.006 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00

Tgt Ion:	142	Resp:	84139
Ion	Ratio	Lower	Upper
142	100		
127	45.9	37.0	55.6
141	14.0	11.5	17.3





#11

Tert butyl alcohol

Concen: 251.888 ug/l

RT: 2.959 min Scan# 307

Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00

Instrument :

MSVOA_X

ClientSampleId :

ICVVX061725

Tgt Ion: 59 Resp: 4038

Ion Ratio Lower Upper

59 100





57 10.4 8.3 12.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#12

1,1-Dichloroethene

Concen: 51.945 ug/l

RT: 2.337 min Scan# 205

Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00




Tgt Ion: 96 Resp: 74097


Ion Ratio Lower Upper

96 100

61 159.8 126.6 189.8

98 63.7 50.7 76.1





#13

Acrolein

Concen: 254.972 ug/l

RT: 2.246 min Scan# 190

Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00

Instrument :

MSVOA_X

ClientSampleId :

ICVX061725

Tgt Ion: 56 Resp: 59544

Ion Ratio Lower Upper

56 100





55 70.0 55.2 82.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#14

Allyl chloride

Concen: 50.922 ug/l

RT: 2.678 min Scan# 261

Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00




Tgt Ion: 41 Resp: 131610


Ion Ratio Lower Upper

41 100

39 66.1 53.6 80.4

76 33.1 26.8 40.2





#15

Acrylonitrile

Concen: 260.219 ug/l

RT: 3.075 min Scan# 31

Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00

Instrument :

MSVOA_X

ClientSampleId :

ICVX061725

Tgt Ion: 53 Resp: 18760

Ion Ratio Lower Upper

53 100

52 82.9 65.1 97.7





51 36.7 28.9 43.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#16

Acetone

Concen: 246.347 ug/l

RT: 2.386 min Scan# 213

Delta R.T. -0.000 min

Lab File: VX046726.D




Acq: 17 Jun 2025 18:00


Tgt Ion: 43 Resp: 146745

Ion Ratio Lower Upper

43 100

58 29.4 24.5 36.7





#17

Carbon Disulfide

Concen: 47.710 ug/l

RT: 2.526 min Scan# 236

Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00

Instrument :

MSVOA_X

ClientSampleId :

ICVVX061725

Tgt Ion: 76 Resp: 205510

Ion Ratio Lower Upper

76 100





78 8.7 7.5 11.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#18

Methyl Acetate

Concen: 52.909 ug/l

RT: 2.715 min Scan# 267

Delta R.T. -0.000 min

Lab File: VX046726.D




Acq: 17 Jun 2025 18:00


Tgt Ion: 43 Resp: 80049

Ion Ratio Lower Upper

43 100

74 21.9 18.4 27.6









#19
Methyl tert-butyl Ether
Concen: 51.008 ug/l
RT: 3.123 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725

Tgt Ion: 73 Resp: 22655
Ion Ratio Lower Upper
73 100
57 22.3 18.2 27.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#20
Methylene Chloride
Concen: 48.696 ug/l
RT: 2.800 min Scan# 281
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00


Tgt Ion: 84 Resp: 80668
Ion Ratio Lower Upper
84 100
49 126.0 101.3 151.9
51 38.3 31.0 46.6
86 65.3 51.1 76.7





#21
trans-1,2-Dichloroethene
Concen: 50.375 ug/l
RT: 3.105 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00




Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725



Tgt Ion: 96 Resp: 76889
Ion Ratio Lower Upper
96 100
61 142.0 112.5 168.7
98 69.1 51.9 77.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#22
Diisopropyl ether
Concen: 51.890 ug/l
RT: 3.770 min Scan# 440
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Tgt Ion: 45 Resp: 257775
Ion Ratio Lower Upper
45 100
43 114.3 91.9 137.9
87 26.4 22.0 33.0
59 11.2 9.1 13.7









#23
Vinyl Acetate
Concen: 264.451 ug/l
RT: 3.733 min Scan# 41
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725

Tgt Ion: 43 Resp: 1044679
Ion Ratio Lower Upper
43 100
86 10.2 8.2 12.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#24
1,1-Dichloroethane
Concen: 50.502 ug/l
RT: 3.623 min Scan# 416
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Tgt Ion: 63 Resp: 142442
Ion Ratio Lower Upper
63 100
98 7.8 4.0 11.9
100 4.7 2.2 6.6





#25

2-Butanone

Concen: 264.542 ug/l

RT: 4.562 min Scan# 51

Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00

Instrument :

MSVOA_X

ClientSampleId :

ICVVX061725

Tgt Ion: 43 Resp: 222870

Ion Ratio Lower Upper

43 100





72 24.1 20.2 30.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#26

2,2-Dichloropropane

Concen: 50.629 ug/l

RT: 4.483 min Scan# 557

Delta R.T. -0.006 min

Lab File: VX046726.D




Acq: 17 Jun 2025 18:00


Tgt Ion: 77 Resp: 104151

Ion Ratio Lower Upper

77 100

97 23.0 11.8 35.3





#27

cis-1,2-Dichloroethene

Concen: 50.003 ug/l

RT: 4.501 min Scan# 50

Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00

Instrument :

MSVOA_X

ClientSampleId :

ICVVX061725

Tgt Ion: 96 Resp: 8966

Ion Ratio Lower Upper

96 100

61 145.1 0.0 291.0





98 64.2 0.0 126.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#28

Bromochloromethane

Concen: 50.308 ug/l

RT: 4.904 min Scan# 626

Delta R.T. -0.006 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00




Tgt Ion: 49 Resp: 64330


Ion Ratio Lower Upper

49 100

129 1.5 0.0 3.8

130 75.6 61.4 92.0





#29

Tetrahydrofuran

Concen: 263.836 ug/l

RT: 5.013 min Scan# 644

Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00

Instrument :

MSVOA_X

ClientSampleId :

ICVX061725

Tgt Ion: 42 Resp: 143570

Ion Ratio Lower Upper

42 100

72 44.3 34.9 52.3





71 40.8 32.2 48.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#30

Chloroform

Concen: 51.089 ug/l

RT: 5.105 min Scan# 659

Delta R.T. -0.000 min

Lab File: VX046726.D




Acq: 17 Jun 2025 18:00


Tgt Ion: 83 Resp: 144108

Ion Ratio Lower Upper

83 100

85 65.4 53.7 80.5





#31

Cyclohexane

Concen: 49.462 ug/l

RT: 5.483 min Scan# 705

Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00

Instrument :

MSVOA_X

ClientSampleId :

ICVX061725

Tgt Ion: 56 Resp: 13230

Ion Ratio Lower Upper

56 100

69 30.8 24.8 37.2





84 86.8 68.2 102.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#32

1,1,1-Trichloroethane

Concen: 50.042 ug/l

RT: 5.391 min Scan# 706

Delta R.T. -0.006 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00




Tgt Ion: 97 Resp: 123796


Ion Ratio Lower Upper

97 100

99 63.5 51.3 76.9

61 45.1 36.4 54.6





#33

1,2-Dichloroethane-d4

Concen: 45.697 ug/l

RT: 5.958 min Scan# 799

Delta R.T. -0.006 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00

Instrument :

MSVOA_X

ClientSampleId :

ICVVX061725

Tgt Ion: 65 Resp: 81629

Ion Ratio Lower Upper

65 100





67 52.5 0.0 105.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.769 min Scan# 932

Delta R.T. 0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00




Tgt Ion:114 Resp: 209541


Ion Ratio Lower Upper

114 100

63 20.5 0.0 44.2

88 16.2 0.0 33.2





#35

Dibromofluoromethane

Concen: 48.072 ug/l

RT: 5.397 min Scan# 707

Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00

Instrument :

MSVOA_X

ClientSampleId :





ICVVX061725

Tgt Ion:	113	Resp:	6664
Ion	Ratio	Lower	Upper
113	100		
111	102.5	82.0	123.0
192	18.9	15.3	22.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#36

1,1-Dichloropropene

Concen: 50.088 ug/l




RT: 5.702 min Scan# 757


Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00

Tgt Ion:	75	Resp:	100241
Ion	Ratio	Lower	Upper
75	100		
110	35.3	17.2	51.5
77	30.7	24.2	36.4





#37

Ethyl Acetate

Concen: 51.401 ug/l

RT: 4.721 min Scan# 596

Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00

Instrument :

MSVOA_X

ClientSampleId :





ICVX061725

Tgt Ion: 43 Resp: 9495

Ion	Ratio	Lower	Upper
43	100		
61	12.6	9.7	14.5
70	10.1	8.1	12.1

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#38

Carbon Tetrachloride

Concen: 50.403 ug/l

RT: 5.690 min Scan# 755




Delta R.T. -0.000 min


Lab File: VX046726.D

Acq: 17 Jun 2025 18:00

Tgt Ion:117 Resp: 109471

Ion	Ratio	Lower	Upper
117	100		
119	96.3	75.2	112.8
121	31.6	24.1	36.1





#39

Methylcyclohexane

Concen: 50.155 ug/l

RT: 7.385 min Scan# 1033

Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00

Instrument :

MSVOA_X

ClientSampleId :

ICVX061725

Tgt Ion: 83 Resp: 132080

Ion Ratio Lower Upper

83 100

55 81.1 63.0 94.4





98 50.5 38.8 58.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#40

Benzene

Concen: 51.385 ug/l

RT: 6.044 min Scan# 813

Delta R.T. -0.000 min

Lab File: VX046726.D




Acq: 17 Jun 2025 18:00


Tgt Ion: 78 Resp: 304346

Ion Ratio Lower Upper

78 100

77 23.4 18.8 28.2





#41

Methacrylonitrile

Concen: 55.341 ug/l

RT: 4.928 min Scan# 61

Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00

Instrument :

MSVOA_X

ClientSampleId :

ICVVX061725

Tgt Ion: 41 Resp: 5450

Ion Ratio Lower Upper

41 100

39 55.5 45.9 68.9

67 72.5 61.2 91.8





52 30.7 25.5 38.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#42

1,2-Dichloroethane

Concen: 50.970 ug/l

RT: 6.099 min Scan# 822

Delta R.T. 0.006 min

Lab File: VX046726.D




Acq: 17 Jun 2025 18:00


Tgt Ion: 62 Resp: 106085

Ion Ratio Lower Upper

62 100


98 9.4 0.0 17.8





#43
Isopropyl Acetate
Concen: 53.888 ug/l
RT: 6.342 min Scan# 862
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00




Instrument :
MSVOA_X
ClientSampleId :
ICVX061725



Tgt Ion: 43 Resp: 15577
Ion Ratio Lower Upper
43 100
61 18.8 15.3 22.9
87 12.2 10.0 15.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#44
Trichloroethene
Concen: 51.000 ug/l
RT: 7.129 min Scan# 991
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Tgt Ion:130 Resp: 76981
Ion Ratio Lower Upper
130 100
95 98.6 0.0 202.6





#45

1,2-Dichloropropane

Concen: 50.940 ug/l

RT: 7.434 min Scan# 1041

Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00

Instrument :

MSVOA_X

ClientSampleId :

ICVX061725

Tgt Ion: 63 Resp: 7477

Ion Ratio Lower Upper

63 100





65 31.9 22.9 34.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#46

Dibromomethane

Concen: 51.329 ug/l

RT: 7.586 min Scan# 1066

Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00




Tgt Ion: 93 Resp: 54149


Ion Ratio Lower Upper

93 100

95 82.7 66.2 99.4

174 96.3 77.0 115.6





#47

Bromodichloromethane

Concen: 51.479 ug/l

RT: 7.824 min Scan# 1105

Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00

Instrument :

MSVOA_X

ClientSampleId :





ICVX061725

Tgt Ion:	83	Resp:	11082
Ion	Ratio	Lower	Upper
83	100		
85	64.7	51.4	77.2
127	8.4	6.9	10.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#48

Methyl methacrylate

Concen: 55.091 ug/l




RT: 7.696 min Scan# 1084


Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00

Tgt Ion:	41	Resp:	79868
Ion	Ratio	Lower	Upper
41	100		
69	80.6	63.4	95.2
39	55.9	45.4	68.0





#49

1,4-Dioxane

Concen: 964.253 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. 0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00

Instrument :

MSVOA_X

ClientSampleId :





ICVVX061725

Tgt Ion:	88	Resp:	20314
Ion	Ratio	Lower	Upper
88	100		
43	38.3	31.2	46.8
58	72.0	54.7	82.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#50

Toluene-d8

Concen: 47.537 ug/l




RT: 8.647 min Scan# 1240


Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00

Tgt Ion:	98	Resp:	236781
Ion	Ratio	Lower	Upper
98	100		
100	66.9	53.0	79.4









#51
4-Methyl-2-Pentanone
Concen: 272.916 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725

Tgt Ion: 43 Resp: 47052
Ion Ratio Lower Upper
43 100
58 38.2 30.8 46.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#52
Toluene
Concen: 51.237 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Tgt Ion: 92 Resp: 188131
Ion Ratio Lower Upper
92 100
91 170.4 134.8 202.2





#53

t-1,3-Dichloropropene

Concen: 52.785 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00

Instrument :

MSVOA_X

ClientSampleId :

ICVVX061725

Tgt Ion: 75 Resp: 105438

Ion Ratio Lower Upper

75 100





77 32.4 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#54

cis-1,3-Dichloropropene

Concen: 51.813 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00




Tgt Ion: 75 Resp: 117487


Ion Ratio Lower Upper

75 100

77 31.5 25.3 37.9


39 52.2 39.2 58.8





#55
1,1,2-Trichloroethane
Concen: 51.698 ug/l
RT: 9.153 min Scan# 1323
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00




Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725



Tgt Ion: 97 Resp: 70739
Ion Ratio Lower Upper
97 100
83 83.8 67.2 100.8
85 55.5 44.1 66.1
99 64.8 50.3 75.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#56
Ethyl methacrylate
Concen: 54.266 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Tgt Ion: 69 Resp: 109612
Ion Ratio Lower Upper
69 100
41 69.3 55.4 83.2
39 36.7 29.7 44.5





#57

1,3-Dichloropropane

Concen: 51.625 ug/l

RT: 9.305 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00

Instrument :

MSVOA_X

ClientSampleId :

ICVVX061725

Tgt Ion: 76 Resp: 121649

Ion Ratio Lower Upper

76 100





78 32.0 25.7 38.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#58

2-Chloroethyl Vinyl ether

Concen: 279.922 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. -0.000 min

Lab File: VX046726.D




Acq: 17 Jun 2025 18:00


Tgt Ion: 63 Resp: 293709

Ion Ratio Lower Upper

63 100

106 28.3 22.6 34.0





#59

2-Hexanone

Concen: 273.982 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00

Instrument :

MSVOA_X

ClientSampleId :

ICVX061725

Tgt Ion: 43 Resp: 326564

Ion Ratio Lower Upper

43 100





58 54.1 26.8 80.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#60

Dibromochloromethane

Concen: 50.997 ug/l

RT: 9.519 min Scan# 1383

Delta R.T. -0.000 min

Lab File: VX046726.D




Acq: 17 Jun 2025 18:00


Tgt Ion:129 Resp: 82234

Ion Ratio Lower Upper

129 100


127 77.7 38.4 115.2





#61
1,2-Dibromoethane
Concen: 52.268 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00




Instrument :
MSVOA_X
ClientSampleId :
ICVX061725



Tgt Ion:107 Resp: 72749
Ion Ratio Lower Upper
107 100
109 92.8 75.8 113.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#62
4-Bromofluorobenzene
Concen: 47.712 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00


Tgt Ion: 95 Resp: 88588
Ion Ratio Lower Upper
95 100
174 75.6 0.0 150.4
176 73.6 0.0 145.0





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00




Instrument : MSVOA_X
ClientSampleId : ICVVX061725



Tgt Ion:117 Resp: 18476
Ion Ratio Lower Upper
117 100
82 55.0 44.6 66.8
119 32.9 25.8 38.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#64
Tetrachloroethene
Concen: 49.172 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Tgt Ion:164 Resp: 62943
Ion Ratio Lower Upper
164 100
166 127.1 104.6 157.0
129 94.2 76.3 114.5
131 91.2 74.1 111.1









#65
Chlorobenzene
Concen: 50.006 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725

Tgt Ion:112 Resp: 20394
Ion Ratio Lower Upper
112 100
114 32.2 25.5 38.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 50.856 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Tgt Ion:131 Resp: 69709
Ion Ratio Lower Upper
131 100
133 95.2 47.3 141.8
119 63.5 32.1 96.3





#67

Ethyl Benzene

Concen: 50.487 ug/l

RT: 10.195 min Scan# 1493

Delta R.T. 0.006 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00

Instrument :

MSVOA_X

ClientSampleId :

ICVVX061725

Tgt Ion: 91 Resp: 359779

Ion Ratio Lower Upper

91 100





106 31.4 24.2 36.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#68

m/p-Xylenes

Concen: 102.789 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. -0.000 min

Lab File: VX046726.D




Acq: 17 Jun 2025 18:00


Tgt Ion:106 Resp: 272935

Ion Ratio Lower Upper

106 100

91 205.2 164.8 247.2





#69

o-Xylene

Concen: 52.635 ug/l

RT: 10.640 min Scan# 1567

Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00

Instrument :

MSVOA_X

ClientSampleId :

ICVX061725

Tgt Ion:106 Resp: 130980

Ion Ratio Lower Upper

106 100





91 218.4 111.3 333.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#70

Styrene

Concen: 52.139 ug/l

RT: 10.653 min Scan# 1569

Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00




Tgt Ion:104 Resp: 227060


Ion Ratio Lower Upper

104 100

78 51.7 41.4 62.2


103 54.1 43.1 64.7





#71
Bromoform
Concen: 51.404 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00




Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725



Tgt Ion:173 Resp: 53660
Ion Ratio Lower Upper
173 100
175 48.6 24.1 72.4
254 0.0 0.1 0.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Tgt Ion:152 Resp: 91992
Ion Ratio Lower Upper
152 100
115 86.9 43.2 129.6
150 174.0 0.0 346.8









#73
Isopropylbenzene
Concen: 51.083 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725

Tgt Ion:105 Resp: 346098
Ion Ratio Lower Upper
105 100
120 26.0 13.0 39.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#74
N-ethyl acetate
Concen: 54.107 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Tgt Ion: 43 Resp: 138890
Ion Ratio Lower Upper
43 100
70 41.5 33.4 50.0
55 23.9 19.1 28.7
61 23.9 19.0 28.4









#75
1,1,2,2-Tetrachloroethane
Concen: 52.310 ug/l
RT: 11.207 min Scan# 1660
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Instrument :
MSVOA_X
ClientSampleId :
ICVX061725

Tgt Ion: 83 Resp: 98960
Ion Ratio Lower Upper
83 100
131 9.9 4.9 14.6
85 64.7 32.0 96.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#76
1,2,3-Trichloropropane
Concen: 48.286 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00


Tgt Ion: 75 Resp: 81355
Ion Ratio Lower Upper
75 100
77 41.9 18.5 55.5





#77
Bromobenzene
Concen: 50.611 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00




Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725



Tgt Ion:156 Resp: 82515
Ion Ratio Lower Upper
156 100
77 167.0 82.8 248.6
158 96.2 47.6 142.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#78
n-propylbenzene
Concen: 51.240 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. 0.006 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Tgt Ion: 91 Resp: 412345
Ion Ratio Lower Upper
91 100
120 22.6 11.4 34.1









#79
2-Chlorotoluene
Concen: 49.931 ug/l
RT: 11.360 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725

Tgt Ion: 91 Resp: 239904
Ion Ratio Lower Upper
91 100
126 33.5 16.6 49.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#80
1,3,5-Trimethylbenzene
Concen: 51.746 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00


Tgt Ion:105 Resp: 288104
Ion Ratio Lower Upper
105 100
120 47.6 24.1 72.2





#81
trans-1,4-Dichloro-2-butene
Concen: 50.210 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00




Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725



Tgt Ion: 75 Resp: 29929
Ion Ratio Lower Upper
75 100
53 109.9 86.8 130.2
89 45.3 36.9 55.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#82
4-Chlorotoluene
Concen: 50.085 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Tgt Ion: 91 Resp: 280848
Ion Ratio Lower Upper
91 100
126 29.8 14.4 43.2





#83

tert-Butylbenzene

Concen: 49.958 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00

Instrument :

MSVOA_X

ClientSampleId :

ICVX061725

Tgt Ion:119 Resp: 28751

Ion Ratio Lower Upper

119 100

91 60.6 29.5 88.6





134 23.5 11.7 35.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#84

1,2,4-Trimethylbenzene

Concen: 51.193 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. -0.000 min

Lab File: VX046726.D




Acq: 17 Jun 2025 18:00


Tgt Ion:105 Resp: 286365

Ion Ratio Lower Upper

105 100


120 44.7 22.1 66.3





#85
sec-Butylbenzene
Concen: 51.079 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00




Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725



Tgt Ion:105 Resp: 37152
Ion Ratio Lower Upper
105 100
134 19.7 9.8 29.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#86
p-Isopropyltoluene
Concen: 50.791 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Tgt Ion:119 Resp: 311698
Ion Ratio Lower Upper
119 100
134 26.0 12.9 38.7
91 24.6 12.3 36.9





#87

1,3-Dichlorobenzene

Concen: 48.501 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00

Instrument :

MSVOA_X

ClientSampleId :

ICVX061725

Tgt Ion:146 Resp: 15418

Ion Ratio Lower Upper

146 100

111 41.8 20.6 61.8





148 64.2 31.6 94.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025

Supervised By :Mahesh Dadoda 06/18/2025



#88

1,4-Dichlorobenzene

Concen: 46.906 ug/l

RT: 12.036 min Scan# 1796

Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00




Tgt Ion:146 Resp: 153158


Ion Ratio Lower Upper

146 100

111 41.0 20.3 61.0

148 64.0 32.1 96.5





#89

n-Butylbenzene

Concen: 50.829 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00

Instrument :

MSVOA_X

ClientSampleId :

ICVVX061725





Tgt Ion: 91 Resp: 296174

Ion	Ratio	Lower	Upper
91	100		
92	54.2	27.1	81.3
134	25.4	12.9	38.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#90

Hexachloroethane

Concen: 49.161 ug/l

RT: 12.536 min Scan# 1878




Delta R.T. -0.000 min


Lab File: VX046726.D

Acq: 17 Jun 2025 18:00

Tgt Ion:117 Resp: 52831


Ion	Ratio	Lower	Upper
117	100		
201	70.3	34.4	103.3





#91
1,2-Dichlorobenzene
Concen: 49.672 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.006 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00




Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725



Tgt Ion:146 Resp: 14756
Ion Ratio Lower Upper
146 100
111 43.4 21.7 65.1
148 63.5 32.0 96.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 53.151 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00


Tgt Ion: 75 Resp: 19851
Ion Ratio Lower Upper
75 100
155 84.6 41.9 125.7
157 107.2 51.4 154.3





#93
1,2,4-Trichlorobenzene
Concen: 49.369 ug/l
RT: 13.585 min Scan# 2050
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00




Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725




Tgt Ion:180 Resp: 10426
Ion Ratio Lower Upper
180 100
182 94.4 46.7 140.1
145 32.5 15.8 47.4

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025




#94
Hexachlorobutadiene
Concen: 51.729 ug/l
RT: 13.719 min Scan# 2072
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00




Tgt Ion:225 Resp: 45958
Ion Ratio Lower Upper
225 100
223 63.4 32.5 97.5
227 62.5 32.2 96.6





#95
Naphthalene
Concen: 52.462 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00




Instrument :
MSVOA_X
ClientSampleId :
ICVX061725



Tgt Ion:128 Resp: 31566
Ion Ratio Lower Upper
128 100
127 12.7 10.1 15.1
129 10.9 8.7 13.1




Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 06/18/2025
Supervised By :Mahesh Dadoda 06/18/2025



#96
1,2,3-Trichlorobenzene
Concen: 51.200 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Tgt Ion:180 Resp: 103451
Ion Ratio Lower Upper
180 100
182 94.5 47.8 143.4
145 33.9 17.3 51.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
 Data File : VX046726.D
 Acq On : 17 Jun 2025 18:00
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICSVX061725

Quant Time: Jun 18 07:50:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	105	0.00
2 T	Dichlorodifluoromethane	0.534	0.585	-9.6	105	0.00
3 P	Chloromethane	0.576	0.592	-2.8	105	0.00
4 C	Vinyl Chloride	0.614	0.636	-3.6#	105	0.00
5 T	Bromomethane	0.346	0.379	-9.5	108	0.00
6 T	Chloroethane	0.372	0.387	-4.0	107	0.00
7 T	Trichlorofluoromethane	0.933	0.958	-2.7	104	0.00
8 T	Diethyl Ether	0.320	0.329	-2.8	107	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.572	0.585	-2.3	106	0.00
10 T	Methyl Iodide	0.642	0.652	-1.6	104	0.00
11 T	Tert butyl alcohol	0.062	0.063	-1.6	106	0.00
12 CM	1,1-Dichloroethene	0.552	0.574	-4.0#	106	0.00
13 T	Acrolein	0.090	0.092	-2.2	111	0.00
14 T	Allyl chloride	1.001	1.019	-1.8	107	0.00
15 T	Acrylonitrile	0.279	0.291	-4.3	107	0.00
16 T	Acetone	0.231	0.227	1.7	108	0.00
17 T	Carbon Disulfide	1.668	1.591	4.6	104	0.00
18 T	Methyl Acetate	0.586	0.620	-5.8	112	0.00
19 T	Methyl tert-butyl Ether	1.720	1.754	-2.0	105	0.00
20 T	Methylene Chloride	0.641	0.625	2.5	107	0.00
21 T	trans-1,2-Dichloroethene	0.591	0.595	-0.7	106	0.00
22 T	Diisopropyl ether	1.923	1.996	-3.8	106	0.00
23 T	Vinyl Acetate	1.529	1.618	-5.8	106	0.00
24 P	1,1-Dichloroethane	1.092	1.103	-1.0	106	0.00
25 T	2-Butanone	0.326	0.345	-5.8	107	0.00
26 T	2,2-Dichloropropane	0.796	0.806	-1.3	111	0.00
27 T	cis-1,2-Dichloroethene	0.694	0.694	0.0	106	0.00
28 T	Bromochloromethane	0.495	0.498	-0.6	107	0.00
29 T	Tetrahydrofuran	0.211	0.222	-5.2	107	0.00
30 C	Chloroform	1.092	1.116	-2.2#	105	0.00
31 T	Cyclohexane	1.036	1.024	1.2	106	0.00
32 T	1,1,1-Trichloroethane	0.958	0.959	-0.1	105	0.00
33 S	1,2-Dichloroethane-d4	0.692	0.632	8.7	102	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	103	0.00
35 S	Dibromofluoromethane	0.331	0.318	3.9	102	0.00
36 T	1,1-Dichloropropene	0.478	0.478	0.0	103	0.00
37 T	Ethyl Acetate	0.441	0.453	-2.7	108	0.00
38 T	Carbon Tetrachloride	0.518	0.522	-0.8	104	0.00
39 T	Methylcyclohexane	0.628	0.630	-0.3	103	0.00
40 TM	Benzene	1.413	1.452	-2.8	105	0.00
41 T	Methacrylonitrile	0.235	0.260	-10.6	112	0.00
42 TM	1,2-Dichloroethane	0.497	0.506	-1.8	105	0.00
43 T	Isopropyl Acetate	0.690	0.743	-7.7	106	0.00
44 TM	Trichloroethene	0.360	0.367	-1.9	106	0.00
45 C	1,2-Dichloropropane	0.350	0.357	-2.0#	106	0.00
46 T	Dibromomethane	0.252	0.258	-2.4	106	0.00
47 T	Bromodichloromethane	0.514	0.529	-2.9	104	0.00
48 T	Methyl methacrylate	0.346	0.381	-10.1	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
 Data File : VX046726.D
 Acq On : 17 Jun 2025 18:00
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICSVVX061725

Quant Time: Jun 18 07:50:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.004	0.005	-25.0	103	0.00
50 S	Toluene-d8	1.189	1.130	5.0	101	0.00
51 T	4-Methyl-2-Pentanone	0.411	0.449	-9.2	108	0.00
52 CM	Toluene	0.876	0.898	-2.5#	104	0.00
53 T	t-1,3-Dichloropropene	0.477	0.503	-5.5	106	0.00
54 T	cis-1,3-Dichloropropene	0.541	0.561	-3.7	105	0.00
55 T	1,1,2-Trichloroethane	0.327	0.338	-3.4	106	0.00
56 T	Ethyl methacrylate	0.482	0.523	-8.5	106	0.00
57 T	1,3-Dichloropropane	0.562	0.581	-3.4	106	0.00
58 T	2-Chloroethyl Vinyl ether	0.250	0.280	-12.0	102	0.00
59 T	2-Hexanone	0.284	0.312	-9.9	108	0.00
60 T	Dibromochloromethane	0.385	0.392	-1.8	104	0.00
61 T	1,2-Dibromoethane	0.332	0.347	-4.5	106	0.00
62 S	4-Bromofluorobenzene	0.443	0.423	4.5	102	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	103	0.00
64 T	Tetrachloroethene	0.346	0.341	1.4	103	0.00
65 PM	Chlorobenzene	1.104	1.104	0.0	104	0.00
66 T	1,1,1,2-Tetrachloroethane	0.371	0.377	-1.6	104	0.00
67 C	Ethyl Benzene	1.929	1.947	-0.9#	104	0.00
68 T	m/p-Xylenes	0.719	0.739	-2.8	105	0.00
69 T	o-Xylene	0.673	0.709	-5.3	106	0.00
70 T	Styrene	1.179	1.229	-4.2	105	0.00
71 P	Bromoform	0.282	0.290	-2.8	104	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
73 T	Isopropylbenzene	3.682	3.762	-2.2	104	0.00
74 T	N-amyl acetate	1.395	1.510	-8.2	106	0.00
75 P	1,1,2,2-Tetrachloroethane	1.028	1.076	-4.7	107	0.00
76 T	1,2,3-Trichloropropane	0.916	0.884	3.5	92	0.00
77 T	Bromobenzene	0.886	0.897	-1.2	104	0.00
78 T	n-propylbenzene	4.374	4.482	-2.5	105	0.00
79 T	2-Chlorotoluene	2.611	2.608	0.1	104	0.00
80 T	1,3,5-Trimethylbenzene	3.026	3.132	-3.5	105	0.00
81 T	trans-1,4-Dichloro-2-butene	0.324	0.325	-0.3	108	0.00
82 T	4-Chlorotoluene	3.048	3.053	-0.2	104	0.00
83 T	tert-Butylbenzene	3.128	3.125	0.1	103	0.00
84 T	1,2,4-Trimethylbenzene	3.040	3.113	-2.4	105	0.00
85 T	sec-Butylbenzene	3.953	4.039	-2.2	104	0.00
86 T	p-Isopropyltoluene	3.336	3.388	-1.6	105	0.00
87 T	1,3-Dichlorobenzene	1.728	1.676	3.0	103	0.00
88 T	1,4-Dichlorobenzene	1.775	1.665	6.2	104	0.00
89 T	n-Butylbenzene	3.167	3.220	-1.7	105	0.00
90 T	Hexachloroethane	0.584	0.574	1.7	102	0.00
91 T	1,2-Dichlorobenzene	1.615	1.604	0.7	104	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.203	0.216	-6.4	106	0.00
93 T	1,2,4-Trichlorobenzene	1.148	1.133	1.3	104	0.00
94 T	Hexachlorobutadiene	0.483	0.500	-3.5	108	0.00
95 T	Naphthalene	3.270	3.431	-4.9	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
Data File : VX046726.D
Acq On : 17 Jun 2025 18:00
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725

Quant Time: Jun 18 07:50:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 03:09:16 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.098	1.125	-2.5	107	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX061725\
 Data File : VX046726.D
 Acq On : 17 Jun 2025 18:00
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICSVX061725

Quant Time: Jun 18 07:50:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	105	0.00
2 T	Dichlorodifluoromethane	50.000	54.798	-9.6	105	0.00
3 P	Chloromethane	50.000	51.393	-2.8	105	0.00
4 C	Vinyl Chloride	50.000	51.768	-3.5#	105	0.00
5 T	Bromomethane	50.000	54.831	-9.7	108	0.00
6 T	Chloroethane	50.000	51.911	-3.8	107	0.00
7 T	Trichlorofluoromethane	50.000	51.363	-2.7	104	0.00
8 T	Diethyl Ether	50.000	51.316	-2.6	107	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	51.089	-2.2	106	0.00
10 T	Methyl Iodide	50.000	45.861	8.3	104	0.00
11 T	Tert butyl alcohol	250.000	251.888	-0.8	106	0.00
12 CM	1,1-Dichloroethene	50.000	51.945	-3.9#	106	0.00
13 T	Acrolein	250.000	254.972	-2.0	111	0.00
14 T	Allyl chloride	50.000	50.922	-1.8	107	0.00
15 T	Acrylonitrile	250.000	260.219	-4.1	107	0.00
16 T	Acetone	250.000	246.347	1.5	108	0.00
17 T	Carbon Disulfide	50.000	47.710	4.6	104	0.00
18 T	Methyl Acetate	50.000	52.909	-5.8	112	0.00
19 T	Methyl tert-butyl Ether	50.000	51.008	-2.0	105	0.00
20 T	Methylene Chloride	50.000	48.696	2.6	107	0.00
21 T	trans-1,2-Dichloroethene	50.000	50.375	-0.8	106	0.00
22 T	Diisopropyl ether	50.000	51.890	-3.8	106	0.00
23 T	Vinyl Acetate	250.000	264.451	-5.8	106	0.00
24 P	1,1-Dichloroethane	50.000	50.502	-1.0	106	0.00
25 T	2-Butanone	250.000	264.542	-5.8	107	0.00
26 T	2,2-Dichloropropane	50.000	50.629	-1.3	111	0.00
27 T	cis-1,2-Dichloroethene	50.000	50.003	-0.0	106	0.00
28 T	Bromochloromethane	50.000	50.308	-0.6	107	0.00
29 T	Tetrahydrofuran	250.000	263.836	-5.5	107	0.00
30 C	Chloroform	50.000	51.089	-2.2#	105	0.00
31 T	Cyclohexane	50.000	49.462	1.1	106	0.00
32 T	1,1,1-Trichloroethane	50.000	50.042	-0.1	105	0.00
33 S	1,2-Dichloroethane-d4	50.000	45.697	8.6	102	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	103	0.00
35 S	Dibromofluoromethane	50.000	48.072	3.9	102	0.00
36 T	1,1-Dichloropropene	50.000	50.088	-0.2	103	0.00
37 T	Ethyl Acetate	50.000	51.401	-2.8	108	0.00
38 T	Carbon Tetrachloride	50.000	50.403	-0.8	104	0.00
39 T	Methylcyclohexane	50.000	50.155	-0.3	103	0.00
40 TM	Benzene	50.000	51.385	-2.8	105	0.00
41 T	Methacrylonitrile	50.000	55.341	-10.7	112	0.00
42 TM	1,2-Dichloroethane	50.000	50.970	-1.9	105	0.00
43 T	Isopropyl Acetate	50.000	53.888	-7.8	106	0.00
44 TM	Trichloroethene	50.000	51.000	-2.0	106	0.00
45 C	1,2-Dichloropropane	50.000	50.940	-1.9#	106	0.00
46 T	Dibromomethane	50.000	51.329	-2.7	106	0.00
47 T	Bromodichloromethane	50.000	51.479	-3.0	104	0.00
48 T	Methyl methacrylate	50.000	55.091	-10.2	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX061725\
 Data File : VX046726.D
 Acq On : 17 Jun 2025 18:00
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICSVX061725

Quant Time: Jun 18 07:50:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	964.253	3.6	103	0.00
50 S	Toluene-d8	50.000	47.537	4.9	101	0.00
51 T	4-Methyl-2-Pentanone	250.000	272.916	-9.2	108	0.00
52 CM	Toluene	50.000	51.237	-2.5#	104	0.00
53 T	t-1,3-Dichloropropene	50.000	52.785	-5.6	106	0.00
54 T	cis-1,3-Dichloropropene	50.000	51.813	-3.6	105	0.00
55 T	1,1,2-Trichloroethane	50.000	51.698	-3.4	106	0.00
56 T	Ethyl methacrylate	50.000	54.266	-8.5	106	0.00
57 T	1,3-Dichloropropane	50.000	51.625	-3.3	106	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	279.922	-12.0	102	0.00
59 T	2-Hexanone	250.000	273.982	-9.6	108	0.00
60 T	Dibromochloromethane	50.000	50.997	-2.0	104	0.00
61 T	1,2-Dibromoethane	50.000	52.268	-4.5	106	0.00
62 S	4-Bromofluorobenzene	50.000	47.712	4.6	102	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	103	0.00
64 T	Tetrachloroethene	50.000	49.172	1.7	103	0.00
65 PM	Chlorobenzene	50.000	50.006	-0.0	104	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	50.856	-1.7	104	0.00
67 C	Ethyl Benzene	50.000	50.487	-1.0#	104	0.00
68 T	m/p-Xylenes	100.000	102.789	-2.8	105	0.00
69 T	o-Xylene	50.000	52.635	-5.3	106	0.00
70 T	Styrene	50.000	52.139	-4.3	105	0.00
71 P	Bromoform	50.000	51.404	-2.8	104	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	103	0.00
73 T	Isopropylbenzene	50.000	51.083	-2.2	104	0.00
74 T	N-ethyl acetate	50.000	54.107	-8.2	106	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	52.310	-4.6	107	0.00
76 T	1,2,3-Trichloropropane	50.000	48.286	3.4	92	0.00
77 T	Bromobenzene	50.000	50.611	-1.2	104	0.00
78 T	n-propylbenzene	50.000	51.240	-2.5	105	0.00
79 T	2-Chlorotoluene	50.000	49.931	0.1	104	0.00
80 T	1,3,5-Trimethylbenzene	50.000	51.746	-3.5	105	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	50.210	-0.4	108	0.00
82 T	4-Chlorotoluene	50.000	50.085	-0.2	104	0.00
83 T	tert-Butylbenzene	50.000	49.958	0.1	103	0.00
84 T	1,2,4-Trimethylbenzene	50.000	51.193	-2.4	105	0.00
85 T	sec-Butylbenzene	50.000	51.079	-2.2	104	0.00
86 T	p-Isopropyltoluene	50.000	50.791	-1.6	105	0.00
87 T	1,3-Dichlorobenzene	50.000	48.501	3.0	103	0.00
88 T	1,4-Dichlorobenzene	50.000	46.906	6.2	104	0.00
89 T	n-Butylbenzene	50.000	50.829	-1.7	105	0.00
90 T	Hexachloroethane	50.000	49.161	1.7	102	0.00
91 T	1,2-Dichlorobenzene	50.000	49.672	0.7	104	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	53.151	-6.3	106	0.00
93 T	1,2,4-Trichlorobenzene	50.000	49.369	1.3	104	0.00
94 T	Hexachlorobutadiene	50.000	51.729	-3.5	108	0.00
95 T	Naphthalene	50.000	52.462	-4.9	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
Data File : VX046726.D
Acq On : 17 Jun 2025 18:00
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725

Quant Time: Jun 18 07:50:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 03:09:16 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	51.200	-2.4	107	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q2361 SAS No.: Q2361 SDG No.: Q2361

Instrument ID: MSVOA_X Calibration Date/Time: 06/19/2025 09:43

Lab File ID: VX046758.D Init. Calib. Date(s): 06/17/2025 06/17/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:19 17:18

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.534	0.565		5.8	20
Chloromethane	0.576	0.595	0.1	3.3	20
Vinyl Chloride	0.614	0.641		4.4	20
Bromomethane	0.346	0.384		10.98	20
Chloroethane	0.372	0.387		4.03	20
Trichlorofluoromethane	0.933	0.957		2.57	20
1,1,2-Trichlorotrifluoroethane	0.572	0.592		3.5	20
1,1-Dichloroethene	0.552	0.572		3.62	20
Acetone	0.231	0.185		-19.91	20
Carbon Disulfide	1.668	1.595		-4.38	20
Methyl tert-butyl Ether	1.720	1.569		-8.78	20
Methyl Acetate	0.586	0.480		-18.09	20
Methylene Chloride	0.641	0.618		-3.59	20
trans-1,2-Dichloroethene	0.591	0.587		-0.68	20
1,1-Dichloroethane	1.092	1.081	0.1	-1.01	20
Cyclohexane	1.036	1.018		-1.74	20
2-Butanone	0.326	0.263		-19.33	20
Carbon Tetrachloride	0.518	0.515		-0.58	20
cis-1,2-Dichloroethene	0.694	0.686		-1.15	20
Bromochloromethane	0.495	0.476		-3.84	20
Chloroform	1.092	1.093		0.09	20
1,1,1-Trichloroethane	0.958	0.939		-1.98	20
Methylcyclohexane	0.628	0.627		-0.32	20
Benzene	1.413	1.415		0.14	20
1,2-Dichloroethane	0.497	0.481		-3.22	20
Trichloroethene	0.360	0.357		-0.83	20
1,2-Dichloropropane	0.350	0.353		0.86	20
Bromodichloromethane	0.514	0.521		1.36	20
4-Methyl-2-Pentanone	0.411	0.339		-17.52	20
Toluene	0.876	0.877		0.11	20
t-1,3-Dichloropropene	0.477	0.481		0.84	20
cis-1,3-Dichloropropene	0.541	0.545		0.74	20
1,1,2-Trichloroethane	0.327	0.311		-4.89	20
2-Hexanone	0.284	0.229		-19.37	20
Dibromochloromethane	0.385	0.377		-2.08	20
1,2-Dibromoethane	0.332	0.317		-4.52	20
Tetrachloroethene	0.346	0.344		-0.58	20
Chlorobenzene	1.104	1.089	0.3	-1.36	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: Q2361 SAS No.: Q2361 SDG No.: Q2361
 Instrument ID: MSVOA_X Calibration Date/Time: 06/19/2025 09:43
 Lab File ID: VX046758.D Init. Calib. Date(s): 06/17/2025 06/17/2025
 Heated Purge: (Y/N) N Init. Calib. Time(s): 11:19 17:18
 GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.929	1.918		-0.57	20
m/p-Xylenes	0.719	0.724		0.69	20
o-Xylene	0.673	0.687		2.08	20
Styrene	1.179	1.192		1.1	20
Bromoform	0.282	0.263	0.1	-6.74	20
Isopropylbenzene	3.682	3.747		1.76	20
1,1,2,2-Tetrachloroethane	1.028	0.933	0.3	-9.24	20
1,3-Dichlorobenzene	1.728	1.670		-3.36	20
1,4-Dichlorobenzene	1.775	1.693		-4.62	20
1,2-Dichlorobenzene	1.615	1.598		-1.05	20
1,2-Dibromo-3-Chloropropane	0.203	0.164		-19.21	20
1,2,4-Trichlorobenzene	1.148	1.111		-3.22	20
1,2,3-Trichlorobenzene	1.098	1.064		-3.1	20
1,2-Dichloroethane-d4	0.692	0.639		-7.66	20
Dibromofluoromethane	0.331	0.327		-1.21	20
Toluene-d8	1.189	1.166		-1.93	20
4-Bromofluorobenzene	0.443	0.433		-2.26	20

All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
 Data File : VX046758.D
 Acq On : 19 Jun 2025 09:43
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/20/2025
 Supervised By : Mahesh Dadoda 06/21/2025

Quant Time: Jun 20 05:14:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	137182	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	224265	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	197740	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	96867	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	87638	46.187	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	92.380%		
35) Dibromofluoromethane	5.391	113	73413	49.476	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	98.960%		
50) Toluene-d8	8.647	98	261576	49.067	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	98.140%		
62) 4-Bromofluorobenzene	11.079	95	97021	48.823	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	97.640%		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.185	85	77571	52.964	ug/l	99
3) Chloromethane	1.307	50	81669	51.662	ug/l	96
4) Vinyl Chloride	1.386	62	87898	52.141	ug/l	96
5) Bromomethane	1.624	94	52745	55.612	ug/l	94
6) Chloroethane	1.703	64	53086	51.957	ug/l	98
7) Trichlorofluoromethane	1.904	101	131261	51.280	ug/l	99
8) Diethyl Ether	2.142	74	44511	50.670	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.343	101	81171	51.691	ug/l	98
10) Methyl Iodide	2.465	142	94903	48.429	ug/l	99
11) Tert butyl alcohol	2.947	59	26783	157.262	ug/l	100
12) 1,1-Dichloroethene	2.337	96	78427	51.759	ug/l	97
13) Acrolein	2.245	56	50116	202.026	ug/l	98
14) Allyl chloride	2.678	41	136507	49.722	ug/l	99
15) Acrylonitrile	3.068	53	162714	212.474	ug/l	99
16) Acetone	2.380	43	127043	200.776	ug/l	99
17) Carbon Disulfide	2.526	76	218810	47.820	ug/l	99
18) Methyl Acetate	2.709	43	65851	40.975	ug/l	99
19) Methyl tert-butyl Ether	3.117	73	215275	45.628	ug/l	100
20) Methylene Chloride	2.800	84	84742	48.158	ug/l	99
21) trans-1,2-Dichloroethene	3.105	96	80577	49.698	ug/l	96
22) Diisopropyl ether	3.763	45	269978	51.162	ug/l	97
23) Vinyl Acetate	3.727	43	1003345	239.105	ug/l	100
24) 1,1-Dichloroethane	3.623	63	148260	49.484	ug/l	100
25) 2-Butanone	4.556	43	180589m	201.795	ug/l	
26) 2,2-Dichloropropane	4.483	77	112913	51.672	ug/l	99
27) cis-1,2-Dichloroethene	4.495	96	94126	49.415	ug/l	99
28) Bromochloromethane	4.904	49	65263	48.047	ug/l	100
29) Tetrahydrofuran	5.007	42	116160	200.958	ug/l	99
30) Chloroform	5.099	83	149922	50.036	ug/l	95
31) Cyclohexane	5.477	56	139626	49.142	ug/l	97
32) 1,1,1-Trichloroethane	5.385	97	128769	49.003	ug/l	100
36) 1,1-Dichloropropene	5.702	75	106011	49.493	ug/l	99
37) Ethyl Acetate	4.715	43	79324	40.121	ug/l	99
38) Carbon Tetrachloride	5.684	117	115565	49.715	ug/l	98
39) Methylcyclohexane	7.379	83	140502	49.848	ug/l	96
40) Benzene	6.044	78	317423	50.074	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
 Data File : VX046758.D
 Acq On : 19 Jun 2025 09:43
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Quant Time: Jun 20 05:14:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/20/2025
 Supervised By : Mahesh Dadoda 06/21/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	47871	45.417	ug/l	95
42) 1,2-Dichloroethane	6.086	62	107931	48.452	ug/l	99
43) Isopropyl Acetate	6.336	43	132833	42.935	ug/l	99
44) Trichloroethene	7.129	130	80127	49.599	ug/l	96
45) 1,2-Dichloropropane	7.427	63	79156	50.385	ug/l	95
46) Dibromomethane	7.580	93	53399	47.295	ug/l	99
47) Bromodichloromethane	7.818	83	116822	50.703	ug/l	100
48) Methyl methacrylate	7.690	41	70314	45.316	ug/l	99
49) 1,4-Dioxane	7.659	88	15950	714.909	ug/l	96
51) 4-Methyl-2-Pentanone	8.568	43	379704	205.777	ug/l	100
52) Toluene	8.714	92	196587	50.025	ug/l	97
53) t-1,3-Dichloropropene	8.976	75	107788	50.420	ug/l	99
54) cis-1,3-Dichloropropene	8.366	75	122151	50.333	ug/l	99
55) 1,1,2-Trichloroethane	9.147	97	69729	47.614	ug/l	98
56) Ethyl methacrylate	9.116	69	99225	45.899	ug/l	100
57) 1,3-Dichloropropane	9.305	76	122083	48.408	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.238	63	279394	248.797	ug/l	100
59) 2-Hexanone	9.427	43	257015	201.475	ug/l	100
60) Dibromochloromethane	9.519	129	84603	49.021	ug/l	99
61) 1,2-Dibromoethane	9.610	107	70987	47.653	ug/l	100
64) Tetrachloroethene	9.275	164	67934	49.587	ug/l	98
65) Chlorobenzene	10.073	112	215423	49.352	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	73797	50.304	ug/l	100
67) Ethyl Benzene	10.189	91	379202	49.719	ug/l	100
68) m/p-Xylenes	10.299	106	286180	100.703	ug/l	99
69) o-Xylene	10.640	106	135837	51.004	ug/l	99
70) Styrene	10.652	104	235709	50.573	ug/l	100
71) Bromoform	10.799	173	51917	46.470	ug/l #	100
73) Isopropylbenzene	10.957	105	362998	50.881	ug/l	100
74) N-amyl acetate	10.841	43	120822	44.699	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	90340	45.350	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	74859m	42.194	ug/l	
77) Bromobenzene	11.195	156	86326	50.284	ug/l	98
78) n-propylbenzene	11.299	91	434668	51.296	ug/l	100
79) 2-Chlorotoluene	11.360	91	250583	49.529	ug/l	99
80) 1,3,5-Trimethylbenzene	11.445	105	294565	50.244	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	27012	43.036	ug/l	98
82) 4-Chlorotoluene	11.451	91	294491	49.875	ug/l	100
83) tert-Butylbenzene	11.713	119	302344	49.891	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	298590	50.692	ug/l	100
85) sec-Butylbenzene	11.890	105	389381	50.841	ug/l	99
86) p-Isopropyltoluene	12.006	119	324490	50.214	ug/l	100
87) 1,3-Dichlorobenzene	11.963	146	161807	48.338	ug/l	100
88) 1,4-Dichlorobenzene	12.036	146	164023	47.705	ug/l	98
89) n-Butylbenzene	12.329	91	310154	50.550	ug/l	100
90) Hexachloroethane	12.536	117	55264	48.837	ug/l	96
91) 1,2-Dichlorobenzene	12.329	146	154832	49.496	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	15868	40.348	ug/l	97
93) 1,2,4-Trichlorobenzene	13.585	180	107660	48.410	ug/l	98
94) Hexachlorobutadiene	13.719	225	46034	49.207	ug/l	99
95) Naphthalene	13.774	128	286491	45.216	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	103088	48.453	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
Data File : VX046758.D
Acq On : 19 Jun 2025 09:43
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025

Quant Time: Jun 20 05:14:38 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 03:09:16 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Manual Integrations


Quant Time: Jun 20 05:14:38 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 03:09:16 2025
Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion:168 Resp: 13718
Ion Ratio Lower Upper
168 100
99 59.5 48.5 72.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#2
Dichlorodifluoromethane
Concen: 52.964 ug/l
RT: 1.185 min Scan# 16
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Tgt Ion: 85 Resp: 77571
Ion Ratio Lower Upper
85 100
87 32.7 16.6 49.8









#3
Chloromethane
Concen: 51.662 ug/l
RT: 1.307 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion: 50 Resp: 81669
Ion Ratio Lower Upper
50 100
52 33.4 24.9 37.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#4
Vinyl Chloride
Concen: 52.141 ug/l
RT: 1.386 min Scan# 49
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Tgt Ion: 62 Resp: 87898
Ion Ratio Lower Upper
62 100
64 31.8 23.9 35.9









#5
Bromomethane
Concen: 55.612 ug/l
RT: 1.624 min Scan# 88
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion: 94 Resp: 52745
Ion Ratio Lower Upper
94 100
96 96.1 72.3 108.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#6
Chloroethane
Concen: 51.957 ug/l
RT: 1.703 min Scan# 101
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43


Tgt Ion: 64 Resp: 53086
Ion Ratio Lower Upper
64 100
66 32.8 27.1 40.7





#7
 Trichlorofluoromethane
 Concen: 51.280 ug/l
 RT: 1.904 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: VX046758.D
 Acq: 19 Jun 2025 09:43




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050




Tgt Ion:101 Resp: 13126
 Ion Ratio Lower Upper
 101 100
 103 65.7 51.8 77.8

Manual Integrations
 APPROVED



Reviewed By :John Carlone 06/20/2025
 Supervised By :Mahesh Dadoda 06/21/2025




#8
 Diethyl Ether
 Concen: 50.670 ug/l
 RT: 2.142 min Scan# 173
 Delta R.T. -0.006 min
 Lab File: VX046758.D
 Acq: 19 Jun 2025 09:43



Tgt Ion: 74 Resp: 44511
 Ion Ratio Lower Upper
 74 100
 45 103.6 52.3 157.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 51.691 ug/l
RT: 2.343 min Scan# 206
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion:101 Resp: 8117
Ion Ratio Lower Upper
101 100
85 42.5 35.4 53.0
151 74.7 61.2 91.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#10
Methyl Iodide
Concen: 48.429 ug/l
RT: 2.465 min Scan# 226
Delta R.T. -0.006 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Tgt Ion:142 Resp: 94903
Ion Ratio Lower Upper
142 100
127 45.2 37.0 55.6
141 14.2 11.5 17.3









#11
Tert butyl alcohol
Concen: 157.262 ug/l
RT: 2.947 min Scan# 307
Delta R.T. -0.012 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion: 59 Resp: 2678
Ion Ratio Lower Upper
59 100
57 10.6 8.3 12.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#12
1,1-Dichloroethene
Concen: 51.759 ug/l
RT: 2.337 min Scan# 205
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43


Tgt Ion: 96 Resp: 78427
Ion Ratio Lower Upper
96 100
61 163.0 126.6 189.8
98 64.7 50.7 76.1





#13
Acrolein
Concen: 202.026 ug/l
RT: 2.245 min Scan# 190
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion: 56 Resp: 50110
Ion Ratio Lower Upper
56 100
55 71.0 55.2 82.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#14
Allyl chloride
Concen: 49.722 ug/l
RT: 2.678 min Scan# 261
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43


Tgt Ion: 41 Resp: 136507
Ion Ratio Lower Upper
41 100
39 66.2 53.6 80.4
76 33.3 26.8 40.2





#15
Acrylonitrile
Concen: 212.474 ug/l
RT: 3.068 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion: 53 Resp: 162714
Ion Ratio Lower Upper
53 100
52 82.0 65.1 97.7
51 35.9 28.9 43.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#16
Acetone
Concen: 200.776 ug/l
RT: 2.380 min Scan# 212
Delta R.T. -0.006 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Tgt Ion: 43 Resp: 127043
Ion Ratio Lower Upper
43 100
58 30.2 24.5 36.7









#17
Carbon Disulfide
Concen: 47.820 ug/l
RT: 2.526 min Scan# 218810
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion: 76 Resp: 218810
Ion Ratio Lower Upper
76 100
78 9.1 7.5 11.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#18
Methyl Acetate
Concen: 40.975 ug/l
RT: 2.709 min Scan# 266
Delta R.T. -0.006 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43


Tgt Ion: 43 Resp: 65851
Ion Ratio Lower Upper
43 100
74 22.4 18.4 27.6





#19
Methyl tert-butyl Ether
Concen: 45.628 ug/l
RT: 3.117 min Scan# 311
Delta R.T. -0.006 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion: 73 Resp: 21527
Ion Ratio Lower Upper
73 100
57 22.7 18.2 27.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#20
Methylene Chloride
Concen: 48.158 ug/l
RT: 2.800 min Scan# 281
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43


Tgt Ion: 84 Resp: 84742
Ion Ratio Lower Upper
84 100
49 127.5 101.3 151.9
51 37.8 31.0 46.6
86 64.8 51.1 76.7





#21
trans-1,2-Dichloroethene
Concen: 49.698 ug/l
RT: 3.105 min Scan# 311
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion: 96 Resp: 8057
Ion Ratio Lower Upper
96 100
61 134.5 112.5 168.7
98 63.6 51.9 77.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#22
Diisopropyl ether
Concen: 51.162 ug/l
RT: 3.763 min Scan# 439
Delta R.T. -0.006 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43


Tgt Ion: 45 Resp: 269978
Ion Ratio Lower Upper
45 100
43 110.7 91.9 137.9
87 25.9 22.0 33.0
59 11.1 9.1 13.7





#23
 Vinyl Acetate
 Concen: 239.105 ug/l
 RT: 3.727 min Scan# 41
 Delta R.T. -0.006 min
 Lab File: VX046758.D
 Acq: 19 Jun 2025 09:43




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050




Tgt Ion: 43 Resp: 100334
 Ion Ratio Lower Upper
 43 100
 86 10.2 8.2 12.2

Manual Integrations
 APPROVED



Reviewed By :John Carlone 06/20/2025
 Supervised By :Mahesh Dadoda 06/21/2025




#24
 1,1-Dichloroethane
 Concen: 49.484 ug/l
 RT: 3.623 min Scan# 416
 Delta R.T. -0.000 min
 Lab File: VX046758.D
 Acq: 19 Jun 2025 09:43



Tgt Ion: 63 Resp: 148260
 Ion Ratio Lower Upper
 63 100
 98 7.8 4.0 11.9
 100 4.6 2.2 6.6





#25

2-Butanone

Concen: 201.795 ug/l m

RT: 4.556 min Scan# 5

Delta R.T. -0.006 min

Lab File: VX046758.D

Acq: 19 Jun 2025 09:43

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 180589

Ion Ratio Lower Upper

43 100





72 23.5 20.2 30.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#26

2,2-Dichloropropane

Concen: 51.672 ug/l

RT: 4.483 min Scan# 557

Delta R.T. -0.006 min

Lab File: VX046758.D




Acq: 19 Jun 2025 09:43


Tgt Ion: 77 Resp: 112913

Ion Ratio Lower Upper

77 100

97 22.8 11.8 35.3





#27
 cis-1,2-Dichloroethene
 Concen: 49.415 ug/l
 RT: 4.495 min Scan# 51
 Delta R.T. -0.006 min
 Lab File: VX046758.D
 Acq: 19 Jun 2025 09:43

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion: 96 Resp: 94120
 Ion Ratio Lower Upper
 96 100
 61 145.1 0.0 291.0
 98 65.4 0.0 126.8

Manual Integrations




APPROVED


Reviewed By :John Carlone 06/20/2025
 Supervised By :Mahesh Dadoda 06/21/2025



#28
 Bromochloromethane
 Concen: 48.047 ug/l
 RT: 4.904 min Scan# 626
 Delta R.T. -0.006 min
 Lab File: VX046758.D
 Acq: 19 Jun 2025 09:43

Tgt Ion: 49 Resp: 65263
 Ion Ratio Lower Upper
 49 100
 129 1.9 0.0 3.8
 130 76.8 61.4 92.0





#29

Tetrahydrofuran

Concen: 200.958 ug/l

RT: 5.007 min Scan# 644

Delta R.T. -0.006 min

Lab File: VX046758.D

Acq: 19 Jun 2025 09:43

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 42 Resp: 116160

Ion Ratio Lower Upper

42 100

72 42.9 34.9 52.3





71 39.1 32.2 48.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#30

Chloroform

Concen: 50.036 ug/l

RT: 5.099 min Scan# 658

Delta R.T. -0.006 min

Lab File: VX046758.D




Acq: 19 Jun 2025 09:43


Tgt Ion: 83 Resp: 149922

Ion Ratio Lower Upper

83 100


85 63.2 53.7 80.5





#31
Cyclohexane
Concen: 49.142 ug/l
RT: 5.477 min Scan# 719
Delta R.T. -0.006 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion: 56 Resp: 139620
Ion Ratio Lower Upper
56 100
69 29.6 24.8 37.2
84 82.9 68.2 102.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#32
1,1,1-Trichloroethane
Concen: 49.003 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.012 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43


Tgt Ion: 97 Resp: 128769
Ion Ratio Lower Upper
97 100
99 64.0 51.3 76.9
61 45.4 36.4 54.6





#33
1,2-Dichloroethane-d4
Concen: 46.187 ug/l
RT: 5.958 min Scan# 799
Delta R.T. -0.006 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050




Tgt Ion: 65 Resp: 87638
Ion Ratio Lower Upper
65 100
67 53.3 0.0 105.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025




#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.763 min Scan# 931
Delta R.T. -0.006 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Tgt Ion:114 Resp: 224265
Ion Ratio Lower Upper
114 100
63 20.7 0.0 44.2
88 16.2 0.0 33.2





#35
Dibromofluoromethane
Concen: 49.476 ug/l
RT: 5.391 min Scan# 707
Delta R.T. -0.006 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion: 113 Resp: 7341
Ion Ratio Lower Upper
113 100
111 102.5 82.0 123.0
192 18.6 15.3 22.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#36
1,1-Dichloropropene
Concen: 49.493 ug/l
RT: 5.702 min Scan# 757
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43


Tgt Ion: 75 Resp: 106011
Ion Ratio Lower Upper
75 100
110 34.8 17.2 51.5
77 31.2 24.2 36.4





#37
Ethyl Acetate
Concen: 40.121 ug/l
RT: 4.715 min Scan# 595
Delta R.T. -0.006 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion: 43 Resp: 79324
Ion Ratio Lower Upper
43 100
61 11.6 9.7 14.5
70 10.2 8.1 12.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#38
Carbon Tetrachloride
Concen: 49.715 ug/l
RT: 5.684 min Scan# 754
Delta R.T. -0.006 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Tgt Ion:117 Resp: 115565
Ion Ratio Lower Upper
117 100
119 92.4 75.2 112.8
121 32.1 24.1 36.1





#39
Methylcyclohexane
Concen: 49.848 ug/l
RT: 7.379 min Scan# 1032
Delta R.T. -0.006 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion: 83 Resp: 140502
Ion Ratio Lower Upper
83 100
55 82.7 63.0 94.4
98 50.0 38.8 58.2

Manual Integrations




APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#40
Benzene
Concen: 50.074 ug/l
RT: 6.044 min Scan# 813
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Tgt Ion: 78 Resp: 317423
Ion Ratio Lower Upper
78 100
77 23.7 18.8 28.2






#41
Methacrylonitrile
Concen: 45.417 ug/l
RT: 4.922 min Scan# 61
Delta R.T. -0.006 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050






Tgt Ion: 41 Resp: 4787
Ion Ratio Lower Upper
41 100
39 54.6 45.9 68.9
67 70.2 61.2 91.8
52 31.4 25.5 38.3

Manual Integrations

APPROVED




Reviewed By :John Carlone 06/20/2025


Supervised By :Mahesh Dadoda 06/21/2025



#42
1,2-Dichloroethane
Concen: 48.452 ug/l
RT: 6.086 min Scan# 820
Delta R.T. -0.006 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43


Tgt Ion: 62 Resp: 107931
Ion Ratio Lower Upper
62 100
98 9.3 0.0 17.8





#43
Isopropyl Acetate
Concen: 42.935 ug/l
RT: 6.336 min Scan# 80
Delta R.T. -0.006 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion: 43 Resp: 13283
Ion Ratio Lower Upper
43 100
61 19.4 15.3 22.9
87 12.3 10.0 15.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#44
Trichloroethene
Concen: 49.599 ug/l
RT: 7.129 min Scan# 991
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Tgt Ion:130 Resp: 80127
Ion Ratio Lower Upper
130 100
95 97.5 0.0 202.6





#45

1,2-Dichloropropane

Concen: 50.385 ug/l

RT: 7.427 min Scan# 1040

Delta R.T. -0.006 min

Lab File: VX046758.D

Acq: 19 Jun 2025 09:43

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 63 Resp: 79150

Ion Ratio Lower Upper

63 100





65 31.3 22.9 34.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#46

Dibromomethane

Concen: 47.295 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. -0.006 min

Lab File: VX046758.D

Acq: 19 Jun 2025 09:43




Tgt Ion: 93 Resp: 53399


Ion Ratio Lower Upper

93 100

95 83.4 66.2 99.4


174 98.2 77.0 115.6





#47
Bromodichloromethane
Concen: 50.703 ug/l
RT: 7.818 min Scan# 1104
Delta R.T. -0.006 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion: 83 Resp: 11682
Ion Ratio Lower Upper
83 100
85 64.1 51.4 77.2
127 8.3 6.9 10.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#48
Methyl methacrylate
Concen: 45.316 ug/l
RT: 7.690 min Scan# 1083
Delta R.T. -0.006 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43


Tgt Ion: 41 Resp: 70314
Ion Ratio Lower Upper
41 100
69 79.5 63.4 95.2
39 55.5 45.4 68.0





#49
1,4-Dioxane
Concen: 714.909 ug/l
RT: 7.659 min Scan# 1078
Delta R.T. 0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion: 88 Resp: 15950
Ion Ratio Lower Upper
88 100
43 37.3 31.2 46.8
58 72.4 54.7 82.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#50
Toluene-d8
Concen: 49.067 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Tgt Ion: 98 Resp: 261576
Ion Ratio Lower Upper
98 100
100 67.6 53.0 79.4









#51
4-Methyl-2-Pentanone
Concen: 205.777 ug/l
RT: 8.568 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion: 43 Resp: 379704
Ion Ratio Lower Upper
43 100
58 38.4 30.8 46.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#52
Toluene
Concen: 50.025 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. -0.006 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Tgt Ion: 92 Resp: 196587
Ion Ratio Lower Upper
92 100
91 171.9 134.8 202.2





#53

t-1,3-Dichloropropene

Concen: 50.420 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046758.D

Acq: 19 Jun 2025 09:43

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 75 Resp: 10778

Ion Ratio Lower Upper

75 100





77 31.3 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#54

cis-1,3-Dichloropropene

Concen: 50.333 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046758.D

Acq: 19 Jun 2025 09:43




Tgt Ion: 75 Resp: 122151


Ion Ratio Lower Upper

75 100

77 31.5 25.3 37.9

39 50.5 39.2 58.8






#55
1,1,2-Trichloroethane
Concen: 47.614 ug/l
RT: 9.147 min Scan# 1323
Delta R.T. -0.006 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050






Tgt Ion	Ratio	Lower	Upper
97	100		
83	86.3	67.2	100.8
85	55.4	44.1	66.1
99	65.3	50.3	75.5

Manual Integrations


APPROVED

Reviewed By :John Carlone 06/20/2025



Supervised By :Mahesh Dadoda 06/21/2025




#56
Ethyl methacrylate
Concen: 45.899 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43



Tgt Ion	Ratio	Lower	Upper
69	100		
41	69.7	55.4	83.2
39	37.1	29.7	44.5





#57

1,3-Dichloropropane

Concen: 48.408 ug/l

RT: 9.305 min Scan# 11208

Delta R.T. -0.000 min

Lab File: VX046758.D

Acq: 19 Jun 2025 09:43

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 76 Resp: 12208

Ion Ratio Lower Upper

76 100





78 32.4 25.7 38.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#58

2-Chloroethyl Vinyl ether

Concen: 248.797 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. -0.006 min

Lab File: VX046758.D




Acq: 19 Jun 2025 09:43


Tgt Ion: 63 Resp: 279394

Ion Ratio Lower Upper

63 100

106 28.2 22.6 34.0









#59
2-Hexanone
Concen: 201.475 ug/l
RT: 9.427 min Scan# 1368
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion: 43 Resp: 257015
Ion Ratio Lower Upper
43 100
58 53.4 26.8 80.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#60
Dibromochloromethane
Concen: 49.021 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43


Tgt Ion:129 Resp: 84603
Ion Ratio Lower Upper
129 100
127 77.6 38.4 115.2





#61
1,2-Dibromoethane
Concen: 47.653 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion:107 Resp: 7098
Ion Ratio Lower Upper
107 100
109 94.8 75.8 113.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#62
4-Bromofluorobenzene
Concen: 48.823 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43


Tgt Ion: 95 Resp: 97021
Ion Ratio Lower Upper
95 100
174 75.1 0.0 150.4
176 73.0 0.0 145.0





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. -0.006 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion:117 Resp: 197740
Ion Ratio Lower Upper
117 100
82 57.8 44.6 66.8
119 32.3 25.8 38.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#64
Tetrachloroethene
Concen: 49.587 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Tgt Ion:164 Resp: 67934
Ion Ratio Lower Upper
164 100
166 128.0 104.6 157.0
129 95.0 76.3 114.5
131 91.1 74.1 111.1









#65
Chlorobenzene
Concen: 49.352 ug/l
RT: 10.073 min Scan# 1475
Delta R.T. -0.006 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion:112 Resp: 21542
Ion Ratio Lower Upper
112 100
114 32.4 25.5 38.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 50.304 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43


Tgt Ion:131 Resp: 73797
Ion Ratio Lower Upper
131 100
133 94.0 47.3 141.8
119 64.5 32.1 96.3





#67
Ethyl Benzene
Concen: 49.719 ug/l
RT: 10.189 min Scan# 1493
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion: 91 Resp: 37920
Ion Ratio Lower Upper
91 100
106 30.4 24.2 36.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#68
m/p-Xylenes
Concen: 100.703 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Tgt Ion:106 Resp: 286180
Ion Ratio Lower Upper
106 100
91 206.9 164.8 247.2









#69
o-Xylene
Concen: 51.004 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion:106 Resp: 13583
Ion Ratio Lower Upper
106 100
91 220.3 111.3 333.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#70
Styrene
Concen: 50.573 ug/l
RT: 10.652 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Tgt Ion:104 Resp: 235709
Ion Ratio Lower Upper
104 100
78 51.9 41.4 62.2
103 54.3 43.1 64.7





#71

Bromoform

Concen: 46.470 ug/l

RT: 10.799 min Scan# 1593

Delta R.T. -0.000 min

Lab File: VX046758.D

Acq: 19 Jun 2025 09:43

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion:173	Resp: 5191
Ion Ratio	Lower Upper
173 100	
175 48.4	24.1 72.4
254 0.0	0.1 0.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 12.018 min Scan# 1793


Delta R.T. -0.000 min

Lab File: VX046758.D

Acq: 19 Jun 2025 09:43


Tgt Ion:152	Resp: 96867
Ion Ratio	Lower Upper
152 100	
115 85.5	43.2 129.6
150 175.8	0.0 346.8





#73
Isopropylbenzene
Concen: 50.881 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050




Tgt Ion:105 Resp: 362998
Ion Ratio Lower Upper
105 100
120 26.2 13.0 39.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025




#74
N-ethyl acetate
Concen: 44.699 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43



Tgt Ion: 43 Resp: 120822
Ion Ratio Lower Upper
43 100
70 41.0 33.4 50.0
55 23.7 19.1 28.7
61 23.4 19.0 28.4









#75
1,1,2,2-Tetrachloroethane
Concen: 45.350 ug/l
RT: 11.207 min Scan# 1660
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion: 83 Resp: 90340
Ion Ratio Lower Upper
83 100
131 10.5 4.9 14.6
85 63.3 32.0 96.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#76
1,2,3-Trichloropropane
Concen: 42.194 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43


Tgt Ion: 75 Resp: 74859
Ion Ratio Lower Upper
75 100
77 40.2 18.5 55.5





#77
Bromobenzene
Concen: 50.284 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion:156 Resp: 86320
Ion Ratio Lower Upper
156 100
77 167.7 82.8 248.6
158 98.3 47.6 142.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#78
n-propylbenzene
Concen: 51.296 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43


Tgt Ion: 91 Resp: 434668
Ion Ratio Lower Upper
91 100
120 22.5 11.4 34.1





#79
2-Chlorotoluene
Concen: 49.529 ug/l
RT: 11.360 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050




Tgt Ion: 91 Resp: 250583
Ion Ratio Lower Upper
91 100
126 33.6 16.6 49.7

Manual Integrations
APPROVED



Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025




#80
1,3,5-Trimethylbenzene
Concen: 50.244 ug/l
RT: 11.445 min Scan# 1699
Delta R.T. -0.006 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43



Tgt Ion:105 Resp: 294565
Ion Ratio Lower Upper
105 100
120 49.1 24.1 72.2









#81
trans-1,4-Dichloro-2-butene
Concen: 43.036 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion: 75 Resp: 27012
Ion Ratio Lower Upper
75 100
53 109.1 86.8 130.2
89 50.3 36.9 55.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#82
4-Chlorotoluene
Concen: 49.875 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43


Tgt Ion: 91 Resp: 294491
Ion Ratio Lower Upper
91 100
126 29.0 14.4 43.2





#83
tert-Butylbenzene
Concen: 49.891 ug/l
RT: 11.713 min Scan# 1174
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050




Tgt Ion:119 Resp: 302344
Ion Ratio Lower Upper
119 100
91 58.1 29.5 88.6
134 23.0 11.7 35.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025




#84
1,2,4-Trimethylbenzene
Concen: 50.692 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Tgt Ion:105 Resp: 298590
Ion Ratio Lower Upper
105 100
120 44.1 22.1 66.3





#85
 sec-Butylbenzene
 Concen: 50.841 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. -0.000 min
 Lab File: VX046758.D
 Acq: 19 Jun 2025 09:43




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050




Tgt Ion:105 Resp: 38938
 Ion Ratio Lower Upper
 105 100
 134 19.9 9.8 29.4

Manual Integrations
 APPROVED



Reviewed By :John Carlone 06/20/2025
 Supervised By :Mahesh Dadoda 06/21/2025




#86
 p-Isopropyltoluene
 Concen: 50.214 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. -0.000 min
 Lab File: VX046758.D
 Acq: 19 Jun 2025 09:43



Tgt Ion:119 Resp: 324490
 Ion Ratio Lower Upper
 119 100
 134 25.7 12.9 38.7
 91 24.3 12.3 36.9





#87

1,3-Dichlorobenzene

Concen: 48.338 ug/l

RT: 11.963 min Scan# 1784

Delta R.T. -0.006 min

Lab File: VX046758.D

Acq: 19 Jun 2025 09:43

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion:146 Resp: 16180

Ion Ratio Lower Upper

146 100

111 41.3 20.6 61.8





148 62.8 31.6 94.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#88

1,4-Dichlorobenzene

Concen: 47.705 ug/l

RT: 12.036 min Scan# 1796

Delta R.T. -0.000 min

Lab File: VX046758.D

Acq: 19 Jun 2025 09:43




Tgt Ion:146 Resp: 164023


Ion Ratio Lower Upper

146 100

111 39.5 20.3 61.0


148 62.4 32.1 96.5





#89
n-Butylbenzene
Concen: 50.550 ug/l
RT: 12.329 min Scan# 1844
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion: 91 Resp: 310154
Ion Ratio Lower Upper
91 100
92 54.1 27.1 81.3
134 25.9 12.9 38.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#90
Hexachloroethane
Concen: 48.837 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43


Tgt Ion:117 Resp: 55264
Ion Ratio Lower Upper
117 100
201 71.8 34.4 103.3





#91
1,2-Dichlorobenzene
Concen: 49.496 ug/l
RT: 12.329 min Scan# 1844
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050




Tgt Ion:146 Resp: 15483
Ion Ratio Lower Upper
146 100
111 42.3 21.7 65.1
148 63.1 32.0 96.2

Manual Integrations
APPROVED



Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 40.348 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43



Tgt Ion: 75 Resp: 15868
Ion Ratio Lower Upper
75 100
155 84.4 41.9 125.7
157 107.9 51.4 154.3





#93

1,2,4-Trichlorobenzene

Concen: 48.410 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. -0.000 min

Lab File: VX046758.D

Acq: 19 Jun 2025 09:43

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion:180 Resp: 107660

Ion Ratio Lower Upper

180 100

182 95.9 46.7 140.1





145 32.7 15.8 47.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#94

Hexachlorobutadiene

Concen: 49.207 ug/l

RT: 13.719 min Scan# 2072

Delta R.T. -0.000 min

Lab File: VX046758.D

Acq: 19 Jun 2025 09:43




Tgt Ion:225 Resp: 46034


Ion Ratio Lower Upper

225 100

223 64.3 32.5 97.5


227 63.2 32.2 96.6





#95
Naphthalene
Concen: 45.216 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion:128 Resp: 28649
Ion Ratio Lower Upper
128 100
127 12.8 10.1 15.1
129 10.7 8.7 13.1




Manual Integrations
APPROVED

Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#96
1,2,3-Trichlorobenzene
Concen: 48.453 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Tgt Ion:180 Resp: 103088
Ion Ratio Lower Upper
180 100
182 94.7 47.8 143.4
145 34.7 17.3 51.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
 Data File : VX046758.D
 Acq On : 19 Jun 2025 09:43
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 20 05:14:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	111	0.00
2 T	Dichlorodifluoromethane	0.534	0.565	-5.8	108	0.00
3 P	Chloromethane	0.576	0.595	-3.3	112	0.00
4 C	Vinyl Chloride	0.614	0.641	-4.4#	112	0.00
5 T	Bromomethane	0.346	0.384	-11.0	116	0.00
6 T	Chloroethane	0.372	0.387	-4.0	114	0.00
7 T	Trichlorofluoromethane	0.933	0.957	-2.6	110	0.00
8 T	Diethyl Ether	0.320	0.324	-1.3	112	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.572	0.592	-3.5	114	0.00
10 T	Methyl Iodide	0.642	0.692	-7.8	118	0.00
11 T	Tert butyl alcohol	0.062	0.039	37.1#	70	-0.01
12 CM	1,1-Dichloroethene	0.552	0.572	-3.6#	112	0.00
13 T	Acrolein	0.090	0.073	18.9	93	0.00
14 T	Allyl chloride	1.001	0.995	0.6	111	0.00
15 T	Acrylonitrile	0.279	0.237	15.1	93	0.00
16 T	Acetone	0.231	0.185	19.9	94	0.00
17 T	Carbon Disulfide	1.668	1.595	4.4	111	0.00
18 T	Methyl Acetate	0.586	0.480	18.1	92	0.00
19 T	Methyl tert-butyl Ether	1.720	1.569	8.8	100	0.00
20 T	Methylene Chloride	0.641	0.618	3.6	113	0.00
21 T	trans-1,2-Dichloroethene	0.591	0.587	0.7	111	0.00
22 T	Diisopropyl ether	1.923	1.968	-2.3	111	0.00
23 T	Vinyl Acetate	1.529	1.463	4.3	102	0.00
24 P	1,1-Dichloroethane	1.092	1.081	1.0	110	0.00
25 T	2-Butanone	0.326	0.263	19.3	87	0.00
26 T	2,2-Dichloropropane	0.796	0.823	-3.4	120	0.00
27 T	cis-1,2-Dichloroethene	0.694	0.686	1.2	111	0.00
28 T	Bromochloromethane	0.495	0.476	3.8	109	0.00
29 T	Tetrahydrofuran	0.211	0.169	19.9	87	0.00
30 C	Chloroform	1.092	1.093	-0.1#	109	0.00
31 T	Cyclohexane	1.036	1.018	1.7	112	0.00
32 T	1,1,1-Trichloroethane	0.958	0.939	2.0	109	-0.01
33 S	1,2-Dichloroethane-d4	0.692	0.639	7.7	109	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	110	0.00
35 S	Dibromofluoromethane	0.331	0.327	1.2	112	0.00
36 T	1,1-Dichloropropene	0.478	0.473	1.0	109	0.00
37 T	Ethyl Acetate	0.441	0.354	19.7	90	0.00
38 T	Carbon Tetrachloride	0.518	0.515	0.6	110	0.00
39 T	Methylcyclohexane	0.628	0.626	0.3	109	0.00
40 TM	Benzene	1.413	1.415	-0.1	110	0.00
41 T	Methacrylonitrile	0.235	0.213	9.4	98	0.00
42 TM	1,2-Dichloroethane	0.497	0.481	3.2	107	0.00
43 T	Isopropyl Acetate	0.690	0.592	14.2	91	0.00
44 TM	Trichloroethene	0.360	0.357	0.8	110	0.00
45 C	1,2-Dichloropropane	0.350	0.353	-0.9#	112	0.00
46 T	Dibromomethane	0.252	0.238	5.6	104	0.00
47 T	Bromodichloromethane	0.514	0.521	-1.4	110	0.00
48 T	Methyl methacrylate	0.346	0.314	9.2	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
 Data File : VX046758.D
 Acq On : 19 Jun 2025 09:43
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 20 05:14:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.004	0.004	0.0	81	0.00
50 S	Toluene-d8	1.189	1.166	1.9	112	0.00
51 T	4-Methyl-2-Pentanone	0.411	0.339	17.5	87	0.00
52 CM	Toluene	0.876	0.877	-0.1#	108	0.00
53 T	t-1,3-Dichloropropene	0.477	0.481	-0.8	108	0.00
54 T	cis-1,3-Dichloropropene	0.541	0.545	-0.7	109	0.00
55 T	1,1,2-Trichloroethane	0.327	0.311	4.9	105	0.00
56 T	Ethyl methacrylate	0.482	0.442	8.3	96	0.00
57 T	1,3-Dichloropropane	0.562	0.544	3.2	106	0.00
58 T	2-Chloroethyl Vinyl ether	0.250	0.249	0.4	97	0.00
59 T	2-Hexanone	0.284	0.229	19.4	85	0.00
60 T	Dibromochloromethane	0.385	0.377	2.1	107	0.00
61 T	1,2-Dibromoethane	0.332	0.317	4.5	104	0.00
62 S	4-Bromofluorobenzene	0.443	0.433	2.3	111	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	110	0.00
64 T	Tetrachloroethene	0.346	0.344	0.6	111	0.00
65 PM	Chlorobenzene	1.104	1.089	1.4	110	0.00
66 T	1,1,1,2-Tetrachloroethane	0.371	0.373	-0.5	110	0.00
67 C	Ethyl Benzene	1.929	1.918	0.6#	109	0.00
68 T	m/p-Xylenes	0.719	0.724	-0.7	110	0.00
69 T	o-Xylene	0.673	0.687	-2.1	109	0.00
70 T	Styrene	1.179	1.192	-1.1	109	0.00
71 P	Bromoform	0.282	0.263	6.7	101	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	109	0.00
73 T	Isopropylbenzene	3.682	3.747	-1.8	110	0.00
74 T	N-amyl acetate	1.395	1.247	10.6	93	0.00
75 P	1,1,2,2-Tetrachloroethane	1.028	0.933	9.2	97	0.00
76 T	1,2,3-Trichloropropane	0.916	0.773	15.6	85	0.00
77 T	Bromobenzene	0.886	0.891	-0.6	109	0.00
78 T	n-propylbenzene	4.374	4.487	-2.6	110	0.00
79 T	2-Chlorotoluene	2.611	2.587	0.9	109	0.00
80 T	1,3,5-Trimethylbenzene	3.026	3.041	-0.5	108	0.00
81 T	trans-1,4-Dichloro-2-butene	0.324	0.279	13.9	98	0.00
82 T	4-Chlorotoluene	3.048	3.040	0.3	109	0.00
83 T	tert-Butylbenzene	3.128	3.121	0.2	109	0.00
84 T	1,2,4-Trimethylbenzene	3.040	3.082	-1.4	110	0.00
85 T	sec-Butylbenzene	3.953	4.020	-1.7	109	0.00
86 T	p-Isopropyltoluene	3.336	3.350	-0.4	109	0.00
87 T	1,3-Dichlorobenzene	1.728	1.670	3.4	108	0.00
88 T	1,4-Dichlorobenzene	1.775	1.693	4.6	111	0.00
89 T	n-Butylbenzene	3.167	3.202	-1.1	110	0.00
90 T	Hexachloroethane	0.584	0.571	2.2	106	0.00
91 T	1,2-Dichlorobenzene	1.615	1.598	1.1	109	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.203	0.164	19.2	85	0.00
93 T	1,2,4-Trichlorobenzene	1.148	1.111	3.2	107	0.00
94 T	Hexachlorobutadiene	0.483	0.475	1.7	108	0.00
95 T	Naphthalene	3.270	2.958	9.5	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
Data File : VX046758.D
Acq On : 19 Jun 2025 09:43
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 20 05:14:38 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 03:09:16 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.098	1.064	3.1	107	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
 Data File : VX046758.D
 Acq On : 19 Jun 2025 09:43
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 20 05:14:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	111	0.00
2 T	Dichlorodifluoromethane	50.000	52.964	-5.9	108	0.00
3 P	Chloromethane	50.000	51.662	-3.3	112	0.00
4 C	Vinyl Chloride	50.000	52.141	-4.3#	112	0.00
5 T	Bromomethane	50.000	55.612	-11.2	116	0.00
6 T	Chloroethane	50.000	51.957	-3.9	114	0.00
7 T	Trichlorofluoromethane	50.000	51.280	-2.6	110	0.00
8 T	Diethyl Ether	50.000	50.670	-1.3	112	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	51.691	-3.4	114	0.00
10 T	Methyl Iodide	50.000	48.429	3.1	118	0.00
11 T	Tert butyl alcohol	250.000	157.262	37.1#	70	-0.01
12 CM	1,1-Dichloroethene	50.000	51.759	-3.5#	112	0.00
13 T	Acrolein	250.000	202.026	19.2	93	0.00
14 T	Allyl chloride	50.000	49.722	0.6	111	0.00
15 T	Acrylonitrile	250.000	212.474	15.0	93	0.00
16 T	Acetone	250.000	200.776	19.7	94	0.00
17 T	Carbon Disulfide	50.000	47.820	4.4	111	0.00
18 T	Methyl Acetate	50.000	40.975	18.0	92	0.00
19 T	Methyl tert-butyl Ether	50.000	45.628	8.7	100	0.00
20 T	Methylene Chloride	50.000	48.158	3.7	113	0.00
21 T	trans-1,2-Dichloroethene	50.000	49.698	0.6	111	0.00
22 T	Diisopropyl ether	50.000	51.162	-2.3	111	0.00
23 T	Vinyl Acetate	250.000	239.105	4.4	102	0.00
24 P	1,1-Dichloroethane	50.000	49.484	1.0	110	0.00
25 T	2-Butanone	250.000	201.795	19.3	87	0.00
26 T	2,2-Dichloropropane	50.000	51.672	-3.3	120	0.00
27 T	cis-1,2-Dichloroethene	50.000	49.415	1.2	111	0.00
28 T	Bromochloromethane	50.000	48.047	3.9	109	0.00
29 T	Tetrahydrofuran	250.000	200.958	19.6	87	0.00
30 C	Chloroform	50.000	50.036	-0.1#	109	0.00
31 T	Cyclohexane	50.000	49.142	1.7	112	0.00
32 T	1,1,1-Trichloroethane	50.000	49.003	2.0	109	-0.01
33 S	1,2-Dichloroethane-d4	50.000	46.187	7.6	109	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	110	0.00
35 S	Dibromofluoromethane	50.000	49.476	1.0	112	0.00
36 T	1,1-Dichloropropene	50.000	49.493	1.0	109	0.00
37 T	Ethyl Acetate	50.000	40.121	19.8	90	0.00
38 T	Carbon Tetrachloride	50.000	49.715	0.6	110	0.00
39 T	Methylcyclohexane	50.000	49.848	0.3	109	0.00
40 TM	Benzene	50.000	50.074	-0.1	110	0.00
41 T	Methacrylonitrile	50.000	45.417	9.2	98	0.00
42 TM	1,2-Dichloroethane	50.000	48.452	3.1	107	0.00
43 T	Isopropyl Acetate	50.000	42.935	14.1	91	0.00
44 TM	Trichloroethene	50.000	49.599	0.8	110	0.00
45 C	1,2-Dichloropropane	50.000	50.385	-0.8#	112	0.00
46 T	Dibromomethane	50.000	47.295	5.4	104	0.00
47 T	Bromodichloromethane	50.000	50.703	-1.4	110	0.00
48 T	Methyl methacrylate	50.000	45.316	9.4	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
 Data File : VX046758.D
 Acq On : 19 Jun 2025 09:43
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 20 05:14:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	714.909	28.5#	81	0.00
50 S	Toluene-d8	50.000	49.067	1.9	112	0.00
51 T	4-Methyl-2-Pentanone	250.000	205.777	17.7	87	0.00
52 CM	Toluene	50.000	50.025	-0.0#	108	0.00
53 T	t-1,3-Dichloropropene	50.000	50.420	-0.8	108	0.00
54 T	cis-1,3-Dichloropropene	50.000	50.333	-0.7	109	0.00
55 T	1,1,2-Trichloroethane	50.000	47.614	4.8	105	0.00
56 T	Ethyl methacrylate	50.000	45.899	8.2	96	0.00
57 T	1,3-Dichloropropane	50.000	48.408	3.2	106	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	248.797	0.5	97	0.00
59 T	2-Hexanone	250.000	201.475	19.4	85	0.00
60 T	Dibromochloromethane	50.000	49.021	2.0	107	0.00
61 T	1,2-Dibromoethane	50.000	47.653	4.7	104	0.00
62 S	4-Bromofluorobenzene	50.000	48.823	2.4	111	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	110	0.00
64 T	Tetrachloroethene	50.000	49.587	0.8	111	0.00
65 PM	Chlorobenzene	50.000	49.352	1.3	110	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	50.304	-0.6	110	0.00
67 C	Ethyl Benzene	50.000	49.719	0.6#	109	0.00
68 T	m/p-Xylenes	100.000	100.703	-0.7	110	0.00
69 T	o-Xylene	50.000	51.004	-2.0	109	0.00
70 T	Styrene	50.000	50.573	-1.1	109	0.00
71 P	Bromoform	50.000	46.470	7.1	101	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	109	0.00
73 T	Isopropylbenzene	50.000	50.881	-1.8	110	0.00
74 T	N-amyl acetate	50.000	44.699	10.6	93	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	45.350	9.3	97	0.00
76 T	1,2,3-Trichloropropane	50.000	42.194	15.6	85	0.00
77 T	Bromobenzene	50.000	50.284	-0.6	109	0.00
78 T	n-propylbenzene	50.000	51.296	-2.6	110	0.00
79 T	2-Chlorotoluene	50.000	49.529	0.9	109	0.00
80 T	1,3,5-Trimethylbenzene	50.000	50.244	-0.5	108	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	43.036	13.9	98	0.00
82 T	4-Chlorotoluene	50.000	49.875	0.3	109	0.00
83 T	tert-Butylbenzene	50.000	49.891	0.2	109	0.00
84 T	1,2,4-Trimethylbenzene	50.000	50.692	-1.4	110	0.00
85 T	sec-Butylbenzene	50.000	50.841	-1.7	109	0.00
86 T	p-Isopropyltoluene	50.000	50.214	-0.4	109	0.00
87 T	1,3-Dichlorobenzene	50.000	48.338	3.3	108	0.00
88 T	1,4-Dichlorobenzene	50.000	47.705	4.6	111	0.00
89 T	n-Butylbenzene	50.000	50.550	-1.1	110	0.00
90 T	Hexachloroethane	50.000	48.837	2.3	106	0.00
91 T	1,2-Dichlorobenzene	50.000	49.496	1.0	109	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	40.348	19.3	85	0.00
93 T	1,2,4-Trichlorobenzene	50.000	48.410	3.2	107	0.00
94 T	Hexachlorobutadiene	50.000	49.207	1.6	108	0.00
95 T	Naphthalene	50.000	45.216	9.6	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
Data File : VX046758.D
Acq On : 19 Jun 2025 09:43
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 20 05:14:38 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 03:09:16 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	48.453	3.1	107	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q2361 SAS No.: Q2361 SDG No.: Q2361

Instrument ID: MSVOA_X Calibration Date/Time: 06/19/2025 17:10

Lab File ID: VX046775.D Init. Calib. Date(s): 06/17/2025 06/17/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:19 17:18

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.534	0.573		7.3	50
Chloromethane	0.576	0.601	0.1	4.34	50
Vinyl Chloride	0.614	0.661		7.66	50
Bromomethane	0.346	0.382		10.4	50
Chloroethane	0.372	0.413		11.02	50
Trichlorofluoromethane	0.933	0.983		5.36	50
1,1,2-Trichlorotrifluoroethane	0.572	0.614		7.34	50
1,1-Dichloroethene	0.552	0.592		7.25	50
Acetone	0.231	0.223		-3.46	50
Carbon Disulfide	1.668	1.588		-4.8	50
Methyl tert-butyl Ether	1.720	1.740		1.16	50
Methyl Acetate	0.586	0.608		3.75	50
Methylene Chloride	0.641	0.645		0.62	50
trans-1,2-Dichloroethene	0.591	0.612		3.55	50
1,1-Dichloroethane	1.092	1.144	0.1	4.76	50
Cyclohexane	1.036	1.061		2.41	50
2-Butanone	0.326	0.334		2.45	50
Carbon Tetrachloride	0.518	0.513		-0.96	50
cis-1,2-Dichloroethene	0.694	0.729		5.04	50
Bromochloromethane	0.495	0.513		3.64	50
Chloroform	1.092	1.155		5.77	50
1,1,1-Trichloroethane	0.958	0.973		1.57	50
Methylcyclohexane	0.628	0.645		2.71	50
Benzene	1.413	1.472		4.18	50
1,2-Dichloroethane	0.497	0.505		1.61	50
Trichloroethene	0.360	0.369		2.5	50
1,2-Dichloropropane	0.350	0.363		3.71	50
Bromodichloromethane	0.514	0.529		2.92	50
4-Methyl-2-Pentanone	0.411	0.419		1.95	50
Toluene	0.876	0.902		2.97	50
t-1,3-Dichloropropene	0.477	0.490		2.72	50
cis-1,3-Dichloropropene	0.541	0.565		4.44	50
1,1,2-Trichloroethane	0.327	0.339		3.67	50
2-Hexanone	0.284	0.287		1.06	50
Dibromochloromethane	0.385	0.396		2.86	50
1,2-Dibromoethane	0.332	0.339		2.11	50
Tetrachloroethene	0.346	0.348		0.58	50
Chlorobenzene	1.104	1.134	0.3	2.72	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: Q2361 SAS No.: Q2361 SDG No.: Q2361
 Instrument ID: MSVOA_X Calibration Date/Time: 06/19/2025 17:10
 Lab File ID: VX046775.D Init. Calib. Date(s): 06/17/2025 06/17/2025
 Heated Purge: (Y/N) N Init. Calib. Time(s): 11:19 17:18
 GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.929	1.995		3.42	50
m/p-Xylenes	0.719	0.745		3.62	50
o-Xylene	0.673	0.716		6.39	50
Styrene	1.179	1.253		6.28	50
Bromoform	0.282	0.283	0.1	0.35	50
Isopropylbenzene	3.682	3.814		3.59	50
1,1,2,2-Tetrachloroethane	1.028	1.047	0.3	1.85	50
1,3-Dichlorobenzene	1.728	1.702		-1.5	50
1,4-Dichlorobenzene	1.775	1.686		-5.01	50
1,2-Dichlorobenzene	1.615	1.619		0.25	50
1,2-Dibromo-3-Chloropropane	0.203	0.194		-4.43	50
1,2,4-Trichlorobenzene	1.148	1.143		-0.44	50
1,2,3-Trichlorobenzene	1.098	1.103		0.46	50
1,2-Dichloroethane-d4	0.692	0.648		-6.36	50
Dibromofluoromethane	0.331	0.320		-3.32	50
Toluene-d8	1.189	1.138		-4.29	50
4-Bromofluorobenzene	0.443	0.431		-2.71	50

All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
 Data File : VX046775.D
 Acq On : 19 Jun 2025 17:10
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/20/2025
 Supervised By : Mahesh Dadoda 06/21/2025

Quant Time: Jun 20 05:22:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	129409	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	216626	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	190720	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	95668	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	83815	46.825	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	93.660%		
35) Dibromofluoromethane	5.397	113	69249	48.316	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.640%		
50) Toluene-d8	8.647	98	246552	47.880	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	95.760%		
62) 4-Bromofluorobenzene	11.079	95	93266	48.588	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	97.180%		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.179	85	74095	53.629	ug/l	99
3) Chloromethane	1.307	50	77768	52.149	ug/l	96
4) Vinyl Chloride	1.386	62	85590	53.822	ug/l	98
5) Bromomethane	1.618	94	49377	55.188	ug/l	95
6) Chloroethane	1.703	64	53474	55.480	ug/l	94
7) Trichlorofluoromethane	1.904	101	127152	52.659	ug/l	99
8) Diethyl Ether	2.148	74	44616	53.840	ug/l	98
9) 1,1,2-Trichlorotrifluo...	2.349	101	79479	53.654	ug/l	98
10) Methyl Iodide	2.465	142	85230	46.314	ug/l	100
11) Tert butyl alcohol	2.959	59	35007	217.898	ug/l	100
12) 1,1-Dichloroethene	2.337	96	76619	53.603	ug/l	98
13) Acrolein	2.245	56	69969	298.999	ug/l	98
14) Allyl chloride	2.678	41	137304	53.017	ug/l	98
15) Acrylonitrile	3.075	53	185932	257.376	ug/l	99
16) Acetone	2.386	43	144349	241.828	ug/l	97
17) Carbon Disulfide	2.526	76	205491	47.607	ug/l	99
18) Methyl Acetate	2.715	43	78667	51.889	ug/l	99
19) Methyl tert-butyl Ether	3.123	73	225140	50.585	ug/l	100
20) Methylene Chloride	2.800	84	83440	50.266	ug/l	99
21) trans-1,2-Dichloroethene	3.105	96	79204	51.786	ug/l	99
22) Diisopropyl ether	3.770	45	270205	54.281	ug/l	98
23) Vinyl Acetate	3.733	43	1055978	266.763	ug/l	99
24) 1,1-Dichloroethane	3.623	63	148012	52.369	ug/l	99
25) 2-Butanone	4.562	43	216354	256.282	ug/l	97
26) 2,2-Dichloropropane	4.489	77	106712	51.768	ug/l	100
27) cis-1,2-Dichloroethene	4.501	96	94359	52.513	ug/l	99
28) Bromochloromethane	4.910	49	66329	51.765	ug/l	99
29) Tetrahydrofuran	5.007	42	139493	255.819	ug/l	99
30) Chloroform	5.105	83	149506	52.894	ug/l	95
31) Cyclohexane	5.477	56	137290	51.222	ug/l	94
32) 1,1,1-Trichloroethane	5.397	97	125889	50.784	ug/l	99
36) 1,1-Dichloropropene	5.702	75	105895	51.182	ug/l	100
37) Ethyl Acetate	4.721	43	92088	48.219	ug/l	98
38) Carbon Tetrachloride	5.690	117	111211	49.529	ug/l	99
39) Methylcyclohexane	7.385	83	139618	51.281	ug/l	99
40) Benzene	6.050	78	318931	52.086	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
 Data File : VX046775.D
 Acq On : 19 Jun 2025 17:10
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/20/2025
 Supervised By : Mahesh Dadoda 06/21/2025

Quant Time: Jun 20 05:22:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.928	41	55058	54.078	ug/l	95
42) 1,2-Dichloroethane	6.092	62	109435	50.860	ug/l	98
43) Isopropyl Acetate	6.342	43	151380	50.655	ug/l	100
44) Trichloroethene	7.135	130	79977	51.252	ug/l	97
45) 1,2-Dichloropropane	7.434	63	78602	51.797	ug/l	94
46) Dibromomethane	7.586	93	54750	50.201	ug/l	99
47) Bromodichloromethane	7.824	83	114561	51.475	ug/l	99
48) Methyl methacrylate	7.696	41	80176	53.494	ug/l	99
49) 1,4-Dioxane	7.665	88	19434	894.415	ug/l	94
51) 4-Methyl-2-Pentanone	8.574	43	453562	254.472	ug/l	99
52) Toluene	8.720	92	195403	51.477	ug/l	97
53) t-1,3-Dichloropropene	8.976	75	106195	51.427	ug/l	98
54) cis-1,3-Dichloropropene	8.366	75	122494	52.254	ug/l	99
55) 1,1,2-Trichloroethane	9.153	97	73528	51.978	ug/l	95
56) Ethyl methacrylate	9.116	69	109987	52.671	ug/l	99
57) 1,3-Dichloropropane	9.305	76	125247	51.414	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.244	63	321699	296.571	ug/l	100
59) 2-Hexanone	9.427	43	310545	252.021	ug/l	100
60) Dibromochloromethane	9.519	129	85854	51.500	ug/l	99
61) 1,2-Dibromoethane	9.610	107	73504	51.083	ug/l	99
64) Tetrachloroethene	9.275	164	66307	50.181	ug/l	97
65) Chlorobenzene	10.079	112	216318	51.382	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	72835	51.476	ug/l	99
67) Ethyl Benzene	10.195	91	380469	51.722	ug/l	99
68) m/p-Xylenes	10.299	106	284342	103.739	ug/l	99
69) o-Xylene	10.640	106	136531	53.151	ug/l	98
70) Styrene	10.653	104	238877	53.139	ug/l	99
71) Bromoform	10.799	173	54055	50.164	ug/l #	100
73) Isopropylbenzene	10.963	105	364838	51.780	ug/l	100
74) N-amyl acetate	10.841	43	134737	50.472	ug/l	100
75) 1,1,2,2-Tetrachloroethane	11.207	83	100156	50.908	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	81274m	46.384	ug/l	
77) Bromobenzene	11.195	156	87850	51.813	ug/l	100
78) n-propylbenzene	11.305	91	432963	51.735	ug/l	99
79) 2-Chlorotoluene	11.360	91	249407	49.914	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	298594	51.570	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	29148	47.021	ug/l	100
82) 4-Chlorotoluene	11.451	91	293281	50.293	ug/l	100
83) tert-Butylbenzene	11.713	119	304194	50.825	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	302076	51.927	ug/l	98
85) sec-Butylbenzene	11.890	105	387895	51.281	ug/l	99
86) p-Isopropyltoluene	12.006	119	325914	51.067	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	162851	49.259	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	161313	47.505	ug/l	100
89) n-Butylbenzene	12.329	91	308844	50.967	ug/l	100
90) Hexachloroethane	12.536	117	54659	48.908	ug/l	98
91) 1,2-Dichlorobenzene	12.335	146	154918	50.145	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	18542	47.738	ug/l	96
93) 1,2,4-Trichlorobenzene	13.585	180	109313	49.769	ug/l	98
94) Hexachlorobutadiene	13.725	225	45830	49.603	ug/l	99
95) Naphthalene	13.774	128	313462	50.093	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	105543	50.228	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
Data File : VX046775.D
Acq On : 19 Jun 2025 17:10
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Manual Integrations
APPROVED

Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025

Quant Time: Jun 20 05:22:53 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 03:09:16 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


```
Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
Data File : VX046775.D
Acq On    : 19 Jun 2025  17:10
Operator  : JC/MD
Sample    : VSTDCCC050
Misc      : 5.0mL/MSVOA_X/WATER
ALS Vial  : 19   Sample Multiplier: 1
```


Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Manual Integrations

Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025

Quant Time: Jun 20 05:22:53 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 03:09:16 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.562 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion:168 Resp: 129409
Ion Ratio Lower Upper
168 100
99 58.5 48.5 72.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#2
Dichlorodifluoromethane
Concen: 53.629 ug/l
RT: 1.179 min Scan# 15
Delta R.T. -0.006 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10


Tgt Ion: 85 Resp: 74095
Ion Ratio Lower Upper
85 100
87 32.6 16.6 49.8





#3
Chloromethane
Concen: 52.149 ug/l
RT: 1.307 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 50 Resp: 7776
Ion Ratio Lower Upper
50 100
52 33.2 24.9 37.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#4
Vinyl Chloride
Concen: 53.822 ug/l
RT: 1.386 min Scan# 49
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10


Tgt Ion: 62 Resp: 85590
Ion Ratio Lower Upper
62 100
64 30.8 23.9 35.9





#5
Bromomethane
Concen: 55.188 ug/l
RT: 1.618 min Scan# 81
Delta R.T. -0.006 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 94 Resp: 4937
Ion Ratio Lower Upper
94 100
96 95.6 72.3 108.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#6
Chloroethane
Concen: 55.480 ug/l
RT: 1.703 min Scan# 101
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10


Tgt Ion: 64 Resp: 53474
Ion Ratio Lower Upper
64 100
66 30.7 27.1 40.7





#7
Trichlorofluoromethane
Concen: 52.659 ug/l
RT: 1.904 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion:101 Resp: 12715
Ion Ratio Lower Upper
101 100
103 65.4 51.8 77.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#8
Diethyl Ether
Concen: 53.840 ug/l
RT: 2.148 min Scan# 174
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10

Tgt Ion: 74 Resp: 44616
Ion Ratio Lower Upper
74 100
45 102.7 52.3 157.1





#9

1,1,2-Trichlorotrifluoroethane

Concen: 53.654 ug/l

RT: 2.349 min Scan# 206

Delta R.T. 0.006 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion:101 Resp: 79479

Ion Ratio Lower Upper

101 100

85 43.2 35.4 53.0





151 74.1 61.2 91.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#10

Methyl Iodide

Concen: 46.314 ug/l

RT: 2.465 min Scan# 226

Delta R.T. -0.006 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10




Tgt Ion:142 Resp: 85230


Ion Ratio Lower Upper

142 100

127 46.1 37.0 55.6

141 14.0 11.5 17.3





#11

Tert butyl alcohol

Concen: 217.898 ug/l

RT: 2.959 min Scan# 307

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 59 Resp: 3500

Ion Ratio Lower Upper

59 100





57 10.5 8.3 12.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#12

1,1-Dichloroethene

Concen: 53.603 ug/l

RT: 2.337 min Scan# 205

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10




Tgt Ion: 96 Resp: 76619


Ion Ratio Lower Upper

96 100

61 161.7 126.6 189.8

98 64.0 50.7 76.1





#13

Acrolein

Concen: 298.999 ug/l

RT: 2.245 min Scan# 190

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 56 Resp: 69969

Ion Ratio Lower Upper

56 100





55 70.7 55.2 82.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#14

Allyl chloride

Concen: 53.017 ug/l

RT: 2.678 min Scan# 261

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10




Tgt Ion: 41 Resp: 137304


Ion Ratio Lower Upper

41 100

39 65.0 53.6 80.4

76 32.3 26.8 40.2





#15

Acrylonitrile

Concen: 257.376 ug/l

RT: 3.075 min Scan# 31

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 53 Resp: 18593

Ion Ratio Lower Upper

53 100

52 82.3 65.1 97.7





51 36.4 28.9 43.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#16

Acetone

Concen: 241.828 ug/l

RT: 2.386 min Scan# 213

Delta R.T. -0.000 min

Lab File: VX046775.D




Acq: 19 Jun 2025 17:10


Tgt Ion: 43 Resp: 144349

Ion Ratio Lower Upper

43 100

58 32.3 24.5 36.7





#17

Carbon Disulfide

Concen: 47.607 ug/l

RT: 2.526 min Scan# 21

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 76 Resp: 20549

Ion Ratio Lower Upper

76 100





78 9.1 7.5 11.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#18

Methyl Acetate

Concen: 51.889 ug/l

RT: 2.715 min Scan# 267

Delta R.T. -0.000 min

Lab File: VX046775.D




Acq: 19 Jun 2025 17:10


Tgt Ion: 43 Resp: 78667

Ion Ratio Lower Upper

43 100

74 22.7 18.4 27.6





#19

Methyl tert-butyl Ether

Concen: 50.585 ug/l

RT: 3.123 min Scan# 31

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 73 Resp: 225140

Ion Ratio Lower Upper

73 100





57 22.9 18.2 27.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#20

Methylene Chloride

Concen: 50.266 ug/l

RT: 2.800 min Scan# 281

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10

Tgt Ion: 84 Resp: 83440




Ion Ratio Lower Upper


84 100

49 127.7 101.3 151.9

51 40.1 31.0 46.6


86 62.9 51.1 76.7





#21
trans-1,2-Dichloroethene
Concen: 51.786 ug/l
RT: 3.105 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 96 Resp: 79204
Ion Ratio Lower Upper
96 100
61 141.7 112.5 168.7
98 65.4 51.9 77.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#22
Diisopropyl ether
Concen: 54.281 ug/l
RT: 3.770 min Scan# 440
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10

Tgt Ion: 45 Resp: 270205
Ion Ratio Lower Upper
45 100
43 117.8 91.9 137.9
87 27.3 22.0 33.0
59 11.3 9.1 13.7





#23

Vinyl Acetate

Concen: 266.763 ug/l

RT: 3.733 min Scan# 41

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10

Tgt Ion: 43 Resp: 1055978

Ion Ratio Lower Upper

43 100

86 10.0 8.2 12.2

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#24

1,1-Dichloroethane

Concen: 52.369 ug/l

RT: 3.623 min Scan# 416

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10




Tgt Ion: 63 Resp: 148012


Ion Ratio Lower Upper

63 100

98 7.3 4.0 11.9

100 4.1 2.2 6.6





#25

2-Butanone

Concen: 256.282 ug/l

RT: 4.562 min Scan# 51

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 43 Resp: 216354

Ion Ratio Lower Upper

43 100





72 23.7 20.2 30.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#26

2,2-Dichloropropane

Concen: 51.768 ug/l

RT: 4.489 min Scan# 558

Delta R.T. -0.000 min

Lab File: VX046775.D




Acq: 19 Jun 2025 17:10


Tgt Ion: 77 Resp: 106712

Ion Ratio Lower Upper

77 100

97 23.6 11.8 35.3





#27

cis-1,2-Dichloroethene

Concen: 52.513 ug/l

RT: 4.501 min Scan# 560

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 96 Resp: 9435

Ion Ratio Lower Upper

96 100

61 143.7 0.0 291.0





98 63.9 0.0 126.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#28

Bromochloromethane

Concen: 51.765 ug/l

RT: 4.910 min Scan# 627

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10




Tgt Ion: 49 Resp: 66329


Ion Ratio Lower Upper

49 100

129 1.9 0.0 3.8

130 75.8 61.4 92.0





#29

Tetrahydrofuran

Concen: 255.819 ug/l

RT: 5.007 min Scan# 644

Delta R.T. -0.006 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 42 Resp: 13949

Ion Ratio Lower Upper

42 100

72 42.7 34.9 52.3





71 39.6 32.2 48.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#30

Chloroform

Concen: 52.894 ug/l

RT: 5.105 min Scan# 659

Delta R.T. -0.000 min

Lab File: VX046775.D




Acq: 19 Jun 2025 17:10


Tgt Ion: 83 Resp: 149506

Ion Ratio Lower Upper

83 100

85 63.1 53.7 80.5





#31

Cyclohexane

Concen: 51.222 ug/l

RT: 5.477 min Scan# 721

Delta R.T. -0.006 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 56 Resp: 137290

Ion Ratio Lower Upper

56 100

69 28.7 24.8 37.2





84 79.7 68.2 102.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#32

1,1,1-Trichloroethane

Concen: 50.784 ug/l

RT: 5.397 min Scan# 707

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10




Tgt Ion: 97 Resp: 125889


Ion Ratio Lower Upper

97 100

99 63.9 51.3 76.9

61 46.1 36.4 54.6





#33

1,2-Dichloroethane-d4

Concen: 46.825 ug/l

RT: 5.964 min Scan# 800

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 65 Resp: 8381

Ion Ratio Lower Upper

65 100





67 52.4 0.0 105.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.769 min Scan# 932

Delta R.T. 0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10




Tgt Ion:114 Resp: 216626


Ion Ratio Lower Upper

114 100

63 19.7 0.0 44.2

88 16.1 0.0 33.2





#35

Dibromofluoromethane

Concen: 48.316 ug/l

RT: 5.397 min Scan# 707

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion:113 Resp: 69249

Ion Ratio Lower Upper

113 100

111 103.9 82.0 123.0





192 18.5 15.3 22.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#36

1,1-Dichloropropene

Concen: 51.182 ug/l

RT: 5.702 min Scan# 757

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10




Tgt Ion: 75 Resp: 105895


Ion Ratio Lower Upper

75 100

110 34.3 17.2 51.5

77 30.8 24.2 36.4





#37

Ethyl Acetate

Concen: 48.219 ug/l

RT: 4.721 min Scan# 596

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 43 Resp: 92088

Ion Ratio Lower Upper

43 100

61 13.5 9.7 14.5





70 10.2 8.1 12.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#38

Carbon Tetrachloride

Concen: 49.529 ug/l

RT: 5.690 min Scan# 755

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10




Tgt Ion:117 Resp: 111211


Ion Ratio Lower Upper

117 100

119 94.7 75.2 112.8

121 30.9 24.1 36.1





#39

Methylcyclohexane

Concen: 51.281 ug/l

RT: 7.385 min Scan# 1033

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 83 Resp: 139618

Ion Ratio Lower Upper

83 100

55 78.5 63.0 94.4





98 49.6 38.8 58.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#40

Benzene

Concen: 52.086 ug/l

RT: 6.050 min Scan# 814

Delta R.T. 0.006 min

Lab File: VX046775.D




Acq: 19 Jun 2025 17:10


Tgt Ion: 78 Resp: 318931

Ion Ratio Lower Upper

78 100


77 23.7 18.8 28.2





#41
Methacrylonitrile
Concen: 54.078 ug/l
RT: 4.928 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 41 Resp: 55058
Ion Ratio Lower Upper
41 100
39 53.7 45.9 68.9
67 71.4 61.2 91.8
52 30.7 25.5 38.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#42
1,2-Dichloroethane
Concen: 50.860 ug/l
RT: 6.092 min Scan# 821
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10

Tgt Ion: 62 Resp: 109435
Ion Ratio Lower Upper
62 100
98 9.5 0.0 17.8





#43

Isopropyl Acetate

Concen: 50.655 ug/l

RT: 6.342 min Scan# 862

Delta R.T. -0.000 min

Lab File: VX046775.D


Acq: 19 Jun 2025 17:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC






Tgt Ion:	43	Resp:	15138
Ion Ratio	Lower	Upper	
43	100		
61	19.2	15.3	22.9
87	12.5	10.0	15.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#44

Trichloroethene


Concen: 51.252 ug/l

RT: 7.135 min Scan# 992



Delta R.T. 0.006 min


Lab File: VX046775.D

Acq: 19 Jun 2025 17:10



Tgt Ion:	130	Resp:	79977
Ion Ratio	Lower	Upper	
130	100		
95	98.6	0.0	202.6





#45

1,2-Dichloropropane

Concen: 51.797 ug/l

RT: 7.434 min Scan# 1041

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 63 Resp: 7860

Ion Ratio Lower Upper

63 100





65 31.9 22.9 34.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#46

Dibromomethane

Concen: 50.201 ug/l

RT: 7.586 min Scan# 1066

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10




Tgt Ion: 93 Resp: 54750


Ion Ratio Lower Upper

93 100

95 83.8 66.2 99.4

174 97.6 77.0 115.6





#47

Bromodichloromethane

Concen: 51.475 ug/l

RT: 7.824 min Scan# 1105

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 83 Resp: 11456

Ion Ratio Lower Upper

83 100

85 65.0 51.4 77.2





127 8.5 6.9 10.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#48

Methyl methacrylate

Concen: 53.494 ug/l

RT: 7.696 min Scan# 1084

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10




Tgt Ion: 41 Resp: 80176


Ion Ratio Lower Upper

41 100

69 79.0 63.4 95.2

39 54.7 45.4 68.0





#49

1,4-Dioxane

Concen: 894.415 ug/l

RT: 7.665 min Scan# 1078

Delta R.T. 0.006 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 88 Resp: 19434

Ion Ratio Lower Upper

88 100

43 36.0 31.2 46.8





58 73.2 54.7 82.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#50

Toluene-d8

Concen: 47.880 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. -0.000 min

Lab File: VX046775.D




Acq: 19 Jun 2025 17:10


Tgt Ion: 98 Resp: 246552

Ion Ratio Lower Upper

98 100

100 67.5 53.0 79.4









#51
4-Methyl-2-Pentanone
Concen: 254.472 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 43 Resp: 453562
Ion Ratio Lower Upper
43 100
58 38.2 30.8 46.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#52
Toluene
Concen: 51.477 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10


Tgt Ion: 92 Resp: 195403
Ion Ratio Lower Upper
92 100
91 172.1 134.8 202.2





#53
t-1,3-Dichloropropene
Concen: 51.427 ug/l
RT: 8.976 min Scan# 1194
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 75 Resp: 10619
Ion Ratio Lower Upper
75 100
77 30.9 25.6 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#54
cis-1,3-Dichloropropene
Concen: 52.254 ug/l
RT: 8.366 min Scan# 1194
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10

Tgt Ion: 75 Resp: 122494
Ion Ratio Lower Upper
75 100
77 31.0 25.3 37.9
39 48.8 39.2 58.8





#55

1,1,2-Trichloroethane

Concen: 51.978 ug/l

RT: 9.153 min Scan# 1323

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 97 Resp: 73528

Ion Ratio Lower Upper

97 100

83 86.6 67.2 100.8

85 57.4 44.1 66.1





99 69.5 50.3 75.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#56

Ethyl methacrylate

Concen: 52.671 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10




Tgt Ion: 69 Resp: 109987


Ion Ratio Lower Upper

69 100

41 69.5 55.4 83.2

39 36.3 29.7 44.5





#57

1,3-Dichloropropane

Concen: 51.414 ug/l

RT: 9.305 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 76 Resp: 12524

Ion Ratio Lower Upper

76 100





78 32.9 25.7 38.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#58

2-Chloroethyl Vinyl ether

Concen: 296.571 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. -0.000 min

Lab File: VX046775.D




Acq: 19 Jun 2025 17:10


Tgt Ion: 63 Resp: 321699

Ion Ratio Lower Upper

63 100

106 28.4 22.6 34.0





#59

2-Hexanone

Concen: 252.021 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 43 Resp: 31054

Ion Ratio Lower Upper

43 100





58 53.5 26.8 80.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#60

Dibromochloromethane

Concen: 51.500 ug/l

RT: 9.519 min Scan# 1383

Delta R.T. -0.000 min

Lab File: VX046775.D




Acq: 19 Jun 2025 17:10


Tgt Ion:129 Resp: 85854

Ion Ratio Lower Upper

129 100


127 75.9 38.4 115.2





#61
1,2-Dibromoethane
Concen: 51.083 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion:107 Resp: 73504
Ion Ratio Lower Upper
107 100
109 94.2 75.8 113.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#62
4-Bromofluorobenzene
Concen: 48.588 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10

Tgt Ion: 95 Resp: 93266
Ion Ratio Lower Upper
95 100
174 75.5 0.0 150.4
176 73.4 0.0 145.0





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.055 min Scan# 1471

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion:117 Resp: 190720

Ion Ratio Lower Upper

117 100

82 54.4 44.6 66.8





119 32.1 25.8 38.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#64

Tetrachloroethene

Concen: 50.181 ug/l

RT: 9.275 min Scan# 1343

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10

Tgt Ion:164 Resp: 66307




Ion Ratio Lower Upper


164 100

166 124.5 104.6 157.0

129 94.6 76.3 114.5

131 89.1 74.1 111.1





#65

Chlorobenzene

Concen: 51.382 ug/l

RT: 10.079 min Scan# 1475

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion:112 Resp: 216318

Ion Ratio Lower Upper

112 100





114 31.5 25.5 38.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#66

1,1,1,2-Tetrachloroethane

Concen: 51.476 ug/l

RT: 10.159 min Scan# 1488

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10




Tgt Ion:131 Resp: 72835


Ion Ratio Lower Upper

131 100

133 96.0 47.3 141.8


119 63.9 32.1 96.3





#67
Ethyl Benzene
Concen: 51.722 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.006 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 91 Resp: 380469
Ion Ratio Lower Upper
91 100
106 30.8 24.2 36.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#68
m/p-Xylenes
Concen: 103.739 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10

Tgt Ion:106 Resp: 284342
Ion Ratio Lower Upper
106 100
91 207.8 164.8 247.2









#69
o-Xylene
Concen: 53.151 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion:106 Resp: 13653
Ion Ratio Lower Upper
106 100
91 219.1 111.3 333.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#70
Styrene
Concen: 53.139 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10


Tgt Ion:104 Resp: 238877
Ion Ratio Lower Upper
104 100
78 51.0 41.4 62.2
103 53.6 43.1 64.7





#71
Bromoform
Concen: 50.164 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion:173 Resp: 5405
Ion Ratio Lower Upper
173 100
175 48.3 24.1 72.4
254 0.0 0.1 0.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10

Tgt Ion:152 Resp: 95668
Ion Ratio Lower Upper
152 100
115 85.8 43.2 129.6
150 173.3 0.0 346.8





#73

Isopropylbenzene

Concen: 51.780 ug/l

RT: 10.963 min Scan# 1

Delta R.T. 0.006 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion:105 Resp: 364838

Ion Ratio Lower Upper

105 100





120 25.9 13.0 39.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#74

N-ethyl acetate

Concen: 50.472 ug/l

RT: 10.841 min Scan# 1600

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10

Tgt Ion: 43 Resp: 134737




Ion Ratio Lower Upper


43 100

70 42.0 33.4 50.0

55 24.0 19.1 28.7

61 24.0 19.0 28.4





#75

1,1,2,2-Tetrachloroethane

Concen: 50.908 ug/l

RT: 11.207 min Scan# 1660

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 83 Resp: 100150

Ion Ratio Lower Upper

83 100

131 10.5 4.9 14.6





85 64.3 32.0 96.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#76

1,2,3-Trichloropropane

Concen: 46.384 ug/l m

RT: 11.238 min Scan# 1665

Delta R.T. -0.000 min

Lab File: VX046775.D




Acq: 19 Jun 2025 17:10


Tgt Ion: 75 Resp: 81274

Ion Ratio Lower Upper

75 100


77 42.5 18.5 55.5





#77
Bromobenzene
Concen: 51.813 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion:156 Resp: 87850
Ion Ratio Lower Upper
156 100
77 164.9 82.8 248.6
158 95.3 47.6 142.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#78
n-propylbenzene
Concen: 51.735 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. 0.006 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10

Tgt Ion: 91 Resp: 432963
Ion Ratio Lower Upper
91 100
120 22.5 11.4 34.1









#79
2-Chlorotoluene
Concen: 49.914 ug/l
RT: 11.360 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 91 Resp: 24940
Ion Ratio Lower Upper
91 100
126 33.4 16.6 49.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#80
1,3,5-Trimethylbenzene
Concen: 51.570 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10


Tgt Ion:105 Resp: 298594
Ion Ratio Lower Upper
105 100
120 48.0 24.1 72.2





#81
trans-1,4-Dichloro-2-butene
Concen: 47.021 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 75 Resp: 2914
Ion Ratio Lower Upper
75 100
53 108.3 86.8 130.2
89 45.6 36.9 55.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#82
4-Chlorotoluene
Concen: 50.293 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10


Tgt Ion: 91 Resp: 293281
Ion Ratio Lower Upper
91 100
126 29.0 14.4 43.2





#83
tert-Butylbenzene
Concen: 50.825 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion:119 Resp: 304194
Ion Ratio Lower Upper
119 100
91 59.5 29.5 88.6
134 23.4 11.7 35.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#84
1,2,4-Trimethylbenzene
Concen: 51.927 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10


Tgt Ion:105 Resp: 302076
Ion Ratio Lower Upper
105 100
120 43.0 22.1 66.3





#85
 sec-Butylbenzene
 Concen: 51.281 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. -0.000 min
 Lab File: VX046775.D
 Acq: 19 Jun 2025 17:10




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC




Tgt Ion:105 Resp: 38789
 Ion Ratio Lower Upper
 105 100
 134 19.8 9.8 29.4

Manual Integrations
 APPROVED



Reviewed By :John Carlone 06/20/2025
 Supervised By :Mahesh Dadoda 06/21/2025




#86
 p-Isopropyltoluene
 Concen: 51.067 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. -0.000 min
 Lab File: VX046775.D
 Acq: 19 Jun 2025 17:10



Tgt Ion:119 Resp: 325914
 Ion Ratio Lower Upper
 119 100
 134 26.0 12.9 38.7
 91 24.2 12.3 36.9





#87

1,3-Dichlorobenzene

Concen: 49.259 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX046775.D


Acq: 19 Jun 2025 17:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC



Tgt Ion:146 Resp: 16285

Ion Ratio Lower Upper

146 100

111 40.8 20.6 61.8




148 63.7 31.6 94.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#88

1,4-Dichlorobenzene


Concen: 47.505 ug/l

RT: 12.036 min Scan# 1796

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10





Tgt Ion:146 Resp: 161313


Ion Ratio Lower Upper

146 100

111 40.7 20.3 61.0

148 64.9 32.1 96.5





#89

n-Butylbenzene

Concen: 50.967 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 91 Resp: 30884

Ion Ratio Lower Upper

91 100

92 54.4 27.1 81.3





134 25.4 12.9 38.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#90

Hexachloroethane

Concen: 48.908 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX046775.D




Acq: 19 Jun 2025 17:10


Tgt Ion:117 Resp: 54659

Ion Ratio Lower Upper

117 100


201 70.7 34.4 103.3





#91
1,2-Dichlorobenzene
Concen: 50.145 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.006 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion:146 Resp: 154918
Ion Ratio Lower Upper
146 100
111 42.6 21.7 65.1
148 64.1 32.0 96.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 47.738 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10


Tgt Ion: 75 Resp: 18542
Ion Ratio Lower Upper
75 100
155 84.1 41.9 125.7
157 109.5 51.4 154.3





#93
1,2,4-Trichlorobenzene
Concen: 49.769 ug/l
RT: 13.585 min Scan# 2050
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion:180 Resp: 10931
Ion Ratio Lower Upper
180 100
182 95.3 46.7 140.1
145 31.5 15.8 47.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#94
Hexachlorobutadiene
Concen: 49.603 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.006 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10


Tgt Ion:225 Resp: 45830
Ion Ratio Lower Upper
225 100
223 64.3 32.5 97.5
227 63.7 32.2 96.6





#95
Naphthalene
Concen: 50.093 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion:128 Resp: 313462
Ion Ratio Lower Upper
128 100
127 12.7 10.1 15.1
129 10.9 8.7 13.1




Manual Integrations
APPROVED

Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#96
1,2,3-Trichlorobenzene
Concen: 50.228 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10

Tgt Ion:180 Resp: 105543
Ion Ratio Lower Upper
180 100
182 95.2 47.8 143.4
145 33.7 17.3 51.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
 Data File : VX046775.D
 Acq On : 19 Jun 2025 17:10
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 20 05:22:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	105	0.00
2 T	Dichlorodifluoromethane	0.534	0.573	-7.3	103	0.00
3 P	Chloromethane	0.576	0.601	-4.3	106	0.00
4 C	Vinyl Chloride	0.614	0.661	-7.7#	109	0.00
5 T	Bromomethane	0.346	0.382	-10.4	108	0.00
6 T	Chloroethane	0.372	0.413	-11.0	115	0.00
7 T	Trichlorofluoromethane	0.933	0.983	-5.4	107	0.00
8 T	Diethyl Ether	0.320	0.345	-7.8	112	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.572	0.614	-7.3	112	0.00
10 T	Methyl Iodide	0.642	0.659	-2.6	106	0.00
11 T	Tert butyl alcohol	0.062	0.054	12.9	92	0.00
12 CM	1,1-Dichloroethene	0.552	0.592	-7.2#	109	0.00
13 T	Acrolein	0.090	0.108	-20.0	130	0.00
14 T	Allyl chloride	1.001	1.061	-6.0	111	0.00
15 T	Acrylonitrile	0.279	0.287	-2.9	106	0.00
16 T	Acetone	0.231	0.223	3.5	106	0.00
17 T	Carbon Disulfide	1.668	1.588	4.8	104	0.00
18 T	Methyl Acetate	0.586	0.608	-3.8	110	0.00
19 T	Methyl tert-butyl Ether	1.720	1.740	-1.2	104	0.00
20 T	Methylene Chloride	0.641	0.645	-0.6	111	0.00
21 T	trans-1,2-Dichloroethene	0.591	0.612	-3.6	109	0.00
22 T	Diisopropyl ether	1.923	2.088	-8.6	111	0.00
23 T	Vinyl Acetate	1.529	1.632	-6.7	107	0.00
24 P	1,1-Dichloroethane	1.092	1.144	-4.8	110	0.00
25 T	2-Butanone	0.326	0.334	-2.5	104	0.00
26 T	2,2-Dichloropropane	0.796	0.825	-3.6	114	0.00
27 T	cis-1,2-Dichloroethene	0.694	0.729	-5.0	111	0.00
28 T	Bromochloromethane	0.495	0.513	-3.6	111	0.00
29 T	Tetrahydrofuran	0.211	0.216	-2.4	104	0.00
30 C	Chloroform	1.092	1.155	-5.8#	109	0.00
31 T	Cyclohexane	1.036	1.061	-2.4	110	0.00
32 T	1,1,1-Trichloroethane	0.958	0.973	-1.6	107	0.00
33 S	1,2-Dichloroethane-d4	0.692	0.648	6.4	105	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	106	0.00
35 S	Dibromofluoromethane	0.331	0.320	3.3	106	0.00
36 T	1,1-Dichloropropene	0.478	0.489	-2.3	109	0.00
37 T	Ethyl Acetate	0.441	0.425	3.6	105	0.00
38 T	Carbon Tetrachloride	0.518	0.513	1.0	106	0.00
39 T	Methylcyclohexane	0.628	0.645	-2.7	109	0.00
40 TM	Benzene	1.413	1.472	-4.2	110	0.00
41 T	Methacrylonitrile	0.235	0.254	-8.1	113	0.00
42 TM	1,2-Dichloroethane	0.497	0.505	-1.6	108	0.00
43 T	Isopropyl Acetate	0.690	0.699	-1.3	103	0.00
44 TM	Trichloroethene	0.360	0.369	-2.5	110	0.00
45 C	1,2-Dichloropropane	0.350	0.363	-3.7#	111	0.00
46 T	Dibromomethane	0.252	0.253	-0.4	107	0.00
47 T	Bromodichloromethane	0.514	0.529	-2.9	107	0.00
48 T	Methyl methacrylate	0.346	0.370	-6.9	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
 Data File : VX046775.D
 Acq On : 19 Jun 2025 17:10
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 20 05:22:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.004	0.004	0.0	98	0.00
50 S	Toluene-d8	1.189	1.138	4.3	105	0.00
51 T	4-Methyl-2-Pentanone	0.411	0.419	-1.9	104	0.00
52 CM	Toluene	0.876	0.902	-3.0#	108	0.00
53 T	t-1,3-Dichloropropene	0.477	0.490	-2.7	106	0.00
54 T	cis-1,3-Dichloropropene	0.541	0.565	-4.4	109	0.00
55 T	1,1,2-Trichloroethane	0.327	0.339	-3.7	110	0.00
56 T	Ethyl methacrylate	0.482	0.508	-5.4	106	0.00
57 T	1,3-Dichloropropane	0.562	0.578	-2.8	109	0.00
58 T	2-Chloroethyl Vinyl ether	0.250	0.297	-18.8	112	0.00
59 T	2-Hexanone	0.284	0.287	-1.1	102	0.00
60 T	Dibromochloromethane	0.385	0.396	-2.9	108	0.00
61 T	1,2-Dibromoethane	0.332	0.339	-2.1	107	0.00
62 S	4-Bromofluorobenzene	0.443	0.431	2.7	107	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	106	0.00
64 T	Tetrachloroethene	0.346	0.348	-0.6	109	0.00
65 PM	Chlorobenzene	1.104	1.134	-2.7	111	0.00
66 T	1,1,1,2-Tetrachloroethane	0.371	0.382	-3.0	109	0.00
67 C	Ethyl Benzene	1.929	1.995	-3.4#	110	0.00
68 T	m/p-Xylenes	0.719	0.745	-3.6	109	0.00
69 T	o-Xylene	0.673	0.716	-6.4	110	0.00
70 T	Styrene	1.179	1.253	-6.3	110	0.00
71 P	Bromoform	0.282	0.283	-0.4	105	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	0.00
73 T	Isopropylbenzene	3.682	3.814	-3.6	110	0.00
74 T	N-amyl acetate	1.395	1.408	-0.9	103	0.00
75 P	1,1,2,2-Tetrachloroethane	1.028	1.047	-1.8	108	0.00
76 T	1,2,3-Trichloropropane	0.916	0.850	7.2	92	0.00
77 T	Bromobenzene	0.886	0.918	-3.6	111	0.00
78 T	n-propylbenzene	4.374	4.526	-3.5	110	0.00
79 T	2-Chlorotoluene	2.611	2.607	0.2	108	0.00
80 T	1,3,5-Trimethylbenzene	3.026	3.121	-3.1	109	0.00
81 T	trans-1,4-Dichloro-2-butene	0.324	0.305	5.9	105	0.00
82 T	4-Chlorotoluene	3.048	3.066	-0.6	109	0.00
83 T	tert-Butylbenzene	3.128	3.180	-1.7	109	0.00
84 T	1,2,4-Trimethylbenzene	3.040	3.158	-3.9	111	0.00
85 T	sec-Butylbenzene	3.953	4.055	-2.6	109	0.00
86 T	p-Isopropyltoluene	3.336	3.407	-2.1	110	0.00
87 T	1,3-Dichlorobenzene	1.728	1.702	1.5	109	0.00
88 T	1,4-Dichlorobenzene	1.775	1.686	5.0	110	0.00
89 T	n-Butylbenzene	3.167	3.228	-1.9	110	0.00
90 T	Hexachloroethane	0.584	0.571	2.2	105	0.00
91 T	1,2-Dichlorobenzene	1.615	1.619	-0.2	109	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.203	0.194	4.4	99	0.00
93 T	1,2,4-Trichlorobenzene	1.148	1.143	0.4	109	0.00
94 T	Hexachlorobutadiene	0.483	0.479	0.8	107	0.00
95 T	Naphthalene	3.270	3.277	-0.2	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
Data File : VX046775.D
Acq On : 19 Jun 2025 17:10
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 20 05:22:53 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 03:09:16 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.098	1.103	-0.5	110	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
Data File : VX046775.D
Acq On : 19 Jun 2025 17:10
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 20 05:22:53 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 03:09:16 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	105	0.00
2 T	Dichlorodifluoromethane	50.000	53.629	-7.3	103	0.00
3 P	Chloromethane	50.000	52.149	-4.3	106	0.00
4 C	Vinyl Chloride	50.000	53.822	-7.6#	109	0.00
5 T	Bromomethane	50.000	55.188	-10.4	108	0.00
6 T	Chloroethane	50.000	55.480	-11.0	115	0.00
7 T	Trichlorofluoromethane	50.000	52.659	-5.3	107	0.00
8 T	Diethyl Ether	50.000	53.840	-7.7	112	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	53.654	-7.3	112	0.00
10 T	Methyl Iodide	50.000	46.314	7.4	106	0.00
11 T	Tert butyl alcohol	250.000	217.898	12.8	92	0.00
12 CM	1,1-Dichloroethene	50.000	53.603	-7.2#	109	0.00
13 T	Acrolein	250.000	298.999	-19.6	130	0.00
14 T	Allyl chloride	50.000	53.017	-6.0	111	0.00
15 T	Acrylonitrile	250.000	257.376	-3.0	106	0.00
16 T	Acetone	250.000	241.828	3.3	106	0.00
17 T	Carbon Disulfide	50.000	47.607	4.8	104	0.00
18 T	Methyl Acetate	50.000	51.889	-3.8	110	0.00
19 T	Methyl tert-butyl Ether	50.000	50.585	-1.2	104	0.00
20 T	Methylene Chloride	50.000	50.266	-0.5	111	0.00
21 T	trans-1,2-Dichloroethene	50.000	51.786	-3.6	109	0.00
22 T	Diisopropyl ether	50.000	54.281	-8.6	111	0.00
23 T	Vinyl Acetate	250.000	266.763	-6.7	107	0.00
24 P	1,1-Dichloroethane	50.000	52.369	-4.7	110	0.00
25 T	2-Butanone	250.000	256.282	-2.5	104	0.00
26 T	2,2-Dichloropropane	50.000	51.768	-3.5	114	0.00
27 T	cis-1,2-Dichloroethene	50.000	52.513	-5.0	111	0.00
28 T	Bromochloromethane	50.000	51.765	-3.5	111	0.00
29 T	Tetrahydrofuran	250.000	255.819	-2.3	104	0.00
30 C	Chloroform	50.000	52.894	-5.8#	109	0.00
31 T	Cyclohexane	50.000	51.222	-2.4	110	0.00
32 T	1,1,1-Trichloroethane	50.000	50.784	-1.6	107	0.00
33 S	1,2-Dichloroethane-d4	50.000	46.825	6.3	105	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	106	0.00
35 S	Dibromofluoromethane	50.000	48.316	3.4	106	0.00
36 T	1,1-Dichloropropene	50.000	51.182	-2.4	109	0.00
37 T	Ethyl Acetate	50.000	48.219	3.6	105	0.00
38 T	Carbon Tetrachloride	50.000	49.529	0.9	106	0.00
39 T	Methylcyclohexane	50.000	51.281	-2.6	109	0.00
40 TM	Benzene	50.000	52.086	-4.2	110	0.00
41 T	Methacrylonitrile	50.000	54.078	-8.2	113	0.00
42 TM	1,2-Dichloroethane	50.000	50.860	-1.7	108	0.00
43 T	Isopropyl Acetate	50.000	50.655	-1.3	103	0.00
44 TM	Trichloroethene	50.000	51.252	-2.5	110	0.00
45 C	1,2-Dichloropropane	50.000	51.797	-3.6#	111	0.00
46 T	Dibromomethane	50.000	50.201	-0.4	107	0.00
47 T	Bromodichloromethane	50.000	51.475	-3.0	107	0.00
48 T	Methyl methacrylate	50.000	53.494	-7.0	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
 Data File : VX046775.D
 Acq On : 19 Jun 2025 17:10
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 20 05:22:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	894.415	10.6	98	0.00
50 S	Toluene-d8	50.000	47.880	4.2	105	0.00
51 T	4-Methyl-2-Pentanone	250.000	254.472	-1.8	104	0.00
52 CM	Toluene	50.000	51.477	-3.0#	108	0.00
53 T	t-1,3-Dichloropropene	50.000	51.427	-2.9	106	0.00
54 T	cis-1,3-Dichloropropene	50.000	52.254	-4.5	109	0.00
55 T	1,1,2-Trichloroethane	50.000	51.978	-4.0	110	0.00
56 T	Ethyl methacrylate	50.000	52.671	-5.3	106	0.00
57 T	1,3-Dichloropropane	50.000	51.414	-2.8	109	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	296.571	-18.6	112	0.00
59 T	2-Hexanone	250.000	252.021	-0.8	102	0.00
60 T	Dibromochloromethane	50.000	51.500	-3.0	108	0.00
61 T	1,2-Dibromoethane	50.000	51.083	-2.2	107	0.00
62 S	4-Bromofluorobenzene	50.000	48.588	2.8	107	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	106	0.00
64 T	Tetrachloroethene	50.000	50.181	-0.4	109	0.00
65 PM	Chlorobenzene	50.000	51.382	-2.8	111	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	51.476	-3.0	109	0.00
67 C	Ethyl Benzene	50.000	51.722	-3.4#	110	0.00
68 T	m/p-Xylenes	100.000	103.739	-3.7	109	0.00
69 T	o-Xylene	50.000	53.151	-6.3	110	0.00
70 T	Styrene	50.000	53.139	-6.3	110	0.00
71 P	Bromoform	50.000	50.164	-0.3	105	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	107	0.00
73 T	Isopropylbenzene	50.000	51.780	-3.6	110	0.00
74 T	N-amyl acetate	50.000	50.472	-0.9	103	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	50.908	-1.8	108	0.00
76 T	1,2,3-Trichloropropane	50.000	46.384	7.2	92	0.00
77 T	Bromobenzene	50.000	51.813	-3.6	111	0.00
78 T	n-propylbenzene	50.000	51.735	-3.5	110	0.00
79 T	2-Chlorotoluene	50.000	49.914	0.2	108	0.00
80 T	1,3,5-Trimethylbenzene	50.000	51.570	-3.1	109	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	47.021	6.0	105	0.00
82 T	4-Chlorotoluene	50.000	50.293	-0.6	109	0.00
83 T	tert-Butylbenzene	50.000	50.825	-1.7	109	0.00
84 T	1,2,4-Trimethylbenzene	50.000	51.927	-3.9	111	0.00
85 T	sec-Butylbenzene	50.000	51.281	-2.6	109	0.00
86 T	p-Isopropyltoluene	50.000	51.067	-2.1	110	0.00
87 T	1,3-Dichlorobenzene	50.000	49.259	1.5	109	0.00
88 T	1,4-Dichlorobenzene	50.000	47.505	5.0	110	0.00
89 T	n-Butylbenzene	50.000	50.967	-1.9	110	0.00
90 T	Hexachloroethane	50.000	48.908	2.2	105	0.00
91 T	1,2-Dichlorobenzene	50.000	50.145	-0.3	109	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	47.738	4.5	99	0.00
93 T	1,2,4-Trichlorobenzene	50.000	49.769	0.5	109	0.00
94 T	Hexachlorobutadiene	50.000	49.603	0.8	107	0.00
95 T	Naphthalene	50.000	50.093	-0.2	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
Data File : VX046775.D
Acq On : 19 Jun 2025 17:10
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 20 05:22:53 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 03:09:16 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	50.228	-0.5	110	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6





QC SAMPLE DATA

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
Data File : VX046715.D
Acq On : 17 Jun 2025 08:46
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Title : SW846 8260
Last Update : Wed Jun 18 03:09:16 2025



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631


Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.7	12764	PASS
75	95	30	60	50.7	34573	PASS
95	95	100	100	100.0	68235	PASS
96	95	5	9	6.5	4402	PASS
173	174	0.00	2	0.7	374	PASS
174	95	50	100	74.8	51016	PASS
175	174	5	9	7.4	3785	PASS
176	174	95	101	96.2	49101	PASS
177	176	5	9	6.0	2956	PASS

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
Data File : VX046757.D
Acq On : 19 Jun 2025 08:42
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Title : SW846 8260
Last Update : Wed Jun 18 03:09:16 2025



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.8	9154	PASS
75	95	30	60	50.7	24627	PASS
95	95	100	100	100.0	48571	PASS
96	95	5	9	6.3	3061	PASS
173	174	0.00	2	1.0	352	PASS
174	95	50	100	72.6	35267	PASS
175	174	5	9	7.8	2745	PASS
176	174	95	101	97.9	34539	PASS
177	176	5	9	6.3	2186	PASS

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0619WBL01		SDG No.:	Q2361
Lab Sample ID:	VX0619WBL01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046760.D	1		06/19/25 10:33	VX061925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0619WBL01		SDG No.:	Q2361
Lab Sample ID:	VX0619WBL01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046760.D	1		06/19/25 10:33	VX061925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	48.3		81 - 118		97%	SPK: 50
1868-53-7	Dibromofluoromethane	49.0		80 - 119		98%	SPK: 50
2037-26-5	Toluene-d8	50.0		89 - 112		100%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.4		85 - 114		99%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	125000	5.562				
540-36-3	1,4-Difluorobenzene	212000	6.769				
3114-55-4	Chlorobenzene-d5	192000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	93000	12.018				



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0619WBL01		SDG No.:	Q2361
Lab Sample ID:	VX0619WBL01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046760.D	1		06/19/25 10:33	VX061925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
Data File : VX046760.D
Acq On : 19 Jun 2025 10:33
Operator : JC/MD
Sample : VX0619WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0619WBL01

Quant Time: Jun 20 05:15:59 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 03:09:16 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	124937	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	212303	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	191555	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	93040	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	83430	48.278	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.560%
35) Dibromofluoromethane	5.397	113	68758	48.950	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.900%
50) Toluene-d8	8.647	98	252224	49.979	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.960%
62) 4-Bromofluorobenzene	11.079	95	92918	49.393	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	98.780%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
Data File : VX046760.D
Acq On : 19 Jun 2025 10:33
Operator : JC/MD
Sample : VX0619WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0619WBL01

Quant Time: Jun 20 05:15:59 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 03:09:16 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.562 min Scan# 71
Delta R.T. 0.000 min
Lab File: VX046760.D
Acq: 19 Jun 2025 10:33




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBL01


Tgt Ion:168 Resp: 124937
Ion Ratio Lower Upper
168 100
99 60.2 48.5 72.7



#33
1,2-Dichloroethane-d4
Concen: 48.278 ug/l
RT: 5.964 min Scan# 800
Delta R.T. -0.000 min
Lab File: VX046760.D
Acq: 19 Jun 2025 10:33

Tgt Ion: 65 Resp: 83430
Ion Ratio Lower Upper
65 100
67 52.4 0.0 105.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.769 min Scan# 91

Delta R.T. 0.000 min

Lab File: VX046760.D

Acq: 19 Jun 2025 10:33

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBL01





Tgt Ion:114 Resp: 212303

Ion Ratio Lower Upper

114 100

63 21.2 0.0 44.2

88 16.0 0.0 33.2



#35

Dibromofluoromethane

Concen: 48.950 ug/l

RT: 5.397 min Scan# 707

Delta R.T. -0.000 min

Lab File: VX046760.D

Acq: 19 Jun 2025 10:33




Tgt Ion:113 Resp: 68758


Ion Ratio Lower Upper

113 100

111 103.2 82.0 123.0

192 19.4 15.3 22.9









#50
Toluene-d8
Concen: 49.979 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. -0.000 min
Lab File: VX046760.D
Acq: 19 Jun 2025 10:33




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBL01


Tgt Ion: 98 Resp: 252224
Ion Ratio Lower Upper
98 100
100 66.3 53.0 79.4



#62
4-Bromofluorobenzene
Concen: 49.393 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046760.D
Acq: 19 Jun 2025 10:33

Tgt Ion: 95 Resp: 92918
Ion Ratio Lower Upper
95 100
174 75.8 0.0 150.4
176 71.2 0.0 145.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX046760.D
Acq: 19 Jun 2025 10:33




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBL01

Tgt Ion:117 Resp: 191555
Ion Ratio Lower Upper
117 100
82 55.3 44.6 66.8
119 31.4 25.8 38.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX046760.D
Acq: 19 Jun 2025 10:33

Tgt Ion:152 Resp: 93040
Ion Ratio Lower Upper
152 100
115 60.1 43.2 129.6
150 157.7 0.0 346.8



Report of Analysis

Client:	AECOM Technical Services, Inc.			Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258			Date Received:	
Client Sample ID:	VX0619WBS01			SDG No.:	Q2361
Lab Sample ID:	VX0619WBS01			Matrix:	Water
Analytical Method:	8260D			% Solid:	0
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000 uL
Soil Aliquot Vol:			uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID :	0.18	Level :	LOW
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046762.D	1		06/19/25 11:24	VX061925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	19.1		0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	18.7		0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	18.9		0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	21.2		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	19.9		0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	18.4		0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	18.5		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	18.6		0.23	0.75	1.00	ug/L
67-64-1	Acetone	74.1		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	17.0		0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	17.0		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	16.0		0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	18.1		0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	18.3		0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	18.5		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	18.8		1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	78.9		0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	18.0		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	18.6		0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	20.8		0.22	0.50	1.00	ug/L
67-66-3	Chloroform	19.0		0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	17.7		0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	18.5		0.16	0.50	1.00	ug/L
71-43-2	Benzene	18.9		0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	18.5		0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	18.4		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	18.4		0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	18.9		0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	81.6		0.68	2.50	5.00	ug/L
108-88-3	Toluene	19.0		0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0619WBS01		SDG No.:	Q2361
Lab Sample ID:	VX0619WBS01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046762.D	1		06/19/25 11:24	VX061925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	18.0		0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	18.2		0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	18.5		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	78.9		0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	18.4		0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	18.0		0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	17.7		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	18.0		0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	18.0		0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	37.2		0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	18.6		0.12	0.50	1.00	ug/L
100-42-5	Styrene	18.3		0.15	0.50	1.00	ug/L
75-25-2	Bromoform	16.7		0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	18.1		0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	17.3		0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	17.8		0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	17.5		0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	18.1		0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	15.1		0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	17.1		0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	17.6		0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	49.8		81 - 118		100%	SPK: 50
1868-53-7	Dibromofluoromethane	52.0		80 - 119		104%	SPK: 50
2037-26-5	Toluene-d8	52.3		89 - 112		105%	SPK: 50
460-00-4	4-Bromofluorobenzene	54.1		85 - 114		108%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	138000	5.562				
540-36-3	1,4-Difluorobenzene	227000	6.769				
3114-55-4	Chlorobenzene-d5	207000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	105000	12.018				



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0619WBS01		SDG No.:	Q2361
Lab Sample ID:	VX0619WBS01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046762.D	1		06/19/25 11:24	VX061925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
 Data File : VX046762.D
 Acq On : 19 Jun 2025 11:24
 Operator : JC/MD
 Sample : VX0619WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0619WBS01

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/20/2025
 Supervised By : Mahesh Dadoda 06/21/2025

Quant Time: Jun 20 05:16:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	137936	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	227140	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	207463	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	104652	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	5.964	65	94946	49.765	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.520%
35) Dibromofluoromethane	5.397	113	78196	52.032	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.060%
50) Toluene-d8	8.647	98	282602	52.340	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.680%
62) 4-Bromofluorobenzene	11.079	95	108869	54.092	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	108.180%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.185	85	28069	19.060	ug/l	97
3) Chloromethane	1.307	50	29736	18.707	ug/l	99
4) Vinyl Chloride	1.386	62	32033	18.898	ug/l	93
5) Bromomethane	1.630	94	20191	21.172	ug/l	93
6) Chloroethane	1.709	64	20468	19.923	ug/l	98
7) Trichlorofluoromethane	1.910	101	47476	18.446	ug/l	99
8) Diethyl Ether	2.148	74	16252	18.400	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.349	101	29254	18.528	ug/l	99
10) Methyl Iodide	2.471	142	29614	18.018	ug/l	99
11) Tert butyl alcohol	2.959	59	10810	63.126	ug/l	97
12) 1,1-Dichloroethene	2.337	96	28292	18.570	ug/l	97
13) Acrolein	2.251	56	18186	72.910	ug/l	98
14) Allyl chloride	2.684	41	49324	17.868	ug/l	99
15) Acrylonitrile	3.075	53	62463	81.119	ug/l	98
16) Acetone	2.386	43	47143	74.096	ug/l	96
17) Carbon Disulfide	2.532	76	78405	17.041	ug/l	98
18) Methyl Acetate	2.715	43	25862	16.004	ug/l	98
19) Methyl tert-butyl Ether	3.123	73	80635	16.997	ug/l	98
20) Methylene Chloride	2.806	84	32048	18.113	ug/l	98
21) trans-1,2-Dichloroethene	3.111	96	29757	18.253	ug/l	99
22) Diisopropyl ether	3.769	45	100392	18.921	ug/l	94
23) Vinyl Acetate	3.733	43	365712	86.676	ug/l	100
24) 1,1-Dichloroethane	3.623	63	55663	18.477	ug/l	99
25) 2-Butanone	4.568	43	70964m	78.864	ug/l	
26) 2,2-Dichloropropane	4.489	77	39442	17.951	ug/l	99
27) cis-1,2-Dichloroethene	4.507	96	35547	18.560	ug/l	98
28) Bromochloromethane	4.910	49	28430	20.816	ug/l	99
29) Tetrahydrofuran	5.013	42	44591	76.721	ug/l	99
30) Chloroform	5.105	83	57187	18.982	ug/l	93
31) Cyclohexane	5.483	56	53811	18.836	ug/l	96
32) 1,1,1-Trichloroethane	5.391	97	46675	17.665	ug/l	98
36) 1,1-Dichloropropene	5.702	75	40684	18.754	ug/l	99
37) Ethyl Acetate	4.727	43	31829	15.895	ug/l	98
38) Carbon Tetrachloride	5.690	117	42451	18.031	ug/l	99
39) Methylcyclohexane	7.385	83	52931	18.541	ug/l	99
40) Benzene	6.050	78	121111	18.864	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
 Data File : VX046762.D
 Acq On : 19 Jun 2025 11:24
 Operator : JC/MD
 Sample : VX0619WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0619WBS01

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/20/2025
 Supervised By : Mahesh Dadoda 06/21/2025

Quant Time: Jun 20 05:16:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.928	41	17227	16.137	ug/l	99
42) 1,2-Dichloroethane	6.092	62	41780	18.518	ug/l	100
43) Isopropyl Acetate	6.348	43	51032	16.286	ug/l	99
44) Trichloroethene	7.135	130	30093	18.392	ug/l	99
45) 1,2-Dichloropropane	7.433	63	29323	18.429	ug/l	95
46) Dibromomethane	7.592	93	20791	18.181	ug/l	99
47) Bromodichloromethane	7.824	83	44034	18.870	ug/l	99
48) Methyl methacrylate	7.696	41	25618	16.301	ug/l	98
49) 1,4-Dioxane	7.659	88	6374	299.149	ug/l	94
51) 4-Methyl-2-Pentanone	8.573	43	152568	81.636	ug/l	98
52) Toluene	8.720	92	75814	19.048	ug/l	98
53) t-1,3-Dichloropropene	8.982	75	38942	17.985	ug/l	99
54) cis-1,3-Dichloropropene	8.366	75	44663	18.171	ug/l	98
55) 1,1,2-Trichloroethane	9.153	97	27505	18.544	ug/l	99
56) Ethyl methacrylate	9.116	69	37095	16.942	ug/l	98
57) 1,3-Dichloropropane	9.311	76	47213	18.484	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.244	63	101869	89.565	ug/l	99
59) 2-Hexanone	9.427	43	101986	78.935	ug/l	100
60) Dibromochloromethane	9.518	129	32154	18.395	ug/l	100
61) 1,2-Dibromoethane	9.610	107	27221	18.042	ug/l	100
64) Tetrachloroethene	9.275	164	25376	17.655	ug/l	95
65) Chlorobenzene	10.079	112	82392	17.991	ug/l	98
66) 1,1,1,2-Tetrachloroethane	10.159	131	27463	17.843	ug/l	98
67) Ethyl Benzene	10.195	91	144047	18.002	ug/l	97
68) m/p-Xylenes	10.299	106	111044	37.244	ug/l	100
69) o-Xylene	10.640	106	51894	18.572	ug/l	98
70) Styrene	10.652	104	89289	18.260	ug/l	99
71) Bromoform	10.799	173	19576	16.701	ug/l #	98
73) Isopropylbenzene	10.957	105	139510	18.100	ug/l	99
74) N-amyl acetate	10.841	43	46412	15.893	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	37312	17.337	ug/l	100
76) 1,2,3-Trichloropropane	11.238	75	29393m	15.335	ug/l	
77) Bromobenzene	11.195	156	33896	18.275	ug/l	99
78) n-propylbenzene	11.299	91	168589	18.415	ug/l	99
79) 2-Chlorotoluene	11.360	91	99687	18.238	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	115409	18.221	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	10102	14.897	ug/l	96
82) 4-Chlorotoluene	11.451	91	116713	18.296	ug/l	99
83) tert-Butylbenzene	11.713	119	118027	18.027	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	116817	18.357	ug/l	99
85) sec-Butylbenzene	11.890	105	153043	18.496	ug/l	100
86) p-Isopropyltoluene	12.006	119	127841	18.312	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	64330	17.788	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	64856	17.460	ug/l	99
89) n-Butylbenzene	12.329	91	118756	17.915	ug/l	100
90) Hexachloroethane	12.536	117	20652	16.893	ug/l	98
91) 1,2-Dichlorobenzene	12.335	146	61156	18.096	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	6425	15.122	ug/l	97
93) 1,2,4-Trichlorobenzene	13.585	180	41109	17.110	ug/l	98
94) Hexachlorobutadiene	13.719	225	17833	17.644	ug/l	96
95) Naphthalene	13.774	128	110507	16.144	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	40546	17.639	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
Data File : VX046762.D
Acq On : 19 Jun 2025 11:24
Operator : JC/MD
Sample : VX0619WBS01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01

Manual Integrations
APPROVED

Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025

Quant Time: Jun 20 05:16:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 03:09:16 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
Data File : VX046762.D
Acq On : 19 Jun 2025 11:24
Operator : JC/MD
Sample : VX0619WBS01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1


Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01

Manual Integrations APPROVED

Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025


Quant Time: Jun 20 05:16:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 03:09:16 2025
Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.562 min Scan# 734
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01



Tgt Ion:168 Resp: 137930
Ion Ratio Lower Upper
168 100
99 58.9 48.5 72.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#2
Dichlorodifluoromethane
Concen: 19.060 ug/l
RT: 1.185 min Scan# 16
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24


Tgt Ion: 85 Resp: 28069
Ion Ratio Lower Upper
85 100
87 31.5 16.6 49.8





#3
Chloromethane
Concen: 18.707 ug/l
RT: 1.307 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01



Tgt Ion: 50 Resp: 29730
Ion Ratio Lower Upper
50 100
52 31.9 24.9 37.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#4
Vinyl Chloride
Concen: 18.898 ug/l
RT: 1.386 min Scan# 49
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Tgt Ion: 62 Resp: 32033
Ion Ratio Lower Upper
62 100
64 33.8 23.9 35.9









#5
Bromomethane
Concen: 21.172 ug/l
RT: 1.630 min Scan# 89
Delta R.T. 0.006 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01

Tgt Ion: 94 Resp: 2019
Ion Ratio Lower Upper
94 100
96 97.2 72.3 108.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#6
Chloroethane
Concen: 19.923 ug/l
RT: 1.709 min Scan# 102
Delta R.T. 0.006 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24


Tgt Ion: 64 Resp: 20468
Ion Ratio Lower Upper
64 100
66 33.0 27.1 40.7





#7
Trichlorofluoromethane
Concen: 18.446 ug/l
RT: 1.910 min Scan# 11
Delta R.T. 0.006 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01



Tgt Ion:101 Resp: 47470
Ion Ratio Lower Upper
101 100
103 65.5 51.8 77.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#8
Diethyl Ether
Concen: 18.400 ug/l
RT: 2.148 min Scan# 174
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Tgt Ion: 74 Resp: 16252
Ion Ratio Lower Upper
74 100
45 104.0 52.3 157.1





#9

1,1,2-Trichlorotrifluoroethane

Concen: 18.528 ug/l

RT: 2.349 min Scan# 206

Delta R.T. 0.006 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBS01





Tgt Ion:101	Resp:	29254
Ion Ratio	Lower	Upper
101	100	
85	43.7	35.4 53.0
151	76.9	61.2 91.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#10

Methyl Iodide

Concen: 18.018 ug/l




RT: 2.471 min Scan# 227


Delta R.T. -0.000 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24

Tgt Ion:142	Resp:	29614
Ion Ratio	Lower	Upper
142	100	
127	45.8	37.0 55.6
141	13.4	11.5 17.3





#11

Tert butyl alcohol

Concen: 63.126 ug/l

RT: 2.959 min Scan# 307

Delta R.T. -0.000 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBS01

Tgt Ion: 59 Resp: 10810

Ion Ratio Lower Upper

59 100





57 9.3 8.3 12.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#12

1,1-Dichloroethene

Concen: 18.570 ug/l

RT: 2.337 min Scan# 205

Delta R.T. -0.000 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24




Tgt Ion: 96 Resp: 28292


Ion Ratio Lower Upper

96 100

61 163.0 126.6 189.8


98 63.7 50.7 76.1





#13
Acrolein
Concen: 72.910 ug/l
RT: 2.251 min Scan# 191
Delta R.T. 0.006 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01



Tgt Ion: 56 Resp: 18180
Ion Ratio Lower Upper
56 100
55 70.9 55.2 82.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#14
Allyl chloride
Concen: 17.868 ug/l
RT: 2.684 min Scan# 262
Delta R.T. 0.006 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Tgt Ion: 41 Resp: 49324
Ion Ratio Lower Upper
41 100
39 67.0 53.6 80.4
76 32.3 26.8 40.2





#15

Acrylonitrile

Concen: 81.119 ug/l

RT: 3.075 min Scan# 31

Delta R.T. -0.000 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBS01

Tgt Ion: 53 Resp: 6246

Ion Ratio Lower Upper

53 100

52 83.7 65.1 97.7





51 36.4 28.9 43.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#16

Acetone

Concen: 74.096 ug/l

RT: 2.386 min Scan# 213

Delta R.T. -0.000 min

Lab File: VX046762.D




Acq: 19 Jun 2025 11:24


Tgt Ion: 43 Resp: 47143

Ion Ratio Lower Upper

43 100

58 33.0 24.5 36.7









#17
Carbon Disulfide
Concen: 17.041 ug/l
RT: 2.532 min Scan# 21
Delta R.T. 0.006 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01

Tgt Ion: 76 Resp: 7840
Ion Ratio Lower Upper
76 100
78 10.2 7.5 11.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#18
Methyl Acetate
Concen: 16.004 ug/l
RT: 2.715 min Scan# 267
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Tgt Ion: 43 Resp: 25862
Ion Ratio Lower Upper
43 100
74 22.2 18.4 27.6





#19

Methyl tert-butyl Ether

Concen: 16.997 ug/l

RT: 3.123 min Scan# 31

Delta R.T. -0.000 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBS01

Tgt Ion: 73 Resp: 8063

Ion Ratio Lower Upper

73 100





57 21.9 18.2 27.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#20

Methylene Chloride

Concen: 18.113 ug/l

RT: 2.806 min Scan# 282

Delta R.T. 0.006 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24

Tgt Ion: 84 Resp: 32048




Ion Ratio Lower Upper


84 100

49 123.7 101.3 151.9

51 38.1 31.0 46.6

86 63.0 51.1 76.7









#21
trans-1,2-Dichloroethene
Concen: 18.253 ug/l
RT: 3.111 min Scan# 311
Delta R.T. 0.006 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01

Tgt Ion: 96 Resp: 2975
Ion Ratio Lower Upper
96 100
61 139.5 112.5 168.7
98 65.5 51.9 77.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#22
Diisopropyl ether
Concen: 18.921 ug/l
RT: 3.769 min Scan# 440
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Tgt Ion: 45 Resp: 100392
Ion Ratio Lower Upper
45 100
43 121.9 91.9 137.9
87 25.0 22.0 33.0
59 11.9 9.1 13.7





#23

Vinyl Acetate

Concen: 86.676 ug/l

RT: 3.733 min Scan# 41

Delta R.T. -0.000 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24

Tgt Ion: 43 Resp: 36571

Ion Ratio Lower Upper

43 100

86 10.1 8.2 12.2

Instrument :

MSVOA_X

ClientSampleId :





VX0619WBS01

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#24

1,1-Dichloroethane

Concen: 18.477 ug/l

RT: 3.623 min Scan# 416

Delta R.T. -0.000 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24




Tgt Ion: 63 Resp: 55663


Ion Ratio Lower Upper

63 100

98 8.5 4.0 11.9

100 4.8 2.2 6.6





#25

2-Butanone

Concen: 78.864 ug/l m

RT: 4.568 min Scan# 51

Delta R.T. 0.006 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBS01

Tgt Ion: 43 Resp: 70964

Ion Ratio Lower Upper

43 100





72 24.6 20.2 30.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#26

2,2-Dichloropropane

Concen: 17.951 ug/l

RT: 4.489 min Scan# 558

Delta R.T. -0.000 min

Lab File: VX046762.D




Acq: 19 Jun 2025 11:24


Tgt Ion: 77 Resp: 39442

Ion Ratio Lower Upper

77 100

97 22.8 11.8 35.3





#27

cis-1,2-Dichloroethene

Concen: 18.560 ug/l

RT: 4.507 min Scan# 560

Delta R.T. 0.006 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBS01

Tgt Ion: 96 Resp: 3554

Ion Ratio Lower Upper

96 100

61 142.4 0.0 291.0





98 64.5 0.0 126.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#28

Bromochloromethane

Concen: 20.816 ug/l

RT: 4.910 min Scan# 627

Delta R.T. -0.000 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24




Tgt Ion: 49 Resp: 28430


Ion Ratio Lower Upper

49 100

129 2.1 0.0 3.8

130 75.5 61.4 92.0





#29

Tetrahydrofuran

Concen: 76.721 ug/l

RT: 5.013 min Scan# 644

Delta R.T. -0.000 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBS01

Tgt Ion: 42 Resp: 4459

Ion Ratio Lower Upper

42 100

72 43.4 34.9 52.3





71 41.1 32.2 48.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#30

Chloroform

Concen: 18.982 ug/l

RT: 5.105 min Scan# 659

Delta R.T. -0.000 min

Lab File: VX046762.D




Acq: 19 Jun 2025 11:24


Tgt Ion: 83 Resp: 57187

Ion Ratio Lower Upper

83 100


85 61.3 53.7 80.5





#31
Cyclohexane
Concen: 18.836 ug/l
RT: 5.483 min Scan# 706
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01



Tgt Ion: 56 Resp: 5381
Ion Ratio Lower Upper
56 100
69 29.2 24.8 37.2
84 81.2 68.2 102.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#32
1,1,1-Trichloroethane
Concen: 17.665 ug/l
RT: 5.391 min Scan# 706
Delta R.T. -0.006 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Tgt Ion: 97 Resp: 46675
Ion Ratio Lower Upper
97 100
99 64.5 51.3 76.9
61 47.4 36.4 54.6





#33

1,2-Dichloroethane-d4

Concen: 49.765 ug/l

RT: 5.964 min Scan# 800

Delta R.T. -0.000 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBS01

Tgt Ion: 65 Resp: 94940

Ion Ratio Lower Upper

65 100





67 51.8 0.0 105.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.769 min Scan# 932

Delta R.T. -0.000 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24




Tgt Ion:114 Resp: 227140


Ion Ratio Lower Upper

114 100

63 20.5 0.0 44.2

88 15.8 0.0 33.2





#35

Dibromofluoromethane

Concen: 52.032 ug/l

RT: 5.397 min Scan# 707

Delta R.T. -0.000 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBS01

Tgt Ion:113 Resp: 78190

Ion Ratio Lower Upper

113 100

111 102.8 82.0 123.0





192 18.8 15.3 22.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#36

1,1-Dichloropropene

Concen: 18.754 ug/l

RT: 5.702 min Scan# 757

Delta R.T. -0.000 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24




Tgt Ion: 75 Resp: 40684


Ion Ratio Lower Upper

75 100

110 34.0 17.2 51.5

77 31.1 24.2 36.4





#37

Ethyl Acetate

Concen: 15.895 ug/l

RT: 4.727 min Scan# 596

Delta R.T. 0.006 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBS01

Tgt Ion: 43 Resp: 31829

Ion Ratio Lower Upper

43 100

61 12.9 9.7 14.5





70 9.5 8.1 12.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#38

Carbon Tetrachloride

Concen: 18.031 ug/l

RT: 5.690 min Scan# 755

Delta R.T. -0.000 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24




Tgt Ion:117 Resp: 42451


Ion Ratio Lower Upper

117 100

119 94.7 75.2 112.8

121 28.0 24.1 36.1





#39

Methylcyclohexane

Concen: 18.541 ug/l

RT: 7.385 min Scan# 1033

Delta R.T. -0.000 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBS01

Tgt Ion: 83 Resp: 5293

Ion Ratio Lower Upper

83 100

55 78.7 63.0 94.4





98 49.6 38.8 58.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#40

Benzene

Concen: 18.864 ug/l

RT: 6.050 min Scan# 814

Delta R.T. 0.006 min

Lab File: VX046762.D




Acq: 19 Jun 2025 11:24


Tgt Ion: 78 Resp: 121111

Ion Ratio Lower Upper

78 100

77 23.0 18.8 28.2





#41

Methacrylonitrile

Concen: 16.137 ug/l

RT: 4.928 min Scan# 61

Delta R.T. -0.000 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBS01

Tgt Ion: 41 Resp: 1722

Ion Ratio Lower Upper

41 100

39 58.4 45.9 68.9

67 77.6 61.2 91.8





52 32.7 25.5 38.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#42

1,2-Dichloroethane

Concen: 18.518 ug/l

RT: 6.092 min Scan# 821

Delta R.T. -0.000 min

Lab File: VX046762.D




Acq: 19 Jun 2025 11:24


Tgt Ion: 62 Resp: 41780

Ion Ratio Lower Upper

62 100

98 9.0 0.0 17.8





#43

Isopropyl Acetate

Concen: 16.286 ug/l

RT: 6.348 min Scan# 80

Delta R.T. 0.006 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBS01

Tgt Ion: 43 Resp: 5103

Ion Ratio Lower Upper

43 100

61 18.7 15.3 22.9





87 12.0 10.0 15.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#44

Trichloroethene

Concen: 18.392 ug/l

RT: 7.135 min Scan# 992

Delta R.T. 0.006 min

Lab File: VX046762.D




Acq: 19 Jun 2025 11:24


Tgt Ion:130 Resp: 30093

Ion Ratio Lower Upper

130 100

95 102.2 0.0 202.6





#45

1,2-Dichloropropane

Concen: 18.429 ug/l

RT: 7.433 min Scan# 1041

Delta R.T. -0.000 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBS01

Tgt Ion: 63 Resp: 2932

Ion Ratio Lower Upper

63 100





65 31.0 22.9 34.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#46

Dibromomethane

Concen: 18.181 ug/l

RT: 7.592 min Scan# 1067

Delta R.T. 0.006 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24




Tgt Ion: 93 Resp: 20791


Ion Ratio Lower Upper

93 100

95 83.0 66.2 99.4

174 97.3 77.0 115.6





#47

Bromodichloromethane

Concen: 18.870 ug/l

RT: 7.824 min Scan# 1105

Delta R.T. -0.000 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBS01





Tgt Ion:	83	Resp:	44034
Ion	Ratio	Lower	Upper
83	100		
85	63.8	51.4	77.2
127	9.3	6.9	10.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#48

Methyl methacrylate

Concen: 16.301 ug/l




RT: 7.696 min Scan# 1084


Delta R.T. -0.000 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24

Tgt Ion:	41	Resp:	25618
Ion	Ratio	Lower	Upper
41	100		
69	82.0	63.4	95.2
39	57.0	45.4	68.0





#49

1,4-Dioxane

Concen: 299.149 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. 0.000 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBS01

Tgt Ion: 88 Resp: 6374

Ion Ratio Lower Upper

88 100

43 38.3 31.2 46.8





58 75.1 54.7 82.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#50

Toluene-d8

Concen: 52.340 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. -0.000 min

Lab File: VX046762.D




Acq: 19 Jun 2025 11:24


Tgt Ion: 98 Resp: 282602

Ion Ratio Lower Upper

98 100

100 66.6 53.0 79.4









#51
4-Methyl-2-Pentanone
Concen: 81.636 ug/l
RT: 8.573 min Scan# 1228
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01

Tgt Ion: 43 Resp: 15256
Ion Ratio Lower Upper
43 100
58 37.5 30.8 46.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#52
Toluene
Concen: 19.048 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Tgt Ion: 92 Resp: 75814
Ion Ratio Lower Upper
92 100
91 170.6 134.8 202.2





#53

t-1,3-Dichloropropene

Concen: 17.985 ug/l

RT: 8.982 min Scan# 1294

Delta R.T. 0.006 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBS01

Tgt Ion: 75 Resp: 3894

Ion Ratio Lower Upper

75 100





77 31.6 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#54

cis-1,3-Dichloropropene

Concen: 18.171 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24




Tgt Ion: 75 Resp: 44663


Ion Ratio Lower Upper

75 100

77 28.9 25.3 37.9

39 48.7 39.2 58.8





#55

1,1,2-Trichloroethane

Concen: 18.544 ug/l

RT: 9.153 min Scan# 1323

Delta R.T. -0.000 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBS01

Tgt Ion: 97 Resp: 2750

Ion Ratio Lower Upper

97 100

83 81.4 67.2 100.8

85 54.6 44.1 66.1





99 63.1 50.3 75.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#56

Ethyl methacrylate

Concen: 16.942 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. -0.000 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24




Tgt Ion: 69 Resp: 37095


Ion Ratio Lower Upper

69 100

41 70.6 55.4 83.2

39 36.2 29.7 44.5





#57

1,3-Dichloropropane

Concen: 18.484 ug/l

RT: 9.311 min Scan# 1349

Delta R.T. 0.006 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBS01

Tgt Ion: 76 Resp: 4721

Ion Ratio Lower Upper

76 100





78 32.0 25.7 38.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#58

2-Chloroethyl Vinyl ether

Concen: 89.565 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. -0.000 min

Lab File: VX046762.D




Acq: 19 Jun 2025 11:24


Tgt Ion: 63 Resp: 101869

Ion Ratio Lower Upper

63 100

106 27.8 22.6 34.0





#59

2-Hexanone

Concen: 78.935 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. -0.000 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBS01

Tgt Ion: 43 Resp: 101980

Ion Ratio Lower Upper

43 100





58 53.7 26.8 80.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#60

Dibromochloromethane

Concen: 18.395 ug/l

RT: 9.518 min Scan# 1383

Delta R.T. -0.000 min

Lab File: VX046762.D




Acq: 19 Jun 2025 11:24


Tgt Ion:129 Resp: 32154

Ion Ratio Lower Upper

129 100


127 76.9 38.4 115.2





#61
1,2-Dibromoethane
Concen: 18.042 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01



Tgt Ion: 107 Resp: 2722
Ion Ratio Lower Upper
107 100
109 94.6 75.8 113.6




Manual Integrations
APPROVED


Reviewed By : John Carlone 06/20/2025
Supervised By : Mahesh Dadoda 06/21/2025



#62
4-Bromofluorobenzene
Concen: 54.092 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24


Tgt Ion: 95 Resp: 108869
Ion Ratio Lower Upper
95 100
174 75.4 0.0 150.4
176 71.8 0.0 145.0





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01



Tgt Ion:117 Resp: 20746
Ion Ratio Lower Upper
117 100
82 55.6 44.6 66.8
119 31.9 25.8 38.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#64
Tetrachloroethene
Concen: 17.655 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Tgt Ion:164 Resp: 25376
Ion Ratio Lower Upper
164 100
166 121.4 104.6 157.0
129 99.6 76.3 114.5
131 93.8 74.1 111.1









#65
Chlorobenzene
Concen: 17.991 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01

Tgt Ion:112 Resp: 82392
Ion Ratio Lower Upper
112 100
114 31.0 25.5 38.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 17.843 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Tgt Ion:131 Resp: 27463
Ion Ratio Lower Upper
131 100
133 96.2 47.3 141.8
119 66.0 32.1 96.3









#67
Ethyl Benzene
Concen: 18.002 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.006 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01

Tgt Ion: 91 Resp: 14404
Ion Ratio Lower Upper
91 100
106 31.7 24.2 36.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#68
m/p-Xylenes
Concen: 37.244 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Tgt Ion:106 Resp: 111044
Ion Ratio Lower Upper
106 100
91 205.3 164.8 247.2









#69
o-Xylene
Concen: 18.572 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01

Tgt Ion:106 Resp: 51894
Ion Ratio Lower Upper
106 100
91 219.2 111.3 333.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#70
Styrene
Concen: 18.260 ug/l
RT: 10.652 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24


Tgt Ion:104 Resp: 89289
Ion Ratio Lower Upper
104 100
78 52.0 41.4 62.2
103 54.7 43.1 64.7





#71
Bromoform
Concen: 16.701 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01



Tgt Ion:173 Resp: 19570
Ion Ratio Lower Upper
173 100
175 50.0 24.1 72.4
254 0.0 0.1 0.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24


Tgt Ion:152 Resp: 104652
Ion Ratio Lower Upper
152 100
115 70.8 43.2 129.6
150 164.1 0.0 346.8





#73
Isopropylbenzene
Concen: 18.100 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01



Tgt Ion:105 Resp: 139510
Ion Ratio Lower Upper
105 100
120 26.3 13.0 39.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#74
N-ethyl acetate
Concen: 15.893 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Tgt Ion: 43 Resp: 46412
Ion Ratio Lower Upper
43 100
70 41.3 33.4 50.0
55 24.2 19.1 28.7
61 23.0 19.0 28.4









#75
1,1,2,2-Tetrachloroethane
Concen: 17.337 ug/l
RT: 11.207 min Scan# 1660
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01

Tgt Ion: 83 Resp: 3731
Ion Ratio Lower Upper
83 100
131 9.7 4.9 14.6
85 63.7 32.0 96.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#76
1,2,3-Trichloropropane
Concen: 15.335 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Tgt Ion: 75 Resp: 29393
Ion Ratio Lower Upper
75 100
77 43.8 18.5 55.5





#77

Bromobenzene

Concen: 18.275 ug/l

RT: 11.195 min Scan# 1658

Delta R.T. -0.000 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBS01

Tgt Ion:156 Resp: 33890

Ion Ratio Lower Upper

156 100

77 164.7 82.8 248.6





158 97.8 47.6 142.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#78

n-propylbenzene

Concen: 18.415 ug/l

RT: 11.299 min Scan# 1675

Delta R.T. -0.000 min

Lab File: VX046762.D




Acq: 19 Jun 2025 11:24


Tgt Ion: 91 Resp: 168589

Ion Ratio Lower Upper

91 100

120 22.3 11.4 34.1









#79
2-Chlorotoluene
Concen: 18.238 ug/l
RT: 11.360 min Scan# 1100
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01

Tgt Ion: 91 Resp: 9968
Ion Ratio Lower Upper
91 100
126 32.5 16.6 49.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#80
1,3,5-Trimethylbenzene
Concen: 18.221 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24


Tgt Ion:105 Resp: 115409
Ion Ratio Lower Upper
105 100
120 48.8 24.1 72.2





#81
trans-1,4-Dichloro-2-butene
Concen: 14.897 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01



Tgt Ion: 75 Resp: 1010
Ion Ratio Lower Upper
75 100
53 112.7 86.8 130.2
89 43.9 36.9 55.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#82
4-Chlorotoluene
Concen: 18.296 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24


Tgt Ion: 91 Resp: 116713
Ion Ratio Lower Upper
91 100
126 29.3 14.4 43.2





#83
tert-Butylbenzene
Concen: 18.027 ug/l
RT: 11.713 min Scan# 1174
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01



Tgt Ion:119 Resp: 11802
Ion Ratio Lower Upper
119 100
91 58.4 29.5 88.6
134 23.0 11.7 35.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#84
1,2,4-Trimethylbenzene
Concen: 18.357 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Tgt Ion:105 Resp: 116817
Ion Ratio Lower Upper
105 100
120 44.6 22.1 66.3









#85
 sec-Butylbenzene
 Concen: 18.496 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. -0.000 min
 Lab File: VX046762.D
 Acq: 19 Jun 2025 11:24

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0619WBS01

Tgt Ion:105 Resp: 15304
 Ion Ratio Lower Upper
 105 100
 134 19.7 9.8 29.4




Manual Integrations
 APPROVED


Reviewed By :John Carlone 06/20/2025
 Supervised By :Mahesh Dadoda 06/21/2025



#86
 p-Isopropyltoluene
 Concen: 18.312 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. -0.000 min
 Lab File: VX046762.D
 Acq: 19 Jun 2025 11:24

Tgt Ion:119 Resp: 127841
 Ion Ratio Lower Upper
 119 100
 134 25.1 12.9 38.7
 91 25.2 12.3 36.9





#87

1,3-Dichlorobenzene

Concen: 17.788 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBS01

Tgt Ion:146 Resp: 64330

Ion Ratio Lower Upper

146 100

111 40.8 20.6 61.8





148 62.5 31.6 94.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#88

1,4-Dichlorobenzene

Concen: 17.460 ug/l

RT: 12.036 min Scan# 1796

Delta R.T. -0.000 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24




Tgt Ion:146 Resp: 64856


Ion Ratio Lower Upper

146 100

111 42.3 20.3 61.0

148 64.0 32.1 96.5





#89

n-Butylbenzene

Concen: 17.915 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. -0.000 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBS01

Tgt Ion: 91 Resp: 118750

Ion Ratio Lower Upper

91 100

92 54.0 27.1 81.3





134 25.7 12.9 38.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#90

Hexachloroethane

Concen: 16.893 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX046762.D




Acq: 19 Jun 2025 11:24


Tgt Ion:117 Resp: 20652

Ion Ratio Lower Upper

117 100


201 70.7 34.4 103.3





#91
1,2-Dichlorobenzene
Concen: 18.096 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.006 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01



Tgt Ion:146 Resp: 61150
Ion Ratio Lower Upper
146 100
111 42.8 21.7 65.1
148 64.4 32.0 96.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 15.122 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Tgt Ion: 75 Resp: 6425
Ion Ratio Lower Upper
75 100
155 80.9 41.9 125.7
157 105.8 51.4 154.3





#93

1,2,4-Trichlorobenzene

Concen: 17.110 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. -0.000 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBS01

Tgt Ion:180 Resp: 41109

Ion Ratio Lower Upper

180 100

182 95.7 46.7 140.1





145 31.1 15.8 47.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#94

Hexachlorobutadiene

Concen: 17.644 ug/l

RT: 13.719 min Scan# 2072

Delta R.T. -0.000 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24




Tgt Ion:225 Resp: 17833


Ion Ratio Lower Upper

225 100

223 62.5 32.5 97.5


227 61.1 32.2 96.6





#95
Naphthalene
Concen: 16.144 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01



Tgt Ion:128 Resp: 11050
Ion Ratio Lower Upper
128 100
127 12.6 10.1 15.1
129 11.0 8.7 13.1




Manual Integrations
APPROVED

Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#96
1,2,3-Trichlorobenzene
Concen: 17.639 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Tgt Ion:180 Resp: 40546
Ion Ratio Lower Upper
180 100
182 94.7 47.8 143.4
145 34.2 17.3 51.9



Report of Analysis

Client:	AECOM Technical Services, Inc.	Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258	Date Received:	
Client Sample ID:	VX0619WBSD01	SDG No.:	Q2361
Lab Sample ID:	VX0619WBSD01	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046763.D	1		06/19/25 11:51	VX061925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	19.8		0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	19.5		0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	19.7		0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	22.2		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	20.6		0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	19.3		0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	19.7		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	19.5		0.23	0.75	1.00	ug/L
67-64-1	Acetone	81.5		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	18.0		0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	18.7		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	17.8		0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	19.3		0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	19.2		0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	19.4		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	19.7		1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	87.0		0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	19.1		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	19.4		0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	23.1		0.22	0.50	1.00	ug/L
67-66-3	Chloroform	20.5		0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	19.3		0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	19.0		0.16	0.50	1.00	ug/L
71-43-2	Benzene	20.0		0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	20.1		0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	19.3		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	20.3		0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	20.1		0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	91.8		0.68	2.50	5.00	ug/L
108-88-3	Toluene	20.3		0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0619WBSD01		SDG No.:	Q2361
Lab Sample ID:	VX0619WBSD01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046763.D	1		06/19/25 11:51	VX061925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	19.5		0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	19.9		0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	20.4		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	88.8		0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	19.9		0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	20.1		0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	18.8		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	19.4		0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	19.1		0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	38.8		0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	19.8		0.12	0.50	1.00	ug/L
100-42-5	Styrene	19.7		0.15	0.50	1.00	ug/L
75-25-2	Bromoform	18.4		0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	18.6		0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	18.4		0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	18.6		0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	17.8		0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	19.0		0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	16.1		0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	18.3		0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	18.2		0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	51.5		81 - 118		103%	SPK: 50
1868-53-7	Dibromofluoromethane	52.8		80 - 119		106%	SPK: 50
2037-26-5	Toluene-d8	52.7		89 - 112		105%	SPK: 50
460-00-4	4-Bromofluorobenzene	54.9		85 - 114		110%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	120000	5.562				
540-36-3	1,4-Difluorobenzene	199000	6.763				
3114-55-4	Chlorobenzene-d5	184000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	94900	12.018				



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0619WBSD01		SDG No.:	Q2361
Lab Sample ID:	VX0619WBSD01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046763.D	1		06/19/25 11:51	VX061925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
 Data File : VX046763.D
 Acq On : 19 Jun 2025 11:51
 Operator : JC/MD
 Sample : VX0619WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0619WBSD01

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/20/2025
 Supervised By : Mahesh Dadoda 06/21/2025

Quant Time: Jun 20 05:17:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	120371	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	199496	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	183807	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	94851	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	85794	51.530	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 103.060%			
35) Dibromofluoromethane	5.391	113	69686	52.795	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 105.600%			
50) Toluene-d8	8.647	98	249730	52.661	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 105.320%			
62) 4-Bromofluorobenzene	11.079	95	97074	54.915	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 109.820%			
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.185	85	25471	19.820	ug/l	99
3) Chloromethane	1.307	50	27061	19.509	ug/l	96
4) Vinyl Chloride	1.386	62	29098	19.672	ug/l	95
5) Bromomethane	1.630	94	18436	22.153	ug/l	93
6) Chloroethane	1.703	64	18444	20.573	ug/l	95
7) Trichlorofluoromethane	1.904	101	43435	19.339	ug/l	97
8) Diethyl Ether	2.148	74	15726	20.402	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.349	101	27115	19.679	ug/l	98
10) Methyl Iodide	2.465	142	30018	20.229	ug/l	98
11) Tert butyl alcohol	2.959	59	10932	73.154	ug/l	96
12) 1,1-Dichloroethene	2.337	96	25867	19.455	ug/l	98
13) Acrolein	2.245	56	18810	86.416	ug/l	97
14) Allyl chloride	2.678	41	45893	19.051	ug/l	100
15) Acrylonitrile	3.068	53	61358	91.312	ug/l	99
16) Acetone	2.386	43	45269	81.534	ug/l	95
17) Carbon Disulfide	2.526	76	72150	17.970	ug/l	99
18) Methyl Acetate	2.709	43	25035	17.753	ug/l	100
19) Methyl tert-butyl Ether	3.117	73	77256	18.661	ug/l	98
20) Methylene Chloride	2.800	84	29836	19.323	ug/l	99
21) trans-1,2-Dichloroethene	3.105	96	27280	19.176	ug/l	97
22) Diisopropyl ether	3.764	45	94092	20.321	ug/l	87
23) Vinyl Acetate	3.733	43	359723	97.697	ug/l	100
24) 1,1-Dichloroethane	3.623	63	50927	19.372	ug/l	100
25) 2-Butanone	4.562	43	68340	87.030	ug/l	97
26) 2,2-Dichloropropane	4.489	77	36206	18.883	ug/l	99
27) cis-1,2-Dichloroethene	4.501	96	32489	19.438	ug/l	100
28) Bromochloromethane	4.910	49	27522	23.092	ug/l	97
29) Tetrahydrofuran	5.013	42	43894	86.542	ug/l	99
30) Chloroform	5.099	83	53771	20.452	ug/l	97
31) Cyclohexane	5.477	56	49014	19.660	ug/l	98
32) 1,1,1-Trichloroethane	5.391	97	44445	19.276	ug/l	99
36) 1,1-Dichloropropene	5.702	75	35819	18.799	ug/l	97
37) Ethyl Acetate	4.721	43	31001	17.627	ug/l	98
38) Carbon Tetrachloride	5.684	117	39531	19.117	ug/l	99
39) Methylcyclohexane	7.385	83	47544	18.962	ug/l	99
40) Benzene	6.044	78	112866	20.015	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
 Data File : VX046763.D
 Acq On : 19 Jun 2025 11:51
 Operator : JC/MD
 Sample : VX0619WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0619WBSD01

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/20/2025
 Supervised By : Mahesh Dadoda 06/21/2025

Quant Time: Jun 20 05:17:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.928	41	17801	18.985	ug/l	96
42) 1,2-Dichloroethane	6.092	62	39829	20.100	ug/l	99
43) Isopropyl Acetate	6.342	43	50813	18.463	ug/l	99
44) Trichloroethene	7.129	130	27765	19.321	ug/l	89
45) 1,2-Dichloropropane	7.427	63	28336	20.276	ug/l	92
46) Dibromomethane	7.580	93	19978	19.891	ug/l	99
47) Bromodichloromethane	7.824	83	41221	20.112	ug/l	99
48) Methyl methacrylate	7.696	41	26092	18.904	ug/l	100
49) 1,4-Dioxane	7.659	88	6068	321.885	ug/l	97
51) 4-Methyl-2-Pentanone	8.574	43	150700	91.811	ug/l	98
52) Toluene	8.720	92	70998	20.310	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	37151	19.536	ug/l	98
54) cis-1,3-Dichloropropene	8.366	75	43044	19.939	ug/l	100
55) 1,1,2-Trichloroethane	9.153	97	26588	20.409	ug/l	98
56) Ethyl methacrylate	9.116	69	35838	18.636	ug/l	97
57) 1,3-Dichloropropane	9.305	76	45490	20.277	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.238	63	100239	100.344	ug/l	99
59) 2-Hexanone	9.427	43	100820	88.846	ug/l	100
60) Dibromochloromethane	9.519	129	30624	19.947	ug/l	99
61) 1,2-Dibromoethane	9.610	107	26614	20.084	ug/l	98
64) Tetrachloroethene	9.275	164	23913	18.778	ug/l	98
65) Chlorobenzene	10.079	112	78801	19.421	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.159	131	26522	19.449	ug/l	100
67) Ethyl Benzene	10.195	91	135702	19.141	ug/l	99
68) m/p-Xylenes	10.299	106	102519	38.810	ug/l	99
69) o-Xylene	10.640	106	49068	19.820	ug/l	99
70) Styrene	10.653	104	85333	19.696	ug/l	98
71) Bromoform	10.799	173	19127	18.418	ug/l #	98
73) Isopropylbenzene	10.957	105	129881	18.592	ug/l	100
74) N-amyl acetate	10.842	43	45562	17.214	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	35939	18.425	ug/l	100
76) 1,2,3-Trichloropropane	11.238	75	29301m	16.867	ug/l	
77) Bromobenzene	11.195	156	31899	18.976	ug/l	98
78) n-propylbenzene	11.299	91	156018	18.803	ug/l	100
79) 2-Chlorotoluene	11.360	91	92310	18.633	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	108639	18.925	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	9561	15.556	ug/l	95
82) 4-Chlorotoluene	11.451	91	108666	18.795	ug/l	99
83) tert-Butylbenzene	11.713	119	108810	18.337	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	108849	18.872	ug/l	100
85) sec-Butylbenzene	11.890	105	141281	18.839	ug/l	100
86) p-Isopropyltoluene	12.006	119	118528	18.732	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	60812	18.553	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	59834	17.772	ug/l	98
89) n-Butylbenzene	12.329	91	111302	18.526	ug/l	99
90) Hexachloroethane	12.536	117	19081	17.220	ug/l	96
91) 1,2-Dichlorobenzene	12.335	146	58188	18.997	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	6206	16.116	ug/l	97
93) 1,2,4-Trichlorobenzene	13.585	180	39871	18.309	ug/l	99
94) Hexachlorobutadiene	13.719	225	17144	18.715	ug/l	96
95) Naphthalene	13.774	128	107578	17.340	ug/l	99
96) 1,2,3-Trichlorobenzene	13.957	180	37893	18.189	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
Data File : VX046763.D
Acq On : 19 Jun 2025 11:51
Operator : JC/MD
Sample : VX0619WBSD01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01

Manual Integrations
APPROVED

Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025

Quant Time: Jun 20 05:17:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 03:09:16 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
Data File : VX046763.D
Acq On : 19 Jun 2025 11:51
Operator : JC/MD
Sample : VX0619WBSD01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1


Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01

Manual Integrations

Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025


Quant Time: Jun 20 05:17:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 03:09:16 2025
Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.562 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01



Tgt Ion:168 Resp: 12037
Ion Ratio Lower Upper
168 100
99 61.3 48.5 72.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#2
Dichlorodifluoromethane
Concen: 19.820 ug/l
RT: 1.185 min Scan# 16
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Tgt Ion: 85 Resp: 25471
Ion Ratio Lower Upper
85 100
87 32.8 16.6 49.8









#3
Chloromethane
Concen: 19.509 ug/l
RT: 1.307 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01

Tgt Ion: 50 Resp: 2706
Ion Ratio Lower Upper
50 100
52 33.3 24.9 37.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#4
Vinyl Chloride
Concen: 19.672 ug/l
RT: 1.386 min Scan# 49
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51


Tgt Ion: 62 Resp: 29098
Ion Ratio Lower Upper
62 100
64 32.5 23.9 35.9





#5
Bromomethane
Concen: 22.153 ug/l
RT: 1.630 min Scan# 89
Delta R.T. 0.006 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01



Tgt Ion: 94 Resp: 18436
Ion Ratio Lower Upper
94 100
96 96.8 72.3 108.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#6
Chloroethane
Concen: 20.573 ug/l
RT: 1.703 min Scan# 101
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51


Tgt Ion: 64 Resp: 18444
Ion Ratio Lower Upper
64 100
66 30.9 27.1 40.7





#7
Trichlorofluoromethane
Concen: 19.339 ug/l
RT: 1.904 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51



Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01



Tgt Ion:101 Resp: 4343
Ion Ratio Lower Upper
101 100
103 62.7 51.8 77.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#8
Diethyl Ether
Concen: 20.402 ug/l
RT: 2.148 min Scan# 174
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Tgt Ion: 74 Resp: 15726
Ion Ratio Lower Upper
74 100
45 106.1 52.3 157.1





#9

1,1,2-Trichlorotrifluoroethane

Concen: 19.679 ug/l

RT: 2.349 min Scan# 206

Delta R.T. 0.006 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBSD01

Tgt Ion:101 Resp: 2711

Ion Ratio Lower Upper

101 100

85 42.4 35.4 53.0





151 75.5 61.2 91.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#10

Methyl Iodide

Concen: 20.229 ug/l

RT: 2.465 min Scan# 226

Delta R.T. -0.006 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51




Tgt Ion:142 Resp: 30018


Ion Ratio Lower Upper

142 100

127 44.9 37.0 55.6

141 14.0 11.5 17.3





#11

Tert butyl alcohol

Concen: 73.154 ug/l

RT: 2.959 min Scan# 307

Delta R.T. -0.000 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBSD01

Tgt Ion: 59 Resp: 1093

Ion Ratio Lower Upper

59 100





57 9.0 8.3 12.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#12

1,1-Dichloroethene

Concen: 19.455 ug/l

RT: 2.337 min Scan# 205

Delta R.T. -0.000 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51




Tgt Ion: 96 Resp: 25867


Ion Ratio Lower Upper

96 100

61 162.1 126.6 189.8

98 63.6 50.7 76.1





#13

Acrolein

Concen: 86.416 ug/l

RT: 2.245 min Scan# 190

Delta R.T. -0.000 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBSD01

Tgt Ion: 56 Resp: 18810

Ion Ratio Lower Upper

56 100





55 71.1 55.2 82.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#14

Allyl chloride

Concen: 19.051 ug/l

RT: 2.678 min Scan# 261

Delta R.T. -0.000 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51




Tgt Ion: 41 Resp: 45893


Ion Ratio Lower Upper

41 100

39 67.1 53.6 80.4

76 33.3 26.8 40.2





#15

Acrylonitrile

Concen: 91.312 ug/l

RT: 3.068 min Scan# 31

Delta R.T. -0.006 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBSD01

Tgt Ion: 53 Resp: 61358

Ion Ratio Lower Upper

53 100

52 82.1 65.1 97.7





51 36.0 28.9 43.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#16

Acetone

Concen: 81.534 ug/l

RT: 2.386 min Scan# 213

Delta R.T. -0.000 min

Lab File: VX046763.D




Acq: 19 Jun 2025 11:51


Tgt Ion: 43 Resp: 45269

Ion Ratio Lower Upper

43 100


58 33.3 24.5 36.7





#17
Carbon Disulfide
Concen: 17.970 ug/l
RT: 2.526 min Scan# 21
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01



Tgt Ion: 76 Resp: 72150
Ion Ratio Lower Upper
76 100
78 9.2 7.5 11.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#18
Methyl Acetate
Concen: 17.753 ug/l
RT: 2.709 min Scan# 266
Delta R.T. -0.006 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Tgt Ion: 43 Resp: 25035
Ion Ratio Lower Upper
43 100
74 23.0 18.4 27.6





#19

Methyl tert-butyl Ether

Concen: 18.661 ug/l

RT: 3.117 min Scan# 31

Delta R.T. -0.006 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBSD01

Tgt Ion: 73 Resp: 77250

Ion Ratio Lower Upper

73 100





57 21.9 18.2 27.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#20

Methylene Chloride

Concen: 19.323 ug/l

RT: 2.800 min Scan# 281

Delta R.T. -0.000 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51

Tgt Ion: 84 Resp: 29836




Ion Ratio Lower Upper


84 100

49 126.4 101.3 151.9

51 38.8 31.0 46.6

86 66.4 51.1 76.7









#21
trans-1,2-Dichloroethene
Concen: 19.176 ug/l
RT: 3.105 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01

Tgt Ion: 96 Resp: 27280
Ion Ratio Lower Upper
96 100
61 138.8 112.5 168.7
98 60.0 51.9 77.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#22
Diisopropyl ether
Concen: 20.321 ug/l
RT: 3.764 min Scan# 439
Delta R.T. -0.006 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Tgt Ion: 45 Resp: 94092
Ion Ratio Lower Upper
45 100
43 133.4 91.9 137.9
87 25.5 22.0 33.0
59 11.3 9.1 13.7





#23

Vinyl Acetate

Concen: 97.697 ug/l

RT: 3.733 min Scan# 41

Delta R.T. -0.000 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51

Tgt Ion: 43 Resp: 35972

Ion Ratio Lower Upper

43 100

86 10.3 8.2 12.2

Instrument :

MSVOA_X

ClientSampleId :





VX0619WBSD01

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#24

1,1-Dichloroethane

Concen: 19.372 ug/l

RT: 3.623 min Scan# 416

Delta R.T. -0.000 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51




Tgt Ion: 63 Resp: 50927


Ion Ratio Lower Upper

63 100

98 8.0 4.0 11.9

100 4.5 2.2 6.6









#25
2-Butanone
Concen: 87.030 ug/l
RT: 4.562 min Scan# 570
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01

Tgt Ion: 43 Resp: 68340
Ion Ratio Lower Upper
43 100
72 24.0 20.2 30.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#26
2,2-Dichloropropane
Concen: 18.883 ug/l
RT: 4.489 min Scan# 558
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Tgt Ion: 77 Resp: 36206
Ion Ratio Lower Upper
77 100
97 22.8 11.8 35.3





#27

cis-1,2-Dichloroethene

Concen: 19.438 ug/l

RT: 4.501 min Scan# 560

Delta R.T. -0.000 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBSD01

Tgt Ion: 96 Resp: 32489

Ion Ratio Lower Upper

96 100

61 146.1 0.0 291.0





98 63.4 0.0 126.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#28

Bromochloromethane

Concen: 23.092 ug/l

RT: 4.910 min Scan# 627

Delta R.T. -0.000 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51




Tgt Ion: 49 Resp: 27522


Ion Ratio Lower Upper

49 100

129 1.6 0.0 3.8

130 73.9 61.4 92.0





#29

Tetrahydrofuran

Concen: 86.542 ug/l

RT: 5.013 min Scan# 644

Delta R.T. -0.000 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBSD01





Tgt Ion:	42	Resp:	43894
Ion Ratio	Lower	Upper	
42	100		
72	44.2	34.9	52.3
71	39.6	32.2	48.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#30

Chloroform

Concen: 20.452 ug/l




RT: 5.099 min Scan# 658


Delta R.T. -0.006 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51

Tgt Ion:	83	Resp:	53771
Ion Ratio	Lower	Upper	
83	100		
85	64.5	53.7	80.5





#31

Cyclohexane

Concen: 19.660 ug/l

RT: 5.477 min Scan# 701

Delta R.T. -0.006 min

Lab File: VX046763.D


Acq: 19 Jun 2025 11:51

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBSD01



Tgt Ion: 56 Resp: 49014

Ion Ratio Lower Upper

56 100

69 27.9 24.8 37.2




84 85.3 68.2 102.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#32

1,1,1-Trichloroethane

Concen: 19.276 ug/l

RT: 5.391 min Scan# 706

Delta R.T. -0.006 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51




Tgt Ion: 97 Resp: 44445


Ion Ratio Lower Upper

97 100

99 64.2 51.3 76.9

61 46.4 36.4 54.6





#33

1,2-Dichloroethane-d4

Concen: 51.530 ug/l

RT: 5.958 min Scan# 799

Delta R.T. -0.006 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBSD01

Tgt Ion: 65 Resp: 85794

Ion Ratio Lower Upper

65 100





67 52.5 0.0 105.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. -0.006 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51




Tgt Ion:114 Resp: 199496


Ion Ratio Lower Upper

114 100

63 20.4 0.0 44.2

88 16.2 0.0 33.2





#35

Dibromofluoromethane

Concen: 52.795 ug/l

RT: 5.391 min Scan# 707

Delta R.T. -0.006 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBSD01

Tgt Ion:113 Resp: 69680

Ion Ratio Lower Upper

113 100

111 102.4 82.0 123.0





192 18.6 15.3 22.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#36

1,1-Dichloropropene

Concen: 18.799 ug/l

RT: 5.702 min Scan# 757

Delta R.T. -0.000 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51




Tgt Ion: 75 Resp: 35819


Ion Ratio Lower Upper

75 100

110 36.2 17.2 51.5

77 31.6 24.2 36.4





#37

Ethyl Acetate

Concen: 17.627 ug/l

RT: 4.721 min Scan# 596

Delta R.T. -0.000 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBSD01

Tgt Ion: 43 Resp: 3100

Ion Ratio Lower Upper

43 100

61 13.3 9.7 14.5





70 9.4 8.1 12.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#38

Carbon Tetrachloride

Concen: 19.117 ug/l

RT: 5.684 min Scan# 754

Delta R.T. -0.006 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51




Tgt Ion:117 Resp: 39531


Ion Ratio Lower Upper

117 100

119 94.8 75.2 112.8

121 29.5 24.1 36.1





#39

Methylcyclohexane

Concen: 18.962 ug/l

RT: 7.385 min Scan# 1033

Delta R.T. -0.000 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBSD01





Tgt Ion:	83	Resp:	47544
Ion Ratio	Lower	Upper	
83	100		
55	79.5	63.0	94.4
98	49.7	38.8	58.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#40

Benzene

Concen: 20.015 ug/l




RT: 6.044 min Scan# 813


Delta R.T. -0.000 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51


Tgt Ion:	78	Resp:	112866
Ion Ratio	Lower	Upper	
78	100		
77	24.3	18.8	28.2





#41
Methacrylonitrile
Concen: 18.985 ug/l
RT: 4.928 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01



Tgt Ion: 41 Resp: 1780
Ion Ratio Lower Upper
41 100
39 54.0 45.9 68.9
67 72.8 61.2 91.8
52 30.2 25.5 38.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#42
1,2-Dichloroethane
Concen: 20.100 ug/l
RT: 6.092 min Scan# 821
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Tgt Ion: 62 Resp: 39829
Ion Ratio Lower Upper
62 100
98 9.3 0.0 17.8





#43

Isopropyl Acetate

Concen: 18.463 ug/l

RT: 6.342 min Scan# 862

Delta R.T. -0.000 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBSD01

Tgt Ion: 43 Resp: 5081

Ion Ratio Lower Upper

43 100

61 18.4 15.3 22.9





87 12.4 10.0 15.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#44

Trichloroethene

Concen: 19.321 ug/l

RT: 7.129 min Scan# 991

Delta R.T. -0.000 min

Lab File: VX046763.D




Acq: 19 Jun 2025 11:51


Tgt Ion:130 Resp: 27765

Ion Ratio Lower Upper

130 100

95 112.5 0.0 202.6





#45

1,2-Dichloropropane

Concen: 20.276 ug/l

RT: 7.427 min Scan# 1040

Delta R.T. -0.006 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBSD01

Tgt Ion: 63 Resp: 28336

Ion Ratio Lower Upper

63 100





65 32.8 22.9 34.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#46

Dibromomethane

Concen: 19.891 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. -0.006 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51




Tgt Ion: 93 Resp: 19978


Ion Ratio Lower Upper

93 100

95 82.3 66.2 99.4

174 97.0 77.0 115.6





#47

Bromodichloromethane

Concen: 20.112 ug/l

RT: 7.824 min Scan# 1105

Delta R.T. -0.000 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBSD01





Tgt Ion:	83	Resp:	4122
Ion	Ratio	Lower	Upper
83	100		
85	65.4	51.4	77.2
127	8.7	6.9	10.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#48

Methyl methacrylate

Concen: 18.904 ug/l




RT: 7.696 min Scan# 1084


Delta R.T. -0.000 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51

Tgt Ion:	41	Resp:	26092
Ion	Ratio	Lower	Upper
41	100		
69	79.1	63.4	95.2
39	56.2	45.4	68.0





#49

1,4-Dioxane

Concen: 321.885 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. 0.000 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBSD01

Tgt Ion: 88 Resp: 6068

Ion Ratio Lower Upper

88 100

43 42.8 31.2 46.8





58 69.2 54.7 82.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#50

Toluene-d8

Concen: 52.661 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. -0.000 min

Lab File: VX046763.D




Acq: 19 Jun 2025 11:51


Tgt Ion: 98 Resp: 249730

Ion Ratio Lower Upper

98 100


100 66.2 53.0 79.4





#51
4-Methyl-2-Pentanone
Concen: 91.811 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01



Tgt Ion: 43 Resp: 150700
Ion Ratio Lower Upper
43 100
58 37.2 30.8 46.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#52
Toluene
Concen: 20.310 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Tgt Ion: 92 Resp: 70998
Ion Ratio Lower Upper
92 100
91 170.3 134.8 202.2





#53

t-1,3-Dichloropropene

Concen: 19.536 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBSD01

Tgt Ion: 75 Resp: 3715

Ion Ratio Lower Upper

75 100





77 30.8 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#54

cis-1,3-Dichloropropene

Concen: 19.939 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51




Tgt Ion: 75 Resp: 43044


Ion Ratio Lower Upper

75 100

77 31.4 25.3 37.9

39 48.8 39.2 58.8





#55

1,1,2-Trichloroethane

Concen: 20.409 ug/l

RT: 9.153 min Scan# 1323

Delta R.T. -0.000 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBSD01

Tgt Ion: 97 Resp: 2658

Ion Ratio Lower Upper

97 100

83 80.6 67.2 100.8

85 55.0 44.1 66.1





99 64.8 50.3 75.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#56

Ethyl methacrylate

Concen: 18.636 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. -0.000 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51




Tgt Ion: 69 Resp: 35838


Ion Ratio Lower Upper

69 100

41 71.4 55.4 83.2

39 38.8 29.7 44.5





#57

1,3-Dichloropropane

Concen: 20.277 ug/l

RT: 9.305 min Scan# 1134

Delta R.T. -0.000 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBSD01

Tgt Ion: 76 Resp: 45490

Ion Ratio Lower Upper

76 100





78 31.6 25.7 38.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#58

2-Chloroethyl Vinyl ether

Concen: 100.344 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. -0.006 min

Lab File: VX046763.D




Acq: 19 Jun 2025 11:51


Tgt Ion: 63 Resp: 100239

Ion Ratio Lower Upper

63 100

106 28.8 22.6 34.0





#59

2-Hexanone

Concen: 88.846 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. -0.000 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBSD01

Tgt Ion: 43 Resp: 100820

Ion Ratio Lower Upper

43 100





58 53.3 26.8 80.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#60

Dibromochloromethane

Concen: 19.947 ug/l

RT: 9.519 min Scan# 1383

Delta R.T. -0.000 min

Lab File: VX046763.D




Acq: 19 Jun 2025 11:51


Tgt Ion:129 Resp: 30624

Ion Ratio Lower Upper

129 100


127 77.9 38.4 115.2





#61
1,2-Dibromoethane
Concen: 20.084 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01



Tgt Ion:107 Resp: 26614
Ion Ratio Lower Upper
107 100
109 93.0 75.8 113.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#62
4-Bromofluorobenzene
Concen: 54.915 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51


Tgt Ion: 95 Resp: 97074
Ion Ratio Lower Upper
95 100
174 74.5 0.0 150.4
176 71.8 0.0 145.0





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01



Tgt Ion:117 Resp: 18380
Ion Ratio Lower Upper
117 100
82 54.4 44.6 66.8
119 31.9 25.8 38.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#64
Tetrachloroethene
Concen: 18.778 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Tgt Ion:164 Resp: 23913
Ion Ratio Lower Upper
164 100
166 128.8 104.6 157.0
129 99.3 76.3 114.5
131 92.5 74.1 111.1









#65
Chlorobenzene
Concen: 19.421 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01

Tgt Ion:112 Resp: 7880
Ion Ratio Lower Upper
112 100
114 32.1 25.5 38.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 19.449 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Tgt Ion:131 Resp: 26522
Ion Ratio Lower Upper
131 100
133 94.3 47.3 141.8
119 64.3 32.1 96.3









#67
Ethyl Benzene
Concen: 19.141 ug/l
RT: 10.195 min Scan# 1493
Delta R.T. 0.006 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01

Tgt Ion: 91 Resp: 135702
Ion Ratio Lower Upper
91 100
106 29.4 24.2 36.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#68
m/p-Xylenes
Concen: 38.810 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Tgt Ion:106 Resp: 102519
Ion Ratio Lower Upper
106 100
91 208.3 164.8 247.2









#69
o-Xylene
Concen: 19.820 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01

Tgt Ion:106 Resp: 49068
Ion Ratio Lower Upper
106 100
91 221.1 111.3 333.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#70
Styrene
Concen: 19.696 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51


Tgt Ion:104 Resp: 85333
Ion Ratio Lower Upper
104 100
78 50.6 41.4 62.2
103 55.0 43.1 64.7





#71
Bromoform
Concen: 18.418 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01



Tgt Ion:173 Resp: 1912
Ion Ratio Lower Upper
173 100
175 49.8 24.1 72.4
254 0.0 0.1 0.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Tgt Ion:152 Resp: 94851
Ion Ratio Lower Upper
152 100
115 70.0 43.2 129.6
150 162.0 0.0 346.8









#73
Isopropylbenzene
Concen: 18.592 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01

Tgt Ion:105 Resp: 12988
Ion Ratio Lower Upper
105 100
120 26.1 13.0 39.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#74
N-ethyl acetate
Concen: 17.214 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Tgt Ion: 43 Resp: 45562
Ion Ratio Lower Upper
43 100
70 41.4 33.4 50.0
55 24.6 19.1 28.7
61 23.8 19.0 28.4





#75

1,1,2,2-Tetrachloroethane

Concen: 18.425 ug/l

RT: 11.207 min Scan# 1660

Delta R.T. -0.000 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBSD01

Tgt Ion: 83 Resp: 35939

Ion Ratio Lower Upper

83 100

131 10.0 4.9 14.6





85 63.9 32.0 96.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#76

1,2,3-Trichloropropane

Concen: 16.867 ug/l m

RT: 11.238 min Scan# 1665

Delta R.T. -0.000 min

Lab File: VX046763.D




Acq: 19 Jun 2025 11:51


Tgt Ion: 75 Resp: 29301

Ion Ratio Lower Upper

75 100


77 41.0 18.5 55.5





#77
Bromobenzene
Concen: 18.976 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01




Tgt Ion:156 Resp: 31899
Ion Ratio Lower Upper
156 100
77 169.0 82.8 248.6
158 96.8 47.6 142.9

Manual Integrations
APPROVED



Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025




#78
n-propylbenzene
Concen: 18.803 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51



Tgt Ion: 91 Resp: 156018
Ion Ratio Lower Upper
91 100
120 22.6 11.4 34.1









#79
2-Chlorotoluene
Concen: 18.633 ug/l
RT: 11.360 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01

Tgt Ion: 91 Resp: 92310
Ion Ratio Lower Upper
91 100
126 33.6 16.6 49.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#80
1,3,5-Trimethylbenzene
Concen: 18.925 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51


Tgt Ion:105 Resp: 108639
Ion Ratio Lower Upper
105 100
120 48.5 24.1 72.2





#81
trans-1,4-Dichloro-2-butene
Concen: 15.556 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01



Tgt Ion: 75 Resp: 956
Ion Ratio Lower Upper
75 100
53 116.3 86.8 130.2
89 46.3 36.9 55.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#82
4-Chlorotoluene
Concen: 18.795 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51


Tgt Ion: 91 Resp: 108666
Ion Ratio Lower Upper
91 100
126 29.3 14.4 43.2





#83
tert-Butylbenzene
Concen: 18.337 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01



Tgt Ion:119 Resp: 108810
Ion Ratio Lower Upper
119 100
91 59.3 29.5 88.6
134 23.4 11.7 35.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#84
1,2,4-Trimethylbenzene
Concen: 18.872 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51


Tgt Ion:105 Resp: 108849
Ion Ratio Lower Upper
105 100
120 44.5 22.1 66.3





#85
 sec-Butylbenzene
 Concen: 18.839 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. -0.000 min
 Lab File: VX046763.D
 Acq: 19 Jun 2025 11:51




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0619WBSD01




Tgt Ion:105 Resp: 14128
 Ion Ratio Lower Upper
 105 100
 134 19.6 9.8 29.4

Manual Integrations
 APPROVED



Reviewed By :John Carlone 06/20/2025
 Supervised By :Mahesh Dadoda 06/21/2025




#86
 p-Isopropyltoluene
 Concen: 18.732 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. -0.000 min
 Lab File: VX046763.D
 Acq: 19 Jun 2025 11:51




Tgt Ion:119 Resp: 118528
 Ion Ratio Lower Upper
 119 100
 134 26.0 12.9 38.7
 91 25.2 12.3 36.9





#87
1,3-Dichlorobenzene
Concen: 18.553 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01



Tgt Ion:146 Resp: 60812
Ion Ratio Lower Upper
146 100
111 40.6 20.6 61.8
148 63.4 31.6 94.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#88
1,4-Dichlorobenzene
Concen: 17.772 ug/l
RT: 12.036 min Scan# 1796
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Tgt Ion:146 Resp: 59834
Ion Ratio Lower Upper
146 100
111 42.6 20.3 61.0
148 65.7 32.1 96.5





#89

n-Butylbenzene

Concen: 18.526 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. -0.000 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBSD01

Tgt Ion: 91 Resp: 11130

Ion Ratio Lower Upper

91 100

92 53.7 27.1 81.3





134 25.4 12.9 38.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#90

Hexachloroethane

Concen: 17.220 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX046763.D




Acq: 19 Jun 2025 11:51


Tgt Ion:117 Resp: 19081

Ion Ratio Lower Upper

117 100


201 72.0 34.4 103.3





#91
1,2-Dichlorobenzene
Concen: 18.997 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.006 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01



Tgt Ion:146 Resp: 5818
Ion Ratio Lower Upper
146 100
111 41.8 21.7 65.1
148 63.9 32.0 96.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 16.116 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Tgt Ion: 75 Resp: 6206
Ion Ratio Lower Upper
75 100
155 85.7 41.9 125.7
157 106.9 51.4 154.3





#93

1,2,4-Trichlorobenzene

Concen: 18.309 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. -0.000 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBSD01

Tgt Ion:180 Resp: 3987

Ion Ratio Lower Upper

180 100

182 94.6 46.7 140.1





145 31.9 15.8 47.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/20/2025

Supervised By :Mahesh Dadoda 06/21/2025



#94

Hexachlorobutadiene

Concen: 18.715 ug/l

RT: 13.719 min Scan# 2072

Delta R.T. -0.000 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51




Tgt Ion:225 Resp: 17144


Ion Ratio Lower Upper

225 100

223 63.4 32.5 97.5


227 60.2 32.2 96.6





#95
Naphthalene
Concen: 17.340 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01



Tgt Ion:128 Resp: 107578
Ion Ratio Lower Upper
128 100
127 12.9 10.1 15.1
129 10.7 8.7 13.1




Manual Integrations
APPROVED

Reviewed By :John Carlone 06/20/2025
Supervised By :Mahesh Dadoda 06/21/2025



#96
1,2,3-Trichlorobenzene
Concen: 18.189 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Tgt Ion:180 Resp: 37893
Ion Ratio Lower Upper
180 100
182 96.0 47.8 143.4
145 33.9 17.3 51.9



Manual Integration Report

Sequence:	vx061725	Instrument	MSVOA_x

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDICC005	VX046718.D	1,2,3-Trichloropropane	Sam	6/18/2025 5:26:54 PM	MMDadoda	6/18/2025 6:21:14 PM	Peak Integrated by Software
VSTDICC020	VX046719.D	1,2,3-Trichloropropane	Sam	6/18/2025 5:27:04 PM	MMDadoda	6/18/2025 6:21:20 PM	Peak Integrated by Software
VSTDICCC050	VX046720.D	1,2,3-Trichloropropane	Sam	6/18/2025 5:27:12 PM	MMDadoda	6/18/2025 6:21:32 PM	Peak Integrated by Software
VSTDICC100	VX046721.D	1,2,3-Trichloropropane	Sam	6/18/2025 5:27:20 PM	MMDadoda	6/18/2025 6:21:40 PM	Peak Integrated by Software
VSTDICC150	VX046722.D	1,2,3-Trichloropropane	Sam	6/18/2025 5:27:27 PM	MMDadoda	6/18/2025 6:21:48 PM	Peak Integrated by Software
VSTDICC001	VX046725.D	1,2,3-Trichloropropane	Sam	6/18/2025 5:27:40 PM	MMDadoda	6/18/2025 6:22:02 PM	Peak Integrated by Software
VSTDICC001	VX046725.D	1,4-Dichlorobenzene	Sam	6/18/2025 5:27:40 PM	MMDadoda	6/18/2025 6:22:02 PM	Peak Integrated by Software
VSTDICC001	VX046725.D	2,2-Dichloropropane	Sam	6/18/2025 5:27:40 PM	MMDadoda	6/18/2025 6:22:02 PM	Peak Integrated by Software
VSTDICC001	VX046725.D	Chloroform	Sam	6/18/2025 5:27:40 PM	MMDadoda	6/18/2025 6:22:02 PM	Peak Integrated by Software
VSTDICC001	VX046725.D	Ethyl Acetate	Sam	6/18/2025 5:27:40 PM	MMDadoda	6/18/2025 6:22:02 PM	Peak Integrated by Software
VSTDICC001	VX046725.D	Methacrylonitrile	Sam	6/18/2025 5:27:40 PM	MMDadoda	6/18/2025 6:22:02 PM	Peak Integrated by Software
VSTDICV050	VX046726.D	1,2,3-Trichloropropane	Sam	6/18/2025 5:27:47 PM	MMDadoda	6/18/2025 6:22:09 PM	Peak Integrated by Software



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:

vx061725

Instrument

MSVOA_x

Sample ID

File ID

Parameter

Review By

Review On

Supervised
By

Supervised On

Reason

Manual Integration Report

Sequence:

VX061925

Instrument

MSVOA_x

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDCCC050	VX046758.D	1,2,3-Trichloropropane	JOHN	6/20/2025 8:42:24 AM	MMdadoda	6/21/2025 12:20:44 AM	Peak Integrated by Software
VSTDCCC050	VX046758.D	2-Butanone	JOHN	6/20/2025 8:42:24 AM	MMdadoda	6/21/2025 12:20:44 AM	Peak Integrated by Software
VX0619WBS01	VX046762.D	1,2,3-Trichloropropane	JOHN	6/20/2025 8:42:30 AM	MMdadoda	6/21/2025 12:20:48 AM	Peak Integrated by Software
VX0619WBS01	VX046762.D	2-Butanone	JOHN	6/20/2025 8:42:30 AM	MMdadoda	6/21/2025 12:20:48 AM	Peak Integrated by Software
VX0619WBSD0 1	VX046763.D	1,2,3-Trichloropropane	JOHN	6/20/2025 8:42:34 AM	MMdadoda	6/21/2025 12:20:52 AM	Peak Integrated by Software
VSTDCCC050	VX046775.D	1,2,3-Trichloropropane	JOHN	6/20/2025 8:42:59 AM	MMdadoda	6/21/2025 12:21:03 AM	Peak Integrated by Software

Instrument ID: **MSVOA_X**

Daily Analysis Runlog For Sequence/QC Batch ID # VX061725

Review By	Semsettin Yesilyurt	Review On	6/18/2025 5:28:16 PM
Supervise By	Mahesh Dadoda	Supervise On	6/18/2025 6:22:48 PM
SubDirectory	VX061725	HP Acquire Method	HP Processing Method 82X061725W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP134389		
Initial Calibration Stds	VP134390,VP134391,VP134392,VP134393,VP134394,VP134395		
CCC			
Internal Standard/PEM			
ICV/I.BLK	VP134396		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VX046715.D	17 Jun 2025 08:46	JC/MD	Ok
2	VSTDIC001	VX046716.D	17 Jun 2025 10:09	JC/MD	Not Ok
3	VSTDIC001	VX046717.D	17 Jun 2025 10:58	JC/MD	ReRun
4	VSTDIC005	VX046718.D	17 Jun 2025 11:19	JC/MD	Ok,M
5	VSTDIC020	VX046719.D	17 Jun 2025 13:59	JC/MD	Ok,M
6	VSTDIC050	VX046720.D	17 Jun 2025 14:20	JC/MD	Ok,M
7	VSTDIC100	VX046721.D	17 Jun 2025 14:41	JC/MD	Ok,M
8	VSTDIC150	VX046722.D	17 Jun 2025 15:02	JC/MD	Ok,M
9	IBLK	VX046723.D	17 Jun 2025 15:23	JC/MD	Ok
10	VSTDIC001	VX046724.D	17 Jun 2025 16:20	JC/MD	Not Ok
11	VSTDIC001	VX046725.D	17 Jun 2025 17:18	JC/MD	Ok,M
12	VSTDICV050	VX046726.D	17 Jun 2025 18:00	JC/MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_X**

Daily Analysis Runlog For Sequence/QC Batch ID # VX061925

Review By	John Carlone	Review On	6/20/2025 8:46:07 AM
Supervise By	Maresh Dadoda	Supervise On	6/21/2025 12:21:16 AM
SubDirectory	VX061925	HP Acquire Method	HP Processing Method 82X061725W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134428		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134429,VP134430		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VX046757.D	19 Jun 2025 08:42	JC/MD	Ok
2	VSTDCCC050	VX046758.D	19 Jun 2025 09:43	JC/MD	Ok,M
3	VX0619MBL01	VX046759.D	19 Jun 2025 10:12	JC/MD	Ok
4	VX0619WBL01	VX046760.D	19 Jun 2025 10:33	JC/MD	Ok
5	Q2345-01	VX046761.D	19 Jun 2025 11:03	JC/MD	Ok
6	VX0619WBS01	VX046762.D	19 Jun 2025 11:24	JC/MD	Ok,M
7	VX0619WBSD01	VX046763.D	19 Jun 2025 11:51	JC/MD	Ok,M
8	Q2361-01	VX046764.D	19 Jun 2025 12:12	JC/MD	Ok
9	Q2334-01	VX046765.D	19 Jun 2025 12:33	JC/MD	Ok
10	Q2334-02	VX046766.D	19 Jun 2025 12:54	JC/MD	Ok
11	Q2334-03	VX046767.D	19 Jun 2025 13:15	JC/MD	Ok
12	Q2363-02	VX046768.D	19 Jun 2025 13:36	JC/MD	Dilution
13	IBLK	VX046769.D	19 Jun 2025 13:57	JC/MD	Ok
14	IBLK	VX046770.D	19 Jun 2025 14:18	JC/MD	Ok
15	Q2351-01	VX046771.D	19 Jun 2025 15:45	JC/MD	Ok
16	Q2372-02	VX046772.D	19 Jun 2025 16:06	JC/MD	Ok
17	Q2372-01	VX046773.D	19 Jun 2025 16:27	JC/MD	Ok
18	IBLK	VX046774.D	19 Jun 2025 16:49	JC/MD	Ok
19	VSTDCCC050	VX046775.D	19 Jun 2025 17:10	JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX061725

Review By	Semsettin Yesilyurt	Review On	6/18/2025 5:28:16 PM
Supervise By	Maresh Dadoda	Supervise On	6/18/2025 6:22:48 PM
SubDirectory	VX061725	HP Acquire Method	HP Processing Method 82X061725W.M

STD. NAME	STD REF.#
Tune/Reschk	VP134389
Initial Calibration Stds	VP134390,VP134391,VP134392,VP134393,VP134394,VP134395
CCC	
Internal Standard/PEM	
ICV/I.BLK	VP134396
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX046715.D	17 Jun 2025 08:46		JC/MD	Ok
2	VSTDIC001	VSTDIC001	VX046716.D	17 Jun 2025 10:09	RRF check	JC/MD	Not Ok
3	VSTDIC001	VSTDIC001	VX046717.D	17 Jun 2025 10:58	low response	JC/MD	ReRun
4	VSTDIC005	VSTDIC005	VX046718.D	17 Jun 2025 11:19		JC/MD	Ok,M
5	VSTDIC020	VSTDIC020	VX046719.D	17 Jun 2025 13:59	LR- 10,49	JC/MD	Ok,M
6	VSTDIC050	VSTDIC050	VX046720.D	17 Jun 2025 14:20		JC/MD	Ok,M
7	VSTDIC100	VSTDIC100	VX046721.D	17 Jun 2025 14:41		JC/MD	Ok,M
8	VSTDIC150	VSTDIC150	VX046722.D	17 Jun 2025 15:02		JC/MD	Ok,M
9	IBLK	IBLK	VX046723.D	17 Jun 2025 15:23		JC/MD	Ok
10	VSTDIC001	VSTDIC001	VX046724.D	17 Jun 2025 16:20	spike error	JC/MD	Not Ok
11	VSTDIC001	VSTDIC001	VX046725.D	17 Jun 2025 17:18		JC/MD	Ok,M
12	VSTDICV050	ICVVX061725	VX046726.D	17 Jun 2025 18:00		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX061925

Review By	John Carlone	Review On	6/20/2025 8:46:07 AM
Supervise By	Maresh Dadoda	Supervise On	6/21/2025 12:21:16 AM
SubDirectory	VX061925	HP Acquire Method	HP Processing Method 82X061725W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134428
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134429,VP134430

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX046757.D	19 Jun 2025 08:42		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX046758.D	19 Jun 2025 09:43	pH#Lot#V12668	JC/MD	Ok,M
3	VX0619MBL01	VX0619MBL01	VX046759.D	19 Jun 2025 10:12		JC/MD	Ok
4	VX0619WBL01	VX0619WBL01	VX046760.D	19 Jun 2025 10:33		JC/MD	Ok
5	Q2345-01	EB02-20250616	VX046761.D	19 Jun 2025 11:03	vial B pH<2 EB	JC/MD	Ok
6	VX0619WBS01	VX0619WBS01	VX046762.D	19 Jun 2025 11:24		JC/MD	Ok,M
7	VX0619WBSD01	VX0619WBSD01	VX046763.D	19 Jun 2025 11:51		JC/MD	Ok,M
8	Q2361-01	TT205S1-20250617	VX046764.D	19 Jun 2025 12:12	vial A pH<2	JC/MD	Ok
9	Q2334-01	MW3	VX046765.D	19 Jun 2025 12:33	vial A pH<2	JC/MD	Ok
10	Q2334-02	MW4	VX046766.D	19 Jun 2025 12:54	vial A pH<2	JC/MD	Ok
11	Q2334-03	GBTW1	VX046767.D	19 Jun 2025 13:15	vial A pH<2	JC/MD	Ok
12	Q2363-02	GCAP1W	VX046768.D	19 Jun 2025 13:36	vial A pH<2 Need 20X	JC/MD	Dilution
13	IBLK	IBLK	VX046769.D	19 Jun 2025 13:57		JC/MD	Ok
14	IBLK	IBLK	VX046770.D	19 Jun 2025 14:18		JC/MD	Ok
15	Q2351-01	FAIRLAWN-SUMP	VX046771.D	19 Jun 2025 15:45	vial A pH<2 oily sample	JC/MD	Ok
16	Q2372-02	GAV2W	VX046772.D	19 Jun 2025 16:06	vial A pH<2	JC/MD	Ok
17	Q2372-01	GAV1W	VX046773.D	19 Jun 2025 16:27	vial A pH<2	JC/MD	Ok
18	IBLK	IBLK	VX046774.D	19 Jun 2025 16:49		JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX061925

Review By	John Carlone	Review On	6/20/2025 8:46:07 AM				
Supervise By	Mahesh Dadoda	Supervise On	6/21/2025 12:21:16 AM				
SubDirectory	VX061925	HP Acquire Method	HP Processing Method		82X061725W.M		
STD. NAME		STD REF.#					
Tune/Reschk Initial Calibration Stds		VP134428					
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard		VP134429,VP134430					
19	VSTDCCC050	VSTDCCC050EC	VX046775.D	19 Jun 2025 17:10		JC/MD	Ok,M

M : Manual Integration

Prep Standard - Chemical Standard Summary

Order ID : Q2361

Test : VOCMS Group1

Prepbatch ID :

Sequence ID/Qc Batch ID: VX061925,

Standard ID :

VP133174,VP133887,VP133935,VP133953,VP133991,VP134142,VP134149,VP134428,VP134429,VP134430,

Chemical ID :

V13391,V13706,V14290,V14432,V14435,V14503,V14504,V14525,V14526,V14613,V14620,V14626,V14636,V14637,V14638,V14639,V14668,V14671,V14673,V14675,V14711,V14717,V14718,V14721,V14749,V14750,V14811,V14812,V14843,V14921,V14929,V14944,V14945,V14946,V14947,W3112,

VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
617	8260 Surrogate, 400PPM	VP133174	02/27/2025	08/27/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
								03/04/2025

FROM 0.40000ml of V13706 + 24.60000ml of V14613 = Final Quantity: 25.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
257	8260 Calibration Working STD Mix-First source, 160PPM	VP133887	05/12/2025	06/23/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
								05/14/2025

FROM 0.40000ml of V14843 + 1.00000ml of V14432 + 1.00000ml of V14435 + 1.00000ml of V14503 + 1.00000ml of V14504 + 1.00000ml of V14525 + 1.00000ml of V14526 + 1.00000ml of V14711 + 1.00000ml of V14717 + 1.00000ml of V14718 + 1.00000ml of V14721 + 1.00000ml of V14749 + 1.00000ml of V14750 + 1.00000ml of V14811 + 1.00000ml of V14812 + 10.60000ml of V14921 = Final Quantity: 25.000 ml

VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
247	8260 Internal Standard, 250PPM	VP133935	05/16/2025	11/12/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
								05/21/2025

FROM 0.25000ml of V14290 + 24.75000ml of V14921 = Final Quantity: 25.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
218	BFB, 25PPM	VP133953	05/19/2025	11/09/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
								05/21/2025

FROM 0.25000ml of V13391 + 24.75000ml of V14626 = Final Quantity: 25.000 ml

[illegible]

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
719	8260 Working STD (BCM)-First source, 400PPM	VP134142	06/06/2025	12/06/2025	Semsettin Yesilyurt	None	None	Mahesh Dadoda 06/10/2025
<u>FROM</u>	1.00000ml of V14668 + 1.00000ml of V14671 + 1.00000ml of V14673 + 1.00000ml of V14675 + 16.00000ml of V14929 = Final Quantity: 20.000 ml							

VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1810	8260 Working Std(2-CVE)-800ppm	VP134149	06/06/2025	12/06/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
								06/10/2025

FROM 1.00000ml of V14636 + 1.00000ml of V14637 + 1.00000ml of V14638 + 1.00000ml of V14639 + 46.00000ml of V14929 = Final Quantity: 50.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
589	BFB TUNE CHECK	VP134428	06/19/2025	06/20/2025	John Carlone	None	None	Maresh Dadoda
								06/23/2025

FROM 39.98400ml of W3112 + 0.01600ml of VP133953 = Final Quantity: 40.000 ml

[illegible]

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
620	50 PPB CCC, 8260-Water	VP134430	06/19/2025	06/19/2025	John Carlone	None	None	Mahesh Dadoda 06/23/2025
<u>FROM</u>	39.94450ml of W3112 + 0.00500ml of VP133174 + 0.00500ml of VP134142 + 0.00800ml of VP133935 + 0.01250ml of VP133887 + 0.01250ml of VP133991 + 0.01250ml of VP134149 = Final Quantity: 40.000 ml							

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30067 / BFB tuneing solution	A0191805	11/22/2025	11/22/2024 / SAM	01/13/2023 / SAM	V13391

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555582 / Custom Mixture, 8260 A/B Surrogate Mix [CS 5179-2]	A0196865	02/27/2026	02/27/2025 / SAM	04/12/2023 / SAM	V13706

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555581 / Custom Standard, 8260 Internal Std [CS 5179-1]	A0210184	12/12/2025	12/12/2024 / SAM	04/15/2024 / SAM	V14290

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30489 / VOA Mix, 8260B Acetates Mix, P&TM, 1mL	A0209618	09/30/2025	05/12/2025 / SAM	08/15/2024 / SAM	V14432

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30489 / VOA Mix, 8260B Acetates Mix, P&TM, 1mL	A0209618	09/20/2025	03/20/2025 / SAM	08/15/2024 / SAM	V14435

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95317 / Universal VOA Mega Mix (Min order = 5)	021624	11/12/2025	05/12/2025 / SAM	09/17/2024 / SAM	V14503

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95317 / Universal VOA Mega Mix (Min order = 5)	021624	11/12/2025	05/12/2025 / SAM	09/17/2024 / SAM	V14504

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95319 / Revised Additions Mix (Min = 5)	091724	11/12/2025	05/12/2025 / SAM	09/18/2024 / SAM	V14525

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95319 / Revised Additions Mix (Min = 5)	091724	11/12/2025	05/12/2025 / SAM	09/18/2024 / SAM	V14526

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	22L0562016	08/27/2025	02/27/2025 / SAM	11/26/2024 / SAM	V14613

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	22L0562016	10/25/2025	05/09/2025 / SAM	11/26/2024 / SAM	V14620

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	23I0762004	11/09/2025	05/09/2025 / SAM	11/26/2024 / SAM	V14626

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	12/06/2025	06/06/2025 / SAM	12/06/2024 / SAM	V14636

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	12/06/2025	06/06/2025 / SAM	12/06/2024 / SAM	V14637

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	12/06/2025	06/06/2025 / SAM	12/06/2024 / SAM	V14638

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	12/06/2025	06/06/2025 / SAM	12/06/2024 / SAM	V14639

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0214960	12/06/2025	06/06/2025 / SAM	12/09/2024 / SAM	V14668

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0214960	12/06/2025	06/06/2025 / SAM	12/09/2024 / SAM	V14671

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0214960	12/06/2025	06/06/2025 / SAM	12/09/2024 / SAM	V14673

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0214960	12/06/2025	06/06/2025 / SAM	12/09/2024 / SAM	V14675

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A02110618	11/12/2025	05/12/2025 / SAM	12/17/2024 / SAM	V14711

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A02110618	11/12/2025	05/12/2025 / SAM	12/17/2024 / SAM	V14717

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A02110618	11/12/2025	05/12/2025 / SAM	12/17/2024 / SAM	V14718

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A02110618	11/12/2025	05/12/2025 / SAM	12/17/2024 / SAM	V14721

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30042 / VOA Mix,500 series method 502.2 Calibration Std #1 gases, 2000uq/ml, PTM, 1ml	A0216826	11/13/2025	05/12/2025 / SAM	12/17/2024 / SAM	V14749

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30042 / VOA Mix,500 series method 502.2 Calibration Std #1 gases, 2000uq/ml, PTM, 1ml	A0216826	11/12/2025	05/12/2025 / SAM	12/17/2024 / SAM	V14750

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555408 / Custom Standard, Vinyl Acetate Standard w/ Grav [CS 5066-6] TWO SEPARATE	A0220471	11/12/2025	05/12/2025 / SAM	01/08/2025 / SAM	V14811

LOTS

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555408 / Custom Standard, Vinyl Acetate Standard w/ Grav [CS 5066-6] TWO SEPARATE	A0220471	06/30/2026	05/12/2025 / SAM	01/08/2025 / SAM	V14812

LOTS

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30470 / VOA Stock Solution, tert-butanol std, 1mL, P&TM	A0217535	11/12/2025	05/12/2025 / SAM	01/21/2025 / SAM	V14843

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	24G0262002	11/12/2025	05/12/2025 / SAM	05/09/2025 / SAM	V14921

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	24G0262002	12/06/2025	06/06/2025 / SAM	05/09/2025 / SAM	V14929

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	051925	06/19/2025	05/22/2025 / SAM	05/21/2025 / SAM	V14944

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	051925	06/19/2025	05/22/2025 / SAM	05/21/2025 / SAM	V14945

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	051925	06/19/2025	05/22/2025 / SAM	05/21/2025 / SAM	V14946

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	051925	06/19/2025	05/22/2025 / SAM	05/21/2025 / SAM	V14947

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / lwona	07/03/2024 / lwona	W3112

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis



Material No.: 9077-02
Batch No.: 23I0762004
Manufactured Date: 2023-08-11
Expiration Date: 2026-08-10
Revision No.: 0

Certificate of Analysis


Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.5 ppm
Titration Acid (μeq/g)	≤ 0.3	0.2
Titration Base (μeq/g)	≤ 0.10	0.01
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis – Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Ken Koehnlein
Sr. Manager, Quality Assurance

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis



Material No.: 9077-02
Batch No.: 22L0562016
Manufactured Date: 2022-10-26
Expiration Date: 2025-10-25
Revision No.: 0

Certificate of Analysis


Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.2 ppm
Titration Acid (μeq/g)	≤ 0.3	0.2
Titration Base (μeq/g)	≤ 0.10	0.03
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis – Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Jamie Ethier
Vice President Global Quality

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis




Material No.: 9077-02
Batch No.: 22L0562016
Manufactured Date: 2022-10-26
Expiration Date: 2025-10-25
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.2 ppm
Titration Acid (μeq/g)	≤ 0.3	0.2
Titration Base (μeq/g)	≤ 0.10	0.03
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis – Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC


Jamie Ethier
Vice President Global Quality



Ree 03117/24

CERTIFIED WEIGHT REPORT

Part Number: 95317

Lot Number: 021624

Description: Universal VOA Megamix

69 components

Expiration Date: 021627

Recommended Storage: Freezer (0 °C)

Nominal Concentration (µg/mL): 2000

NIST Test ID#: BUTB

Solvent(s): Lot#
Methanol EG359-USQ12Weight(s) shown below were combined and diluted to (mL): 100.0 0.021 Balance Uncertainty
Flask Uncertainty


Formulated By:	Prashant Chauhan	021624
Reviewed By:	Pedro L. Renteria	021624
DATE		

Compound	(KMe) Part Number	Lot Number	Dir. Factor	Initial Vol. (mL)	Initial Conc. (µg/mL)	Nominal Conc. (µg/mL)	Purity (%)	Purity Uncertainty	Uncertainty Pipette (mL)	Target Weight(g)	Actual Weight(g)	Actual Conc. (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
														CAS#	OSHA PEL (TWA)	LD50
1. Acetonitrile	(0324)	021644	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20020	2001.3	8.1	75-05-8	40 ppm (70mg/m3/8H)	or-rat 2400mg/kg
2. Allyl chloride (3-Chloropropene)	(0325)	102396	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	107-05-1	1 ppm (3mg/m3/8H)	or-rat 700mg/kg
3. Carbon disulfide	(0060)	MKCR8561	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20023	2001.6	8.1	75-15-0	4 ppm (12mg/m3) (skin)	or-rat 1200mg/kg
4. cis-1,4-Dichloro-2-butene	(1196)	14718EF	NA	NA	NA	2000	95	0.2	NA	0.21056	0.21069	2001.1	8.5	1478-11-6	N/A	N/A
5. trans-1,4-Dichloro-2-butene	(0486)	MKBP9041V	NA	NA	NA	2000	96.5	0.2	NA	0.20731	0.20746	2001.7	8.4	110-57-6	N/A	N/A
6. Diethyl ether	(0153)	IK18CAS000C	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	60-29-7	N/A	N/A
7. Ethyl methacrylate	(0381)	06128PX	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20230	2002.3	8.2	97-63-2	N/A	N/A
8. Iodomethane	(0489)	SHBF8718V	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20121	2001.5	8.2	74-88-4	5 ppm (28mg/m3/8H) (skin)	or-rat 14800mg/kg
9. 2-Methyl-1-propanol	(0445)	15241EB	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20120	2001.4	8.1	78-83-1	50 ppm (150mg/m3/8H)	or-rat 3460mg/kg
10. Methacrylonitrile	(0442)	00427ET	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	128-98-7	1 ppm (3mg/m3/8H) (skin)	or-rat 120mg/kg
11. Methyl acrylate	(0755)	SHBK0679	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	96-33-3	10 ppm (35mg/m3/8H) (skin)	or-rat 277mg/kg
12. Methyl methacrylate	(0404)	MKGW6137V	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20041	2001.6	8.1	80-62-6	100 ppm (410mg/m3/8H)	or-rat 7872mg/kg
13. Nitrobenzene	(0228)	01213TV	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20220	2001.3	8.2	98-95-3	1 ppm (5mg/m3/8H) (skin)	or-rat 780mg/kg
14. 2-Nitropropane	(0461)	HG02JX	NA	NA	NA	2000	97.3	0.2	NA	0.20560	0.20577	2001.6	8.3	78-46-9	10 ppm (35mg/m3/8H)	or-rat 720mg/kg
15. Pentachloroethane	(0460)	HGA01	NA	NA	NA	2000	98	0.2	NA	0.20413	0.20430	2001.6	8.3	78-01-7	N/A	N/A
16. 1,1,2-Trichlorotrifluoroethane	(0474)	18990	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20225	2001.8	8.2	78-13-1	1000 ppm (7600mg/m3/8H)	or-rat 43g/kg
17. Bromodichloromethane	35171	101623	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-27-4	NA	or-rat 915mg/kg
18. Dibromochloromethane	35171	101623	0.05	5.00	40002.1	2000	NA	NA	0.017	NA	NA	1999.6	23.0	124-48-1	N/A	or-rat 848mg/kg
19. cis-1,2-Dichloroethene	35171	101623	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.7	22.9	156-59-2	N/A	N/A
20. trans-1,2-Dichloroethene	35171	101623	0.05	5.00	40003.4	2000	NA	NA	0.017	NA	NA	1999.6	23.0	156-60-5	N/A	or-rat 1235mg/kg
21. Methylene chloride	35171	101623	0.05	5.00	40002.8	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-09-2	500 ppm	or-rat 820mg/kg
22. 1,1-Dichloroethane	32251	102023	0.10	10.00	20001.6	2000	NA	NA	0.042	NA	NA	1999.7	20.4	75-35-4	1 ppm (4mg/m3/8H)	or-rat 200mg/kg
23. Bromoform	95321	020724	0.10	10.00	20003.2	2000	NA	NA	0.042	NA	NA	1999.8	20.5	75-25-2	0.5 ppm (5mg/m3) (skin)	or-rat 930mg/kg
24. Carbon tetrachloride	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	56-23-5	2 ppm (12.6mg/m3/8H)	or-rat 2350mg/kg
25. Chloroform	95321	020724	0.10	10.00	20004.0	2000	NA	NA	0.042	NA	NA	1999.8	20.5	67-68-3	60 ppm (240mg/m3) (CL)	or-rat 908mg/kg
26. Dibromomethane	95321	020724	0.10	10.00	20002.9	2000	NA	NA	0.042	NA	NA	1999.8	20.5	74-95-3	N/A	or-rat 108mg/kg
27. 1,1-Dichloroethane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	594-20-7	N/A	N/A
28. 2,2-Dichloropropane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	594-20-7	N/A	N/A
29. Trichloroethene	95321	020724	0.10	10.00	20201.1	2000	NA	NA	0.042	NA	NA	2019.8	20.8	127-18-4	25 ppm (170mg/m3/8H) (final)	or-rat 2629mg/kg
30. 1,1,1-Trichloroethane	95321	020724	0.10	10.00	20003.0	2000	NA	NA	0.042	NA	NA	1999.8	20.5	71-55-6	350 ppm (1900mg/m3/8H)	or-rat 10300mg/kg
31. 1,2-Dibromo-3-chloropropane	35161	112322	0.05	5.00	40016.5	2000	NA	NA	0.017	NA	NA	2000.3	22.9	96-12-8	0.001 ppm	or-rat 170mg/kg
32. 1,2-Dibromoethane	35161	112322	0.05	5.00	40024.8	2000	NA	NA	0.017	NA	NA	2000.7	22.9	106-93-4	20 ppm (8H)	or-rat 108mg/kg
33. 1,2-Dichloroethane	35161	112322	0.05	5.00	40018.0	2000	NA	NA	0.017	NA	NA	2000.4	22.9	107-06-2	50 ppm (8H)	or-rat 670mg/kg
34. 1,2-Dichloropropane	35161	112322	0.05	5.00	40051.0	2000	NA	NA	0.017	NA	NA	2002.0	22.9	78-87-5	75 ppm (350mg/m3/8H)	or-rat 1847mg/kg
35. 1,3-Dichloropropane	35161	112322	0.05	5.00	40005.9	2000	NA	NA	0.017	NA	NA	1999.8	22.9	142-28-9	N/A	or-rat 3600mg/kg
36. 1,1-Dichloropropene	35161	112322	0.05	5.00	40012.1	2000	NA	NA	0.017	NA	NA	2000.1	28.7	563-58-6	N/A	N/A
37. cis-1,3-Dichloropropene	35161	112322	0.05	5.00	40010.0	2000	NA	NA	0.017	NA	NA	2000.0	23.0	10061-01-5	N/A	N/A
38. trans-1,3-Dichloropropene	35161	112322	0.05	5.00	40017.6	2000	NA	NA	0.017	NA	NA	2000.4	23.0	10061-02-6	N/A	N/A
39. Hexachloro-1,3-butadiene	35161	112322	0.05	5.00	40021.9	2000	NA	NA	0.017	NA	NA	2000.6	29.7	87-68-3	0.02 ppm (0.24mg/m3/8H)	or-rat 82mg/kg
40. 1,1,1,2-Tetrachloroethane	35161	112322	0.05	5.00	40011.9	2000	NA	NA	0.017	NA	NA	2000.1	22.9	830-20-6	N/A	or-rat 670mg/kg
41. 1,1,2,2-Tetrachloroethane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	79-34-5	5 ppm (35mg/m3/8H) (skin)	or-rat 800mg/kg
42. 1,1,2-Trichloroethane	35161	112322	0.05	5.00	40006.6	2000	NA	NA	0.017	NA	NA	1999.8	23.0	79-00-5	10 ppm (45mg/m3/8H) (skin)	or-rat 936mg/kg
43. Trichloroethene	35161	112322	0.05	5.00	40029.0	2000	NA	NA	0.017	NA	NA	2000.9	22.9	79-01-6	50 ppm (270mg/m3/8H)	or-mus 2402mg/kg
44. 1,2,3-Trichloropropane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	96-18-4	10 ppm (60mg/m3/8H)	or-rat 149.8mg/kg
45. Benzene	35162	050823	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.7	22.9	71-43-2	1 ppm	or-rat 4894mg/kg
46. Bromobenzene	35162	050823	0.05	5.00	40006.0	2000	NA	NA	0.017	NA	NA	1999.7	22.9	104-51-8	N/A	N/A
47. n-Butyl benzene	35162	050823	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-41-4	100 ppm (435mg/m3/8H)	or-rat >2000mg/kg
48. Ethyl benzene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	99-87-6	N/A	or-rat 4750mg/kg
49. p-Isopropyl toluene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	91-20-3	10 ppm (50mg/m3/8H)	or-rat 490mg/kg
50. Naphthalene	35162	050823	0.05	5.00	40005.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	100-42-5	100 ppm	or-rat 5000mg/kg
51. Styrene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	108-98-3	200 ppm	or-rat 5000mg/kg
52. Toluene	35162	050823	0.05	5.00	40006.2	2000	NA	NA	0.017	NA	NA	1999.7	22.9	87-81-6	N/A	or-rat 1390mg/kg
53. 1,2,3-Trichlorobenzene	35162	050823	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.8	22.9	120-82-1	5 ppm (CL) (40mg/m3)	or-rat 756mg/kg
54. 1,2,4-Trichlorobenzene	35162	050823	0.05	5.00	40006.8	2000	NA	NA	0.017	NA	NA	1999.8	23.0	95-63-6	N/A	or-rat 5g/kg
55. 1,2,4-Trimethylbenzene	35162	050823	0.05	5.00	40001.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-87-6	N/A	or-rat 5000mg/kg
56. 1,3,5-Trimethylbenzene	35162	050823	0.05	5.00	40006.7	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-38-3	100 ppm (435mg/m3/8H)	or-rat 5g/kg
57. m-Xylene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	96-06-6	N/A	N/A
58. tert-Butyl benzene	35163	101923	0.05	5.00	40001.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	135-98-6	N/A	or-rat 2240mg/kg
59. sec-Butyl benzene	35163	101923	0.05	5.00	40002.4	2000	NA	NA	0.017	NA	NA	1999.7	22.9	108-90-7	75 ppm (350mg/m3/8H)	or-rat 2290mg/kg
60. Chlorobenzene	35163	101923	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.5	22.9	95-49-8	50 ppm (250mg/m3/8H)	or-rat 3900mg/kg
61. 2-Chlorotoluene	35163	101923	0.05	5.00	40003.3	2000	NA	NA	0.017	NA	NA	1999.7	22.9	106-43-4	N/A	or-rat 2100mg/kg
62. 4-Chlorotoluene	35163	101923	0.05	5.00	40003.3	2000	NA	NA	0.017	NA	NA	1999.7	22.9	95-50-1	50 ppm (300mg/m3) (CL)	or-rat 500mg/kg
63. 1,2-Dichlorobenzene	35163	101923	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	23.0	541-73-1	N/A	or-mus 1060mg/kg
64. 1,3-Dichlorobenzene	35163	101923	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	22.9	108-46-7	75 ppm (450mg/m3/8H)	or-r

Run 16, "P95317 L021624 (2000µg/mL in MeOH)"

Run Length: 60.00 min, 35998 points at 10 points/second.
Created: Sat, Feb 17, 2024 at 8:56:46 AM.
Sampled: Sequence "021624-GC5M1", Method "GC5-M1".
Analyzed using Method "GC5-M1".

Comments
GC5-M1 Analysis by Candice Warren
Column ID SPB-Vocol 105 meter X 0.53mm X 3.0µm film thickness
Flow rates: Total flow=290mL/min., Helium (carrier)=10mL/min.,
Helium(make-up)=10mL/min., Hydrogen(make-up)=40mL/min., Air(make-up)=230mL/min.
Oven Profile: Temp. 1=35°C (Time 1=10 min.), Temp 2=200°C (Time 2=8.75 min.),
Rate = 4°C/min., Total run time=60 min. Injector temp.=200°C, FID Temp.=200°C.
FID Signal = Edaq Channel 1
Standard injection = 0.5µL, Range=3



Peak #	Name	FID RT (min.)
1	Ether	9.97
2	1,1,2-Trichloro-1,2,2-trifluoroethane	10.53
3	1,1-Dichloroethane	11.10
4	Acetonitrile	12.60
5	Iodomethane	12.91
6	Allyl chloride	12.96
7	Carbon disulfide/Methylene chloride	13.04
8	trans-1,2-Dichloroethene	14.07
9	1,1-Dichloroethane	15.74
10	2,2-Dichloropropane	17.74
11	cis-1,2-Dichloroethene	18.00
12	Methoxyvinylbenzene/Methyl acrylate/Chloroform	18.49
13	Isobutene/1,1,1-Trichloroethane	18.91
14	1,1-Dichloropropane	20.46
15	Carbon tetrachloride	20.79
16	Benzene/1,2-Dichloroethane	21.48
17	Trichloroethene	23.88
18	1,2-Dichloropropane	24.52
19	Methyl methacrylate	25.13
20	Bromochloropropane	25.46
21	Dibromomethane/2-Pentene	27.07
22	cis-1,2-Dichloroethene	28.03
23	Toluene	28.73
24	Ethyl methacrylate/trans-1,2-Dichloroethene	29.34
25	1,1,2-Trichloroethane	30.34
26	Tetrachloroethane/1,2-Dichloropropane	31.16
27	Dibromochloromethane	31.84
28	1,2-Dibromomethane	33.06
29	Chlorobenzene	33.40
30	Ethylbenzene/1,1,1,2-Tetrachloroethane	33.85
31	m-Xylene/p-Xylene	35.33
32	o-Xylene	35.70
33	Styrene	35.70
34	Isopropylbenzene/Bromofarm	36.48
35	cis-1,4-Dichloro-3-butene	36.80
36	1,1,2-Trichloropropane	37.23
37	1,2,2-Trichloropropane	37.77
38	n-Propylbenzene	37.92
39	trans-1,4-Dichloro-3-butene	38.05
40	Bromobenzene	38.14
41	1,3,5-Trimethylbenzene	38.90
42	Chlorobenzene	38.93
43	4-Chlorobenzene	38.97
44	tert-Butylbenzene	39.76
45	1,2,4-Trimethylbenzene	39.91
46	Pentachloroethane	40.17
47	sec-Butylbenzene	40.52
48	p-Isopropylbenzene	41.02
49	1,3-Trichlorobenzene	41.42
50	1,4-Dichlorobenzene	41.83
51	n-Butylbenzene	42.18
52	1,2-Dichlorobenzene	42.18
53	1,2-Dibromo-3-chloropropane	46.12
54	Nitrobenzene	46.40
55	1,2,4-Trifluorobenzene	49.28
56	Hexachlorocyclopentadiene	49.72
57	Naphthalene	50.56
58	1,2,3-Trichlorobenzene	51.16

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr. Hamden CT, 06514	Emergency Telephone International	1-352-323-3500
		Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225 H370 P271 P302,332	Highly Flammable Liquid and Vapor Cause damage to organs Use in ventilated area If on skin, wash with soap and water	H301, 311, 331 H351 P280 P305,351,338	Toxic if swallowed, skin contact, inhaled Suspected of causing cancer Use gloves, eye protection/face shield If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))	CAS#:	% (optional)
Methanol METHYL ALCOHOL	67-56-1	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist.
Storage Conditions	Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge. Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1 TWA 200 ppm
Skin notation	TWA 200 ppm
Potential for skin absorption, ingestion and inhalation.	
Personal protective equipment	Respiratory protection Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.	

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)
UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol

IATA
UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Ree 03117/24

CERTIFIED WEIGHT REPORT

Part Number: 95317

Lot Number: 021624

Description: Universal VOA Megamix

69 components

Expiration Date: 021627

Recommended Storage: Freezer (0 °C)

Nominal Concentration (µg/mL): 2000

NIST Test ID#: BUTB

Solvent(s): Lot#
Methanol EG359-USQ12Weight(s) shown below were combined and diluted to (mL): 100.0 0.021 Balance Uncertainty
Flask Uncertainty


Formulated By: Prashant Chauhan		021624
Reviewed By: Pedro L. Renteria		021624
DATE		DATE

Compound	(KMe)	Lot	DR	Initial	Initial	Nominal	Purity	Purity	Uncertainty	Target	Actual	Actual	Expanded	SDS Information			
	Part Number	Number	Factor	Vol. (mL)	Conc.(µg/mL)	Conc (µg/mL)	(%)	Uncertainty	Pipette (mL)	Weight(g)	Weight(g)	Conc (µg/mL)	Uncertainty	(+/-) (µg/mL)	CASE	OSHA PEL (TWA)	LD50
1. Acetonitrile	(0324)	021644	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20020	2001.3	8.1	75-05-8	40 ppm (70mg/m3/8H)	or-rat 2450mg/kg	
2. Allyl chloride (3-Chloropropene)	(0325)	102396	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	107-05-1	1 ppm (3mg/m3/8H)	or-rat 700mg/kg	
3. Carbon disulfide	(0060)	MKCR8561	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20023	2001.6	8.1	75-15-0	4 ppm (12mg/m3) (skin)	or-rat 1200mg/kg	
4. cis-1,4-Dichloro-2-butene	(1196)	14718EF	NA	NA	NA	2000	95	0.2	NA	0.21056	0.21069	2001.1	8.5	1478-11-5	NA	N/A	
5. trans-1,4-Dichloro-2-butene	(0486)	MKBP9041V	NA	NA	NA	2000	96.5	0.2	NA	0.20731	0.20746	2001.7	8.4	110-57-6	NA	N/A	
6. Diethyl ether	(0153)	IK18CAS000C	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	60-29-7	NA	N/A	
7. Ethyl methacrylate	(0381)	06128PX	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20230	2002.3	8.2	97-63-2	NA	N/A	or-rat 14800mg/kg
8. Iodomethane	(0489)	SHBF8718V	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20121	2001.5	8.2	74-88-4	5 ppm(28mg/m3/8H)(skin)	or-rat 76mg/kg	
9. 2-Methyl-1-propanol	(0445)	15241EB	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20120	2001.4	8.1	78-83-1	50 ppm (150mg/m3/8H)	or-rat 2460mg/kg	
10. Methacrylonitrile	(0472)	00427ET	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	128-98-7	1 ppm (3mg/m3/8H)(skin)	or-rat 120mg/kg	
11. Methyl acrylate	(1075)	SHBK0679	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	96-33-3	10 ppm(35mg/m3/8H)(skin)	or-rat 277mg/kg	
12. Methyl methacrylate	(0404)	MKGW6137V	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20041	2001.6	8.1	80-62-6	100 ppm (410mg/m3/8H)	or-rat 7872mg/kg	
13. Nitrobenzene	(0228)	01213TV	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20220	2001.3	8.2	98-95-3	1 ppm (5mg/m3/8H)(skin)	or-rat 780mg/kg	
14. 2-Nitropropane	(0461)	14002JK	NA	NA	NA	2000	97.3	0.2	NA	0.20560	0.20577	2001.6	8.3	78-46-9	10 ppm (30mg/m3/8H)	or-rat 720mg/kg	
15. Perchloroethane	(0460)	HGA01	NA	NA	NA	2000	98	0.2	NA	0.20413	0.20430	2001.6	8.3	78-01-7	NA	N/A	
16. 1,1,2-Trichlorotrifluoroethane	(0474)	18930	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20225	2001.8	8.2	76-13-1	1000 ppm (7600mg/m3/8H)	or-rat 43g/kg	
17. Bromodichloromethane	35171	101623	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-27-4	NA	or-rat 915mg/kg	
18. Dibromochloromethane	35171	101623	0.05	5.00	40002.1	2000	NA	NA	0.017	NA	NA	1999.6	23.0	124-48-1	NA	or-rat 848mg/kg	
19. cis-1,2-Dichloroethene	35171	101623	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.7	22.9	156-59-2	NA	N/A	
20. trans-1,2-Dichloroethene	35171	101623	0.05	5.00	40003.2	2000	NA	NA	0.017	NA	NA	1999.6	22.9	156-60-5	NA	N/A	
21. Methylene chloride	35171	101623	0.05	5.00	40002.8	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-09-2	500 ppm	or-rat 1235mg/kg	
22. 1,1-Dichloroethane	32261	102023	0.10	10.00	20001.6	2000	NA	NA	0.042	NA	NA	1999.8	20.5	78-25-2	0.5 ppm (5mg/m3) (skin)	or-rat 933mg/kg	
23. Bromoform	95321	020724	0.10	10.00	20003.2	2000	NA	NA	0.042	NA	NA	1999.6	20.4	56-23-5	2 ppm (12.6mg/m3/8H)	or-rat 2350mg/kg	
24. Carbon tetrachloride	95321	020724	0.10	10.00	20024.0	2000	NA	NA	0.042	NA	NA	2001.3	20.5	67-68-3	60 ppm (240mg/m3) (CL)	or-rat 908mg/kg	
25. Chloroform	95321	020724	0.10	10.00	20002.9	2000	NA	NA	0.042	NA	NA	1999.8	20.5	74-95-3	100 ppm	or-rat 725mg/kg	
26. Dibromomethane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.5	594-20-7	NA	N/A	
27. 1,1-Dichloroethane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.5	74-95-3	NA	N/A	
28. 2,2-Dichloropropane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.5	75-34-3	100 ppm	or-rat 108mg/kg	
29. Trichloroethene	95321	020724	0.10	10.00	20201.1	2000	NA	NA	0.042	NA	NA	2019.8	20.8	127-18-4	25 ppm (170mg/m3/8H)(final)	or-rat 2629mg/kg	
30. 1,1,1-Trichloroethane	95321	020724	0.10	10.00	20003.0	2000	NA	NA	0.042	NA	NA	1999.8	20.5	71-55-6	350 ppm (1900mg/m3/8H)	or-rat 10300mg/kg	
31. 1,2-Dibromo-3-chloropropane	35161	112322	0.05	5.00	40018.5	2000	NA	NA	0.017	NA	NA	2000.3	22.9	96-12-8	0.001 ppm	or-rat 170mg/kg	
32. 1,2-Dibromoethane	35161	112322	0.05	5.00	40024.8	2000	NA	NA	0.017	NA	NA	2000.7	22.9	106-93-4	20 ppm (8H)	or-rat 108mg/kg	
33. 1,2-Dichloroethane	35161	112322	0.05	5.00	40018.0	2000	NA	NA	0.017	NA	NA	2000.4	22.9	107-06-2	50 ppm (8H)	or-rat 670mg/kg	
34. 1,2-Dichloropropane	35161	112322	0.05	5.00	40051.0	2000	NA	NA	0.017	NA	NA	2002.0	22.9	78-87-5	75 ppm (350mg/m3/8H)	or-rat 1947mg/kg	
35. 1,3-Dichloropropane	35161	112322	0.05	5.00	40005.9	2000	NA	NA	0.017	NA	NA	1999.8	22.9	142-28-9	NA	or-mus 3600mg/kg	
36. 1,1-Dichloropropane	35161	112322	0.05	5.00	40012.1	2000	NA	NA	0.017	NA	NA	2000.1	28.7	563-58-6	NA	N/A	
37. cis-1,3-Dichloropropane	35161	112322	0.05	5.00	40010.0	2000	NA	NA	0.017	NA	NA	2000.0	23.0	10061-01-5	NA	N/A	
38. trans-1,3-Dichloropropane	35161	112322	0.05	5.00	40017.6	2000	NA	NA	0.017	NA	NA	2000.4	23.0	10061-02-6	NA	N/A	
39. Hexachloro-1,3-butadiene	35161	112322	0.05	5.00	40021.9	2000	NA	NA	0.017	NA	NA	2000.6	29.7	87-68-3	0.02 ppm (0.24mg/m3/8H)	or-rat 82mg/kg	
40. 1,1,1,2-Tetrachloroethane	35161	112322	0.05	5.00	40011.9	2000	NA	NA	0.017	NA	NA	2000.1	22.9	830-20-6	NA	or-rat 670mg/kg	
41. 1,1,2,2-Tetrachloroethane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	79-34-5	5 ppm (35mg/m3/8H)(skin)	or-rat 800mg/kg	
42. 1,1,2-Trichloroethane	35161	112322	0.05	5.00	40006.6	2000	NA	NA	0.017	NA	NA	1999.8	23.0	79-00-5	10 ppm (45mg/m3/8H)(skin)	or-rat 936mg/kg	
43. Trichloroethene	35161	112322	0.05	5.00	40029.0	2000	NA	NA	0.017	NA	NA	2000.9	22.9	79-01-6	50 ppm (270mg/m3/8H)	or-mus 2402mg/kg	
44. 1,2,3-Trichloropropane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	96-18-4	10 ppm (60mg/m3/8H)	or-rat 149.8mg/kg	
45. Benzene	35162	050823	0.05	5.00	40006.0	2000	NA	NA	0.017	NA	NA	1999.7	22.9	71-43-2	1 ppm	or-rat 4894mg/kg	
46. Bromobenzene	35162	050823	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	104-51-8	NA	or-rat 2699mg/kg	
47. n-Butyl benzene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-41-4	100 ppm (435mg/m3/8H)	or-rat >2000mg/kg	
48. Ethyl benzene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	99-87-6	NA	or-rat 4750mg/kg	
49. p-Isopropyl toluene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	91-20-3	10 ppm (50mg/m3/8H)	or-rat 490mg/kg	
50. Naphthalene	35162	050823	0.05	5.00	40006.2	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-42-5	100 ppm	or-rat 5000mg/kg	
51. Styrene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-88-3	200 ppm	or-rat 5000mg/kg	
52. Toluene	35162	050823	0.05	5.00	40008.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	87-61-6	NA	or-mus 1390mg/kg	
53. 1,2,3-Trichlorobenzene	35162	050823	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.8	22.9	120-82-1	5 ppm (CL) (40mg/m3)	or-rat 756mg/kg	
54. 1,2,4-Trichlorobenzene	35162	050823	0.05	5.00	40001.8	2000	NA	NA	0.017	NA	NA	1999.6	23.0	95-63-6	NA	or-rat 5g/kg	
55. 1,3,5-Trimethylbenzene	35162	050823	0.05	5.00	40006.7	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-87-8	NA	or-rat 5000mg/kg	
56. m-Xylene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-38-3	100 ppm (435mg/m3/8H)	or-rat 5g/kg	
57. tert-Butyl benzene	35163	101923	0.05	5.00	40001.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	98-06-6	NA	N/A	
58. sec-Butyl benzene	35163	101923	0.05	5.00	40002.4	2000	NA	NA	0.017	NA	NA	1999.8	22.9	135-98-8	NA	N/A	
59. Chlorobenzene	35163	101923	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	108-90-7	75 ppm (350mg/m3/8H)	or-rat 2240mg/kg	
60. 2-Chlorotoluene	35163	101923	0.05	5.00	40000.3	2000	NA	NA	0.017	NA	NA	1999.5	22.9	95-49-8	50 ppm (250mg/m3/8H)	or-rat 2290mg/kg	
61. 4-Chlorotoluene	35163	101923	0.05	5.00	40003.3	2000	NA	NA	0.017	NA	NA	1999.7	22.9	106-43-4	NA	or-rat 3900mg/kg	
62. 1,2-Dichlorobenzene	35163	101923	0.05	5.00	40003.3	2000	NA	NA	0.017	NA	NA	1999.7	22.9	106-43-4	NA	or-rat 2100mg/kg	
63. 1,3-Dichlorobenzene	35163	101923	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.7	22.9	95-50-1	50 ppm (300mg/m3) (CL)	or-rat 500mg/kg	

Run 16, "P95317 L021624 (2000µg/mL in MeOH)"

Run Length: 60.00 min, 35998 points at 10 points/second.
Created: Sat, Feb 17, 2024 at 8:56:46 AM.
Sampled: Sequence "021624-GC5M1", Method "GC5-M1".
Analyzed using Method "GC5-M1".

Comments
GC5-M1 Analysis by Candice Warren
Column ID SPB-Vocol 105 meter X 0.53mm X 3.0µm film thickness
Flow rates: Total flow=290mL/min., Helium (carrier)=10mL/min.,
Helium(make-up)=10mL/min., Hydrogen(make-up)=40mL/min., Air(make-up)=230mL/min.
Oven Profile: Temp. 1=35°C (Time 1=10 min.), Temp 2=200°C (Time 2=8.75 min.),
Rate = 4°C/min., Total run time=60 min. Injector temp.=200°C, FID Temp.=200°C.
FID Signal = Edaq Channel 1
Standard injection = 0.5µL, Range=3



Peak #	Name	FID RT (min.)
1	Ether	9.97
2	1,1,2-Trichloro-1,2,2-trifluoroethane	10.53
3	1,1-Dichloroethane	11.10
4	Acetonitrile	12.60
5	Iodomethane	12.91
6	Allyl chloride	12.96
7	Carbon disulfide/Methylene chloride	13.04
8	trans-1,2-Dichloroethene	14.07
9	1,1-Dichloroethane	15.74
10	2,2-Dichloropropane	17.74
11	cis-1,2-Dichloroethene	18.00
12	Hexamethylcyclotriphosphazene/Hexyl acrylate/Chloroform	18.49
13	Isobutanol/1,1,1-Trichloroethane	18.91
14	1,1-Dichloropropane	20.46
15	Carbon tetrachloride	20.79
16	Benzene/1,2-Dichloroethane	21.48
17	Trichloroethene	23.88
18	1,2-Dichloropropane	24.52
19	Methyl methacrylate	25.13
20	Bromodichloromethane	25.46
21	Dibromomethane/2-Pentene	27.07
22	cis-1,2-Dichloroethene	28.03
23	Toluene	28.73
24	Ethyl methacrylate/Trans-1,2-Dichloroethene	29.34
25	1,1,2-Trichloroethane	30.34
26	Tetrachloroethane/1,2-Dichloropropane	31.16
27	Dibromochloromethane	31.84
28	1,2-Dibromomethane	33.06
29	Chlorobenzene	33.40
30	Ethylbenzene/1,1,1,2-Tetrachloroethane	33.85
31	m-Xylene/p-Xylene	35.33
32	o-Xylene	35.70
33	Styrene	35.70
34	Isopropylbenzene/Bromofarm	36.48
35	cis-1,4-Dichloro-3-butene	36.80
36	1,1,2-Trichloropropane	37.23
37	n-Propylbenzene	37.77
38	trans-1,4-Dichloro-3-butene	37.92
39	Bromobenzene	38.05
40	1,3,5-Trimethylbenzene	38.14
41	2-Chlorobenzene	38.50
42	4-Chlorobenzene	38.63
43	tert-Butylbenzene	38.77
44	sec-Butylbenzene	39.76
45	1,2,4-Trimethylbenzene	40.17
46	1,2,4-Trimethylbenzene	40.52
47	p-Isopropylbenzene	41.02
48	1,3-Dichlorobenzene	41.42
49	1,4-Dichlorobenzene	41.83
50	n-Butylbenzene	42.52
51	1,2-Dichloropropane	42.18
52	1,2-Dichloropropane	46.17
53	1,2-Dichloro-3-methylpropane	46.40
54	Nitrobenzene	49.26
55	1,2,4-Trichlorobenzene	49.72
56	Hexachlorocyclopentadiene	50.56
57	Naphthalene	51.16
58	1,2,3-Trichlorobenzene	

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr. Hamden CT, 06514	Emergency Telephone International	1-352-323-3500
		Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225 H370 P271 P302,332	Highly Flammable Liquid and Vapor Cause damage to organs Use in ventilated area If on skin, wash with soap and water	H301, 311, 331 H351 P280 P305,351,338	Toxic if swallowed, skin contact, inhaled Suspected of causing cancer Use gloves, eye protection/face shield If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))	CAS#	% (optional)
Methanol METHYL ALCOHOL	67-56-1	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist.
Storage Conditions	Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge. Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1 TWA 200 ppm
Skin notation	TWA 200 ppm
Potential for skin absorption, ingestion and inhalation.	
Personal protective equipment	Respiratory protection Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.	

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)
UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol

IATA
UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Certified Reference Material CRM

Dec 09/17/24

2 vial

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT

Part Number: **91980**
Lot Number: **091424**
Description: **Acrolein**Solvent(s):
Water Lot#
072324Q

<i>Justin Dippold</i>		091424
Formulated By:	Justin Dippold	DATE
<i>Pedro L. Rentas</i>		091424
Reviewed By:	Pedro L. Rentas	DATE

Expiration Date: 101424

Recommended Storage: Refrigerate (4 °C)

Nominal Concentration (µg/mL): 5000

NIST Test ID#: 6UTB

5E-05 Balance Uncertainty

Weight(s) shown below were combined and diluted to (mL): 10.0 0.001 Flask Uncertainty

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05175	5008.9	52.5	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.), Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately. Long term storage is not recommended. Please contact our technical department if further information is required.

Abundance TIC: [BSB2]79005.D

Abundance

Scan 232 (8.927 min): [BSB2]79005.D

250000 8.93

200000

150000

100000

50000

Time--> 0

10.00 15.00 20.00 25.00 30.00 35.00 40.00 45.00 50.00 55.00 60.00

m/z--> 0

20 30 40 50 60 70 80 90 100 110 120 130 140 150 160 170

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).



Certified Reference Material CRM

Dec 09/17/24

2 vial

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT

Part Number: **91980**
Lot Number: **091424**
Description: **Acrolein**Solvent(s):
Water Lot#
072324Q

<i>Justin Dippold</i>		091424
Formulated By:	Justin Dippold	DATE
<i>Pedro L. Rentas</i>		091424
Reviewed By:	Pedro L. Rentas	DATE

Expiration Date: 101424

Recommended Storage: Refrigerate (4 °C)

Nominal Concentration (µg/mL): 5000

NIST Test ID#: 6UTB

5E-05 Balance Uncertainty

Weight(s) shown below were combined and diluted to (mL): 10.0 0.001 Flask Uncertainty

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05175	5008.9	52.5	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.), Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately. Long term storage is not recommended. Please contact our technical department if further information is required.

Abundance TIC: [BSB2]79005.D

Abundance

Scan 232 (8.927 min): [BSB2]79005.D

250000 8.93

200000

150000

100000

50000

Time--> 0

10.00 15.00 20.00 25.00 30.00 35.00 40.00 45.00 50.00 55.00 60.00

m/z--> 0

20 30 40 50 60 70 80 90 100 110 120 130 140 150 160 170

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).



Certified Reference Material CRM



CERTIFIED WEIGHT REPORT

Part Number: **95318**
Lot Number: **120524**
Description: **2-Chloroethyl vinyl ether**

Solvent(s):
Methanol

Lot#
EJ143-US

Expiration Date: **120527**
Recommended Storage: **Refrigerate (4 °C)**
Nominal Concentration (µg/mL): **10000**
NIST Test ID#: **6UTB**

See 1216124
20 vial
✓ 14630 to
✓ 14649



Weight(s) shown below were combined and diluted to (mL): **50.0**

5E-05 Balance Uncertainty
0.001 Flask Uncertainty

<i>Prashant Chauhan</i>		120524
Formulated By:	Prashant Chauhan	DATE
<i>Pedro L. Rentas</i>		120524
Reviewed By:	Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	40.5	110-75-8	N/A	ori-rat 250mg/kg

Method: GC6MSD-1.M. Detector: MSD. Column: (60m X 0.25mm X 1.5 µm). Oven Profile: Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min., Injector B Temp = 200°C, Detector B Temp = 220°C. Analyst: Candice Warren.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr.	Emergency Telephone International	1-352-323-3500
	Hamden CT, 06514	Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	Highly Flammable Liquid and Vapor	H301, 311, 331	Toxic if swallowed, skin contact, inhaled
H370	Cause damage to organs	H351	Suspected of causing cancer
P271	Use in ventilated area	P280	Use gloves, eye protection/face shield
P302,332	If on skin, wash with soap and water	P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))		% (optional)
Methanol	METHYL ALCOHOL	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1	67-56-1	TWA 200 ppm
Skin notation			TWA 200 ppm
Potential for skin absorption, ingestion and inhalation.			
Personal protective equipment	Respiratory protection	Handle with gloves. Gloves must be inspected prior to use.	Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.			

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Certified Reference Material CRM



CERTIFIED WEIGHT REPORT

Part Number: **95318**
Lot Number: **120524**
Description: **2-Chloroethyl vinyl ether**

Solvent(s):
Methanol

Lot#
EJ143-US

Expiration Date: **120527**
Recommended Storage: **Refrigerate (4 °C)**
Nominal Concentration (µg/mL): **10000**
NIST Test ID#: **6UTB**

See 1216124
20 vial
✓ 14630 to
✓ 14649



Weight(s) shown below were combined and diluted to (mL): **50.0**

5E-05 Balance Uncertainty
0.001 Flask Uncertainty

<i>Prashant Chauhan</i>		120524
Formulated By:	Prashant Chauhan	DATE
<i>Pedro L. Rentas</i>		120524
Reviewed By:	Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	40.5	110-75-8	N/A	ori-rat 250mg/kg

Method: GC6MSD-1.M. Detector: MSD. Column: (60m X 0.25mm X 1.5 µm). Oven Profile: Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min., Injector B Temp = 200°C, Detector B Temp = 220°C. Analyst: Candice Warren.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr.	Emergency Telephone International	1-352-323-3500
	Hamden CT, 06514	Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	Highly Flammable Liquid and Vapor	H301, 311, 331	Toxic if swallowed, skin contact, inhaled
H370	Cause damage to organs	H351	Suspected of causing cancer
P271	Use in ventilated area	P280	Use gloves, eye protection/face shield
P302,332	If on skin, wash with soap and water	P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))		% (optional)
Methanol	METHYL ALCOHOL	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1	TWA 200 ppm
Skin notation		TWA 200 ppm
Potential for skin absorption, ingestion and inhalation.		
Personal protective equipment	Respiratory protection	Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.		

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Certified Reference Material CRM



CERTIFIED WEIGHT REPORT

Part Number: **95318**
Lot Number: **120524**
Description: **2-Chloroethyl vinyl ether**

Solvent(s):
Methanol

Lot#
EJ143-US

Expiration Date: **120527**
Recommended Storage: **Refrigerate (4 °C)**
Nominal Concentration (µg/mL): **10000**
NIST Test ID#: **6UTB**

See 1216124
20 vial
✓ 14630 to
✓ 14649



Weight(s) shown below were combined and diluted to (mL): **50.0**

5E-05 Balance Uncertainty
0.001 Flask Uncertainty

<i>Prashant Chauhan</i>		120524
Formulated By:	Prashant Chauhan	DATE
<i>Pedro L. Rentas</i>		120524
Reviewed By:	Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	40.5	110-75-8	N/A	ori-rat 250mg/kg

Method: GC6MSD-1.M. Detector: MSD. Column: (60m X 0.25mm X 1.5 µm). Oven Profile: Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min., Injector B Temp = 200°C, Detector B Temp = 220°C. Analyst: Candice Warren.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr.	Emergency Telephone International	1-352-323-3500
	Hamden CT, 06514	Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	Highly Flammable Liquid and Vapor	H301, 311, 331	Toxic if swallowed, skin contact, inhaled
H370	Cause damage to organs	H351	Suspected of causing cancer
P271	Use in ventilated area	P280	Use gloves, eye protection/face shield
P302,332	If on skin, wash with soap and water	P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))		% (optional)
Methanol	METHYL ALCOHOL	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1	67-56-1	TWA 200 ppm
Skin notation			TWA 200 ppm
Potential for skin absorption, ingestion and inhalation.			
Personal protective equipment	Respiratory protection	Handle with gloves. Gloves must be inspected prior to use.	Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.			

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Certified Reference Material CRM



CERTIFIED WEIGHT REPORT

Part Number: **95318**
Lot Number: **120524**
Description: **2-Chloroethyl vinyl ether**

Solvent(s):
Methanol

Lot#
EJ143-US

Expiration Date: **120527**
Recommended Storage: **Refrigerate (4 °C)**
Nominal Concentration (µg/mL): **10000**
NIST Test ID#: **6UTB**

See 1216124
20 vial
✓ 14630 to
✓ 14649



Weight(s) shown below were combined and diluted to (mL): **50.0**

5E-05 Balance Uncertainty
0.001 Flask Uncertainty

<i>Prashant Chauhan</i>		120524
Formulated By:	Prashant Chauhan	DATE
<i>Pedro L. Rentas</i>		120524
Reviewed By:	Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	40.5	110-75-8	N/A	ori-rat 250mg/kg

Method: GC6MSD-1.M. Detector: MSD. Column: (60m X 0.25mm X 1.5 µm). Oven Profile: Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min., Injector B Temp = 200°C, Detector B Temp = 220°C. Analyst: Candice Warren.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr.	Emergency Telephone International	1-352-323-3500
	Hamden CT, 06514	Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	Highly Flammable Liquid and Vapor	H301, 311, 331	Toxic if swallowed, skin contact, inhaled
H370	Cause damage to organs	H351	Suspected of causing cancer
P271	Use in ventilated area	P280	Use gloves, eye protection/face shield
P302,332	If on skin, wash with soap and water	P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))		% (optional)
Methanol	METHYL ALCOHOL	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1	67-56-1	TWA 200 ppm
Skin notation			TWA 200 ppm
Potential for skin absorption, ingestion and inhalation.			
Personal protective equipment	Respiratory protection	Handle with gloves. Gloves must be inspected prior to use.	Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.			

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30067 **Lot No.:** A0191805

Description : 4-Bromofluorobenzene Standard

4-Bromofluorobenzene Standard 2,500µg/mL, P&T Methanol,
1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : November 30, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1-Bromo-4-fluorobenzene (BFB)	460-00-4	184975	99%	2,483.9 µg/mL	+/- 139.5488

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:


40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Alicia Leathers - Operation Technician I

Date Mixed: 17-Nov-2022

Balance Serial # B251644995


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 21-Nov-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ μ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555582 **Lot No.:** A0196865

Description : Custom 8260A/B Surrogate Mix

Custom 8260A/B Surrogate Mix 25,000µg/mL, P&T Methanol,
1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2-Dichloroethane-d4	17060-07-0	PR-32845	99%	25,036.0 µg/mL	+/- 1,417.9179
2	1-Bromo-4-fluorobenzene (BFB)	460-00-4	184975	99%	25,132.0 µg/mL	+/- 1,423.3549
3	Dibromofluoromethane	1868-53-7	022013	99%	25,040.0 µg/mL	+/- 1,418.1445
4	Toluene-d8	2037-26-5	PR-33397	99%	25,028.0 µg/mL	+/- 1,417.4648

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Russ Bookhamer - Operations Technician I

Date Mixed: 11-Apr-2023

Balance: 1127510105

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30489 **Lot No.:** A0209618

Description : 8260B Acetates Mix

8260B Acetates Mix 2,000 µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : September 30, 2025 **Storage:** -20°C or colder

Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Methyl acetate	79-20-9	SHBP3100	99%	2,019.3 µg/mL	+/- 69.7974
2	Vinyl acetate	108-05-4	RP231030CTH	98%	2,016.8 µg/mL	+/- 69.7112
3	Ethyl acetate	141-78-6	SHBQ9682	99%	2,010.7 µg/mL	+/- 69.4979
4	Isopropyl acetate	108-21-4	BCCG7069	99%	2,016.0 µg/mL	+/- 69.6822
5	Propyl acetate	109-60-4	P8XLN	99%	2,008.0 µg/mL	+/- 69.4057
6	Butyl acetate	123-86-4	SHBP6314	99%	2,007.3 µg/mL	+/- 69.3826
7	Amyl acetate	628-63-7	41325/1	97%	2,004.7 µg/mL	+/- 69.2905

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods. Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this

reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler
Sam Moodler - Operations Tech I

Date Mixed: 28-Mar-2024

Balance Serial # B707717271

Dillon Murphy
Dillon Murphy - Operations Technician I

Date Passed: 01-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30489 **Lot No.:** A0209618

Description : 8260B Acetates Mix

8260B Acetates Mix 2,000 µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : September 30, 2025 **Storage:** -20°C or colder

Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Methyl acetate	79-20-9	SHBP3100	99%	2,019.3 µg/mL	+/- 69.7974
2	Vinyl acetate	108-05-4	RP231030CTH	98%	2,016.8 µg/mL	+/- 69.7112
3	Ethyl acetate	141-78-6	SHBQ9682	99%	2,010.7 µg/mL	+/- 69.4979
4	Isopropyl acetate	108-21-4	BCCG7069	99%	2,016.0 µg/mL	+/- 69.6822
5	Propyl acetate	109-60-4	P8XLN	99%	2,008.0 µg/mL	+/- 69.4057
6	Butyl acetate	123-86-4	SHBP6314	99%	2,007.3 µg/mL	+/- 69.3826
7	Amyl acetate	628-63-7	41325/1	97%	2,004.7 µg/mL	+/- 69.2905

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods. Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this

reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler
Sam Moodler - Operations Tech I

Date Mixed: 28-Mar-2024

Balance Serial # B707717271

Dillon Murphy
Dillon Murphy - Operations Technician I

Date Passed: 01-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555581 **Lot No.:** A0210184

Description : Custom 8260 Internal Standard Mix
Custom 8260 Internal Standard Mix 25,000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL


Expiration Date : April 30, 2027 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dichlorobenzene-d4	3855-82-1	PR-30447	99%	25,212.0 µg/mL	+/- 1,427.8857
2	1,4-Difluorobenzene	540-36-3	MKCS8657	99%	25,220.0 µg/mL	+/- 1,428.3388
3	Chlorobenzene-d5	3114-55-4	PR-31132	99%	25,116.0 µg/mL	+/- 1,422.4487
4	Pentafluorobenzene	363-72-4	MKCR9383	99%	25,180.0 µg/mL	+/- 1,426.0734

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%


John Friedline - Operations Technician I

Date Mixed: 11-Apr-2024

Balance: 1127510105

REVISED
By Restek Review of SDS per April 18, 2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Dec 12/17/24
30 v. 4
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus
V14697-to-14726



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30006 **Lot No.:** A0210618
Description : VOA Calibration Mix #1
VOA Calibration Mix #1 5,000µg/mL, P&T Methanol/Water(90:10), 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : July 31, 2027 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Acetone	67-64-1	SHBQ8504	99%	5,014.8 µg/mL	+/- 173.2883
2	2-Butanone (MEK)	78-93-3	SHBQ4704	99%	5,012.4 µg/mL	+/- 173.2054
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP9200	99%	5,011.6 µg/mL	+/- 173.1777
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,013.0 µg/mL	+/- 173.2261

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol/Water (90:10)
CAS # 67-56-1/7732-18-5
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Dakota Parson - Operations Technician I

Date Mixed: 22-Apr-2024

Balance Serial # B707717271


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Dec 12/17/24
30 v. 4
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus
V14697-to-14726



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30006 **Lot No.:** A0210618
Description : VOA Calibration Mix #1
VOA Calibration Mix #1 5,000µg/mL, P&T Methanol/Water(90:10), 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : July 31, 2027 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Acetone	67-64-1	SHBQ8504	99%	5,014.8 µg/mL	+/- 173.2883
2	2-Butanone (MEK)	78-93-3	SHBQ4704	99%	5,012.4 µg/mL	+/- 173.2054
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP9200	99%	5,011.6 µg/mL	+/- 173.1777
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,013.0 µg/mL	+/- 173.2261

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol/Water (90:10)
CAS # 67-56-1/7732-18-5
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Dakota Parson - Operations Technician I

Date Mixed: 22-Apr-2024

Balance Serial # B707717271


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Dec 12/17/24
30 v. 4
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

✓ 14697-to-14726



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30006 **Lot No.:** A0210618

Description : VOA Calibration Mix #1

VOA Calibration Mix #1 5,000µg/mL, P&T Methanol/Water(90:10), 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Acetone	67-64-1	SHBQ8504	99%	5,014.8 µg/mL	+/- 173.2883
2	2-Butanone (MEK)	78-93-3	SHBQ4704	99%	5,012.4 µg/mL	+/- 173.2054
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP9200	99%	5,011.6 µg/mL	+/- 173.1777
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,013.0 µg/mL	+/- 173.2261

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol/Water (90:10)
CAS # 67-56-1/7732-18-5
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Dakota Parson - Operations Technician I

Date Mixed: 22-Apr-2024

Balance Serial # B707717271


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Dec 12/17/24
30 v. 4
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus
V14697-to-14726



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30006 **Lot No.:** A0210618
Description : VOA Calibration Mix #1
VOA Calibration Mix #1 5,000µg/mL, P&T Methanol/Water(90:10), 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : July 31, 2027 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Acetone	67-64-1	SHBQ8504	99%	5,014.8 µg/mL	+/- 173.2883
2	2-Butanone (MEK)	78-93-3	SHBQ4704	99%	5,012.4 µg/mL	+/- 173.2054
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP9200	99%	5,011.6 µg/mL	+/- 173.1777
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,013.0 µg/mL	+/- 173.2261

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol/Water (90:10)
CAS # 67-56-1/7732-18-5
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Dakota Parson - Operations Technician I

Date Mixed: 22-Apr-2024

Balance Serial # B707717271


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

10 vial.
Dec: 12/09/24
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus
V14667-5
V14676



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225 **Lot No.:** A0214960
Description : Bromochloromethane Standard
Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : August 31, 2029 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	SYN240416CTH	99%	2,012.0 µg/mL	+/- 113.0519

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Stacey Wanner - Operations Technician I

Date Mixed: 08-Aug-2024

Balance Serial # 1127510105

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 14-Aug-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ μ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

10 vial.
Dec: 12/09/24
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus
V14667
V14676



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225 **Lot No.:** A0214960
Description : Bromochloromethane Standard
Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : August 31, 2029 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	SYN240416CTH	99%	2,012.0 µg/mL	+/- 113.0519

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Stacey Wanner - Operations Technician I

Date Mixed: 08-Aug-2024

Balance Serial # 1127510105

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 14-Aug-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ μ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

10 vial.
Dec: 12/09/24
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus

V14667-5
V14676



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225 **Lot No.:** A0214960
Description : Bromochloromethane Standard
Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : August 31, 2029 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	SYN240416CTH	99%	2,012.0 µg/mL	+/- 113.0519

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Stacey Wanner - Operations Technician I

Date Mixed: 08-Aug-2024

Balance Serial # 1127510105

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 14-Aug-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ μ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

10 vial.
Dec: 12/09/24
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus
V14667-5
V14676



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225 **Lot No.:** A0214960
Description : Bromochloromethane Standard
Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : August 31, 2029 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	SYN240416CTH	99%	2,012.0 µg/mL	+/- 113.0519

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Stacey Wanner - Operations Technician I

Date Mixed: 08-Aug-2024

Balance Serial # 1127510105

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 14-Aug-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ μ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Rec 12/17/24
30 ml
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus

V14727 to
V14756



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30042 **Lot No.:** A0216826

Description : 502.2 Calibration Mix #1
502.2 Calibration Mix #1 2,000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : May 31, 2031 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Dichlorodifluoromethane (CFC-12)	75-71-8	00022922	99%	2,000.9 µg/mL	+/- 112.4144
2	Chloromethane (methyl chloride)	74-87-3	00022694	99%	2,000.7 µg/mL	+/- 112.3998
3	Vinyl chloride	75-01-4	00015559	99%	2,000.3 µg/mL	+/- 112.3779
4	Bromomethane (methyl bromide)	74-83-9	00017022	99%	2,001.8 µg/mL	+/- 112.4650
5	Chloroethane (ethyl chloride)	75-00-3	107-401039114-1	99%	2,000.1 µg/mL	+/- 112.3700
6	Trichlorofluoromethane (CFC-11)	75-69-4	MKCJ8658	99%	2,000.7 µg/mL	+/- 112.3992

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

60m x 0.25mm x 1.4µm
Rtx-502.2 (cat.#10916)

Carrier Gas:

helium-constant flow 2.0 mL/min.

Temp. Program:

40°C (hold 6 min.) to 100°C
@ 6°C/min.

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


MSD

Split Vent:

Split ratio 10:1

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Tom Suckal - Mix Technician

Date Mixed: 23-Sep-2024

Balance Serial # B707717271

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 04-Oct-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Rec 12/17/24
CERTIFIED REFERENCE MATERIAL

30 ml
Certificate of Analysis
chromatographic plus

V14727 to
V14756



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30042 **Lot No.:** A0216826
Description : 502.2 Calibration Mix #1
502.2 Calibration Mix #1 2,000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : May 31, 2031 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Dichlorodifluoromethane (CFC-12)	75-71-8	00022922	99%	2,000.9 µg/mL	+/- 112.4144
2	Chloromethane (methyl chloride)	74-87-3	00022694	99%	2,000.7 µg/mL	+/- 112.3998
3	Vinyl chloride	75-01-4	00015559	99%	2,000.3 µg/mL	+/- 112.3779
4	Bromomethane (methyl bromide)	74-83-9	00017022	99%	2,001.8 µg/mL	+/- 112.4650
5	Chloroethane (ethyl chloride)	75-00-3	107-401039114-1	99%	2,000.1 µg/mL	+/- 112.3700
6	Trichlorofluoromethane (CFC-11)	75-69-4	MKCJ8658	99%	2,000.7 µg/mL	+/- 112.3992

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

60m x 0.25mm x 1.4µm
Rtx-502.2 (cat.#10916)

Carrier Gas:

helium-constant flow 2.0 mL/min.

Temp. Program:

40°C (hold 6 min.) to 100°C
@ 6°C/min.

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


MSD

Split Vent:

Split ratio 10:1

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Tom Suckal - Mix Technician

Date Mixed: 23-Sep-2024

Balance Serial # B707717271

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 04-Oct-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

✓ 14842 to 14846



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30470 **Lot No.:** A0217535

Description : tert-Butanol Standard

tert-Butanol Std 50,000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : October 31, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	tert-Butanol (TBA)	75-65-0	SHBQ8002-1	99%	50,007.5 µg/mL	+/- 717.6137

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

A. O. E.
Aaron Enyart - Operations Tech I

Date Mixed: 07-Oct-2024

Balance Serial # B251644995

Brittany Federinko
Brittany Federinko - Operations Tech I

Date Passed: 09-Oct-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ μ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

2014 Dec 01/08/21
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic

V14803 - V14822



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555408-SL **Lot No.:** A0220471
Description : Custom Vinyl Acetate Standard
Custom Vinyl Acetate Standard 8,000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : June 30, 2026 **Storage:** -20°C or colder
Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Vinyl acetate	108-05-4	RD240423RSR	99%	8,066.0 µg/mL	+/- 278.7979

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods. Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Ethan Winiarski - Operations Tech I

Date Mixed: 24-Dec-2024

Balance Serial # 1127510105

Dillan Murphy - Operations Technician I

Date Passed: 02-Jan-2025

REVIEWED
By Jennifer Pollock at 7:12 am, Jan 05, 2025

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

2014 Dec 01/08/21
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic

V14803 - V14822



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555408-SL **Lot No.:** A0220471
Description : Custom Vinyl Acetate Standard
Custom Vinyl Acetate Standard 8,000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : June 30, 2026 **Storage:** -20°C or colder
Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Vinyl acetate	108-05-4	RD240423RSR	99%	8,066.0 µg/mL	+/- 278.7979

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods. Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Ethan Winiarski - Operations Tech I

Date Mixed: 24-Dec-2024

Balance Serial # 1127510105

Dillan Murphy - Operations Technician I

Date Passed: 02-Jan-2025

REVIEWED
By Jennifer Pollock at 7:12 am, Jan 05, 2025

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis

Rev 05/09/25
Avantor™



V 14921 to
V 14938

Material No.: 9077-02
Batch No.: 24G0262002
Manufactured Date: 2024-05-14
Expiration Date: 2027-05-14
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.3 ppm
Titration Acid (μeq/g)	≤ 0.3	0.3
Titration Base (μeq/g)	≤ 0.10	0.03
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis – Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Jamie Croak
Director Quality Operations, Bioscience Production

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis

Rev 05/09/25
Avantor™



V 14921 to
V 14938

Material No.: 9077-02
Batch No.: 24G0262002
Manufactured Date: 2024-05-14
Expiration Date: 2027-05-14
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.3 ppm
Titration Acid (μeq/g)	≤ 0.3	0.3
Titration Base (μeq/g)	≤ 0.10	0.03
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis – Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Jamie Croak
Director Quality Operations, Bioscience Production



Certified Reference Material CRM

see 05/21/25



CERTIFIED WEIGHT REPORT

Part Number: **91980**
Lot Number: **051925**
Description: **Acrolein**

Solvent(s):
Water Lot# 041725Q

5.019

114944-114948

		051925
Formulated By:	Lawrence Barry	DATE
		051925
Reviewed By:	Pedro L. Rentas	DATE



Expiration Date: 061925
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 5000
NIST Test ID#: 6UTB

5E-05 Balance Uncertainty
0.001 Flask Uncertainty

Weight(s) shown below were combined and diluted to (mL): 10.0

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05170	5004.1	52.5	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.) Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately. Long term storage is not recommended. Please contact our technical department if further information is required.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
 - Standards are prepared gravimetrically using balances that are calibrated by an ISO 17025 certified organization with weights traceable through NIST to the SI kilogram (see above).
 - Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
 - All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
 - Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).
- Rev 1.0, 2/25/2025



Certified Reference Material CRM

see 05/21/25



CERTIFIED WEIGHT REPORT

Part Number: **91980**
Lot Number: **051925**
Description: **Acrolein**

Solvent(s):
Water

Lot#
041725Q

5.019

114944-114948

		051925
Formulated By:	Lawrence Barry	DATE
		051925
Reviewed By:	Pedro L. Rentas	DATE



Expiration Date: **061925**
Recommended Storage: **Refrigerate (4 °C)**
Nominal Concentration (µg/mL): **5000**
NIST Test ID#: **6UTB**

5E-05 Balance Uncertainty
0.001 Flask Uncertainty

Weight(s) shown below were combined and diluted to (mL): **10.0**

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05170	5004.1	52.5	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.) Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately. Long term storage is not recommended. Please contact our technical department if further information is required.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
 - Standards are prepared gravimetrically using balances that are calibrated by an ISO 17025 certified organization with weights traceable through NIST to the SI kilogram (see above).
 - Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
 - All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
 - Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).
- Rev 1.0, 2/25/2025



Certified Reference Material CRM

see 05/21/25



CERTIFIED WEIGHT REPORT

Part Number: **91980**
Lot Number: **051925**
Description: **Acrolein**

Solvent(s):
Water Lot# 041725Q

5.019

114944-114948

		051925
Formulated By:	Lawrence Barry	DATE
		051925
Reviewed By:	Pedro L. Rentas	DATE



Expiration Date: 061925
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 5000
NIST Test ID#: 6UTB

5E-05 Balance Uncertainty
0.001 Flask Uncertainty

Weight(s) shown below were combined and diluted to (mL): 10.0

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05170	5004.1	52.5	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.) Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately. Long term storage is not recommended. Please contact our technical department if further information is required.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
 - Standards are prepared gravimetrically using balances that are calibrated by an ISO 17025 certified organization with weights traceable through NIST to the SI kilogram (see above).
 - Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
 - All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
 - Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).
- Rev 1.0, 2/25/2025



Certified Reference Material CRM

See 05/21/25



CERTIFIED WEIGHT REPORT

Part Number: **91980**
Lot Number: **051925**
Description: **Acrolein**

Solvent(s):
Water Lot# 041725Q

5.019

114944-114948

		051925
Formulated By:	Lawrence Barry	DATE
		051925
Reviewed By:	Pedro L. Rentas	DATE



Expiration Date: 061925
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 5000
NIST Test ID#: 6UTB

5E-05 Balance Uncertainty
0.001 Flask Uncertainty

Weight(s) shown below were combined and diluted to (mL): 10.0

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05170	5004.1	52.5	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.) Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately. Long term storage is not recommended. Please contact our technical department if further information is required.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
 - Standards are prepared gravimetrically using balances that are calibrated by an ISO 17025 certified organization with weights traceable through NIST to the SI kilogram (see above).
 - Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
 - All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
 - Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).
- Rev 1.0, 2/25/2025



SHIPPING DOCUMENTS

CLIENT INFORMATION

REPORT TO BE SENT TO:

COMPANY: Resolution Consultants
ADDRESS:
CITY STATE ZIP:
ATTENTION: Eleanor V.
PHONE: FAX:

CLIENT PROJECT INFORMATION

PROJECT NAME: NWIRP Bethpage
PROJECT NO.: 6073/872 LOCATION: Bethpage
PROJECT MANAGER: Eleanor Vivian
e-mail: Eleanor.Vivian@alliance.com
PHONE: FAX:

CLIENT BILLING INFORMATION

BILL TO: Eleanor V. PO#:
ADDRESS: 27 Ellis place
CITY Ossining STATE: NY ZIP: 10562
ATTENTION: PHONE:

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) DAYS*
HARDCOPY (DATA PACKAGE): Standard DAYS*
EDD: Standard DAYS*
*TO BE APPROVED BY CHEMTECH
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☐ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B
+ Raw Data ☐ Other
☐ EDD FORMAT

VOCs 8260
1,4 dioxane 8270511

PRESERVATIVES

COMMENTS

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES										← Specify Preservatives A-HCl D-NaOH B-HNO3 E-ICE C-H2SO4 F-OTHER
			COMP	GRAB	DATE	TIME		A	F								
1.	TT20551-20250617	GW		X	6/17	1240	3	2	1								
2.																	
3.																	
4.																	
5.																	
6.																	
7.																	
8.																	
9.																	
10.																	

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER: 1. <u>Memo Adams</u>	DATE/TIME: <u>6-18-25</u>	RECEIVED BY: <u>[Signature]</u> <u>6-18-25</u>	Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP <u>2.30</u> °C
RELINQUISHED BY SAMPLER: 2.	DATE/TIME:	RECEIVED BY: <u>[Signature]</u>	Comments:
RELINQUISHED BY SAMPLER: 3. <u>[Signature]</u>	DATE/TIME: <u>6-18-25</u>	RECEIVED BY: 3.	Page ____ of

CLIENT: ☐ Hand Delivered ☐ Other

Shipment Complete

☐ YES ☐ NO

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q2361 AECO15

Order Date : 6/18/2025 3:09:00 PM

Project Mgr : NYS ASP B

Client Name : AECOM Technical Services

Project Name : NAVFAC NWIRP Bethpage

Report Type : Results Only ~~Level 1~~ 2/1

Client Contact : Eleanor Vivadou

Receive DateTime : 6/18/2025 12:00:00 PM 16:30 2/1

EDD Type : EQUIS

Invoice Name : AECOM Technical Services

Purchase Order :

Hard Copy Date :

Invoice Contact : Eleanor Vivadou

Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q2361-01	TT205S1-20250617	Water	06/17/2025	12:40	VOCMS Group1		8260-Low	10 Bus. Days	

Relinquished By :

Date / Time : 6/19/25 0730

Received By :

Date / Time : 06/19/25 7:30 Pg # 4

Storage Area : VOA Refridgerator Room

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
 Data File : VX046718.D
 Acq On : 17 Jun 2025 11:19
 Operator : JC/MD
 Sample : VSTDICC005
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC005

Quant Time: Jun 18 02:37:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 02:34:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	130530	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	217019	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	195782	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	100990	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	10456	5.791	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	11.580%#
35) Dibromofluoromethane	5.403	113	7865	5.478	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	10.960%#
50) Toluene-d8	8.653	98	29120	5.645	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	11.280%#
62) 4-Bromofluorobenzene	11.079	95	11137	5.791	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	11.580%#
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.179	85	6051	4.342	ug/l	99
3) Chloromethane	1.307	50	7066	4.698	ug/l	90
4) Vinyl Chloride	1.386	62	7425	4.629	ug/l	99
5) Bromomethane	1.617	94	4837	5.360	ug/l	95
6) Chloroethane	1.703	64	4575	4.706	ug/l	92
7) Trichlorofluoromethane	1.904	101	11714	4.810	ug/l	92
8) Diethyl Ether	2.148	74	4038	4.831	ug/l	98
9) 1,1,2-Trichlorotrifluo...	2.349	101	7353	4.921	ug/l	96
10) Methyl Iodide	2.465	142	5830	7.181	ug/l	99
11) Tert butyl alcohol	2.959	59	3134	19.340	ug/l	99
12) 1,1-Dichloroethene	2.337	96	6884	4.775	ug/l	91
13) Acrolein	2.251	56	6208	26.301	ug/l	100
14) Allyl chloride	2.678	41	12195	4.668	ug/l	99
15) Acrylonitrile	3.081	53	17611	24.169	ug/l	99
16) Acetone	2.392	43	14275	23.710	ug/l	98
17) Carbon Disulfide	2.526	76	19624	4.507	ug/l	100
18) Methyl Acetate	2.721	43	7530	4.924	ug/l	100
19) Methyl tert-butyl Ether	3.123	73	21254	4.734	ug/l	97
20) Methylene Chloride	2.800	84	8556	5.110	ug/l	97
21) trans-1,2-Dichloroethene	3.111	96	7425	4.813	ug/l	97
22) Diisopropyl ether	3.769	45	24580	4.895	ug/l #	67
23) Vinyl Acetate	3.739	43	92889	23.264	ug/l	98
24) 1,1-Dichloroethane	3.623	63	13761	4.827	ug/l	99
25) 2-Butanone	4.580	43	20821	24.452	ug/l	97
26) 2,2-Dichloropropane	4.483	77	10020	4.819	ug/l	96
27) cis-1,2-Dichloroethene	4.507	96	8891	4.906	ug/l	99
28) Bromochloromethane	4.910	49	6884	5.326	ug/l #	96
29) Tetrahydrofuran	5.025	42	13402	24.367	ug/l	99
30) Chloroform	5.099	83	14385	5.046	ug/l	86
31) Cyclohexane	5.483	56	12835	4.748	ug/l	94
32) 1,1,1-Trichloroethane	5.397	97	12150	4.859	ug/l	97
36) 1,1-Dichloropropene	5.702	75	10359	4.998	ug/l	99
37) Ethyl Acetate	4.733	43	8763	4.580	ug/l	96
38) Carbon Tetrachloride	5.690	117	11043	4.909	ug/l	98
39) Methylcyclohexane	7.391	83	13704	5.024	ug/l #	87
40) Benzene	6.050	78	31061	5.064	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
 Data File : VX046718.D
 Acq On : 17 Jun 2025 11:19
 Operator : JC/MD
 Sample : VSTDICC005
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC005

Quant Time: Jun 18 02:37:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 02:34:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.940	41	4488	4.400	ug/l #	92
42) 1,2-Dichloroethane	6.104	62	11029	5.116	ug/l	98
43) Isopropyl Acetate	6.348	43	14711	4.914	ug/l	97
44) Trichloroethene	7.135	130	7703	4.927	ug/l	93
45) 1,2-Dichloropropane	7.433	63	7529	4.952	ug/l	97
46) Dibromomethane	7.592	93	5267	4.821	ug/l	95
47) Bromodichloromethane	7.824	83	11072	4.966	ug/l	96
48) Methyl methacrylate	7.702	41	6864	4.571	ug/l	99
49) 1,4-Dioxane	7.671	88	1949	113.091	ug/l	95
51) 4-Methyl-2-Pentanone	8.574	43	44282	24.800	ug/l	97
52) Toluene	8.720	92	19190	5.046	ug/l	99
53) t-1,3-Dichloropropene	8.982	75	9497	4.591	ug/l	99
54) cis-1,3-Dichloropropene	8.372	75	11188	4.764	ug/l	95
55) 1,1,2-Trichloroethane	9.153	97	7151	5.046	ug/l	95
56) Ethyl methacrylate	9.122	69	10101	4.828	ug/l	99
57) 1,3-Dichloropropane	9.311	76	12483	5.115	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.244	63	26158	24.071	ug/l	100
59) 2-Hexanone	9.433	43	30938	25.062	ug/l	99
60) Dibromochloromethane	9.518	129	8249	4.939	ug/l	99
61) 1,2-Dibromoethane	9.610	107	7221	5.009	ug/l	98
64) Tetrachloroethene	9.275	164	6755	4.980	ug/l	97
65) Chlorobenzene	10.079	112	21807	5.046	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.165	131	7006	4.823	ug/l	98
67) Ethyl Benzene	10.195	91	37295	4.939	ug/l	98
68) m/p-Xylenes	10.299	106	27609	9.812	ug/l	96
69) o-Xylene	10.640	106	13421	5.090	ug/l	97
70) Styrene	10.659	104	22403	4.855	ug/l	100
71) Bromoform	10.799	173	5248	4.744	ug/l #	100
73) Isopropylbenzene	10.963	105	36281	4.878	ug/l	99
74) N-amyl acetate	10.841	43	12496	4.434	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.213	83	10468	5.040	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	10741	5.236	ug/l	90
77) Bromobenzene	11.201	156	8694	4.857	ug/l	95
78) n-propylbenzene	11.305	91	42848	4.850	ug/l	100
79) 2-Chlorotoluene	11.366	91	26221	4.971	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	29167	4.772	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	2663	4.070	ug/l	94
82) 4-Chlorotoluene	11.457	91	30097	4.889	ug/l	99
83) tert-Butylbenzene	11.713	119	30728	4.864	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	30213	4.920	ug/l	98
85) sec-Butylbenzene	11.890	105	39588	4.958	ug/l	100
86) p-Isopropyltoluene	12.006	119	33076	4.909	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	17198	4.928	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	17600	4.910	ug/l	95
89) n-Butylbenzene	12.329	91	29990	4.688	ug/l	98
90) Hexachloroethane	12.536	117	5285	4.480	ug/l	97
91) 1,2-Dichlorobenzene	12.335	146	16203	4.968	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	1982	4.834	ug/l	94
93) 1,2,4-Trichlorobenzene	13.585	180	10507	4.532	ug/l	99
94) Hexachlorobutadiene	13.725	225	4953	5.078	ug/l	95
95) Naphthalene	13.774	128	29158	4.414	ug/l	99
96) 1,2,3-Trichlorobenzene	13.963	180	10725	4.835	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
Data File : VX046718.D
Acq On : 17 Jun 2025 11:19
Operator : JC/MD
Sample : VSTDICC005
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1


Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005


Quant Time: Jun 18 02:37:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 02:34:13 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Quant Time: Jun 18 02:37:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 02:34:13 2025
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.562 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005







Tgt Ion:168 Resp: 130530
Ion Ratio Lower Upper
168 100
99 60.0 48.5 72.7



#2
Dichlorodifluoromethane
Concen: 4.342 ug/l
RT: 1.179 min Scan# 15
Delta R.T. -0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Tgt Ion: 85 Resp: 6051
Ion Ratio Lower Upper
85 100
87 32.7 16.6 49.8









#3
Chloromethane
Concen: 4.698 ug/l
RT: 1.307 min Scan# 36
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005


Tgt Ion: 50 Resp: 7066
Ion Ratio Lower Upper
50 100
52 36.4 24.9 37.3



#4
Vinyl Chloride
Concen: 4.629 ug/l
RT: 1.386 min Scan# 49
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Tgt Ion: 62 Resp: 7425
Ion Ratio Lower Upper
62 100
64 29.3 23.9 35.9









#5
Bromomethane
Concen: 5.360 ug/l
RT: 1.617 min Scan# 81
Delta R.T. -0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion: 94 Resp: 4837
Ion Ratio Lower Upper
94 100
96 95.3 72.3 108.5



#6
Chloroethane
Concen: 4.706 ug/l
RT: 1.703 min Scan# 101
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Tgt Ion: 64 Resp: 4575
Ion Ratio Lower Upper
64 100
66 29.4 27.1 40.7









#7
 Trichlorofluoromethane
 Concen: 4.810 ug/l
 RT: 1.904 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: VX046718.D
 Acq: 17 Jun 2025 11:19




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC005


Tgt Ion:101 Resp: 11714
 Ion Ratio Lower Upper
 101 100
 103 70.9 51.8 77.8



#8
 Diethyl Ether
 Concen: 4.831 ug/l
 RT: 2.148 min Scan# 174
 Delta R.T. -0.000 min
 Lab File: VX046718.D
 Acq: 17 Jun 2025 11:19

Tgt Ion: 74 Resp: 4038
 Ion Ratio Lower Upper
 74 100
 45 106.5 52.3 157.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 4.921 ug/l
RT: 2.349 min Scan# 206
Delta R.T. 0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion:101 Resp: 7353
Ion Ratio Lower Upper
101 100
85 43.2 35.4 53.0
151 71.5 61.2 91.8



#10
Methyl Iodide
Concen: 7.181 ug/l
RT: 2.465 min Scan# 226
Delta R.T. -0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Tgt Ion:142 Resp: 5830
Ion Ratio Lower Upper
142 100
127 45.5 37.0 55.6
141 14.6 11.5 17.3









#11
Tert butyl alcohol
Concen: 19.340 ug/l
RT: 2.959 min Scan# 307
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion: 59 Resp: 3134
Ion Ratio Lower Upper
59 100
57 10.9 8.3 12.5



#12
1,1-Dichloroethene
Concen: 4.775 ug/l
RT: 2.337 min Scan# 205
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Tgt Ion: 96 Resp: 6884
Ion Ratio Lower Upper
96 100
61 171.4 126.6 189.8
98 68.2 50.7 76.1





#13

Acrolein

Concen: 26.301 ug/l

RT: 2.251 min Scan# 190

Delta R.T. 0.006 min

Lab File: VX046718.D

Acq: 17 Jun 2025 11:19

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC005

Tgt Ion: 56 Resp: 6208

Ion Ratio Lower Upper

56 100

55 68.8 55.2 82.8



#14

Allyl chloride

Concen: 4.668 ug/l

RT: 2.678 min Scan# 261

Delta R.T. -0.000 min

Lab File: VX046718.D

Acq: 17 Jun 2025 11:19




Tgt Ion: 41 Resp: 12195


Ion Ratio Lower Upper

41 100

39 68.1 53.6 80.4

76 33.8 26.8 40.2









#15
Acrylonitrile
Concen: 24.169 ug/l
RT: 3.081 min Scan# 31
Delta R.T. 0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005


Tgt Ion: 53 Resp: 17611
Ion Ratio Lower Upper
53 100
52 81.8 65.1 97.7
51 35.6 28.9 43.3



#16
Acetone
Concen: 23.710 ug/l
RT: 2.392 min Scan# 214
Delta R.T. 0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Tgt Ion: 43 Resp: 14275
Ion Ratio Lower Upper
43 100
58 31.7 24.5 36.7





#17

Carbon Disulfide

Concen: 4.507 ug/l

RT: 2.526 min Scan# 21

Delta R.T. -0.000 min

Lab File: VX046718.D

Acq: 17 Jun 2025 11:19

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC005

Tgt Ion: 76 Resp: 19624

Ion Ratio Lower Upper

76 100

78 9.3 7.5 11.3



#18

Methyl Acetate

Concen: 4.924 ug/l

RT: 2.721 min Scan# 268

Delta R.T. 0.006 min

Lab File: VX046718.D




Acq: 17 Jun 2025 11:19


Tgt Ion: 43 Resp: 7530

Ion Ratio Lower Upper

43 100

74 22.8 18.4 27.6









#19
Methyl tert-butyl Ether
Concen: 4.734 ug/l
RT: 3.123 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion: 73 Resp: 21254
Ion Ratio Lower Upper
73 100
57 24.0 18.2 27.2



#20
Methylene Chloride
Concen: 5.110 ug/l
RT: 2.800 min Scan# 281
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Tgt Ion: 84 Resp: 8556
Ion Ratio Lower Upper
84 100
49 130.6 101.3 151.9
51 37.6 31.0 46.6
86 65.2 51.1 76.7









#21
trans-1,2-Dichloroethene
Concen: 4.813 ug/l
RT: 3.111 min Scan# 311
Delta R.T. 0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion: 96 Resp: 7425
Ion Ratio Lower Upper
96 100
61 144.2 112.5 168.7
98 68.5 51.9 77.9



#22
Diisopropyl ether
Concen: 4.895 ug/l
RT: 3.769 min Scan# 440
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Tgt Ion: 45 Resp: 24580
Ion Ratio Lower Upper
45 100
43 160.9 91.9 137.9#
87 25.1 22.0 33.0
59 11.3 9.1 13.7









#23
 Vinyl Acetate
 Concen: 23.264 ug/l
 RT: 3.739 min Scan# 41
 Delta R.T. 0.006 min
 Lab File: VX046718.D
 Acq: 17 Jun 2025 11:19




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC005


Tgt Ion: 43 Resp: 92889
 Ion Ratio Lower Upper
 43 100
 86 9.5 8.2 12.2



#24
 1,1-Dichloroethane
 Concen: 4.827 ug/l
 RT: 3.623 min Scan# 416
 Delta R.T. -0.000 min
 Lab File: VX046718.D
 Acq: 17 Jun 2025 11:19

Tgt Ion: 63 Resp: 13761
 Ion Ratio Lower Upper
 63 100
 98 8.5 4.0 11.9
 100 4.2 2.2 6.6









#25
2-Butanone
Concen: 24.452 ug/l
RT: 4.580 min Scan# 51
Delta R.T. 0.018 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion: 43 Resp: 20821
Ion Ratio Lower Upper
43 100
72 26.8 20.2 30.4



#26
2,2-Dichloropropane
Concen: 4.819 ug/l
RT: 4.483 min Scan# 557
Delta R.T. -0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Tgt Ion: 77 Resp: 10020
Ion Ratio Lower Upper
77 100
97 21.7 11.8 35.3





#27

cis-1,2-Dichloroethene

Concen: 4.906 ug/l

RT: 4.507 min Scan# 560

Delta R.T. 0.006 min

Lab File: VX046718.D

Acq: 17 Jun 2025 11:19

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion: 96 Resp: 8891

Ion Ratio Lower Upper

96 100

61 146.2 0.0 291.0

98 64.6 0.0 126.8



#28

Bromochloromethane

Concen: 5.326 ug/l

RT: 4.910 min Scan# 627

Delta R.T. -0.000 min

Lab File: VX046718.D

Acq: 17 Jun 2025 11:19




Tgt Ion: 49 Resp: 6884


Ion Ratio Lower Upper

49 100

129 0.0 0.0 3.8

130 73.3 61.4 92.0





#29

Tetrahydrofuran

Concen: 24.367 ug/l

RT: 5.025 min Scan# 64

Delta R.T. 0.012 min

Lab File: VX046718.D

Acq: 17 Jun 2025 11:19

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC005





Tgt Ion: 42 Resp: 13402

Ion Ratio Lower Upper

42 100

72 44.7 34.9 52.3

71 40.5 32.2 48.2



#30

Chloroform

Concen: 5.046 ug/l

RT: 5.099 min Scan# 658

Delta R.T. -0.006 min

Lab File: VX046718.D




Acq: 17 Jun 2025 11:19


Tgt Ion: 83 Resp: 14385

Ion Ratio Lower Upper

83 100

85 56.2 53.7 80.5









#31
Cyclohexane
Concen: 4.748 ug/l
RT: 5.483 min Scan# 721
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion:	56	Resp:	12835
Ion	Ratio	Lower	Upper
56	100		
69	28.2	24.8	37.2
84	79.6	68.2	102.2



#32
1,1,1-Trichloroethane
Concen: 4.859 ug/l
RT: 5.397 min Scan# 707
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Tgt Ion:	97	Resp:	12150
Ion	Ratio	Lower	Upper
97	100		
99	65.5	51.3	76.9
61	47.8	36.4	54.6









#33
 1,2-Dichloroethane-d4
 Concen: 5.791 ug/l
 RT: 5.964 min Scan# 800
 Delta R.T. -0.000 min
 Lab File: VX046718.D
 Acq: 17 Jun 2025 11:19




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC005


Tgt Ion: 65 Resp: 10456
 Ion Ratio Lower Upper
 65 100
 67 50.9 0.0 105.4



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.769 min Scan# 932
 Delta R.T. -0.000 min
 Lab File: VX046718.D
 Acq: 17 Jun 2025 11:19

Tgt Ion: 114 Resp: 217019
 Ion Ratio Lower Upper
 114 100
 63 21.8 0.0 44.2
 88 16.4 0.0 33.2









#35
Dibromofluoromethane
Concen: 5.478 ug/l
RT: 5.403 min Scan# 707
Delta R.T. 0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion:113 Resp: 7865
Ion Ratio Lower Upper
113 100
111 104.7 82.0 123.0
192 16.4 15.3 22.9



#36
1,1-Dichloropropene
Concen: 4.998 ug/l
RT: 5.702 min Scan# 757
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Tgt Ion: 75 Resp: 10359
Ion Ratio Lower Upper
75 100
110 34.8 17.2 51.5
77 30.4 24.2 36.4





#37

Ethyl Acetate

Concen: 4.580 ug/l

RT: 4.733 min Scan# 596

Delta R.T. 0.012 min

Lab File: VX046718.D

Acq: 17 Jun 2025 11:19

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion: 43 Resp: 8763

Ion Ratio Lower Upper

43 100

61 10.1 9.7 14.5

70 8.9 8.1 12.1



#38

Carbon Tetrachloride

Concen: 4.909 ug/l

RT: 5.690 min Scan# 755

Delta R.T. -0.000 min

Lab File: VX046718.D

Acq: 17 Jun 2025 11:19




Tgt Ion: 117 Resp: 11043


Ion Ratio Lower Upper

117 100

119 95.6 75.2 112.8

121 31.6 24.1 36.1









#39
Methylcyclohexane
Concen: 5.024 ug/l
RT: 7.391 min Scan# 1034
Delta R.T. 0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion: 83 Resp: 13704
Ion Ratio Lower Upper
83 100
55 69.0 63.0 94.4
98 38.4 38.8 58.2#



#40
Benzene
Concen: 5.064 ug/l
RT: 6.050 min Scan# 814
Delta R.T. 0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Tgt Ion: 78 Resp: 31061
Ion Ratio Lower Upper
78 100
77 25.1 18.8 28.2









#41
Methacrylonitrile
Concen: 4.400 ug/l
RT: 4.940 min Scan# 61
Delta R.T. 0.012 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion: 41 Resp: 4488
Ion Ratio Lower Upper
41 100
39 61.2 45.9 68.9
67 74.8 61.2 91.8
52 17.5 25.5 38.3#



#42
1,2-Dichloroethane
Concen: 5.116 ug/l
RT: 6.104 min Scan# 823
Delta R.T. 0.012 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Tgt Ion: 62 Resp: 11029
Ion Ratio Lower Upper
62 100
98 8.1 0.0 17.8









#43
Isopropyl Acetate
Concen: 4.914 ug/l
RT: 6.348 min Scan# 863
Delta R.T. 0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion: 43 Resp: 14711
Ion Ratio Lower Upper
43 100
61 17.3 15.3 22.9
87 11.6 10.0 15.0



#44
Trichloroethene
Concen: 4.927 ug/l
RT: 7.135 min Scan# 992
Delta R.T. 0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Tgt Ion:130 Resp: 7703
Ion Ratio Lower Upper
130 100
95 94.6 0.0 202.6





#45

1,2-Dichloropropane

Concen: 4.952 ug/l

RT: 7.433 min Scan# 1041

Delta R.T. -0.000 min

Lab File: VX046718.D

Acq: 17 Jun 2025 11:19

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC005

Tgt Ion: 63 Resp: 7529

Ion Ratio Lower Upper

63 100

65 30.2 22.9 34.3



#46

Dibromomethane

Concen: 4.821 ug/l

RT: 7.592 min Scan# 1067

Delta R.T. 0.006 min

Lab File: VX046718.D

Acq: 17 Jun 2025 11:19




Tgt Ion: 93 Resp: 5267


Ion Ratio Lower Upper

93 100

95 86.9 66.2 99.4

174 102.0 77.0 115.6









#47
Bromodichloromethane
Concen: 4.966 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion:	83	Resp:	11072
Ion	Ratio	Lower	Upper
83	100		
85	61.1	51.4	77.2
127	7.9	6.9	10.3



#48
Methyl methacrylate
Concen: 4.571 ug/l
RT: 7.702 min Scan# 1085
Delta R.T. 0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Tgt Ion:	41	Resp:	6864
Ion	Ratio	Lower	Upper
41	100		
69	79.8	63.4	95.2
39	58.5	45.4	68.0





#49

1,4-Dioxane

Concen: 113.091 ug/l

RT: 7.671 min Scan# 1080

Delta R.T. 0.012 min

Lab File: VX046718.D

Acq: 17 Jun 2025 11:19

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion: 88 Resp: 1949

Ion Ratio Lower Upper

88 100

43 44.9 31.2 46.8

58 70.8 54.7 82.1



#50

Toluene-d8

Concen: 5.645 ug/l

RT: 8.653 min Scan# 1241

Delta R.T. 0.006 min

Lab File: VX046718.D




Acq: 17 Jun 2025 11:19


Tgt Ion: 98 Resp: 29120

Ion Ratio Lower Upper

98 100

100 68.4 53.0 79.4









#51
4-Methyl-2-Pentanone
Concen: 24.800 ug/l
RT: 8.574 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion: 43 Resp: 44282
Ion Ratio Lower Upper
43 100
58 36.8 30.8 46.2



#52
Toluene
Concen: 5.046 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Tgt Ion: 92 Resp: 19190
Ion Ratio Lower Upper
92 100
91 170.2 134.8 202.2





#53

t-1,3-Dichloropropene

Concen: 4.591 ug/l

RT: 8.982 min Scan# 1195

Delta R.T. 0.006 min

Lab File: VX046718.D

Acq: 17 Jun 2025 11:19

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC005

Tgt Ion: 75 Resp: 9497

Ion Ratio Lower Upper

75 100

77 32.7 25.6 38.4



#54

cis-1,3-Dichloropropene

Concen: 4.764 ug/l

RT: 8.372 min Scan# 1195

Delta R.T. 0.006 min

Lab File: VX046718.D

Acq: 17 Jun 2025 11:19




Tgt Ion: 75 Resp: 11188


Ion Ratio Lower Upper

75 100

77 31.7 25.3 37.9

39 54.9 39.2 58.8









#55
1,1,2-Trichloroethane
Concen: 5.046 ug/l
RT: 9.153 min Scan# 1318
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005


Tgt Ion: 97	Resp: 7151
Ion Ratio	Lower Upper
97 100	
83 90.8	67.2 100.8
85 55.8	44.1 66.1
99 58.5	50.3 75.5



#56
Ethyl methacrylate
Concen: 4.828 ug/l
RT: 9.122 min Scan# 1318
Delta R.T. 0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Tgt Ion: 69	Resp: 10101
Ion Ratio	Lower Upper
69 100	
41 68.6	55.4 83.2
39 36.5	29.7 44.5





#57

1,3-Dichloropropane

Concen: 5.115 ug/l

RT: 9.311 min Scan# 11349

Delta R.T. 0.006 min

Lab File: VX046718.D

Acq: 17 Jun 2025 11:19

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC005

Tgt Ion: 76 Resp: 12483

Ion Ratio Lower Upper

76 100

78 31.7 25.7 38.5



#58

2-Chloroethyl Vinyl ether

Concen: 24.071 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. -0.000 min

Lab File: VX046718.D




Acq: 17 Jun 2025 11:19


Tgt Ion: 63 Resp: 26158

Ion Ratio Lower Upper

63 100

106 28.2 22.6 34.0









#59
2-Hexanone
Concen: 25.062 ug/l
RT: 9.433 min Scan# 1369
Delta R.T. 0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005


Tgt Ion: 43 Resp: 30938
Ion Ratio Lower Upper
43 100
58 52.6 26.8 80.4



#60
Dibromochloromethane
Concen: 4.939 ug/l
RT: 9.518 min Scan# 1383
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Tgt Ion: 129 Resp: 8249
Ion Ratio Lower Upper
129 100
127 76.2 38.4 115.2









#61
1,2-Dibromoethane
Concen: 5.009 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion:107 Resp: 7221
Ion Ratio Lower Upper
107 100
109 92.8 75.8 113.6



#62
4-Bromofluorobenzene
Concen: 5.791 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Tgt Ion: 95 Resp: 11137
Ion Ratio Lower Upper
95 100
174 77.4 0.0 150.4
176 73.9 0.0 145.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion:117 Resp: 195782
Ion Ratio Lower Upper
117 100
82 55.2 44.6 66.8
119 32.2 25.8 38.8



#64
Tetrachloroethene
Concen: 4.980 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Tgt Ion:164 Resp: 6755
Ion Ratio Lower Upper
164 100
166 125.4 104.6 157.0
129 94.1 76.3 114.5
131 89.4 74.1 111.1









#65
Chlorobenzene
Concen: 5.046 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion:112 Resp: 21807
Ion Ratio Lower Upper
112 100
114 31.6 25.5 38.3



#66
1,1,1,2-Tetrachloroethane
Concen: 4.823 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. 0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Tgt Ion:131 Resp: 7006
Ion Ratio Lower Upper
131 100
133 94.8 47.3 141.8
119 68.5 32.1 96.3





#67

Ethyl Benzene

Concen: 4.939 ug/l

RT: 10.195 min Scan# 1493

Delta R.T. 0.006 min

Lab File: VX046718.D

Acq: 17 Jun 2025 11:19

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC005

Tgt Ion: 91 Resp: 37295

Ion Ratio Lower Upper

91 100

106 29.0 24.2 36.2



#68

m/p-Xylenes

Concen: 9.812 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. -0.000 min

Lab File: VX046718.D




Acq: 17 Jun 2025 11:19


Tgt Ion:106 Resp: 27609

Ion Ratio Lower Upper

106 100

91 212.5 164.8 247.2









#69
o-Xylene
Concen: 5.090 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion:106 Resp: 13421
Ion Ratio Lower Upper
106 100
91 218.1 111.3 333.9



#70
Styrene
Concen: 4.855 ug/l
RT: 10.659 min Scan# 1570
Delta R.T. 0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Tgt Ion:104 Resp: 22403
Ion Ratio Lower Upper
104 100
78 52.1 41.4 62.2
103 54.0 43.1 64.7









#71
Bromoform
Concen: 4.744 ug/l
RT: 10.799 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005


Tgt Ion:173 Resp: 5248
Ion Ratio Lower Upper
173 100
175 48.3 24.1 72.4
254 0.0 0.1 0.1#



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Tgt Ion:152 Resp: 100990
Ion Ratio Lower Upper
152 100
115 64.0 43.2 129.6
150 156.3 0.0 346.8









#73
Isopropylbenzene
Concen: 4.878 ug/l
RT: 10.963 min Scan# 1620
Delta R.T. 0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion:105 Resp: 36281
Ion Ratio Lower Upper
105 100
120 26.3 13.0 39.0



#74
N-ethyl acetate
Concen: 4.434 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Tgt Ion: 43 Resp: 12496
Ion Ratio Lower Upper
43 100
70 41.0 33.4 50.0
55 23.9 19.1 28.7
61 22.9 19.0 28.4





#75

1,1,2,2-Tetrachloroethane

Concen: 5.040 ug/l

RT: 11.213 min Scan# 1

Delta R.T. 0.006 min

Lab File: VX046718.D

Acq: 17 Jun 2025 11:19

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion: 83 Resp: 10468

Ion Ratio Lower Upper

83 100

131 9.0 4.9 14.6

85 64.4 32.0 96.0



#76

1,2,3-Trichloropropane

Concen: 5.236 ug/l

RT: 11.238 min Scan# 1665

Delta R.T. -0.000 min

Lab File: VX046718.D




Acq: 17 Jun 2025 11:19


Tgt Ion: 75 Resp: 10741

Ion Ratio Lower Upper

75 100

77 30.9 18.5 55.5





#77

Bromobenzene

Concen: 4.857 ug/l

RT: 11.201 min Scan# 1659

Delta R.T. 0.006 min

Lab File: VX046718.D

Acq: 17 Jun 2025 11:19

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion:156 Resp: 8694

Ion Ratio Lower Upper

156 100

77 173.3 82.8 248.6

158 99.4 47.6 142.9



#78

n-propylbenzene

Concen: 4.850 ug/l

RT: 11.305 min Scan# 1676

Delta R.T. 0.006 min

Lab File: VX046718.D




Acq: 17 Jun 2025 11:19


Tgt Ion: 91 Resp: 42848

Ion Ratio Lower Upper

91 100

120 22.7 11.4 34.1





#79

2-Chlorotoluene

Concen: 4.971 ug/l

RT: 11.366 min Scan# 1

Delta R.T. 0.006 min

Lab File: VX046718.D

Acq: 17 Jun 2025 11:19

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC005

Tgt Ion: 91 Resp: 26221

Ion Ratio Lower Upper

91 100

126 33.1 16.6 49.7



#80

1,3,5-Trimethylbenzene

Concen: 4.772 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. -0.000 min

Lab File: VX046718.D




Acq: 17 Jun 2025 11:19


Tgt Ion:105 Resp: 29167

Ion Ratio Lower Upper

105 100

120 47.7 24.1 72.2









#81
trans-1,4-Dichloro-2-butene
Concen: 4.070 ug/l
RT: 11.018 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion: 75 Resp: 2663
Ion Ratio Lower Upper
75 100
53 114.8 86.8 130.2
89 42.7 36.9 55.3



#82
4-Chlorotoluene
Concen: 4.889 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.006 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Tgt Ion: 91 Resp: 30097
Ion Ratio Lower Upper
91 100
126 28.4 14.4 43.2









#83
tert-Butylbenzene
Concen: 4.864 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion:119 Resp: 30728
Ion Ratio Lower Upper
119 100
91 60.3 29.5 88.6
134 23.0 11.7 35.0



#84
1,2,4-Trimethylbenzene
Concen: 4.920 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Tgt Ion:105 Resp: 30213
Ion Ratio Lower Upper
105 100
120 45.2 22.1 66.3









#85
 sec-Butylbenzene
 Concen: 4.958 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. -0.000 min
 Lab File: VX046718.D
 Acq: 17 Jun 2025 11:19




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC005


Tgt Ion:105 Resp: 39588
 Ion Ratio Lower Upper
 105 100
 134 19.6 9.8 29.4



#86
 p-Isopropyltoluene
 Concen: 4.909 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. -0.000 min
 Lab File: VX046718.D
 Acq: 17 Jun 2025 11:19

Tgt Ion:119 Resp: 33076
 Ion Ratio Lower Upper
 119 100
 134 26.2 12.9 38.7
 91 23.9 12.3 36.9





#87

1,3-Dichlorobenzene

Concen: 4.928 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX046718.D

Acq: 17 Jun 2025 11:19

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion:146 Resp: 17198

Ion Ratio Lower Upper

146 100

111 42.0 20.6 61.8

148 63.3 31.6 94.8



#88

1,4-Dichlorobenzene

Concen: 4.910 ug/l

RT: 12.036 min Scan# 1796

Delta R.T. -0.000 min

Lab File: VX046718.D

Acq: 17 Jun 2025 11:19




Tgt Ion:146 Resp: 17600


Ion Ratio Lower Upper

146 100

111 46.6 20.3 61.0

148 66.0 32.1 96.5









#89
n-Butylbenzene
Concen: 4.688 ug/l
RT: 12.329 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion: 91 Resp: 29990
Ion Ratio Lower Upper
91 100
92 55.6 27.1 81.3
134 26.3 12.9 38.6



#90
Hexachloroethane
Concen: 4.480 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. -0.000 min
Lab File: VX046718.D
Acq: 17 Jun 2025 11:19

Tgt Ion: 117 Resp: 5285
Ion Ratio Lower Upper
117 100
201 71.1 34.4 103.3









#91
 1,2-Dichlorobenzene
 Concen: 4.968 ug/l
 RT: 12.335 min Scan# 1845
 Delta R.T. 0.006 min
 Lab File: VX046718.D
 Acq: 17 Jun 2025 11:19




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC005


Tgt Ion:146 Resp: 16203
 Ion Ratio Lower Upper
 146 100
 111 44.1 21.7 65.1
 148 64.0 32.0 96.2



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 4.834 ug/l
 RT: 12.939 min Scan# 1944
 Delta R.T. -0.000 min
 Lab File: VX046718.D
 Acq: 17 Jun 2025 11:19

Tgt Ion: 75 Resp: 1982
 Ion Ratio Lower Upper
 75 100
 155 72.2 41.9 125.7
 157 102.4 51.4 154.3





#93

1,2,4-Trichlorobenzene

Concen: 4.532 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. -0.000 min

Lab File: VX046718.D

Acq: 17 Jun 2025 11:19

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion:180 Resp: 10507

Ion Ratio Lower Upper

180 100

182 94.5 46.7 140.1

145 31.8 15.8 47.4



#94

Hexachlorobutadiene

Concen: 5.078 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. 0.006 min

Lab File: VX046718.D

Acq: 17 Jun 2025 11:19




Tgt Ion:225 Resp: 4953


Ion Ratio Lower Upper

225 100

223 61.5 32.5 97.5

227 60.3 32.2 96.6





#95

Naphthalene

Concen: 4.414 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. -0.000 min

Lab File: VX046718.D

Acq: 17 Jun 2025 11:19

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion:128 Resp: 29158

Ion Ratio Lower Upper

128 100

127 12.6 10.1 15.1

129 10.3 8.7 13.1



#96

1,2,3-Trichlorobenzene

Concen: 4.835 ug/l

RT: 13.963 min Scan# 2112

Delta R.T. 0.006 min

Lab File: VX046718.D

Acq: 17 Jun 2025 11:19




Tgt Ion:180 Resp: 10725

Ion Ratio Lower Upper

180 100

182 92.3 47.8 143.4

145 32.0 17.3 51.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
 Data File : VX046719.D
 Acq On : 17 Jun 2025 13:59
 Operator : JC/MD
 Sample : VSTDIC020
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC020

Quant Time: Jun 18 02:38:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 02:34:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	134732	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	226567	50.000	ug/l	-0.01
63) Chlorobenzene-d5	10.049	117	197947	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	98285	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	30582	16.410	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	32.820%#
35) Dibromofluoromethane	5.385	113	24156	16.114	ug/l	-0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	32.220%#
50) Toluene-d8	8.647	98	88197	16.376	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	32.760%#
62) 4-Bromofluorobenzene	11.079	95	32067	15.973	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	31.940%#
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.185	85	27297	18.977	ug/l	97
3) Chloromethane	1.307	50	30731	19.793	ug/l	98
4) Vinyl Chloride	1.386	62	34049	20.565	ug/l	96
5) Bromomethane	1.618	94	19805	21.261	ug/l	91
6) Chloroethane	1.697	64	20911	20.838	ug/l	97
7) Trichlorofluoromethane	1.904	101	52480	20.875	ug/l	99
8) Diethyl Ether	2.142	74	18139	21.024	ug/l	100
9) 1,1,2-Trichlorotrifluo...	2.343	101	32778	21.253	ug/l	98
10) Methyl Iodide	2.465	142	30525	18.775	ug/l	99
11) Tert butyl alcohol	2.953	59	16360	97.808	ug/l	96
12) 1,1-Dichloroethene	2.331	96	30924	20.780	ug/l	92
13) Acrolein	2.246	56	20036	82.237	ug/l	98
14) Allyl chloride	2.672	41	56147	20.823	ug/l	99
15) Acrylonitrile	3.069	53	77278	102.746	ug/l	99
16) Acetone	2.386	43	59747	96.140	ug/l	97
17) Carbon Disulfide	2.526	76	88625	19.721	ug/l	99
18) Methyl Acetate	2.709	43	31812	20.154	ug/l	99
19) Methyl tert-butyl Ether	3.117	73	97187	20.974	ug/l	95
20) Methylene Chloride	2.800	84	35302	20.426	ug/l	99
21) trans-1,2-Dichloroethene	3.105	96	33796	21.224	ug/l	95
22) Diisopropyl ether	3.764	45	111712	21.555	ug/l	95
23) Vinyl Acetate	3.727	43	433313	105.140	ug/l	100
24) 1,1-Dichloroethane	3.617	63	62164	21.126	ug/l	98
25) 2-Butanone	4.556	43	87671	99.748	ug/l	99
26) 2,2-Dichloropropane	4.483	77	42992	20.032	ug/l	99
27) cis-1,2-Dichloroethene	4.495	96	39374	21.047	ug/l	99
28) Bromochloromethane	4.904	49	24738	18.544	ug/l	100
29) Tetrahydrofuran	5.007	42	56765	99.990	ug/l	98
30) Chloroform	5.093	83	64109	21.785	ug/l	100
31) Cyclohexane	5.477	56	58552	20.982	ug/l	98
32) 1,1,1-Trichloroethane	5.385	97	54542	21.133	ug/l	99
36) 1,1-Dichloropropene	5.696	75	44023	20.344	ug/l	99
37) Ethyl Acetate	4.715	43	39916	19.984	ug/l	99
38) Carbon Tetrachloride	5.684	117	48694	20.735	ug/l	97
39) Methylcyclohexane	7.379	83	58170	20.428	ug/l	97
40) Benzene	6.038	78	133872	20.904	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
 Data File : VX046719.D
 Acq On : 17 Jun 2025 13:59
 Operator : JC/MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC020

Quant Time: Jun 18 02:38:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 02:34:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	21970	20.632	ug/l	97
42) 1,2-Dichloroethane	6.086	62	47430	21.076	ug/l	99
43) Isopropyl Acetate	6.342	43	62432	19.975	ug/l	100
44) Trichloroethene	7.129	130	33927	20.788	ug/l	100
45) 1,2-Dichloropropane	7.428	63	33688	21.226	ug/l	94
46) Dibromomethane	7.580	93	23569	20.663	ug/l	99
47) Bromodichloromethane	7.818	83	49411	21.227	ug/l	98
48) Methyl methacrylate	7.690	41	31856	20.322	ug/l	98
49) 1,4-Dioxane	7.659	88	8340	382.035	ug/l	93
51) 4-Methyl-2-Pentanone	8.568	43	187509	100.586	ug/l	99
52) Toluene	8.714	92	84030	21.166	ug/l	100
53) t-1,3-Dichloropropene	8.976	75	43852	20.304	ug/l	99
54) cis-1,3-Dichloropropene	8.366	75	50326	20.526	ug/l	98
55) 1,1,2-Trichloroethane	9.147	97	30865	20.862	ug/l	94
56) Ethyl methacrylate	9.116	69	44702	20.468	ug/l	99
57) 1,3-Dichloropropane	9.305	76	52181	20.480	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.238	63	106505	93.878	ug/l	99
59) 2-Hexanone	9.427	43	129218	100.265	ug/l	99
60) Dibromochloromethane	9.519	129	36392	20.872	ug/l	99
61) 1,2-Dibromoethane	9.604	107	31507	20.936	ug/l	100
64) Tetrachloroethene	9.269	164	27988	20.408	ug/l	97
65) Chlorobenzene	10.073	112	90868	20.796	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.159	131	31508	21.455	ug/l	100
67) Ethyl Benzene	10.189	91	160761	21.056	ug/l	99
68) m/p-Xylenes	10.299	106	121014	42.539	ug/l	99
69) o-Xylene	10.640	106	57689	21.638	ug/l	100
70) Styrene	10.653	104	99462	21.318	ug/l	99
71) Bromoform	10.799	173	23372	20.898	ug/l #	100
73) Isopropylbenzene	10.957	105	153861	21.255	ug/l	99
74) N-amyl acetate	10.842	43	57185	20.851	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	42272	20.914	ug/l	100
76) 1,2,3-Trichloropropane	11.238	75	44152	22.117	ug/l	93
77) Bromobenzene	11.195	156	37464	21.507	ug/l	99
78) n-propylbenzene	11.299	91	184078	21.410	ug/l	100
79) 2-Chlorotoluene	11.360	91	107552	20.952	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	129635	21.793	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	11694	18.362	ug/l	98
82) 4-Chlorotoluene	11.451	91	127954	21.358	ug/l	100
83) tert-Butylbenzene	11.713	119	130359	21.201	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	128566	21.512	ug/l	99
85) sec-Butylbenzene	11.890	105	165047	21.239	ug/l	99
86) p-Isopropyltoluene	12.006	119	138623	21.142	ug/l	100
87) 1,3-Dichlorobenzene	11.969	146	70243	20.681	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	71953	20.625	ug/l	99
89) n-Butylbenzene	12.329	91	129101	20.738	ug/l	100
90) Hexachloroethane	12.536	117	23645	20.594	ug/l	97
91) 1,2-Dichlorobenzene	12.329	146	66883	21.073	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	8071	20.226	ug/l	98
93) 1,2,4-Trichlorobenzene	13.585	180	44726	19.821	ug/l	97
94) Hexachlorobutadiene	13.725	225	19698	20.752	ug/l	99
95) Naphthalene	13.774	128	129768	20.186	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	44380	20.558	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
Data File : VX046719.D
Acq On : 17 Jun 2025 13:59
Operator : JC/MD
Sample : VSTDICC020
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1


Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020


Quant Time: Jun 18 02:38:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 02:34:13 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jun 18 02:38:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 02:34:13 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion:168 Resp: 134732
Ion Ratio Lower Upper
168 100
99 58.4 48.5 72.7



#2
Dichlorodifluoromethane
Concen: 18.977 ug/l
RT: 1.185 min Scan# 16
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Tgt Ion: 85 Resp: 27297
Ion Ratio Lower Upper
85 100
87 31.6 16.6 49.8









#3
Chloromethane
Concen: 19.793 ug/l
RT: 1.307 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020


Tgt Ion: 50 Resp: 30731
Ion Ratio Lower Upper
50 100
52 32.1 24.9 37.3



#4
Vinyl Chloride
Concen: 20.565 ug/l
RT: 1.386 min Scan# 49
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Tgt Ion: 62 Resp: 34049
Ion Ratio Lower Upper
62 100
64 32.0 23.9 35.9









#5
Bromomethane
Concen: 21.261 ug/l
RT: 1.618 min Scan# 81
Delta R.T. -0.006 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion: 94 Resp: 19805
Ion Ratio Lower Upper
94 100
96 99.2 72.3 108.5



#6
Chloroethane
Concen: 20.838 ug/l
RT: 1.697 min Scan# 100
Delta R.T. -0.006 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Tgt Ion: 64 Resp: 20911
Ion Ratio Lower Upper
64 100
66 32.4 27.1 40.7









#7
Trichlorofluoromethane
Concen: 20.875 ug/l
RT: 1.904 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion:101 Resp: 52480
Ion Ratio Lower Upper
101 100
103 65.7 51.8 77.8



#8
Diethyl Ether
Concen: 21.024 ug/l
RT: 2.142 min Scan# 173
Delta R.T. -0.006 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Tgt Ion: 74 Resp: 18139
Ion Ratio Lower Upper
74 100
45 104.8 52.3 157.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 21.253 ug/l
RT: 2.343 min Scan# 206
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion:101 Resp: 32778
Ion Ratio Lower Upper
101 100
85 43.4 35.4 53.0
151 74.1 61.2 91.8



#10
Methyl Iodide
Concen: 18.775 ug/l
RT: 2.465 min Scan# 226
Delta R.T. -0.006 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Tgt Ion:142 Resp: 30525
Ion Ratio Lower Upper
142 100
127 46.7 37.0 55.6
141 13.7 11.5 17.3









#11
Tert butyl alcohol
Concen: 97.808 ug/l
RT: 2.953 min Scan# 307
Delta R.T. -0.006 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion: 59 Resp: 16360
Ion Ratio Lower Upper
59 100
57 8.9 8.3 12.5



#12
1,1-Dichloroethene
Concen: 20.780 ug/l
RT: 2.331 min Scan# 204
Delta R.T. -0.006 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Tgt Ion: 96 Resp: 30924
Ion Ratio Lower Upper
96 100
61 171.1 126.6 189.8
98 64.9 50.7 76.1





#13

Acrolein

Concen: 82.237 ug/l

RT: 2.246 min Scan# 190

Delta R.T. -0.000 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC020

Tgt Ion: 56 Resp: 20036

Ion Ratio Lower Upper

56 100

55 70.4 55.2 82.8



#14

Allyl chloride

Concen: 20.823 ug/l

RT: 2.672 min Scan# 260

Delta R.T. -0.006 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59




Tgt Ion: 41 Resp: 56147


Ion Ratio Lower Upper

41 100

39 66.0 53.6 80.4

76 33.0 26.8 40.2





#15

Acrylonitrile

Concen: 102.746 ug/l

RT: 3.069 min Scan# 31

Delta R.T. -0.006 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020





Tgt Ion: 53 Resp: 77278

Ion Ratio Lower Upper

53 100

52 82.5 65.1 97.7

51 35.1 28.9 43.3



#16

Acetone

Concen: 96.140 ug/l

RT: 2.386 min Scan# 213

Delta R.T. -0.000 min

Lab File: VX046719.D




Acq: 17 Jun 2025 13:59


Tgt Ion: 43 Resp: 59747

Ion Ratio Lower Upper

43 100

58 28.7 24.5 36.7









#17
Carbon Disulfide
Concen: 19.721 ug/l
RT: 2.526 min Scan# 21
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion: 76 Resp: 88625
Ion Ratio Lower Upper
76 100
78 8.9 7.5 11.3



#18
Methyl Acetate
Concen: 20.154 ug/l
RT: 2.709 min Scan# 266
Delta R.T. -0.006 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Tgt Ion: 43 Resp: 31812
Ion Ratio Lower Upper
43 100
74 22.7 18.4 27.6









#19
Methyl tert-butyl Ether
Concen: 20.974 ug/l
RT: 3.117 min Scan# 311
Delta R.T. -0.006 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion: 73 Resp: 97187
Ion Ratio Lower Upper
73 100
57 20.2 18.2 27.2



#20
Methylene Chloride
Concen: 20.426 ug/l
RT: 2.800 min Scan# 281
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Tgt Ion: 84 Resp: 35302
Ion Ratio Lower Upper
84 100
49 127.8 101.3 151.9
51 39.4 31.0 46.6
86 66.0 51.1 76.7









#21
trans-1,2-Dichloroethene
Concen: 21.224 ug/l
RT: 3.105 min Scan# 311
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion: 96 Resp: 33796
Ion Ratio Lower Upper
96 100
61 133.3 112.5 168.7
98 62.1 51.9 77.9



#22
Diisopropyl ether
Concen: 21.555 ug/l
RT: 3.764 min Scan# 439
Delta R.T. -0.006 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Tgt Ion: 45 Resp: 111712
Ion Ratio Lower Upper
45 100
43 120.9 91.9 137.9
87 25.2 22.0 33.0
59 11.1 9.1 13.7









#23
 Vinyl Acetate
 Concen: 105.140 ug/l
 RT: 3.727 min Scan# 41
 Delta R.T. -0.006 min
 Lab File: VX046719.D
 Acq: 17 Jun 2025 13:59




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC020


Tgt Ion: 43 Resp: 433313
 Ion Ratio Lower Upper
 43 100
 86 10.2 8.2 12.2



#24
 1,1-Dichloroethane
 Concen: 21.126 ug/l
 RT: 3.617 min Scan# 415
 Delta R.T. -0.006 min
 Lab File: VX046719.D
 Acq: 17 Jun 2025 13:59

Tgt Ion: 63 Resp: 62164
 Ion Ratio Lower Upper
 63 100
 98 7.1 4.0 11.9
 100 4.5 2.2 6.6





#25

2-Butanone

Concen: 99.748 ug/l

RT: 4.556 min Scan# 5

Delta R.T. -0.006 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC020

Tgt Ion: 43 Resp: 87671

Ion Ratio Lower Upper

43 100

72 24.9 20.2 30.4



#26

2,2-Dichloropropane

Concen: 20.032 ug/l

RT: 4.483 min Scan# 557

Delta R.T. -0.006 min

Lab File: VX046719.D




Acq: 17 Jun 2025 13:59


Tgt Ion: 77 Resp: 42992

Ion Ratio Lower Upper

77 100

97 22.9 11.8 35.3





#27

cis-1,2-Dichloroethene

Concen: 21.047 ug/l

RT: 4.495 min Scan# 51

Delta R.T. -0.006 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020





Tgt Ion: 96 Resp: 39374

Ion Ratio Lower Upper

96 100

61 144.0 0.0 291.0

98 62.5 0.0 126.8



#28

Bromochloromethane

Concen: 18.544 ug/l

RT: 4.904 min Scan# 626

Delta R.T. -0.006 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59




Tgt Ion: 49 Resp: 24738


Ion Ratio Lower Upper

49 100

129 1.7 0.0 3.8

130 76.8 61.4 92.0





#29

Tetrahydrofuran

Concen: 99.990 ug/l

RT: 5.007 min Scan# 64

Delta R.T. -0.006 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020





Tgt Ion: 42 Resp: 56765

Ion Ratio Lower Upper

42 100

72 42.0 34.9 52.3

71 40.8 32.2 48.2



#30

Chloroform

Concen: 21.785 ug/l

RT: 5.093 min Scan# 657

Delta R.T. -0.012 min

Lab File: VX046719.D




Acq: 17 Jun 2025 13:59


Tgt Ion: 83 Resp: 64109

Ion Ratio Lower Upper

83 100

85 67.0 53.7 80.5









#31
Cyclohexane
Concen: 20.982 ug/l
RT: 5.477 min Scan# 705
Delta R.T. -0.006 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020


Tgt Ion: 56 Resp: 58552
Ion Ratio Lower Upper
56 100
69 28.3 24.8 37.2
84 84.5 68.2 102.2



#32
1,1,1-Trichloroethane
Concen: 21.133 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.012 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Tgt Ion: 97 Resp: 54542
Ion Ratio Lower Upper
97 100
99 63.4 51.3 76.9
61 46.4 36.4 54.6









#33
 1,2-Dichloroethane-d4
 Concen: 16.410 ug/l
 RT: 5.964 min Scan# 800
 Delta R.T. -0.000 min
 Lab File: VX046719.D
 Acq: 17 Jun 2025 13:59




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC020


Tgt Ion: 65 Resp: 30582
 Ion Ratio Lower Upper
 65 100
 67 52.5 0.0 105.4



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.757 min Scan# 930
 Delta R.T. -0.012 min
 Lab File: VX046719.D
 Acq: 17 Jun 2025 13:59

Tgt Ion: 114 Resp: 226567
 Ion Ratio Lower Upper
 114 100
 63 21.2 0.0 44.2
 88 15.5 0.0 33.2









#35
Dibromofluoromethane
Concen: 16.114 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.012 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020


Tgt Ion:113 Resp: 24156
Ion Ratio Lower Upper
113 100
111 104.6 82.0 123.0
192 19.1 15.3 22.9



#36
1,1-Dichloropropene
Concen: 20.344 ug/l
RT: 5.696 min Scan# 756
Delta R.T. -0.006 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Tgt Ion: 75 Resp: 44023
Ion Ratio Lower Upper
75 100
110 34.9 17.2 51.5
77 30.6 24.2 36.4





#37

Ethyl Acetate

Concen: 19.984 ug/l

RT: 4.715 min Scan# 595

Delta R.T. -0.006 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020





Tgt Ion: 43 Resp: 39916

Ion Ratio Lower Upper

43 100

61 11.7 9.7 14.5

70 9.7 8.1 12.1



#38

Carbon Tetrachloride

Concen: 20.735 ug/l

RT: 5.684 min Scan# 754

Delta R.T. -0.006 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59




Tgt Ion: 117 Resp: 48694


Ion Ratio Lower Upper

117 100

119 90.9 75.2 112.8

121 31.8 24.1 36.1





#39

Methylcyclohexane

Concen: 20.428 ug/l

RT: 7.379 min Scan# 1

Delta R.T. -0.006 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020





Tgt Ion: 83 Resp: 58170

Ion Ratio Lower Upper

83 100

55 76.1 63.0 94.4

98 50.7 38.8 58.2



#40

Benzene

Concen: 20.904 ug/l

RT: 6.038 min Scan# 812

Delta R.T. -0.006 min

Lab File: VX046719.D




Acq: 17 Jun 2025 13:59


Tgt Ion: 78 Resp: 133872

Ion Ratio Lower Upper

78 100

77 23.3 18.8 28.2









#41
Methacrylonitrile
Concen: 20.632 ug/l
RT: 4.922 min Scan# 61
Delta R.T. -0.006 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020

Tgt Ion: 41 Resp: 21970




Ion	Ratio	Lower	Upper
41	100		
39	54.8	45.9	68.9
67	73.2	61.2	91.8
52	30.8	25.5	38.3




#42
1,2-Dichloroethane
Concen: 21.076 ug/l
RT: 6.086 min Scan# 820
Delta R.T. -0.006 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Tgt Ion: 62 Resp: 47430

Ion	Ratio	Lower	Upper
62	100		
98	8.7	0.0	17.8









#43
Isopropyl Acetate
Concen: 19.975 ug/l
RT: 6.342 min Scan# 862
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion: 43 Resp: 62432
Ion Ratio Lower Upper
43 100
61 19.2 15.3 22.9
87 12.5 10.0 15.0



#44
Trichloroethene
Concen: 20.788 ug/l
RT: 7.129 min Scan# 991
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Tgt Ion:130 Resp: 33927
Ion Ratio Lower Upper
130 100
95 101.1 0.0 202.6





#45

1,2-Dichloropropane

Concen: 21.226 ug/l

RT: 7.428 min Scan# 1040

Delta R.T. -0.006 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC020

Tgt Ion: 63 Resp: 33688

Ion Ratio Lower Upper

63 100

65 31.6 22.9 34.3



#46

Dibromomethane

Concen: 20.663 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. -0.006 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59




Tgt Ion: 93 Resp: 23569


Ion Ratio Lower Upper

93 100

95 84.4 66.2 99.4

174 97.1 77.0 115.6









#47
Bromodichloromethane
Concen: 21.227 ug/l
RT: 7.818 min Scan# 1104
Delta R.T. -0.006 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion: 83 Resp: 49411
Ion Ratio Lower Upper
83 100
85 65.7 51.4 77.2
127 8.5 6.9 10.3



#48
Methyl methacrylate
Concen: 20.322 ug/l
RT: 7.690 min Scan# 1083
Delta R.T. -0.006 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Tgt Ion: 41 Resp: 31856
Ion Ratio Lower Upper
41 100
69 82.0 63.4 95.2
39 57.2 45.4 68.0









#49
1,4-Dioxane
Concen: 382.035 ug/l
RT: 7.659 min Scan# 1078
Delta R.T. 0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020


Tgt Ion: 88 Resp: 8340
Ion Ratio Lower Upper
88 100
43 37.1 31.2 46.8
58 75.7 54.7 82.1



#50
Toluene-d8
Concen: 16.376 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Tgt Ion: 98 Resp: 88197
Ion Ratio Lower Upper
98 100
100 67.7 53.0 79.4





#51

4-Methyl-2-Pentanone

Concen: 100.586 ug/l

RT: 8.568 min Scan# 11

Delta R.T. -0.006 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC020

Tgt Ion: 43 Resp: 187509

Ion Ratio Lower Upper

43 100

58 37.8 30.8 46.2



#52

Toluene

Concen: 21.166 ug/l

RT: 8.714 min Scan# 1251

Delta R.T. -0.006 min

Lab File: VX046719.D




Acq: 17 Jun 2025 13:59


Tgt Ion: 92 Resp: 84030

Ion Ratio Lower Upper

92 100

91 169.0 134.8 202.2





#53

t-1,3-Dichloropropene

Concen: 20.304 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC020

Tgt Ion: 75 Resp: 43852

Ion Ratio Lower Upper

75 100

77 31.4 25.6 38.4



#54

cis-1,3-Dichloropropene

Concen: 20.526 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59




Tgt Ion: 75 Resp: 50326


Ion Ratio Lower Upper

75 100

77 30.9 25.3 37.9

39 47.9 39.2 58.8









#55
1,1,2-Trichloroethane
Concen: 20.862 ug/l
RT: 9.147 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020

Tgt Ion: 97 Resp: 30865




Ion	Ratio	Lower	Upper
97	100		
83	91.5	67.2	100.8
85	53.2	44.1	66.1
99	68.3	50.3	75.5




#56
Ethyl methacrylate
Concen: 20.468 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Tgt Ion: 69 Resp: 44702

Ion	Ratio	Lower	Upper
69	100		
41	70.7	55.4	83.2
39	37.3	29.7	44.5





#57

1,3-Dichloropropane

Concen: 20.480 ug/l

RT: 9.305 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC020

Tgt Ion: 76 Resp: 52181

Ion Ratio Lower Upper

76 100

78 32.6 25.7 38.5



#58

2-Chloroethyl Vinyl ether

Concen: 93.878 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. -0.006 min

Lab File: VX046719.D




Acq: 17 Jun 2025 13:59


Tgt Ion: 63 Resp: 106505

Ion Ratio Lower Upper

63 100

106 27.9 22.6 34.0





#59

2-Hexanone

Concen: 100.265 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. -0.000 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC020

Tgt Ion: 43 Resp: 129218

Ion Ratio Lower Upper

43 100

58 52.8 26.8 80.4



#60

Dibromochloromethane

Concen: 20.872 ug/l

RT: 9.519 min Scan# 1383

Delta R.T. -0.000 min

Lab File: VX046719.D




Acq: 17 Jun 2025 13:59


Tgt Ion: 129 Resp: 36392

Ion Ratio Lower Upper

129 100

127 78.0 38.4 115.2









#61
1,2-Dibromoethane
Concen: 20.936 ug/l
RT: 9.604 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion:107 Resp: 31507
Ion Ratio Lower Upper
107 100
109 95.0 75.8 113.6



#62
4-Bromofluorobenzene
Concen: 15.973 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Tgt Ion: 95 Resp: 32067
Ion Ratio Lower Upper
95 100
174 78.1 0.0 150.4
176 75.2 0.0 145.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. -0.006 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020


Tgt Ion:117 Resp: 197947
Ion Ratio Lower Upper
117 100
82 58.9 44.6 66.8
119 33.0 25.8 38.8



#64
Tetrachloroethene
Concen: 20.408 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. -0.006 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Tgt Ion:164 Resp: 27988
Ion Ratio Lower Upper
164 100
166 133.4 104.6 157.0
129 102.4 76.3 114.5
131 93.3 74.1 111.1









#65
Chlorobenzene
Concen: 20.796 ug/l
RT: 10.073 min Scan# 1475
Delta R.T. -0.006 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion:112 Resp: 90868
Ion Ratio Lower Upper
112 100
114 31.9 25.5 38.3



#66
1,1,1,2-Tetrachloroethane
Concen: 21.455 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Tgt Ion:131 Resp: 31508
Ion Ratio Lower Upper
131 100
133 94.3 47.3 141.8
119 63.4 32.1 96.3





#67

Ethyl Benzene

Concen: 21.056 ug/l

RT: 10.189 min Scan# 1493

Delta R.T. -0.000 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC020

Tgt Ion: 91 Resp: 160761

Ion Ratio Lower Upper

91 100

106 30.5 24.2 36.2



#68

m/p-Xylenes

Concen: 42.539 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. -0.000 min

Lab File: VX046719.D




Acq: 17 Jun 2025 13:59


Tgt Ion: 106 Resp: 121014

Ion Ratio Lower Upper

106 100

91 207.7 164.8 247.2









#69
o-Xylene
Concen: 21.638 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion:106 Resp: 57689
Ion Ratio Lower Upper
106 100
91 222.2 111.3 333.9



#70
Styrene
Concen: 21.318 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Tgt Ion:104 Resp: 99462
Ion Ratio Lower Upper
104 100
78 52.6 41.4 62.2
103 54.4 43.1 64.7





#71

Bromoform

Concen: 20.898 ug/l

RT: 10.799 min Scan# 1593

Delta R.T. -0.000 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020





Tgt Ion:173 Resp: 23372

Ion Ratio Lower Upper

173 100

175 48.0 24.1 72.4

254 0.0 0.1 0.1#



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 1793

Delta R.T. -0.000 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59




Tgt Ion:152 Resp: 98285


Ion Ratio Lower Upper

152 100

115 71.7 43.2 129.6

150 166.5 0.0 346.8









#73
Isopropylbenzene
Concen: 21.255 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion:105 Resp: 153861
Ion Ratio Lower Upper
105 100
120 25.6 13.0 39.0



#74
N-ethyl acetate
Concen: 20.851 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Tgt Ion: 43 Resp: 57185
Ion Ratio Lower Upper
43 100
70 40.3 33.4 50.0
55 23.9 19.1 28.7
61 23.3 19.0 28.4









#75
1,1,2,2-Tetrachloroethane
Concen: 20.914 ug/l
RT: 11.207 min Scan# 1660
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion: 83 Resp: 42272
Ion Ratio Lower Upper
83 100
131 10.0 4.9 14.6
85 64.0 32.0 96.0



#76
1,2,3-Trichloropropane
Concen: 22.117 ug/l
RT: 11.238 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Tgt Ion: 75 Resp: 44152
Ion Ratio Lower Upper
75 100
77 32.9 18.5 55.5









#77
Bromobenzene
Concen: 21.507 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion:156 Resp: 37464
Ion Ratio Lower Upper
156 100
77 163.6 82.8 248.6
158 96.6 47.6 142.9



#78
n-propylbenzene
Concen: 21.410 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Tgt Ion: 91 Resp: 184078
Ion Ratio Lower Upper
91 100
120 22.5 11.4 34.1









#79
2-Chlorotoluene
Concen: 20.952 ug/l
RT: 11.360 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion: 91 Resp: 107552
Ion Ratio Lower Upper
91 100
126 33.4 16.6 49.7



#80
1,3,5-Trimethylbenzene
Concen: 21.793 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Tgt Ion: 105 Resp: 129635
Ion Ratio Lower Upper
105 100
120 47.6 24.1 72.2









#81
 trans-1,4-Dichloro-2-butene
 Concen: 18.362 ug/l
 RT: 11.018 min Scan# 1629
 Delta R.T. -0.000 min
 Lab File: VX046719.D
 Acq: 17 Jun 2025 13:59




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC020


Tgt Ion: 75 Resp: 11694
 Ion Ratio Lower Upper
 75 100
 53 111.5 86.8 130.2
 89 46.7 36.9 55.3



#82
 4-Chlorotoluene
 Concen: 21.358 ug/l
 RT: 11.451 min Scan# 1700
 Delta R.T. -0.000 min
 Lab File: VX046719.D
 Acq: 17 Jun 2025 13:59

Tgt Ion: 91 Resp: 127954
 Ion Ratio Lower Upper
 91 100
 126 28.6 14.4 43.2





#83

tert-Butylbenzene

Concen: 21.201 ug/l

RT: 11.713 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020





Tgt Ion:119 Resp: 130359

Ion Ratio Lower Upper

119 100

91 58.4 29.5 88.6

134 23.0 11.7 35.0



#84

1,2,4-Trimethylbenzene

Concen: 21.512 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. -0.000 min

Lab File: VX046719.D




Acq: 17 Jun 2025 13:59


Tgt Ion:105 Resp: 128566

Ion Ratio Lower Upper

105 100

120 43.5 22.1 66.3









#85
 sec-Butylbenzene
 Concen: 21.239 ug/l
 RT: 11.890 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: VX046719.D
 Acq: 17 Jun 2025 13:59




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC020


Tgt Ion:105 Resp: 165047
 Ion Ratio Lower Upper
 105 100
 134 20.0 9.8 29.4



#86
 p-Isopropyltoluene
 Concen: 21.142 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. -0.000 min
 Lab File: VX046719.D
 Acq: 17 Jun 2025 13:59

Tgt Ion:119 Resp: 138623
 Ion Ratio Lower Upper
 119 100
 134 25.9 12.9 38.7
 91 24.8 12.3 36.9









#87
1,3-Dichlorobenzene
Concen: 20.681 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020


Tgt Ion:146 Resp: 70243
Ion Ratio Lower Upper
146 100
111 40.5 20.6 61.8
148 62.4 31.6 94.8



#88
1,4-Dichlorobenzene
Concen: 20.625 ug/l
RT: 12.036 min Scan# 1796
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Tgt Ion:146 Resp: 71953
Ion Ratio Lower Upper
146 100
111 40.7 20.3 61.0
148 62.6 32.1 96.5





#89

n-Butylbenzene

Concen: 20.738 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. -0.000 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020





Tgt Ion: 91 Resp: 129101

Ion Ratio Lower Upper

91 100

92 54.2 27.1 81.3

134 25.8 12.9 38.6



#90

Hexachloroethane

Concen: 20.594 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX046719.D




Acq: 17 Jun 2025 13:59


Tgt Ion: 117 Resp: 23645

Ion Ratio Lower Upper

117 100

201 71.6 34.4 103.3









#91
1,2-Dichlorobenzene
Concen: 21.073 ug/l
RT: 12.329 min Scan# 1844
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion:146 Resp: 66883
Ion Ratio Lower Upper
146 100
111 43.0 21.7 65.1
148 63.5 32.0 96.2



#92
1,2-Dibromo-3-Chloropropane
Concen: 20.226 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Tgt Ion: 75 Resp: 8071
Ion Ratio Lower Upper
75 100
155 79.6 41.9 125.7
157 102.7 51.4 154.3





#93

1,2,4-Trichlorobenzene

Concen: 19.821 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. -0.000 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020





Tgt Ion:180 Resp: 44726

Ion Ratio Lower Upper

180 100

182 96.1 46.7 140.1

145 32.8 15.8 47.4



#94

Hexachlorobutadiene

Concen: 20.752 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. 0.006 min

Lab File: VX046719.D

Acq: 17 Jun 2025 13:59




Tgt Ion:225 Resp: 19698


Ion Ratio Lower Upper

225 100

223 66.0 32.5 97.5

227 65.3 32.2 96.6









#95
Naphthalene
Concen: 20.186 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020

Tgt Ion:128 Resp: 129768
Ion Ratio Lower Upper
128 100
127 12.7 10.1 15.1
129 10.8 8.7 13.1



#96
1,2,3-Trichlorobenzene
Concen: 20.558 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. -0.000 min
Lab File: VX046719.D
Acq: 17 Jun 2025 13:59

Tgt Ion:180 Resp: 44380
Ion Ratio Lower Upper
180 100
182 95.4 47.8 143.4
145 34.4 17.3 51.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
 Data File : VX046720.D
 Acq On : 17 Jun 2025 14:20
 Operator : JC/MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050

Quant Time: Jun 18 02:39:23 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 02:34:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	123412	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	204400	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	179407	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	89016	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	80133	46.944	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	93.880%		
35) Dibromofluoromethane	5.397	113	65447	48.394	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.780%		
50) Toluene-d8	8.647	98	233825	48.124	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	96.240%		
62) 4-Bromofluorobenzene	11.079	95	87131	48.107	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	96.220%		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.185	85	71910	54.577	ug/l	100
3) Chloromethane	1.307	50	73175	51.454	ug/l	100
4) Vinyl Chloride	1.386	62	78242	51.592	ug/l	100
5) Bromomethane	1.624	94	45535	53.367	ug/l	100
6) Chloroethane	1.703	64	46570	50.665	ug/l	100
7) Trichlorofluoromethane	1.904	101	119335	51.823	ug/l	100
8) Diethyl Ether	2.148	74	39712	50.251	ug/l	100
9) 1,1,2-Trichlorotrifluo...	2.343	101	71037	50.285	ug/l	100
10) Methyl Iodide	2.471	142	80710	46.019	ug/l	100
11) Tert butyl alcohol	2.959	59	38256	249.692	ug/l	100
12) 1,1-Dichloroethene	2.337	96	70202	51.500	ug/l	100
13) Acrolein	2.246	56	53791	241.036	ug/l	100
14) Allyl chloride	2.678	41	123265	49.909	ug/l	100
15) Acrylonitrile	3.075	53	175846	255.243	ug/l	100
16) Acetone	2.386	43	135683	238.356	ug/l	100
17) Carbon Disulfide	2.526	76	197344	47.941	ug/l	100
18) Methyl Acetate	2.715	43	71685	49.582	ug/l	100
19) Methyl tert-butyl Ether	3.123	73	216266	50.953	ug/l	100
20) Methylene Chloride	2.800	84	75315	47.576	ug/l	100
21) trans-1,2-Dichloroethene	3.105	96	72648	49.807	ug/l	100
22) Diisopropyl ether	3.770	45	242789	51.143	ug/l	100
23) Vinyl Acetate	3.733	43	985040	260.935	ug/l	100
24) 1,1-Dichloroethane	3.623	63	134324	49.835	ug/l	100
25) 2-Butanone	4.562	43	208346	258.788	ug/l	100
26) 2,2-Dichloropropane	4.489	77	93956	47.794	ug/l	100
27) cis-1,2-Dichloroethene	4.501	96	84638	49.392	ug/l	100
28) Bromochloromethane	4.910	49	59857	48.984	ug/l	100
29) Tetrahydrofuran	5.013	42	133555	256.831	ug/l	100
30) Chloroform	5.105	83	137459	50.996	ug/l	100
31) Cyclohexane	5.483	56	125014	48.909	ug/l	100
32) 1,1,1-Trichloroethane	5.397	97	117954	49.895	ug/l	100
36) 1,1-Dichloropropene	5.702	75	97413	49.899	ug/l	100
37) Ethyl Acetate	4.721	43	88076	48.877	ug/l	100
38) Carbon Tetrachloride	5.690	117	105202	49.656	ug/l	100
39) Methylcyclohexane	7.385	83	128553	50.041	ug/l	100
40) Benzene	6.044	78	289724	50.146	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
 Data File : VX046720.D
 Acq On : 17 Jun 2025 14:20
 Operator : JC/MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050

Quant Time: Jun 18 02:39:23 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 02:34:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.928	41	48765	50.762	ug/l	100
42) 1,2-Dichloroethane	6.092	62	101218	49.855	ug/l	100
43) Isopropyl Acetate	6.342	43	146377	51.911	ug/l	100
44) Trichloroethene	7.129	130	72796	49.441	ug/l	100
45) 1,2-Dichloropropane	7.434	63	70702	49.378	ug/l	100
46) Dibromomethane	7.586	93	51288	49.840	ug/l	100
47) Bromodichloromethane	7.824	83	106651	50.787	ug/l	100
48) Methyl methacrylate	7.696	41	75708	53.535	ug/l	100
49) 1,4-Dioxane	7.659	88	19802	962.534	ug/l	100
51) 4-Methyl-2-Pentanone	8.574	43	435000	258.656	ug/l	100
52) Toluene	8.720	92	181414	50.650	ug/l	100
53) t-1,3-Dichloropropene	8.976	75	99727	51.183	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	111917	50.598	ug/l	100
55) 1,1,2-Trichloroethane	9.153	97	66688	49.963	ug/l	100
56) Ethyl methacrylate	9.116	69	103697	52.629	ug/l	100
57) 1,3-Dichloropropane	9.305	76	114824	49.955	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.245	63	287531	280.927	ug/l	100
59) 2-Hexanone	9.427	43	303322	260.883	ug/l	100
60) Dibromochloromethane	9.519	129	79227	50.368	ug/l	100
61) 1,2-Dibromoethane	9.610	107	68515	50.464	ug/l	100
64) Tetrachloroethene	9.275	164	61048	49.115	ug/l	100
65) Chlorobenzene	10.080	112	195752	49.429	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.159	131	67004	50.341	ug/l	100
67) Ethyl Benzene	10.189	91	346812	50.119	ug/l	100
68) m/p-Xylenes	10.299	106	259899	100.800	ug/l	100
69) o-Xylene	10.640	106	124120	51.366	ug/l	100
70) Styrene	10.653	104	216738	51.254	ug/l	100
71) Bromoform	10.799	173	51485	50.792	ug/l #	100
73) Isopropylbenzene	10.957	105	331391	50.548	ug/l	100
74) N-amyl acetate	10.842	43	130606	52.581	ug/l	100
75) 1,1,2,2-Tetrachloroethane	11.207	83	92893	50.745	ug/l	100
76) 1,2,3-Trichloropropane	11.238	75	99621	55.099	ug/l	93
77) Bromobenzene	11.195	156	79404	50.331	ug/l	100
78) n-propylbenzene	11.299	91	394375	50.646	ug/l	100
79) 2-Chlorotoluene	11.360	91	230178	49.509	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	273477	50.761	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	27677	47.984	ug/l	100
82) 4-Chlorotoluene	11.451	91	269198	49.613	ug/l	100
83) tert-Butylbenzene	11.713	119	277915	49.905	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	272510	50.345	ug/l	100
85) sec-Butylbenzene	11.890	105	355947	50.574	ug/l	100
86) p-Isopropyltoluene	12.006	119	297314	50.067	ug/l	100
87) 1,3-Dichlorobenzene	11.969	146	149402	48.568	ug/l	100
88) 1,4-Dichlorobenzene	12.037	146	147146	46.571	ug/l	100
89) n-Butylbenzene	12.329	91	280794	49.801	ug/l	100
90) Hexachloroethane	12.536	117	51961	49.968	ug/l	100
91) 1,2-Dichlorobenzene	12.329	146	141743	49.309	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	12.939	75	18747	51.873	ug/l	100
93) 1,2,4-Trichlorobenzene	13.585	180	100346	49.101	ug/l	100
94) Hexachlorobutadiene	13.719	225	42659	49.621	ug/l	100
95) Naphthalene	13.774	128	296912	50.994	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	96318	49.263	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
Data File : VX046720.D
Acq On : 17 Jun 2025 14:20
Operator : JC/MD
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Quant Time: Jun 18 02:39:23 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 02:34:13 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jun 18 02:39:23 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 02:34:13 2025
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.562 min Scan# 71

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion:168 Resp: 123412

Ion Ratio Lower Upper

168 100

99 60.6 48.5 72.7



#2

Dichlorodifluoromethane

Concen: 54.577 ug/l

RT: 1.185 min Scan# 16

Delta R.T. 0.000 min

Lab File: VX046720.D




Acq: 17 Jun 2025 14:20


Tgt Ion: 85 Resp: 71910

Ion Ratio Lower Upper

85 100

87 33.2 16.6 49.8





#3

Chloromethane

Concen: 51.454 ug/l

RT: 1.307 min Scan# 30

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion: 50 Resp: 73175

Ion Ratio Lower Upper

50 100

52 31.1 24.9 37.3



#4

Vinyl Chloride

Concen: 51.592 ug/l

RT: 1.386 min Scan# 49

Delta R.T. 0.000 min

Lab File: VX046720.D




Acq: 17 Jun 2025 14:20


Tgt Ion: 62 Resp: 78242

Ion Ratio Lower Upper

62 100

64 29.9 23.9 35.9









#5
Bromomethane
Concen: 53.367 ug/l
RT: 1.624 min Scan# 88
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 94 Resp: 45535
Ion Ratio Lower Upper
94 100
96 90.4 72.3 108.5



#6
Chloroethane
Concen: 50.665 ug/l
RT: 1.703 min Scan# 101
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Tgt Ion: 64 Resp: 46570
Ion Ratio Lower Upper
64 100
66 33.9 27.1 40.7









#7
 Trichlorofluoromethane
 Concen: 51.823 ug/l
 RT: 1.904 min Scan# 11
 Delta R.T. 0.000 min
 Lab File: VX046720.D
 Acq: 17 Jun 2025 14:20




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050


Tgt Ion:101 Resp: 119335
 Ion Ratio Lower Upper
 101 100
 103 64.8 51.8 77.8



#8
 Diethyl Ether
 Concen: 50.251 ug/l
 RT: 2.148 min Scan# 174
 Delta R.T. 0.000 min
 Lab File: VX046720.D
 Acq: 17 Jun 2025 14:20

Tgt Ion: 74 Resp: 39712
 Ion Ratio Lower Upper
 74 100
 45 104.7 52.3 157.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 50.285 ug/l
RT: 2.343 min Scan# 206
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:101 Resp: 71037
Ion Ratio Lower Upper
101 100
85 44.2 35.4 53.0
151 76.5 61.2 91.8



#10
Methyl Iodide
Concen: 46.019 ug/l
RT: 2.471 min Scan# 227
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Tgt Ion:142 Resp: 80710
Ion Ratio Lower Upper
142 100
127 46.3 37.0 55.6
141 14.4 11.5 17.3









#11
Tert butyl alcohol
Concen: 249.692 ug/l
RT: 2.959 min Scan# 307
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 59 Resp: 38256
Ion Ratio Lower Upper
59 100
57 10.4 8.3 12.5



#12
1,1-Dichloroethene
Concen: 51.500 ug/l
RT: 2.337 min Scan# 205
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Tgt Ion: 96 Resp: 70202
Ion Ratio Lower Upper
96 100
61 158.2 126.6 189.8
98 63.4 50.7 76.1





#13

Acrolein

Concen: 241.036 ug/l

RT: 2.246 min Scan# 190

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion: 56 Resp: 53791

Ion Ratio Lower Upper

56 100

55 69.0 55.2 82.8



#14

Allyl chloride

Concen: 49.909 ug/l

RT: 2.678 min Scan# 261

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20




Tgt Ion: 41 Resp: 123265


Ion Ratio Lower Upper

41 100

39 67.0 53.6 80.4

76 33.5 26.8 40.2





#15

Acrylonitrile

Concen: 255.243 ug/l

RT: 3.075 min Scan# 31

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion: 53 Resp: 175846

Ion Ratio Lower Upper

53 100

52 81.4 65.1 97.7

51 36.1 28.9 43.3



#16

Acetone

Concen: 238.356 ug/l

RT: 2.386 min Scan# 213

Delta R.T. 0.000 min

Lab File: VX046720.D




Acq: 17 Jun 2025 14:20


Tgt Ion: 43 Resp: 135683

Ion Ratio Lower Upper

43 100

58 30.6 24.5 36.7









#17
Carbon Disulfide
Concen: 47.941 ug/l
RT: 2.526 min Scan# 21
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 76 Resp: 197344
Ion Ratio Lower Upper
76 100
78 9.4 7.5 11.3



#18
Methyl Acetate
Concen: 49.582 ug/l
RT: 2.715 min Scan# 267
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Tgt Ion: 43 Resp: 71685
Ion Ratio Lower Upper
43 100
74 23.0 18.4 27.6









#19
Methyl tert-butyl Ether
Concen: 50.953 ug/l
RT: 3.123 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 73 Resp: 216266
Ion Ratio Lower Upper
73 100
57 22.7 18.2 27.2



#20
Methylene Chloride
Concen: 47.576 ug/l
RT: 2.800 min Scan# 281
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Tgt Ion: 84 Resp: 75315
Ion Ratio Lower Upper
84 100
49 126.6 101.3 151.9
51 38.8 31.0 46.6
86 63.9 51.1 76.7









#21
trans-1,2-Dichloroethene
Concen: 49.807 ug/l
RT: 3.105 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 96 Resp: 72648
Ion Ratio Lower Upper
96 100
61 140.6 112.5 168.7
98 64.9 51.9 77.9



#22
Diisopropyl ether
Concen: 51.143 ug/l
RT: 3.770 min Scan# 440
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Tgt Ion: 45 Resp: 242789
Ion Ratio Lower Upper
45 100
43 114.9 91.9 137.9
87 27.5 22.0 33.0
59 11.4 9.1 13.7









#23
 Vinyl Acetate
 Concen: 260.935 ug/l
 RT: 3.733 min Scan# 411
 Delta R.T. 0.000 min
 Lab File: VX046720.D
 Acq: 17 Jun 2025 14:20




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050


Tgt Ion: 43 Resp: 985040
 Ion Ratio Lower Upper
 43 100
 86 10.2 8.2 12.2



#24
 1,1-Dichloroethane
 Concen: 49.835 ug/l
 RT: 3.623 min Scan# 416
 Delta R.T. 0.000 min
 Lab File: VX046720.D
 Acq: 17 Jun 2025 14:20

Tgt Ion: 63 Resp: 134324
 Ion Ratio Lower Upper
 63 100
 98 7.9 4.0 11.9
 100 4.4 2.2 6.6





#25

2-Butanone

Concen: 258.788 ug/l

RT: 4.562 min Scan# 570

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion: 43 Resp: 208346

Ion Ratio Lower Upper

43 100

72 25.3 20.2 30.4



#26

2,2-Dichloropropane

Concen: 47.794 ug/l

RT: 4.489 min Scan# 558

Delta R.T. 0.000 min

Lab File: VX046720.D




Acq: 17 Jun 2025 14:20


Tgt Ion: 77 Resp: 93956

Ion Ratio Lower Upper

77 100

97 23.5 11.8 35.3





#27

cis-1,2-Dichloroethene

Concen: 49.392 ug/l

RT: 4.501 min Scan# 560

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion: 96 Resp: 84638

Ion Ratio Lower Upper

96 100

61 145.5 0.0 291.0

98 63.4 0.0 126.8



#28

Bromochloromethane

Concen: 48.984 ug/l

RT: 4.910 min Scan# 627

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20




Tgt Ion: 49 Resp: 59857


Ion Ratio Lower Upper

49 100

129 1.9 0.0 3.8

130 76.7 61.4 92.0





#29

Tetrahydrofuran

Concen: 256.831 ug/l

RT: 5.013 min Scan# 644

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion: 42 Resp: 133555

Ion Ratio Lower Upper

42 100

72 43.6 34.9 52.3

71 40.2 32.2 48.2



#30

Chloroform

Concen: 50.996 ug/l

RT: 5.105 min Scan# 659

Delta R.T. 0.000 min

Lab File: VX046720.D




Acq: 17 Jun 2025 14:20


Tgt Ion: 83 Resp: 137459

Ion Ratio Lower Upper

83 100

85 67.1 53.7 80.5









#31
Cyclohexane
Concen: 48.909 ug/l
RT: 5.483 min Scan# 721
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 56 Resp: 125014
Ion Ratio Lower Upper
56 100
69 31.0 24.8 37.2
84 85.2 68.2 102.2



#32
1,1,1-Trichloroethane
Concen: 49.895 ug/l
RT: 5.397 min Scan# 707
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Tgt Ion: 97 Resp: 117954
Ion Ratio Lower Upper
97 100
99 64.1 51.3 76.9
61 45.5 36.4 54.6





#33

1,2-Dichloroethane-d4

Concen: 46.944 ug/l

RT: 5.964 min Scan# 800

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion: 65 Resp: 80133

Ion Ratio Lower Upper

65 100

67 52.7 0.0 105.4



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.769 min Scan# 932

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20




Tgt Ion: 114 Resp: 204400


Ion Ratio Lower Upper

114 100

63 22.1 0.0 44.2

88 16.6 0.0 33.2









#35
Dibromofluoromethane
Concen: 48.394 ug/l
RT: 5.397 min Scan# 707
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:113 Resp: 65447
Ion Ratio Lower Upper
113 100
111 102.5 82.0 123.0
192 19.1 15.3 22.9



#36
1,1-Dichloropropene
Concen: 49.899 ug/l
RT: 5.702 min Scan# 757
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Tgt Ion: 75 Resp: 97413
Ion Ratio Lower Upper
75 100
110 34.3 17.2 51.5
77 30.3 24.2 36.4









#37
Ethyl Acetate
Concen: 48.877 ug/l
RT: 4.721 min Scan# 596
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 43 Resp: 88076
Ion Ratio Lower Upper
43 100
61 12.1 9.7 14.5
70 10.1 8.1 12.1



#38
Carbon Tetrachloride
Concen: 49.656 ug/l
RT: 5.690 min Scan# 755
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Tgt Ion: 117 Resp: 105202
Ion Ratio Lower Upper
117 100
119 94.0 75.2 112.8
121 30.1 24.1 36.1









#39
Methylcyclohexane
Concen: 50.041 ug/l
RT: 7.385 min Scan# 1033
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 83 Resp: 128553
Ion Ratio Lower Upper
83 100
55 78.7 63.0 94.4
98 48.5 38.8 58.2



#40
Benzene
Concen: 50.146 ug/l
RT: 6.044 min Scan# 813
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Tgt Ion: 78 Resp: 289724
Ion Ratio Lower Upper
78 100
77 23.5 18.8 28.2









#41
Methacrylonitrile
Concen: 50.762 ug/l
RT: 4.928 min Scan# 61
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:	41	Resp:	48765
Ion	Ratio	Lower	Upper
41	100		
39	57.4	45.9	68.9
67	76.5	61.2	91.8
52	31.9	25.5	38.3



#42
1,2-Dichloroethane
Concen: 49.855 ug/l
RT: 6.092 min Scan# 821
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Tgt Ion:	62	Resp:	101218
Ion	Ratio	Lower	Upper
62	100		
98	8.9	0.0	17.8





#43

Isopropyl Acetate

Concen: 51.911 ug/l

RT: 6.342 min Scan# 862

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion: 43 Resp: 146377

Ion Ratio Lower Upper

43 100

61 19.1 15.3 22.9

87 12.5 10.0 15.0



#44

Trichloroethene

Concen: 49.441 ug/l

RT: 7.129 min Scan# 991

Delta R.T. 0.000 min

Lab File: VX046720.D




Acq: 17 Jun 2025 14:20


Tgt Ion:130 Resp: 72796

Ion Ratio Lower Upper

130 100

95 101.3 0.0 202.6





#45

1,2-Dichloropropane

Concen: 49.378 ug/l

RT: 7.434 min Scan# 1041

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion: 63 Resp: 70702

Ion Ratio Lower Upper

63 100

65 28.6 22.9 34.3



#46

Dibromomethane

Concen: 49.840 ug/l

RT: 7.586 min Scan# 1066

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20




Tgt Ion: 93 Resp: 51288


Ion Ratio Lower Upper

93 100

95 82.8 66.2 99.4

174 96.3 77.0 115.6









#47
Bromodichloromethane
Concen: 50.787 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 83 Resp: 106651
Ion Ratio Lower Upper
83 100
85 64.3 51.4 77.2
127 8.6 6.9 10.3



#48
Methyl methacrylate
Concen: 53.535 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Tgt Ion: 41 Resp: 75708
Ion Ratio Lower Upper
41 100
69 79.3 63.4 95.2
39 56.7 45.4 68.0





#49

1,4-Dioxane

Concen: 962.534 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion: 88 Resp: 19802

Ion Ratio Lower Upper

88 100

43 39.0 31.2 46.8

58 68.4 54.7 82.1



#50

Toluene-d8

Concen: 48.124 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. 0.000 min

Lab File: VX046720.D




Acq: 17 Jun 2025 14:20


Tgt Ion: 98 Resp: 233825

Ion Ratio Lower Upper

98 100

100 66.2 53.0 79.4









#51
4-Methyl-2-Pentanone
Concen: 258.656 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 43 Resp: 435000
Ion Ratio Lower Upper
43 100
58 38.5 30.8 46.2



#52
Toluene
Concen: 50.650 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Tgt Ion: 92 Resp: 181414
Ion Ratio Lower Upper
92 100
91 168.5 134.8 202.2





#53

t-1,3-Dichloropropene

Concen: 51.183 ug/l

RT: 8.976 min Scan# 1129

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion: 75 Resp: 99727

Ion Ratio Lower Upper

75 100

77 32.0 25.6 38.4



#54

cis-1,3-Dichloropropene

Concen: 50.598 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20




Tgt Ion: 75 Resp: 111917


Ion Ratio Lower Upper

75 100

77 31.6 25.3 37.9

39 49.0 39.2 58.8









#55
1,1,2-Trichloroethane
Concen: 49.963 ug/l
RT: 9.153 min Scan# 1323
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:	97	Resp:	66688
Ion	Ratio	Lower	Upper
97	100		
83	84.0	67.2	100.8
85	55.1	44.1	66.1
99	62.9	50.3	75.5



#56
Ethyl methacrylate
Concen: 52.629 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Tgt Ion:	69	Resp:	103697
Ion	Ratio	Lower	Upper
69	100		
41	69.3	55.4	83.2
39	37.1	29.7	44.5





#57

1,3-Dichloropropane

Concen: 49.955 ug/l

RT: 9.305 min Scan# 114824

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion: 76 Resp: 114824

Ion Ratio Lower Upper

76 100

78 32.1 25.7 38.5



#58

2-Chloroethyl Vinyl ether

Concen: 280.927 ug/l

RT: 8.245 min Scan# 1174

Delta R.T. 0.000 min

Lab File: VX046720.D




Acq: 17 Jun 2025 14:20


Tgt Ion: 63 Resp: 287531

Ion Ratio Lower Upper

63 100

106 28.3 22.6 34.0





#59

2-Hexanone

Concen: 260.883 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion: 43 Resp: 303322

Ion Ratio Lower Upper

43 100

58 53.6 26.8 80.4



#60

Dibromochloromethane

Concen: 50.368 ug/l

RT: 9.519 min Scan# 1383

Delta R.T. 0.000 min

Lab File: VX046720.D




Acq: 17 Jun 2025 14:20


Tgt Ion: 129 Resp: 79227

Ion Ratio Lower Upper

129 100

127 76.8 38.4 115.2









#61
1,2-Dibromoethane
Concen: 50.464 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:107 Resp: 68515
Ion Ratio Lower Upper
107 100
109 94.7 75.8 113.6



#62
4-Bromofluorobenzene
Concen: 48.107 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Tgt Ion: 95 Resp: 87131
Ion Ratio Lower Upper
95 100
174 75.2 0.0 150.4
176 72.5 0.0 145.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:117 Resp: 179407
Ion Ratio Lower Upper
117 100
82 55.7 44.6 66.8
119 32.3 25.8 38.8



#64
Tetrachloroethene
Concen: 49.115 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Tgt Ion:164 Resp: 61048
Ion Ratio Lower Upper
164 100
166 130.8 104.6 157.0
129 95.4 76.3 114.5
131 92.6 74.1 111.1









#65
Chlorobenzene
Concen: 49.429 ug/l
RT: 10.080 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:112 Resp: 195752
Ion Ratio Lower Upper
112 100
114 31.9 25.5 38.3



#66
1,1,1,2-Tetrachloroethane
Concen: 50.341 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Tgt Ion:131 Resp: 67004
Ion Ratio Lower Upper
131 100
133 94.5 47.3 141.8
119 64.2 32.1 96.3





#67

Ethyl Benzene

Concen: 50.119 ug/l

RT: 10.189 min Scan# 1493

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion: 91 Resp: 346812

Ion Ratio Lower Upper

91 100

106 30.2 24.2 36.2



#68

m/p-Xylenes

Concen: 100.800 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. 0.000 min

Lab File: VX046720.D




Acq: 17 Jun 2025 14:20


Tgt Ion:106 Resp: 259899

Ion Ratio Lower Upper

106 100

91 206.0 164.8 247.2









#69
o-Xylene
Concen: 51.366 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:106 Resp: 124120
Ion Ratio Lower Upper
106 100
91 222.6 111.3 333.9



#70
Styrene
Concen: 51.254 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Tgt Ion:104 Resp: 216738
Ion Ratio Lower Upper
104 100
78 51.8 41.4 62.2
103 53.9 43.1 64.7





#71

Bromoform

Concen: 50.792 ug/l

RT: 10.799 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion:173 Resp: 51485

Ion Ratio Lower Upper

173 100

175 48.3 24.1 72.4

254 0.1 0.1 0.1#



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 1793

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20




Tgt Ion:152 Resp: 89016


Ion Ratio Lower Upper

152 100

115 86.4 43.2 129.6

150 173.4 0.0 346.8









#73
Isopropylbenzene
Concen: 50.548 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:105 Resp: 331391
Ion Ratio Lower Upper
105 100
120 26.0 13.0 39.0



#74
N-ethyl acetate
Concen: 52.581 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Tgt Ion: 43 Resp: 130606
Ion Ratio Lower Upper
43 100
70 41.7 33.4 50.0
55 23.9 19.1 28.7
61 23.7 19.0 28.4





#75

1,1,2-Tetrachloroethane

Concen: 50.745 ug/l

RT: 11.207 min Scan# 1660

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion: 83 Resp: 92893

Ion Ratio Lower Upper

83 100

131 9.7 4.9 14.6

85 64.0 32.0 96.0



#76

1,2,3-Trichloropropane

Concen: 55.099 ug/l

RT: 11.238 min Scan# 1665

Delta R.T. 0.000 min

Lab File: VX046720.D




Acq: 17 Jun 2025 14:20


Tgt Ion: 75 Resp: 99621

Ion Ratio Lower Upper

75 100

77 32.7 18.5 55.5





#77

Bromobenzene

Concen: 50.331 ug/l

RT: 11.195 min Scan# 1658

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion:156 Resp: 79404

Ion Ratio Lower Upper

156 100

77 165.7 82.8 248.6

158 95.3 47.6 142.9



#78

n-propylbenzene

Concen: 50.646 ug/l

RT: 11.299 min Scan# 1675

Delta R.T. 0.000 min

Lab File: VX046720.D




Acq: 17 Jun 2025 14:20


Tgt Ion: 91 Resp: 394375

Ion Ratio Lower Upper

91 100

120 22.7 11.4 34.1









#79
2-Chlorotoluene
Concen: 49.509 ug/l
RT: 11.360 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 91 Resp: 230178
Ion Ratio Lower Upper
91 100
126 33.1 16.6 49.7



#80
1,3,5-Trimethylbenzene
Concen: 50.761 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Tgt Ion: 105 Resp: 273477
Ion Ratio Lower Upper
105 100
120 48.1 24.1 72.2









#81
trans-1,4-Dichloro-2-butene
Concen: 47.984 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 75 Resp: 27677
Ion Ratio Lower Upper
75 100
53 108.5 86.8 130.2
89 46.1 36.9 55.3



#82
4-Chlorotoluene
Concen: 49.613 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Tgt Ion: 91 Resp: 269198
Ion Ratio Lower Upper
91 100
126 28.8 14.4 43.2





#83

tert-Butylbenzene

Concen: 49.905 ug/l

RT: 11.713 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion:119 Resp: 277915

Ion Ratio Lower Upper

119 100

91 59.1 29.5 88.6

134 23.3 11.7 35.0



#84

1,2,4-Trimethylbenzene

Concen: 50.345 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. 0.000 min

Lab File: VX046720.D




Acq: 17 Jun 2025 14:20


Tgt Ion:105 Resp: 272510

Ion Ratio Lower Upper

105 100

120 44.2 22.1 66.3









#85
 sec-Butylbenzene
 Concen: 50.574 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. 0.000 min
 Lab File: VX046720.D
 Acq: 17 Jun 2025 14:20




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050


Tgt Ion:105 Resp: 355947
 Ion Ratio Lower Upper
 105 100
 134 19.6 9.8 29.4



#86
 p-Isopropyltoluene
 Concen: 50.067 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. 0.000 min
 Lab File: VX046720.D
 Acq: 17 Jun 2025 14:20

Tgt Ion:119 Resp: 297314
 Ion Ratio Lower Upper
 119 100
 134 25.8 12.9 38.7
 91 24.6 12.3 36.9









#87
1,3-Dichlorobenzene
Concen: 48.568 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:146 Resp: 149402
Ion Ratio Lower Upper
146 100
111 41.2 20.6 61.8
148 63.2 31.6 94.8



#88
1,4-Dichlorobenzene
Concen: 46.571 ug/l
RT: 12.037 min Scan# 1796
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Tgt Ion:146 Resp: 147146
Ion Ratio Lower Upper
146 100
111 40.7 20.3 61.0
148 64.3 32.1 96.5





#89

n-Butylbenzene

Concen: 49.801 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. 0.000 min

Lab File: VX046720.D

Acq: 17 Jun 2025 14:20

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion: 91 Resp: 280794

Ion Ratio Lower Upper

91 100

92 54.2 27.1 81.3

134 25.7 12.9 38.6



#90

Hexachloroethane

Concen: 49.968 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. 0.000 min

Lab File: VX046720.D




Acq: 17 Jun 2025 14:20


Tgt Ion: 117 Resp: 51961

Ion Ratio Lower Upper

117 100

201 68.9 34.4 103.3









#91
 1,2-Dichlorobenzene
 Concen: 49.309 ug/l
 RT: 12.329 min Scan# 1844
 Delta R.T. 0.000 min
 Lab File: VX046720.D
 Acq: 17 Jun 2025 14:20




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050


Tgt Ion:146 Resp: 141743
 Ion Ratio Lower Upper
 146 100
 111 43.4 21.7 65.1
 148 64.1 32.0 96.2



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 51.873 ug/l
 RT: 12.939 min Scan# 1944
 Delta R.T. 0.000 min
 Lab File: VX046720.D
 Acq: 17 Jun 2025 14:20

Tgt Ion: 75 Resp: 18747
 Ion Ratio Lower Upper
 75 100
 155 83.8 41.9 125.7
 157 102.9 51.4 154.3









#93
 1,2,4-Trichlorobenzene
 Concen: 49.101 ug/l
 RT: 13.585 min Scan# 2050
 Delta R.T. 0.000 min
 Lab File: VX046720.D
 Acq: 17 Jun 2025 14:20




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050


Tgt Ion:180 Resp: 100346
 Ion Ratio Lower Upper
 180 100
 182 93.4 46.7 140.1
 145 31.6 15.8 47.4



#94
 Hexachlorobutadiene
 Concen: 49.621 ug/l
 RT: 13.719 min Scan# 2072
 Delta R.T. 0.000 min
 Lab File: VX046720.D
 Acq: 17 Jun 2025 14:20

Tgt Ion:225 Resp: 42659
 Ion Ratio Lower Upper
 225 100
 223 65.0 32.5 97.5
 227 64.4 32.2 96.6









#95
Naphthalene
Concen: 50.994 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion:128 Resp: 296912
Ion Ratio Lower Upper
128 100
127 12.6 10.1 15.1
129 10.9 8.7 13.1



#96
1,2,3-Trichlorobenzene
Concen: 49.263 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. 0.000 min
Lab File: VX046720.D
Acq: 17 Jun 2025 14:20

Tgt Ion:180 Resp: 96318
Ion Ratio Lower Upper
180 100
182 95.6 47.8 143.4
145 34.6 17.3 51.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
 Data File : VX046721.D
 Acq On : 17 Jun 2025 14:41
 Operator : JC/MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC100

Quant Time: Jun 18 02:40:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 02:34:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	110482	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	183300	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	160869	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	78522	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	160514	105.037	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 210.080%#	
35) Dibromofluoromethane	5.391	113	129932	107.136	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 214.280%#	
50) Toluene-d8	8.647	98	458174	105.153	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 210.300%#	
62) 4-Bromofluorobenzene	11.079	95	170795	105.156	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 210.320%#	

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.179	85	138589	117.494	ug/l	99
3) Chloromethane	1.307	50	141568	111.195	ug/l	99
4) Vinyl Chloride	1.386	62	151879	111.868	ug/l	97
5) Bromomethane	1.611	94	75361	98.660	ug/l	94
6) Chloroethane	1.697	64	89532	108.805	ug/l	96
7) Trichlorofluoromethane	1.898	101	227688	110.449	ug/l	98
8) Diethyl Ether	2.148	74	77791	109.956	ug/l	96
9) 1,1,2-Trichlorotrifluo...	2.343	101	137759	108.929	ug/l	99
10) Methyl Iodide	2.465	142	174182	104.824	ug/l	99
11) Tert butyl alcohol	2.965	59	78366	571.345	ug/l	99
12) 1,1-Dichloroethene	2.331	96	134660	110.347	ug/l	98
13) Acrolein	2.246	56	108796	544.566	ug/l	99
14) Allyl chloride	2.678	41	243477	110.119	ug/l	100
15) Acrylonitrile	3.075	53	344981	559.348	ug/l	100
16) Acetone	2.386	43	263024	516.133	ug/l	99
17) Carbon Disulfide	2.526	76	383348	104.026	ug/l	98
18) Methyl Acetate	2.709	43	143551	110.909	ug/l	100
19) Methyl tert-butyl Ether	3.123	73	419445	110.387	ug/l	98
20) Methylene Chloride	2.800	84	147144	103.828	ug/l	98
21) trans-1,2-Dichloroethene	3.105	96	140792	107.824	ug/l	97
22) Diisopropyl ether	3.770	45	466831	109.846	ug/l	98
23) Vinyl Acetate	3.733	43	1930136	571.126	ug/l	99
24) 1,1-Dichloroethane	3.623	63	258546	107.149	ug/l	98
25) 2-Butanone	4.562	43	409960	568.810	ug/l	99
26) 2,2-Dichloropropane	4.489	77	194006	110.239	ug/l	99
27) cis-1,2-Dichloroethene	4.501	96	163672	106.691	ug/l	99
28) Bromochloromethane	4.910	49	114720	104.869	ug/l	99
29) Tetrahydrofuran	5.013	42	265233	569.746	ug/l	100
30) Chloroform	5.105	83	260888	108.113	ug/l	96
31) Cyclohexane	5.477	56	237424	103.757	ug/l	96
32) 1,1,1-Trichloroethane	5.391	97	231101	109.198	ug/l	99
36) 1,1-Dichloropropene	5.702	75	183355	104.733	ug/l	98
37) Ethyl Acetate	4.721	43	171444	106.094	ug/l	97
38) Carbon Tetrachloride	5.690	117	200914	105.749	ug/l	98
39) Methylcyclohexane	7.385	83	246277	106.902	ug/l	99
40) Benzene	6.044	78	553336	106.798	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
 Data File : VX046721.D
 Acq On : 17 Jun 2025 14:41
 Operator : JC/MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC100

Quant Time: Jun 18 02:40:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 02:34:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.928	41	97782	113.502	ug/l	97
42) 1,2-Dichloroethane	6.092	62	193498	106.278	ug/l	99
43) Isopropyl Acetate	6.342	43	291467	115.264	ug/l	100
44) Trichloroethene	7.129	130	142659	108.043	ug/l	96
45) 1,2-Dichloropropane	7.434	63	137291	106.920	ug/l	96
46) Dibromomethane	7.586	93	98964	107.240	ug/l	99
47) Bromodichloromethane	7.824	83	205870	109.319	ug/l	100
48) Methyl methacrylate	7.696	41	150002	118.279	ug/l	99
49) 1,4-Dioxane	7.659	88	38203	2040.447	ug/l	95
51) 4-Methyl-2-Pentanone	8.574	43	843944	559.584	ug/l	100
52) Toluene	8.720	92	339735	105.772	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	203470	116.448	ug/l	99
54) cis-1,3-Dichloropropene	8.366	75	221241	111.537	ug/l	98
55) 1,1,2-Trichloroethane	9.153	97	127195	106.264	ug/l	99
56) Ethyl methacrylate	9.116	69	205153	116.106	ug/l	100
57) 1,3-Dichloropropane	9.305	76	220021	106.740	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.244	63	541734	590.219	ug/l	100
59) 2-Hexanone	9.427	43	587221	563.200	ug/l	100
60) Dibromochloromethane	9.519	129	153731	108.983	ug/l	99
61) 1,2-Dibromoethane	9.610	107	133253	109.444	ug/l	99
64) Tetrachloroethene	9.275	164	115774	103.876	ug/l	98
65) Chlorobenzene	10.079	112	374913	105.577	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.159	131	131134	109.876	ug/l	99
67) Ethyl Benzene	10.189	91	663369	106.914	ug/l	100
68) m/p-Xylenes	10.299	106	492448	213.003	ug/l	100
69) o-Xylene	10.640	106	237631	109.675	ug/l	99
70) Styrene	10.653	104	407566	107.488	ug/l	99
71) Bromoform	10.799	173	101329	111.485	ug/l #	99
73) Isopropylbenzene	10.957	105	628803	108.731	ug/l	100
74) N-amyl acetate	10.842	43	260620	118.945	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	176523	109.317	ug/l	100
76) 1,2,3-Trichloropropane	11.238	75	194835	122.161	ug/l	92
77) Bromobenzene	11.195	156	150110	107.864	ug/l	99
78) n-propylbenzene	11.299	91	743979	108.310	ug/l	100
79) 2-Chlorotoluene	11.360	91	434989	106.065	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	517286	108.848	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	57608	113.225	ug/l	96
82) 4-Chlorotoluene	11.451	91	506852	105.896	ug/l	99
83) tert-Butylbenzene	11.713	119	533235	108.548	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	518524	108.598	ug/l	100
85) sec-Butylbenzene	11.890	105	670896	108.062	ug/l	99
86) p-Isopropyltoluene	12.006	119	561258	107.145	ug/l	100
87) 1,3-Dichlorobenzene	11.969	146	280540	103.388	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	280884	100.779	ug/l	99
89) n-Butylbenzene	12.329	91	536268	107.822	ug/l	100
90) Hexachloroethane	12.536	117	101647	110.811	ug/l	96
91) 1,2-Dichlorobenzene	12.335	146	267432	105.466	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	36858	115.616	ug/l	95
93) 1,2,4-Trichlorobenzene	13.585	180	196715	109.119	ug/l	99
94) Hexachlorobutadiene	13.725	225	80407	106.029	ug/l	98
95) Naphthalene	13.774	128	587914	114.468	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	189473	109.861	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
Data File : VX046721.D
Acq On : 17 Jun 2025 14:41
Operator : JC/MD
Sample : VSTDICC100
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Quant Time: Jun 18 02:40:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 02:34:13 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
 Data File : VX046721.D
 Acq On : 17 Jun 2025 14:41
 Operator : JC/MD
 Sample : VSTDIC100
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC100

Quant Time: Jun 18 02:40:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 02:34:13 2025
 Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion:168 Resp: 110482
Ion Ratio Lower Upper
168 100
99 61.4 48.5 72.7



#2
Dichlorodifluoromethane
Concen: 117.494 ug/l
RT: 1.179 min Scan# 15
Delta R.T. -0.006 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Tgt Ion: 85 Resp: 138589
Ion Ratio Lower Upper
85 100
87 32.6 16.6 49.8









#3
Chloromethane
Concen: 111.195 ug/l
RT: 1.307 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100


Tgt Ion: 50 Resp: 141568
Ion Ratio Lower Upper
50 100
52 31.8 24.9 37.3



#4
Vinyl Chloride
Concen: 111.868 ug/l
RT: 1.386 min Scan# 49
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Tgt Ion: 62 Resp: 151879
Ion Ratio Lower Upper
62 100
64 31.7 23.9 35.9









#5
Bromomethane
Concen: 98.660 ug/l
RT: 1.611 min Scan# 80
Delta R.T. -0.012 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 94 Resp: 75361
Ion Ratio Lower Upper
94 100
96 95.8 72.3 108.5



#6
Chloroethane
Concen: 108.805 ug/l
RT: 1.697 min Scan# 100
Delta R.T. -0.006 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Tgt Ion: 64 Resp: 89532
Ion Ratio Lower Upper
64 100
66 31.8 27.1 40.7









#7
 Trichlorofluoromethane
 Concen: 110.449 ug/l
 RT: 1.898 min Scan# 11
 Delta R.T. -0.006 min
 Lab File: VX046721.D
 Acq: 17 Jun 2025 14:41




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC100


Tgt Ion:101 Resp: 227688
 Ion Ratio Lower Upper
 101 100
 103 63.5 51.8 77.8



#8
 Diethyl Ether
 Concen: 109.956 ug/l
 RT: 2.148 min Scan# 174
 Delta R.T. -0.000 min
 Lab File: VX046721.D
 Acq: 17 Jun 2025 14:41

Tgt Ion: 74 Resp: 77791
 Ion Ratio Lower Upper
 74 100
 45 100.7 52.3 157.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 108.929 ug/l
RT: 2.343 min Scan# 206
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion:101 Resp: 137759
Ion Ratio Lower Upper
101 100
85 43.8 35.4 53.0
151 76.0 61.2 91.8



#10
Methyl Iodide
Concen: 104.824 ug/l
RT: 2.465 min Scan# 226
Delta R.T. -0.006 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Tgt Ion:142 Resp: 174182
Ion Ratio Lower Upper
142 100
127 45.7 37.0 55.6
141 14.0 11.5 17.3









#11
Tert butyl alcohol
Concen: 571.345 ug/l
RT: 2.965 min Scan# 307
Delta R.T. 0.006 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 59 Resp: 78366
Ion Ratio Lower Upper
59 100
57 10.0 8.3 12.5



#12
1,1-Dichloroethene
Concen: 110.347 ug/l
RT: 2.331 min Scan# 204
Delta R.T. -0.006 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Tgt Ion: 96 Resp: 134660
Ion Ratio Lower Upper
96 100
61 161.9 126.6 189.8
98 63.0 50.7 76.1









#13
Acrolein
Concen: 544.566 ug/l
RT: 2.246 min Scan# 190
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 56 Resp: 108796
Ion Ratio Lower Upper
56 100
55 69.7 55.2 82.8



#14
Allyl chloride
Concen: 110.119 ug/l
RT: 2.678 min Scan# 261
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Tgt Ion: 41 Resp: 243477
Ion Ratio Lower Upper
41 100
39 66.5 53.6 80.4
76 33.6 26.8 40.2









#15
Acrylonitrile
Concen: 559.348 ug/l
RT: 3.075 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100


Tgt Ion: 53 Resp: 344981
Ion Ratio Lower Upper
53 100
52 81.9 65.1 97.7
51 36.0 28.9 43.3



#16
Acetone
Concen: 516.133 ug/l
RT: 2.386 min Scan# 213
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Tgt Ion: 43 Resp: 263024
Ion Ratio Lower Upper
43 100
58 30.2 24.5 36.7









#17
Carbon Disulfide
Concen: 104.026 ug/l
RT: 2.526 min Scan# 21
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 76 Resp: 383348
Ion Ratio Lower Upper
76 100
78 8.8 7.5 11.3



#18
Methyl Acetate
Concen: 110.909 ug/l
RT: 2.709 min Scan# 266
Delta R.T. -0.006 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Tgt Ion: 43 Resp: 143551
Ion Ratio Lower Upper
43 100
74 23.2 18.4 27.6









#19
Methyl tert-butyl Ether
Concen: 110.387 ug/l
RT: 3.123 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 73 Resp: 419445
Ion Ratio Lower Upper
73 100
57 21.8 18.2 27.2



#20
Methylene Chloride
Concen: 103.828 ug/l
RT: 2.800 min Scan# 281
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Tgt Ion: 84 Resp: 147144
Ion Ratio Lower Upper
84 100
49 122.3 101.3 151.9
51 38.8 31.0 46.6
86 63.5 51.1 76.7









#21
trans-1,2-Dichloroethene
Concen: 107.824 ug/l
RT: 3.105 min Scan# 311
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 96 Resp: 140792
Ion Ratio Lower Upper
96 100
61 136.9 112.5 168.7
98 63.1 51.9 77.9



#22
Diisopropyl ether
Concen: 109.846 ug/l
RT: 3.770 min Scan# 440
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Tgt Ion: 45 Resp: 466831
Ion Ratio Lower Upper
45 100
43 112.3 91.9 137.9
87 26.8 22.0 33.0
59 11.4 9.1 13.7









#23
 Vinyl Acetate
 Concen: 571.126 ug/l
 RT: 3.733 min Scan# 41
 Delta R.T. -0.000 min
 Lab File: VX046721.D
 Acq: 17 Jun 2025 14:41




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC100


Tgt Ion: 43 Resp: 1930136
 Ion Ratio Lower Upper
 43 100
 86 10.4 8.2 12.2



#24
 1,1-Dichloroethane
 Concen: 107.149 ug/l
 RT: 3.623 min Scan# 416
 Delta R.T. -0.000 min
 Lab File: VX046721.D
 Acq: 17 Jun 2025 14:41

Tgt Ion: 63 Resp: 258546
 Ion Ratio Lower Upper
 63 100
 98 7.0 4.0 11.9
 100 4.6 2.2 6.6





#25

2-Butanone

Concen: 568.810 ug/l

RT: 4.562 min Scan# 5

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC100

Tgt Ion: 43 Resp: 409960

Ion Ratio Lower Upper

43 100

72 24.6 20.2 30.4



#26

2,2-Dichloropropane

Concen: 110.239 ug/l

RT: 4.489 min Scan# 558

Delta R.T. -0.000 min

Lab File: VX046721.D




Acq: 17 Jun 2025 14:41


Tgt Ion: 77 Resp: 194006

Ion Ratio Lower Upper

77 100

97 22.8 11.8 35.3





#27

cis-1,2-Dichloroethene

Concen: 106.691 ug/l

RT: 4.501 min Scan# 560

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100





Tgt Ion: 96 Resp: 163672

Ion Ratio Lower Upper

96 100

61 144.8 0.0 291.0

98 64.6 0.0 126.8



#28

Bromochloromethane

Concen: 104.869 ug/l

RT: 4.910 min Scan# 627

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41




Tgt Ion: 49 Resp: 114720


Ion Ratio Lower Upper

49 100

129 1.9 0.0 3.8

130 77.6 61.4 92.0





#29

Tetrahydrofuran

Concen: 569.746 ug/l

RT: 5.013 min Scan# 644

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100





Tgt Ion: 42 Resp: 265233

Ion Ratio Lower Upper

42 100

72 43.5 34.9 52.3

71 40.2 32.2 48.2



#30

Chloroform

Concen: 108.113 ug/l

RT: 5.105 min Scan# 659

Delta R.T. -0.000 min

Lab File: VX046721.D




Acq: 17 Jun 2025 14:41


Tgt Ion: 83 Resp: 260888

Ion Ratio Lower Upper

83 100

85 63.6 53.7 80.5









#31
Cyclohexane
Concen: 103.757 ug/l
RT: 5.477 min Scan# 719
Delta R.T. -0.006 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 56 Resp: 237424
Ion Ratio Lower Upper
56 100
69 28.9 24.8 37.2
84 81.9 68.2 102.2



#32
1,1,1-Trichloroethane
Concen: 109.198 ug/l
RT: 5.391 min Scan# 706
Delta R.T. -0.006 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Tgt Ion: 97 Resp: 231101
Ion Ratio Lower Upper
97 100
99 63.5 51.3 76.9
61 44.1 36.4 54.6





#33

1,2-Dichloroethane-d4

Concen: 105.037 ug/l

RT: 5.958 min Scan# 799

Delta R.T. -0.006 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC100

Tgt Ion: 65 Resp: 160514

Ion Ratio Lower Upper

65 100

67 52.1 0.0 105.4



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. -0.006 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41




Tgt Ion: 114 Resp: 183300


Ion Ratio Lower Upper

114 100

63 21.0 0.0 44.2

88 16.4 0.0 33.2





#35

Dibromofluoromethane

Concen: 107.136 ug/l

RT: 5.391 min Scan# 707

Delta R.T. -0.006 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100





Tgt Ion:113 Resp: 129932

Ion Ratio Lower Upper

113 100

111 102.2 82.0 123.0

192 18.7 15.3 22.9



#36

1,1-Dichloropropene

Concen: 104.733 ug/l

RT: 5.702 min Scan# 757

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41




Tgt Ion: 75 Resp: 183355


Ion Ratio Lower Upper

75 100

110 35.3 17.2 51.5

77 31.2 24.2 36.4









#37
Ethyl Acetate
Concen: 106.094 ug/l
RT: 4.721 min Scan# 596
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 43 Resp: 171444
Ion Ratio Lower Upper
43 100
61 13.8 9.7 14.5
70 10.4 8.1 12.1



#38
Carbon Tetrachloride
Concen: 105.749 ug/l
RT: 5.690 min Scan# 755
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Tgt Ion: 117 Resp: 200914
Ion Ratio Lower Upper
117 100
119 96.0 75.2 112.8
121 30.8 24.1 36.1





#39

Methylcyclohexane

Concen: 106.902 ug/l

RT: 7.385 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100





Tgt Ion: 83 Resp: 246277

Ion Ratio Lower Upper

83 100

55 77.9 63.0 94.4

98 48.8 38.8 58.2



#40

Benzene

Concen: 106.798 ug/l

RT: 6.044 min Scan# 813

Delta R.T. -0.000 min

Lab File: VX046721.D




Acq: 17 Jun 2025 14:41


Tgt Ion: 78 Resp: 553336

Ion Ratio Lower Upper

78 100

77 23.8 18.8 28.2









#41
Methacrylonitrile
Concen: 113.502 ug/l
RT: 4.928 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion:	41	Resp:	97782
Ion	Ratio	Lower	Upper
41	100		
39	54.8	45.9	68.9
67	73.6	61.2	91.8
52	31.6	25.5	38.3



#42
1,2-Dichloroethane
Concen: 106.278 ug/l
RT: 6.092 min Scan# 821
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Tgt Ion:	62	Resp:	193498
Ion	Ratio	Lower	Upper
62	100		
98	9.1	0.0	17.8





#43

Isopropyl Acetate

Concen: 115.264 ug/l

RT: 6.342 min Scan# 862

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100





Tgt Ion: 43 Resp: 291467

Ion Ratio Lower Upper

43 100

61 19.4 15.3 22.9

87 12.5 10.0 15.0



#44

Trichloroethene

Concen: 108.043 ug/l

RT: 7.129 min Scan# 991

Delta R.T. -0.000 min

Lab File: VX046721.D




Acq: 17 Jun 2025 14:41


Tgt Ion:130 Resp: 142659

Ion Ratio Lower Upper

130 100

95 97.1 0.0 202.6





#45

1,2-Dichloropropane

Concen: 106.920 ug/l

RT: 7.434 min Scan# 1041

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC100

Tgt Ion: 63 Resp: 137291

Ion Ratio Lower Upper

63 100

65 30.8 22.9 34.3



#46

Dibromomethane

Concen: 107.240 ug/l

RT: 7.586 min Scan# 1066

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41




Tgt Ion: 93 Resp: 98964


Ion Ratio Lower Upper

93 100

95 82.5 66.2 99.4

174 97.4 77.0 115.6









#47
Bromodichloromethane
Concen: 109.319 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 83 Resp: 205870
Ion Ratio Lower Upper
83 100
85 64.4 51.4 77.2
127 8.8 6.9 10.3



#48
Methyl methacrylate
Concen: 118.279 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Tgt Ion: 41 Resp: 150002
Ion Ratio Lower Upper
41 100
69 80.4 63.4 95.2
39 56.2 45.4 68.0









#49
1,4-Dioxane
Concen: 2040.447 ug/l
RT: 7.659 min Scan# 1078
Delta R.T. 0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 88 Resp: 38203
Ion Ratio Lower Upper
88 100
43 37.8 31.2 46.8
58 73.7 54.7 82.1



#50
Toluene-d8
Concen: 105.153 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Tgt Ion: 98 Resp: 458174
Ion Ratio Lower Upper
98 100
100 66.7 53.0 79.4









#51
4-Methyl-2-Pentanone
Concen: 559.584 ug/l
RT: 8.574 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100


Tgt Ion: 43 Resp: 843944
Ion Ratio Lower Upper
43 100
58 38.5 30.8 46.2



#52
Toluene
Concen: 105.772 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Tgt Ion: 92 Resp: 339735
Ion Ratio Lower Upper
92 100
91 170.4 134.8 202.2





#53

t-1,3-Dichloropropene

Concen: 116.448 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC100

Tgt Ion: 75 Resp: 203470

Ion Ratio Lower Upper

75 100

77 31.6 25.6 38.4



#54

cis-1,3-Dichloropropene

Concen: 111.537 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41




Tgt Ion: 75 Resp: 221241


Ion Ratio Lower Upper

75 100

77 32.0 25.3 37.9

39 51.0 39.2 58.8









#55
1,1,2-Trichloroethane
Concen: 106.264 ug/l
RT: 9.153 min Scan# 1323
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100


Tgt Ion:	97	Resp:	127195
Ion	Ratio	Lower	Upper
97	100		
83	84.7	67.2	100.8
85	55.4	44.1	66.1
99	62.7	50.3	75.5



#56
Ethyl methacrylate
Concen: 116.106 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Tgt Ion:	69	Resp:	205153
Ion	Ratio	Lower	Upper
69	100		
41	69.6	55.4	83.2
39	36.9	29.7	44.5





#57

1,3-Dichloropropane

Concen: 106.740 ug/l

RT: 9.305 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC100

Tgt Ion: 76 Resp: 220021

Ion Ratio Lower Upper

76 100

78 32.1 25.7 38.5



#58

2-Chloroethyl Vinyl ether

Concen: 590.219 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. -0.000 min

Lab File: VX046721.D




Acq: 17 Jun 2025 14:41


Tgt Ion: 63 Resp: 541734

Ion Ratio Lower Upper

63 100

106 28.2 22.6 34.0









#59
2-Hexanone
Concen: 563.200 ug/l
RT: 9.427 min Scan# 1368
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 43 Resp: 587221
Ion Ratio Lower Upper
43 100
58 54.0 26.8 80.4



#60
Dibromochloromethane
Concen: 108.983 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Tgt Ion: 129 Resp: 153731
Ion Ratio Lower Upper
129 100
127 77.6 38.4 115.2





#61

1,2-Dibromoethane

Concen: 109.444 ug/l

RT: 9.610 min Scan# 1398

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC100

Tgt Ion:107 Resp: 133253

Ion Ratio Lower Upper

107 100

109 93.8 75.8 113.6



#62

4-Bromofluorobenzene

Concen: 105.156 ug/l

RT: 11.079 min Scan# 1639

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41




Tgt Ion: 95 Resp: 170795


Ion Ratio Lower Upper

95 100

174 76.4 0.0 150.4

176 72.7 0.0 145.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion:117 Resp: 160869
Ion Ratio Lower Upper
117 100
82 54.4 44.6 66.8
119 32.1 25.8 38.8



#64
Tetrachloroethene
Concen: 103.876 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Tgt Ion:164 Resp: 115774
Ion Ratio Lower Upper
164 100
166 128.0 104.6 157.0
129 96.2 76.3 114.5
131 91.4 74.1 111.1









#65
Chlorobenzene
Concen: 105.577 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion:112 Resp: 374913
Ion Ratio Lower Upper
112 100
114 31.7 25.5 38.3



#66
1,1,1,2-Tetrachloroethane
Concen: 109.876 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Tgt Ion:131 Resp: 131134
Ion Ratio Lower Upper
131 100
133 93.8 47.3 141.8
119 63.0 32.1 96.3





#67

Ethyl Benzene

Concen: 106.914 ug/l

RT: 10.189 min Scan# 1493

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC100

Tgt Ion: 91 Resp: 663369

Ion Ratio Lower Upper

91 100

106 30.4 24.2 36.2



#68

m/p-Xylenes

Concen: 213.003 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. -0.000 min

Lab File: VX046721.D




Acq: 17 Jun 2025 14:41


Tgt Ion:106 Resp: 492448

Ion Ratio Lower Upper

106 100

91 206.0 164.8 247.2









#69
o-Xylene
Concen: 109.675 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion:106 Resp: 237631
Ion Ratio Lower Upper
106 100
91 220.5 111.3 333.9



#70
Styrene
Concen: 107.488 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Tgt Ion:104 Resp: 407566
Ion Ratio Lower Upper
104 100
78 52.3 41.4 62.2
103 54.3 43.1 64.7









#71
Bromoform
Concen: 111.485 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion:173 Resp: 101329
Ion Ratio Lower Upper
173 100
175 48.9 24.1 72.4
254 0.1 0.1 0.1#



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Tgt Ion:152 Resp: 78522
Ion Ratio Lower Upper
152 100
115 115.4 43.2 129.6
150 191.6 0.0 346.8





#73

Isopropylbenzene

Concen: 108.731 ug/l

RT: 10.957 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC100

Tgt Ion:105 Resp: 628803

Ion Ratio Lower Upper

105 100

120 26.1 13.0 39.0



#74

N-ethyl acetate

Concen: 118.945 ug/l

RT: 10.842 min Scan# 1600

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41

Tgt Ion: 43 Resp: 260620




Ion Ratio Lower Upper


43 100

70 40.8 33.4 50.0

55 23.2 19.1 28.7

61 24.1 19.0 28.4









#75
1,1,2,2-Tetrachloroethane
Concen: 109.317 ug/l
RT: 11.207 min Scan# 1660
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100


Tgt Ion: 83 Resp: 176523
Ion Ratio Lower Upper
83 100
131 10.3 4.9 14.6
85 63.8 32.0 96.0



#76
1,2,3-Trichloropropane
Concen: 122.161 ug/l
RT: 11.238 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Tgt Ion: 75 Resp: 194835
Ion Ratio Lower Upper
75 100
77 32.2 18.5 55.5





#77

Bromobenzene

Concen: 107.864 ug/l

RT: 11.195 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100





Tgt Ion:156 Resp: 150110

Ion Ratio Lower Upper

156 100

77 166.5 82.8 248.6

158 97.7 47.6 142.9



#78

n-propylbenzene

Concen: 108.310 ug/l

RT: 11.299 min Scan# 1675

Delta R.T. -0.000 min

Lab File: VX046721.D




Acq: 17 Jun 2025 14:41


Tgt Ion: 91 Resp: 743979

Ion Ratio Lower Upper

91 100

120 22.6 11.4 34.1









#79
2-Chlorotoluene
Concen: 106.065 ug/l
RT: 11.360 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 91 Resp: 434989
Ion Ratio Lower Upper
91 100
126 33.2 16.6 49.7



#80
1,3,5-Trimethylbenzene
Concen: 108.848 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Tgt Ion: 105 Resp: 517286
Ion Ratio Lower Upper
105 100
120 48.0 24.1 72.2









#81
trans-1,4-Dichloro-2-butene
Concen: 113.225 ug/l
RT: 11.018 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 75 Resp: 57608
Ion Ratio Lower Upper
75 100
53 103.4 86.8 130.2
89 46.9 36.9 55.3



#82
4-Chlorotoluene
Concen: 105.896 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046721.D
Acq: 17 Jun 2025 14:41

Tgt Ion: 91 Resp: 506852
Ion Ratio Lower Upper
91 100
126 29.1 14.4 43.2









#83
 tert-Butylbenzene
 Concen: 108.548 ug/l
 RT: 11.713 min Scan# 1743
 Delta R.T. -0.000 min
 Lab File: VX046721.D
 Acq: 17 Jun 2025 14:41




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC100


Tgt Ion:119 Resp: 533235
 Ion Ratio Lower Upper
 119 100
 91 59.1 29.5 88.6
 134 22.7 11.7 35.0



#84
 1,2,4-Trimethylbenzene
 Concen: 108.598 ug/l
 RT: 11.750 min Scan# 1749
 Delta R.T. -0.000 min
 Lab File: VX046721.D
 Acq: 17 Jun 2025 14:41

Tgt Ion:105 Resp: 518524
 Ion Ratio Lower Upper
 105 100
 120 44.0 22.1 66.3









#85
 sec-Butylbenzene
 Concen: 108.062 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. -0.000 min
 Lab File: VX046721.D
 Acq: 17 Jun 2025 14:41




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC100


Tgt Ion:105 Resp: 670896
 Ion Ratio Lower Upper
 105 100
 134 20.0 9.8 29.4



#86
 p-Isopropyltoluene
 Concen: 107.145 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. -0.000 min
 Lab File: VX046721.D
 Acq: 17 Jun 2025 14:41

Tgt Ion:119 Resp: 561258
 Ion Ratio Lower Upper
 119 100
 134 25.7 12.9 38.7
 91 24.7 12.3 36.9





#87

1,3-Dichlorobenzene

Concen: 103.388 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100





Tgt Ion:146 Resp: 280540

Ion Ratio Lower Upper

146 100

111 41.7 20.6 61.8

148 63.9 31.6 94.8



#88

1,4-Dichlorobenzene

Concen: 100.779 ug/l

RT: 12.036 min Scan# 1796

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41




Tgt Ion:146 Resp: 280884


Ion Ratio Lower Upper

146 100

111 39.5 20.3 61.0

148 63.9 32.1 96.5





#89

n-Butylbenzene

Concen: 107.822 ug/l

RT: 12.329 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100





Tgt Ion: 91 Resp: 536268

Ion Ratio Lower Upper

91 100

92 54.5 27.1 81.3

134 25.6 12.9 38.6



#90

Hexachloroethane

Concen: 110.811 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX046721.D




Acq: 17 Jun 2025 14:41


Tgt Ion: 117 Resp: 101647

Ion Ratio Lower Upper

117 100

201 72.5 34.4 103.3





#91

1,2-Dichlorobenzene

Concen: 105.466 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.006 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100





Tgt Ion:146 Resp: 267432

Ion Ratio Lower Upper

146 100

111 43.2 21.7 65.1

148 63.5 32.0 96.2



#92

1,2-Dibromo-3-Chloropropane

Concen: 115.616 ug/l

RT: 12.939 min Scan# 1944

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41




Tgt Ion: 75 Resp: 36858


Ion Ratio Lower Upper

75 100

155 85.8 41.9 125.7

157 110.6 51.4 154.3





#93

1,2,4-Trichlorobenzene

Concen: 109.119 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100





Tgt Ion:180 Resp: 196715

Ion Ratio Lower Upper

180 100

182 94.3 46.7 140.1

145 32.0 15.8 47.4



#94

Hexachlorobutadiene

Concen: 106.029 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. 0.006 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41




Tgt Ion:225 Resp: 80407


Ion Ratio Lower Upper

225 100

223 62.8 32.5 97.5

227 63.0 32.2 96.6





#95

Naphthalene

Concen: 114.468 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100





Tgt Ion:128 Resp: 587914

Ion Ratio Lower Upper

128 100

127 12.7 10.1 15.1

129 10.9 8.7 13.1



#96

1,2,3-Trichlorobenzene

Concen: 109.861 ug/l

RT: 13.957 min Scan# 2111

Delta R.T. -0.000 min

Lab File: VX046721.D

Acq: 17 Jun 2025 14:41




Tgt Ion:180 Resp: 189473

Ion Ratio Lower Upper

180 100

182 96.0 47.8 143.4

145 34.4 17.3 51.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
 Data File : VX046722.D
 Acq On : 17 Jun 2025 15:02
 Operator : JC/MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC150

Quant Time: Jun 18 02:41:13 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 02:34:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	107848	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	178346	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	156970	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	75829	50.000	ug/l	# 0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	230915	154.797	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	309.600%#
35) Dibromofluoromethane	5.397	113	187536	158.930	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	317.860%#
50) Toluene-d8	8.653	98	660221	155.733	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	311.460%#
62) 4-Bromofluorobenzene	11.079	95	244007	154.404	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	308.800%#
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.179	85	193321	167.898	ug/l	99
3) Chloromethane	1.307	50	202755	163.144	ug/l	97
4) Vinyl Chloride	1.386	62	208285	157.161	ug/l	96
5) Bromomethane	1.612	94	90708	121.652	ug/l	97
6) Chloroethane	1.691	64	123323	153.530	ug/l	96
7) Trichlorofluoromethane	1.898	101	314552	156.313	ug/l	98
8) Diethyl Ether	2.142	74	108801	157.544	ug/l	97
9) 1,1,2-Trichlorotrifluo...	2.343	101	192071	155.583	ug/l	99
10) Methyl Iodide	2.465	142	243423	148.201	ug/l	99
11) Tert butyl alcohol	2.971	59	111150	830.155	ug/l	99
12) 1,1-Dichloroethene	2.331	96	188770	158.466	ug/l	97
13) Acrolein	2.246	56	156844	804.238	ug/l	99
14) Allyl chloride	2.678	41	340117	157.583	ug/l	99
15) Acrylonitrile	3.075	53	477449	793.036	ug/l	99
16) Acetone	2.386	43	378554	760.980	ug/l	99
17) Carbon Disulfide	2.526	76	535657	148.907	ug/l	99
18) Methyl Acetate	2.715	43	209349	165.695	ug/l	99
19) Methyl tert-butyl Ether	3.123	73	585091	157.742	ug/l	99
20) Methylene Chloride	2.800	84	201983	146.004	ug/l	99
21) trans-1,2-Dichloroethene	3.105	96	192490	151.016	ug/l	99
22) Diisopropyl ether	3.770	45	638477	153.904	ug/l	99
23) Vinyl Acetate	3.733	43	2691267	815.794	ug/l	99
24) 1,1-Dichloroethane	3.623	63	360395	153.006	ug/l	98
25) 2-Butanone	4.562	43	570672	811.132	ug/l	99
26) 2,2-Dichloropropane	4.483	77	286713	166.896	ug/l	99
27) cis-1,2-Dichloroethene	4.495	96	227788	152.112	ug/l	99
28) Bromochloromethane	4.910	49	161801	151.520	ug/l	100
29) Tetrahydrofuran	5.013	42	370760	815.879	ug/l	100
30) Chloroform	5.105	83	358782	152.312	ug/l	99
31) Cyclohexane	5.483	56	330235	147.842	ug/l	98
32) 1,1,1-Trichloroethane	5.397	97	320382	155.082	ug/l	99
36) 1,1-Dichloropropene	5.702	75	256619	150.653	ug/l	99
37) Ethyl Acetate	4.721	43	245724	156.284	ug/l	99
38) Carbon Tetrachloride	5.684	117	284066	153.668	ug/l	99
39) Methylcyclohexane	7.385	83	346118	154.413	ug/l	100
40) Benzene	6.044	78	763341	151.423	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
 Data File : VX046722.D
 Acq On : 17 Jun 2025 15:02
 Operator : JC/MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC150

Quant Time: Jun 18 02:41:13 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 02:34:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.928	41	142283	169.745	ug/l	94
42) 1,2-Dichloroethane	6.092	62	265740	150.011	ug/l	99
43) Isopropyl Acetate	6.342	43	412067	167.483	ug/l	99
44) Trichloroethene	7.129	130	195943	152.520	ug/l	99
45) 1,2-Dichloropropane	7.434	63	191181	153.025	ug/l	95
46) Dibromomethane	7.580	93	138541	154.297	ug/l	99
47) Bromodichloromethane	7.824	83	288826	157.630	ug/l	100
48) Methyl methacrylate	7.696	41	211908	171.734	ug/l	99
49) 1,4-Dioxane	7.659	88	54646	2987.386	ug/l	96
51) 4-Methyl-2-Pentanone	8.574	43	1173666	799.825	ug/l	100
52) Toluene	8.720	92	471273	150.800	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	288893	169.929	ug/l	99
54) cis-1,3-Dichloropropene	8.366	75	314965	163.198	ug/l	98
55) 1,1,2-Trichloroethane	9.153	97	175973	151.099	ug/l	99
56) Ethyl methacrylate	9.116	69	288720	167.940	ug/l	99
57) 1,3-Dichloropropane	9.305	76	301418	150.290	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.244	63	754730	845.118	ug/l	100
59) 2-Hexanone	9.433	43	816655	805.005	ug/l	99
60) Dibromochloromethane	9.519	129	215347	156.904	ug/l	99
61) 1,2-Dibromoethane	9.610	107	185268	156.392	ug/l	99
64) Tetrachloroethene	9.275	164	159752	146.895	ug/l	98
65) Chlorobenzene	10.080	112	517859	149.453	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.165	131	181981	156.268	ug/l	99
67) Ethyl Benzene	10.195	91	915724	151.251	ug/l	98
68) m/p-Xylenes	10.299	106	676085	299.697	ug/l	99
69) o-Xylene	10.640	106	328782	155.514	ug/l	98
70) Styrene	10.653	104	566606	153.143	ug/l	100
71) Bromoform	10.799	173	143229	161.498	ug/l	100
73) Isopropylbenzene	10.964	105	867728	155.374	ug/l	100
74) N-amyl acetate	10.842	43	369757	174.748	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	244406	156.730	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	277006	179.850	ug/l	91
77) Bromobenzene	11.195	156	207348	154.285	ug/l	98
78) n-propylbenzene	11.305	91	1025742	154.633	ug/l	100
79) 2-Chlorotoluene	11.366	91	595555	150.373	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	707005	154.053	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	86677	176.408	ug/l	93
82) 4-Chlorotoluene	11.451	91	697556	150.915	ug/l	99
83) tert-Butylbenzene	11.713	119	735750	155.093	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	719283	155.994	ug/l	100
85) sec-Butylbenzene	11.890	105	916769	152.910	ug/l	99
86) p-Isopropyltoluene	12.006	119	771998	152.610	ug/l	100
87) 1,3-Dichlorobenzene	11.969	146	391049	149.232	ug/l	100
88) 1,4-Dichlorobenzene	12.036	146	387139	143.836	ug/l	99
89) n-Butylbenzene	12.329	91	726563	151.271	ug/l	100
90) Hexachloroethane	12.536	117	142663	161.049	ug/l	98
91) 1,2-Dichlorobenzene	12.329	146	370234	151.192	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	12.939	75	54014	175.448	ug/l	97
93) 1,2,4-Trichlorobenzene	13.585	180	263803	151.531	ug/l	97
94) Hexachlorobutadiene	13.725	225	95520	130.431	ug/l	100
95) Naphthalene	13.774	128	846206	170.609	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	262348	157.517	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
Data File : VX046722.D
Acq On : 17 Jun 2025 15:02
Operator : JC/MD
Sample : VSTDICC150
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Quant Time: Jun 18 02:41:13 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 02:34:13 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
 Data File : VX046722.D
 Acq On : 17 Jun 2025 15:02
 Operator : JC/MD
 Sample : VSTDIC150
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC150

Quant Time: Jun 18 02:41:13 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 02:34:13 2025
 Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.562 min Scan# 71
Delta R.T. 0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150


Tgt Ion:168 Resp: 107848
Ion Ratio Lower Upper
168 100
99 59.3 48.5 72.7



#2
Dichlorodifluoromethane
Concen: 167.898 ug/l
RT: 1.179 min Scan# 15
Delta R.T. -0.006 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion: 85 Resp: 193321
Ion Ratio Lower Upper
85 100
87 32.6 16.6 49.8









#3
Chloromethane
Concen: 163.144 ug/l
RT: 1.307 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 50 Resp: 202755
Ion Ratio Lower Upper
50 100
52 32.6 24.9 37.3



#4
Vinyl Chloride
Concen: 157.161 ug/l
RT: 1.386 min Scan# 49
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion: 62 Resp: 208285
Ion Ratio Lower Upper
62 100
64 31.9 23.9 35.9









#5
Bromomethane
Concen: 121.652 ug/l
RT: 1.612 min Scan# 80
Delta R.T. -0.012 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 94 Resp: 90708
Ion Ratio Lower Upper
94 100
96 92.9 72.3 108.5



#6
Chloroethane
Concen: 153.530 ug/l
RT: 1.691 min Scan# 99
Delta R.T. -0.012 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion: 64 Resp: 123323
Ion Ratio Lower Upper
64 100
66 31.7 27.1 40.7









#7
 Trichlorofluoromethane
 Concen: 156.313 ug/l
 RT: 1.898 min Scan# 11
 Delta R.T. -0.006 min
 Lab File: VX046722.D
 Acq: 17 Jun 2025 15:02




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC150


Tgt Ion:101 Resp: 314552
 Ion Ratio Lower Upper
 101 100
 103 63.4 51.8 77.8



#8
 Diethyl Ether
 Concen: 157.544 ug/l
 RT: 2.142 min Scan# 173
 Delta R.T. -0.006 min
 Lab File: VX046722.D
 Acq: 17 Jun 2025 15:02

Tgt Ion: 74 Resp: 108801
 Ion Ratio Lower Upper
 74 100
 45 101.4 52.3 157.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 155.583 ug/l
RT: 2.343 min Scan# 206
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150


Tgt Ion:101 Resp: 192071
Ion Ratio Lower Upper
101 100
85 43.7 35.4 53.0
151 75.4 61.2 91.8



#10
Methyl Iodide
Concen: 148.201 ug/l
RT: 2.465 min Scan# 226
Delta R.T. -0.006 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion:142 Resp: 243423
Ion Ratio Lower Upper
142 100
127 45.5 37.0 55.6
141 13.6 11.5 17.3









#11
Tert butyl alcohol
Concen: 830.155 ug/l
RT: 2.971 min Scan# 307
Delta R.T. 0.012 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 59 Resp: 111150
Ion Ratio Lower Upper
59 100
57 10.7 8.3 12.5



#12
1,1-Dichloroethene
Concen: 158.466 ug/l
RT: 2.331 min Scan# 204
Delta R.T. -0.006 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion: 96 Resp: 188770
Ion Ratio Lower Upper
96 100
61 163.4 126.6 189.8
98 63.6 50.7 76.1









#13
Acrolein
Concen: 804.238 ug/l
RT: 2.246 min Scan# 190
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 56 Resp: 156844
Ion Ratio Lower Upper
56 100
55 69.8 55.2 82.8



#14
Allyl chloride
Concen: 157.583 ug/l
RT: 2.678 min Scan# 261
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion: 41 Resp: 340117
Ion Ratio Lower Upper
41 100
39 65.9 53.6 80.4
76 33.6 26.8 40.2









#15
Acrylonitrile
Concen: 793.036 ug/l
RT: 3.075 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 53 Resp: 477449
Ion Ratio Lower Upper
53 100
52 82.8 65.1 97.7
51 36.3 28.9 43.3



#16
Acetone
Concen: 760.980 ug/l
RT: 2.386 min Scan# 213
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion: 43 Resp: 378554
Ion Ratio Lower Upper
43 100
58 30.2 24.5 36.7









#17
Carbon Disulfide
Concen: 148.907 ug/l
RT: 2.526 min Scan# 21
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 76 Resp: 535657
Ion Ratio Lower Upper
76 100
78 9.0 7.5 11.3



#18
Methyl Acetate
Concen: 165.695 ug/l
RT: 2.715 min Scan# 267
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion: 43 Resp: 209349
Ion Ratio Lower Upper
43 100
74 22.4 18.4 27.6









#19
Methyl tert-butyl Ether
Concen: 157.742 ug/l
RT: 3.123 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 73 Resp: 585091
Ion Ratio Lower Upper
73 100
57 22.1 18.2 27.2



#20
Methylene Chloride
Concen: 146.004 ug/l
RT: 2.800 min Scan# 281
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion: 84 Resp: 201983
Ion Ratio Lower Upper
84 100
49 126.6 101.3 151.9
51 39.0 31.0 46.6
86 66.5 51.1 76.7









#21
trans-1,2-Dichloroethene
Concen: 151.016 ug/l
RT: 3.105 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 96 Resp: 192490
Ion Ratio Lower Upper
96 100
61 139.6 112.5 168.7
98 62.5 51.9 77.9



#22
Diisopropyl ether
Concen: 153.904 ug/l
RT: 3.770 min Scan# 440
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion: 45 Resp: 638477
Ion Ratio Lower Upper
45 100
43 115.7 91.9 137.9
87 26.6 22.0 33.0
59 11.1 9.1 13.7









#23
 Vinyl Acetate
 Concen: 815.794 ug/l
 RT: 3.733 min Scan# 41
 Delta R.T. -0.000 min
 Lab File: VX046722.D
 Acq: 17 Jun 2025 15:02




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC150


Tgt Ion: 43 Resp: 2691267
 Ion Ratio Lower Upper
 43 100
 86 10.4 8.2 12.2



#24
 1,1-Dichloroethane
 Concen: 153.006 ug/l
 RT: 3.623 min Scan# 416
 Delta R.T. -0.000 min
 Lab File: VX046722.D
 Acq: 17 Jun 2025 15:02

Tgt Ion: 63 Resp: 360395
 Ion Ratio Lower Upper
 63 100
 98 7.1 4.0 11.9
 100 4.3 2.2 6.6









#25
2-Butanone
Concen: 811.132 ug/l
RT: 4.562 min Scan# 570
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 43 Resp: 570672
Ion Ratio Lower Upper
43 100
72 24.7 20.2 30.4



#26
2,2-Dichloropropane
Concen: 166.896 ug/l
RT: 4.483 min Scan# 557
Delta R.T. -0.006 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion: 77 Resp: 286713
Ion Ratio Lower Upper
77 100
97 22.9 11.8 35.3





#27

cis-1,2-Dichloroethene

Concen: 152.112 ug/l

RT: 4.495 min Scan# 51

Delta R.T. -0.006 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion: 96 Resp: 227788

Ion Ratio Lower Upper

96 100

61 145.0 0.0 291.0

98 64.7 0.0 126.8



#28

Bromochloromethane

Concen: 151.520 ug/l

RT: 4.910 min Scan# 627

Delta R.T. -0.000 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02




Tgt Ion: 49 Resp: 161801


Ion Ratio Lower Upper

49 100

129 2.0 0.0 3.8

130 76.9 61.4 92.0





#29

Tetrahydrofuran

Concen: 815.879 ug/l

RT: 5.013 min Scan# 644

Delta R.T. -0.000 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion: 42 Resp: 370760

Ion Ratio Lower Upper

42 100

72 43.8 34.9 52.3

71 40.1 32.2 48.2



#30

Chloroform

Concen: 152.312 ug/l

RT: 5.105 min Scan# 659

Delta R.T. -0.000 min

Lab File: VX046722.D




Acq: 17 Jun 2025 15:02


Tgt Ion: 83 Resp: 358782

Ion Ratio Lower Upper

83 100

85 66.2 53.7 80.5









#31
Cyclohexane
Concen: 147.842 ug/l
RT: 5.483 min Scan# 721
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 56 Resp: 330235
Ion Ratio Lower Upper
56 100
69 29.4 24.8 37.2
84 83.6 68.2 102.2



#32
1,1,1-Trichloroethane
Concen: 155.082 ug/l
RT: 5.397 min Scan# 707
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion: 97 Resp: 320382
Ion Ratio Lower Upper
97 100
99 64.4 51.3 76.9
61 44.3 36.4 54.6









#33
 1,2-Dichloroethane-d4
 Concen: 154.797 ug/l
 RT: 5.964 min Scan# 800
 Delta R.T. -0.000 min
 Lab File: VX046722.D
 Acq: 17 Jun 2025 15:02




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC150


Tgt Ion: 65 Resp: 230915
 Ion Ratio Lower Upper
 65 100
 67 52.1 0.0 105.4



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.769 min Scan# 932
 Delta R.T. -0.000 min
 Lab File: VX046722.D
 Acq: 17 Jun 2025 15:02

Tgt Ion:114 Resp: 178346
 Ion Ratio Lower Upper
 114 100
 63 21.5 0.0 44.2
 88 16.1 0.0 33.2





#35

Dibromofluoromethane

Concen: 158.930 ug/l

RT: 5.397 min Scan# 707

Delta R.T. -0.000 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion:113 Resp: 187536

Ion Ratio Lower Upper

113 100

111 101.9 82.0 123.0

192 19.2 15.3 22.9



#36

1,1-Dichloropropene

Concen: 150.653 ug/l

RT: 5.702 min Scan# 757

Delta R.T. -0.000 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02




Tgt Ion: 75 Resp: 256619


Ion Ratio Lower Upper

75 100

110 35.0 17.2 51.5

77 31.0 24.2 36.4









#37
Ethyl Acetate
Concen: 156.284 ug/l
RT: 4.721 min Scan# 596
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 43 Resp: 245724
Ion Ratio Lower Upper
43 100
61 12.3 9.7 14.5
70 10.4 8.1 12.1



#38
Carbon Tetrachloride
Concen: 153.668 ug/l
RT: 5.684 min Scan# 754
Delta R.T. -0.006 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion: 117 Resp: 284066
Ion Ratio Lower Upper
117 100
119 94.6 75.2 112.8
121 30.9 24.1 36.1









#39
Methylcyclohexane
Concen: 154.413 ug/l
RT: 7.385 min Scan# 1033
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 83 Resp: 346118
Ion Ratio Lower Upper
83 100
55 78.6 63.0 94.4
98 48.5 38.8 58.2



#40
Benzene
Concen: 151.423 ug/l
RT: 6.044 min Scan# 813
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion: 78 Resp: 763341
Ion Ratio Lower Upper
78 100
77 23.5 18.8 28.2









#41
Methacrylonitrile
Concen: 169.745 ug/l
RT: 4.928 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150

Tgt Ion: 41 Resp: 142283




Ion	Ratio	Lower	Upper
41	100		
39	52.0	45.9	68.9
67	71.3	61.2	91.8
52	29.4	25.5	38.3




#42
1,2-Dichloroethane
Concen: 150.011 ug/l
RT: 6.092 min Scan# 821
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion: 62 Resp: 265740

Ion	Ratio	Lower	Upper
62	100		
98	9.3	0.0	17.8









#43
Isopropyl Acetate
Concen: 167.483 ug/l
RT: 6.342 min Scan# 862
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 43 Resp: 412067
Ion Ratio Lower Upper
43 100
61 19.6 15.3 22.9
87 12.8 10.0 15.0



#44
Trichloroethene
Concen: 152.520 ug/l
RT: 7.129 min Scan# 991
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion:130 Resp: 195943
Ion Ratio Lower Upper
130 100
95 100.8 0.0 202.6





#45

1,2-Dichloropropane

Concen: 153.025 ug/l

RT: 7.434 min Scan# 1041

Delta R.T. -0.000 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC150

Tgt Ion: 63 Resp: 191181

Ion Ratio Lower Upper

63 100

65 31.4 22.9 34.3



#46

Dibromomethane

Concen: 154.297 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. -0.006 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02




Tgt Ion: 93 Resp: 138541


Ion Ratio Lower Upper

93 100

95 82.2 66.2 99.4

174 97.0 77.0 115.6









#47
Bromodichloromethane
Concen: 157.630 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 83 Resp: 288826
Ion Ratio Lower Upper
83 100
85 64.0 51.4 77.2
127 8.8 6.9 10.3



#48
Methyl methacrylate
Concen: 171.734 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion: 41 Resp: 211908
Ion Ratio Lower Upper
41 100
69 80.7 63.4 95.2
39 56.3 45.4 68.0





#49

1,4-Dioxane

Concen: 2987.386 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. 0.000 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion: 88 Resp: 54646

Ion Ratio Lower Upper

88 100

43 37.6 31.2 46.8

58 71.8 54.7 82.1



#50

Toluene-d8

Concen: 155.733 ug/l

RT: 8.653 min Scan# 1241

Delta R.T. 0.006 min

Lab File: VX046722.D




Acq: 17 Jun 2025 15:02


Tgt Ion: 98 Resp: 660221

Ion Ratio Lower Upper

98 100

100 66.6 53.0 79.4









#51
4-Methyl-2-Pentanone
Concen: 799.825 ug/l
RT: 8.574 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 43 Resp: 1173666
Ion Ratio Lower Upper
43 100
58 38.8 30.8 46.2



#52
Toluene
Concen: 150.800 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion: 92 Resp: 471273
Ion Ratio Lower Upper
92 100
91 169.7 134.8 202.2





#53

t-1,3-Dichloropropene

Concen: 169.929 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC150

Tgt Ion: 75 Resp: 288893

Ion Ratio Lower Upper

75 100

77 31.5 25.6 38.4



#54

cis-1,3-Dichloropropene

Concen: 163.198 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02




Tgt Ion: 75 Resp: 314965


Ion Ratio Lower Upper

75 100

77 32.0 25.3 37.9

39 51.4 39.2 58.8









#55
1,1,2-Trichloroethane
Concen: 151.099 ug/l
RT: 9.153 min Scan# 1323
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150

Tgt Ion: 97 Resp: 175973




Ion	Ratio	Lower	Upper
97	100		
83	85.2	67.2	100.8
85	55.5	44.1	66.1
99	62.6	50.3	75.5




#56
Ethyl methacrylate
Concen: 167.940 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion: 69 Resp: 288720

Ion	Ratio	Lower	Upper
69	100		
41	68.8	55.4	83.2
39	37.3	29.7	44.5





#57

1,3-Dichloropropane

Concen: 150.290 ug/l

RT: 9.305 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC150

Tgt Ion: 76 Resp: 301418

Ion Ratio Lower Upper

76 100

78 32.4 25.7 38.5



#58

2-Chloroethyl Vinyl ether

Concen: 845.118 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. -0.000 min

Lab File: VX046722.D




Acq: 17 Jun 2025 15:02


Tgt Ion: 63 Resp: 754730

Ion Ratio Lower Upper

63 100

106 28.4 22.6 34.0









#59
2-Hexanone
Concen: 805.005 ug/l
RT: 9.433 min Scan# 11
Delta R.T. 0.006 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 43 Resp: 816655
Ion Ratio Lower Upper
43 100
58 54.1 26.8 80.4



#60
Dibromochloromethane
Concen: 156.904 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion: 129 Resp: 215347
Ion Ratio Lower Upper
129 100
127 77.3 38.4 115.2









#61
1,2-Dibromoethane
Concen: 156.392 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion:107 Resp: 185268
Ion Ratio Lower Upper
107 100
109 93.6 75.8 113.6



#62
4-Bromofluorobenzene
Concen: 154.404 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion: 95 Resp: 244007
Ion Ratio Lower Upper
95 100
174 75.6 0.0 150.4
176 72.7 0.0 145.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion:117 Resp: 156970
Ion Ratio Lower Upper
117 100
82 56.4 44.6 66.8
119 33.1 25.8 38.8



#64
Tetrachloroethene
Concen: 146.895 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion:164 Resp: 159752
Ion Ratio Lower Upper
164 100
166 125.9 104.6 157.0
129 97.1 76.3 114.5
131 91.9 74.1 111.1









#65
Chlorobenzene
Concen: 149.453 ug/l
RT: 10.080 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion:112 Resp: 517859
Ion Ratio Lower Upper
112 100
114 31.7 25.5 38.3



#66
1,1,1,2-Tetrachloroethane
Concen: 156.268 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. 0.006 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion:131 Resp: 181981
Ion Ratio Lower Upper
131 100
133 95.5 47.3 141.8
119 64.7 32.1 96.3









#67
Ethyl Benzene
Concen: 151.251 ug/l
RT: 10.195 min Scan# 1493
Delta R.T. 0.006 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 91 Resp: 915724
Ion Ratio Lower Upper
91 100
106 31.0 24.2 36.2



#68
m/p-Xylenes
Concen: 299.697 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion: 106 Resp: 676085
Ion Ratio Lower Upper
106 100
91 208.3 164.8 247.2









#69
o-Xylene
Concen: 155.514 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion:106 Resp: 328782
Ion Ratio Lower Upper
106 100
91 220.1 111.3 333.9



#70
Styrene
Concen: 153.143 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion:104 Resp: 566606
Ion Ratio Lower Upper
104 100
78 52.2 41.4 62.2
103 54.1 43.1 64.7









#71
Bromoform
Concen: 161.498 ug/l
RT: 10.799 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion:173 Resp: 143229
Ion Ratio Lower Upper
173 100
175 48.6 24.1 72.4
254 0.1 0.1 0.1



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion:152 Resp: 75829
Ion Ratio Lower Upper
152 100
115 139.0 43.2 129.6#
150 208.8 0.0 346.8









#73
Isopropylbenzene
Concen: 155.374 ug/l
RT: 10.964 min Scan# 1620
Delta R.T. 0.006 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion:105 Resp: 867728
Ion Ratio Lower Upper
105 100
120 26.0 13.0 39.0



#74
N-ethyl acetate
Concen: 174.748 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion: 43 Resp: 369757
Ion Ratio Lower Upper
43 100
70 41.3 33.4 50.0
55 23.3 19.1 28.7
61 23.9 19.0 28.4





#75

1,1,2,2-Tetrachloroethane

Concen: 156.730 ug/l

RT: 11.207 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion: 83 Resp: 244406

Ion Ratio Lower Upper

83 100

131 10.1 4.9 14.6

85 64.4 32.0 96.0



#76

1,2,3-Trichloropropane

Concen: 179.850 ug/l

RT: 11.238 min Scan# 1665

Delta R.T. -0.000 min

Lab File: VX046722.D




Acq: 17 Jun 2025 15:02


Tgt Ion: 75 Resp: 277006

Ion Ratio Lower Upper

75 100

77 31.9 18.5 55.5









#77
Bromobenzene
Concen: 154.285 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion:156 Resp: 207348
Ion Ratio Lower Upper
156 100
77 167.9 82.8 248.6
158 96.7 47.6 142.9



#78
n-propylbenzene
Concen: 154.633 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. 0.006 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion: 91 Resp: 1025742
Ion Ratio Lower Upper
91 100
120 22.6 11.4 34.1









#79
2-Chlorotoluene
Concen: 150.373 ug/l
RT: 11.366 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 91 Resp: 595555
Ion Ratio Lower Upper
91 100
126 33.4 16.6 49.7



#80
1,3,5-Trimethylbenzene
Concen: 154.053 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion: 105 Resp: 707005
Ion Ratio Lower Upper
105 100
120 48.5 24.1 72.2









#81
trans-1,4-Dichloro-2-butene
Concen: 176.408 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 75 Resp: 86677
Ion Ratio Lower Upper
75 100
53 98.9 86.8 130.2
89 45.8 36.9 55.3



#82
4-Chlorotoluene
Concen: 150.915 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion: 91 Resp: 697556
Ion Ratio Lower Upper
91 100
126 29.1 14.4 43.2









#83
tert-Butylbenzene
Concen: 155.093 ug/l
RT: 11.713 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion:119 Resp: 735750
Ion Ratio Lower Upper
119 100
91 59.0 29.5 88.6
134 22.8 11.7 35.0



#84
1,2,4-Trimethylbenzene
Concen: 155.994 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion:105 Resp: 719283
Ion Ratio Lower Upper
105 100
120 44.4 22.1 66.3









#85
 sec-Butylbenzene
 Concen: 152.910 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. -0.000 min
 Lab File: VX046722.D
 Acq: 17 Jun 2025 15:02




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC150


Tgt Ion:105 Resp: 916769
 Ion Ratio Lower Upper
 105 100
 134 19.9 9.8 29.4



#86
 p-Isopropyltoluene
 Concen: 152.610 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. -0.000 min
 Lab File: VX046722.D
 Acq: 17 Jun 2025 15:02

Tgt Ion:119 Resp: 771998
 Ion Ratio Lower Upper
 119 100
 134 25.9 12.9 38.7
 91 24.4 12.3 36.9





#87

1,3-Dichlorobenzene

Concen: 149.232 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion:146 Resp: 391049

Ion Ratio Lower Upper

146 100

111 41.3 20.6 61.8

148 63.2 31.6 94.8



#88

1,4-Dichlorobenzene

Concen: 143.836 ug/l

RT: 12.036 min Scan# 1796

Delta R.T. -0.000 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02




Tgt Ion:146 Resp: 387139


Ion Ratio Lower Upper

146 100

111 40.3 20.3 61.0

148 63.5 32.1 96.5





#89

n-Butylbenzene

Concen: 151.271 ug/l

RT: 12.329 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion: 91 Resp: 726563

Ion Ratio Lower Upper

91 100

92 54.7 27.1 81.3

134 25.6 12.9 38.6



#90

Hexachloroethane

Concen: 161.049 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX046722.D




Acq: 17 Jun 2025 15:02


Tgt Ion: 117 Resp: 142663

Ion Ratio Lower Upper

117 100

201 70.7 34.4 103.3





#91

1,2-Dichlorobenzene

Concen: 151.192 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. -0.000 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion:146 Resp: 370234

Ion Ratio Lower Upper

146 100

111 43.3 21.7 65.1

148 63.8 32.0 96.2



#92

1,2-Dibromo-3-Chloropropane

Concen: 175.448 ug/l

RT: 12.939 min Scan# 1944

Delta R.T. -0.000 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02




Tgt Ion: 75 Resp: 54014


Ion Ratio Lower Upper

75 100

155 84.2 41.9 125.7

157 108.5 51.4 154.3





#93

1,2,4-Trichlorobenzene

Concen: 151.531 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. -0.000 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion:180 Resp: 263803

Ion Ratio Lower Upper

180 100

182 96.9 46.7 140.1

145 32.2 15.8 47.4



#94

Hexachlorobutadiene

Concen: 130.431 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. 0.006 min

Lab File: VX046722.D

Acq: 17 Jun 2025 15:02




Tgt Ion:225 Resp: 95520


Ion Ratio Lower Upper

225 100

223 64.5 32.5 97.5

227 64.2 32.2 96.6









#95
Naphthalene
Concen: 170.609 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150

Tgt Ion:128 Resp: 846206
Ion Ratio Lower Upper
128 100
127 12.7 10.1 15.1
129 10.9 8.7 13.1



#96
1,2,3-Trichlorobenzene
Concen: 157.517 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. -0.000 min
Lab File: VX046722.D
Acq: 17 Jun 2025 15:02

Tgt Ion:180 Resp: 262348
Ion Ratio Lower Upper
180 100
182 93.7 47.8 143.4
145 33.8 17.3 51.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
 Data File : VX046725.D
 Acq On : 17 Jun 2025 17:18
 Operator : JC/MD
 Sample : VSTDICC001
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC001

Quant Time: Jun 18 02:42:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 02:34:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	136219	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	226094	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	205370	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	103593	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	1977	1.049	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	2.100%#
35) Dibromofluoromethane	5.403	113	1491	0.997	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	2.000%#
50) Toluene-d8	8.653	98	5563	1.035	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	2.080%#
62) 4-Bromofluorobenzene	11.085	95	2763	1.379	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	2.760%#
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.179	85	1159	0.797	ug/l	83
3) Chloromethane	1.307	50	1322	0.842	ug/l	93
4) Vinyl Chloride	1.386	62	1419	0.848	ug/l	89
5) Bromomethane	1.636	94	1325	1.407	ug/l	95
6) Chloroethane	1.703	64	905	0.892	ug/l	90
7) Trichlorofluoromethane	1.904	101	2062	0.811	ug/l	97
8) Diethyl Ether	2.148	74	722	0.828	ug/l	87
9) 1,1,2-Trichlorotrifluo...	2.343	101	1280	0.821	ug/l	99
10) Methyl Iodide	2.465	142	1790	5.171	ug/l	91
11) Tert butyl alcohol	2.953	59	606	3.583	ug/l #	72
12) 1,1-Dichloroethene	2.337	96	1228	0.816	ug/l #	76
13) Acrolein	2.252	56	1219	4.949	ug/l	99
14) Allyl chloride	2.672	41	2386	0.875	ug/l	96
15) Acrylonitrile	3.081	53	3075	4.044	ug/l	96
16) Acetone	2.386	43	3424	5.449	ug/l #	80
17) Carbon Disulfide	2.526	76	5092	1.121	ug/l	96
18) Methyl Acetate	2.721	43	1280	0.802	ug/l #	87
19) Methyl tert-butyl Ether	3.129	73	3888	0.830	ug/l #	89
20) Methylene Chloride	2.800	84	1736	0.994	ug/l	91
21) trans-1,2-Dichloroethene	3.111	96	1441	0.895	ug/l #	81
22) Diisopropyl ether	3.770	45	4170	0.796	ug/l #	71
23) Vinyl Acetate	3.745	43	15507	3.722	ug/l	95
24) 1,1-Dichloroethane	3.617	63	2648	0.890	ug/l #	79
25) 2-Butanone	4.593	43	3422	3.851	ug/l	93
26) 2,2-Dichloropropane	4.477	77	891	0.411	ug/l	94
27) cis-1,2-Dichloroethene	4.507	96	1698	0.898	ug/l	97
28) Bromochloromethane	4.922	49	1307	0.969	ug/l #	91
29) Tetrahydrofuran	5.032	42	2212	3.854	ug/l #	54
30) Chloroform	5.123	83	700	0.235	ug/l	86
31) Cyclohexane	5.477	56	2208	0.783	ug/l	86
32) 1,1,1-Trichloroethane	5.391	97	2212	0.848	ug/l #	50
36) 1,1-Dichloropropene	5.708	75	2016	0.934	ug/l	90
37) Ethyl Acetate	4.739	43	806	0.404	ug/l #	71
38) Carbon Tetrachloride	5.684	117	2124	0.906	ug/l	92
39) Methylcyclohexane	7.379	83	2485	0.875	ug/l	87
40) Benzene	6.044	78	5507	0.862	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
 Data File : VX046725.D
 Acq On : 17 Jun 2025 17:18
 Operator : JC/MD
 Sample : VSTDICC001
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC001

Quant Time: Jun 18 02:42:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 02:34:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.934	41	311	0.293	ug/l #	51
42) 1,2-Dichloroethane	6.098	62	1938	0.863	ug/l #	75
43) Isopropyl Acetate	6.367	43	2218	0.711	ug/l #	83
44) Trichloroethene	7.135	130	1448	0.889	ug/l	87
45) 1,2-Dichloropropane	7.434	63	1380	0.871	ug/l #	89
46) Dibromomethane	7.592	93	1030	0.905	ug/l	89
47) Bromodichloromethane	7.824	83	1825	0.786	ug/l #	82
48) Methyl methacrylate	7.720	41	1050	0.671	ug/l #	80
49) 1,4-Dioxane	7.671	88	56	28.694	ug/l #	77
51) 4-Methyl-2-Pentanone	8.580	43	7273	3.910	ug/l	97
52) Toluene	8.726	92	3393	0.856	ug/l	96
53) t-1,3-Dichloropropene	8.994	75	1607	0.746	ug/l #	87
54) cis-1,3-Dichloropropene	8.372	75	1971	0.806	ug/l #	57
55) 1,1,2-Trichloroethane	9.153	97	1297	0.878	ug/l #	83
56) Ethyl methacrylate	9.122	69	1477	0.678	ug/l	93
57) 1,3-Dichloropropane	9.311	76	2249	0.885	ug/l	95
58) 2-Chloroethyl Vinyl ether	8.251	63	3778	3.337	ug/l	92
59) 2-Hexanone	9.439	43	4833	3.758	ug/l	94
60) Dibromochloromethane	9.525	129	1436	0.825	ug/l	99
61) 1,2-Dibromoethane	9.610	107	1209	0.805	ug/l	90
64) Tetrachloroethene	9.275	164	1399	0.983	ug/l	95
65) Chlorobenzene	10.073	112	4127	0.910	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	1242	0.815	ug/l #	64
67) Ethyl Benzene	10.195	91	6967	0.880	ug/l	98
68) m/p-Xylenes	10.305	106	5214	1.767	ug/l	98
69) o-Xylene	10.646	106	2045	0.739	ug/l	80
70) Styrene	10.659	104	4077	0.842	ug/l	98
71) Bromoform	10.799	173	927	0.799	ug/l #	96
73) Isopropylbenzene	10.963	105	6314	0.828	ug/l	97
74) N-amyl acetate	10.848	43	1921	0.665	ug/l #	85
75) 1,1,2,2-Tetrachloroethane	11.213	83	1690	0.793	ug/l	93
76) 1,2,3-Trichloropropane	11.238	75	1746	0.830	ug/l	77
77) Bromobenzene	11.201	156	1541	0.839	ug/l	95
78) n-propylbenzene	11.305	91	7545	0.833	ug/l	98
79) 2-Chlorotoluene	11.366	91	4896	0.905	ug/l	93
80) 1,3,5-Trimethylbenzene	11.451	105	5174	0.825	ug/l	94
81) trans-1,4-Dichloro-2-b...	11.031	75	391	0.582	ug/l #	81
82) 4-Chlorotoluene	11.457	91	5664	0.897	ug/l	99
83) tert-Butylbenzene	11.713	119	5507	0.850	ug/l	95
84) 1,2,4-Trimethylbenzene	11.750	105	5087	0.808	ug/l	99
85) sec-Butylbenzene	11.890	105	6839	0.835	ug/l	98
86) p-Isopropyltoluene	12.006	119	6018	0.871	ug/l	92
87) 1,3-Dichlorobenzene	11.969	146	3509	0.980	ug/l	96
88) 1,4-Dichlorobenzene	11.969	146	3509	0.954	ug/l	96
89) n-Butylbenzene	12.335	91	6186	0.943	ug/l	96
90) Hexachloroethane	12.536	117	1081	0.893	ug/l	90
91) 1,2-Dichlorobenzene	12.335	146	3024	0.904	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.951	75	277	0.659	ug/l	95
93) 1,2,4-Trichlorobenzene	13.591	180	2424	1.019	ug/l	99
94) Hexachlorobutadiene	13.725	225	1025	1.025	ug/l	97
95) Naphthalene	13.780	128	5461	0.806	ug/l	96
96) 1,2,3-Trichlorobenzene	13.963	180	1982	0.871	ug/l	96


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
Data File : VX046725.D
Acq On : 17 Jun 2025 17:18
Operator : JC/MD
Sample : VSTDICC001
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Quant Time: Jun 18 02:42:58 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 02:34:13 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion:168 Resp: 136219
Ion Ratio Lower Upper
168 100
99 59.2 48.5 72.7



#2
Dichlorodifluoromethane
Concen: 0.797 ug/l
RT: 1.179 min Scan# 15
Delta R.T. -0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Tgt Ion: 85 Resp: 1159
Ion Ratio Lower Upper
85 100
87 23.5 16.6 49.8









#3
Chloromethane
Concen: 0.842 ug/l
RT: 1.307 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion: 50 Resp: 1322
Ion Ratio Lower Upper
50 100
52 35.2 24.9 37.3



#4
Vinyl Chloride
Concen: 0.848 ug/l
RT: 1.386 min Scan# 49
Delta R.T. -0.000 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Tgt Ion: 62 Resp: 1419
Ion Ratio Lower Upper
62 100
64 35.7 23.9 35.9









#5
Bromomethane
Concen: 1.407 ug/l
RT: 1.636 min Scan# 90
Delta R.T. 0.012 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion: 94 Resp: 1325
Ion Ratio Lower Upper
94 100
96 85.5 72.3 108.5



#6
Chloroethane
Concen: 0.892 ug/l
RT: 1.703 min Scan# 101
Delta R.T. -0.000 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Tgt Ion: 64 Resp: 905
Ion Ratio Lower Upper
64 100
66 28.0 27.1 40.7









#7
 Trichlorofluoromethane
 Concen: 0.811 ug/l
 RT: 1.904 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: VX046725.D
 Acq: 17 Jun 2025 17:18




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC001


Tgt Ion:101 Resp: 2062
 Ion Ratio Lower Upper
 101 100
 103 67.4 51.8 77.8



#8
 Diethyl Ether
 Concen: 0.828 ug/l
 RT: 2.148 min Scan# 174
 Delta R.T. -0.000 min
 Lab File: VX046725.D
 Acq: 17 Jun 2025 17:18

Tgt Ion: 74 Resp: 722
 Ion Ratio Lower Upper
 74 100
 45 118.6 52.3 157.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 0.821 ug/l
RT: 2.343 min Scan# 206
Delta R.T. -0.000 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion:	101	Resp:	1280
Ion	Ratio	Lower	Upper
101	100		
85	45.9	35.4	53.0
151	76.5	61.2	91.8



#10
Methyl Iodide
Concen: 5.171 ug/l
RT: 2.465 min Scan# 226
Delta R.T. -0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Tgt Ion:	142	Resp:	1790
Ion	Ratio	Lower	Upper
142	100		
127	40.0	37.0	55.6
141	12.0	11.5	17.3









#11
Tert butyl alcohol
Concen: 3.583 ug/l
RT: 2.953 min Scan# 307
Delta R.T. -0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion: 59 Resp: 606
Ion Ratio Lower Upper
59 100
57 0.0 8.3 12.5#



#12
1,1-Dichloroethene
Concen: 0.816 ug/l
RT: 2.337 min Scan# 205
Delta R.T. -0.000 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Tgt Ion: 96 Resp: 1228
Ion Ratio Lower Upper
96 100
61 200.9 126.6 189.8#
98 61.6 50.7 76.1





#13

Acrolein

Concen: 4.949 ug/l

RT: 2.252 min Scan# 19

Delta R.T. 0.006 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC001

Tgt Ion: 56 Resp: 1219

Ion Ratio Lower Upper

56 100

55 69.6 55.2 82.8



#14

Allyl chloride

Concen: 0.875 ug/l

RT: 2.672 min Scan# 260

Delta R.T. -0.006 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18




Tgt Ion: 41 Resp: 2386


Ion Ratio Lower Upper

41 100

39 64.2 53.6 80.4

76 31.4 26.8 40.2





#15

Acrylonitrile

Concen: 4.044 ug/l

RT: 3.081 min Scan# 31

Delta R.T. 0.006 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC001





Tgt Ion: 53 Resp: 3075

Ion Ratio Lower Upper

53 100

52 79.2 65.1 97.7

51 31.7 28.9 43.3



#16

Acetone

Concen: 5.449 ug/l

RT: 2.386 min Scan# 213

Delta R.T. -0.000 min

Lab File: VX046725.D




Acq: 17 Jun 2025 17:18


Tgt Ion: 43 Resp: 3424

Ion Ratio Lower Upper

43 100

58 19.6 24.5 36.7#









#17
Carbon Disulfide
Concen: 1.121 ug/l
RT: 2.526 min Scan# 21
Delta R.T. -0.000 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion: 76 Resp: 5092
Ion Ratio Lower Upper
76 100
78 8.1 7.5 11.3



#18
Methyl Acetate
Concen: 0.802 ug/l
RT: 2.721 min Scan# 268
Delta R.T. 0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Tgt Ion: 43 Resp: 1280
Ion Ratio Lower Upper
43 100
74 16.9 18.4 27.6#









#19
Methyl tert-butyl Ether
Concen: 0.830 ug/l
RT: 3.129 min Scan# 31
Delta R.T. 0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion: 73 Resp: 3888
Ion Ratio Lower Upper
73 100
57 17.3 18.2 27.2#



#20
Methylene Chloride
Concen: 0.994 ug/l
RT: 2.800 min Scan# 281
Delta R.T. -0.000 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Tgt Ion: 84 Resp: 1736
Ion Ratio Lower Upper
84 100
49 142.6 101.3 151.9
51 38.8 31.0 46.6
86 66.8 51.1 76.7









#21
trans-1,2-Dichloroethene
Concen: 0.895 ug/l
RT: 3.111 min Scan# 311
Delta R.T. 0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion: 96 Resp: 1441
Ion Ratio Lower Upper
96 100
61 118.7 112.5 168.7
98 47.7 51.9 77.9#



#22
Diisopropyl ether
Concen: 0.796 ug/l
RT: 3.770 min Scan# 440
Delta R.T. -0.000 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Tgt Ion: 45 Resp: 4170
Ion Ratio Lower Upper
45 100
43 155.2 91.9 137.9#
87 24.5 22.0 33.0
59 12.4 9.1 13.7









#23
 Vinyl Acetate
 Concen: 3.722 ug/l
 RT: 3.745 min Scan# 41
 Delta R.T. 0.012 min
 Lab File: VX046725.D
 Acq: 17 Jun 2025 17:18




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC001


Tgt Ion: 43 Resp: 15507
 Ion Ratio Lower Upper
 43 100
 86 12.0 8.2 12.2



#24
 1,1-Dichloroethane
 Concen: 0.890 ug/l
 RT: 3.617 min Scan# 415
 Delta R.T. -0.006 min
 Lab File: VX046725.D
 Acq: 17 Jun 2025 17:18

Tgt Ion: 63 Resp: 2648
 Ion Ratio Lower Upper
 63 100
 98 0.0 4.0 11.9#
 100 10.4 2.2 6.6#









#25
2-Butanone
Concen: 3.851 ug/l
RT: 4.593 min Scan# 51
Delta R.T. 0.030 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion: 43 Resp: 3422
Ion Ratio Lower Upper
43 100
72 22.0 20.2 30.4



#26
2,2-Dichloropropane
Concen: 0.411 ug/l
RT: 4.477 min Scan# 556
Delta R.T. -0.012 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Tgt Ion: 77 Resp: 891
Ion Ratio Lower Upper
77 100
97 20.7 11.8 35.3





#27

cis-1,2-Dichloroethene

Concen: 0.898 ug/l

RT: 4.507 min Scan# 560

Delta R.T. 0.006 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001





Tgt Ion: 96 Resp: 1698

Ion Ratio Lower Upper

96 100

61 143.8 0.0 291.0

98 69.4 0.0 126.8



#28

Bromochloromethane

Concen: 0.969 ug/l

RT: 4.922 min Scan# 629

Delta R.T. 0.012 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18




Tgt Ion: 49 Resp: 1307


Ion Ratio Lower Upper

49 100

129 0.0 0.0 3.8

130 68.6 61.4 92.0





#29

Tetrahydrofuran

Concen: 3.854 ug/l

RT: 5.032 min Scan# 64

Delta R.T. 0.018 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001





Tgt Ion: 42 Resp: 2212

Ion Ratio Lower Upper

42 100

72 0.0 34.9 52.3#

71 26.1 32.2 48.2#



#30

Chloroform

Concen: 0.235 ug/l

RT: 5.123 min Scan# 662

Delta R.T. 0.018 min

Lab File: VX046725.D




Acq: 17 Jun 2025 17:18


Tgt Ion: 83 Resp: 700

Ion Ratio Lower Upper

83 100

85 78.2 53.7 80.5









#31
Cyclohexane
Concen: 0.783 ug/l
RT: 5.477 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion: 56 Resp: 2208
Ion Ratio Lower Upper
56 100
69 28.7 24.8 37.2
84 69.3 68.2 102.2



#32
1,1,1-Trichloroethane
Concen: 0.848 ug/l
RT: 5.391 min Scan# 706
Delta R.T. -0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Tgt Ion: 97 Resp: 2212
Ion Ratio Lower Upper
97 100
99 0.0 51.3 76.9#
61 41.9 36.4 54.6





#33

1,2-Dichloroethane-d4

Concen: 1.049 ug/l

RT: 5.964 min Scan# 800

Delta R.T. -0.000 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC001

Tgt Ion: 65 Resp: 1977

Ion Ratio Lower Upper

65 100

67 55.4 0.0 105.4



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. -0.006 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18




Tgt Ion: 114 Resp: 226094


Ion Ratio Lower Upper

114 100

63 21.2 0.0 44.2

88 16.3 0.0 33.2









#35
Dibromofluoromethane
Concen: 0.997 ug/l
RT: 5.403 min Scan# 707
Delta R.T. 0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion:113 Resp: 1491
Ion Ratio Lower Upper
113 100
111 101.5 82.0 123.0
192 12.9 15.3 22.9#



#36
1,1-Dichloropropene
Concen: 0.934 ug/l
RT: 5.708 min Scan# 758
Delta R.T. 0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Tgt Ion: 75 Resp: 2016
Ion Ratio Lower Upper
75 100
110 27.6 17.2 51.5
77 25.5 24.2 36.4





#37

Ethyl Acetate

Concen: 0.404 ug/l

RT: 4.739 min Scan# 599

Delta R.T. 0.018 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001





Tgt Ion: 43 Resp: 806

Ion Ratio Lower Upper

43 100

61 0.0 9.7 14.5#

70 0.0 8.1 12.1#



#38

Carbon Tetrachloride

Concen: 0.906 ug/l

RT: 5.684 min Scan# 754

Delta R.T. -0.006 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18




Tgt Ion: 117 Resp: 2124


Ion Ratio Lower Upper

117 100

119 102.3 75.2 112.8

121 32.4 24.1 36.1





#39

Methylcyclohexane

Concen: 0.875 ug/l

RT: 7.379 min Scan# 1032

Delta R.T. -0.006 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001





Tgt Ion: 83 Resp: 2485

Ion Ratio Lower Upper

83 100

55 91.5 63.0 94.4

98 55.1 38.8 58.2



#40

Benzene

Concen: 0.862 ug/l

RT: 6.044 min Scan# 813

Delta R.T. -0.000 min

Lab File: VX046725.D




Acq: 17 Jun 2025 17:18


Tgt Ion: 78 Resp: 5507

Ion Ratio Lower Upper

78 100

77 22.0 18.8 28.2









#41
Methacrylonitrile
Concen: 0.293 ug/l
RT: 4.934 min Scan# 61
Delta R.T. 0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion: 41	Resp:	311
Ion	Ratio	Lower Upper
41	100	
39	0.0	45.9 68.9#
67	56.3	61.2 91.8#
52	0.0	25.5 38.3#



#42
1,2-Dichloroethane
Concen: 0.863 ug/l
RT: 6.098 min Scan# 822
Delta R.T. 0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Tgt Ion: 62	Resp:	1938
Ion	Ratio	Lower Upper
62	100	
98	0.0	0.0 17.8









#43
Isopropyl Acetate
Concen: 0.711 ug/l
RT: 6.367 min Scan# 866
Delta R.T. 0.024 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion: 43 Resp: 2218
Ion Ratio Lower Upper
43 100
61 9.3 15.3 22.9#
87 8.6 10.0 15.0#



#44
Trichloroethene
Concen: 0.889 ug/l
RT: 7.135 min Scan# 992
Delta R.T. 0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Tgt Ion:130 Resp: 1448
Ion Ratio Lower Upper
130 100
95 114.0 0.0 202.6





#45

1,2-Dichloropropane

Concen: 0.871 ug/l

RT: 7.434 min Scan# 1041

Delta R.T. -0.000 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC001

Tgt Ion: 63 Resp: 1380

Ion Ratio Lower Upper

63 100

65 34.6 22.9 34.3#



#46

Dibromomethane

Concen: 0.905 ug/l

RT: 7.592 min Scan# 1067

Delta R.T. 0.006 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18




Tgt Ion: 93 Resp: 1030


Ion Ratio Lower Upper

93 100

95 71.1 66.2 99.4

174 86.7 77.0 115.6









#47
Bromodichloromethane
Concen: 0.786 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.000 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion: 83 Resp: 1825
Ion Ratio Lower Upper
83 100
85 80.3 51.4 77.2#
127 7.7 6.9 10.3



#48
Methyl methacrylate
Concen: 0.671 ug/l
RT: 7.720 min Scan# 1088
Delta R.T. 0.024 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Tgt Ion: 41 Resp: 1050
Ion Ratio Lower Upper
41 100
69 71.8 63.4 95.2
39 30.3 45.4 68.0#





#49

1,4-Dioxane

Concen: 28.694 ug/l

RT: 7.671 min Scan# 1080

Delta R.T. 0.012 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001





Tgt Ion: 88 Resp: 56

Ion Ratio Lower Upper

88 100

43 75.0 31.2 46.8#

58 66.1 54.7 82.1



#50

Toluene-d8

Concen: 1.035 ug/l

RT: 8.653 min Scan# 1241

Delta R.T. 0.006 min

Lab File: VX046725.D




Acq: 17 Jun 2025 17:18


Tgt Ion: 98 Resp: 5563

Ion Ratio Lower Upper

98 100

100 67.9 53.0 79.4









#51
4-Methyl-2-Pentanone
Concen: 3.910 ug/l
RT: 8.580 min Scan# 1229
Delta R.T. 0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion: 43 Resp: 7273
Ion Ratio Lower Upper
43 100
58 36.9 30.8 46.2



#52
Toluene
Concen: 0.856 ug/l
RT: 8.726 min Scan# 1253
Delta R.T. 0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Tgt Ion: 92 Resp: 3393
Ion Ratio Lower Upper
92 100
91 163.1 134.8 202.2





#53

t-1,3-Dichloropropene

Concen: 0.746 ug/l

RT: 8.994 min Scan# 1195

Delta R.T. 0.018 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC001

Tgt Ion: 75 Resp: 1607

Ion Ratio Lower Upper

75 100

77 39.3 25.6 38.4#



#54

cis-1,3-Dichloropropene

Concen: 0.806 ug/l

RT: 8.372 min Scan# 1195

Delta R.T. 0.006 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18




Tgt Ion: 75 Resp: 1971


Ion Ratio Lower Upper

75 100

77 30.3 25.3 37.9

39 96.1 39.2 58.8#









#55
1,1,2-Trichloroethane
Concen: 0.878 ug/l
RT: 9.153 min Scan# 1318
Delta R.T. -0.000 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion:	97	Resp:	1297
Ion	Ratio	Lower	Upper
97	100		
83	66.9	67.2	100.8#
85	42.3	44.1	66.1#
99	53.1	50.3	75.5



#56
Ethyl methacrylate
Concen: 0.678 ug/l
RT: 9.122 min Scan# 1318
Delta R.T. 0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Tgt Ion:	69	Resp:	1477
Ion	Ratio	Lower	Upper
69	100		
41	63.2	55.4	83.2
39	41.0	29.7	44.5









#57
1,3-Dichloropropane
Concen: 0.885 ug/l
RT: 9.311 min Scan# 1175
Delta R.T. 0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion: 76 Resp: 2249
Ion Ratio Lower Upper
76 100
78 29.5 25.7 38.5



#58
2-Chloroethyl Vinyl ether
Concen: 3.337 ug/l
RT: 8.251 min Scan# 1175
Delta R.T. 0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Tgt Ion: 63 Resp: 3778
Ion Ratio Lower Upper
63 100
106 24.2 22.6 34.0





#59

2-Hexanone

Concen: 3.758 ug/l

RT: 9.439 min Scan# 11

Delta R.T. 0.012 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC001

Tgt Ion: 43 Resp: 4833

Ion Ratio Lower Upper

43 100

58 49.0 26.8 80.4



#60

Dibromochloromethane

Concen: 0.825 ug/l

RT: 9.525 min Scan# 1384

Delta R.T. 0.006 min

Lab File: VX046725.D




Acq: 17 Jun 2025 17:18


Tgt Ion:129 Resp: 1436

Ion Ratio Lower Upper

129 100

127 75.8 38.4 115.2





#61

1,2-Dibromoethane

Concen: 0.805 ug/l

RT: 9.610 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC001

Tgt Ion:107 Resp: 1209

Ion Ratio Lower Upper

107 100

109 104.6 75.8 113.6



#62

4-Bromofluorobenzene

Concen: 1.379 ug/l

RT: 11.085 min Scan# 1640

Delta R.T. 0.006 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18




Tgt Ion: 95 Resp: 2763


Ion Ratio Lower Upper

95 100

174 72.0 0.0 150.4

176 77.1 0.0 145.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. -0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion:117 Resp: 205370
Ion Ratio Lower Upper
117 100
82 58.0 44.6 66.8
119 32.1 25.8 38.8



#64
Tetrachloroethene
Concen: 0.983 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Tgt Ion:164 Resp: 1399
Ion Ratio Lower Upper
164 100
166 119.7 104.6 157.0
129 96.4 76.3 114.5
131 95.9 74.1 111.1









#65
Chlorobenzene
Concen: 0.910 ug/l
RT: 10.073 min Scan# 1474
Delta R.T. -0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion:112 Resp: 4127
Ion Ratio Lower Upper
112 100
114 31.3 25.5 38.3



#66
1,1,1,2-Tetrachloroethane
Concen: 0.815 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Tgt Ion:131 Resp: 1242
Ion Ratio Lower Upper
131 100
133 89.0 47.3 141.8
119 0.0 32.1 96.3#









#67
Ethyl Benzene
Concen: 0.880 ug/l
RT: 10.195 min Scan# 1493
Delta R.T. 0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion: 91 Resp: 6967
Ion Ratio Lower Upper
91 100
106 29.3 24.2 36.2



#68
m/p-Xylenes
Concen: 1.767 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. 0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Tgt Ion: 106 Resp: 5214
Ion Ratio Lower Upper
106 100
91 202.7 164.8 247.2





#69

o-Xylene

Concen: 0.739 ug/l

RT: 10.646 min Scan# 1

Delta R.T. 0.006 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC001

Tgt Ion:106 Resp: 2045

Ion Ratio Lower Upper

106 100

91 255.0 111.3 333.9



#70

Styrene

Concen: 0.842 ug/l

RT: 10.659 min Scan# 1570

Delta R.T. 0.006 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18




Tgt Ion:104 Resp: 4077


Ion Ratio Lower Upper

104 100

78 50.7 41.4 62.2

103 52.8 43.1 64.7









#71
Bromoform
Concen: 0.799 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion:173 Resp: 927
Ion Ratio Lower Upper
173 100
175 45.4 24.1 72.4
254 0.0 0.1 0.1#



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Tgt Ion:152 Resp: 103593
Ion Ratio Lower Upper
152 100
115 61.3 43.2 129.6
150 159.2 0.0 346.8









#73
Isopropylbenzene
Concen: 0.828 ug/l
RT: 10.963 min Scan# 1620
Delta R.T. 0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion:105 Resp: 6314
Ion Ratio Lower Upper
105 100
120 27.6 13.0 39.0



#74
N-ethyl acetate
Concen: 0.665 ug/l
RT: 10.848 min Scan# 1601
Delta R.T. 0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Tgt Ion: 43 Resp: 1921
Ion Ratio Lower Upper
43 100
70 38.4 33.4 50.0
55 20.1 19.1 28.7
61 5.0 19.0 28.4#









#75
 1,1,2,2-Tetrachloroethane
 Concen: 0.793 ug/l
 RT: 11.213 min Scan# 1661
 Delta R.T. 0.006 min
 Lab File: VX046725.D
 Acq: 17 Jun 2025 17:18




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC001


Tgt Ion: 83 Resp: 1690
 Ion Ratio Lower Upper
 83 100
 131 10.8 4.9 14.6
 85 70.3 32.0 96.0



#76
 1,2,3-Trichloropropane
 Concen: 0.830 ug/l
 RT: 11.238 min Scan# 1665
 Delta R.T. -0.000 min
 Lab File: VX046725.D
 Acq: 17 Jun 2025 17:18

Tgt Ion: 75 Resp: 1746
 Ion Ratio Lower Upper
 75 100
 77 23.4 18.5 55.5









#77
Bromobenzene
Concen: 0.839 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion:156 Resp: 1541
Ion Ratio Lower Upper
156 100
77 169.3 82.8 248.6
158 103.2 47.6 142.9



#78
n-propylbenzene
Concen: 0.833 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. 0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Tgt Ion: 91 Resp: 7545
Ion Ratio Lower Upper
91 100
120 23.6 11.4 34.1





#79

2-Chlorotoluene

Concen: 0.905 ug/l

RT: 11.366 min Scan# 1

Delta R.T. 0.006 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC001

Tgt Ion: 91 Resp: 4896

Ion Ratio Lower Upper

91 100

126 28.9 16.6 49.7



#80

1,3,5-Trimethylbenzene

Concen: 0.825 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. -0.000 min

Lab File: VX046725.D




Acq: 17 Jun 2025 17:18


Tgt Ion: 105 Resp: 5174

Ion Ratio Lower Upper

105 100

120 44.2 24.1 72.2









#81
trans-1,4-Dichloro-2-butene
Concen: 0.582 ug/l
RT: 11.031 min Scan# 1
Delta R.T. 0.012 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion: 75 Resp: 391
Ion Ratio Lower Upper
75 100
53 127.4 86.8 130.2
89 59.8 36.9 55.3#



#82
4-Chlorotoluene
Concen: 0.897 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Tgt Ion: 91 Resp: 5664
Ion Ratio Lower Upper
91 100
126 29.2 14.4 43.2









#83
tert-Butylbenzene
Concen: 0.850 ug/l
RT: 11.713 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion:119 Resp: 5507
Ion Ratio Lower Upper
119 100
91 55.1 29.5 88.6
134 21.6 11.7 35.0



#84
1,2,4-Trimethylbenzene
Concen: 0.808 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Tgt Ion:105 Resp: 5087
Ion Ratio Lower Upper
105 100
120 43.6 22.1 66.3









#85
sec-Butylbenzene
Concen: 0.835 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. -0.000 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion:105 Resp: 6839
Ion Ratio Lower Upper
105 100
134 20.3 9.8 29.4



#86
p-Isopropyltoluene
Concen: 0.871 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Tgt Ion:119 Resp: 6018
Ion Ratio Lower Upper
119 100
134 24.5 12.9 38.7
91 31.1 12.3 36.9





#87

1,3-Dichlorobenzene

Concen: 0.980 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001





Tgt Ion:146 Resp: 3509

Ion Ratio Lower Upper

146 100

111 40.8 20.6 61.8

148 58.8 31.6 94.8



#88

1,4-Dichlorobenzene

Concen: 0.954 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.067 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18




Tgt Ion:146 Resp: 3509


Ion Ratio Lower Upper

146 100

111 40.8 20.3 61.0

148 58.8 32.1 96.5





#89

n-Butylbenzene

Concen: 0.943 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.006 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001





Tgt Ion: 91 Resp: 6186

Ion Ratio Lower Upper

91 100

92 51.3 27.1 81.3

134 24.5 12.9 38.6



#90

Hexachloroethane

Concen: 0.893 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX046725.D




Acq: 17 Jun 2025 17:18


Tgt Ion: 117 Resp: 1081

Ion Ratio Lower Upper

117 100

201 60.5 34.4 103.3









#91
 1,2-Dichlorobenzene
 Concen: 0.904 ug/l
 RT: 12.335 min Scan# 1845
 Delta R.T. 0.006 min
 Lab File: VX046725.D
 Acq: 17 Jun 2025 17:18




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC001


Tgt Ion:146 Resp: 3024
 Ion Ratio Lower Upper
 146 100
 111 44.1 21.7 65.1
 148 63.1 32.0 96.2



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 0.659 ug/l
 RT: 12.951 min Scan# 1946
 Delta R.T. 0.012 min
 Lab File: VX046725.D
 Acq: 17 Jun 2025 17:18

Tgt Ion: 75 Resp: 277
 Ion Ratio Lower Upper
 75 100
 155 73.6 41.9 125.7
 157 102.5 51.4 154.3





#93

1,2,4-Trichlorobenzene

Concen: 1.019 ug/l

RT: 13.591 min Scan# 2051

Delta R.T. 0.006 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001





Tgt Ion:180 Resp: 2424

Ion Ratio Lower Upper

180 100

182 93.2 46.7 140.1

145 30.0 15.8 47.4



#94

Hexachlorobutadiene

Concen: 1.025 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. 0.006 min

Lab File: VX046725.D

Acq: 17 Jun 2025 17:18




Tgt Ion:225 Resp: 1025


Ion Ratio Lower Upper

225 100

223 62.1 32.5 97.5

227 62.4 32.2 96.6









#95
Naphthalene
Concen: 0.806 ug/l
RT: 13.780 min Scan# 2081
Delta R.T. 0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001

Tgt Ion:	128	Resp:	5461
Ion	Ratio	Lower	Upper
128	100		
127	11.8	10.1	15.1
129	8.9	8.7	13.1



#96
1,2,3-Trichlorobenzene
Concen: 0.871 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. 0.006 min
Lab File: VX046725.D
Acq: 17 Jun 2025 17:18

Tgt Ion:	180	Resp:	1982
Ion	Ratio	Lower	Upper
180	100		
182	99.9	47.8	143.4
145	34.1	17.3	51.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
 Data File : VX046726.D
 Acq On : 17 Jun 2025 18:00
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX061725

Quant Time: Jun 18 07:54:49 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	129144	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	209541	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	184761	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	91992	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	81629	45.697	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.400%
35) Dibromofluoromethane	5.397	113	66647	48.072	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.140%
50) Toluene-d8	8.647	98	236781	47.537	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.080%
62) 4-Bromofluorobenzene	11.079	95	88588	47.712	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	95.420%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.179	85	75554	54.798	ug/l	100
3) Chloromethane	1.307	50	76483	51.393	ug/l	99
4) Vinyl Chloride	1.386	62	82156	51.768	ug/l	95
5) Bromomethane	1.618	94	48957	54.831	ug/l	96
6) Chloroethane	1.703	64	49931	51.911	ug/l	99
7) Trichlorofluoromethane	1.904	101	123768	51.363	ug/l	98
8) Diethyl Ether	2.148	74	42437	51.316	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.343	101	75525	51.089	ug/l	99
10) Methyl Iodide	2.465	142	84139	45.861	ug/l	99
11) Tert butyl alcohol	2.959	59	40385	251.888	ug/l	100
12) 1,1-Dichloroethene	2.337	96	74097	51.945	ug/l	99
13) Acrolein	2.246	56	59544	254.972	ug/l	99
14) Allyl chloride	2.678	41	131610	50.922	ug/l	99
15) Acrylonitrile	3.075	53	187601	260.219	ug/l	98
16) Acetone	2.386	43	146745	246.347	ug/l	98
17) Carbon Disulfide	2.526	76	205516	47.710	ug/l	98
18) Methyl Acetate	2.715	43	80049	52.909	ug/l	98
19) Methyl tert-butyl Ether	3.123	73	226557	51.008	ug/l	99
20) Methylene Chloride	2.800	84	80668	48.696	ug/l	99
21) trans-1,2-Dichloroethene	3.105	96	76889	50.375	ug/l	98
22) Diisopropyl ether	3.770	45	257775	51.890	ug/l	99
23) Vinyl Acetate	3.733	43	1044679	264.451	ug/l	100
24) 1,1-Dichloroethane	3.623	63	142442	50.502	ug/l	99
25) 2-Butanone	4.562	43	222870	264.542	ug/l	98
26) 2,2-Dichloropropane	4.483	77	104151	50.629	ug/l	99
27) cis-1,2-Dichloroethene	4.501	96	89665	50.003	ug/l	99
28) Bromochloromethane	4.904	49	64330	50.308	ug/l	99
29) Tetrahydrofuran	5.013	42	143570	263.836	ug/l	99
30) Chloroform	5.105	83	144108	51.089	ug/l	98
31) Cyclohexane	5.483	56	132301	49.462	ug/l	99
32) 1,1,1-Trichloroethane	5.391	97	123796	50.042	ug/l	99
36) 1,1-Dichloropropene	5.702	75	100241	50.088	ug/l	99
37) Ethyl Acetate	4.721	43	94953	51.401	ug/l	99
38) Carbon Tetrachloride	5.690	117	109471	50.403	ug/l	98
39) Methylcyclohexane	7.385	83	132086	50.155	ug/l	97
40) Benzene	6.044	78	304346	51.385	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
 Data File : VX046726.D
 Acq On : 17 Jun 2025 18:00
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX061725

Quant Time: Jun 18 07:54:49 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.928	41	54501	55.341	ug/l	97
42) 1,2-Dichloroethane	6.099	62	106085	50.970	ug/l	99
43) Isopropyl Acetate	6.342	43	155773	53.888	ug/l	99
44) Trichloroethene	7.129	130	76981	51.000	ug/l	97
45) 1,2-Dichloropropane	7.434	63	74773	50.940	ug/l	94
46) Dibromomethane	7.586	93	54149	51.329	ug/l	100
47) Bromodichloromethane	7.824	83	110823	51.479	ug/l	100
48) Methyl methacrylate	7.696	41	79868	55.091	ug/l	99
49) 1,4-Dioxane	7.659	88	20314	964.253	ug/l	97
51) 4-Methyl-2-Pentanone	8.574	43	470527	272.916	ug/l	99
52) Toluene	8.720	92	188131	51.237	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	105435	52.785	ug/l	99
54) cis-1,3-Dichloropropene	8.366	75	117487	51.813	ug/l	97
55) 1,1,2-Trichloroethane	9.153	97	70739	51.698	ug/l	99
56) Ethyl methacrylate	9.116	69	109612	54.266	ug/l	100
57) 1,3-Dichloropropane	9.305	76	121649	51.625	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.244	63	293709	279.922	ug/l	100
59) 2-Hexanone	9.427	43	326564	273.982	ug/l	99
60) Dibromochloromethane	9.519	129	82234	50.997	ug/l	99
61) 1,2-Dibromoethane	9.610	107	72749	52.268	ug/l	98
64) Tetrachloroethene	9.275	164	62943	49.172	ug/l	98
65) Chlorobenzene	10.079	112	203947	50.006	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	69709	50.856	ug/l	99
67) Ethyl Benzene	10.195	91	359779	50.487	ug/l	98
68) m/p-Xylenes	10.299	106	272935	102.789	ug/l	99
69) o-Xylene	10.640	106	130980	52.635	ug/l	97
70) Styrene	10.653	104	227060	52.139	ug/l	100
71) Bromoform	10.799	173	53660	51.404	ug/l #	100
73) Isopropylbenzene	10.957	105	346098	51.083	ug/l	100
74) N-amyl acetate	10.842	43	138890	54.107	ug/l	100
75) 1,1,2,2-Tetrachloroethane	11.207	83	98960	52.310	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	106102	62.974	ug/l	92
77) Bromobenzene	11.195	156	82515	50.611	ug/l	99
78) n-propylbenzene	11.305	91	412345	51.240	ug/l	100
79) 2-Chlorotoluene	11.360	91	239904	49.931	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	288104	51.746	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	29929	50.210	ug/l	99
82) 4-Chlorotoluene	11.451	91	280848	50.085	ug/l	98
83) tert-Butylbenzene	11.713	119	287513	49.958	ug/l	98
84) 1,2,4-Trimethylbenzene	11.750	105	286365	51.193	ug/l	99
85) sec-Butylbenzene	11.890	105	371523	51.079	ug/l	100
86) p-Isopropyltoluene	12.006	119	311698	50.791	ug/l	100
87) 1,3-Dichlorobenzene	11.969	146	154181	48.501	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	153158	46.906	ug/l	100
89) n-Butylbenzene	12.329	91	296174	50.829	ug/l	100
90) Hexachloroethane	12.536	117	52831	49.161	ug/l	98
91) 1,2-Dichlorobenzene	12.335	146	147561	49.672	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	12.939	75	19851	53.151	ug/l	97
93) 1,2,4-Trichlorobenzene	13.585	180	104267	49.369	ug/l	99
94) Hexachlorobutadiene	13.719	225	45958	51.729	ug/l	98
95) Naphthalene	13.774	128	315669	52.462	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	103451	51.200	ug/l	99


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
Data File : VX046726.D
Acq On : 17 Jun 2025 18:00
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725

Quant Time: Jun 18 07:54:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 03:09:16 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.562 min Scan# 71
Delta R.T. 0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00




Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725


Tgt Ion:168 Resp: 129144
Ion Ratio Lower Upper
168 100
99 61.8 48.5 72.7



#2
Dichlorodifluoromethane
Concen: 54.798 ug/l
RT: 1.179 min Scan# 15
Delta R.T. -0.006 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Tgt Ion: 85 Resp: 75554
Ion Ratio Lower Upper
85 100
87 33.0 16.6 49.8









#3
Chloromethane
Concen: 51.393 ug/l
RT: 1.307 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00




Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725


Tgt Ion: 50 Resp: 76483
Ion Ratio Lower Upper
50 100
52 31.8 24.9 37.3



#4
Vinyl Chloride
Concen: 51.768 ug/l
RT: 1.386 min Scan# 49
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Tgt Ion: 62 Resp: 82156
Ion Ratio Lower Upper
62 100
64 32.7 23.9 35.9









#5
Bromomethane
Concen: 54.831 ug/l
RT: 1.618 min Scan# 81
Delta R.T. -0.006 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00




Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725


Tgt Ion: 94 Resp: 48957
Ion Ratio Lower Upper
94 100
96 94.4 72.3 108.5



#6
Chloroethane
Concen: 51.911 ug/l
RT: 1.703 min Scan# 101
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Tgt Ion: 64 Resp: 49931
Ion Ratio Lower Upper
64 100
66 33.5 27.1 40.7









#7
 Trichlorofluoromethane
 Concen: 51.363 ug/l
 RT: 1.904 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: VX046726.D
 Acq: 17 Jun 2025 18:00




Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX061725


Tgt Ion:101 Resp: 123768
 Ion Ratio Lower Upper
 101 100
 103 63.4 51.8 77.8



#8
 Diethyl Ether
 Concen: 51.316 ug/l
 RT: 2.148 min Scan# 174
 Delta R.T. -0.000 min
 Lab File: VX046726.D
 Acq: 17 Jun 2025 18:00

Tgt Ion: 74 Resp: 42437
 Ion Ratio Lower Upper
 74 100
 45 103.7 52.3 157.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 51.089 ug/l
RT: 2.343 min Scan# 206
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00




Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725


Tgt Ion:101 Resp: 75525
Ion Ratio Lower Upper
101 100
85 43.6 35.4 53.0
151 75.9 61.2 91.8



#10
Methyl Iodide
Concen: 45.861 ug/l
RT: 2.465 min Scan# 226
Delta R.T. -0.006 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Tgt Ion:142 Resp: 84139
Ion Ratio Lower Upper
142 100
127 45.9 37.0 55.6
141 14.0 11.5 17.3









#11
Tert butyl alcohol
Concen: 251.888 ug/l
RT: 2.959 min Scan# 307
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00




Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725


Tgt Ion: 59 Resp: 40385
Ion Ratio Lower Upper
59 100
57 10.4 8.3 12.5



#12
1,1-Dichloroethene
Concen: 51.945 ug/l
RT: 2.337 min Scan# 205
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Tgt Ion: 96 Resp: 74097
Ion Ratio Lower Upper
96 100
61 159.8 126.6 189.8
98 63.7 50.7 76.1





#13

Acrolein

Concen: 254.972 ug/l

RT: 2.246 min Scan# 190

Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00

Instrument :

MSVOA_X

ClientSampleId :





ICVVX061725

Tgt Ion: 56 Resp: 59544

Ion Ratio Lower Upper

56 100

55 70.0 55.2 82.8



#14

Allyl chloride

Concen: 50.922 ug/l

RT: 2.678 min Scan# 261

Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00




Tgt Ion: 41 Resp: 131610


Ion Ratio Lower Upper

41 100

39 66.1 53.6 80.4

76 33.1 26.8 40.2









#15
Acrylonitrile
Concen: 260.219 ug/l
RT: 3.075 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00




Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725


Tgt Ion: 53 Resp: 187601
Ion Ratio Lower Upper
53 100
52 82.9 65.1 97.7
51 36.7 28.9 43.3



#16
Acetone
Concen: 246.347 ug/l
RT: 2.386 min Scan# 213
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Tgt Ion: 43 Resp: 146745
Ion Ratio Lower Upper
43 100
58 29.4 24.5 36.7









#17
Carbon Disulfide
Concen: 47.710 ug/l
RT: 2.526 min Scan# 21
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00




Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725


Tgt Ion: 76 Resp: 205516
Ion Ratio Lower Upper
76 100
78 8.7 7.5 11.3



#18
Methyl Acetate
Concen: 52.909 ug/l
RT: 2.715 min Scan# 267
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Tgt Ion: 43 Resp: 80049
Ion Ratio Lower Upper
43 100
74 21.9 18.4 27.6









#19
Methyl tert-butyl Ether
Concen: 51.008 ug/l
RT: 3.123 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00




Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725


Tgt Ion: 73 Resp: 226557
Ion Ratio Lower Upper
73 100
57 22.3 18.2 27.2



#20
Methylene Chloride
Concen: 48.696 ug/l
RT: 2.800 min Scan# 281
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Tgt Ion: 84 Resp: 80668
Ion Ratio Lower Upper
84 100
49 126.0 101.3 151.9
51 38.3 31.0 46.6
86 65.3 51.1 76.7









#21
trans-1,2-Dichloroethene
Concen: 50.375 ug/l
RT: 3.105 min Scan# 311
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00




Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725


Tgt Ion: 96 Resp: 76889
Ion Ratio Lower Upper
96 100
61 142.0 112.5 168.7
98 69.1 51.9 77.9



#22
Diisopropyl ether
Concen: 51.890 ug/l
RT: 3.770 min Scan# 440
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Tgt Ion: 45 Resp: 257775
Ion Ratio Lower Upper
45 100
43 114.3 91.9 137.9
87 26.4 22.0 33.0
59 11.2 9.1 13.7









#23
 Vinyl Acetate
 Concen: 264.451 ug/l
 RT: 3.733 min Scan# 41
 Delta R.T. -0.000 min
 Lab File: VX046726.D
 Acq: 17 Jun 2025 18:00




Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX061725


Tgt Ion: 43 Resp: 1044679
 Ion Ratio Lower Upper
 43 100
 86 10.2 8.2 12.2



#24
 1,1-Dichloroethane
 Concen: 50.502 ug/l
 RT: 3.623 min Scan# 416
 Delta R.T. -0.000 min
 Lab File: VX046726.D
 Acq: 17 Jun 2025 18:00

Tgt Ion: 63 Resp: 142442
 Ion Ratio Lower Upper
 63 100
 98 7.8 4.0 11.9
 100 4.7 2.2 6.6





#25

2-Butanone

Concen: 264.542 ug/l

RT: 4.562 min Scan# 51

Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00

Instrument :

MSVOA_X

ClientSampleId :





ICVVX061725

Tgt Ion: 43 Resp: 222870

Ion Ratio Lower Upper

43 100

72 24.1 20.2 30.4



#26

2,2-Dichloropropane

Concen: 50.629 ug/l

RT: 4.483 min Scan# 557

Delta R.T. -0.006 min

Lab File: VX046726.D




Acq: 17 Jun 2025 18:00


Tgt Ion: 77 Resp: 104151

Ion Ratio Lower Upper

77 100

97 23.0 11.8 35.3









#27
 cis-1,2-Dichloroethene
 Concen: 50.003 ug/l
 RT: 4.501 min Scan# 560
 Delta R.T. -0.000 min
 Lab File: VX046726.D
 Acq: 17 Jun 2025 18:00




Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX061725


Tgt Ion: 96 Resp: 89665
 Ion Ratio Lower Upper
 96 100
 61 145.1 0.0 291.0
 98 64.2 0.0 126.8



#28
 Bromochloromethane
 Concen: 50.308 ug/l
 RT: 4.904 min Scan# 626
 Delta R.T. -0.006 min
 Lab File: VX046726.D
 Acq: 17 Jun 2025 18:00

Tgt Ion: 49 Resp: 64330
 Ion Ratio Lower Upper
 49 100
 129 1.5 0.0 3.8
 130 75.6 61.4 92.0





#29

Tetrahydrofuran

Concen: 263.836 ug/l

RT: 5.013 min Scan# 644

Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00

Instrument :

MSVOA_X

ClientSampleId :

ICVVX061725





Tgt Ion: 42 Resp: 143570

Ion Ratio Lower Upper

42 100

72 44.3 34.9 52.3

71 40.8 32.2 48.2



#30

Chloroform

Concen: 51.089 ug/l

RT: 5.105 min Scan# 659

Delta R.T. -0.000 min

Lab File: VX046726.D




Acq: 17 Jun 2025 18:00


Tgt Ion: 83 Resp: 144108

Ion Ratio Lower Upper

83 100

85 65.4 53.7 80.5









#31
Cyclohexane
Concen: 49.462 ug/l
RT: 5.483 min Scan# 701
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00




Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725


Tgt Ion: 56 Resp: 132301
Ion Ratio Lower Upper
56 100
69 30.8 24.8 37.2
84 86.8 68.2 102.2



#32
1,1,1-Trichloroethane
Concen: 50.042 ug/l
RT: 5.391 min Scan# 706
Delta R.T. -0.006 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Tgt Ion: 97 Resp: 123796
Ion Ratio Lower Upper
97 100
99 63.5 51.3 76.9
61 45.1 36.4 54.6









#33
 1,2-Dichloroethane-d4
 Concen: 45.697 ug/l
 RT: 5.958 min Scan# 799
 Delta R.T. -0.006 min
 Lab File: VX046726.D
 Acq: 17 Jun 2025 18:00




Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX061725


Tgt Ion: 65 Resp: 81629
 Ion Ratio Lower Upper
 65 100
 67 52.5 0.0 105.4



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.769 min Scan# 932
 Delta R.T. 0.000 min
 Lab File: VX046726.D
 Acq: 17 Jun 2025 18:00

Tgt Ion: 114 Resp: 209541
 Ion Ratio Lower Upper
 114 100
 63 20.5 0.0 44.2
 88 16.2 0.0 33.2









#35
Dibromofluoromethane
Concen: 48.072 ug/l
RT: 5.397 min Scan# 707
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00




Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725


Tgt Ion:113 Resp: 66647
Ion Ratio Lower Upper
113 100
111 102.5 82.0 123.0
192 18.9 15.3 22.9



#36
1,1-Dichloropropene
Concen: 50.088 ug/l
RT: 5.702 min Scan# 757
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Tgt Ion: 75 Resp: 100241
Ion Ratio Lower Upper
75 100
110 35.3 17.2 51.5
77 30.7 24.2 36.4









#37
Ethyl Acetate
Concen: 51.401 ug/l
RT: 4.721 min Scan# 596
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00




Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725


Tgt Ion: 43 Resp: 94953
Ion Ratio Lower Upper
43 100
61 12.6 9.7 14.5
70 10.1 8.1 12.1



#38
Carbon Tetrachloride
Concen: 50.403 ug/l
RT: 5.690 min Scan# 755
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Tgt Ion: 117 Resp: 109471
Ion Ratio Lower Upper
117 100
119 96.3 75.2 112.8
121 31.6 24.1 36.1









#39
Methylcyclohexane
Concen: 50.155 ug/l
RT: 7.385 min Scan# 1033
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00




Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725


Tgt Ion: 83 Resp: 132086
Ion Ratio Lower Upper
83 100
55 81.1 63.0 94.4
98 50.5 38.8 58.2



#40
Benzene
Concen: 51.385 ug/l
RT: 6.044 min Scan# 813
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Tgt Ion: 78 Resp: 304346
Ion Ratio Lower Upper
78 100
77 23.4 18.8 28.2









#41
Methacrylonitrile
Concen: 55.341 ug/l
RT: 4.928 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00




Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725


Tgt Ion:	41	Resp:	54501
Ion	Ratio	Lower	Upper
41	100		
39	55.5	45.9	68.9
67	72.5	61.2	91.8
52	30.7	25.5	38.3



#42
1,2-Dichloroethane
Concen: 50.970 ug/l
RT: 6.099 min Scan# 822
Delta R.T. 0.006 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Tgt Ion:	62	Resp:	106085
Ion	Ratio	Lower	Upper
62	100		
98	9.4	0.0	17.8









#43
Isopropyl Acetate
Concen: 53.888 ug/l
RT: 6.342 min Scan# 862
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00




Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725


Tgt Ion: 43 Resp: 155773
Ion Ratio Lower Upper
43 100
61 18.8 15.3 22.9
87 12.2 10.0 15.0



#44
Trichloroethene
Concen: 51.000 ug/l
RT: 7.129 min Scan# 991
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Tgt Ion:130 Resp: 76981
Ion Ratio Lower Upper
130 100
95 98.6 0.0 202.6









#45
1,2-Dichloropropane
Concen: 50.940 ug/l
RT: 7.434 min Scan# 1041
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00




Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725


Tgt Ion: 63 Resp: 74773
Ion Ratio Lower Upper
63 100
65 31.9 22.9 34.3



#46
Dibromomethane
Concen: 51.329 ug/l
RT: 7.586 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Tgt Ion: 93 Resp: 54149
Ion Ratio Lower Upper
93 100
95 82.7 66.2 99.4
174 96.3 77.0 115.6









#47
Bromodichloromethane
Concen: 51.479 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00




Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725


Tgt Ion: 83 Resp: 110823
Ion Ratio Lower Upper
83 100
85 64.7 51.4 77.2
127 8.4 6.9 10.3



#48
Methyl methacrylate
Concen: 55.091 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Tgt Ion: 41 Resp: 79868
Ion Ratio Lower Upper
41 100
69 80.6 63.4 95.2
39 55.9 45.4 68.0





#49

1,4-Dioxane

Concen: 964.253 ug/l

RT: 7.659 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00

Instrument :

MSVOA_X

ClientSampleId :

ICVVX061725





Tgt Ion: 88 Resp: 20314

Ion Ratio Lower Upper

88 100

43 38.3 31.2 46.8

58 72.0 54.7 82.1



#50

Toluene-d8

Concen: 47.537 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. -0.000 min

Lab File: VX046726.D




Acq: 17 Jun 2025 18:00


Tgt Ion: 98 Resp: 236781

Ion Ratio Lower Upper

98 100

100 66.9 53.0 79.4









#51
4-Methyl-2-Pentanone
Concen: 272.916 ug/l
RT: 8.574 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00




Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725


Tgt Ion: 43 Resp: 470527
Ion Ratio Lower Upper
43 100
58 38.2 30.8 46.2



#52
Toluene
Concen: 51.237 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Tgt Ion: 92 Resp: 188131
Ion Ratio Lower Upper
92 100
91 170.4 134.8 202.2





#53

t-1,3-Dichloropropene

Concen: 52.785 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00

Instrument :

MSVOA_X

ClientSampleId :





ICVVX061725

Tgt Ion: 75 Resp: 105435

Ion Ratio Lower Upper

75 100

77 32.4 25.6 38.4



#54

cis-1,3-Dichloropropene

Concen: 51.813 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00




Tgt Ion: 75 Resp: 117487


Ion Ratio Lower Upper

75 100

77 31.5 25.3 37.9

39 52.2 39.2 58.8









#55
1,1,2-Trichloroethane
Concen: 51.698 ug/l
RT: 9.153 min Scan# 1323
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00




Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725


Tgt Ion:	97	Resp:	70739
Ion	Ratio	Lower	Upper
97	100		
83	83.8	67.2	100.8
85	55.5	44.1	66.1
99	64.8	50.3	75.5



#56
Ethyl methacrylate
Concen: 54.266 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Tgt Ion:	69	Resp:	109612
Ion	Ratio	Lower	Upper
69	100		
41	69.3	55.4	83.2
39	36.7	29.7	44.5





#57

1,3-Dichloropropane

Concen: 51.625 ug/l

RT: 9.305 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00

Instrument :

MSVOA_X

ClientSampleId :





ICVVX061725

Tgt Ion: 76 Resp: 121649

Ion Ratio Lower Upper

76 100

78 32.0 25.7 38.5



#58

2-Chloroethyl Vinyl ether

Concen: 279.922 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. -0.000 min

Lab File: VX046726.D




Acq: 17 Jun 2025 18:00


Tgt Ion: 63 Resp: 293709

Ion Ratio Lower Upper

63 100

106 28.3 22.6 34.0









#59
2-Hexanone
Concen: 273.982 ug/l
RT: 9.427 min Scan# 1368
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00




Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725


Tgt Ion: 43 Resp: 326564
Ion Ratio Lower Upper
43 100
58 54.1 26.8 80.4



#60
Dibromochloromethane
Concen: 50.997 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Tgt Ion: 129 Resp: 82234
Ion Ratio Lower Upper
129 100
127 77.7 38.4 115.2









#61
1,2-Dibromoethane
Concen: 52.268 ug/l
RT: 9.610 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00




Instrument :
MSVOA_X
ClientSampleId :
ICVX061725


Tgt Ion:107 Resp: 72749
Ion Ratio Lower Upper
107 100
109 92.8 75.8 113.6



#62
4-Bromofluorobenzene
Concen: 47.712 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Tgt Ion: 95 Resp: 88588
Ion Ratio Lower Upper
95 100
174 75.6 0.0 150.4
176 73.6 0.0 145.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00




Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725


Tgt Ion:117 Resp: 184761
Ion Ratio Lower Upper
117 100
82 55.0 44.6 66.8
119 32.9 25.8 38.8



#64
Tetrachloroethene
Concen: 49.172 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Tgt Ion:164 Resp: 62943
Ion Ratio Lower Upper
164 100
166 127.1 104.6 157.0
129 94.2 76.3 114.5
131 91.2 74.1 111.1









#65
Chlorobenzene
Concen: 50.006 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00




Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725


Tgt Ion:112 Resp: 203947
Ion Ratio Lower Upper
112 100
114 32.2 25.5 38.3



#66
1,1,1,2-Tetrachloroethane
Concen: 50.856 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Tgt Ion:131 Resp: 69709
Ion Ratio Lower Upper
131 100
133 95.2 47.3 141.8
119 63.5 32.1 96.3





#67

Ethyl Benzene

Concen: 50.487 ug/l

RT: 10.195 min Scan# 1493

Delta R.T. 0.006 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00

Instrument :

MSVOA_X

ClientSampleId :





ICVVX061725

Tgt Ion: 91 Resp: 359779

Ion Ratio Lower Upper

91 100

106 31.4 24.2 36.2



#68

m/p-Xylenes

Concen: 102.789 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. -0.000 min

Lab File: VX046726.D




Acq: 17 Jun 2025 18:00


Tgt Ion: 106 Resp: 272935

Ion Ratio Lower Upper

106 100

91 205.2 164.8 247.2









#69
o-Xylene
Concen: 52.635 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00




Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725


Tgt Ion:106 Resp: 130980
Ion Ratio Lower Upper
106 100
91 218.4 111.3 333.9



#70
Styrene
Concen: 52.139 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Tgt Ion:104 Resp: 227060
Ion Ratio Lower Upper
104 100
78 51.7 41.4 62.2
103 54.1 43.1 64.7









#71
Bromoform
Concen: 51.404 ug/l
RT: 10.799 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00




Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725


Tgt Ion:173 Resp: 53660
Ion Ratio Lower Upper
173 100
175 48.6 24.1 72.4
254 0.0 0.1 0.1#



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Tgt Ion:152 Resp: 91992
Ion Ratio Lower Upper
152 100
115 86.9 43.2 129.6
150 174.0 0.0 346.8









#73
Isopropylbenzene
Concen: 51.083 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00




Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725


Tgt Ion:105 Resp: 346098
Ion Ratio Lower Upper
105 100
120 26.0 13.0 39.0



#74
N-ethyl acetate
Concen: 54.107 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Tgt Ion: 43 Resp: 138890
Ion Ratio Lower Upper
43 100
70 41.5 33.4 50.0
55 23.9 19.1 28.7
61 23.9 19.0 28.4





#75

1,1,2,2-Tetrachloroethane

Concen: 52.310 ug/l

RT: 11.207 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00

Instrument :

MSVOA_X

ClientSampleId :

ICVVX061725





Tgt Ion: 83 Resp: 98960

Ion Ratio Lower Upper

83 100

131 9.9 4.9 14.6

85 64.7 32.0 96.0



#76

1,2,3-Trichloropropane

Concen: 62.974 ug/l

RT: 11.238 min Scan# 1665

Delta R.T. -0.000 min

Lab File: VX046726.D




Acq: 17 Jun 2025 18:00


Tgt Ion: 75 Resp: 106102

Ion Ratio Lower Upper

75 100

77 32.1 18.5 55.5





#77

Bromobenzene

Concen: 50.611 ug/l

RT: 11.195 min Scan# 1658

Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00

Instrument :

MSVOA_X

ClientSampleId :

ICVVX061725





Tgt Ion:156 Resp: 82515

Ion Ratio Lower Upper

156 100

77 167.0 82.8 248.6

158 96.2 47.6 142.9



#78

n-propylbenzene

Concen: 51.240 ug/l

RT: 11.305 min Scan# 1676

Delta R.T. 0.006 min

Lab File: VX046726.D




Acq: 17 Jun 2025 18:00


Tgt Ion: 91 Resp: 412345

Ion Ratio Lower Upper

91 100

120 22.6 11.4 34.1









#79
2-Chlorotoluene
Concen: 49.931 ug/l
RT: 11.360 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00




Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725


Tgt Ion: 91 Resp: 239904
Ion Ratio Lower Upper
91 100
126 33.5 16.6 49.7



#80
1,3,5-Trimethylbenzene
Concen: 51.746 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Tgt Ion: 105 Resp: 288104
Ion Ratio Lower Upper
105 100
120 47.6 24.1 72.2









#81
trans-1,4-Dichloro-2-butene
Concen: 50.210 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00




Instrument :
MSVOA_X
ClientSampleId :
ICVVX061725


Tgt Ion: 75 Resp: 29929
Ion Ratio Lower Upper
75 100
53 109.9 86.8 130.2
89 45.3 36.9 55.3



#82
4-Chlorotoluene
Concen: 50.085 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046726.D
Acq: 17 Jun 2025 18:00

Tgt Ion: 91 Resp: 280848
Ion Ratio Lower Upper
91 100
126 29.8 14.4 43.2





#83

tert-Butylbenzene

Concen: 49.958 ug/l

RT: 11.713 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00

Instrument :

MSVOA_X

ClientSampleId :

ICVX061725





Tgt Ion:119 Resp: 287513

Ion Ratio Lower Upper

119 100

91 60.6 29.5 88.6

134 23.5 11.7 35.0



#84

1,2,4-Trimethylbenzene

Concen: 51.193 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. -0.000 min

Lab File: VX046726.D




Acq: 17 Jun 2025 18:00


Tgt Ion:105 Resp: 286365

Ion Ratio Lower Upper

105 100

120 44.7 22.1 66.3









#85
 sec-Butylbenzene
 Concen: 51.079 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. -0.000 min
 Lab File: VX046726.D
 Acq: 17 Jun 2025 18:00




Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX061725


Tgt Ion:105 Resp: 371523
 Ion Ratio Lower Upper
 105 100
 134 19.7 9.8 29.4



#86
 p-Isopropyltoluene
 Concen: 50.791 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. -0.000 min
 Lab File: VX046726.D
 Acq: 17 Jun 2025 18:00

Tgt Ion:119 Resp: 311698
 Ion Ratio Lower Upper
 119 100
 134 26.0 12.9 38.7
 91 24.6 12.3 36.9





#87

1,3-Dichlorobenzene

Concen: 48.501 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00

Instrument :

MSVOA_X

ClientSampleId :

ICVVX061725





Tgt Ion:146 Resp: 154181

Ion Ratio Lower Upper

146 100

111 41.8 20.6 61.8

148 64.2 31.6 94.8



#88

1,4-Dichlorobenzene

Concen: 46.906 ug/l

RT: 12.036 min Scan# 1796

Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00




Tgt Ion:146 Resp: 153158


Ion Ratio Lower Upper

146 100

111 41.0 20.3 61.0

148 64.0 32.1 96.5





#89

n-Butylbenzene

Concen: 50.829 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00

Instrument :

MSVOA_X

ClientSampleId :

ICVVX061725





Tgt Ion: 91 Resp: 296174

Ion Ratio Lower Upper

91 100

92 54.2 27.1 81.3

134 25.4 12.9 38.6



#90

Hexachloroethane

Concen: 49.161 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX046726.D




Acq: 17 Jun 2025 18:00


Tgt Ion: 117 Resp: 52831

Ion Ratio Lower Upper

117 100

201 70.3 34.4 103.3









#91
 1,2-Dichlorobenzene
 Concen: 49.672 ug/l
 RT: 12.335 min Scan# 1845
 Delta R.T. 0.006 min
 Lab File: VX046726.D
 Acq: 17 Jun 2025 18:00




Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX061725


Tgt Ion:146 Resp: 147561
 Ion Ratio Lower Upper
 146 100
 111 43.4 21.7 65.1
 148 63.5 32.0 96.2



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 53.151 ug/l
 RT: 12.939 min Scan# 1944
 Delta R.T. -0.000 min
 Lab File: VX046726.D
 Acq: 17 Jun 2025 18:00

Tgt Ion: 75 Resp: 19851
 Ion Ratio Lower Upper
 75 100
 155 84.6 41.9 125.7
 157 107.2 51.4 154.3





#93

1,2,4-Trichlorobenzene

Concen: 49.369 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00

Instrument :

MSVOA_X

ClientSampleId :

ICVVX061725





Tgt Ion:180 Resp: 104267

Ion Ratio Lower Upper

180 100

182 94.4 46.7 140.1

145 32.5 15.8 47.4



#94

Hexachlorobutadiene

Concen: 51.729 ug/l

RT: 13.719 min Scan# 2072

Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00




Tgt Ion:225 Resp: 45958


Ion Ratio Lower Upper

225 100

223 63.4 32.5 97.5

227 62.5 32.2 96.6





#95

Naphthalene

Concen: 52.462 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00

Instrument :

MSVOA_X

ClientSampleId :

ICVVX061725





Tgt Ion:128 Resp: 315669

Ion Ratio Lower Upper

128 100

127 12.7 10.1 15.1

129 10.9 8.7 13.1



#96

1,2,3-Trichlorobenzene

Concen: 51.200 ug/l

RT: 13.957 min Scan# 2111

Delta R.T. -0.000 min

Lab File: VX046726.D

Acq: 17 Jun 2025 18:00




Tgt Ion:180 Resp: 103451

Ion Ratio Lower Upper

180 100

182 94.5 47.8 143.4

145 33.9 17.3 51.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
 Data File : VX046758.D
 Acq On : 19 Jun 2025 09:43
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Quant Time: Jun 20 05:14:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	137182	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	224265	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	197740	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	96867	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	87638	46.187	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.380%
35) Dibromofluoromethane	5.391	113	73413	49.476	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.960%
50) Toluene-d8	8.647	98	261576	49.067	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.140%
62) 4-Bromofluorobenzene	11.079	95	97021	48.823	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	97.640%

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.185	85	77571	52.964	ug/l	99
3) Chloromethane	1.307	50	81669	51.662	ug/l	96
4) Vinyl Chloride	1.386	62	87898	52.141	ug/l	96
5) Bromomethane	1.624	94	52745	55.612	ug/l	94
6) Chloroethane	1.703	64	53086	51.957	ug/l	98
7) Trichlorofluoromethane	1.904	101	131261	51.280	ug/l	99
8) Diethyl Ether	2.142	74	44511	50.670	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.343	101	81171	51.691	ug/l	98
10) Methyl Iodide	2.465	142	94903	48.429	ug/l	99
11) Tert butyl alcohol	2.947	59	26783	157.262	ug/l	100
12) 1,1-Dichloroethene	2.337	96	78427	51.759	ug/l	97
13) Acrolein	2.245	56	50116	202.026	ug/l	98
14) Allyl chloride	2.678	41	136507	49.722	ug/l	99
15) Acrylonitrile	3.068	53	162714	212.474	ug/l	99
16) Acetone	2.380	43	127043	200.776	ug/l	99
17) Carbon Disulfide	2.526	76	218810	47.820	ug/l	99
18) Methyl Acetate	2.709	43	65851	40.975	ug/l	99
19) Methyl tert-butyl Ether	3.117	73	215275	45.628	ug/l	100
20) Methylene Chloride	2.800	84	84742	48.158	ug/l	99
21) trans-1,2-Dichloroethene	3.105	96	80577	49.698	ug/l	96
22) Diisopropyl ether	3.763	45	269978	51.162	ug/l	97
23) Vinyl Acetate	3.727	43	1003345	239.105	ug/l	100
24) 1,1-Dichloroethane	3.623	63	148260	49.484	ug/l	100
25) 2-Butanone	4.556	43	177052	197.843	ug/l	95
26) 2,2-Dichloropropane	4.483	77	112913	51.672	ug/l	99
27) cis-1,2-Dichloroethene	4.495	96	94126	49.415	ug/l	99
28) Bromochloromethane	4.904	49	65263	48.047	ug/l	100
29) Tetrahydrofuran	5.007	42	116160	200.958	ug/l	99
30) Chloroform	5.099	83	149922	50.036	ug/l	95
31) Cyclohexane	5.477	56	139626	49.142	ug/l	97
32) 1,1,1-Trichloroethane	5.385	97	128769	49.003	ug/l	100
36) 1,1-Dichloropropene	5.702	75	106011	49.493	ug/l	99
37) Ethyl Acetate	4.715	43	79324	40.121	ug/l	99
38) Carbon Tetrachloride	5.684	117	115565	49.715	ug/l	98
39) Methylcyclohexane	7.379	83	140502	49.848	ug/l	96
40) Benzene	6.044	78	317423	50.074	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
 Data File : VX046758.D
 Acq On : 19 Jun 2025 09:43
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Quant Time: Jun 20 05:14:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	47871	45.417	ug/l	95
42) 1,2-Dichloroethane	6.086	62	107931	48.452	ug/l	99
43) Isopropyl Acetate	6.336	43	132833	42.935	ug/l	99
44) Trichloroethene	7.129	130	80127	49.599	ug/l	96
45) 1,2-Dichloropropane	7.427	63	79156	50.385	ug/l	95
46) Dibromomethane	7.580	93	53399	47.295	ug/l	99
47) Bromodichloromethane	7.818	83	116822	50.703	ug/l	100
48) Methyl methacrylate	7.690	41	70314	45.316	ug/l	99
49) 1,4-Dioxane	7.659	88	15950	714.909	ug/l	96
51) 4-Methyl-2-Pentanone	8.568	43	379704	205.777	ug/l	100
52) Toluene	8.714	92	196587	50.025	ug/l	97
53) t-1,3-Dichloropropene	8.976	75	107788	50.420	ug/l	99
54) cis-1,3-Dichloropropene	8.366	75	122151	50.333	ug/l	99
55) 1,1,2-Trichloroethane	9.147	97	69729	47.614	ug/l	98
56) Ethyl methacrylate	9.116	69	99225	45.899	ug/l	100
57) 1,3-Dichloropropane	9.305	76	122083	48.408	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.238	63	279394	248.797	ug/l	100
59) 2-Hexanone	9.427	43	257015	201.475	ug/l	100
60) Dibromochloromethane	9.519	129	84603	49.021	ug/l	99
61) 1,2-Dibromoethane	9.610	107	70987	47.653	ug/l	100
64) Tetrachloroethene	9.275	164	67934	49.587	ug/l	98
65) Chlorobenzene	10.073	112	215423	49.352	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	73797	50.304	ug/l	100
67) Ethyl Benzene	10.189	91	379202	49.719	ug/l	100
68) m/p-Xylenes	10.299	106	286180	100.703	ug/l	99
69) o-Xylene	10.640	106	135837	51.004	ug/l	99
70) Styrene	10.652	104	235709	50.573	ug/l	100
71) Bromoform	10.799	173	51917	46.470	ug/l #	100
73) Isopropylbenzene	10.957	105	362998	50.881	ug/l	100
74) N-amyl acetate	10.841	43	120822	44.699	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	90340	45.350	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	95570	53.868	ug/l	91
77) Bromobenzene	11.195	156	86326	50.284	ug/l	98
78) n-propylbenzene	11.299	91	434668	51.296	ug/l	100
79) 2-Chlorotoluene	11.360	91	250583	49.529	ug/l	99
80) 1,3,5-Trimethylbenzene	11.445	105	294565	50.244	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	27012	43.036	ug/l	98
82) 4-Chlorotoluene	11.451	91	294491	49.875	ug/l	100
83) tert-Butylbenzene	11.713	119	302344	49.891	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	298590	50.692	ug/l	100
85) sec-Butylbenzene	11.890	105	389381	50.841	ug/l	99
86) p-Isopropyltoluene	12.006	119	324490	50.214	ug/l	100
87) 1,3-Dichlorobenzene	11.963	146	161807	48.338	ug/l	100
88) 1,4-Dichlorobenzene	12.036	146	164023	47.705	ug/l	98
89) n-Butylbenzene	12.329	91	310154	50.550	ug/l	100
90) Hexachloroethane	12.536	117	55264	48.837	ug/l	96
91) 1,2-Dichlorobenzene	12.329	146	154832	49.496	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	15868	40.348	ug/l	97
93) 1,2,4-Trichlorobenzene	13.585	180	107660	48.410	ug/l	98
94) Hexachlorobutadiene	13.719	225	46034	49.207	ug/l	99
95) Naphthalene	13.774	128	286491	45.216	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	103088	48.453	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
Data File : VX046758.D
Acq On : 19 Jun 2025 09:43
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1


Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Quant Time: Jun 20 05:14:38 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 03:09:16 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jun 20 05:14:38 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 03:09:16 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:168 Resp: 137182
Ion Ratio Lower Upper
168 100
99 59.5 48.5 72.7



#2
Dichlorodifluoromethane
Concen: 52.964 ug/l
RT: 1.185 min Scan# 16
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Tgt Ion: 85 Resp: 77571
Ion Ratio Lower Upper
85 100
87 32.7 16.6 49.8









#3
Chloromethane
Concen: 51.662 ug/l
RT: 1.307 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 50 Resp: 81669
Ion Ratio Lower Upper
50 100
52 33.4 24.9 37.3



#4
Vinyl Chloride
Concen: 52.141 ug/l
RT: 1.386 min Scan# 49
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Tgt Ion: 62 Resp: 87898
Ion Ratio Lower Upper
62 100
64 31.8 23.9 35.9









#5
Bromomethane
Concen: 55.612 ug/l
RT: 1.624 min Scan# 88
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 94 Resp: 52745
Ion Ratio Lower Upper
94 100
96 96.1 72.3 108.5



#6
Chloroethane
Concen: 51.957 ug/l
RT: 1.703 min Scan# 101
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Tgt Ion: 64 Resp: 53086
Ion Ratio Lower Upper
64 100
66 32.8 27.1 40.7









#7
 Trichlorofluoromethane
 Concen: 51.280 ug/l
 RT: 1.904 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: VX046758.D
 Acq: 19 Jun 2025 09:43




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tgt Ion:101 Resp: 131261
 Ion Ratio Lower Upper
 101 100
 103 65.7 51.8 77.8



#8
 Diethyl Ether
 Concen: 50.670 ug/l
 RT: 2.142 min Scan# 173
 Delta R.T. -0.006 min
 Lab File: VX046758.D
 Acq: 19 Jun 2025 09:43

Tgt Ion: 74 Resp: 44511
 Ion Ratio Lower Upper
 74 100
 45 103.6 52.3 157.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 51.691 ug/l
RT: 2.343 min Scan# 206
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:101 Resp: 81171
Ion Ratio Lower Upper
101 100
85 42.5 35.4 53.0
151 74.7 61.2 91.8



#10
Methyl Iodide
Concen: 48.429 ug/l
RT: 2.465 min Scan# 226
Delta R.T. -0.006 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Tgt Ion:142 Resp: 94903
Ion Ratio Lower Upper
142 100
127 45.2 37.0 55.6
141 14.2 11.5 17.3









#11
Tert butyl alcohol
Concen: 157.262 ug/l
RT: 2.947 min Scan# 307
Delta R.T. -0.012 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 59 Resp: 26783
Ion Ratio Lower Upper
59 100
57 10.6 8.3 12.5



#12
1,1-Dichloroethene
Concen: 51.759 ug/l
RT: 2.337 min Scan# 205
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Tgt Ion: 96 Resp: 78427
Ion Ratio Lower Upper
96 100
61 163.0 126.6 189.8
98 64.7 50.7 76.1





#13

Acrolein

Concen: 202.026 ug/l

RT: 2.245 min Scan# 190

Delta R.T. -0.000 min

Lab File: VX046758.D

Acq: 19 Jun 2025 09:43

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion: 56 Resp: 50116

Ion Ratio Lower Upper

56 100

55 71.0 55.2 82.8



#14

Allyl chloride

Concen: 49.722 ug/l

RT: 2.678 min Scan# 261

Delta R.T. -0.000 min

Lab File: VX046758.D

Acq: 19 Jun 2025 09:43




Tgt Ion: 41 Resp: 136507


Ion Ratio Lower Upper

41 100

39 66.2 53.6 80.4

76 33.3 26.8 40.2









#15
Acrylonitrile
Concen: 212.474 ug/l
RT: 3.068 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 53 Resp: 162714
Ion Ratio Lower Upper
53 100
52 82.0 65.1 97.7
51 35.9 28.9 43.3



#16
Acetone
Concen: 200.776 ug/l
RT: 2.380 min Scan# 212
Delta R.T. -0.006 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Tgt Ion: 43 Resp: 127043
Ion Ratio Lower Upper
43 100
58 30.2 24.5 36.7









#17
Carbon Disulfide
Concen: 47.820 ug/l
RT: 2.526 min Scan# 21
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 76 Resp: 218810
Ion Ratio Lower Upper
76 100
78 9.1 7.5 11.3



#18
Methyl Acetate
Concen: 40.975 ug/l
RT: 2.709 min Scan# 266
Delta R.T. -0.006 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Tgt Ion: 43 Resp: 65851
Ion Ratio Lower Upper
43 100
74 22.4 18.4 27.6









#19
Methyl tert-butyl Ether
Concen: 45.628 ug/l
RT: 3.117 min Scan# 311
Delta R.T. -0.006 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 73 Resp: 215275
Ion Ratio Lower Upper
73 100
57 22.7 18.2 27.2



#20
Methylene Chloride
Concen: 48.158 ug/l
RT: 2.800 min Scan# 281
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Tgt Ion: 84 Resp: 84742
Ion Ratio Lower Upper
84 100
49 127.5 101.3 151.9
51 37.8 31.0 46.6
86 64.8 51.1 76.7









#21
trans-1,2-Dichloroethene
Concen: 49.698 ug/l
RT: 3.105 min Scan# 311
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 96 Resp: 80577
Ion Ratio Lower Upper
96 100
61 134.5 112.5 168.7
98 63.6 51.9 77.9



#22
Diisopropyl ether
Concen: 51.162 ug/l
RT: 3.763 min Scan# 439
Delta R.T. -0.006 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Tgt Ion: 45 Resp: 269978
Ion Ratio Lower Upper
45 100
43 110.7 91.9 137.9
87 25.9 22.0 33.0
59 11.1 9.1 13.7









#23
 Vinyl Acetate
 Concen: 239.105 ug/l
 RT: 3.727 min Scan# 411
 Delta R.T. -0.006 min
 Lab File: VX046758.D
 Acq: 19 Jun 2025 09:43




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tgt Ion: 43 Resp: 1003345
 Ion Ratio Lower Upper
 43 100
 86 10.2 8.2 12.2



#24
 1,1-Dichloroethane
 Concen: 49.484 ug/l
 RT: 3.623 min Scan# 416
 Delta R.T. -0.000 min
 Lab File: VX046758.D
 Acq: 19 Jun 2025 09:43

Tgt Ion: 63 Resp: 148260
 Ion Ratio Lower Upper
 63 100
 98 7.8 4.0 11.9
 100 4.6 2.2 6.6





#25

2-Butanone

Concen: 197.843 ug/l

RT: 4.556 min Scan# 5

Delta R.T. -0.006 min

Lab File: VX046758.D

Acq: 19 Jun 2025 09:43

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion: 43 Resp: 177052

Ion Ratio Lower Upper

43 100

72 23.0 20.2 30.4



#26

2,2-Dichloropropane

Concen: 51.672 ug/l

RT: 4.483 min Scan# 557

Delta R.T. -0.006 min

Lab File: VX046758.D




Acq: 19 Jun 2025 09:43


Tgt Ion: 77 Resp: 112913

Ion Ratio Lower Upper

77 100

97 22.8 11.8 35.3









#27
 cis-1,2-Dichloroethene
 Concen: 49.415 ug/l
 RT: 4.495 min Scan# 51
 Delta R.T. -0.006 min
 Lab File: VX046758.D
 Acq: 19 Jun 2025 09:43




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tgt Ion: 96 Resp: 94126
 Ion Ratio Lower Upper
 96 100
 61 145.1 0.0 291.0
 98 65.4 0.0 126.8



#28
 Bromochloromethane
 Concen: 48.047 ug/l
 RT: 4.904 min Scan# 626
 Delta R.T. -0.006 min
 Lab File: VX046758.D
 Acq: 19 Jun 2025 09:43

Tgt Ion: 49 Resp: 65263
 Ion Ratio Lower Upper
 49 100
 129 1.9 0.0 3.8
 130 76.8 61.4 92.0





#29

Tetrahydrofuran

Concen: 200.958 ug/l

RT: 5.007 min Scan# 64

Delta R.T. -0.006 min

Lab File: VX046758.D

Acq: 19 Jun 2025 09:43

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion: 42 Resp: 116160

Ion Ratio Lower Upper

42 100

72 42.9 34.9 52.3

71 39.1 32.2 48.2



#30

Chloroform

Concen: 50.036 ug/l

RT: 5.099 min Scan# 658

Delta R.T. -0.006 min

Lab File: VX046758.D




Acq: 19 Jun 2025 09:43


Tgt Ion: 83 Resp: 149922

Ion Ratio Lower Upper

83 100

85 63.2 53.7 80.5









#31
Cyclohexane
Concen: 49.142 ug/l
RT: 5.477 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 56 Resp: 139626
Ion Ratio Lower Upper
56 100
69 29.6 24.8 37.2
84 82.9 68.2 102.2



#32
1,1,1-Trichloroethane
Concen: 49.003 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.012 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Tgt Ion: 97 Resp: 128769
Ion Ratio Lower Upper
97 100
99 64.0 51.3 76.9
61 45.4 36.4 54.6









#33
1,2-Dichloroethane-d4
Concen: 46.187 ug/l
RT: 5.958 min Scan# 799
Delta R.T. -0.006 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 65 Resp: 87638
Ion Ratio Lower Upper
65 100
67 53.3 0.0 105.4



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.763 min Scan# 931
Delta R.T. -0.006 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Tgt Ion: 114 Resp: 224265
Ion Ratio Lower Upper
114 100
63 20.7 0.0 44.2
88 16.2 0.0 33.2





#35

Dibromofluoromethane

Concen: 49.476 ug/l

RT: 5.391 min Scan# 707

Delta R.T. -0.006 min

Lab File: VX046758.D

Acq: 19 Jun 2025 09:43

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion:113 Resp: 73413

Ion Ratio Lower Upper

113 100

111 102.5 82.0 123.0

192 18.6 15.3 22.9



#36

1,1-Dichloropropene

Concen: 49.493 ug/l

RT: 5.702 min Scan# 757

Delta R.T. -0.000 min

Lab File: VX046758.D

Acq: 19 Jun 2025 09:43




Tgt Ion: 75 Resp: 106011


Ion Ratio Lower Upper

75 100

110 34.8 17.2 51.5

77 31.2 24.2 36.4





#37

Ethyl Acetate

Concen: 40.121 ug/l

RT: 4.715 min Scan# 595

Delta R.T. -0.006 min

Lab File: VX046758.D

Acq: 19 Jun 2025 09:43

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion: 43 Resp: 79324

Ion Ratio Lower Upper

43 100

61 11.6 9.7 14.5

70 10.2 8.1 12.1



#38

Carbon Tetrachloride

Concen: 49.715 ug/l

RT: 5.684 min Scan# 754

Delta R.T. -0.006 min

Lab File: VX046758.D

Acq: 19 Jun 2025 09:43




Tgt Ion: 117 Resp: 115565


Ion Ratio Lower Upper

117 100

119 92.4 75.2 112.8

121 32.1 24.1 36.1





#39

Methylcyclohexane

Concen: 49.848 ug/l

RT: 7.379 min Scan# 1

Delta R.T. -0.006 min

Lab File: VX046758.D

Acq: 19 Jun 2025 09:43

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion: 83 Resp: 140502

Ion Ratio Lower Upper

83 100

55 82.7 63.0 94.4

98 50.0 38.8 58.2



#40

Benzene

Concen: 50.074 ug/l

RT: 6.044 min Scan# 813

Delta R.T. -0.000 min

Lab File: VX046758.D




Acq: 19 Jun 2025 09:43


Tgt Ion: 78 Resp: 317423

Ion Ratio Lower Upper

78 100

77 23.7 18.8 28.2









#41
Methacrylonitrile
Concen: 45.417 ug/l
RT: 4.922 min Scan# 61
Delta R.T. -0.006 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion: 41 Resp: 47871




Ion	Ratio	Lower	Upper
41	100		
39	54.6	45.9	68.9
67	70.2	61.2	91.8
52	31.4	25.5	38.3




#42
1,2-Dichloroethane
Concen: 48.452 ug/l
RT: 6.086 min Scan# 820
Delta R.T. -0.006 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Tgt Ion: 62 Resp: 107931

Ion	Ratio	Lower	Upper
62	100		
98	9.3	0.0	17.8









#43
Isopropyl Acetate
Concen: 42.935 ug/l
RT: 6.336 min Scan# 861
Delta R.T. -0.006 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 132833
Ion Ratio Lower Upper
43 100
61 19.4 15.3 22.9
87 12.3 10.0 15.0



#44
Trichloroethene
Concen: 49.599 ug/l
RT: 7.129 min Scan# 991
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Tgt Ion:130 Resp: 80127
Ion Ratio Lower Upper
130 100
95 97.5 0.0 202.6





#45

1,2-Dichloropropane

Concen: 50.385 ug/l

RT: 7.427 min Scan# 1040

Delta R.T. -0.006 min

Lab File: VX046758.D

Acq: 19 Jun 2025 09:43

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion: 63 Resp: 79156

Ion Ratio Lower Upper

63 100

65 31.3 22.9 34.3



#46

Dibromomethane

Concen: 47.295 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. -0.006 min

Lab File: VX046758.D

Acq: 19 Jun 2025 09:43




Tgt Ion: 93 Resp: 53399


Ion Ratio Lower Upper

93 100

95 83.4 66.2 99.4

174 98.2 77.0 115.6









#47
Bromodichloromethane
Concen: 50.703 ug/l
RT: 7.818 min Scan# 1104
Delta R.T. -0.006 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 83 Resp: 116822
Ion Ratio Lower Upper
83 100
85 64.1 51.4 77.2
127 8.3 6.9 10.3



#48
Methyl methacrylate
Concen: 45.316 ug/l
RT: 7.690 min Scan# 1083
Delta R.T. -0.006 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Tgt Ion: 41 Resp: 70314
Ion Ratio Lower Upper
41 100
69 79.5 63.4 95.2
39 55.5 45.4 68.0





#49

1,4-Dioxane

Concen: 714.909 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. 0.000 min

Lab File: VX046758.D

Acq: 19 Jun 2025 09:43

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion: 88 Resp: 15950

Ion Ratio Lower Upper

88 100

43 37.3 31.2 46.8

58 72.4 54.7 82.1



#50

Toluene-d8

Concen: 49.067 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. -0.000 min

Lab File: VX046758.D




Acq: 19 Jun 2025 09:43


Tgt Ion: 98 Resp: 261576

Ion Ratio Lower Upper

98 100

100 67.6 53.0 79.4









#51
4-Methyl-2-Pentanone
Concen: 205.777 ug/l
RT: 8.568 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 379704
Ion Ratio Lower Upper
43 100
58 38.4 30.8 46.2



#52
Toluene
Concen: 50.025 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. -0.006 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Tgt Ion: 92 Resp: 196587
Ion Ratio Lower Upper
92 100
91 171.9 134.8 202.2





#53

t-1,3-Dichloropropene

Concen: 50.420 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046758.D

Acq: 19 Jun 2025 09:43

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion: 75 Resp: 107788

Ion Ratio Lower Upper

75 100

77 31.3 25.6 38.4



#54

cis-1,3-Dichloropropene

Concen: 50.333 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046758.D

Acq: 19 Jun 2025 09:43




Tgt Ion: 75 Resp: 122151


Ion Ratio Lower Upper

75 100

77 31.5 25.3 37.9

39 50.5 39.2 58.8









#55
1,1,2-Trichloroethane
Concen: 47.614 ug/l
RT: 9.147 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion: 97 Resp: 69729




Ion	Ratio	Lower	Upper
97	100		
83	86.3	67.2	100.8
85	55.4	44.1	66.1
99	65.3	50.3	75.5




#56
Ethyl methacrylate
Concen: 45.899 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Tgt Ion: 69 Resp: 99225

Ion	Ratio	Lower	Upper
69	100		
41	69.7	55.4	83.2
39	37.1	29.7	44.5





#57

1,3-Dichloropropane

Concen: 48.408 ug/l

RT: 9.305 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046758.D

Acq: 19 Jun 2025 09:43

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion: 76 Resp: 122083

Ion Ratio Lower Upper

76 100

78 32.4 25.7 38.5



#58

2-Chloroethyl Vinyl ether

Concen: 248.797 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. -0.006 min

Lab File: VX046758.D




Acq: 19 Jun 2025 09:43


Tgt Ion: 63 Resp: 279394

Ion Ratio Lower Upper

63 100

106 28.2 22.6 34.0









#59
2-Hexanone
Concen: 201.475 ug/l
RT: 9.427 min Scan# 1368
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 257015
Ion Ratio Lower Upper
43 100
58 53.4 26.8 80.4



#60
Dibromochloromethane
Concen: 49.021 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Tgt Ion:129 Resp: 84603
Ion Ratio Lower Upper
129 100
127 77.6 38.4 115.2









#61
1,2-Dibromoethane
Concen: 47.653 ug/l
RT: 9.610 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:107 Resp: 70987
Ion Ratio Lower Upper
107 100
109 94.8 75.8 113.6



#62
4-Bromofluorobenzene
Concen: 48.823 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Tgt Ion: 95 Resp: 97021
Ion Ratio Lower Upper
95 100
174 75.1 0.0 150.4
176 73.0 0.0 145.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1471
Delta R.T. -0.006 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:117 Resp: 197740
Ion Ratio Lower Upper
117 100
82 57.8 44.6 66.8
119 32.3 25.8 38.8



#64
Tetrachloroethene
Concen: 49.587 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Tgt Ion:164 Resp: 67934
Ion Ratio Lower Upper
164 100
166 128.0 104.6 157.0
129 95.0 76.3 114.5
131 91.1 74.1 111.1









#65
Chlorobenzene
Concen: 49.352 ug/l
RT: 10.073 min Scan# 1474
Delta R.T. -0.006 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:112 Resp: 215423
Ion Ratio Lower Upper
112 100
114 32.4 25.5 38.3



#66
1,1,1,2-Tetrachloroethane
Concen: 50.304 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Tgt Ion:131 Resp: 73797
Ion Ratio Lower Upper
131 100
133 94.0 47.3 141.8
119 64.5 32.1 96.3





#67

Ethyl Benzene

Concen: 49.719 ug/l

RT: 10.189 min Scan# 1493

Delta R.T. -0.000 min

Lab File: VX046758.D

Acq: 19 Jun 2025 09:43

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion: 91 Resp: 379202

Ion Ratio Lower Upper

91 100

106 30.4 24.2 36.2



#68

m/p-Xylenes

Concen: 100.703 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. -0.000 min

Lab File: VX046758.D




Acq: 19 Jun 2025 09:43


Tgt Ion:106 Resp: 286180

Ion Ratio Lower Upper

106 100

91 206.9 164.8 247.2









#69
o-Xylene
Concen: 51.004 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:106 Resp: 135837
Ion Ratio Lower Upper
106 100
91 220.3 111.3 333.9



#70
Styrene
Concen: 50.573 ug/l
RT: 10.652 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Tgt Ion:104 Resp: 235709
Ion Ratio Lower Upper
104 100
78 51.9 41.4 62.2
103 54.3 43.1 64.7





#71

Bromoform

Concen: 46.470 ug/l

RT: 10.799 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046758.D

Acq: 19 Jun 2025 09:43

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion:173 Resp: 51917

Ion Ratio Lower Upper

173 100

175 48.4 24.1 72.4

254 0.0 0.1 0.1#



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 1793

Delta R.T. -0.000 min

Lab File: VX046758.D

Acq: 19 Jun 2025 09:43




Tgt Ion:152 Resp: 96867


Ion Ratio Lower Upper

152 100

115 85.5 43.2 129.6

150 175.8 0.0 346.8









#73
Isopropylbenzene
Concen: 50.881 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:105 Resp: 362998
Ion Ratio Lower Upper
105 100
120 26.2 13.0 39.0



#74
N-ethyl acetate
Concen: 44.699 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Tgt Ion: 43 Resp: 120822
Ion Ratio Lower Upper
43 100
70 41.0 33.4 50.0
55 23.7 19.1 28.7
61 23.4 19.0 28.4









#75
 1,1,2,2-Tetrachloroethane
 Concen: 45.350 ug/l
 RT: 11.207 min Scan# 1660
 Delta R.T. -0.000 min
 Lab File: VX046758.D
 Acq: 19 Jun 2025 09:43




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tgt Ion: 83 Resp: 90340
 Ion Ratio Lower Upper
 83 100
 131 10.5 4.9 14.6
 85 63.3 32.0 96.0



#76
 1,2,3-Trichloropropane
 Concen: 53.868 ug/l
 RT: 11.238 min Scan# 1665
 Delta R.T. -0.000 min
 Lab File: VX046758.D
 Acq: 19 Jun 2025 09:43

Tgt Ion: 75 Resp: 95570
 Ion Ratio Lower Upper
 75 100
 77 31.5 18.5 55.5









#77
Bromobenzene
Concen: 50.284 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:156 Resp: 86326
Ion Ratio Lower Upper
156 100
77 167.7 82.8 248.6
158 98.3 47.6 142.9



#78
n-propylbenzene
Concen: 51.296 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Tgt Ion: 91 Resp: 434668
Ion Ratio Lower Upper
91 100
120 22.5 11.4 34.1









#79
2-Chlorotoluene
Concen: 49.529 ug/l
RT: 11.360 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 91 Resp: 250583
Ion Ratio Lower Upper
91 100
126 33.6 16.6 49.7



#80
1,3,5-Trimethylbenzene
Concen: 50.244 ug/l
RT: 11.445 min Scan# 1699
Delta R.T. -0.006 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Tgt Ion: 105 Resp: 294565
Ion Ratio Lower Upper
105 100
120 49.1 24.1 72.2









#81
trans-1,4-Dichloro-2-butene
Concen: 43.036 ug/l
RT: 11.018 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 75 Resp: 27012
Ion Ratio Lower Upper
75 100
53 109.1 86.8 130.2
89 50.3 36.9 55.3



#82
4-Chlorotoluene
Concen: 49.875 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Tgt Ion: 91 Resp: 294491
Ion Ratio Lower Upper
91 100
126 29.0 14.4 43.2









#83
 tert-Butylbenzene
 Concen: 49.891 ug/l
 RT: 11.713 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: VX046758.D
 Acq: 19 Jun 2025 09:43




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tgt Ion:119 Resp: 302344
 Ion Ratio Lower Upper
 119 100
 91 58.1 29.5 88.6
 134 23.0 11.7 35.0



#84
 1,2,4-Trimethylbenzene
 Concen: 50.692 ug/l
 RT: 11.750 min Scan# 1749
 Delta R.T. -0.000 min
 Lab File: VX046758.D
 Acq: 19 Jun 2025 09:43

Tgt Ion:105 Resp: 298590
 Ion Ratio Lower Upper
 105 100
 120 44.1 22.1 66.3









#85
 sec-Butylbenzene
 Concen: 50.841 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. -0.000 min
 Lab File: VX046758.D
 Acq: 19 Jun 2025 09:43




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tgt Ion:105 Resp: 389381
 Ion Ratio Lower Upper
 105 100
 134 19.9 9.8 29.4



#86
 p-Isopropyltoluene
 Concen: 50.214 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. -0.000 min
 Lab File: VX046758.D
 Acq: 19 Jun 2025 09:43

Tgt Ion:119 Resp: 324490
 Ion Ratio Lower Upper
 119 100
 134 25.7 12.9 38.7
 91 24.3 12.3 36.9





#87

1,3-Dichlorobenzene

Concen: 48.338 ug/l

RT: 11.963 min Scan# 1

Delta R.T. -0.006 min

Lab File: VX046758.D

Acq: 19 Jun 2025 09:43

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion:146 Resp: 161807

Ion Ratio Lower Upper

146 100

111 41.3 20.6 61.8

148 62.8 31.6 94.8



#88

1,4-Dichlorobenzene

Concen: 47.705 ug/l

RT: 12.036 min Scan# 1796

Delta R.T. -0.000 min

Lab File: VX046758.D

Acq: 19 Jun 2025 09:43




Tgt Ion:146 Resp: 164023


Ion Ratio Lower Upper

146 100

111 39.5 20.3 61.0

148 62.4 32.1 96.5





#89

n-Butylbenzene

Concen: 50.550 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. -0.000 min

Lab File: VX046758.D

Acq: 19 Jun 2025 09:43

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion: 91 Resp: 310154

Ion Ratio Lower Upper

91 100

92 54.1 27.1 81.3

134 25.9 12.9 38.6



#90

Hexachloroethane

Concen: 48.837 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX046758.D




Acq: 19 Jun 2025 09:43


Tgt Ion: 117 Resp: 55264

Ion Ratio Lower Upper

117 100

201 71.8 34.4 103.3









#91
1,2-Dichlorobenzene
Concen: 49.496 ug/l
RT: 12.329 min Scan# 1844
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:146 Resp: 154832
Ion Ratio Lower Upper
146 100
111 42.3 21.7 65.1
148 63.1 32.0 96.2



#92
1,2-Dibromo-3-Chloropropane
Concen: 40.348 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Tgt Ion: 75 Resp: 15868
Ion Ratio Lower Upper
75 100
155 84.4 41.9 125.7
157 107.9 51.4 154.3





#93

1,2,4-Trichlorobenzene

Concen: 48.410 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. -0.000 min

Lab File: VX046758.D

Acq: 19 Jun 2025 09:43

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion:180 Resp: 107660

Ion Ratio Lower Upper

180 100

182 95.9 46.7 140.1

145 32.7 15.8 47.4



#94

Hexachlorobutadiene

Concen: 49.207 ug/l

RT: 13.719 min Scan# 2072

Delta R.T. -0.000 min

Lab File: VX046758.D

Acq: 19 Jun 2025 09:43




Tgt Ion:225 Resp: 46034


Ion Ratio Lower Upper

225 100

223 64.3 32.5 97.5

227 63.2 32.2 96.6









#95
Naphthalene
Concen: 45.216 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion:128 Resp: 286491
Ion Ratio Lower Upper
128 100
127 12.8 10.1 15.1
129 10.7 8.7 13.1



#96
1,2,3-Trichlorobenzene
Concen: 48.453 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. -0.000 min
Lab File: VX046758.D
Acq: 19 Jun 2025 09:43

Tgt Ion:180 Resp: 103088
Ion Ratio Lower Upper
180 100
182 94.7 47.8 143.4
145 34.7 17.3 51.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
 Data File : VX046762.D
 Acq On : 19 Jun 2025 11:24
 Operator : JC/MD
 Sample : VX0619WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0619WBS01

Quant Time: Jun 20 05:16:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	137936	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	227140	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	207463	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	104652	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	94946	49.765	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.520%
35) Dibromofluoromethane	5.397	113	78196	52.032	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.060%
50) Toluene-d8	8.647	98	282602	52.340	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.680%
62) 4-Bromofluorobenzene	11.079	95	108869	54.092	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	108.180%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.185	85	28069	19.060	ug/l	97
3) Chloromethane	1.307	50	29736	18.707	ug/l	99
4) Vinyl Chloride	1.386	62	32033	18.898	ug/l	93
5) Bromomethane	1.630	94	20191	21.172	ug/l	93
6) Chloroethane	1.709	64	20468	19.923	ug/l	98
7) Trichlorofluoromethane	1.910	101	47476	18.446	ug/l	99
8) Diethyl Ether	2.148	74	16252	18.400	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.349	101	29254	18.528	ug/l	99
10) Methyl Iodide	2.471	142	29614	18.018	ug/l	99
11) Tert butyl alcohol	2.959	59	10810	63.126	ug/l	97
12) 1,1-Dichloroethene	2.337	96	28292	18.570	ug/l	97
13) Acrolein	2.251	56	18186	72.910	ug/l	98
14) Allyl chloride	2.684	41	49324	17.868	ug/l	99
15) Acrylonitrile	3.075	53	62463	81.119	ug/l	98
16) Acetone	2.386	43	47143	74.096	ug/l	96
17) Carbon Disulfide	2.532	76	78405	17.041	ug/l	98
18) Methyl Acetate	2.715	43	25862	16.004	ug/l	98
19) Methyl tert-butyl Ether	3.123	73	80635	16.997	ug/l	98
20) Methylene Chloride	2.806	84	32048	18.113	ug/l	98
21) trans-1,2-Dichloroethene	3.111	96	29757	18.253	ug/l	99
22) Diisopropyl ether	3.769	45	100392	18.921	ug/l	94
23) Vinyl Acetate	3.733	43	365712	86.676	ug/l	100
24) 1,1-Dichloroethane	3.623	63	55663	18.477	ug/l	99
25) 2-Butanone	4.568	43	69661	77.416	ug/l	100
26) 2,2-Dichloropropane	4.489	77	39442	17.951	ug/l	99
27) cis-1,2-Dichloroethene	4.507	96	35547	18.560	ug/l	98
28) Bromochloromethane	4.910	49	28430	20.816	ug/l	99
29) Tetrahydrofuran	5.013	42	44591	76.721	ug/l	99
30) Chloroform	5.105	83	57187	18.982	ug/l	93
31) Cyclohexane	5.483	56	53811	18.836	ug/l	96
32) 1,1,1-Trichloroethane	5.391	97	46675	17.665	ug/l	98
36) 1,1-Dichloropropene	5.702	75	40684	18.754	ug/l	99
37) Ethyl Acetate	4.727	43	31829	15.895	ug/l	98
38) Carbon Tetrachloride	5.690	117	42451	18.031	ug/l	99
39) Methylcyclohexane	7.385	83	52931	18.541	ug/l	99
40) Benzene	6.050	78	121111	18.864	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
 Data File : VX046762.D
 Acq On : 19 Jun 2025 11:24
 Operator : JC/MD
 Sample : VX0619WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0619WBS01

Quant Time: Jun 20 05:16:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.928	41	17227	16.137	ug/l	99
42) 1,2-Dichloroethane	6.092	62	41780	18.518	ug/l	100
43) Isopropyl Acetate	6.348	43	51032	16.286	ug/l	99
44) Trichloroethene	7.135	130	30093	18.392	ug/l	99
45) 1,2-Dichloropropane	7.433	63	29323	18.429	ug/l	95
46) Dibromomethane	7.592	93	20791	18.181	ug/l	99
47) Bromodichloromethane	7.824	83	44034	18.870	ug/l	99
48) Methyl methacrylate	7.696	41	25618	16.301	ug/l	98
49) 1,4-Dioxane	7.659	88	6374	299.149	ug/l	94
51) 4-Methyl-2-Pentanone	8.573	43	152568	81.636	ug/l	98
52) Toluene	8.720	92	75814	19.048	ug/l	98
53) t-1,3-Dichloropropene	8.982	75	38942	17.985	ug/l	99
54) cis-1,3-Dichloropropene	8.366	75	44663	18.171	ug/l	98
55) 1,1,2-Trichloroethane	9.153	97	27505	18.544	ug/l	99
56) Ethyl methacrylate	9.116	69	37095	16.942	ug/l	98
57) 1,3-Dichloropropane	9.311	76	47213	18.484	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.244	63	101869	89.565	ug/l	99
59) 2-Hexanone	9.427	43	101986	78.935	ug/l	100
60) Dibromochloromethane	9.518	129	32154	18.395	ug/l	100
61) 1,2-Dibromoethane	9.610	107	27221	18.042	ug/l	100
64) Tetrachloroethene	9.275	164	25376	17.655	ug/l	95
65) Chlorobenzene	10.079	112	82392	17.991	ug/l	98
66) 1,1,1,2-Tetrachloroethane	10.159	131	27463	17.843	ug/l	98
67) Ethyl Benzene	10.195	91	144047	18.002	ug/l	97
68) m/p-Xylenes	10.299	106	111044	37.244	ug/l	100
69) o-Xylene	10.640	106	51894	18.572	ug/l	98
70) Styrene	10.652	104	89289	18.260	ug/l	99
71) Bromoform	10.799	173	19576	16.701	ug/l #	98
73) Isopropylbenzene	10.957	105	139510	18.100	ug/l	99
74) N-amyl acetate	10.841	43	46412	15.893	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	37312	17.337	ug/l	100
76) 1,2,3-Trichloropropane	11.238	75	36095	18.832	ug/l	98
77) Bromobenzene	11.195	156	33896	18.275	ug/l	99
78) n-propylbenzene	11.299	91	168589	18.415	ug/l	99
79) 2-Chlorotoluene	11.360	91	99687	18.238	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	115409	18.221	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	10102	14.897	ug/l	96
82) 4-Chlorotoluene	11.451	91	116713	18.296	ug/l	99
83) tert-Butylbenzene	11.713	119	118027	18.027	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	116817	18.357	ug/l	99
85) sec-Butylbenzene	11.890	105	153043	18.496	ug/l	100
86) p-Isopropyltoluene	12.006	119	127841	18.312	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	64330	17.788	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	64856	17.460	ug/l	99
89) n-Butylbenzene	12.329	91	118756	17.915	ug/l	100
90) Hexachloroethane	12.536	117	20652	16.893	ug/l	98
91) 1,2-Dichlorobenzene	12.335	146	61156	18.096	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	6425	15.122	ug/l	97
93) 1,2,4-Trichlorobenzene	13.585	180	41109	17.110	ug/l	98
94) Hexachlorobutadiene	13.719	225	17833	17.644	ug/l	96
95) Naphthalene	13.774	128	110507	16.144	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	40546	17.639	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
Data File : VX046762.D
Acq On : 19 Jun 2025 11:24
Operator : JC/MD
Sample : VX0619WBS01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01


Quant Time: Jun 20 05:16:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 03:09:16 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jun 20 05:16:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 03:09:16 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.562 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01


Tgt Ion:168 Resp: 137936
Ion Ratio Lower Upper
168 100
99 58.9 48.5 72.7



#2
Dichlorodifluoromethane
Concen: 19.060 ug/l
RT: 1.185 min Scan# 16
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Tgt Ion: 85 Resp: 28069
Ion Ratio Lower Upper
85 100
87 31.5 16.6 49.8









#3
Chloromethane
Concen: 18.707 ug/l
RT: 1.307 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01


Tgt Ion: 50 Resp: 29736
Ion Ratio Lower Upper
50 100
52 31.9 24.9 37.3



#4
Vinyl Chloride
Concen: 18.898 ug/l
RT: 1.386 min Scan# 49
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Tgt Ion: 62 Resp: 32033
Ion Ratio Lower Upper
62 100
64 33.8 23.9 35.9









#5
Bromomethane
Concen: 21.172 ug/l
RT: 1.630 min Scan# 89
Delta R.T. 0.006 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01


Tgt Ion: 94 Resp: 20191
Ion Ratio Lower Upper
94 100
96 97.2 72.3 108.5



#6
Chloroethane
Concen: 19.923 ug/l
RT: 1.709 min Scan# 102
Delta R.T. 0.006 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Tgt Ion: 64 Resp: 20468
Ion Ratio Lower Upper
64 100
66 33.0 27.1 40.7









#7
 Trichlorofluoromethane
 Concen: 18.446 ug/l
 RT: 1.910 min Scan# 11
 Delta R.T. 0.006 min
 Lab File: VX046762.D
 Acq: 19 Jun 2025 11:24




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0619WBS01


Tgt Ion:101 Resp: 47476
 Ion Ratio Lower Upper
 101 100
 103 65.5 51.8 77.8



#8
 Diethyl Ether
 Concen: 18.400 ug/l
 RT: 2.148 min Scan# 174
 Delta R.T. -0.000 min
 Lab File: VX046762.D
 Acq: 19 Jun 2025 11:24

Tgt Ion: 74 Resp: 16252
 Ion Ratio Lower Upper
 74 100
 45 104.0 52.3 157.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 18.528 ug/l
RT: 2.349 min Scan# 206
Delta R.T. 0.006 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01


Tgt Ion:101 Resp: 29254
Ion Ratio Lower Upper
101 100
85 43.7 35.4 53.0
151 76.9 61.2 91.8



#10
Methyl Iodide
Concen: 18.018 ug/l
RT: 2.471 min Scan# 227
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Tgt Ion:142 Resp: 29614
Ion Ratio Lower Upper
142 100
127 45.8 37.0 55.6
141 13.4 11.5 17.3









#11
Tert butyl alcohol
Concen: 63.126 ug/l
RT: 2.959 min Scan# 307
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01


Tgt Ion: 59 Resp: 10810
Ion Ratio Lower Upper
59 100
57 9.3 8.3 12.5



#12
1,1-Dichloroethene
Concen: 18.570 ug/l
RT: 2.337 min Scan# 205
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Tgt Ion: 96 Resp: 28292
Ion Ratio Lower Upper
96 100
61 163.0 126.6 189.8
98 63.7 50.7 76.1









#13
Acrolein
Concen: 72.910 ug/l
RT: 2.251 min Scan# 191
Delta R.T. 0.006 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01


Tgt Ion: 56 Resp: 18186
Ion Ratio Lower Upper
56 100
55 70.9 55.2 82.8



#14
Allyl chloride
Concen: 17.868 ug/l
RT: 2.684 min Scan# 262
Delta R.T. 0.006 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Tgt Ion: 41 Resp: 49324
Ion Ratio Lower Upper
41 100
39 67.0 53.6 80.4
76 32.3 26.8 40.2









#15
Acrylonitrile
Concen: 81.119 ug/l
RT: 3.075 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01


Tgt Ion: 53 Resp: 62463
Ion Ratio Lower Upper
53 100
52 83.7 65.1 97.7
51 36.4 28.9 43.3



#16
Acetone
Concen: 74.096 ug/l
RT: 2.386 min Scan# 213
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Tgt Ion: 43 Resp: 47143
Ion Ratio Lower Upper
43 100
58 33.0 24.5 36.7









#17
Carbon Disulfide
Concen: 17.041 ug/l
RT: 2.532 min Scan# 21
Delta R.T. 0.006 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01


Tgt Ion: 76 Resp: 78405
Ion Ratio Lower Upper
76 100
78 10.2 7.5 11.3



#18
Methyl Acetate
Concen: 16.004 ug/l
RT: 2.715 min Scan# 267
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Tgt Ion: 43 Resp: 25862
Ion Ratio Lower Upper
43 100
74 22.2 18.4 27.6









#19
Methyl tert-butyl Ether
Concen: 16.997 ug/l
RT: 3.123 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01


Tgt Ion: 73 Resp: 80635
Ion Ratio Lower Upper
73 100
57 21.9 18.2 27.2



#20
Methylene Chloride
Concen: 18.113 ug/l
RT: 2.806 min Scan# 282
Delta R.T. 0.006 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Tgt Ion: 84 Resp: 32048
Ion Ratio Lower Upper
84 100
49 123.7 101.3 151.9
51 38.1 31.0 46.6
86 63.0 51.1 76.7









#21
trans-1,2-Dichloroethene
Concen: 18.253 ug/l
RT: 3.111 min Scan# 311
Delta R.T. 0.006 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01


Tgt Ion: 96 Resp: 29757
Ion Ratio Lower Upper
96 100
61 139.5 112.5 168.7
98 65.5 51.9 77.9



#22
Diisopropyl ether
Concen: 18.921 ug/l
RT: 3.769 min Scan# 440
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Tgt Ion: 45 Resp: 100392
Ion Ratio Lower Upper
45 100
43 121.9 91.9 137.9
87 25.0 22.0 33.0
59 11.9 9.1 13.7









#23
 Vinyl Acetate
 Concen: 86.676 ug/l
 RT: 3.733 min Scan# 41
 Delta R.T. -0.000 min
 Lab File: VX046762.D
 Acq: 19 Jun 2025 11:24




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0619WBS01


Tgt Ion: 43 Resp: 365712
 Ion Ratio Lower Upper
 43 100
 86 10.1 8.2 12.2



#24
 1,1-Dichloroethane
 Concen: 18.477 ug/l
 RT: 3.623 min Scan# 416
 Delta R.T. -0.000 min
 Lab File: VX046762.D
 Acq: 19 Jun 2025 11:24

Tgt Ion: 63 Resp: 55663
 Ion Ratio Lower Upper
 63 100
 98 8.5 4.0 11.9
 100 4.8 2.2 6.6





#25

2-Butanone

Concen: 77.416 ug/l

RT: 4.568 min Scan# 5

Delta R.T. 0.006 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :





VX0619WBS01

Tgt Ion: 43 Resp: 69661

Ion Ratio Lower Upper

43 100

72 25.2 20.2 30.4



#26

2,2-Dichloropropane

Concen: 17.951 ug/l

RT: 4.489 min Scan# 558

Delta R.T. -0.000 min

Lab File: VX046762.D




Acq: 19 Jun 2025 11:24


Tgt Ion: 77 Resp: 39442

Ion Ratio Lower Upper

77 100

97 22.8 11.8 35.3





#27

cis-1,2-Dichloroethene

Concen: 18.560 ug/l

RT: 4.507 min Scan# 560

Delta R.T. 0.006 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBS01





Tgt Ion: 96 Resp: 35547

Ion Ratio Lower Upper

96 100

61 142.4 0.0 291.0

98 64.5 0.0 126.8



#28

Bromochloromethane

Concen: 20.816 ug/l

RT: 4.910 min Scan# 627

Delta R.T. -0.000 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24




Tgt Ion: 49 Resp: 28430


Ion Ratio Lower Upper

49 100

129 2.1 0.0 3.8

130 75.5 61.4 92.0





#29

Tetrahydrofuran

Concen: 76.721 ug/l

RT: 5.013 min Scan# 644

Delta R.T. -0.000 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBS01





Tgt Ion: 42 Resp: 44591

Ion Ratio Lower Upper

42 100

72 43.4 34.9 52.3

71 41.1 32.2 48.2



#30

Chloroform

Concen: 18.982 ug/l

RT: 5.105 min Scan# 659

Delta R.T. -0.000 min

Lab File: VX046762.D




Acq: 19 Jun 2025 11:24


Tgt Ion: 83 Resp: 57187

Ion Ratio Lower Upper

83 100

85 61.3 53.7 80.5









#31
Cyclohexane
Concen: 18.836 ug/l
RT: 5.483 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01


Tgt Ion:	56	Resp:	53811
Ion	Ratio	Lower	Upper
56	100		
69	29.2	24.8	37.2
84	81.2	68.2	102.2



#32
1,1,1-Trichloroethane
Concen: 17.665 ug/l
RT: 5.391 min Scan# 706
Delta R.T. -0.006 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Tgt Ion:	97	Resp:	46675
Ion	Ratio	Lower	Upper
97	100		
99	64.5	51.3	76.9
61	47.4	36.4	54.6









#33
 1,2-Dichloroethane-d4
 Concen: 49.765 ug/l
 RT: 5.964 min Scan# 800
 Delta R.T. -0.000 min
 Lab File: VX046762.D
 Acq: 19 Jun 2025 11:24




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0619WBS01


Tgt Ion: 65 Resp: 94946
 Ion Ratio Lower Upper
 65 100
 67 51.8 0.0 105.4



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.769 min Scan# 932
 Delta R.T. -0.000 min
 Lab File: VX046762.D
 Acq: 19 Jun 2025 11:24

Tgt Ion: 114 Resp: 227140
 Ion Ratio Lower Upper
 114 100
 63 20.5 0.0 44.2
 88 15.8 0.0 33.2





#35

Dibromofluoromethane

Concen: 52.032 ug/l

RT: 5.397 min Scan# 707

Delta R.T. -0.000 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBS01





Tgt Ion:113 Resp: 78196

Ion Ratio Lower Upper

113 100

111 102.8 82.0 123.0

192 18.8 15.3 22.9



#36

1,1-Dichloropropene

Concen: 18.754 ug/l

RT: 5.702 min Scan# 757

Delta R.T. -0.000 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24




Tgt Ion: 75 Resp: 40684


Ion Ratio Lower Upper

75 100

110 34.0 17.2 51.5

77 31.1 24.2 36.4





#37

Ethyl Acetate

Concen: 15.895 ug/l

RT: 4.727 min Scan# 597

Delta R.T. 0.006 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBS01





Tgt Ion: 43 Resp: 31829

Ion Ratio Lower Upper

43 100

61 12.9 9.7 14.5

70 9.5 8.1 12.1



#38

Carbon Tetrachloride

Concen: 18.031 ug/l

RT: 5.690 min Scan# 755

Delta R.T. -0.000 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24




Tgt Ion: 117 Resp: 42451


Ion Ratio Lower Upper

117 100

119 94.7 75.2 112.8

121 28.0 24.1 36.1





#39

Methylcyclohexane

Concen: 18.541 ug/l

RT: 7.385 min Scan# 1033

Delta R.T. -0.000 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBS01





Tgt Ion: 83 Resp: 52931

Ion Ratio Lower Upper

83 100

55 78.7 63.0 94.4

98 49.6 38.8 58.2



#40

Benzene

Concen: 18.864 ug/l

RT: 6.050 min Scan# 814

Delta R.T. 0.006 min

Lab File: VX046762.D




Acq: 19 Jun 2025 11:24


Tgt Ion: 78 Resp: 121111

Ion Ratio Lower Upper

78 100

77 23.0 18.8 28.2









#41
Methacrylonitrile
Concen: 16.137 ug/l
RT: 4.928 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01


Tgt Ion:	41	Resp:	17227
Ion	Ratio	Lower	Upper
41	100		
39	58.4	45.9	68.9
67	77.6	61.2	91.8
52	32.7	25.5	38.3



#42
1,2-Dichloroethane
Concen: 18.518 ug/l
RT: 6.092 min Scan# 821
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Tgt Ion:	62	Resp:	41780
Ion	Ratio	Lower	Upper
62	100		
98	9.0	0.0	17.8





#43

Isopropyl Acetate

Concen: 16.286 ug/l

RT: 6.348 min Scan# 862

Delta R.T. 0.006 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBS01





Tgt Ion: 43 Resp: 51032

Ion Ratio Lower Upper

43 100

61 18.7 15.3 22.9

87 12.0 10.0 15.0



#44

Trichloroethene

Concen: 18.392 ug/l

RT: 7.135 min Scan# 992

Delta R.T. 0.006 min

Lab File: VX046762.D




Acq: 19 Jun 2025 11:24


Tgt Ion:130 Resp: 30093

Ion Ratio Lower Upper

130 100

95 102.2 0.0 202.6





#45

1,2-Dichloropropane

Concen: 18.429 ug/l

RT: 7.433 min Scan# 1041

Delta R.T. -0.000 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :





VX0619WBS01

Tgt Ion: 63 Resp: 29323

Ion Ratio Lower Upper

63 100

65 31.0 22.9 34.3



#46

Dibromomethane

Concen: 18.181 ug/l

RT: 7.592 min Scan# 1067

Delta R.T. 0.006 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24




Tgt Ion: 93 Resp: 20791


Ion Ratio Lower Upper

93 100

95 83.0 66.2 99.4

174 97.3 77.0 115.6









#47
Bromodichloromethane
Concen: 18.870 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01


Tgt Ion:	83	Resp:	44034
Ion	Ratio	Lower	Upper
83	100		
85	63.8	51.4	77.2
127	9.3	6.9	10.3



#48
Methyl methacrylate
Concen: 16.301 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Tgt Ion:	41	Resp:	25618
Ion	Ratio	Lower	Upper
41	100		
69	82.0	63.4	95.2
39	57.0	45.4	68.0









#49
1,4-Dioxane
Concen: 299.149 ug/l
RT: 7.659 min Scan# 1078
Delta R.T. 0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01


Tgt Ion:	88	Resp:	6374
Ion	Ratio	Lower	Upper
88	100		
43	38.3	31.2	46.8
58	75.1	54.7	82.1



#50
Toluene-d8
Concen: 52.340 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Tgt Ion:	98	Resp:	282602
Ion	Ratio	Lower	Upper
98	100		
100	66.6	53.0	79.4









#51
4-Methyl-2-Pentanone
Concen: 81.636 ug/l
RT: 8.573 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01


Tgt Ion: 43 Resp: 152568
Ion Ratio Lower Upper
43 100
58 37.5 30.8 46.2



#52
Toluene
Concen: 19.048 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Tgt Ion: 92 Resp: 75814
Ion Ratio Lower Upper
92 100
91 170.6 134.8 202.2





#53

t-1,3-Dichloropropene

Concen: 17.985 ug/l

RT: 8.982 min Scan# 1129

Delta R.T. 0.006 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :





VX0619WBS01

Tgt Ion: 75 Resp: 38942

Ion Ratio Lower Upper

75 100

77 31.6 25.6 38.4



#54

cis-1,3-Dichloropropene

Concen: 18.171 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24




Tgt Ion: 75 Resp: 44663


Ion Ratio Lower Upper

75 100

77 28.9 25.3 37.9

39 48.7 39.2 58.8









#55
1,1,2-Trichloroethane
Concen: 18.544 ug/l
RT: 9.153 min Scan# 1323
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01


Tgt Ion:	97	Resp:	27505
Ion	Ratio	Lower	Upper
97	100		
83	81.4	67.2	100.8
85	54.6	44.1	66.1
99	63.1	50.3	75.5



#56
Ethyl methacrylate
Concen: 16.942 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Tgt Ion:	69	Resp:	37095
Ion	Ratio	Lower	Upper
69	100		
41	70.6	55.4	83.2
39	36.2	29.7	44.5





#57

1,3-Dichloropropane

Concen: 18.484 ug/l

RT: 9.311 min Scan# 11

Delta R.T. 0.006 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :





VX0619WBS01

Tgt Ion: 76 Resp: 47213

Ion Ratio Lower Upper

76 100

78 32.0 25.7 38.5



#58

2-Chloroethyl Vinyl ether

Concen: 89.565 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. -0.000 min

Lab File: VX046762.D




Acq: 19 Jun 2025 11:24


Tgt Ion: 63 Resp: 101869

Ion Ratio Lower Upper

63 100

106 27.8 22.6 34.0









#59
2-Hexanone
Concen: 78.935 ug/l
RT: 9.427 min Scan# 1368
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01


Tgt Ion: 43 Resp: 101986
Ion Ratio Lower Upper
43 100
58 53.7 26.8 80.4



#60
Dibromochloromethane
Concen: 18.395 ug/l
RT: 9.518 min Scan# 1383
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Tgt Ion: 129 Resp: 32154
Ion Ratio Lower Upper
129 100
127 76.9 38.4 115.2









#61
1,2-Dibromoethane
Concen: 18.042 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01


Tgt Ion:107 Resp: 27221
Ion Ratio Lower Upper
107 100
109 94.6 75.8 113.6



#62
4-Bromofluorobenzene
Concen: 54.092 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Tgt Ion: 95 Resp: 108869
Ion Ratio Lower Upper
95 100
174 75.4 0.0 150.4
176 71.8 0.0 145.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01


Tgt Ion:117 Resp: 207463
Ion Ratio Lower Upper
117 100
82 55.6 44.6 66.8
119 31.9 25.8 38.8



#64
Tetrachloroethene
Concen: 17.655 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Tgt Ion:164 Resp: 25376
Ion Ratio Lower Upper
164 100
166 121.4 104.6 157.0
129 99.6 76.3 114.5
131 93.8 74.1 111.1









#65
Chlorobenzene
Concen: 17.991 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01


Tgt Ion:112 Resp: 82392
Ion Ratio Lower Upper
112 100
114 31.0 25.5 38.3



#66
1,1,1,2-Tetrachloroethane
Concen: 17.843 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Tgt Ion:131 Resp: 27463
Ion Ratio Lower Upper
131 100
133 96.2 47.3 141.8
119 66.0 32.1 96.3









#67
Ethyl Benzene
Concen: 18.002 ug/l
RT: 10.195 min Scan# 1493
Delta R.T. 0.006 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01


Tgt Ion: 91 Resp: 144047
Ion Ratio Lower Upper
91 100
106 31.7 24.2 36.2



#68
m/p-Xylenes
Concen: 37.244 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Tgt Ion:106 Resp: 111044
Ion Ratio Lower Upper
106 100
91 205.3 164.8 247.2









#69
o-Xylene
Concen: 18.572 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01


Tgt Ion:106 Resp: 51894
Ion Ratio Lower Upper
106 100
91 219.2 111.3 333.9



#70
Styrene
Concen: 18.260 ug/l
RT: 10.652 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Tgt Ion:104 Resp: 89289
Ion Ratio Lower Upper
104 100
78 52.0 41.4 62.2
103 54.7 43.1 64.7









#71
Bromoform
Concen: 16.701 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01


Tgt Ion:173 Resp: 19576
Ion Ratio Lower Upper
173 100
175 50.0 24.1 72.4
254 0.0 0.1 0.1#



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Tgt Ion:152 Resp: 104652
Ion Ratio Lower Upper
152 100
115 70.8 43.2 129.6
150 164.1 0.0 346.8









#73
Isopropylbenzene
Concen: 18.100 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01


Tgt Ion:105 Resp: 139510
Ion Ratio Lower Upper
105 100
120 26.3 13.0 39.0



#74
N-ethyl acetate
Concen: 15.893 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Tgt Ion: 43 Resp: 46412
Ion Ratio Lower Upper
43 100
70 41.3 33.4 50.0
55 24.2 19.1 28.7
61 23.0 19.0 28.4









#75
 1,1,2,2-Tetrachloroethane
 Concen: 17.337 ug/l
 RT: 11.207 min Scan# 1660
 Delta R.T. -0.000 min
 Lab File: VX046762.D
 Acq: 19 Jun 2025 11:24




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0619WBS01


Tgt Ion: 83 Resp: 37312
 Ion Ratio Lower Upper
 83 100
 131 9.7 4.9 14.6
 85 63.7 32.0 96.0



#76
 1,2,3-Trichloropropane
 Concen: 18.832 ug/l
 RT: 11.238 min Scan# 1665
 Delta R.T. -0.000 min
 Lab File: VX046762.D
 Acq: 19 Jun 2025 11:24

Tgt Ion: 75 Resp: 36095
 Ion Ratio Lower Upper
 75 100
 77 35.6 18.5 55.5









#77
Bromobenzene
Concen: 18.275 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01


Tgt Ion:156 Resp: 33896
Ion Ratio Lower Upper
156 100
77 164.7 82.8 248.6
158 97.8 47.6 142.9



#78
n-propylbenzene
Concen: 18.415 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Tgt Ion: 91 Resp: 168589
Ion Ratio Lower Upper
91 100
120 22.3 11.4 34.1









#79
2-Chlorotoluene
Concen: 18.238 ug/l
RT: 11.360 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01


Tgt Ion: 91 Resp: 99687
Ion Ratio Lower Upper
91 100
126 32.5 16.6 49.7



#80
1,3,5-Trimethylbenzene
Concen: 18.221 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Tgt Ion: 105 Resp: 115409
Ion Ratio Lower Upper
105 100
120 48.8 24.1 72.2









#81
trans-1,4-Dichloro-2-butene
Concen: 14.897 ug/l
RT: 11.018 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01


Tgt Ion: 75 Resp: 10102
Ion Ratio Lower Upper
75 100
53 112.7 86.8 130.2
89 43.9 36.9 55.3



#82
4-Chlorotoluene
Concen: 18.296 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Tgt Ion: 91 Resp: 116713
Ion Ratio Lower Upper
91 100
126 29.3 14.4 43.2





#83

tert-Butylbenzene

Concen: 18.027 ug/l

RT: 11.713 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBS01





Tgt Ion:119 Resp: 118027

Ion Ratio Lower Upper

119 100

91 58.4 29.5 88.6

134 23.0 11.7 35.0



#84

1,2,4-Trimethylbenzene

Concen: 18.357 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. -0.000 min

Lab File: VX046762.D




Acq: 19 Jun 2025 11:24


Tgt Ion:105 Resp: 116817

Ion Ratio Lower Upper

105 100

120 44.6 22.1 66.3









#85
 sec-Butylbenzene
 Concen: 18.496 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. -0.000 min
 Lab File: VX046762.D
 Acq: 19 Jun 2025 11:24




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0619WBS01


Tgt Ion:105 Resp: 153043
 Ion Ratio Lower Upper
 105 100
 134 19.7 9.8 29.4



#86
 p-Isopropyltoluene
 Concen: 18.312 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. -0.000 min
 Lab File: VX046762.D
 Acq: 19 Jun 2025 11:24

Tgt Ion:119 Resp: 127841
 Ion Ratio Lower Upper
 119 100
 134 25.1 12.9 38.7
 91 25.2 12.3 36.9





#87

1,3-Dichlorobenzene

Concen: 17.788 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBS01





Tgt Ion:146 Resp: 64330

Ion Ratio Lower Upper

146 100

111 40.8 20.6 61.8

148 62.5 31.6 94.8



#88

1,4-Dichlorobenzene

Concen: 17.460 ug/l

RT: 12.036 min Scan# 1796

Delta R.T. -0.000 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24




Tgt Ion:146 Resp: 64856


Ion Ratio Lower Upper

146 100

111 42.3 20.3 61.0

148 64.0 32.1 96.5





#89

n-Butylbenzene

Concen: 17.915 ug/l

RT: 12.329 min Scan# 118756

Delta R.T. -0.000 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBS01





Tgt Ion: 91 Resp: 118756

Ion Ratio Lower Upper

91 100

92 54.0 27.1 81.3

134 25.7 12.9 38.6



#90

Hexachloroethane

Concen: 16.893 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX046762.D




Acq: 19 Jun 2025 11:24


Tgt Ion: 117 Resp: 20652

Ion Ratio Lower Upper

117 100

201 70.7 34.4 103.3





#91

1,2-Dichlorobenzene

Concen: 18.096 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.006 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBS01





Tgt Ion:146 Resp: 61156

Ion Ratio Lower Upper

146 100

111 42.8 21.7 65.1

148 64.4 32.0 96.2



#92

1,2-Dibromo-3-Chloropropane

Concen: 15.122 ug/l

RT: 12.939 min Scan# 1944

Delta R.T. -0.000 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24




Tgt Ion: 75 Resp: 6425


Ion Ratio Lower Upper

75 100

155 80.9 41.9 125.7

157 105.8 51.4 154.3





#93

1,2,4-Trichlorobenzene

Concen: 17.110 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. -0.000 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBS01





Tgt Ion:180 Resp: 41109

Ion Ratio Lower Upper

180 100

182 95.7 46.7 140.1

145 31.1 15.8 47.4



#94

Hexachlorobutadiene

Concen: 17.644 ug/l

RT: 13.719 min Scan# 2072

Delta R.T. -0.000 min

Lab File: VX046762.D

Acq: 19 Jun 2025 11:24




Tgt Ion:225 Resp: 17833


Ion Ratio Lower Upper

225 100

223 62.5 32.5 97.5

227 61.1 32.2 96.6









#95
Naphthalene
Concen: 16.144 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBS01

Tgt Ion:128 Resp: 110507
Ion Ratio Lower Upper
128 100
127 12.6 10.1 15.1
129 11.0 8.7 13.1



#96
1,2,3-Trichlorobenzene
Concen: 17.639 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. -0.000 min
Lab File: VX046762.D
Acq: 19 Jun 2025 11:24

Tgt Ion:180 Resp: 40546
Ion Ratio Lower Upper
180 100
182 94.7 47.8 143.4
145 34.2 17.3 51.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
 Data File : VX046763.D
 Acq On : 19 Jun 2025 11:51
 Operator : JC/MD
 Sample : VX0619WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0619WBSD01

Quant Time: Jun 20 05:17:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	120371	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	199496	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	183807	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	94851	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	85794	51.530	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 103.060%			
35) Dibromofluoromethane	5.391	113	69686	52.795	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 105.600%			
50) Toluene-d8	8.647	98	249730	52.661	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 105.320%			
62) 4-Bromofluorobenzene	11.079	95	97074	54.915	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 109.820%			
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.185	85	25471	19.820	ug/l	99
3) Chloromethane	1.307	50	27061	19.509	ug/l	96
4) Vinyl Chloride	1.386	62	29098	19.672	ug/l	95
5) Bromomethane	1.630	94	18436	22.153	ug/l	93
6) Chloroethane	1.703	64	18444	20.573	ug/l	95
7) Trichlorofluoromethane	1.904	101	43435	19.339	ug/l	97
8) Diethyl Ether	2.148	74	15726	20.402	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.349	101	27115	19.679	ug/l	98
10) Methyl Iodide	2.465	142	30018	20.229	ug/l	98
11) Tert butyl alcohol	2.959	59	10932	73.154	ug/l	96
12) 1,1-Dichloroethene	2.337	96	25867	19.455	ug/l	98
13) Acrolein	2.245	56	18810	86.416	ug/l	97
14) Allyl chloride	2.678	41	45893	19.051	ug/l	100
15) Acrylonitrile	3.068	53	61358	91.312	ug/l	99
16) Acetone	2.386	43	45269	81.534	ug/l	95
17) Carbon Disulfide	2.526	76	72150	17.970	ug/l	99
18) Methyl Acetate	2.709	43	25035	17.753	ug/l	100
19) Methyl tert-butyl Ether	3.117	73	77256	18.661	ug/l	98
20) Methylene Chloride	2.800	84	29836	19.323	ug/l	99
21) trans-1,2-Dichloroethene	3.105	96	27280	19.176	ug/l	97
22) Diisopropyl ether	3.764	45	94092	20.321	ug/l	87
23) Vinyl Acetate	3.733	43	359723	97.697	ug/l	100
24) 1,1-Dichloroethane	3.623	63	50927	19.372	ug/l	100
25) 2-Butanone	4.562	43	68340	87.030	ug/l	97
26) 2,2-Dichloropropane	4.489	77	36206	18.883	ug/l	99
27) cis-1,2-Dichloroethene	4.501	96	32489	19.438	ug/l	100
28) Bromochloromethane	4.910	49	27522	23.092	ug/l	97
29) Tetrahydrofuran	5.013	42	43894	86.542	ug/l	99
30) Chloroform	5.099	83	53771	20.452	ug/l	97
31) Cyclohexane	5.477	56	49014	19.660	ug/l	98
32) 1,1,1-Trichloroethane	5.391	97	44445	19.276	ug/l	99
36) 1,1-Dichloropropene	5.702	75	35819	18.799	ug/l	97
37) Ethyl Acetate	4.721	43	31001	17.627	ug/l	98
38) Carbon Tetrachloride	5.684	117	39531	19.117	ug/l	99
39) Methylcyclohexane	7.385	83	47544	18.962	ug/l	99
40) Benzene	6.044	78	112866	20.015	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
 Data File : VX046763.D
 Acq On : 19 Jun 2025 11:51
 Operator : JC/MD
 Sample : VX0619WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0619WBSD01

Quant Time: Jun 20 05:17:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.928	41	17801	18.985	ug/l	96
42) 1,2-Dichloroethane	6.092	62	39829	20.100	ug/l	99
43) Isopropyl Acetate	6.342	43	50813	18.463	ug/l	99
44) Trichloroethene	7.129	130	27765	19.321	ug/l	89
45) 1,2-Dichloropropane	7.427	63	28336	20.276	ug/l	92
46) Dibromomethane	7.580	93	19978	19.891	ug/l	99
47) Bromodichloromethane	7.824	83	41221	20.112	ug/l	99
48) Methyl methacrylate	7.696	41	26092	18.904	ug/l	100
49) 1,4-Dioxane	7.659	88	6068	321.885	ug/l	97
51) 4-Methyl-2-Pentanone	8.574	43	150700	91.811	ug/l	98
52) Toluene	8.720	92	70998	20.310	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	37151	19.536	ug/l	98
54) cis-1,3-Dichloropropene	8.366	75	43044	19.939	ug/l	100
55) 1,1,2-Trichloroethane	9.153	97	26588	20.409	ug/l	98
56) Ethyl methacrylate	9.116	69	35838	18.636	ug/l	97
57) 1,3-Dichloropropane	9.305	76	45490	20.277	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.238	63	100239	100.344	ug/l	99
59) 2-Hexanone	9.427	43	100820	88.846	ug/l	100
60) Dibromochloromethane	9.519	129	30624	19.947	ug/l	99
61) 1,2-Dibromoethane	9.610	107	26614	20.084	ug/l	98
64) Tetrachloroethene	9.275	164	23913	18.778	ug/l	98
65) Chlorobenzene	10.079	112	78801	19.421	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.159	131	26522	19.449	ug/l	100
67) Ethyl Benzene	10.195	91	135702	19.141	ug/l	99
68) m/p-Xylenes	10.299	106	102519	38.810	ug/l	99
69) o-Xylene	10.640	106	49068	19.820	ug/l	99
70) Styrene	10.653	104	85333	19.696	ug/l	98
71) Bromoform	10.799	173	19127	18.418	ug/l #	98
73) Isopropylbenzene	10.957	105	129881	18.592	ug/l	100
74) N-amyl acetate	10.842	43	45562	17.214	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	35939	18.425	ug/l	100
76) 1,2,3-Trichloropropane	11.238	75	36174	20.823	ug/l	94
77) Bromobenzene	11.195	156	31899	18.976	ug/l	98
78) n-propylbenzene	11.299	91	156018	18.803	ug/l	100
79) 2-Chlorotoluene	11.360	91	92310	18.633	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	108639	18.925	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	9561	15.556	ug/l	95
82) 4-Chlorotoluene	11.451	91	108666	18.795	ug/l	99
83) tert-Butylbenzene	11.713	119	108810	18.337	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	108849	18.872	ug/l	100
85) sec-Butylbenzene	11.890	105	141281	18.839	ug/l	100
86) p-Isopropyltoluene	12.006	119	118528	18.732	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	60812	18.553	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	59834	17.772	ug/l	98
89) n-Butylbenzene	12.329	91	111302	18.526	ug/l	99
90) Hexachloroethane	12.536	117	19081	17.220	ug/l	96
91) 1,2-Dichlorobenzene	12.335	146	58188	18.997	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	6206	16.116	ug/l	97
93) 1,2,4-Trichlorobenzene	13.585	180	39871	18.309	ug/l	99
94) Hexachlorobutadiene	13.719	225	17144	18.715	ug/l	96
95) Naphthalene	13.774	128	107578	17.340	ug/l	99
96) 1,2,3-Trichlorobenzene	13.957	180	37893	18.189	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
Data File : VX046763.D
Acq On : 19 Jun 2025 11:51
Operator : JC/MD
Sample : VX0619WBSD01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01


Quant Time: Jun 20 05:17:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 03:09:16 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01

Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.562 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01


Tgt Ion:168 Resp: 120371
Ion Ratio Lower Upper
168 100
99 61.3 48.5 72.7



#2
Dichlorodifluoromethane
Concen: 19.820 ug/l
RT: 1.185 min Scan# 16
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Tgt Ion: 85 Resp: 25471
Ion Ratio Lower Upper
85 100
87 32.8 16.6 49.8









#3
Chloromethane
Concen: 19.509 ug/l
RT: 1.307 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01


Tgt Ion: 50 Resp: 27061
Ion Ratio Lower Upper
50 100
52 33.3 24.9 37.3



#4
Vinyl Chloride
Concen: 19.672 ug/l
RT: 1.386 min Scan# 49
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Tgt Ion: 62 Resp: 29098
Ion Ratio Lower Upper
62 100
64 32.5 23.9 35.9









#5
Bromomethane
Concen: 22.153 ug/l
RT: 1.630 min Scan# 89
Delta R.T. 0.006 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01


Tgt Ion: 94 Resp: 18436
Ion Ratio Lower Upper
94 100
96 96.8 72.3 108.5



#6
Chloroethane
Concen: 20.573 ug/l
RT: 1.703 min Scan# 101
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Tgt Ion: 64 Resp: 18444
Ion Ratio Lower Upper
64 100
66 30.9 27.1 40.7









#7
Trichlorofluoromethane
Concen: 19.339 ug/l
RT: 1.904 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01


Tgt Ion:101 Resp: 43435
Ion Ratio Lower Upper
101 100
103 62.7 51.8 77.8



#8
Diethyl Ether
Concen: 20.402 ug/l
RT: 2.148 min Scan# 174
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Tgt Ion: 74 Resp: 15726
Ion Ratio Lower Upper
74 100
45 106.1 52.3 157.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 19.679 ug/l
RT: 2.349 min Scan# 206
Delta R.T. 0.006 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01


Tgt Ion:101 Resp: 27115
Ion Ratio Lower Upper
101 100
85 42.4 35.4 53.0
151 75.5 61.2 91.8



#10
Methyl Iodide
Concen: 20.229 ug/l
RT: 2.465 min Scan# 226
Delta R.T. -0.006 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Tgt Ion:142 Resp: 30018
Ion Ratio Lower Upper
142 100
127 44.9 37.0 55.6
141 14.0 11.5 17.3









#11
Tert butyl alcohol
Concen: 73.154 ug/l
RT: 2.959 min Scan# 307
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01


Tgt Ion: 59 Resp: 10932
Ion Ratio Lower Upper
59 100
57 9.0 8.3 12.5



#12
1,1-Dichloroethene
Concen: 19.455 ug/l
RT: 2.337 min Scan# 205
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Tgt Ion: 96 Resp: 25867
Ion Ratio Lower Upper
96 100
61 162.1 126.6 189.8
98 63.6 50.7 76.1





#13

Acrolein

Concen: 86.416 ug/l

RT: 2.245 min Scan# 190

Delta R.T. -0.000 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51

Instrument :

MSVOA_X

ClientSampleId :





VX0619WBSD01

Tgt Ion: 56 Resp: 18810

Ion Ratio Lower Upper

56 100

55 71.1 55.2 82.8



#14

Allyl chloride

Concen: 19.051 ug/l

RT: 2.678 min Scan# 261

Delta R.T. -0.000 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51




Tgt Ion: 41 Resp: 45893


Ion Ratio Lower Upper

41 100

39 67.1 53.6 80.4

76 33.3 26.8 40.2









#15
Acrylonitrile
Concen: 91.312 ug/l
RT: 3.068 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01


Tgt Ion: 53 Resp: 61358
Ion Ratio Lower Upper
53 100
52 82.1 65.1 97.7
51 36.0 28.9 43.3



#16
Acetone
Concen: 81.534 ug/l
RT: 2.386 min Scan# 213
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Tgt Ion: 43 Resp: 45269
Ion Ratio Lower Upper
43 100
58 33.3 24.5 36.7





#17

Carbon Disulfide

Concen: 17.970 ug/l

RT: 2.526 min Scan# 21

Delta R.T. -0.000 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51

Instrument :

MSVOA_X

ClientSampleId :





VX0619WBSD01

Tgt Ion: 76 Resp: 72150

Ion Ratio Lower Upper

76 100

78 9.2 7.5 11.3



#18

Methyl Acetate

Concen: 17.753 ug/l

RT: 2.709 min Scan# 266

Delta R.T. -0.006 min

Lab File: VX046763.D




Acq: 19 Jun 2025 11:51


Tgt Ion: 43 Resp: 25035

Ion Ratio Lower Upper

43 100

74 23.0 18.4 27.6









#19
Methyl tert-butyl Ether
Concen: 18.661 ug/l
RT: 3.117 min Scan# 311
Delta R.T. -0.006 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01


Tgt Ion: 73 Resp: 77256
Ion Ratio Lower Upper
73 100
57 21.9 18.2 27.2



#20
Methylene Chloride
Concen: 19.323 ug/l
RT: 2.800 min Scan# 281
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Tgt Ion: 84 Resp: 29836
Ion Ratio Lower Upper
84 100
49 126.4 101.3 151.9
51 38.8 31.0 46.6
86 66.4 51.1 76.7









#21
trans-1,2-Dichloroethene
Concen: 19.176 ug/l
RT: 3.105 min Scan# 311
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01


Tgt Ion: 96 Resp: 27280
Ion Ratio Lower Upper
96 100
61 138.8 112.5 168.7
98 60.0 51.9 77.9



#22
Diisopropyl ether
Concen: 20.321 ug/l
RT: 3.764 min Scan# 439
Delta R.T. -0.006 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Tgt Ion: 45 Resp: 94092
Ion Ratio Lower Upper
45 100
43 133.4 91.9 137.9
87 25.5 22.0 33.0
59 11.3 9.1 13.7









#23
 Vinyl Acetate
 Concen: 97.697 ug/l
 RT: 3.733 min Scan# 41
 Delta R.T. -0.000 min
 Lab File: VX046763.D
 Acq: 19 Jun 2025 11:51




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0619WBSD01


Tgt Ion: 43 Resp: 359723
 Ion Ratio Lower Upper
 43 100
 86 10.3 8.2 12.2



#24
 1,1-Dichloroethane
 Concen: 19.372 ug/l
 RT: 3.623 min Scan# 416
 Delta R.T. -0.000 min
 Lab File: VX046763.D
 Acq: 19 Jun 2025 11:51

Tgt Ion: 63 Resp: 50927
 Ion Ratio Lower Upper
 63 100
 98 8.0 4.0 11.9
 100 4.5 2.2 6.6





#25

2-Butanone

Concen: 87.030 ug/l

RT: 4.562 min Scan# 51

Delta R.T. -0.000 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51

Instrument :

MSVOA_X

ClientSampleId :





VX0619WBSD01

Tgt Ion: 43 Resp: 68340

Ion Ratio Lower Upper

43 100

72 24.0 20.2 30.4



#26

2,2-Dichloropropane

Concen: 18.883 ug/l

RT: 4.489 min Scan# 558

Delta R.T. -0.000 min

Lab File: VX046763.D




Acq: 19 Jun 2025 11:51


Tgt Ion: 77 Resp: 36206

Ion Ratio Lower Upper

77 100

97 22.8 11.8 35.3





#27

cis-1,2-Dichloroethene

Concen: 19.438 ug/l

RT: 4.501 min Scan# 560

Delta R.T. -0.000 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBSD01





Tgt Ion: 96 Resp: 32489

Ion Ratio Lower Upper

96 100

61 146.1 0.0 291.0

98 63.4 0.0 126.8



#28

Bromochloromethane

Concen: 23.092 ug/l

RT: 4.910 min Scan# 627

Delta R.T. -0.000 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51




Tgt Ion: 49 Resp: 27522


Ion Ratio Lower Upper

49 100

129 1.6 0.0 3.8

130 73.9 61.4 92.0





#29

Tetrahydrofuran

Concen: 86.542 ug/l

RT: 5.013 min Scan# 64

Delta R.T. -0.000 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBSD01





Tgt Ion: 42 Resp: 43894

Ion Ratio Lower Upper

42 100

72 44.2 34.9 52.3

71 39.6 32.2 48.2



#30

Chloroform

Concen: 20.452 ug/l

RT: 5.099 min Scan# 658

Delta R.T. -0.006 min

Lab File: VX046763.D




Acq: 19 Jun 2025 11:51


Tgt Ion: 83 Resp: 53771

Ion Ratio Lower Upper

83 100

85 64.5 53.7 80.5









#31
Cyclohexane
Concen: 19.660 ug/l
RT: 5.477 min Scan# 719
Delta R.T. -0.006 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01


Tgt Ion: 56 Resp: 49014
Ion Ratio Lower Upper
56 100
69 27.9 24.8 37.2
84 85.3 68.2 102.2



#32
1,1,1-Trichloroethane
Concen: 19.276 ug/l
RT: 5.391 min Scan# 706
Delta R.T. -0.006 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Tgt Ion: 97 Resp: 44445
Ion Ratio Lower Upper
97 100
99 64.2 51.3 76.9
61 46.4 36.4 54.6





#33

1,2-Dichloroethane-d4

Concen: 51.530 ug/l

RT: 5.958 min Scan# 799

Delta R.T. -0.006 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51

Instrument :

MSVOA_X

ClientSampleId :





VX0619WBSD01

Tgt Ion: 65 Resp: 85794

Ion Ratio Lower Upper

65 100

67 52.5 0.0 105.4



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. -0.006 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51




Tgt Ion: 114 Resp: 199496


Ion Ratio Lower Upper

114 100

63 20.4 0.0 44.2

88 16.2 0.0 33.2









#35
Dibromofluoromethane
Concen: 52.795 ug/l
RT: 5.391 min Scan# 707
Delta R.T. -0.006 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01


Tgt Ion:113 Resp: 69686
Ion Ratio Lower Upper
113 100
111 102.4 82.0 123.0
192 18.6 15.3 22.9



#36
1,1-Dichloropropene
Concen: 18.799 ug/l
RT: 5.702 min Scan# 757
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Tgt Ion: 75 Resp: 35819
Ion Ratio Lower Upper
75 100
110 36.2 17.2 51.5
77 31.6 24.2 36.4









#37
Ethyl Acetate
Concen: 17.627 ug/l
RT: 4.721 min Scan# 596
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01


Tgt Ion: 43 Resp: 31001
Ion Ratio Lower Upper
43 100
61 13.3 9.7 14.5
70 9.4 8.1 12.1



#38
Carbon Tetrachloride
Concen: 19.117 ug/l
RT: 5.684 min Scan# 754
Delta R.T. -0.006 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Tgt Ion: 117 Resp: 39531
Ion Ratio Lower Upper
117 100
119 94.8 75.2 112.8
121 29.5 24.1 36.1









#39
Methylcyclohexane
Concen: 18.962 ug/l
RT: 7.385 min Scan# 1033
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01


Tgt Ion: 83 Resp: 47544
Ion Ratio Lower Upper
83 100
55 79.5 63.0 94.4
98 49.7 38.8 58.2



#40
Benzene
Concen: 20.015 ug/l
RT: 6.044 min Scan# 813
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Tgt Ion: 78 Resp: 112866
Ion Ratio Lower Upper
78 100
77 24.3 18.8 28.2









#41
Methacrylonitrile
Concen: 18.985 ug/l
RT: 4.928 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01


Tgt Ion:	41	Resp:	17801
Ion	Ratio	Lower	Upper
41	100		
39	54.0	45.9	68.9
67	72.8	61.2	91.8
52	30.2	25.5	38.3



#42
1,2-Dichloroethane
Concen: 20.100 ug/l
RT: 6.092 min Scan# 821
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Tgt Ion:	62	Resp:	39829
Ion	Ratio	Lower	Upper
62	100		
98	9.3	0.0	17.8









#43
Isopropyl Acetate
Concen: 18.463 ug/l
RT: 6.342 min Scan# 862
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01


Tgt Ion: 43 Resp: 50813
Ion Ratio Lower Upper
43 100
61 18.4 15.3 22.9
87 12.4 10.0 15.0



#44
Trichloroethene
Concen: 19.321 ug/l
RT: 7.129 min Scan# 991
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Tgt Ion:130 Resp: 27765
Ion Ratio Lower Upper
130 100
95 112.5 0.0 202.6





#45

1,2-Dichloropropane

Concen: 20.276 ug/l

RT: 7.427 min Scan# 1040

Delta R.T. -0.006 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51

Instrument :

MSVOA_X

ClientSampleId :





VX0619WBSD01

Tgt Ion: 63 Resp: 28336

Ion Ratio Lower Upper

63 100

65 32.8 22.9 34.3



#46

Dibromomethane

Concen: 19.891 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. -0.006 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51




Tgt Ion: 93 Resp: 19978


Ion Ratio Lower Upper

93 100

95 82.3 66.2 99.4

174 97.0 77.0 115.6









#47
Bromodichloromethane
Concen: 20.112 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01


Tgt Ion: 83 Resp: 41221
Ion Ratio Lower Upper
83 100
85 65.4 51.4 77.2
127 8.7 6.9 10.3



#48
Methyl methacrylate
Concen: 18.904 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Tgt Ion: 41 Resp: 26092
Ion Ratio Lower Upper
41 100
69 79.1 63.4 95.2
39 56.2 45.4 68.0









#49
1,4-Dioxane
Concen: 321.885 ug/l
RT: 7.659 min Scan# 1078
Delta R.T. 0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01


Tgt Ion: 88 Resp: 6068
Ion Ratio Lower Upper
88 100
43 42.8 31.2 46.8
58 69.2 54.7 82.1



#50
Toluene-d8
Concen: 52.661 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Tgt Ion: 98 Resp: 249730
Ion Ratio Lower Upper
98 100
100 66.2 53.0 79.4









#51
4-Methyl-2-Pentanone
Concen: 91.811 ug/l
RT: 8.574 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01


Tgt Ion: 43 Resp: 150700
Ion Ratio Lower Upper
43 100
58 37.2 30.8 46.2



#52
Toluene
Concen: 20.310 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Tgt Ion: 92 Resp: 70998
Ion Ratio Lower Upper
92 100
91 170.3 134.8 202.2





#53

t-1,3-Dichloropropene

Concen: 19.536 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51

Instrument :

MSVOA_X

ClientSampleId :





VX0619WBSD01

Tgt Ion: 75 Resp: 37151

Ion Ratio Lower Upper

75 100

77 30.8 25.6 38.4



#54

cis-1,3-Dichloropropene

Concen: 19.939 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51




Tgt Ion: 75 Resp: 43044


Ion Ratio Lower Upper

75 100

77 31.4 25.3 37.9

39 48.8 39.2 58.8









#55
 1,1,2-Trichloroethane
 Concen: 20.409 ug/l
 RT: 9.153 min Scan# 1323
 Delta R.T. -0.000 min
 Lab File: VX046763.D
 Acq: 19 Jun 2025 11:51




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0619WBSD01


Tgt Ion:	97	Resp:	26588
Ion	Ratio	Lower	Upper
97	100		
83	80.6	67.2	100.8
85	55.0	44.1	66.1
99	64.8	50.3	75.5



#56
 Ethyl methacrylate
 Concen: 18.636 ug/l
 RT: 9.116 min Scan# 1317
 Delta R.T. -0.000 min
 Lab File: VX046763.D
 Acq: 19 Jun 2025 11:51

Tgt Ion:	69	Resp:	35838
Ion	Ratio	Lower	Upper
69	100		
41	71.4	55.4	83.2
39	38.8	29.7	44.5





#57

1,3-Dichloropropane

Concen: 20.277 ug/l

RT: 9.305 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51

Instrument :

MSVOA_X

ClientSampleId :





VX0619WBSD01

Tgt Ion: 76 Resp: 45490

Ion Ratio Lower Upper

76 100

78 31.6 25.7 38.5



#58

2-Chloroethyl Vinyl ether

Concen: 100.344 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. -0.006 min

Lab File: VX046763.D




Acq: 19 Jun 2025 11:51


Tgt Ion: 63 Resp: 100239

Ion Ratio Lower Upper

63 100

106 28.8 22.6 34.0









#59
2-Hexanone
Concen: 88.846 ug/l
RT: 9.427 min Scan# 1368
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01


Tgt Ion: 43 Resp: 100820
Ion Ratio Lower Upper
43 100
58 53.3 26.8 80.4



#60
Dibromochloromethane
Concen: 19.947 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Tgt Ion:129 Resp: 30624
Ion Ratio Lower Upper
129 100
127 77.9 38.4 115.2









#61
1,2-Dibromoethane
Concen: 20.084 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01


Tgt Ion:107 Resp: 26614
Ion Ratio Lower Upper
107 100
109 93.0 75.8 113.6



#62
4-Bromofluorobenzene
Concen: 54.915 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Tgt Ion: 95 Resp: 97074
Ion Ratio Lower Upper
95 100
174 74.5 0.0 150.4
176 71.8 0.0 145.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01


Tgt Ion:117 Resp: 183807
Ion Ratio Lower Upper
117 100
82 54.4 44.6 66.8
119 31.9 25.8 38.8



#64
Tetrachloroethene
Concen: 18.778 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Tgt Ion:164 Resp: 23913
Ion Ratio Lower Upper
164 100
166 128.8 104.6 157.0
129 99.3 76.3 114.5
131 92.5 74.1 111.1









#65
Chlorobenzene
Concen: 19.421 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01


Tgt Ion:112 Resp: 78801
Ion Ratio Lower Upper
112 100
114 32.1 25.5 38.3



#66
1,1,1,2-Tetrachloroethane
Concen: 19.449 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Tgt Ion:131 Resp: 26522
Ion Ratio Lower Upper
131 100
133 94.3 47.3 141.8
119 64.3 32.1 96.3





#67

Ethyl Benzene

Concen: 19.141 ug/l

RT: 10.195 min Scan# 1493

Delta R.T. 0.006 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51

Instrument :

MSVOA_X

ClientSampleId :





VX0619WBSD01

Tgt Ion: 91 Resp: 135702

Ion Ratio Lower Upper

91 100

106 29.4 24.2 36.2



#68

m/p-Xylenes

Concen: 38.810 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. -0.000 min

Lab File: VX046763.D




Acq: 19 Jun 2025 11:51


Tgt Ion: 106 Resp: 102519

Ion Ratio Lower Upper

106 100

91 208.3 164.8 247.2









#69
o-Xylene
Concen: 19.820 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01


Tgt Ion:106 Resp: 49068
Ion Ratio Lower Upper
106 100
91 221.1 111.3 333.9



#70
Styrene
Concen: 19.696 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Tgt Ion:104 Resp: 85333
Ion Ratio Lower Upper
104 100
78 50.6 41.4 62.2
103 55.0 43.1 64.7





#71

Bromoform

Concen: 18.418 ug/l

RT: 10.799 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBSD01





Tgt Ion:173 Resp: 19127

Ion Ratio Lower Upper

173 100

175 49.8 24.1 72.4

254 0.0 0.1 0.1#



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 1793

Delta R.T. -0.000 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51




Tgt Ion:152 Resp: 94851


Ion Ratio Lower Upper

152 100

115 70.0 43.2 129.6

150 162.0 0.0 346.8









#73
Isopropylbenzene
Concen: 18.592 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01


Tgt Ion:105 Resp: 129881
Ion Ratio Lower Upper
105 100
120 26.1 13.0 39.0



#74
N-ethyl acetate
Concen: 17.214 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Tgt Ion: 43 Resp: 45562
Ion Ratio Lower Upper
43 100
70 41.4 33.4 50.0
55 24.6 19.1 28.7
61 23.8 19.0 28.4









#75
1,1,2-Tetrachloroethane
Concen: 18.425 ug/l
RT: 11.207 min Scan# 1660
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01


Tgt Ion: 83 Resp: 35939
Ion Ratio Lower Upper
83 100
131 10.0 4.9 14.6
85 63.9 32.0 96.0



#76
1,2,3-Trichloropropane
Concen: 20.823 ug/l
RT: 11.238 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Tgt Ion: 75 Resp: 36174
Ion Ratio Lower Upper
75 100
77 33.2 18.5 55.5









#77
Bromobenzene
Concen: 18.976 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01


Tgt Ion:156 Resp: 31899
Ion Ratio Lower Upper
156 100
77 169.0 82.8 248.6
158 96.8 47.6 142.9



#78
n-propylbenzene
Concen: 18.803 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Tgt Ion: 91 Resp: 156018
Ion Ratio Lower Upper
91 100
120 22.6 11.4 34.1









#79
2-Chlorotoluene
Concen: 18.633 ug/l
RT: 11.360 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01


Tgt Ion: 91 Resp: 92310
Ion Ratio Lower Upper
91 100
126 33.6 16.6 49.7



#80
1,3,5-Trimethylbenzene
Concen: 18.925 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Tgt Ion: 105 Resp: 108639
Ion Ratio Lower Upper
105 100
120 48.5 24.1 72.2









#81
trans-1,4-Dichloro-2-butene
Concen: 15.556 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01


Tgt Ion: 75 Resp: 9561
Ion Ratio Lower Upper
75 100
53 116.3 86.8 130.2
89 46.3 36.9 55.3



#82
4-Chlorotoluene
Concen: 18.795 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Tgt Ion: 91 Resp: 108666
Ion Ratio Lower Upper
91 100
126 29.3 14.4 43.2









#83
tert-Butylbenzene
Concen: 18.337 ug/l
RT: 11.713 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01


Tgt Ion:119 Resp: 108810
Ion Ratio Lower Upper
119 100
91 59.3 29.5 88.6
134 23.4 11.7 35.0



#84
1,2,4-Trimethylbenzene
Concen: 18.872 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Tgt Ion:105 Resp: 108849
Ion Ratio Lower Upper
105 100
120 44.5 22.1 66.3









#85
 sec-Butylbenzene
 Concen: 18.839 ug/l
 RT: 11.890 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: VX046763.D
 Acq: 19 Jun 2025 11:51




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0619WBSD01


Tgt Ion:105 Resp: 141281
 Ion Ratio Lower Upper
 105 100
 134 19.6 9.8 29.4



#86
 p-Isopropyltoluene
 Concen: 18.732 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. -0.000 min
 Lab File: VX046763.D
 Acq: 19 Jun 2025 11:51

Tgt Ion:119 Resp: 118528
 Ion Ratio Lower Upper
 119 100
 134 26.0 12.9 38.7
 91 25.2 12.3 36.9





#87

1,3-Dichlorobenzene

Concen: 18.553 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBSD01





Tgt Ion:146 Resp: 60812

Ion Ratio Lower Upper

146 100

111 40.6 20.6 61.8

148 63.4 31.6 94.8



#88

1,4-Dichlorobenzene

Concen: 17.772 ug/l

RT: 12.036 min Scan# 1796

Delta R.T. -0.000 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51




Tgt Ion:146 Resp: 59834


Ion Ratio Lower Upper

146 100

111 42.6 20.3 61.0

148 65.7 32.1 96.5









#89
n-Butylbenzene
Concen: 18.526 ug/l
RT: 12.329 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01


Tgt Ion: 91 Resp: 111302
Ion Ratio Lower Upper
91 100
92 53.7 27.1 81.3
134 25.4 12.9 38.6



#90
Hexachloroethane
Concen: 17.220 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Tgt Ion: 117 Resp: 19081
Ion Ratio Lower Upper
117 100
201 72.0 34.4 103.3









#91
1,2-Dichlorobenzene
Concen: 18.997 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.006 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51




Instrument :
MSVOA_X
ClientSampleId :
VX0619WBSD01


Tgt Ion:146 Resp: 58188
Ion Ratio Lower Upper
146 100
111 41.8 21.7 65.1
148 63.9 32.0 96.2



#92
1,2-Dibromo-3-Chloropropane
Concen: 16.116 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. -0.000 min
Lab File: VX046763.D
Acq: 19 Jun 2025 11:51

Tgt Ion: 75 Resp: 6206
Ion Ratio Lower Upper
75 100
155 85.7 41.9 125.7
157 106.9 51.4 154.3





#93

1,2,4-Trichlorobenzene

Concen: 18.309 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. -0.000 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBSD01





Tgt Ion:180 Resp: 39871

Ion Ratio Lower Upper

180 100

182 94.6 46.7 140.1

145 31.9 15.8 47.4



#94

Hexachlorobutadiene

Concen: 18.715 ug/l

RT: 13.719 min Scan# 2072

Delta R.T. -0.000 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51




Tgt Ion:225 Resp: 17144


Ion Ratio Lower Upper

225 100

223 63.4 32.5 97.5

227 60.2 32.2 96.6





#95

Naphthalene

Concen: 17.340 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. -0.000 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51

Instrument :

MSVOA_X

ClientSampleId :

VX0619WBSD01





Tgt Ion:128 Resp: 107578

Ion Ratio Lower Upper

128 100

127 12.9 10.1 15.1

129 10.7 8.7 13.1



#96

1,2,3-Trichlorobenzene

Concen: 18.189 ug/l

RT: 13.957 min Scan# 2111

Delta R.T. -0.000 min

Lab File: VX046763.D

Acq: 19 Jun 2025 11:51




Tgt Ion:180 Resp: 37893

Ion Ratio Lower Upper

180 100

182 96.0 47.8 143.4

145 33.9 17.3 51.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
 Data File : VX046775.D
 Acq On : 19 Jun 2025 17:10
 Operator : JC/MD
 Sample : VSTDC0050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDC0050EC

Quant Time: Jun 20 05:22:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	129409	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	216626	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	190720	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	95668	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	83815	46.825	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.660%
35) Dibromofluoromethane	5.397	113	69249	48.316	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.640%
50) Toluene-d8	8.647	98	246552	47.880	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.760%
62) 4-Bromofluorobenzene	11.079	95	93266	48.588	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	97.180%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.179	85	74095	53.629	ug/l	99
3) Chloromethane	1.307	50	77768	52.149	ug/l	96
4) Vinyl Chloride	1.386	62	85590	53.822	ug/l	98
5) Bromomethane	1.618	94	49377	55.188	ug/l	95
6) Chloroethane	1.703	64	53474	55.480	ug/l	94
7) Trichlorofluoromethane	1.904	101	127152	52.659	ug/l	99
8) Diethyl Ether	2.148	74	44616	53.840	ug/l	98
9) 1,1,2-Trichlorotrifluo...	2.349	101	79479	53.654	ug/l	98
10) Methyl Iodide	2.465	142	85230	46.314	ug/l	100
11) Tert butyl alcohol	2.959	59	35007	217.898	ug/l	100
12) 1,1-Dichloroethene	2.337	96	76619	53.603	ug/l	98
13) Acrolein	2.245	56	69969	298.999	ug/l	98
14) Allyl chloride	2.678	41	137304	53.017	ug/l	98
15) Acrylonitrile	3.075	53	185932	257.376	ug/l	99
16) Acetone	2.386	43	144349	241.828	ug/l	97
17) Carbon Disulfide	2.526	76	205491	47.607	ug/l	99
18) Methyl Acetate	2.715	43	78667	51.889	ug/l	99
19) Methyl tert-butyl Ether	3.123	73	225140	50.585	ug/l	100
20) Methylene Chloride	2.800	84	83440	50.266	ug/l	99
21) trans-1,2-Dichloroethene	3.105	96	79204	51.786	ug/l	99
22) Diisopropyl ether	3.770	45	270205	54.281	ug/l	98
23) Vinyl Acetate	3.733	43	1055978	266.763	ug/l	99
24) 1,1-Dichloroethane	3.623	63	148012	52.369	ug/l	99
25) 2-Butanone	4.562	43	216354	256.282	ug/l	97
26) 2,2-Dichloropropane	4.489	77	106712	51.768	ug/l	100
27) cis-1,2-Dichloroethene	4.501	96	94359	52.513	ug/l	99
28) Bromochloromethane	4.910	49	66329	51.765	ug/l	99
29) Tetrahydrofuran	5.007	42	139493	255.819	ug/l	99
30) Chloroform	5.105	83	149506	52.894	ug/l	95
31) Cyclohexane	5.477	56	137290	51.222	ug/l	94
32) 1,1,1-Trichloroethane	5.397	97	125889	50.784	ug/l	99
36) 1,1-Dichloropropene	5.702	75	105895	51.182	ug/l	100
37) Ethyl Acetate	4.721	43	92088	48.219	ug/l	98
38) Carbon Tetrachloride	5.690	117	111211	49.529	ug/l	99
39) Methylcyclohexane	7.385	83	139618	51.281	ug/l	99
40) Benzene	6.050	78	318931	52.086	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
 Data File : VX046775.D
 Acq On : 19 Jun 2025 17:10
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Quant Time: Jun 20 05:22:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.928	41	55058	54.078	ug/l	95
42) 1,2-Dichloroethane	6.092	62	109435	50.860	ug/l	98
43) Isopropyl Acetate	6.342	43	151380	50.655	ug/l	100
44) Trichloroethene	7.135	130	79977	51.252	ug/l	97
45) 1,2-Dichloropropane	7.434	63	78602	51.797	ug/l	94
46) Dibromomethane	7.586	93	54750	50.201	ug/l	99
47) Bromodichloromethane	7.824	83	114561	51.475	ug/l	99
48) Methyl methacrylate	7.696	41	80176	53.494	ug/l	99
49) 1,4-Dioxane	7.665	88	19434	894.415	ug/l	94
51) 4-Methyl-2-Pentanone	8.574	43	453562	254.472	ug/l	99
52) Toluene	8.720	92	195403	51.477	ug/l	97
53) t-1,3-Dichloropropene	8.976	75	106195	51.427	ug/l	98
54) cis-1,3-Dichloropropene	8.366	75	122494	52.254	ug/l	99
55) 1,1,2-Trichloroethane	9.153	97	73528	51.978	ug/l	95
56) Ethyl methacrylate	9.116	69	109987	52.671	ug/l	99
57) 1,3-Dichloropropane	9.305	76	125247	51.414	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.244	63	321699	296.571	ug/l	100
59) 2-Hexanone	9.427	43	310545	252.021	ug/l	100
60) Dibromochloromethane	9.519	129	85854	51.500	ug/l	99
61) 1,2-Dibromoethane	9.610	107	73504	51.083	ug/l	99
64) Tetrachloroethene	9.275	164	66307	50.181	ug/l	97
65) Chlorobenzene	10.079	112	216318	51.382	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	72835	51.476	ug/l	99
67) Ethyl Benzene	10.195	91	380469	51.722	ug/l	99
68) m/p-Xylenes	10.299	106	284342	103.739	ug/l	99
69) o-Xylene	10.640	106	136531	53.151	ug/l	98
70) Styrene	10.653	104	238877	53.139	ug/l	99
71) Bromoform	10.799	173	54055	50.164	ug/l #	100
73) Isopropylbenzene	10.963	105	364838	51.780	ug/l	100
74) N-amyl acetate	10.841	43	134737	50.472	ug/l	100
75) 1,1,2,2-Tetrachloroethane	11.207	83	100156	50.908	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	105868	60.421	ug/l	93
77) Bromobenzene	11.195	156	87850	51.813	ug/l	100
78) n-propylbenzene	11.305	91	432963	51.735	ug/l	99
79) 2-Chlorotoluene	11.360	91	249407	49.914	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	298594	51.570	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	29148	47.021	ug/l	100
82) 4-Chlorotoluene	11.451	91	293281	50.293	ug/l	100
83) tert-Butylbenzene	11.713	119	304194	50.825	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	302076	51.927	ug/l	98
85) sec-Butylbenzene	11.890	105	387895	51.281	ug/l	99
86) p-Isopropyltoluene	12.006	119	325914	51.067	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	162851	49.259	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	161313	47.505	ug/l	100
89) n-Butylbenzene	12.329	91	308844	50.967	ug/l	100
90) Hexachloroethane	12.536	117	54659	48.908	ug/l	98
91) 1,2-Dichlorobenzene	12.335	146	154918	50.145	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	18542	47.738	ug/l	96
93) 1,2,4-Trichlorobenzene	13.585	180	109313	49.769	ug/l	98
94) Hexachlorobutadiene	13.725	225	45830	49.603	ug/l	99
95) Naphthalene	13.774	128	313462	50.093	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	105543	50.228	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
Data File : VX046775.D
Acq On : 19 Jun 2025 17:10
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Quant Time: Jun 20 05:22:53 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 03:09:16 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061925\
Data File : VX046775.D
Acq On : 19 Jun 2025 17:10
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Quant Time: Jun 20 05:22:53 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 03:09:16 2025
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.562 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC







Tgt Ion:168 Resp: 129409
Ion Ratio Lower Upper
168 100
99 58.5 48.5 72.7



#2
Dichlorodifluoromethane
Concen: 53.629 ug/l
RT: 1.179 min Scan# 15
Delta R.T. -0.006 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10

Tgt Ion: 85 Resp: 74095
Ion Ratio Lower Upper
85 100
87 32.6 16.6 49.8









#3
Chloromethane
Concen: 52.149 ug/l
RT: 1.307 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 50 Resp: 77768
Ion Ratio Lower Upper
50 100
52 33.2 24.9 37.3



#4
Vinyl Chloride
Concen: 53.822 ug/l
RT: 1.386 min Scan# 49
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10

Tgt Ion: 62 Resp: 85590
Ion Ratio Lower Upper
62 100
64 30.8 23.9 35.9









#5
Bromomethane
Concen: 55.188 ug/l
RT: 1.618 min Scan# 81
Delta R.T. -0.006 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 94 Resp: 49377
Ion Ratio Lower Upper
94 100
96 95.6 72.3 108.5



#6
Chloroethane
Concen: 55.480 ug/l
RT: 1.703 min Scan# 101
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10

Tgt Ion: 64 Resp: 53474
Ion Ratio Lower Upper
64 100
66 30.7 27.1 40.7









#7
Trichlorofluoromethane
Concen: 52.659 ug/l
RT: 1.904 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:101 Resp: 127152
Ion Ratio Lower Upper
101 100
103 65.4 51.8 77.8



#8
Diethyl Ether
Concen: 53.840 ug/l
RT: 2.148 min Scan# 174
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10

Tgt Ion: 74 Resp: 44616
Ion Ratio Lower Upper
74 100
45 102.7 52.3 157.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 53.654 ug/l
RT: 2.349 min Scan# 206
Delta R.T. 0.006 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:101 Resp: 79479
Ion Ratio Lower Upper
101 100
85 43.2 35.4 53.0
151 74.1 61.2 91.8



#10
Methyl Iodide
Concen: 46.314 ug/l
RT: 2.465 min Scan# 226
Delta R.T. -0.006 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10

Tgt Ion:142 Resp: 85230
Ion Ratio Lower Upper
142 100
127 46.1 37.0 55.6
141 14.0 11.5 17.3









#11
Tert butyl alcohol
Concen: 217.898 ug/l
RT: 2.959 min Scan# 307
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 59 Resp: 35007
Ion Ratio Lower Upper
59 100
57 10.5 8.3 12.5



#12
1,1-Dichloroethene
Concen: 53.603 ug/l
RT: 2.337 min Scan# 205
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10

Tgt Ion: 96 Resp: 76619
Ion Ratio Lower Upper
96 100
61 161.7 126.6 189.8
98 64.0 50.7 76.1









#13
Acrolein
Concen: 298.999 ug/l
RT: 2.245 min Scan# 190
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 56 Resp: 69969
Ion Ratio Lower Upper
56 100
55 70.7 55.2 82.8



#14
Allyl chloride
Concen: 53.017 ug/l
RT: 2.678 min Scan# 261
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10

Tgt Ion: 41 Resp: 137304
Ion Ratio Lower Upper
41 100
39 65.0 53.6 80.4
76 32.3 26.8 40.2









#15
Acrylonitrile
Concen: 257.376 ug/l
RT: 3.075 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 53 Resp: 185932
Ion Ratio Lower Upper
53 100
52 82.3 65.1 97.7
51 36.4 28.9 43.3



#16
Acetone
Concen: 241.828 ug/l
RT: 2.386 min Scan# 213
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10

Tgt Ion: 43 Resp: 144349
Ion Ratio Lower Upper
43 100
58 32.3 24.5 36.7









#17
Carbon Disulfide
Concen: 47.607 ug/l
RT: 2.526 min Scan# 21
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 76 Resp: 205491
Ion Ratio Lower Upper
76 100
78 9.1 7.5 11.3



#18
Methyl Acetate
Concen: 51.889 ug/l
RT: 2.715 min Scan# 267
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10

Tgt Ion: 43 Resp: 78667
Ion Ratio Lower Upper
43 100
74 22.7 18.4 27.6









#19
Methyl tert-butyl Ether
Concen: 50.585 ug/l
RT: 3.123 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 73 Resp: 225140
Ion Ratio Lower Upper
73 100
57 22.9 18.2 27.2



#20
Methylene Chloride
Concen: 50.266 ug/l
RT: 2.800 min Scan# 281
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10

Tgt Ion: 84 Resp: 83440
Ion Ratio Lower Upper
84 100
49 127.7 101.3 151.9
51 40.1 31.0 46.6
86 62.9 51.1 76.7









#21
trans-1,2-Dichloroethene
Concen: 51.786 ug/l
RT: 3.105 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 96 Resp: 79204
Ion Ratio Lower Upper
96 100
61 141.7 112.5 168.7
98 65.4 51.9 77.9



#22
Diisopropyl ether
Concen: 54.281 ug/l
RT: 3.770 min Scan# 440
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10

Tgt Ion: 45 Resp: 270205
Ion Ratio Lower Upper
45 100
43 117.8 91.9 137.9
87 27.3 22.0 33.0
59 11.3 9.1 13.7









#23
 Vinyl Acetate
 Concen: 266.763 ug/l
 RT: 3.733 min Scan# 41
 Delta R.T. -0.000 min
 Lab File: VX046775.D
 Acq: 19 Jun 2025 17:10




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion: 43 Resp: 1055978
 Ion Ratio Lower Upper
 43 100
 86 10.0 8.2 12.2



#24
 1,1-Dichloroethane
 Concen: 52.369 ug/l
 RT: 3.623 min Scan# 416
 Delta R.T. -0.000 min
 Lab File: VX046775.D
 Acq: 19 Jun 2025 17:10

Tgt Ion: 63 Resp: 148012
 Ion Ratio Lower Upper
 63 100
 98 7.3 4.0 11.9
 100 4.1 2.2 6.6





#25

2-Butanone

Concen: 256.282 ug/l

RT: 4.562 min Scan# 5

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 43 Resp: 216354

Ion Ratio Lower Upper

43 100

72 23.7 20.2 30.4



#26

2,2-Dichloropropane

Concen: 51.768 ug/l

RT: 4.489 min Scan# 558

Delta R.T. -0.000 min

Lab File: VX046775.D




Acq: 19 Jun 2025 17:10


Tgt Ion: 77 Resp: 106712

Ion Ratio Lower Upper

77 100

97 23.6 11.8 35.3





#27

cis-1,2-Dichloroethene

Concen: 52.513 ug/l

RT: 4.501 min Scan# 560

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 96 Resp: 94359

Ion Ratio Lower Upper

96 100

61 143.7 0.0 291.0

98 63.9 0.0 126.8



#28

Bromochloromethane

Concen: 51.765 ug/l

RT: 4.910 min Scan# 627

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10




Tgt Ion: 49 Resp: 66329


Ion Ratio Lower Upper

49 100

129 1.9 0.0 3.8

130 75.8 61.4 92.0





#29

Tetrahydrofuran

Concen: 255.819 ug/l

RT: 5.007 min Scan# 644

Delta R.T. -0.006 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 42 Resp: 139493

Ion Ratio Lower Upper

42 100

72 42.7 34.9 52.3

71 39.6 32.2 48.2



#30

Chloroform

Concen: 52.894 ug/l

RT: 5.105 min Scan# 659

Delta R.T. -0.000 min

Lab File: VX046775.D




Acq: 19 Jun 2025 17:10


Tgt Ion: 83 Resp: 149506

Ion Ratio Lower Upper

83 100


85 63.1 53.7 80.5








#31
Cyclohexane
Concen: 51.222 ug/l
RT: 5.477 min Scan# 720
Delta R.T. -0.006 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 56 Resp: 137290




Ion	Ratio	Lower	Upper
56	100		
69	28.7	24.8	37.2
84	79.7	68.2	102.2




#32
1,1,1-Trichloroethane
Concen: 50.784 ug/l
RT: 5.397 min Scan# 707
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10

Tgt Ion: 97 Resp: 125889

Ion	Ratio	Lower	Upper
97	100		
99	63.9	51.3	76.9
61	46.1	36.4	54.6









#33
 1,2-Dichloroethane-d4
 Concen: 46.825 ug/l
 RT: 5.964 min Scan# 800
 Delta R.T. -0.000 min
 Lab File: VX046775.D
 Acq: 19 Jun 2025 17:10




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion: 65 Resp: 83815
 Ion Ratio Lower Upper
 65 100
 67 52.4 0.0 105.4



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.769 min Scan# 932
 Delta R.T. 0.000 min
 Lab File: VX046775.D
 Acq: 19 Jun 2025 17:10

Tgt Ion: 114 Resp: 216626
 Ion Ratio Lower Upper
 114 100
 63 19.7 0.0 44.2
 88 16.1 0.0 33.2









#35
Dibromofluoromethane
Concen: 48.316 ug/l
RT: 5.397 min Scan# 707
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:113 Resp: 69249
Ion Ratio Lower Upper
113 100
111 103.9 82.0 123.0
192 18.5 15.3 22.9



#36
1,1-Dichloropropene
Concen: 51.182 ug/l
RT: 5.702 min Scan# 757
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10

Tgt Ion: 75 Resp: 105895
Ion Ratio Lower Upper
75 100
110 34.3 17.2 51.5
77 30.8 24.2 36.4





#37

Ethyl Acetate

Concen: 48.219 ug/l

RT: 4.721 min Scan# 596

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 43 Resp: 92088

Ion Ratio Lower Upper

43 100

61 13.5 9.7 14.5

70 10.2 8.1 12.1



#38

Carbon Tetrachloride

Concen: 49.529 ug/l

RT: 5.690 min Scan# 755

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10




Tgt Ion: 117 Resp: 111211


Ion Ratio Lower Upper

117 100

119 94.7 75.2 112.8

121 30.9 24.1 36.1









#39
Methylcyclohexane
Concen: 51.281 ug/l
RT: 7.385 min Scan# 1033
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 83 Resp: 139618
Ion Ratio Lower Upper
83 100
55 78.5 63.0 94.4
98 49.6 38.8 58.2



#40
Benzene
Concen: 52.086 ug/l
RT: 6.050 min Scan# 814
Delta R.T. 0.006 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10

Tgt Ion: 78 Resp: 318931
Ion Ratio Lower Upper
78 100
77 23.7 18.8 28.2









#41
Methacrylonitrile
Concen: 54.078 ug/l
RT: 4.928 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:	41	Resp:	55058
Ion	Ratio	Lower	Upper
41	100		
39	53.7	45.9	68.9
67	71.4	61.2	91.8
52	30.7	25.5	38.3



#42
1,2-Dichloroethane
Concen: 50.860 ug/l
RT: 6.092 min Scan# 821
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10

Tgt Ion:	62	Resp:	109435
Ion	Ratio	Lower	Upper
62	100		
98	9.5	0.0	17.8









#43
Isopropyl Acetate
Concen: 50.655 ug/l
RT: 6.342 min Scan# 862
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 151380
Ion Ratio Lower Upper
43 100
61 19.2 15.3 22.9
87 12.5 10.0 15.0



#44
Trichloroethene
Concen: 51.252 ug/l
RT: 7.135 min Scan# 992
Delta R.T. 0.006 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10

Tgt Ion:130 Resp: 79977
Ion Ratio Lower Upper
130 100
95 98.6 0.0 202.6





#45

1,2-Dichloropropane

Concen: 51.797 ug/l

RT: 7.434 min Scan# 1041

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 63 Resp: 78602

Ion Ratio Lower Upper

63 100

65 31.9 22.9 34.3



#46

Dibromomethane

Concen: 50.201 ug/l

RT: 7.586 min Scan# 1066

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10




Tgt Ion: 93 Resp: 54750


Ion Ratio Lower Upper

93 100

95 83.8 66.2 99.4

174 97.6 77.0 115.6









#47
Bromodichloromethane
Concen: 51.475 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 83 Resp: 114561
Ion Ratio Lower Upper
83 100
85 65.0 51.4 77.2
127 8.5 6.9 10.3



#48
Methyl methacrylate
Concen: 53.494 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10

Tgt Ion: 41 Resp: 80176
Ion Ratio Lower Upper
41 100
69 79.0 63.4 95.2
39 54.7 45.4 68.0









#49
1,4-Dioxane
Concen: 894.415 ug/l
RT: 7.665 min Scan# 1078
Delta R.T. 0.006 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 88 Resp: 19434
Ion Ratio Lower Upper
88 100
43 36.0 31.2 46.8
58 73.2 54.7 82.1



#50
Toluene-d8
Concen: 47.880 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10

Tgt Ion: 98 Resp: 246552
Ion Ratio Lower Upper
98 100
100 67.5 53.0 79.4









#51
4-Methyl-2-Pentanone
Concen: 254.472 ug/l
RT: 8.574 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 453562
Ion Ratio Lower Upper
43 100
58 38.2 30.8 46.2



#52
Toluene
Concen: 51.477 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10

Tgt Ion: 92 Resp: 195403
Ion Ratio Lower Upper
92 100
91 172.1 134.8 202.2





#53

t-1,3-Dichloropropene

Concen: 51.427 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 75 Resp: 106195

Ion Ratio Lower Upper

75 100

77 30.9 25.6 38.4



#54

cis-1,3-Dichloropropene

Concen: 52.254 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10




Tgt Ion: 75 Resp: 122494


Ion Ratio Lower Upper

75 100

77 31.0 25.3 37.9

39 48.8 39.2 58.8









#55
1,1,2-Trichloroethane
Concen: 51.978 ug/l
RT: 9.153 min Scan# 1323
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:	97	Resp:	73528
Ion	Ratio	Lower	Upper
97	100		
83	86.6	67.2	100.8
85	57.4	44.1	66.1
99	69.5	50.3	75.5



#56
Ethyl methacrylate
Concen: 52.671 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10

Tgt Ion:	69	Resp:	109987
Ion	Ratio	Lower	Upper
69	100		
41	69.5	55.4	83.2
39	36.3	29.7	44.5





#57

1,3-Dichloropropane

Concen: 51.414 ug/l

RT: 9.305 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 76 Resp: 125247

Ion Ratio Lower Upper

76 100

78 32.9 25.7 38.5



#58

2-Chloroethyl Vinyl ether

Concen: 296.571 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. -0.000 min

Lab File: VX046775.D




Acq: 19 Jun 2025 17:10


Tgt Ion: 63 Resp: 321699

Ion Ratio Lower Upper

63 100

106 28.4 22.6 34.0









#59
2-Hexanone
Concen: 252.021 ug/l
RT: 9.427 min Scan# 1368
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 310545
Ion Ratio Lower Upper
43 100
58 53.5 26.8 80.4



#60
Dibromochloromethane
Concen: 51.500 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10

Tgt Ion: 129 Resp: 85854
Ion Ratio Lower Upper
129 100
127 75.9 38.4 115.2









#61
1,2-Dibromoethane
Concen: 51.083 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:107 Resp: 73504
Ion Ratio Lower Upper
107 100
109 94.2 75.8 113.6



#62
4-Bromofluorobenzene
Concen: 48.588 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10

Tgt Ion: 95 Resp: 93266
Ion Ratio Lower Upper
95 100
174 75.5 0.0 150.4
176 73.4 0.0 145.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:117 Resp: 190720
Ion Ratio Lower Upper
117 100
82 54.4 44.6 66.8
119 32.1 25.8 38.8



#64
Tetrachloroethene
Concen: 50.181 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10

Tgt Ion:164 Resp: 66307
Ion Ratio Lower Upper
164 100
166 124.5 104.6 157.0
129 94.6 76.3 114.5
131 89.1 74.1 111.1





#65

Chlorobenzene

Concen: 51.382 ug/l

RT: 10.079 min Scan# 1475

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion:112 Resp: 216318

Ion Ratio Lower Upper

112 100

114 31.5 25.5 38.3



#66

1,1,1,2-Tetrachloroethane

Concen: 51.476 ug/l

RT: 10.159 min Scan# 1488

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10




Tgt Ion:131 Resp: 72835


Ion Ratio Lower Upper

131 100

133 96.0 47.3 141.8

119 63.9 32.1 96.3





#67

Ethyl Benzene

Concen: 51.722 ug/l

RT: 10.195 min Scan# 1493

Delta R.T. 0.006 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 91 Resp: 380469

Ion Ratio Lower Upper

91 100

106 30.8 24.2 36.2



#68

m/p-Xylenes

Concen: 103.739 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. -0.000 min

Lab File: VX046775.D




Acq: 19 Jun 2025 17:10


Tgt Ion:106 Resp: 284342

Ion Ratio Lower Upper

106 100

91 207.8 164.8 247.2









#69
o-Xylene
Concen: 53.151 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:106 Resp: 136531
Ion Ratio Lower Upper
106 100
91 219.1 111.3 333.9



#70
Styrene
Concen: 53.139 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10

Tgt Ion:104 Resp: 238877
Ion Ratio Lower Upper
104 100
78 51.0 41.4 62.2
103 53.6 43.1 64.7





#71

Bromoform

Concen: 50.164 ug/l

RT: 10.799 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:173 Resp: 54055

Ion Ratio Lower Upper

173 100

175 48.3 24.1 72.4

254 0.0 0.1 0.1#



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 1793

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10




Tgt Ion:152 Resp: 95668


Ion Ratio Lower Upper

152 100

115 85.8 43.2 129.6

150 173.3 0.0 346.8









#73
Isopropylbenzene
Concen: 51.780 ug/l
RT: 10.963 min Scan# 1620
Delta R.T. 0.006 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:105 Resp: 364838
Ion Ratio Lower Upper
105 100
120 25.9 13.0 39.0



#74
N-ethyl acetate
Concen: 50.472 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10

Tgt Ion: 43 Resp: 134737
Ion Ratio Lower Upper
43 100
70 42.0 33.4 50.0
55 24.0 19.1 28.7
61 24.0 19.0 28.4









#75
1,1,2,2-Tetrachloroethane
Concen: 50.908 ug/l
RT: 11.207 min Scan# 1660
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 83 Resp: 100156
Ion Ratio Lower Upper
83 100
131 10.5 4.9 14.6
85 64.3 32.0 96.0



#76
1,2,3-Trichloropropane
Concen: 60.421 ug/l
RT: 11.238 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10

Tgt Ion: 75 Resp: 105868
Ion Ratio Lower Upper
75 100
77 32.6 18.5 55.5





#77

Bromobenzene

Concen: 51.813 ug/l

RT: 11.195 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:156 Resp: 87850

Ion Ratio Lower Upper

156 100

77 164.9 82.8 248.6

158 95.3 47.6 142.9



#78

n-propylbenzene

Concen: 51.735 ug/l

RT: 11.305 min Scan# 1676

Delta R.T. 0.006 min

Lab File: VX046775.D




Acq: 19 Jun 2025 17:10


Tgt Ion: 91 Resp: 432963

Ion Ratio Lower Upper

91 100

120 22.5 11.4 34.1









#79
2-Chlorotoluene
Concen: 49.914 ug/l
RT: 11.360 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 91 Resp: 249407
Ion Ratio Lower Upper
91 100
126 33.4 16.6 49.7



#80
1,3,5-Trimethylbenzene
Concen: 51.570 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10

Tgt Ion: 105 Resp: 298594
Ion Ratio Lower Upper
105 100
120 48.0 24.1 72.2









#81
trans-1,4-Dichloro-2-butene
Concen: 47.021 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 75 Resp: 29148
Ion Ratio Lower Upper
75 100
53 108.3 86.8 130.2
89 45.6 36.9 55.3



#82
4-Chlorotoluene
Concen: 50.293 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10

Tgt Ion: 91 Resp: 293281
Ion Ratio Lower Upper
91 100
126 29.0 14.4 43.2









#83
 tert-Butylbenzene
 Concen: 50.825 ug/l
 RT: 11.713 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: VX046775.D
 Acq: 19 Jun 2025 17:10




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion:119 Resp: 304194
 Ion Ratio Lower Upper
 119 100
 91 59.5 29.5 88.6
 134 23.4 11.7 35.0



#84
 1,2,4-Trimethylbenzene
 Concen: 51.927 ug/l
 RT: 11.750 min Scan# 1749
 Delta R.T. -0.000 min
 Lab File: VX046775.D
 Acq: 19 Jun 2025 17:10

Tgt Ion:105 Resp: 302076
 Ion Ratio Lower Upper
 105 100
 120 43.0 22.1 66.3









#85
sec-Butylbenzene
Concen: 51.281 ug/l
RT: 11.890 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:105 Resp: 387895
Ion Ratio Lower Upper
105 100
134 19.8 9.8 29.4



#86
p-Isopropyltoluene
Concen: 51.067 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10

Tgt Ion:119 Resp: 325914
Ion Ratio Lower Upper
119 100
134 26.0 12.9 38.7
91 24.2 12.3 36.9





#87

1,3-Dichlorobenzene

Concen: 49.259 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:146 Resp: 162851

Ion Ratio Lower Upper

146 100

111 40.8 20.6 61.8

148 63.7 31.6 94.8



#88

1,4-Dichlorobenzene

Concen: 47.505 ug/l

RT: 12.036 min Scan# 1796

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10




Tgt Ion:146 Resp: 161313


Ion Ratio Lower Upper

146 100

111 40.7 20.3 61.0

148 64.9 32.1 96.5





#89

n-Butylbenzene

Concen: 50.967 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 91 Resp: 308844

Ion Ratio Lower Upper

91 100

92 54.4 27.1 81.3

134 25.4 12.9 38.6



#90

Hexachloroethane

Concen: 48.908 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX046775.D




Acq: 19 Jun 2025 17:10


Tgt Ion: 117 Resp: 54659

Ion Ratio Lower Upper

117 100

201 70.7 34.4 103.3





#91

1,2-Dichlorobenzene

Concen: 50.145 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.006 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:146 Resp: 154918

Ion Ratio Lower Upper

146 100

111 42.6 21.7 65.1

148 64.1 32.0 96.2



#92

1,2-Dibromo-3-Chloropropane

Concen: 47.738 ug/l

RT: 12.939 min Scan# 1944

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10




Tgt Ion: 75 Resp: 18542


Ion Ratio Lower Upper

75 100

155 84.1 41.9 125.7

157 109.5 51.4 154.3





#93

1,2,4-Trichlorobenzene

Concen: 49.769 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. -0.000 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:180 Resp: 109313

Ion Ratio Lower Upper

180 100

182 95.3 46.7 140.1

145 31.5 15.8 47.4



#94

Hexachlorobutadiene

Concen: 49.603 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. 0.006 min

Lab File: VX046775.D

Acq: 19 Jun 2025 17:10




Tgt Ion:225 Resp: 45830


Ion Ratio Lower Upper

225 100

223 64.3 32.5 97.5

227 63.7 32.2 96.6









#95
Naphthalene
Concen: 50.093 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion:128 Resp: 313462
Ion Ratio Lower Upper
128 100
127 12.7 10.1 15.1
129 10.9 8.7 13.1



#96
1,2,3-Trichlorobenzene
Concen: 50.228 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. -0.000 min
Lab File: VX046775.D
Acq: 19 Jun 2025 17:10

Tgt Ion:180 Resp: 105543
Ion Ratio Lower Upper
180 100
182 95.2 47.8 143.4
145 33.7 17.3 51.9

