

Surrogate Summary

SDG No.: Q2361

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2361-01	TT205S1-20250617	1,2-Dichloroethane-d4	50	47.5	95	81	118
		Dibromofluoromethane	50	49.1	98	80	119
		Toluene-d8	50	50.5	101	89	112
		4-Bromofluorobenzene	50	48.7	97	85	114
VX0619WBL01	VX0619WBL01	1,2-Dichloroethane-d4	50	48.3	97	81	118
		Dibromofluoromethane	50	49.0	98	80	119
		Toluene-d8	50	50.0	100	89	112
		4-Bromofluorobenzene	50	49.4	99	85	114
VX0619WBS01	VX0619WBS01	1,2-Dichloroethane-d4	50	49.8	100	81	118
		Dibromofluoromethane	50	52.0	104	80	119
		Toluene-d8	50	52.3	105	89	112
		4-Bromofluorobenzene	50	54.1	108	85	114
VX0619WBSD0	VX0619WBSD01	1,2-Dichloroethane-d4	50	51.5	103	81	118
		Dibromofluoromethane	50	52.8	106	80	119
		Toluene-d8	50	52.7	105	89	112
		4-Bromofluorobenzene	50	54.9	110	85	114