

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: Q2361 SAS No.: Q2361 SDG NO.: Q2361
Lab File ID: VX046715.D BFB Injection Date: 06/17/2025
Instrument ID: MSVOA_X BFB Injection Time: 08:46
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	18.7
75	30.0 - 60.0% of mass 95	50.7
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.5
173	Less than 2.0% of mass 174	0.5 (0.7) 1
174	50.0 - 100.0% of mass 95	74.8
175	5.0 - 9.0% of mass 174	5.5 (7.4) 1
176	95.0 - 101.0% of mass 174	72 (96.2) 1
177	5.0 - 9.0% of mass 176	4.3 (6) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC005	VSTDIC005	VX046718.D	06/17/2025	11:19
VSTDIC020	VSTDIC020	VX046719.D	06/17/2025	13:59
VSTDIC050	VSTDIC050	VX046720.D	06/17/2025	14:20
VSTDIC100	VSTDIC100	VX046721.D	06/17/2025	14:41
VSTDIC150	VSTDIC150	VX046722.D	06/17/2025	15:02
VSTDIC001	VSTDIC001	VX046725.D	06/17/2025	17:18