

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: Q2361 SAS No.: Q2361 SDG NO.: Q2361
Lab File ID: VX046757.D BFB Injection Date: 06/19/2025
Instrument ID: MSVOA_X BFB Injection Time: 08:42
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	18.8
75	30.0 - 60.0% of mass 95	50.7
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.3
173	Less than 2.0% of mass 174	0.7 (1) 1
174	50.0 - 100.0% of mass 95	72.6
175	5.0 - 9.0% of mass 174	5.7 (7.8) 1
176	95.0 - 101.0% of mass 174	71.1 (97.9) 1
177	5.0 - 9.0% of mass 176	4.5 (6.3) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX046758.D	06/19/2025	09:43
VX0619WBL01	VX0619WBL01	VX046760.D	06/19/2025	10:33
VX0619WBS01	VX0619WBS01	VX046762.D	06/19/2025	11:24
VX0619WBSD01	VX0619WBSD01	VX046763.D	06/19/2025	11:51
TT205S1-20250617	Q2361-01	VX046764.D	06/19/2025	12:12
VSTDCCC050EC	VSTDCCC050	VX046775.D	06/19/2025	17:10