

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental **SDG No.:** Q2363
Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q2363 **SAS No.:** Q2363
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Lead	3950	4000	99	90 - 110	P	06/23/2025	12:34	LB136236

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental **SDG No.:** Q2363
Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q2363 **SAS No.:** Q2363
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Lead	14.0	12.0	117	80 - 120	P	06/23/2025	13:09	LB136236

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental **SDG No.:** Q2363
Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q2363 **SAS No.:** Q2363
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Lead	5160	5000	103	90 - 110	P	06/23/2025	13:56	LB136236
CCV02	Lead	5120	5000	102	90 - 110	P	06/23/2025	14:54	LB136236
CCV03	Lead	5190	5000	104	90 - 110	P	06/23/2025	15:47	LB136236
CCV04	Lead	5230	5000	105	90 - 110	P	06/23/2025	16:51	LB136236
CCV05	Lead	5210	5000	104	90 - 110	P	06/23/2025	17:45	LB136236
CCV06	Lead	5210	5000	104	90 - 110	P	06/23/2025	18:39	LB136236
CCV07	Lead	5240	5000	105	90 - 110	P	06/23/2025	19:24	LB136236