

Initial Calibration Report for SequenceID : FC061825AL

AreaCount

Parameter Range	FC069221.D	FC069222.D	FC069223.D	FC069224.D	FC069225.D	
Aliphatic C9-C12	42176071.000	20640977.000	8581963.000	4632454.000	2366353.000	
Aliphatic C12-C16	29735945.000	14914656.000	6266673.000	3402693.000	1714249.000	
Aliphatic C16-C21	42009240.000	21363842.000	9096606.000	5044763.000	2549195.000	
Aliphatic C21-C28	49153709.000	24681333.000	10474538.000	5764217.000	2970075.000	
Aliphatic C28-C40	51466619.000	25624553.000	10688039.000	5834277.000	3093728.000	
Aliphatic EPH	214541584.000	107225361.000	45107819.000	24678404.000	12693600.000	

AVG Response Factor

Parameter Range	AVG RF	% RSD				
Aliphatic C9-C12	146679.6266662	6.053				
Aliphatic C12-C16	159210.532	6.932				
Aliphatic C16-C21	154434.3226664	9.092				
Aliphatic C21-C28	133966.3675	8.814				
Aliphatic C28-C40	92124.4166662	8.439				
Aliphatic EPH	128354.1601108	7.942				

Concentration

Parameter Range	FC069221.D	FC069222.D	FC069223.D	FC069224.D	FC069225.D	
Aliphatic C9-C12	300.000	150.000	60.000	30.000	15.000	
Aliphatic C12-C16	200.000	100.000	40.000	20.000	10.000	
Aliphatic C16-C21	300.000	150.000	60.000	30.000	15.000	
Aliphatic C21-C28	400.000	200.000	80.000	40.000	20.000	
Aliphatic C28-C40	600.000	300.000	120.000	60.000	30.000	
Aliphatic EPH	1800.000	900.000	360.000	180.000	90.000	

Response Factor

Parameter Range	FC069221.D	FC069222.D	FC069223.D	FC069224.D	FC069225.D	
Aliphatic C9-C12	140586.903333	137606.513333	143032.716666	154415.133333	157756.866666	
Aliphatic C12-C16	148679.725000	149146.560000	156666.825000	170134.650000	171424.900000	
Aliphatic C16-C21	140030.800000	142425.613333	151610.100000	168158.766666	169946.333333	

Initial Calibration Report for SequenceID : FC061825AL

Aliphatic C21-C28	122884.272500	123406.665000	130931.725000	144105.425000	148503.750000	
Aliphatic C28-C40	85777.698333	85415.176666	89066.991666	97237.950000	103124.266666	
Aliphatic EPH	119189.768888	119139.290000	125299.497222	137102.244444	141040.000000	

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC061825AL\
 Data File : FC069221.D
 Signal(s) : FID1A.ch
 Acq On : 18 Jun 2025 11:17
 Operator : YP/AJ
 Sample : 100 PPM ALIPHATIC HC STD1
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :
 100 PPM ALIPHATIC HC STD1

Integration File: autoint1.e
 Quant Time: Jun 18 13:03:07 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 061825.M
 Quant Title : GC Extractables
 QLast Update : Wed Jun 18 13:00:39 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.693	15986001	97.180 ug/ml
Spiked Amount 50.000		Recovery =	194.36%
12) S 1-chlorooctadecane (S...	13.126	11916196	96.429 ug/ml
Spiked Amount 50.000		Recovery =	192.86%
Target Compounds			
1) T n-Nonane (C9)	3.414	13645652	99.307 ug/ml
2) T n-Decane (C10)	4.488	14045970	99.234 ug/ml
3) T A~Naphthalene (C11.7)	6.084	15351850	98.877 ug/ml
4) T n-Dodecane (C12)	6.515	14484449	98.886 ug/ml
5) T A~2-methylnaphthalene...	7.146	15098330	98.803 ug/ml
6) T n-Tetradecane (C14)	8.315	14646111	97.954 ug/ml
7) T n-Hexadecane (C16)	9.922	15089834	96.838 ug/ml
8) T n-Octadecane (C18)	11.369	14862064	96.162 ug/ml
10) T n-Eicosane (C20)	12.680	13798752	95.895 ug/ml
11) T n-Heneicosane (C21)	13.294	13348424	96.021 ug/ml
13) T n-Docosane (C22)	13.881	13057210	96.175 ug/ml
14) T n-Tetracosane (C24)	14.986	12613064	96.664 ug/ml
15) T n-Hexacosane (C26)	16.008	12075116	97.240 ug/ml
16) T n-Octacosane (C28)	16.960	11408319	97.337 ug/ml
17) T n-Tricontane (C30)	17.849	10984628	97.402 ug/ml
18) T n-Dotriacontane (C32)	18.683	10087033	97.565 ug/ml
19) T n-Tetratriacontane (C34)	19.468	8976674	97.988 ug/ml
20) T n-Hexatriacontane (C36)	20.209	7711070	98.564 ug/ml
21) T n-Octatriacontane (C38)	20.980	7003730	98.486 ug/ml
22) T n-Tetracontane (C40)	21.951	6703484	99.440 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC061825AL\
Data File : FC069221.D
Signal(s) : FID1A.ch
Acq On : 18 Jun 2025 11:17
Operator : YP/AJ
Sample : 100 PPM ALIPHATIC HC STD1
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :
100 PPM ALIPHATIC HC STD1

Integration File: autoint1.e
Quant Time: Jun 18 13:03:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 061825.M
Quant Title : GC Extractables
QLast Update : Wed Jun 18 13:00:39 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#1 n-Nonane (C9)


R.T.: 3.414 min
Delta R.T.: 0.000 min
Response: 13645652
Conc: 99.31 ug/ml

Instrument :
FID_C
ClientSampleId :
100 PPM ALIPHATIC HC STD1




#2 n-Decane (C10)

R.T.: 4.488 min
Delta R.T.: 0.000 min
Response: 14045970
Conc: 99.23 ug/ml




#3 A~Naphthalene (C11.7)

R.T.: 6.084 min
Delta R.T.: 0.000 min
Response: 15351850
Conc: 98.88 ug/ml



#4 n-Dodecane (C12)


R.T.: 6.515 min
Delta R.T.: 0.000 min
Response: 14484449
Conc: 98.89 ug/ml



#5 A~2-methylnaphthalene (C12.89)


R.T.: 7.146 min
Delta R.T.: 0.000 min
Response: 15098330
Conc: 98.80 ug/ml

Instrument :
FID_C
ClientSampleId :
100 PPM ALIPHATIC HC STD1




#6 n-Tetradecane (C14)

R.T.: 8.315 min
Delta R.T.: 0.000 min
Response: 14646111
Conc: 97.95 ug/ml




#7 n-Hexadecane (C16)

R.T.: 9.922 min
Delta R.T.: 0.000 min
Response: 15089834
Conc: 96.84 ug/ml



#8 n-Octadecane (C18)


R.T.: 11.369 min
Delta R.T.: 0.000 min
Response: 14862064
Conc: 96.16 ug/ml



#9 ortho-Terphenyl (SURR)


R.T.: 11.693 min
Delta R.T.: 0.000 min
Response: 15986001
Conc: 97.18 ug/ml

Instrument :
FID_C
ClientSampleId :
100 PPM ALIPHATIC HC STD1




#10 n-Eicosane (C20)

R.T.: 12.680 min
Delta R.T.: 0.000 min
Response: 13798752
Conc: 95.90 ug/ml




#11 n-Heneicosane (C21)

R.T.: 13.294 min
Delta R.T.: 0.000 min
Response: 13348424
Conc: 96.02 ug/ml



#12 1-chlorooctadecane (SURR)


R.T.: 13.126 min
Delta R.T.: 0.000 min
Response: 11916196
Conc: 96.43 ug/ml



#13 n-Docosane (C22)


R.T.: 13.881 min
Delta R.T.: 0.000 min
Response: 13057210
Conc: 96.17 ug/ml

Instrument :
FID_C
ClientSampleId :
100 PPM ALIPHATIC HC STD1




#14 n-Tetracosane (C24)

R.T.: 14.986 min
Delta R.T.: 0.000 min
Response: 12613064
Conc: 96.66 ug/ml




#15 n-Hexacosane (C26)

R.T.: 16.008 min
Delta R.T.: 0.000 min
Response: 12075116
Conc: 97.24 ug/ml



#16 n-Octacosane (C28)


R.T.: 16.960 min
Delta R.T.: 0.000 min
Response: 11408319
Conc: 97.34 ug/ml



#17 n-Tricontane (C30)


R.T.: 17.849 min
Delta R.T.: 0.000 min
Response: 10984628
Conc: 97.40 ug/ml

Instrument :
FID_C
ClientSampleId :
100 PPM ALIPHATIC HC STD1




#18 n-Dotriacontane (C32)

R.T.: 18.683 min
Delta R.T.: 0.000 min
Response: 10087033
Conc: 97.56 ug/ml




#19 n-Tetatriacontane (C34)

R.T.: 19.468 min
Delta R.T.: 0.000 min
Response: 8976674
Conc: 97.99 ug/ml



#20 n-Hexatriacontane (C36)

R.T.: 20.209 min
Delta R.T.: 0.000 min
Response: 7711070
Conc: 98.56 ug/ml



#21 n-Octatriacontane (C38)


R.T.: 20.980 min
Delta R.T.: 0.000 min
Response: 7003730
Conc: 98.49 ug/ml

Instrument :

FID_C

ClientSampleId :

100 PPM ALIPHATIC HC STD1



#22 n-Tetracontane (C40)

R.T.: 21.951 min
Delta R.T.: 0.000 min
Response: 6703484
Conc: 99.44 ug/ml

rteres

Area Percent Report

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC061825AL\
Data File : FC069221.D
Signal(s) : FID1A.ch
Acq On : 18 Jun 2025 11:17
Sample : 100 PPM ALIPHATIC HC STD1
Misc :
ALS Vial : 11 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 061825.M
Title : GC Extractables

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.414	3.375	3.470	BB	1589425	13645652	85.36%	5.000%
2	4.488	4.447	4.543	BB	1501566	14045970	87.86%	5.147%
3	6.084	6.038	6.112	BV	1547258	15351850	96.03%	5.626%
4	6.515	6.468	6.558	BB	1360162	14484449	90.61%	5.308%
5	7.146	7.093	7.225	BB	1411463	15098330	94.45%	5.533%
6	8.315	8.248	8.382	BB	1279328	14646111	91.62%	5.367%
7	9.922	9.878	9.975	BB	1299249	15089834	94.39%	5.530%
8	11.369	11.310	11.412	BB	1250407	14862064	92.97%	5.446%
9	11.693	11.613	11.748	BB	1361720	15986001	100.00%	5.858%
10	12.680	12.632	12.725	BB	1105551	13798752	86.32%	5.056%
11	13.126	13.050	13.188	BB	986163	11916196	74.54%	4.367%
12	13.294	13.208	13.338	BB	1091366	13348424	83.50%	4.891%
13	13.881	13.833	13.935	BB	997311	13057210	81.68%	4.785%
14	14.986	14.935	15.037	BB	992019	12613064	78.90%	4.622%
15	16.008	15.847	16.060	BB	889459	12075116	75.54%	4.425%
16	16.960	16.858	17.012	BB	811298	11408319	71.36%	4.180%
17	17.849	17.800	17.895	BB	772421	10984628	68.71%	4.025%
18	18.683	18.608	18.728	BB	706729	10087033	63.10%	3.696%
19	19.468	19.420	19.510	BB	581252	8976674	56.15%	3.289%
20	20.209	20.160	20.270	BB	522339	7711070	48.24%	2.826%
21	20.980	20.897	21.043	BB	389701	7003730	43.81%	2.566%
22	21.951	21.880	22.023	BB	262169	6703484	41.93%	2.456%
Sum of corrected areas:						272893962		

Aliphatic EPH 061825.M Wed Jun 18 14:39:06 2025

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC061825AL\
 Data File : FC069222.D
 Signal(s) : FID1A.ch
 Acq On : 18 Jun 2025 11:58
 Operator : YP/AJ
 Sample : 50 PPM ALIPHATIC HC STD2
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :
 50 PPM ALIPHATIC HC STD2

Integration File: autoint1.e
 Quant Time: Jun 18 13:04:49 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 061825.M
 Quant Title : GC Extractables
 QLast Update : Wed Jun 18 13:00:39 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.686	8035113	49.225 ug/ml
Spiked Amount 50.000		Recovery =	98.45%
12) S 1-chlorooctadecane (S...	13.120	6029490	49.188 ug/ml
Spiked Amount 50.000		Recovery =	98.38%
Target Compounds			
1) T n-Nonane (C9)	3.410	6643725	48.888 ug/ml
2) T n-Decane (C10)	4.483	6866535	48.998 ug/ml
3) T A~Naphthalene (C11.7)	6.079	7502425	48.868 ug/ml
4) T n-Dodecane (C12)	6.510	7130717	49.113 ug/ml
5) T A~2-methylnaphthalene...	7.140	7422117	49.038 ug/ml
6) T n-Tetradecane (C14)	8.310	7295732	49.190 ug/ml
7) T n-Hexadecane (C16)	9.916	7618924	49.257 ug/ml
8) T n-Octadecane (C18)	11.362	7565304	49.295 ug/ml
10) T n-Eicosane (C20)	12.673	7024996	49.208 ug/ml
11) T n-Heneicosane (C21)	13.287	6773542	49.143 ug/ml
13) T n-Docosane (C22)	13.874	6604673	49.090 ug/ml
14) T n-Tetracosane (C24)	14.980	6337022	49.034 ug/ml
15) T n-Hexacosane (C26)	16.002	6031220	49.037 ug/ml
16) T n-Octacosane (C28)	16.954	5708418	49.129 ug/ml
17) T n-Tricontane (C30)	17.843	5495136	49.144 ug/ml
18) T n-Dotriacontane (C32)	18.677	5055272	49.259 ug/ml
19) T n-Tetratriacontane (C34)	19.461	4482554	49.282 ug/ml
20) T n-Hexatriacontane (C36)	20.203	3834019	49.334 ug/ml
21) T n-Octatriacontane (C38)	20.973	3467597	49.167 ug/ml
22) T n-Tetracontane (C40)	21.942	3289975	49.196 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC061825AL\
Data File : FC069222.D
Signal(s) : FID1A.ch
Acq On : 18 Jun 2025 11:58
Operator : YP/AJ
Sample : 50 PPM ALIPHATIC HC STD2
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :
50 PPM ALIPHATIC HC STD2

Integration File: autoint1.e
Quant Time: Jun 18 13:04:49 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 061825.M
Quant Title : GC Extractables
QLast Update : Wed Jun 18 13:00:39 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#1 n-Nonane (C9)


R.T.: 3.410 min
Delta R.T.: 0.000 min
Response: 6643725
Conc: 48.89 ug/ml

Instrument :
FID_C
ClientSampleId :
50 PPM ALIPHATIC HC STD2




#2 n-Decane (C10)

R.T.: 4.483 min
Delta R.T.: 0.000 min
Response: 6866535
Conc: 49.00 ug/ml




#3 A~Naphthalene (C11.7)

R.T.: 6.079 min
Delta R.T.: 0.000 min
Response: 7502425
Conc: 48.87 ug/ml



#4 n-Dodecane (C12)


R.T.: 6.510 min
Delta R.T.: 0.000 min
Response: 7130717
Conc: 49.11 ug/ml



#5 A~2-methylnaphthalene (C12.89)


R.T.: 7.140 min
Delta R.T.: 0.000 min
Response: 7422117
Conc: 49.04 ug/ml

Instrument :
FID_C
ClientSampleId :
50 PPM ALIPHATIC HC STD2




#6 n-Tetradecane (C14)

R.T.: 8.310 min
Delta R.T.: 0.000 min
Response: 7295732
Conc: 49.19 ug/ml




#7 n-Hexadecane (C16)

R.T.: 9.916 min
Delta R.T.: 0.000 min
Response: 7618924
Conc: 49.26 ug/ml



#8 n-Octadecane (C18)


R.T.: 11.362 min
Delta R.T.: 0.000 min
Response: 7565304
Conc: 49.29 ug/ml



#9 ortho-Terphenyl (SURR)


R.T.: 11.686 min
Delta R.T.: 0.000 min
Response: 8035113
Conc: 49.22 ug/ml

Instrument :
FID_C
ClientSampleId :
50 PPM ALIPHATIC HC STD2




#10 n-Eicosane (C20)

R.T.: 12.673 min
Delta R.T.: 0.000 min
Response: 7024996
Conc: 49.21 ug/ml




#11 n-Heneicosane (C21)

R.T.: 13.287 min
Delta R.T.: 0.000 min
Response: 6773542
Conc: 49.14 ug/ml



#12 1-chlorooctadecane (SURR)


R.T.: 13.120 min
Delta R.T.: 0.000 min
Response: 6029490
Conc: 49.19 ug/ml



#13 n-Docosane (C22)


R.T.: 13.874 min
Delta R.T.: 0.000 min
Response: 6604673
Conc: 49.09 ug/ml

Instrument :
FID_C
ClientSampleId :
50 PPM ALIPHATIC HC STD2




#14 n-Tetracosane (C24)

R.T.: 14.980 min
Delta R.T.: 0.000 min
Response: 6337022
Conc: 49.03 ug/ml




#15 n-Hexacosane (C26)

R.T.: 16.002 min
Delta R.T.: 0.000 min
Response: 6031220
Conc: 49.04 ug/ml



#16 n-Octacosane (C28)


R.T.: 16.954 min
Delta R.T.: 0.000 min
Response: 5708418
Conc: 49.13 ug/ml



#17 n-Tricontane (C30)


R.T.: 17.843 min
Delta R.T.: 0.000 min
Response: 5495136
Conc: 49.14 ug/ml

Instrument :
FID_C
ClientSampleId :
50 PPM ALIPHATIC HC STD2




#18 n-Dotriacontane (C32)

R.T.: 18.677 min
Delta R.T.: 0.000 min
Response: 5055272
Conc: 49.26 ug/ml




#19 n-Tetatriacontane (C34)

R.T.: 19.461 min
Delta R.T.: 0.000 min
Response: 4482554
Conc: 49.28 ug/ml



#20 n-Hexatriacontane (C36)


R.T.: 20.203 min
Delta R.T.: 0.000 min
Response: 3834019
Conc: 49.33 ug/ml



#21 n-Octatriacontane (C38)

R.T.: 20.973 min
Delta R.T.: 0.000 min
Response: 3467597
Conc: 49.17 ug/ml

Instrument :
FID_C
ClientSampleId :
50 PPM ALIPHATIC HC STD2



#22 n-Tetracontane (C40)

R.T.: 21.942 min
Delta R.T.: 0.000 min
Response: 3289975
Conc: 49.20 ug/ml

rteres

Area Percent Report

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC061825AL\
Data File : FC069222.D
Signal(s) : FID1A.ch
Acq On : 18 Jun 2025 11:58
Sample : 50 PPM ALIPHATIC HC STD2
Misc :
ALS Vial : 12 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 061825.M
Title : GC Extractables

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.410	3.375	3.450	BB	795544	6643725	82.68%	4.877%
2	4.483	4.442	4.544	BB	766141	6866535	85.46%	5.041%
3	6.079	6.042	6.112	BV	806752	7502425	93.37%	5.508%
4	6.510	6.472	6.540	BB	707214	7130717	88.74%	5.235%
5	7.140	7.089	7.235	BB	725635	7422117	92.37%	5.449%
6	8.310	8.272	8.382	BB	687407	7295732	90.80%	5.356%
7	9.916	9.874	9.984	BB	709129	7618924	94.82%	5.593%
8	11.362	11.310	11.399	BB	636170	7565304	94.15%	5.554%
9	11.686	11.625	11.752	BB	723509	8035113	100.00%	5.899%
10	12.673	12.630	12.724	BB	588442	7024996	87.43%	5.157%
11	13.120	13.079	13.169	BB	499809	6029490	75.04%	4.426%
12	13.287	13.222	13.339	BB	572639	6773542	84.30%	4.973%
13	13.874	13.800	13.930	BB	524181	6604673	82.20%	4.849%
14	14.980	14.907	15.037	BB	498100	6337022	78.87%	4.652%
15	16.002	15.827	16.050	BB	468957	6031220	75.06%	4.428%
16	16.954	16.782	17.004	BB	434550	5708418	71.04%	4.191%
17	17.843	17.800	17.897	BB	398209	5495136	68.39%	4.034%
18	18.677	18.605	18.720	BB	369601	5055272	62.91%	3.711%
19	19.461	19.420	19.502	BB	293442	4482554	55.79%	3.291%
20	20.203	20.160	20.242	BB	267926	3834019	47.72%	2.815%
21	20.973	20.895	21.037	BB	193429	3467597	43.16%	2.546%
22	21.942	21.880	22.019	BB	132645	3289975	40.94%	2.415%
Sum of corrected areas:						136214504		

Aliphatic EPH 061825.M Wed Jun 18 14:38:30 2025

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC061825AL\
 Data File : FC069223.D
 Signal(s) : FID1A.ch
 Acq On : 18 Jun 2025 12:39
 Operator : YP/AJ
 Sample : 20 PPM ALIPHATIC HC STD3
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :
 20 PPM ALIPHATIC HC STD3

Integration File: autoint1.e
 Quant Time: Jun 18 13:01:01 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 061825.M
 Quant Title : GC Extractables
 QLast Update : Wed Jun 18 13:00:39 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.680	3382762	20.000 ug/ml
Spiked Amount 50.000		Recovery =	40.00%
12) S 1-chlorooctadecane (S...	13.116	2559732	20.000 ug/ml
Spiked Amount 50.000		Recovery =	40.00%
Target Compounds			
1) T n-Nonane (C9)	3.410	2767240	20.000 ug/ml
2) T n-Decane (C10)	4.482	2852554	20.000 ug/ml
3) T A~Naphthalene (C11.7)	6.077	3140084	20.000 ug/ml
4) T n-Dodecane (C12)	6.507	2962169	20.000 ug/ml
5) T A~2-methylnaphthalene...	7.136	3092803	20.000 ug/ml
6) T n-Tetradecane (C14)	8.307	3051599	20.000 ug/ml
7) T n-Hexadecane (C16)	9.912	3215074	20.000 ug/ml
8) T n-Octadecane (C18)	11.357	3209683	20.000 ug/ml
10) T n-Eicosane (C20)	12.670	2996001	20.000 ug/ml
11) T n-Heneicosane (C21)	13.283	2890922	20.000 ug/ml
13) T n-Docosane (C22)	13.871	2819173	20.000 ug/ml
14) T n-Tetracosane (C24)	14.976	2696756	20.000 ug/ml
15) T n-Hexacosane (C26)	15.999	2552119	20.000 ug/ml
16) T n-Octacosane (C28)	16.950	2406490	20.000 ug/ml
17) T n-Tricontane (C30)	17.839	2314105	20.000 ug/ml
18) T n-Dotriacontane (C32)	18.674	2118123	20.000 ug/ml
19) T n-Tetratriacontane (C34)	19.459	1869078	20.000 ug/ml
20) T n-Hexatriacontane (C36)	20.200	1587141	20.000 ug/ml
21) T n-Octatriacontane (C38)	20.969	1443800	20.000 ug/ml
22) T n-Tetracontane (C40)	21.938	1355792	20.000 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC061825AL\
Data File : FC069223.D
Signal(s) : FID1A.ch
Acq On : 18 Jun 2025 12:39
Operator : YP/AJ
Sample : 20 PPM ALIPHATIC HC STD3
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :
20 PPM ALIPHATIC HC STD3

Integration File: autoint1.e
Quant Time: Jun 18 13:01:01 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 061825.M
Quant Title : GC Extractables
QLast Update : Wed Jun 18 13:00:39 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#1 n-Nonane (C9)


R.T.: 3.410 min
Delta R.T.: 0.000 min
Response: 2767240
Conc: 20.00 ug/ml

Instrument :
FID_C
ClientSampleId :
20 PPM ALIPHATIC HC STD3




#2 n-Decane (C10)

R.T.: 4.482 min
Delta R.T.: 0.000 min
Response: 2852554
Conc: 20.00 ug/ml




#3 A~Naphthalene (C11.7)

R.T.: 6.077 min
Delta R.T.: 0.000 min
Response: 3140084
Conc: 20.00 ug/ml



#4 n-Dodecane (C12)


R.T.: 6.507 min
Delta R.T.: 0.000 min
Response: 2962169
Conc: 20.00 ug/ml



#5 A~2-methylnaphthalene (C12.89)


R.T.: 7.136 min
Delta R.T.: 0.000 min
Response: 3092803
Conc: 20.00 ug/ml

Instrument :
FID_C
ClientSampleId :
20 PPM ALIPHATIC HC STD3




#6 n-Tetradecane (C14)

R.T.: 8.307 min
Delta R.T.: 0.000 min
Response: 3051599
Conc: 20.00 ug/ml




#7 n-Hexadecane (C16)

R.T.: 9.912 min
Delta R.T.: 0.000 min
Response: 3215074
Conc: 20.00 ug/ml



#8 n-Octadecane (C18)


R.T.: 11.357 min
Delta R.T.: 0.000 min
Response: 3209683
Conc: 20.00 ug/ml



#9 ortho-Terphenyl (SURR)


R.T.: 11.680 min
Delta R.T.: 0.000 min
Response: 3382762
Conc: 20.00 ug/ml

Instrument :
FID_C
ClientSampleId :
20 PPM ALIPHATIC HC STD3




#10 n-Eicosane (C20)

R.T.: 12.670 min
Delta R.T.: 0.000 min
Response: 2996001
Conc: 20.00 ug/ml




#11 n-Heneicosane (C21)

R.T.: 13.283 min
Delta R.T.: 0.000 min
Response: 2890922
Conc: 20.00 ug/ml



#12 1-chlorooctadecane (SURR)


R.T.: 13.116 min
Delta R.T.: 0.000 min
Response: 2559732
Conc: 20.00 ug/ml



#13 n-Docosane (C22)


R.T.: 13.871 min
Delta R.T.: 0.000 min
Response: 2819173
Conc: 20.00 ug/ml

Instrument :
FID_C
ClientSampleId :
20 PPM ALIPHATIC HC STD3




#14 n-Tetracosane (C24)

R.T.: 14.976 min
Delta R.T.: 0.000 min
Response: 2696756
Conc: 20.00 ug/ml




#15 n-Hexacosane (C26)

R.T.: 15.999 min
Delta R.T.: 0.000 min
Response: 2552119
Conc: 20.00 ug/ml



#16 n-Octacosane (C28)


R.T.: 16.950 min
Delta R.T.: 0.000 min
Response: 2406490
Conc: 20.00 ug/ml



#17 n-Tricontane (C30)


R.T.: 17.839 min
Delta R.T.: 0.000 min
Response: 2314105
Conc: 20.00 ug/ml

Instrument :
FID_C
ClientSampleId :
20 PPM ALIPHATIC HC STD3




#18 n-Dotriacontane (C32)

R.T.: 18.674 min
Delta R.T.: 0.000 min
Response: 2118123
Conc: 20.00 ug/ml




#19 n-Tetatriacontane (C34)

R.T.: 19.459 min
Delta R.T.: 0.000 min
Response: 1869078
Conc: 20.00 ug/ml



#20 n-Hexatriacontane (C36)


R.T.: 20.200 min
Delta R.T.: 0.000 min
Response: 1587141
Conc: 20.00 ug/ml



#21 n-Octatriacontane (C38)

R.T.: 20.969 min
Delta R.T.: 0.000 min
Response: 1443800
Conc: 20.00 ug/ml

Instrument :
FID_C
ClientSampleId :
20 PPM ALIPHATIC HC STD3



#22 n-Tetracontane (C40)

R.T.: 21.938 min
Delta R.T.: 0.000 min
Response: 1355792
Conc: 20.00 ug/ml

rteres

Area Percent Report

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC061825AL\
 Data File : FC069223.D
 Signal(s) : FID1A.ch
 Acq On : 18 Jun 2025 12:39
 Sample : 20 PPM ALIPHATIC HC STD3
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 061825.M
 Title : GC Extractables

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.410	3.374	3.450	BB	328194	2767240	81.80%	4.831%
2	4.482	4.452	4.537	BB	315257	2852554	84.33%	4.980%
3	6.077	6.042	6.147	BB	337847	3140084	92.83%	5.482%
4	6.507	6.459	6.540	BB	299443	2962169	87.57%	5.171%
5	7.136	7.089	7.214	BB	315538	3092803	91.43%	5.399%
6	8.307	8.270	8.372	BB	288724	3051599	90.21%	5.327%
7	9.912	9.874	9.980	BB	300265	3215074	95.04%	5.613%
8	11.357	11.310	11.402	BB	287165	3209683	94.88%	5.603%
9	11.680	11.642	11.739	BB	314534	3382762	100.00%	5.905%
10	12.670	12.632	12.714	BB	249206	2996001	88.57%	5.230%
11	13.116	13.079	13.150	BB	218895	2559732	75.67%	4.469%
12	13.283	13.245	13.332	BB	240870	2890922	85.46%	5.047%
13	13.871	13.832	13.915	BB	241488	2819173	83.34%	4.921%
14	14.976	14.925	15.029	BB	224439	2696756	79.72%	4.708%
15	15.999	15.937	16.037	BB	197748	2552119	75.44%	4.455%
16	16.950	16.879	17.005	BB	170629	2406490	71.14%	4.201%
17	17.839	17.800	17.887	BB	169041	2314105	68.41%	4.040%
18	18.674	18.602	18.722	BB	149490	2118123	62.62%	3.698%
19	19.459	19.420	19.497	BB	125143	1869078	55.25%	3.263%
20	20.200	20.160	20.237	BB	112095	1587141	46.92%	2.771%
21	20.969	20.894	21.044	BB	75910	1443800	42.68%	2.520%
22	21.938	21.880	22.022	BB	54065	1355792	40.08%	2.367%
Sum of corrected areas:							57283198	

Aliphatic EPH 061825.M Wed Jun 18 14:38:01 2025

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC061825AL\
 Data File : FC069224.D
 Signal(s) : FID1A.ch
 Acq On : 18 Jun 2025 13:20
 Operator : YP/AJ
 Sample : 10 PPM ALIPHATIC HC STD4
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :
 10 PPM ALIPHATIC HC STD4

Integration File: autoint1.e
 Quant Time: Jun 18 13:45:26 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 061825.M
 Quant Title : GC Extractables
 QLast Update : Wed Jun 18 13:45:19 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.678	1856143	10.994 ug/ml
Spiked Amount 50.000		Recovery =	21.99%
12) S 1-chlorooctadecane (S...	13.115	1417487	11.129 ug/ml
Spiked Amount 50.000		Recovery =	22.26%
Target Compounds			
1) T n-Nonane (C9)	3.409	1496071	10.738 ug/ml
2) T n-Decane (C10)	4.481	1544377	10.746 ug/ml
3) T A~Naphthalene (C11.7)	6.076	1699596	10.782 ug/ml
4) T n-Dodecane (C12)	6.506	1592006	10.707 ug/ml
5) T A~2-methylnaphthalene...	7.135	1665892	10.736 ug/ml
6) T n-Tetradecane (C14)	8.306	1646260	10.803 ug/ml
7) T n-Hexadecane (C16)	9.911	1756433	10.983 ug/ml
8) T n-Octadecane (C18)	11.356	1771170	11.113 ug/ml
10) T n-Eicosane (C20)	12.668	1665008	11.197 ug/ml
11) T n-Heneicosane (C21)	13.281	1608585	11.203 ug/ml
13) T n-Docosane (C22)	13.869	1565876	11.181 ug/ml
14) T n-Tetracosane (C24)	14.974	1491873	11.115 ug/ml
15) T n-Hexacosane (C26)	15.998	1397927	10.991 ug/ml
16) T n-Octacosane (C28)	16.949	1308541	10.917 ug/ml
17) T n-Tricontane (C30)	17.839	1257631	10.907 ug/ml
18) T n-Dotriacontane (C32)	18.672	1158654	10.937 ug/ml
19) T n-Tetratriacontane (C34)	19.459	1018840	10.875 ug/ml
20) T n-Hexatriacontane (C36)	20.200	869747	10.868 ug/ml
21) T n-Octatriacontane (C38)	20.969	782717	10.802 ug/ml
22) T n-Tetracontane (C40)	21.938	746688	10.849 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC061825AL\
Data File : FC069224.D
Signal(s) : FID1A.ch
Acq On : 18 Jun 2025 13:20
Operator : YP/AJ
Sample : 10 PPM ALIPHATIC HC STD4
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :
10 PPM ALIPHATIC HC STD4

Integration File: autoint1.e
Quant Time: Jun 18 13:45:26 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 061825.M
Quant Title : GC Extractables
QLast Update : Wed Jun 18 13:45:19 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#1 n-Nonane (C9)


R.T.: 3.409 min
Delta R.T.: 0.000 min
Response: 1496071
Conc: 10.74 ug/ml

Instrument :
FID_C
ClientSampleId :
10 PPM ALIPHATIC HC STD4




#2 n-Decane (C10)

R.T.: 4.481 min
Delta R.T.: 0.000 min
Response: 1544377
Conc: 10.75 ug/ml




#3 A~Naphthalene (C11.7)

R.T.: 6.076 min
Delta R.T.: 0.000 min
Response: 1699596
Conc: 10.78 ug/ml



#4 n-Dodecane (C12)


R.T.: 6.506 min
Delta R.T.: 0.000 min
Response: 1592006
Conc: 10.71 ug/ml



#5 A~2-methylnaphthalene (C12.89)


R.T.: 7.135 min
Delta R.T.: 0.000 min
Response: 1665892
Conc: 10.74 ug/ml

Instrument :
FID_C
ClientSampleId :
10 PPM ALIPHATIC HC STD4




#6 n-Tetradecane (C14)

R.T.: 8.306 min
Delta R.T.: 0.000 min
Response: 1646260
Conc: 10.80 ug/ml




#7 n-Hexadecane (C16)

R.T.: 9.911 min
Delta R.T.: 0.000 min
Response: 1756433
Conc: 10.98 ug/ml



#8 n-Octadecane (C18)


R.T.: 11.356 min
Delta R.T.: 0.000 min
Response: 1771170
Conc: 11.11 ug/ml



#9 ortho-Terphenyl (SURR)


R.T.: 11.678 min
Delta R.T.: 0.000 min
Response: 1856143
Conc: 10.99 ug/ml

Instrument :
FID_C
ClientSampleId :
10 PPM ALIPHATIC HC STD4




#10 n-Eicosane (C20)

R.T.: 12.668 min
Delta R.T.: 0.000 min
Response: 1665008
Conc: 11.20 ug/ml




#11 n-Heneicosane (C21)

R.T.: 13.281 min
Delta R.T.: 0.000 min
Response: 1608585
Conc: 11.20 ug/ml



#12 1-chlorooctadecane (SURR)


R.T.: 13.115 min
Delta R.T.: 0.000 min
Response: 1417487
Conc: 11.13 ug/ml



#13 n-Docosane (C22)


R.T.: 13.869 min
Delta R.T.: 0.000 min
Response: 1565876
Conc: 11.18 ug/ml

Instrument :
FID_C
ClientSampleId :
10 PPM ALIPHATIC HC STD4




#14 n-Tetracosane (C24)

R.T.: 14.974 min
Delta R.T.: 0.000 min
Response: 1491873
Conc: 11.11 ug/ml




#15 n-Hexacosane (C26)

R.T.: 15.998 min
Delta R.T.: 0.000 min
Response: 1397927
Conc: 10.99 ug/ml



#16 n-Octacosane (C28)


R.T.: 16.949 min
Delta R.T.: 0.000 min
Response: 1308541
Conc: 10.92 ug/ml



#17 n-Tricontane (C30)


R.T.: 17.839 min
Delta R.T.: 0.000 min
Response: 1257631
Conc: 10.91 ug/ml

Instrument :
FID_C
ClientSampleId :
10 PPM ALIPHATIC HC STD4




#18 n-Dotriacontane (C32)

R.T.: 18.672 min
Delta R.T.: 0.000 min
Response: 1158654
Conc: 10.94 ug/ml




#19 n-Tetatriacontane (C34)

R.T.: 19.459 min
Delta R.T.: 0.000 min
Response: 1018840
Conc: 10.87 ug/ml



#20 n-Hexatriacontane (C36)


R.T.: 20.200 min
Delta R.T.: 0.000 min
Response: 869747
Conc: 10.87 ug/ml



#21 n-Octatriacontane (C38)

R.T.: 20.969 min
Delta R.T.: 0.000 min
Response: 782717
Conc: 10.80 ug/ml

Instrument :
FID_C
ClientSampleId :
10 PPM ALIPHATIC HC STD4



#22 n-Tetracontane (C40)

R.T.: 21.938 min
Delta R.T.: 0.000 min
Response: 746688
Conc: 10.85 ug/ml

rteres

Area Percent Report

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC061825AL\
 Data File : FC069224.D
 Signal(s) : FID1A.ch
 Acq On : 18 Jun 2025 13:20
 Sample : 10 PPM ALIPHATIC HC STD4
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 061825.M
 Title : GC Extractables

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.409	3.350	3.454	BB	177619	1496071	80.60%	4.777%
2	4.481	4.422	4.539	BB	169567	1544377	83.20%	4.931%
3	6.076	6.024	6.145	BB	178321	1699596	91.57%	5.427%
4	6.506	6.467	6.542	BB	161845	1592006	85.77%	5.083%
5	7.135	7.074	7.205	BB	166692	1665892	89.75%	5.319%
6	8.306	8.269	8.370	BB	156558	1646260	88.69%	5.257%
7	9.911	9.860	9.974	BB	157792	1756433	94.63%	5.608%
8	11.356	11.310	11.402	BB	153146	1771170	95.42%	5.656%
9	11.678	11.630	11.722	BB	167827	1856143	100.00%	5.927%
10	12.668	12.629	12.705	BB	147990	1665008	89.70%	5.317%
11	13.115	13.080	13.157	BB	122636	1417487	76.37%	4.526%
12	13.281	13.245	13.327	BB	133774	1608585	86.66%	5.136%
13	13.869	13.827	13.917	BB	133414	1565876	84.36%	5.000%
14	14.974	14.927	15.030	BB	118969	1491873	80.37%	4.764%
15	15.998	15.959	16.050	BB	111607	1397927	75.31%	4.464%
16	16.949	16.892	16.995	BB	100098	1308541	70.50%	4.178%
17	17.839	17.800	17.882	BB	95148	1257631	67.76%	4.016%
18	18.672	18.610	18.730	BB	81586	1158654	62.42%	3.700%
19	19.459	19.420	19.500	BB	71711	1018840	54.89%	3.253%
20	20.200	20.160	20.244	BB	62876	869747	46.86%	2.777%
21	20.969	20.904	21.035	BB	44513	782717	42.17%	2.499%
22	21.938	21.880	22.020	BB	30555	746688	40.23%	2.384%
Sum of corrected areas:						31317521		

Aliphatic EPH 061825.M Wed Jun 18 14:37:36 2025

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC061825AL\
 Data File : FC069225.D
 Signal(s) : FID1A.ch
 Acq On : 18 Jun 2025 14:03
 Operator : YP/AJ
 Sample : 5 PPM ALIPHATIC HC STD5
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :
 5 PPM ALIPHATIC HC STD5

Integration File: autoint1.e
 Quant Time: Jun 18 14:23:55 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 061825.M
 Quant Title : GC Extractables
 QLast Update : Wed Jun 18 14:23:47 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.678	942393	5.455 ug/ml
Spiked Amount 50.000		Recovery =	10.91%
12) S 1-chlorooctadecane (S...	13.114	718614	5.501 ug/ml
Spiked Amount 50.000		Recovery =	11.00%
Target Compounds			
1) T n-Nonane (C9)	3.410	768967	5.407 ug/ml
2) T n-Decane (C10)	4.481	785603	5.366 ug/ml
3) T A~Naphthalene (C11.7)	6.076	869708	5.405 ug/ml
4) T n-Dodecane (C12)	6.506	811783	5.361 ug/ml
5) T A~2-methylnaphthalene...	7.135	852964	5.390 ug/ml
6) T n-Tetradecane (C14)	8.306	827220	5.337 ug/ml
7) T n-Hexadecane (C16)	9.911	887029	5.428 ug/ml
8) T n-Octadecane (C18)	11.356	894124	5.476 ug/ml
10) T n-Eicosane (C20)	12.668	842048	5.517 ug/ml
11) T n-Heneicosane (C21)	13.281	813023	5.516 ug/ml
13) T n-Docosane (C22)	13.868	796002	5.532 ug/ml
14) T n-Tetracosane (C24)	14.974	763797	5.538 ug/ml
15) T n-Hexacosane (C26)	15.997	724210	5.540 ug/ml
16) T n-Octacosane (C28)	16.948	686066	5.563 ug/ml
17) T n-Tricontane (C30)	17.838	666002	5.602 ug/ml
18) T n-Dotriacontane (C32)	18.673	614700	5.622 ug/ml
19) T n-Tetratriacontane (C34)	19.458	540381	5.596 ug/ml
20) T n-Hexatriacontane (C36)	20.200	459698	5.578 ug/ml
21) T n-Octatriacontane (C38)	20.969	419007	5.607 ug/ml
22) T n-Tetracontane (C40)	21.937	393940	5.563 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC061825AL\
Data File : FC069225.D
Signal(s) : FID1A.ch
Acq On : 18 Jun 2025 14:03
Operator : YP/AJ
Sample : 5 PPM ALIPHATIC HC STD5
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :
5 PPM ALIPHATIC HC STD5

Integration File: autoint1.e
Quant Time: Jun 18 14:23:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 061825.M
Quant Title : GC Extractables
QLast Update : Wed Jun 18 14:23:47 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#1 n-Nonane (C9)


R.T.: 3.410 min
Delta R.T.: 0.000 min
Response: 768967
Conc: 5.41 ug/ml

Instrument :
FID_C
ClientSampleId :
5 PPM ALIPHATIC HC STD5




#2 n-Decane (C10)

R.T.: 4.481 min
Delta R.T.: 0.000 min
Response: 785603
Conc: 5.37 ug/ml




#3 A~Naphthalene (C11.7)

R.T.: 6.076 min
Delta R.T.: 0.000 min
Response: 869708
Conc: 5.41 ug/ml



#4 n-Dodecane (C12)


R.T.: 6.506 min
Delta R.T.: 0.000 min
Response: 811783
Conc: 5.36 ug/ml



#5 A~2-methylnaphthalene (C12.89)


R.T.: 7.135 min
Delta R.T.: 0.000 min
Response: 852964
Conc: 5.39 ug/ml

Instrument :
FID_C
ClientSampleId :
5 PPM ALIPHATIC HC STD5




#6 n-Tetradecane (C14)

R.T.: 8.306 min
Delta R.T.: 0.000 min
Response: 827220
Conc: 5.34 ug/ml




#7 n-Hexadecane (C16)

R.T.: 9.911 min
Delta R.T.: 0.000 min
Response: 887029
Conc: 5.43 ug/ml



#8 n-Octadecane (C18)


R.T.: 11.356 min
Delta R.T.: 0.000 min
Response: 894124
Conc: 5.48 ug/ml



#9 ortho-Terphenyl (SURR)


R.T.: 11.678 min
Delta R.T.: 0.000 min
Response: 942393
Conc: 5.45 ug/ml

Instrument :
FID_C
ClientSampleId :
5 PPM ALIPHATIC HC STD5




#10 n-Eicosane (C20)

R.T.: 12.668 min
Delta R.T.: 0.000 min
Response: 842048
Conc: 5.52 ug/ml




#11 n-Heneicosane (C21)

R.T.: 13.281 min
Delta R.T.: 0.000 min
Response: 813023
Conc: 5.52 ug/ml



#12 1-chlorooctadecane (SURR)


R.T.: 13.114 min
Delta R.T.: 0.000 min
Response: 718614
Conc: 5.50 ug/ml



#13 n-Docosane (C22)


R.T.: 13.868 min
Delta R.T.: 0.000 min
Response: 796002
Conc: 5.53 ug/ml

Instrument :
FID_C
ClientSampleId :
5 PPM ALIPHATIC HC STD5




#14 n-Tetracosane (C24)

R.T.: 14.974 min
Delta R.T.: 0.000 min
Response: 763797
Conc: 5.54 ug/ml




#15 n-Hexacosane (C26)

R.T.: 15.997 min
Delta R.T.: 0.000 min
Response: 724210
Conc: 5.54 ug/ml



#16 n-Octacosane (C28)


R.T.: 16.948 min
Delta R.T.: 0.000 min
Response: 686066
Conc: 5.56 ug/ml



#17 n-Tricontane (C30)


R.T.: 17.838 min
Delta R.T.: 0.000 min
Response: 666002
Conc: 5.60 ug/ml

Instrument :
FID_C
ClientSampleId :
5 PPM ALIPHATIC HC STD5




#18 n-Dotriacontane (C32)

R.T.: 18.673 min
Delta R.T.: 0.000 min
Response: 614700
Conc: 5.62 ug/ml




#19 n-Tetratriacontane (C34)

R.T.: 19.458 min
Delta R.T.: 0.000 min
Response: 540381
Conc: 5.60 ug/ml



#20 n-Hexatriacontane (C36)


R.T.: 20.200 min
Delta R.T.: 0.000 min
Response: 459698
Conc: 5.58 ug/ml



#21 n-Octatriacontane (C38)

R.T.: 20.969 min
Delta R.T.: 0.000 min
Response: 419007
Conc: 5.61 ug/ml

Instrument :
FID_C
ClientSampleId :
5 PPM ALIPHATIC HC STD5



#22 n-Tetracontane (C40)

R.T.: 21.937 min
Delta R.T.: 0.000 min
Response: 393940
Conc: 5.56 ug/ml

rteres

Area Percent Report

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC061825AL\
 Data File : FC069225.D
 Signal(s) : FID1A.ch
 Acq On : 18 Jun 2025 14:03
 Sample : 5 PPM ALIPHATIC HC STD5
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 061825.M
 Title : GC Extractables

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.410	3.380	3.449	BB	91878	768967	81.60%	4.783%
2	4.481	4.437	4.527	BB	87247	785603	83.36%	4.886%
3	6.076	6.047	6.139	BB	92272	869708	92.29%	5.410%
4	6.506	6.455	6.549	BB	81230	811783	86.14%	5.049%
5	7.135	7.100	7.204	BB	85753	852964	90.51%	5.305%
6	8.306	8.274	8.350	BB	77049	827220	87.78%	5.145%
7	9.911	9.867	9.960	BB	80293	887029	94.13%	5.517%
8	11.356	11.320	11.394	BB	77410	894124	94.88%	5.561%
9	11.678	11.640	11.729	BB	84784	942393	100.00%	5.862%
10	12.668	12.632	12.709	BB	73078	842048	89.35%	5.238%
11	13.114	13.082	13.157	BB	59088	718614	76.25%	4.470%
12	13.281	13.245	13.315	BB	70672	813023	86.27%	5.057%
13	13.868	13.829	13.905	BB	64412	796002	84.47%	4.951%
14	14.974	14.934	15.010	BB	59988	763797	81.05%	4.751%
15	15.997	15.952	16.049	BB	54380	724210	76.85%	4.505%
16	16.948	16.885	16.997	BB	50496	686066	72.80%	4.267%
17	17.838	17.800	17.884	BB	48943	666002	70.67%	4.143%
18	18.673	18.615	18.715	BB	41575	614700	65.23%	3.823%
19	19.458	19.420	19.502	BB	36225	540381	57.34%	3.361%
20	20.200	20.160	20.237	BB	32066	459698	48.78%	2.859%
21	20.969	20.902	21.032	BB	21311	419007	44.46%	2.606%
22	21.937	21.880	22.024	BB	15337	393940	41.80%	2.450%
Sum of corrected areas:							16077276	

Aliphatic EPH 061825.M Wed Jun 18 14:37:19 2025

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC061825AL\
 Data File : FC069226.D
 Signal(s) : FID1A.ch
 Acq On : 18 Jun 2025 14:45
 Operator : YP/AJ
 Sample : 20 PPM ALIPHATIC HC STD ICV
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :
 20 PPM ALIPHATIC HC STD ICV

Integration File: autoint1.e
 Quant Time: Jun 18 15:23:51 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 061825.M
 Quant Title : GC Extractables
 QLast Update : Wed Jun 18 14:24:27 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.680	3253610	18.833 ug/ml
Spiked Amount 50.000		Recovery =	37.67%
12) S 1-chlorooctadecane (S...	13.116	2474788	18.943 ug/ml
Spiked Amount 50.000		Recovery =	37.89%
Target Compounds			
1) T n-Nonane (C9)	3.410	2571511	18.081 ug/ml
2) T n-Decane (C10)	4.482	2650200	18.103 ug/ml
3) T A~Naphthalene (C11.7)	6.077	2926181	18.187 ug/ml
4) T n-Dodecane (C12)	6.507	2762531	18.244 ug/ml
5) T A~2-methylnaphthalene...	7.136	2898798	18.318 ug/ml
6) T n-Tetradecane (C14)	8.307	2886495	18.622 ug/ml
7) T n-Hexadecane (C16)	9.912	3074968	18.817 ug/ml
8) T n-Octadecane (C18)	11.357	3098893	18.980 ug/ml
10) T n-Eicosane (C20)	12.670	2897467	18.982 ug/ml
11) T n-Heneicosane (C21)	13.283	2797161	18.978 ug/ml
13) T n-Docosane (C22)	13.871	2718940	18.897 ug/ml
14) T n-Tetracosane (C24)	14.975	2594124	18.807 ug/ml
15) T n-Hexacosane (C26)	15.998	2449139	18.735 ug/ml
16) T n-Octacosane (C28)	16.949	2300993	18.657 ug/ml
17) T n-Tricontane (C30)	17.839	2215902	18.639 ug/ml
18) T n-Dotriacontane (C32)	18.675	2037377	18.634 ug/ml
19) T n-Tetratriacontane (C34)	19.460	1803580	18.677 ug/ml
20) T n-Hexatriacontane (C36)	20.201	1557268	18.896 ug/ml
21) T n-Octatriacontane (C38)	20.971	1411730	18.891 ug/ml
22) T n-Tetracontane (C40)	21.939	1331528	18.803 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC061825AL\
Data File : FC069226.D
Signal(s) : FID1A.ch
Acq On : 18 Jun 2025 14:45
Operator : YP/AJ
Sample : 20 PPM ALIPHATIC HC STD ICV
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :
20 PPM ALIPHATIC HC STD ICV

Integration File: autoint1.e
Quant Time: Jun 18 15:23:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 061825.M
Quant Title : GC Extractables
QLast Update : Wed Jun 18 14:24:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um





#1 n-Nonane (C9)


R.T.: 3.410 min
Delta R.T.: 0.000 min
Response: 2571511
Conc: 18.08 ug/ml

Instrument :

FID_C


ClientSampleId :

20 PPM ALIPHATIC HC STD ICV




#2 n-Decane (C10)

R.T.: 4.482 min
Delta R.T.: 0.000 min
Response: 2650200
Conc: 18.10 ug/ml




#3 A~Naphthalene (C11.7)

R.T.: 6.077 min
Delta R.T.: 0.000 min
Response: 2926181
Conc: 18.19 ug/ml



#4 n-Dodecane (C12)


R.T.: 6.507 min
Delta R.T.: 0.000 min
Response: 2762531
Conc: 18.24 ug/ml



#5 A~2-methylnaphthalene (C12.89)


R.T.: 7.136 min
Delta R.T.: 0.000 min
Response: 2898798
Conc: 18.32 ug/ml

Instrument :
FID_C
ClientSampleId :
20 PPM ALIPHATIC HC STD ICV




#6 n-Tetradecane (C14)

R.T.: 8.307 min
Delta R.T.: 0.000 min
Response: 2886495
Conc: 18.62 ug/ml




#7 n-Hexadecane (C16)

R.T.: 9.912 min
Delta R.T.: 0.002 min
Response: 3074968
Conc: 18.82 ug/ml



#8 n-Octadecane (C18)


R.T.: 11.357 min
Delta R.T.: 0.001 min
Response: 3098893
Conc: 18.98 ug/ml



#9 ortho-Terphenyl (SURR)


R.T.: 11.680 min
Delta R.T.: 0.003 min
Response: 3253610
Conc: 18.83 ug/ml

Instrument :
FID_C
ClientSampleId :
20 PPM ALIPHATIC HC STD ICV




#10 n-Eicosane (C20)

R.T.: 12.670 min
Delta R.T.: 0.002 min
Response: 2897467
Conc: 18.98 ug/ml




#11 n-Heneicosane (C21)

R.T.: 13.283 min
Delta R.T.: 0.002 min
Response: 2797161
Conc: 18.98 ug/ml



#12 1-chlorooctadecane (SURR)

R.T.: 13.116 min
Delta R.T.: 0.002 min
Response: 2474788
Conc: 18.94 ug/ml



#13 n-Docosane (C22)


R.T.: 13.871 min
Delta R.T.: 0.002 min
Response: 2718940
Conc: 18.90 ug/ml

Instrument :

FID_C


ClientSampleId :

20 PPM ALIPHATIC HC STD ICV




#14 n-Tetracosane (C24)

R.T.: 14.975 min
Delta R.T.: 0.001 min
Response: 2594124
Conc: 18.81 ug/ml




#15 n-Hexacosane (C26)

R.T.: 15.998 min
Delta R.T.: 0.001 min
Response: 2449139
Conc: 18.74 ug/ml



#16 n-Octacosane (C28)

R.T.: 16.949 min
Delta R.T.: 0.000 min
Response: 2300993
Conc: 18.66 ug/ml



#17 n-Tricontane (C30)


R.T.: 17.839 min
Delta R.T.: 0.001 min
Response: 2215902
Conc: 18.64 ug/ml

Instrument :

FID_C


ClientSampleId :

20 PPM ALIPHATIC HC STD ICV




#18 n-Dotriacontane (C32)

R.T.: 18.675 min
Delta R.T.: 0.002 min
Response: 2037377
Conc: 18.63 ug/ml




#19 n-Tetratriacontane (C34)

R.T.: 19.460 min
Delta R.T.: 0.002 min
Response: 1803580
Conc: 18.68 ug/ml



#20 n-Hexatriacontane (C36)


R.T.: 20.201 min
Delta R.T.: 0.001 min
Response: 1557268
Conc: 18.90 ug/ml



#21 n-Octatriacontane (C38)

R.T.: 20.971 min
Delta R.T.: 0.002 min
Response: 1411730
Conc: 18.89 ug/ml

Instrument :
FID_C
ClientSampleId :
20 PPM ALIPHATIC HC STD ICV



#22 n-Tetracontane (C40)

R.T.: 21.939 min
Delta R.T.: 0.001 min
Response: 1331528
Conc: 18.80 ug/ml

rteres

Area Percent Report

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC061825AL\
 Data File : FC069226.D
 Signal(s) : FID1A.ch
 Acq On : 18 Jun 2025 14:45
 Sample : 20 PPM ALIPHATIC HC STD ICV
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 061825.M
 Title : GC Extractables

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.410	3.357	3.462	BB	310610	2571511	79.04%	4.700%
2	4.482	4.425	4.527	BB	299540	2650200	81.45%	4.844%
3	6.077	6.015	6.145	BB	314091	2926181	89.94%	5.348%
4	6.507	6.453	6.542	BB	280576	2762531	84.91%	5.049%
5	7.136	7.088	7.225	BB	293786	2898798	89.09%	5.298%
6	8.307	8.245	8.378	BB	273480	2886495	88.72%	5.276%
7	9.912	9.872	9.977	BB	276956	3074968	94.51%	5.620%
8	11.357	11.310	11.403	BB	270716	3098893	95.24%	5.664%
9	11.680	11.625	11.738	BB	303465	3253610	100.00%	5.947%
10	12.670	12.630	12.708	BB	242161	2897467	89.05%	5.296%
11	13.116	13.078	13.160	BB	213636	2474788	76.06%	4.523%
12	13.283	13.243	13.335	BB	224409	2797161	85.97%	5.112%
13	13.871	13.828	13.915	BB	223018	2718940	83.57%	4.969%
14	14.975	14.925	15.015	BB	206118	2594124	79.73%	4.741%
15	15.998	15.947	16.048	BB	201397	2449139	75.27%	4.476%
16	16.949	16.883	17.007	BB	168971	2300993	70.72%	4.206%
17	17.839	17.800	17.897	BB	161622	2215902	68.11%	4.050%
18	18.675	18.612	18.727	BB	152633	2037377	62.62%	3.724%
19	19.460	19.420	19.502	BB	128446	1803580	55.43%	3.296%
20	20.201	20.160	20.260	BB	111015	1557268	47.86%	2.846%
21	20.971	20.893	21.050	BB	75755	1411730	43.39%	2.580%
22	21.939	21.880	22.010	BB	52545	1331528	40.92%	2.434%
Sum of corrected areas:							54713182	

Aliphatic EPH 061825.M Wed Jun 18 15:24:52 2025