

Continuing Calibration Report for SequenceID : FC062425AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC069266.D

Aliphatic C9-C12	8878943.000	60.000	3.308	6.606	147982.383	146679.627	-0.888
Aliphatic C12-C16	6370959.000	40.000	6.607	10.010	159273.975	159210.532	-0.040
Aliphatic C16-C21	8920053.000	60.000	10.011	13.381	148667.550	154434.323	3.734
Aliphatic C21-C28	10569984.000	80.000	13.382	17.047	132124.800	133966.368	1.375
Aliphatic C28-C40	11710804.000	120.000	17.048	22.029	97590.033	92124.417	-5.933
Aliphatic EPH	46450743.000	360.000	3.308	22.029	129029.842	128354.160	-0.526