

Continuing Calibration Report for SequenceID : FC062425AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC069274.D

Aliphatic C9-C12	8882441.000	60.000	3.308	6.606	148040.683	146679.627	-0.928
Aliphatic C12-C16	6279167.000	40.000	6.607	10.010	156979.175	159210.532	1.402
Aliphatic C16-C21	8783859.000	60.000	10.011	13.381	146397.650	154434.323	5.204
Aliphatic C21-C28	10450768.000	80.000	13.382	17.047	130634.600	133966.368	2.487
Aliphatic C28-C40	11618177.000	120.000	17.048	22.029	96818.142	92124.417	-5.095
Aliphatic EPH	46014412.000	360.000	3.308	22.029	127817.811	128354.160	0.418