

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN060625

Review By	John Carlone	Review On	6/9/2025 8:08:23 AM
Supervise By	Mahesh Dadoda	Supervise On	6/9/2025 1:13:51 PM
SubDirectory	VN060625	HP Acquire Method	HP Processing Method 82N060625W.M

STD. NAME	STD REF.#
Tune/Reschk	VP134155
Initial Calibration Stds	VP134242,VP134243,VP134244,VP134245,VP134246,VP134247
CCC	VP134156
Internal Standard/PEM	
ICV/I.BLK	VP134248
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN086861.D	06 Jun 2025 07:59		JC\MD	Ok
2	VSTDIC001	VSTDIC001	VN086862.D	06 Jun 2025 12:44	Method failed for com.#13	JC\MD	Ok,M
3	VSTDIC005	VSTDIC005	VN086863.D	06 Jun 2025 13:17		JC\MD	Ok,M
4	VSTDIC020	VSTDIC020	VN086864.D	06 Jun 2025 13:40		JC\MD	Ok,M
5	VSTDIC050	VSTDIC050	VN086865.D	06 Jun 2025 14:03		JC\MD	Ok,M
6	VSTDIC100	VSTDIC100	VN086866.D	06 Jun 2025 14:26		JC\MD	Ok,M
7	VSTDIC150	VSTDIC150	VN086867.D	06 Jun 2025 14:49		JC\MD	Ok,M
8	IBLK	IBLK	VN086868.D	06 Jun 2025 15:12		JC\MD	Ok
9	VSTDICV050	ICVVN060625	VN086869.D	06 Jun 2025 15:54		JC\MD	Ok,M
10	VN0606WBL01	VN0606WBL01	VN086870.D	06 Jun 2025 16:47		JC\MD	Ok
11	VN0606WBL02	VN0606WBL02	VN086871.D	06 Jun 2025 17:10		JC\MD	Ok
12	VN0606WBS01	VN0606WBS01	VN086872.D	06 Jun 2025 17:33		JC\MD	Ok,M
13	VN0606WBSD01	VN0606WBSD01	VN086873.D	06 Jun 2025 17:56		JC\MD	Ok,M
14	Q2254-01	BP-VPB-182-GW-810-8	VN086874.D	06 Jun 2025 18:19	vial A pH<2 endccc out of tune	JC\MD	Not Ok
15	Q2237-02	TW-WTS-10	VN086875.D	06 Jun 2025 18:42	vial A pH<2	JC\MD	Ok
16	Q2216-02	3887	VN086876.D	06 Jun 2025 19:05	vial A pH<2	JC\MD	Ok
17	Q2216-03	3888	VN086877.D	06 Jun 2025 19:28	vial A pH<2	JC\MD	Ok
18	Q2216-04	3864	VN086878.D	06 Jun 2025 19:51	vial A pH<2	JC\MD	Ok

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN060625

Review By	John Carlone	Review On	6/9/2025 8:08:23 AM
Supervise By	Maresh Dadoda	Supervise On	6/9/2025 1:13:51 PM
SubDirectory	VN060625	HP Acquire Method	HP Processing Method 82N060625W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP134155		
Initial Calibration Stds	VP134242,VP134243,VP134244,VP134245,VP134246,VP134247		
CCC	VP134156		
Internal Standard/PEM			
ICV/I.BLK	VP134248		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q2216-05	3865	VN086879.D	06 Jun 2025 20:13	vial A pH<2 Out of Tune	JC\MD	Not Ok
20	Q2216-06	3851	VN086880.D	06 Jun 2025 20:36	vial A pH<2 Out of Tune	JC\MD	Not Ok
21	Q2206-04	TP-1	VN086881.D	06 Jun 2025 20:59	vial A pH<2 Out of Tune	JC\MD	Not Ok
22	Q2242-04	TP09-MHJ	VN086882.D	06 Jun 2025 21:21	vial A pH<2 Out of Tune	JC\MD	Not Ok
23	Q2192-01	SB-1	VN086883.D	06 Jun 2025 21:44	vial A pH<2 Out of Tune	JC\MD	Not Ok
24	Q2198-02	B-202-SB02	VN086884.D	06 Jun 2025 22:07	vial A pH<2 Out of Tune	JC\MD	Not Ok
25	Q2198-04	B-207-SB02	VN086885.D	06 Jun 2025 22:29	vial A pH<2 Out of Tune	JC\MD	Not Ok
26	VSTDCCC050	VSTDCCC050EC	VN086886.D	06 Jun 2025 22:52	Out of Tune	JC\MD	Not Ok

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN062025

Review By	John Carlone	Review On	6/23/2025 9:08:58 AM
Supervise By	Maresh Dadoda	Supervise On	6/23/2025 1:18:33 PM
SubDirectory	VN062025	HP Acquire Method	HP Processing Method 82N060625W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134448
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134449,VP134450

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN087121.D	20 Jun 2025 08:18		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN087122.D	20 Jun 2025 08:51	pH#Lot#V12668	JC\MD	Ok,M
3	VN0620MBL01	VN0620MBL01	VN087123.D	20 Jun 2025 09:25		JC\MD	Ok
4	VN0620WBL01	VN0620WBL01	VN087124.D	20 Jun 2025 09:46		JC\MD	Ok
5	VN0620WBS01	VN0620WBS01	VN087125.D	20 Jun 2025 10:06		JC\MD	Ok,M
6	VN0620WBSD01	VN0620WBSD01	VN087126.D	20 Jun 2025 10:40		JC\MD	Ok,M
7	Q2363-02DL	GCAP1WDL	VN087127.D	20 Jun 2025 11:01	vial B pH<2	JC\MD	Ok
8	PB168545TB	PB168545TB	VN087128.D	20 Jun 2025 11:21		JC\MD	Ok
9	Q2354-02	3321	VN087129.D	20 Jun 2025 11:42	vial A pH#5.0	JC\MD	Ok
10	Q2355-02	3897	VN087130.D	20 Jun 2025 12:03	vial A pH#5.0	JC\MD	Ok
11	Q2362-04	TP-11	VN087131.D	20 Jun 2025 12:23	vial A pH#5.0	JC\MD	Ok
12	Q2362-08	EP-6	VN087132.D	20 Jun 2025 12:44	vial A pH#5.0	JC\MD	Ok
13	Q2367-04	EP-4	VN087133.D	20 Jun 2025 13:04	vial A pH#5.0	JC\MD	Ok
14	Q2367-08	EP-5	VN087134.D	20 Jun 2025 13:25	vial A pH#5.0	JC\MD	Ok
15	IBLK	IBLK	VN087135.D	20 Jun 2025 13:46		JC\MD	Ok
16	Q2377-03	TB01-061925	VN087136.D	20 Jun 2025 14:06	524 Method on login	JC\MD	Not Ok
17	Q2377-01	PW-B6-L66-061925	VN087137.D	20 Jun 2025 14:27	524 Method on login	JC\MD	Not Ok

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN062025

Review By	John Carlone	Review On	6/23/2025 9:08:58 AM
Supervise By	Maresh Dadoda	Supervise On	6/23/2025 1:18:33 PM
SubDirectory	VN062025	HP Acquire Method	HP Processing Method 82N060625W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134448		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134449,VP134450		

18	VN0620MBS01	VN0620MBS01	VN087138.D	20 Jun 2025 14:48		JC\MD	Ok,M
19	VN0620MBSD01	VN0620MBSD01	VN087139.D	20 Jun 2025 15:08		JC\MD	Ok,M
20	Q2363-01ME	GCAP1ME	VN087140.D	20 Jun 2025 15:29	Surrogate fail	JC\MD	Ok
21	Q2355-01	3897	VN087141.D	20 Jun 2025 15:50	BS fail Low ;BSD fail Low; cloth material	JC\MD	Not Ok
22	Q2354-01	3321	VN087142.D	20 Jun 2025 16:11	BS fail Low;BSD fail Low;Run @ Lower Dilution	JC\MD	Not Ok
23	Q2355-03	20071	VN087143.D	20 Jun 2025 16:31	BS fail Low ;BSD fail Low;Need Soil Run	JC\MD	Not Ok
24	IBLK	IBLK	VN087144.D	20 Jun 2025 16:52		JC\MD	Ok
25	IBLK	IBLK	VN087145.D	20 Jun 2025 17:13		JC\MD	Ok
26	Q2377-01	PW-B6-L66-061925	VN087146.D	20 Jun 2025 17:34	524 Method on login	JC\MD	Not Ok
27	VSTDCCC050	VSTDCCC050EC	VN087147.D	20 Jun 2025 17:54		JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX061725

Review By	Semsettin Yesilyurt	Review On	6/18/2025 5:28:16 PM
Supervise By	Maresh Dadoda	Supervise On	6/18/2025 6:22:48 PM
SubDirectory	VX061725	HP Acquire Method	HP Processing Method 82X061725W.M

STD. NAME	STD REF.#
Tune/Reschk	VP134389
Initial Calibration Stds	VP134390,VP134391,VP134392,VP134393,VP134394,VP134395
CCC	
Internal Standard/PEM	
ICV/I.BLK	VP134396
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX046715.D	17 Jun 2025 08:46		JC/MD	Ok
2	VSTDIC001	VSTDIC001	VX046716.D	17 Jun 2025 10:09	RRF check	JC/MD	Not Ok
3	VSTDIC001	VSTDIC001	VX046717.D	17 Jun 2025 10:58	low response	JC/MD	ReRun
4	VSTDIC005	VSTDIC005	VX046718.D	17 Jun 2025 11:19		JC/MD	Ok,M
5	VSTDIC020	VSTDIC020	VX046719.D	17 Jun 2025 13:59	LR- 10,49	JC/MD	Ok,M
6	VSTDIC050	VSTDIC050	VX046720.D	17 Jun 2025 14:20		JC/MD	Ok,M
7	VSTDIC100	VSTDIC100	VX046721.D	17 Jun 2025 14:41		JC/MD	Ok,M
8	VSTDIC150	VSTDIC150	VX046722.D	17 Jun 2025 15:02		JC/MD	Ok,M
9	IBLK	IBLK	VX046723.D	17 Jun 2025 15:23		JC/MD	Ok
10	VSTDIC001	VSTDIC001	VX046724.D	17 Jun 2025 16:20	spike error	JC/MD	Not Ok
11	VSTDIC001	VSTDIC001	VX046725.D	17 Jun 2025 17:18		JC/MD	Ok,M
12	VSTDICV050	ICVVX061725	VX046726.D	17 Jun 2025 18:00		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX061925

Review By	John Carlone	Review On	6/20/2025 8:46:07 AM
Supervise By	Maresh Dadoda	Supervise On	6/21/2025 12:21:16 AM
SubDirectory	VX061925	HP Acquire Method	HP Processing Method 82X061725W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134428
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134429,VP134430

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX046757.D	19 Jun 2025 08:42		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX046758.D	19 Jun 2025 09:43	pH#Lot#V12668	JC/MD	Ok,M
3	VX0619MBL01	VX0619MBL01	VX046759.D	19 Jun 2025 10:12		JC/MD	Ok
4	VX0619WBL01	VX0619WBL01	VX046760.D	19 Jun 2025 10:33		JC/MD	Ok
5	Q2345-01	EB02-20250616	VX046761.D	19 Jun 2025 11:03	vial B pH<2 EB	JC/MD	Ok
6	VX0619WBS01	VX0619WBS01	VX046762.D	19 Jun 2025 11:24		JC/MD	Ok,M
7	VX0619WBSD01	VX0619WBSD01	VX046763.D	19 Jun 2025 11:51		JC/MD	Ok,M
8	Q2361-01	TT205S1-20250617	VX046764.D	19 Jun 2025 12:12	vial A pH<2	JC/MD	Ok
9	Q2334-01	MW3	VX046765.D	19 Jun 2025 12:33	vial A pH<2	JC/MD	Ok
10	Q2334-02	MW4	VX046766.D	19 Jun 2025 12:54	vial A pH<2	JC/MD	Ok
11	Q2334-03	GBTW1	VX046767.D	19 Jun 2025 13:15	vial A pH<2	JC/MD	Ok
12	Q2363-02	GCAP1W	VX046768.D	19 Jun 2025 13:36	vial A pH<2 Need 20X	JC/MD	Dilution
13	IBLK	IBLK	VX046769.D	19 Jun 2025 13:57		JC/MD	Ok
14	IBLK	IBLK	VX046770.D	19 Jun 2025 14:18		JC/MD	Ok
15	Q2351-01	FAIRLAWN-SUMP	VX046771.D	19 Jun 2025 15:45	vial A pH<2 oily sample	JC/MD	Ok
16	Q2372-02	GAV2W	VX046772.D	19 Jun 2025 16:06	vial A pH<2	JC/MD	Ok
17	Q2372-01	GAV1W	VX046773.D	19 Jun 2025 16:27	vial A pH<2	JC/MD	Ok
18	IBLK	IBLK	VX046774.D	19 Jun 2025 16:49		JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX061925

Review By	John Carlone	Review On	6/20/2025 8:46:07 AM
Supervise By	Maresh Dadoda	Supervise On	6/21/2025 12:21:16 AM
SubDirectory	VX061925	HP Acquire Method	HP Processing Method 82X061725W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134428
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134429,VP134430

19	VSTDCCC050	VSTDCCC050EC	VX046775.D	19 Jun 2025 17:10		JC/MD	Ok,M
----	------------	--------------	------------	-------------------	--	-------	------

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY062325

Review By	Mahesh Dadoda	Review On	6/24/2025 8:24:26 AM		
Supervise By	Semsettin Yesilyurt	Supervise On	6/24/2025 8:29:32 AM		
SubDirectory	VY062325	HP Acquire Method	MSVOA_Y	HP Processing Method	82y062325s.m

STD. NAME	STD REF.#
Tune/Reschk	VP134461
Initial Calibration Stds	VP134462,VP134463,VP134464,VP134465,VP134466,VP134467
CCC	
Internal Standard/PEM	VP133934
ICV/I.BLK	VP134468
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022775.D	23 Jun 2025 10:17		SY/MD	Ok
2	VSTDIC005	VSTDIC005	VY022776.D	23 Jun 2025 13:38	LR- 58	SY/MD	Ok,M
3	VSTDIC010	VSTDIC010	VY022777.D	23 Jun 2025 14:00		SY/MD	Ok,M
4	VSTDIC020	VSTDIC020	VY022778.D	23 Jun 2025 14:23		SY/MD	Ok,M
5	VSTDIC050	VSTDIC050	VY022779.D	23 Jun 2025 14:46		SY/MD	Ok,M
6	VSTDIC100	VSTDIC100	VY022780.D	23 Jun 2025 15:08		SY/MD	Ok,M
7	VSTDIC150	VSTDIC150	VY022781.D	23 Jun 2025 15:31		SY/MD	Ok,M
8	VIBLK	VIBLK	VY022782.D	23 Jun 2025 15:54		SY/MD	Ok
9	VSTDICV050	ICVVY062325	VY022783.D	23 Jun 2025 16:17		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY062425

Review By	Mahesh Dadoda	Review On	6/25/2025 12:32:22 PM
Supervise By	Semsettin Yesilyurt	Supervise On	6/25/2025 12:35:55 PM
SubDirectory	VY062425	HP Acquire Method	MSVOA_Y HP Processing Method 82y062325s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134483
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134483,VP134485 VP133934

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022784.D	24 Jun 2025 08:49		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY022785.D	24 Jun 2025 09:20		SY/MD	Ok,M
3	VY0624SBL01	VY0624SBL01	VY022786.D	24 Jun 2025 10:25		SY/MD	Ok
4	VY0624SBS01	VY0624SBS01	VY022787.D	24 Jun 2025 10:58		SY/MD	Ok,M
5	VY0624SBSD01	VY0624SBSD01	VY022788.D	24 Jun 2025 11:21		SY/MD	Ok,M
6	Q2386-01	SU-1-062025	VY022789.D	24 Jun 2025 11:58	vial-A Internal Standard Fail	SY/MD	ReRun
7	Q2383-01RE	RT-5376RE	VY022790.D	24 Jun 2025 12:22	vial-B Internal standard fail;Not match with first run	SY/MD	Not Ok
8	Q2384-01RE	OK-01-6202025RE	VY022791.D	24 Jun 2025 12:45	vial-B Internal Standard Fail	SY/MD	Confirms
9	Q2355-03	20071	VY022792.D	24 Jun 2025 13:09	vial-A not purge	SY/MD	Not Ok
10	Q2363-01	GCAP1	VY022793.D	24 Jun 2025 13:32	vial-A Surrogate fail ;Need MeOH	SY/MD	Dilution
11	Q2392-02	S-1A	VY022794.D	24 Jun 2025 13:55	vial-A	SY/MD	Ok
12	Q2393-02	PARK AVE -1	VY022795.D	24 Jun 2025 14:19	vial-A	SY/MD	Ok
13	Q2393-04	PARK AVE -2	VY022796.D	24 Jun 2025 14:42	vial-A	SY/MD	Ok
14	Q2393-06	PARK AVE -3	VY022797.D	24 Jun 2025 15:06	vial-A	SY/MD	Ok
15	Q2393-08	PARK AVE -4	VY022798.D	24 Jun 2025 15:29	vial-A	SY/MD	Ok
16	Q2393-10	PARK AVE -5	VY022799.D	24 Jun 2025 15:52	vial-A	SY/MD	Ok
17	Q2393-12	PARK AVE -6	VY022800.D	24 Jun 2025 16:16	vial-A	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY062425

Review By	Mahesh Dadoda	Review On	6/25/2025 12:32:22 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	6/25/2025 12:35:55 PM		
SubDirectory	VY062425	HP Acquire Method	MSVOA_Y	HP Processing Method	82y062325s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP134483			
CCC		VP134483,VP134485			
Internal Standard/PEM		VP133934			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	Q2394-03	MH-K/L VOC	VY022801.D	24 Jun 2025 16:39	vial-A	SY/MD	Ok
19	Q2399-03	TP-13-VOC	VY022802.D	24 Jun 2025 17:03	vial-A	SY/MD	Ok
20	Q2399-07	EP-7-VOC	VY022803.D	24 Jun 2025 17:26	vial-A	SY/MD	Ok
21	Q2403-03	LAW-25-0093	VY022804.D	24 Jun 2025 17:50	vial-A Internal Standard Fail	SY/MD	ReRun
22	Q2403-07	ARS 20-0006	VY022805.D	24 Jun 2025 18:13	vial-A	SY/MD	Ok
23	Q2398-03	M00-25-0179	VY022806.D	24 Jun 2025 18:36	vial-A	SY/MD	Ok
24	Q2398-06	ARS20-0036	VY022807.D	24 Jun 2025 19:00	vial-A Not purge	SY/MD	Not Ok
25	VIBLK	VIBLK	VY022808.D	24 Jun 2025 19:23		SY/MD	Ok
26	VSTDCCC050	VSTDCCC050EC	VY022809.D	24 Jun 2025 20:09		SY/MD	Ok,M

M : Manual Integration