

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q2363 SAS No.: Q2363 SDG NO.: Q2363
 Lab File ID: VN087122.D Date Analyzed: 06/20/2025
 Instrument ID: MSVOA_N Time Analyzed: 08:51
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	203626	8.23	354629	9.11	307356	11.87
UPPER LIMIT	407252	8.73	709258	9.606	614712	12.365
LOWER LIMIT	101813	7.73	177315	8.606	153678	11.365
EPA SAMPLE NO.						
GCAP1ME	160799	8.22	252217	9.10	281251	11.87
VN0620MBL01	161603	8.23	319855	9.11	284735	11.87
VN0620MBS01	157684	8.23	276593	9.11	261641	11.87
GCAP1WDL	192434	8.23	385508	9.11	346901	11.87
VN0620WBL01	186699	8.23	379621	9.11	343369	11.87
VN0620WBS01	175525	8.23	309350	9.11	276352	11.87

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.