

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q2363 SAS No.: Q2363 SDG NO.: Q2363
 Lab File ID: VY022785.D Date Analyzed: 06/24/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:20
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	480569	7.71	788269	8.62	678315	11.41
UPPER LIMIT	961138	8.207	1576540	9.116	1356630	11.914
LOWER LIMIT	240285	7.207	394135	8.116	339158	10.914
EPA SAMPLE NO.						
GCAP1	325185	7.70	571084	8.61	576784	11.41
VY0624SBL01	300507	7.71	535641	8.61	452693	11.41
VY0624SBS01	456381	7.70	765818	8.61	658494	11.41
VY0624SBSD01	454676	7.71	751533	8.61	644021	11.41

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.