

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q2363 SAS No.: Q2363 SDG NO.: Q2363
 Lab File ID: VN087122.D Date Analyzed: 06/20/2025
 Instrument ID: MSVOA_N Time Analyzed: 08:51
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	151554	13.788				
UPPER LIMIT	303108	14.288				
LOWER LIMIT	75777	13.288				
EPA SAMPLE NO.						
GCAP1ME	129659	13.79				
VN0620MBL01	131966	13.79				
VN0620MBS01	136861	13.79				
GCAP1WDL	163544	13.79				
VN0620WBL01	159510	13.79				
VN0620WBS01	139787	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.