

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: GENV01
Lab Code: CHEM Case No.: Q2363 SAS No.: Q2363 SDG NO.: Q2363
Lab File ID: VY022775.D BFB Injection Date: 06/23/2025
Instrument ID: MSVOA_Y BFB Injection Time: 10:17
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	22.8
75	30.0 - 60.0% of mass 95	56.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.8
173	Less than 2.0% of mass 174	0.9 (1.1) 1
174	50.0 - 100.0% of mass 95	81.9
175	5.0 - 9.0% of mass 174	6 (7.4) 1
176	95.0 - 101.0% of mass 174	78.2 (95.5) 1
177	5.0 - 9.0% of mass 176	5.1 (6.5) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC005	VSTDIC005	VY022776.D	06/23/2025	13:38
VSTDIC010	VSTDIC010	VY022777.D	06/23/2025	14:00
VSTDIC020	VSTDIC020	VY022778.D	06/23/2025	14:23
VSTDIC050	VSTDIC050	VY022779.D	06/23/2025	14:46
VSTDIC100	VSTDIC100	VY022780.D	06/23/2025	15:08
VSTDIC150	VSTDIC150	VY022781.D	06/23/2025	15:31