



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

G Environmental

Project Name: Ave B

Project # N/A

Order ID # Q2364

Test Name: Mercury, Metals ICP-TAL

A. Number of Samples and Date of Receipt:

4 Solid samples were received on 06/18/2025.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: Mercury, Metals ICP-TAL, METALS-TAL, PCB and VOC-TCLVOA-10. This data package contains results for Mercury, Metals ICP-TAL.

C. Analytical Techniques:

The analysis of Metals ICP-TAL was based on method 6010D, digestion based on method 3050 (soils). The analysis and digestion of Mercury was based on method 7471B.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

Sample GAV2B was diluted due to high concentrations for Mercury, Zinc.

The Blank Spike met requirements for all parameters.

The Duplicate (GAV1BDUP) analysis met criteria for all parameters except for Cadmium, Calcium, Cobalt, Copper, Lead, Manganese, Zinc due to matrix interference.

The Duplicate (GAV1BMSD) analysis met criteria for all parameters except for Copper due to matrix interference.

The Matrix Spike (GAV1BMS) analysis met criteria for all parameters except for Antimony, Barium, Manganese, Sodium due to matrix interference.

The Matrix Spike Duplicate (GAV1BMSD) analysis met criteria for all parameters except for Antimony, Barium, Sodium due to matrix interference.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

The Serial Dilution met the acceptable requirements.



E. Additional Comments:

The Post Digest Spike (GAV1BA) analysis met criteria for all parameters except for Manganese due to unknown chemical interferences of matrix with the addition of spike amount after digestion and before analysis, matrix has suppression effect during addition of spike.

In analytical sequence LB136273, the concentration was outside of acceptance limit for Chromium of CCB05 which is not associated to any sample of this project.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____