


LB136216

7471B

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	060
0.0	0.000	0.020	0.020	-47	0.000	-47					
0.2	0.200	0.213	0.013	989	0.0 %	989					
2.5	2.500	2.440	-0.060	12901	0.0 %	12901					
5.0	5.000	4.959	-0.041	26375	0.0 %	26375					
7.5	7.500	7.643	0.143	40732	0.0 %	40732					
10.0	10.000	9.924	-0.076	52934	0.0 %	52934					

LB136216 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 20 Jun 2025 13:25:31

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-47	-	0.0000	7471B	PPB	20 Jun 2025 13:31:53	S	Std
0.2 - 1	50-2	989	-	0.2000	7471B	PPB	20 Jun 2025 13:37:12	S	Std
2.5 - 1	52.5	12901	-	2.5000	7471B	PPB	20 Jun 2025 13:39:29	S	Std
5.0 - 1	75	26375	-	5.0000	7471B	PPB	20 Jun 2025 13:41:45	S	Std
7.5 - 1	75.5	40732	-	7.5000	7471B	PPB	20 Jun 2025 13:47:09	S	Std
10.0 - 1	710	52934	-	10.0000	7471B	PPB	20 Jun 2025 13:49:25	S	Std
ICV47 - 1	ICV47	21240	3.9992	-	7471B	PPB	20 Jun 2025 13:52:25	U	SMPL
ICB47 - 1	ICB47	-248	-0.0180	-	7471B	PPB	20 Jun 2025 13:54:41	U	SMPL
CCV69 - 1	CCV69	26700	5.0200	-	7471B	PPB	20 Jun 2025 13:56:59	U	SMPL
CCB69 - 1	CCB69	-291	-0.0260	-	7471B	PPB	20 Jun 2025 13:59:14	U	SMPL
CRA - 1	CRA	936	0.2034	-	7471B	PPB	20 Jun 2025 14:01:33	U	SMPL
HighStd - 1	HighStd	52756	9.8912	-	7471B	PPB	20 Jun 2025 14:03:47	U	SMPL
ChkStd - 1	ChkStd	38160	7.1624	-	7471B	PPB	20 Jun 2025 14:06:05	U	SMPL
PB168569BL - 1	PBS	-325	-0.0324	-	7471B	PPB	20 Jun 2025 14:08:31	U	SMPL
PB168569BS - 1	LCSS	22400	4.2161	-	7471B	PPB	20 Jun 2025 14:10:53	U	SMPL
Q2362-01 - 1	TP-11	51	0.0379	-	7471B	PPB	20 Jun 2025 14:13:10	U	SMPL
Q2362-05 - 1	EP-6	394	0.1020	-	7471B	PPB	20 Jun 2025 14:15:29	U	SMPL
Q2364-01 - 1	GAV1A	31501	5.9175	-	7471B	PPB	20 Jun 2025 14:17:48	U	SMPL
Q2364-02 - 1	GAV1B	7197	1.3739	-	7471B	PPB	20 Jun 2025 14:20:04	U	SMPL
Q2364-03 - 1	GAV2A	28407	5.3391	-	7471B	PPB	20 Jun 2025 14:22:28	U	SMPL
CCV70 - 1	CCV70	25912	4.8726	-	7471B	PPB	20 Jun 2025 14:24:45	U	SMPL
CCB70 - 1	CCB70	-291	-0.0260	-	7471B	PPB	20 Jun 2025 14:27:11	U	SMPL
Q2364-04 - 1	GAV2B	149581	27.9926	-	7471B	PPB	20 Jun 2025 14:29:31	U	SMPL
Q2367-01 - 1	EP-4	557	0.1325	-	7471B	PPB	20 Jun 2025 14:31:46	U	SMPL
Q2367-05 - 1	EP-5	300	0.0845	-	7471B	PPB	20 Jun 2025 14:34:33	U	SMPL
Q2367-05DUP - 1	EP-5DUP	355	0.0948	-	7471B	PPB	20 Jun 2025 14:36:49	U	SMPL
Q2367-05MS - 1	EP-5MS	21267	4.0043	-	7471B	PPB	20 Jun 2025 14:39:05	U	SMPL
Q2367-05MSD - 1	EP-5MSD	19666	3.7050	-	7471B	PPB	20 Jun 2025 14:41:22	U	SMPL
Q2367-05LX5 - 1		-181	-0.0055	-	7471B	PPB	20 Jun 2025 14:43:41	U	SMPL
Q2367-05A - 1		21489	4.0458	-	7471B	PPB	20 Jun 2025 14:46:02	U	SMPL
Q2364-04DLX5 - 1		29478	5.5393	-	7471B	PPB	20 Jun 2025 14:48:19	U	SMPL
CCV71 - 1	CCV71	26100	4.9078	-	7471B	PPB	20 Jun 2025 14:50:38	U	SMPL
CCB71 - 1	CCB71	-401	-0.0466	-	7471B	PPB	20 Jun 2025 14:53:00	U	SMPL