

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB136216

Review By	jaswal	Review On	6/23/2025 11:31:03 PM
Supervise By	Janvi	Supervise On	6/25/2025 8:44:08 AM

STD. NAME	STD REF.#
ICAL Standard	MP836020,MP836021,MP836022,MP836023,MP836024,MP836025
ICV Standard	MP836026
CCV Standard	MP836028
ICSA Standard	
CRI Standard	MP836030
LCS Standard	
Chk Standard	MP836027,MP836029,MP836031,MP836033

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	06/20/25 13:31		MOHAN	OK
2	S0.2	S0.2	CAL2	06/20/25 13:37		MOHAN	OK
3	S2.5	S2.5	CAL3	06/20/25 13:39		MOHAN	OK
4	S5	S5	CAL4	06/20/25 13:41		MOHAN	OK
5	S7.5	S7.5	CAL5	06/20/25 13:47		MOHAN	OK
6	S10	S10	CAL6	06/20/25 13:49		MOHAN	OK
7	ICV47	ICV47	ICV	06/20/25 13:52		MOHAN	OK
8	ICB47	ICB47	ICB	06/20/25 13:54		MOHAN	OK
9	CCV69	CCV69	CCV	06/20/25 13:56		MOHAN	OK
10	CCB69	CCB69	CCB	06/20/25 13:59		MOHAN	OK
11	CRA	CRA	CRDL	06/20/25 14:01		MOHAN	OK
12	HighStd	HighStd	HIGH STD	06/20/25 14:03		MOHAN	OK
13	ChkStd	ChkStd	SAM	06/20/25 14:06		MOHAN	OK
14	PB168569BL	PB168569BL	MB	06/20/25 14:08		MOHAN	OK
15	PB168569BS	PB168569BS	LCS	06/20/25 14:10		MOHAN	OK
16	Q2362-01	TP-11	SAM	06/20/25 14:13		MOHAN	OK
17	Q2362-05	EP-6	SAM	06/20/25 14:15		MOHAN	OK
18	Q2364-01	GAV1A	SAM	06/20/25 14:17		MOHAN	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB136216

Review By	jaswal	Review On	6/23/2025 11:31:03 PM
Supervise By	Janvi	Supervise On	6/25/2025 8:44:08 AM

STD. NAME	STD REF.#
ICAL Standard	MP836020,MP836021,MP836022,MP836023,MP836024,MP836025
ICV Standard	MP836026
CCV Standard	MP836028
ICSA Standard	
CRI Standard	MP836030
LCS Standard	
Chk Standard	MP836027,MP836029,MP836031,MP836033

19	Q2364-02	GAV1B	SAM	06/20/25 14:20		MOHAN	OK
20	Q2364-03	GAV2A	SAM	06/20/25 14:22		MOHAN	OK
21	CCV70	CCV70	CCV	06/20/25 14:24		MOHAN	OK
22	CCB70	CCB70	CCB	06/20/25 14:27		MOHAN	OK
23	Q2364-04	GAV2B	SAM	06/20/25 14:29	Hg high	MOHAN	Dilution
24	Q2367-01	EP-4	SAM	06/20/25 14:31		MOHAN	OK
25	Q2367-05	EP-5	SAM	06/20/25 14:34		MOHAN	OK
26	Q2367-05DUP	EP-5DUP	DUP	06/20/25 14:36		MOHAN	OK
27	Q2367-05MS	EP-5MS	MS	06/20/25 14:39		MOHAN	OK
28	Q2367-05MSD	EP-5MSD	MSD	06/20/25 14:41		MOHAN	OK
29	Q2367-05L	EP-5L	SD	06/20/25 14:43		MOHAN	OK
30	Q2367-05A	EP-5A	PS	06/20/25 14:46		MOHAN	OK
31	Q2364-04DL	GAV2BDL	SAM	06/20/25 14:48	5X for Hg	MOHAN	Confirms
32	CCV71	CCV71	CCV	06/20/25 14:50		MOHAN	OK
33	CCB71	CCB71	CCB	06/20/25 14:53		MOHAN	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136236

Review By	jaswal	Review On	6/24/2025 4:50:49 AM
Supervise By	Janvi	Supervise On	6/25/2025 8:43:58 AM

STD. NAME	STD REF.#
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868
ICV Standard	MP85872,MP85897
CCV Standard	MP85875
ICSA Standard	MP85873,MP85874
CRI Standard	MP85897
LCS Standard	
Chk Standard	MP85876,MP85877

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	06/23/25 11:53		jaswal	OK
2	S1	S1	CAL2	06/23/25 11:58		jaswal	OK
3	S2	S2	CAL3	06/23/25 12:02		jaswal	OK
4	S3	S3	CAL4	06/23/25 12:07		jaswal	OK
5	S4	S4	CAL5	06/23/25 12:11		jaswal	OK
6	S5	S5	CAL6	06/23/25 12:15		jaswal	OK
7	ICV01	ICV01	ICV	06/23/25 12:34		jaswal	OK
8	LLICV01	LLICV01	LLICV	06/23/25 13:09		jaswal	OK
9	ICB01	ICB01	ICB	06/23/25 13:19		jaswal	OK
10	CRI01	CRI01	CRDL	06/23/25 13:23		jaswal	OK
11	ICSA01	ICSA01	ICSA	06/23/25 13:34		jaswal	OK
12	ICSAB01	ICSAB01	ICSAB	06/23/25 13:43		jaswal	OK
13	ICSADL	ICSADL	ICSA	06/23/25 13:47		jaswal	OK
14	ICSABDL	ICSABDL	ICSAB	06/23/25 13:52		jaswal	OK
15	CCV01	CCV01	CCV	06/23/25 13:56		jaswal	OK
16	CCB01	CCB01	CCB	06/23/25 14:05		jaswal	OK
17	Q2347-01	TP-10	SAM	06/23/25 14:09		jaswal	OK
18	Q2354-01	3321	SAM	06/23/25 14:14		jaswal	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136236

Review By	jaswal	Review On	6/24/2025 4:50:49 AM
Supervise By	Janvi	Supervise On	6/25/2025 8:43:58 AM

STD. NAME	STD REF.#
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868
ICV Standard	MP85872,MP85897
CCV Standard	MP85875
ICSA Standard	MP85873,MP85874
CRI Standard	MP85897
LCS Standard	
Chk Standard	MP85876,MP85877

19	Q2355-01	3897	SAM	06/23/25 14:18		jaswal	OK
20	Q2355-03	20071	SAM	06/23/25 14:22		jaswal	OK
21	PB168553BL	PB168553BL	MB	06/23/25 14:27		jaswal	OK
22	PB168553BS	PB168553BS	LCS	06/23/25 14:31		jaswal	OK
23	PB168535BL	PB168535BL	MB	06/23/25 14:36		jaswal	OK
24	PB168535BS	PB168535BS	LCS	06/23/25 14:40		jaswal	OK
25	PB168514TB	PB168514TB	MB	06/23/25 14:44		jaswal	OK
26	PB168544TB	PB168544TB	MB	06/23/25 14:49		jaswal	OK
27	CCV02	CCV02	CCV	06/23/25 14:54		jaswal	OK
28	CCB02	CCB02	CCB	06/23/25 14:58		jaswal	OK
29	PB168565BL	PB168565BL	MB	06/23/25 15:02		jaswal	OK
30	PB168565BS	PB168565BS	LCS	06/23/25 15:07		jaswal	OK
31	PB168564BL	PB168564BL	MB	06/23/25 15:11		jaswal	OK
32	PB168564BS	PB168564BS	LCS	06/23/25 15:16		jaswal	OK
33	Q2354-02	3321	SAM	06/23/25 15:20		jaswal	OK
34	Q2355-02	3897	SAM	06/23/25 15:25		jaswal	OK
35	Q2362-04	TP-11	SAM	06/23/25 15:29	Not required	jaswal	Not Ok
36	Q2362-08	EP-6	SAM	06/23/25 15:34		jaswal	OK
37	Q2367-04	EP-4	SAM	06/23/25 15:38	Not Required	jaswal	Not Ok
38	Q2367-08	EP-5	SAM	06/23/25 15:43		jaswal	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136236

Review By	jaswal	Review On	6/24/2025 4:50:49 AM
Supervise By	Janvi	Supervise On	6/25/2025 8:43:58 AM
STD. NAME	STD REF.#		
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868		
ICV Standard	MP85872,MP85897		
CCV Standard	MP85875		
ICSA Standard	MP85873,MP85874		
CRI Standard	MP85897		
LCS Standard			
Chk Standard	MP85876,MP85877		

39	CCV03	CCV03	CCV	06/23/25 15:47		jaswal	OK
40	CCB03	CCB03	CCB	06/23/25 15:52		jaswal	OK
41	Q2367-08DUP	EP-5DUP	DUP	06/23/25 15:57		jaswal	OK
42	Q2367-08L	EP-5L	SD	06/23/25 16:02		jaswal	OK
43	Q2367-08MS	EP-5MS	MS	06/23/25 16:06		jaswal	OK
44	Q2367-08MSD	EP-5MSD	MSD	06/23/25 16:11		jaswal	OK
45	Q2367-08A	EP-5A	PS	06/23/25 16:15		jaswal	OK
46	Q2362-01	TP-11	SAM	06/23/25 16:20		jaswal	OK
47	Q2362-05	EP-6	SAM	06/23/25 16:24		jaswal	OK
48	Q2363-01	GCAP1	SAM	06/23/25 16:28		jaswal	OK
49	Q2364-01	GAV1A	SAM	06/23/25 16:33		jaswal	OK
50	Q2364-02	GAV1B	SAM	06/23/25 16:37		jaswal	OK
51	CCV04	CCV04	CCV	06/23/25 16:51		jaswal	OK
52	CCB04	CCB04	CCB	06/23/25 16:56		jaswal	OK
53	Q2364-02DUP	GAV1BDUP	DUP	06/23/25 17:01		jaswal	OK
54	Q2364-02L	GAV1BL	SD	06/23/25 17:06		jaswal	OK
55	Q2364-02MS	GAV1BMS	MS	06/23/25 17:10		jaswal	OK
56	Q2364-02MSD	GAV1BMSD	MSD	06/23/25 17:14		jaswal	OK
57	Q2364-02A	GAV1BA	PS	06/23/25 17:19		jaswal	OK
58	Q2364-03	GAV2A	SAM	06/23/25 17:23		jaswal	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136236

Review By	jaswal	Review On	6/24/2025 4:50:49 AM
Supervise By	Janvi	Supervise On	6/25/2025 8:43:58 AM
STD. NAME	STD REF.#		
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868		
ICV Standard	MP85872,MP85897		
CCV Standard	MP85875		
ICSA Standard	MP85873,MP85874		
CRI Standard	MP85897		
LCS Standard			
Chk Standard	MP85876,MP85877		

59	Q2364-04	GAV2B	SAM	06/23/25 17:27	Zn high	jaswal	Dilution
60	Q2367-01	EP-4	SAM	06/23/25 17:32		jaswal	OK
61	Q2367-05	EP-5	SAM	06/23/25 17:36		jaswal	OK
62	Q2371-04	RPXY42025	SAM	06/23/25 17:41		jaswal	OK
63	CCV05	CCV05	CCV	06/23/25 17:45		jaswal	OK
64	CCB05	CCB05	CCB	06/23/25 17:49		jaswal	OK
65	Q2371-05	BBX42025	SAM	06/23/25 17:54		jaswal	OK
66	PB168551BL	PB168551BL	MB	06/23/25 17:58	NOT USE	jaswal	Not Ok
67	PB168551BS	PB168551BS	LCS	06/23/25 18:03	NOT USE	jaswal	Not Ok
68	Q2351-01	FAIRLAWN-SUMP	SAM	06/23/25 18:07		jaswal	OK
69	Q2354-05	3311	SAM	06/23/25 18:12	K oversaturated	jaswal	Dilution
70	Q2344-01	MW-1	SAM	06/23/25 18:16		jaswal	OK
71	Q2344-02	MW-1	SAM	06/23/25 18:20		jaswal	OK
72	Q2344-03	MW-3	SAM	06/23/25 18:25		jaswal	OK
73	Q2344-04	MW-3	SAM	06/23/25 18:30		jaswal	OK
74	Q2344-05	MW-4	SAM	06/23/25 18:34		jaswal	OK
75	CCV06	CCV06	CCV	06/23/25 18:39		jaswal	OK
76	CCB06	CCB06	CCB	06/23/25 18:43		jaswal	OK
77	Q2344-06	MW-4	SAM	06/23/25 18:48		jaswal	OK
78	Q2344-07	MW-2	SAM	06/23/25 18:52		jaswal	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QCBatch ID # LB136236

Review By	jaswal	Review On	6/24/2025 4:50:49 AM
Supervise By	Janvi	Supervise On	6/25/2025 8:43:58 AM
STD. NAME	STD REF.#		
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868		
ICV Standard	MP85872,MP85897		
CCV Standard	MP85875		
ICSA Standard	MP85873,MP85874		
CRI Standard	MP85897		
LCS Standard			
Chk Standard	MP85876,MP85877		

79	Q2344-07DUP	MW-2DUP	DUP	06/23/25 18:57		jaswal	OK
80	Q2344-07L	MW-2L	SD	06/23/25 19:02		jaswal	OK
81	Q2344-07MS	MW-2MS	MS	06/23/25 19:06		jaswal	OK
82	Q2344-07MSD	MW-2MSD	MSD	06/23/25 19:10		jaswal	OK
83	Q2344-07A	MW-2A	PS	06/23/25 19:15		jaswal	OK
84	Q2344-08	MW-2	SAM	06/23/25 19:19		jaswal	OK
85	CCV07	CCV07	CCV	06/23/25 19:24		jaswal	OK
86	CCB07	CCB07	CCB	06/23/25 19:28		jaswal	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136273

Review By	Janvi	Review On	6/26/2025 11:24:19 AM
Supervise By	jaswal	Supervise On	6/26/2025 3:29:22 PM

STD. NAME	STD REF.#
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868
ICV Standard	MP85934
CCV Standard	MP85875
ICSA Standard	MP85873,MP85874
CRI Standard	MP85897
LCS Standard	
Chk Standard	MP85876,MP85877

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	06/25/25 12:23		Jaswal	OK
2	S1	S1	CAL2	06/25/25 12:27		Jaswal	OK
3	S2	S2	CAL3	06/25/25 12:32		Jaswal	OK
4	S3	S3	CAL4	06/25/25 12:36		Jaswal	OK
5	S4	S4	CAL5	06/25/25 12:40		Jaswal	OK
6	S5	S5	CAL6	06/25/25 12:44		Jaswal	OK
7	ICV01	ICV01	ICV	06/25/25 13:08	ICV fail for Sb,Cu (200.7) (95-105)	Jaswal	OK
8	LLICV01	LLICV01	LLICV	06/25/25 13:20		Jaswal	OK
9	ICB01	ICB01	ICB	06/25/25 14:53		Jaswal	OK
10	CRI01	CRI01	CRDL	06/25/25 15:24		Jaswal	OK
11	ICSA01	ICSA01	ICSA	06/25/25 15:29		Jaswal	OK
12	ICSAB01	ICSAB01	ICSAB	06/25/25 15:51		Jaswal	OK
13	ICSADL	ICSADL	ICSA	06/25/25 15:55		Jaswal	OK
14	ICSABDL	ICSABDL	ICSAB	06/25/25 15:59		Jaswal	OK
15	CCV01	CCV01	CCV	06/25/25 16:04		Jaswal	OK
16	CCB01	CCB01	CCB	06/25/25 16:09		Jaswal	OK
17	Q2385-01	A5311	SAM	06/25/25 16:13		Jaswal	OK
18	Q2385-01DUP	A5311DUP	DUP	06/25/25 16:17		Jaswal	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136273

Review By	Janvi	Review On	6/26/2025 11:24:19 AM
Supervise By	jaswal	Supervise On	6/26/2025 3:29:22 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868		
ICV Standard	MP85934		
CCV Standard	MP85875		
ICSA Standard	MP85873,MP85874		
CRI Standard	MP85897		
LCS Standard			
Chk Standard	MP85876,MP85877		

19	Q2385-01L	A5311L	SD	06/25/25 16:22		Jaswal	OK
20	Q2385-01MS	A5311MS	MS	06/25/25 16:26		Jaswal	OK
21	Q2385-01MSD	A5311MSD	MSD	06/25/25 16:31		Jaswal	OK
22	Q2385-01A	A5311A	PS	06/25/25 16:35		Jaswal	OK
23	Q2354-05DL	3311DL	SAM	06/25/25 16:39	5X for K	Jaswal	Confirms
24	Q2355-02	3897	SAM	06/25/25 16:44		Jaswal	OK
25	Q2362-04	TP-11	SAM	06/25/25 16:48		Jaswal	OK
26	Q2364-04DL	GAV2BDL	SAM	06/25/25 16:53	5X for Zn, Still high	Jaswal	Dilution
27	CCV02	CCV02	CCV	06/25/25 16:57		Jaswal	OK
28	CCB02	CCB02	CCB	06/25/25 17:09		Jaswal	OK
29	Q2364-04DL2	GAV2BDL2	SAM	06/25/25 17:15	10X for Zn	Jaswal	Confirms
30	Q2380-03DL	VNJ-211DL	SAM	06/25/25 17:24	5X for Na	Jaswal	Confirms
31	Q2380-05DL	RBR-200059DL	SAM	06/25/25 17:28	5X for Na	Jaswal	Confirms
32	Q2381-01	291431	SAM	06/25/25 17:33		Jaswal	Not Ok
33	Q2381-01DUP	291431DUP	DUP	06/25/25 17:37	Na oversaturated	Jaswal	Not Ok
34	Q2381-01L	291431L	SD	06/25/25 17:41	Bad Injection	Jaswal	Not Ok
35	Q2381-01MS	291431MS	MS	06/25/25 17:46	Bad Injection	Jaswal	Not Ok
36	Q2381-01MSD	291431MSD	MSD	06/25/25 17:50		Jaswal	Not Ok
37	Q2381-01A	291431A	PS	06/25/25 17:55	Bad Injection	Jaswal	Not Ok
38	CCV03	CCV03	CCV	06/25/25 18:00		Jaswal	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136273

Review By	Janvi	Review On	6/26/2025 11:24:19 AM
Supervise By	jaswal	Supervise On	6/26/2025 3:29:22 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868		
ICV Standard	MP85934		
CCV Standard	MP85875		
ICSA Standard	MP85873,MP85874		
CRI Standard	MP85897		
LCS Standard			
Chk Standard	MP85876,MP85877		

39	CCB03	CCB03	CCB	06/25/25 18:07		Jaswal	OK
40	Q2367-04	EP-4	SAM	06/25/25 18:12		Jaswal	OK
41	Q2386-02	SU-1-062025	SAM	06/25/25 18:16		Jaswal	OK
42	Q2392-01	S-1	SAM	06/25/25 18:21		Jaswal	OK
43	Q2381-01DL	291431DL	SAM	06/25/25 18:29	Not use	Jaswal	Not Ok
44	Q2381-01DUPDL	291431DUPDL	DUP	06/25/25 18:34	Not use	Jaswal	Not Ok
45	Q2381-01LDL	291431LDL	SD	06/25/25 18:38	Not use	Jaswal	Not Ok
46	Q2381-01MSDL	291431MSDL	MS	06/25/25 18:42	Not use	Jaswal	Not Ok
47	Q2381-01MSDDL	291431MSDDL	MSD	06/25/25 18:47	Not use	Jaswal	Not Ok
48	Q2381-01ADL	291431ADL	PS	06/25/25 18:51	Not use	Jaswal	Not Ok
49	CCV04	CCV04	CCV	06/25/25 18:55		Jaswal	OK
50	CCB04	CCB04	CCB	06/25/25 18:59		Jaswal	OK
51	Q2426-01	001-WILLETS-PT-BL	SAM	06/25/25 19:04		Jaswal	OK
52	Q2426-02	002-35th-Ave(May)	SAM	06/25/25 19:09		Jaswal	OK
53	Q2388-04	TP-12	SAM	06/25/25 19:13		Jaswal	OK
54	Q2389-04	MH-J-I	SAM	06/25/25 19:18	Bad Injection	Jaswal	Not Ok
55	Q2391-01	AUD-1623	SAM	06/25/25 19:22		Jaswal	OK
56	Q2394-04	MH-K/L	SAM	06/25/25 19:29	Bad Injection	Jaswal	Not Ok
57	Q2399-04	TP-13	SAM	06/25/25 19:34		Jaswal	OK
58	Q2399-08	EP-13	SAM	06/25/25 19:39		Jaswal	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136273

Review By	Janvi	Review On	6/26/2025 11:24:19 AM
Supervise By	jaswal	Supervise On	6/26/2025 3:29:22 PM

STD. NAME	STD REF.#
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868
ICV Standard	MP85934
CCV Standard	MP85875
ICSA Standard	MP85873,MP85874
CRI Standard	MP85897
LCS Standard	
Chk Standard	MP85876,MP85877

59	Q2405-04	MH-M/N	SAM	06/25/25 19:43		Jaswal	OK
60	PB168571TB	PB168571TB	MB	06/25/25 19:48		Jaswal	OK
61	CCV05	CCV05	CCV	06/25/25 19:52		Jaswal	OK
62	CCB05	CCB05	CCB	06/25/25 19:57		Jaswal	OK
63	PB168613BL	PB168613BL	MB	06/25/25 20:01	CCB fail for Cr	Jaswal	Not Ok
64	PB168613BS	PB168613BS	LCS	06/25/25 20:06	CCB fail for Cr	Jaswal	Not Ok
65	Q2409-02	COP-SOIL-PILE	SAM	06/25/25 20:10	CCB fail for Cr	Jaswal	Not Ok
66	Q2409-02DUP	COP-SOIL-PILEDUP	DUP	06/25/25 20:14	CCB fail for Cr	Jaswal	Not Ok
67	Q2409-02L	COP-SOIL-PILEL	SD	06/25/25 20:19	CCB fail for Cr	Jaswal	Not Ok
68	Q2409-02MS	COP-SOIL-PILEMS	MS	06/25/25 20:23	CCB fail for Cr	Jaswal	Not Ok
69	Q2409-02MSD	COP-SOIL-PILEMSD	MSD	06/25/25 20:28	CCB fail for Cr	Jaswal	Not Ok
70	Q2409-02A	COP-SOIL-PILEA	PS	06/25/25 20:32	CCB fail for Cr	Jaswal	Not Ok
71	Q2381-01DL2	291431DL2	SAM	06/25/25 20:37	CCB fail for Cr	Jaswal	Not Ok
72	Q2381-01DUPDL2	291431DUPDL2	DUP	06/25/25 20:41	CCB fail for Cr	Jaswal	Not Ok
73	CCV06	CCV06	CCV	06/25/25 20:46		Jaswal	OK
74	CCB06	CCB06	CCB	06/25/25 20:50	Fail for Cr	Jaswal	OK
75	Q2381-01LDL2	291431LDL2	SD	06/25/25 20:54	CCB fail for Cr	Jaswal	Not Ok
76	Q2381-01MSDL2	291431MSDL2	MS	06/25/25 20:59	CCB fail for Cr	Jaswal	Not Ok
77	Q2381-01MSDDL2	291431MSDDL2	MSD	06/25/25 21:03	CCB fail for Cr	Jaswal	Not Ok
78	Q2381-01ADL2	291431ADL2	PS	06/25/25 21:07	CCB fail for Cr	Jaswal	Not Ok

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136273

Review By	Janvi	Review On	6/26/2025 11:24:19 AM
Supervise By	jaswal	Supervise On	6/26/2025 3:29:22 PM

STD. NAME	STD REF.#
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868
ICV Standard	MP85934
CCV Standard	MP85875
ICSA Standard	MP85873,MP85874
CRI Standard	MP85897
LCS Standard	
Chk Standard	MP85876,MP85877

79	CCV07	CCV07	CCV	06/25/25 21:12		Jaswal	OK
80	CCB07	CCB07	CCB	06/25/25 21:16		Jaswal	OK