



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: G Environmental SDG No.: Q2364
Contract: GENV01 Lab Code: CHEM Case No.: Q2364 SAS No.: Q2364
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.20	0.2	102	70 - 130	CV	06/20/2025	14:01	LB136216
CRI01	Aluminum	109	100	109	65 - 135	P	06/23/2025	13:23	LB136236
	Antimony	52.9	50.0	106	65 - 135	P	06/23/2025	13:23	LB136236
	Arsenic	18.0	20.0	90	65 - 135	P	06/23/2025	13:23	LB136236
	Barium	100	100	100	65 - 135	P	06/23/2025	13:23	LB136236
	Beryllium	6.35	6.0	106	65 - 135	P	06/23/2025	13:23	LB136236
	Cadmium	5.84	6.0	97	65 - 135	P	06/23/2025	13:23	LB136236
	Calcium	2020	2000	101	65 - 135	P	06/23/2025	13:23	LB136236
	Chromium	10.1	10.0	101	65 - 135	P	06/23/2025	13:23	LB136236
	Cobalt	31.2	30.0	104	65 - 135	P	06/23/2025	13:23	LB136236
	Copper	23.3	20.0	116	65 - 135	P	06/23/2025	13:23	LB136236
	Iron	94.8	100	95	65 - 135	P	06/23/2025	13:23	LB136236
	Lead	11.8	12.0	98	65 - 135	P	06/23/2025	13:23	LB136236
	Magnesium	2120	2000	106	65 - 135	P	06/23/2025	13:23	LB136236
	Manganese	20.4	20.0	102	65 - 135	P	06/23/2025	13:23	LB136236
	Nickel	40.7	40.0	102	65 - 135	P	06/23/2025	13:23	LB136236
	Potassium	1670	2000	83	65 - 135	P	06/23/2025	13:23	LB136236
	Selenium	17.5	20.0	88	65 - 135	P	06/23/2025	13:23	LB136236
	Silver	9.75	10.0	98	65 - 135	P	06/23/2025	13:23	LB136236
	Sodium	1870	2000	93	65 - 135	P	06/23/2025	13:23	LB136236
	Thallium	40.9	40.0	102	65 - 135	P	06/23/2025	13:23	LB136236
	Vanadium	34.3	40.0	86	65 - 135	P	06/23/2025	13:23	LB136236
	Zinc	40.3	40.0	101	65 - 135	P	06/23/2025	13:23	LB136236
CRI01	Aluminum	103	100	103	65 - 135	P	06/25/2025	15:24	LB136273
	Antimony	49.6	50.0	99	65 - 135	P	06/25/2025	15:24	LB136273
	Arsenic	16.5	20.0	82	65 - 135	P	06/25/2025	15:24	LB136273
	Barium	99.4	100	99	65 - 135	P	06/25/2025	15:24	LB136273
	Beryllium	6.43	6.0	107	65 - 135	P	06/25/2025	15:24	LB136273
	Cadmium	5.59	6.0	93	65 - 135	P	06/25/2025	15:24	LB136273
	Calcium	2000	2000	100	65 - 135	P	06/25/2025	15:24	LB136273
	Chromium	8.98	10.0	90	65 - 135	P	06/25/2025	15:24	LB136273
	Cobalt	29.9	30.0	100	65 - 135	P	06/25/2025	15:24	LB136273
	Copper	22.1	20.0	110	65 - 135	P	06/25/2025	15:24	LB136273
	Iron	107	100	107	65 - 135	P	06/25/2025	15:24	LB136273
	Lead	13.2	12.0	110	65 - 135	P	06/25/2025	15:24	LB136273

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: G Environmental **SDG No.:** Q2364
Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q2364 **SAS No.:** Q2364
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Magnesium	2150	2000	108	65 - 135	P	06/25/2025	15:24	LB136273
	Manganese	24.3	20.0	122	65 - 135	P	06/25/2025	15:24	LB136273
	Nickel	38.6	40.0	96	65 - 135	P	06/25/2025	15:24	LB136273
	Potassium	1940	2000	97	65 - 135	P	06/25/2025	15:24	LB136273
	Selenium	20.3	20.0	102	65 - 135	P	06/25/2025	15:24	LB136273
	Silver	9.71	10.0	97	65 - 135	P	06/25/2025	15:24	LB136273
	Sodium	1870	2000	94	65 - 135	P	06/25/2025	15:24	LB136273
	Thallium	36.8	40.0	92	65 - 135	P	06/25/2025	15:24	LB136273
	Vanadium	42.8	40.0	107	65 - 135	P	06/25/2025	15:24	LB136273
	Zinc	39.7	40.0	99	65 - 135	P	06/25/2025	15:24	LB136273