

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: GENV01

Lab Code: CHEM **Case No.:** Q2364 **SAS No.:** Q2364 **SDG NO.:** Q2364

Instrument ID: ECD_P **Calibration Date(s):** 06/17/2025 06/17/2025
Calibration Times: 10:04 20:10

GC Column: ZB-MR1 **ID:** 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PP072991.D</u>		CF 750 = <u>PP072992.D</u>			
CF 500 = <u>PP072993.D</u>		CF 250 = <u>PP072994.D</u>		CF 050 = <u>PP072995.D</u>			
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	63671603	67644257	70148492	76963128	84930220	12
Aroclor-1016-2	(2)	97440661	102354848	107008936	114458460	98840440	7
Aroclor-1016-3	(3)	59983524	63267185	65041372	73554480	64895380	8
Aroclor-1016-4	(4)	50589003	52912481	54400336	56287652	47214800	7
Aroclor-1016-5	(5)	46240143	47840919	49349310	51433860	55161280	7
Aroclor-1260-1	(1)	82538787	86792979	90284630	94211564	87238820	5
Aroclor-1260-2	(2)	125600539	132072107	137363486	143645412	140737120	5
Aroclor-1260-3	(3)	106201830	111185451	115341698	120182220	117849880	5
Aroclor-1260-4	(4)	98670506	102821608	101520658	108282220	103932620	3
Aroclor-1260-5	(5)	227637841	236485680	245198018	255372264	238418920	4
Decachlorobiphenyl		1620648950	1677444467	1749703060	1784234960	1712434400	4
Tetrachloro-m-xylene		1977356170	2059482187	2109826840	2243117400	2164249200	5
Aroclor-1242-1	(1)	54005130	54512615	57421654	60290804	65956000	8
Aroclor-1242-2	(2)	81931377	79616400	88109514	98852440	84261940	9
Aroclor-1242-3	(3)	50398728	48446856	53645312	57446220	53601300	7
Aroclor-1242-4	(4)	42240385	41710237	44319270	46660108	44861080	5
Aroclor-1242-5	(5)	45003164	45972225	47192024	56790384	52840600	10
Decachlorobiphenyl		1612495640	1627929573	1725341220	1814607120	1610121000	5
Tetrachloro-m-xylene		1888746510	1833283947	1976454780	2024098800	2031619200	4
Aroclor-1248-1	(1)	41858490	44170731	45862520	49041552	56908940	12
Aroclor-1248-2	(2)	55380415	58022209	55617524	59041032	64693660	6
Aroclor-1248-3	(3)	63266998	66239924	62737624	66758520	67243400	3
Aroclor-1248-4	(4)	77728954	82015129	82895984	86036548	96645880	8
Aroclor-1248-5	(5)	75041779	78134308	81070348	82544344	101675060	12
Decachlorobiphenyl		1620420410	1689508547	1707113880	1772211040	1987340600	8
Tetrachloro-m-xylene		1885218090	1960849747	1904121260	1960400040	2044577000	3
Aroclor-1254-1	(1)	73763285	78885443	81757104	88891384	97572240	11
Aroclor-1254-2	(2)	111451506	118674340	123357828	133620984	139899720	9
Aroclor-1254-3	(3)	119235877	126285657	130365552	141133040	139034740	7
Aroclor-1254-4	(4)	108754762	114515340	117946542	128101128	129602040	7
Aroclor-1254-5	(5)	106940669	112254864	115991910	123909512	112185340	6
Decachlorobiphenyl		1648218000	1739306053	1775656020	1935470960	1659491600	7
Tetrachloro-m-xylene		1872735760	1960277133	2019595300	2159825040	2074265000	5
Aroclor-1268-1	(1)	342609292	345928729	355896714	398597476	345083220	7

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	290418323	293649853	302966618	340918088	297353440	305061264	7
Aroclor-1268-3	(3)	249992898	253511760	262173960	295232256	266657480	265513671	7
Aroclor-1268-4	(4)	114648839	115165531	118059850	128871696	108376940	117024571	6
Aroclor-1268-5	(5)	713831298	735909031	738672980	819932440	689883900	739645930	7
Decachlorobiphenyl		3002859930	3106573307	3204862440	3588507520	3062996400	3193159919	7
Tetrachloro-m-xylene		1965399560	2013897333	2066727860	2334276840	2005844200	2077229159	7

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: GENV01

Lab Code: CHEM **Case No.:** Q2364 **SAS No.:** Q2364 **SDG NO.:** Q2364

Instrument ID: ECD_P **Calibration Date(s):** 06/17/2025 06/17/2025
Calibration Times: 10:04 20:10

GC Column: ZB-MR2 **ID:** 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PP072991.D</u>		CF 750 = <u>PP072992.D</u>			
CF 500 = <u>PP072993.D</u>		CF 250 = <u>PP072994.D</u>		CF 050 = <u>PP072995.D</u>			
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	58734267	64000205	67674152	72695560	80484240	12
Aroclor-1016-2	(2)	90382349	94344453	98038690	104433472	115993580	10
Aroclor-1016-3	(3)	48136215	50748025	53595712	56975444	64125620	11
Aroclor-1016-4	(4)	38107893	40907291	43245840	46954372	53532680	13
Aroclor-1016-5	(5)	49203503	51615780	54910048	58505108	62177080	9
Aroclor-1260-1	(1)	84892100	88579871	95448406	100236824	104495860	9
Aroclor-1260-2	(2)	105588352	108160505	115763440	121715240	130953840	9
Aroclor-1260-3	(3)	96095206	97763095	104346708	109521728	113078520	7
Aroclor-1260-4	(4)	77788306	79915727	86740682	92092068	102155800	11
Aroclor-1260-5	(5)	196041650	198840268	209548214	218270728	244301260	9
Decachlorobiphenyl		1225771100	1248497173	1342901640	1435827000	1420581600	7
Tetrachloro-m-xylene		1662896120	1767446653	1788437460	1893713440	1896490200	5
Aroclor-1242-1	(1)	50070930	49549424	54443442	59214632	62816600	10
Aroclor-1242-2	(2)	73623300	71492977	79469612	84317900	88528280	9
Aroclor-1242-3	(3)	39446084	38282884	42982138	46709656	49104180	11
Aroclor-1242-4	(4)	37721961	36558999	41186366	44840456	47211420	11
Aroclor-1242-5	(5)	49158591	46923105	52837542	55505240	58508140	9
Decachlorobiphenyl		1205418560	1226140693	1281128260	1336855440	1341504000	5
Tetrachloro-m-xylene		1683517200	1544398853	1783586240	1693726480	1780668000	6
Aroclor-1248-1	(1)	39929038	42320627	40002718	45282436	50417100	10
Aroclor-1248-2	(2)	53044396	57410791	53304096	60483108	68735640	11
Aroclor-1248-3	(3)	55822275	59851821	54787908	62834628	69424220	10
Aroclor-1248-4	(4)	64712793	69142612	62749900	72698224	80808780	10
Aroclor-1248-5	(5)	65436447	68813545	68635516	72459776	79814180	8
Decachlorobiphenyl		1226985710	1250965200	1294060880	1349363760	1435080400	6
Tetrachloro-m-xylene		1707872610	1775213400	1595172200	1742416320	1838339400	5
Aroclor-1254-1	(1)	94131500	102369100	108358230	120483728	122907600	11
Aroclor-1254-2	(2)	80645865	88030495	92596756	103733056	107725280	12
Aroclor-1254-3	(3)	128776319	138178049	147317104	161124892	162339800	10
Aroclor-1254-4	(4)	83562949	90268741	96675708	106896836	108101020	11
Aroclor-1254-5	(5)	116424396	122854644	129598642	142067368	138868340	8
Decachlorobiphenyl		1234350420	1314370427	1354328000	1438665160	1427909200	6
Tetrachloro-m-xylene		1670033590	1775346533	1836392320	1937020080	1788575000	5
Aroclor-1268-1	(1)	263809069	277704169	280800266	313024424	300977600	7

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	236468443	251452185	251704404	280300416	265662900	257117670	6
Aroclor-1268-3	(3)	197435098	210533244	211556648	231588560	228588480	215940406	7
Aroclor-1268-4	(4)	83966367	90997261	93004838	103484904	99474300	94185534	8
Aroclor-1268-5	(5)	544182935	572449871	587387342	623593632	566405340	578803824	5
Decachlorobiphenyl		2216214780	2314031280	2438938620	2677731000	2452104000	2419803936	7
Tetrachloro-m-xylene		1798342890	1816167067	1835922080	2018457320	1810038400	1855785551	5