

CALIBRATION VERIFICATION SUMMARY

Contract: GENV01

Lab Code: CHEM Case No.: Q2364 SAS No.: Q2364 SDG NO.: Q2364

Continuing Calib Date: 06/20/2025 Initial Calibration Date(s): 06/17/2025 06/17/2025

Continuing Calib Time: 15:56 Initial Calibration Time(s): 10:04 20:10

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1248-1	(1)	5.64	5.65	5.55	5.75	0.01
Aroclor-1248-2	(2)	5.91	5.92	5.82	6.02	0.01
Aroclor-1248-3	(3)	6.12	6.12	6.02	6.22	0.01
Aroclor-1248-4	(4)	6.51	6.52	6.42	6.62	0.01
Aroclor-1248-5	(5)	6.55	6.56	6.46	6.66	0.01
Tetrachloro-m-xylene		4.49	4.49	4.39	4.59	0.00
Decachlorobiphenyl		10.18	10.19	10.09	10.29	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: GENV01

Lab Code: CHEM Case No.: Q2364 SAS No.: Q2364 SDG NO.: Q2364

Continuing Calib Date: 06/20/2025 Initial Calibration Date(s): 06/17/2025 06/17/2025

Continuing Calib Time: 15:56 Initial Calibration Time(s): 10:04 20:10

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	4.87	4.87	4.77	4.97	0.00
Aroclor-1248-2 (2)	5.10	5.10	5.00	5.20	0.00
Aroclor-1248-3 (3)	5.15	5.15	5.05	5.25	0.01
Aroclor-1248-4 (4)	5.32	5.32	5.22	5.42	0.00
Aroclor-1248-5 (5)	5.71	5.71	5.61	5.81	0.00
Tetrachloro-m-xylene	3.79	3.79	3.69	3.89	0.00
Decachlorobiphenyl	8.80	8.80	8.70	8.90	0.00