

CALIBRATION VERIFICATION SUMMARY

Contract: GENV01

Lab Code: CHEM Case No.: Q2364 SAS No.: Q2364 SDG NO.: Q2364

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL01 Date Analyzed: 06/20/2025

Lab Sample No.: AR1660CCC500 Data File : PP073117.D Time Analyzed: 09:16

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.643	5.546	5.746	469.830	500.000	-6.0
Aroclor-1016-2	5.664	5.567	5.767	512.290	500.000	2.5
Aroclor-1016-3	5.726	5.629	5.829	503.120	500.000	0.6
Aroclor-1016-4	5.824	5.727	5.927	526.750	500.000	5.4
Aroclor-1016-5	6.116	6.019	6.219	498.320	500.000	-0.3
Aroclor-1260-1	7.233	7.136	7.336	519.110	500.000	3.8
Aroclor-1260-2	7.487	7.389	7.589	506.380	500.000	1.3
Aroclor-1260-3	7.845	7.747	7.947	513.930	500.000	2.8
Aroclor-1260-4	8.069	7.971	8.171	529.950	500.000	6.0
Aroclor-1260-5	8.387	8.288	8.488	518.800	500.000	3.8
Decachlorobiphenyl	10.184	10.087	10.287	52.820	50.000	5.6
Tetrachloro-m-xylene	4.491	4.394	4.594	49.340	50.000	-1.3

CALIBRATION VERIFICATION SUMMARY

Contract: GENV01

Lab Code: CHEM Case No.: Q2364 SAS No.: Q2364 SDG NO.: Q2364

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL01 Date Analyzed: 06/20/2025

Lab Sample No.: AR1660CCC500 Data File : PP073117.D Time Analyzed: 09:16

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.866	4.768	4.968	459.740	500.000	-8.1
Aroclor-1016-2	4.884	4.786	4.986	465.280	500.000	-6.9
Aroclor-1016-3	5.061	4.962	5.162	470.750	500.000	-5.9
Aroclor-1016-4	5.102	5.004	5.204	470.970	500.000	-5.8
Aroclor-1016-5	5.316	5.218	5.418	475.800	500.000	-4.8
Aroclor-1260-1	6.346	6.249	6.449	452.490	500.000	-9.5
Aroclor-1260-2	6.535	6.438	6.638	454.940	500.000	-9.0
Aroclor-1260-3	6.688	6.589	6.789	460.220	500.000	-8.0
Aroclor-1260-4	7.157	7.060	7.260	458.490	500.000	-8.3
Aroclor-1260-5	7.400	7.302	7.502	457.330	500.000	-8.5
Decachlorobiphenyl	8.796	8.697	8.897	50.260	50.000	0.5
Tetrachloro-m-xylene	3.786	3.688	3.888	46.960	50.000	-6.1