

CALIBRATION VERIFICATION SUMMARY

Contract: GENV01

Lab Code: CHEM Case No.: Q2364 SAS No.: Q2364 SDG NO.: Q2364

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL02 Date Analyzed: 06/20/2025

Lab Sample No.: AR1660CCC500 Data File : PP073132.D Time Analyzed: 15:23

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.644	5.546	5.746	468.760	500.000	-6.2
Aroclor-1016-2	5.666	5.567	5.767	504.590	500.000	0.9
Aroclor-1016-3	5.728	5.629	5.829	502.440	500.000	0.5
Aroclor-1016-4	5.825	5.727	5.927	527.390	500.000	5.5
Aroclor-1016-5	6.118	6.019	6.219	491.830	500.000	-1.6
Aroclor-1260-1	7.235	7.136	7.336	517.500	500.000	3.5
Aroclor-1260-2	7.488	7.389	7.589	495.250	500.000	-1.0
Aroclor-1260-3	7.847	7.747	7.947	502.020	500.000	0.4
Aroclor-1260-4	8.070	7.971	8.171	517.610	500.000	3.5
Aroclor-1260-5	8.388	8.288	8.488	507.960	500.000	1.6
Decachlorobiphenyl	10.185	10.087	10.287	52.640	50.000	5.3
Tetrachloro-m-xylene	4.493	4.394	4.594	48.680	50.000	-2.6

CALIBRATION VERIFICATION SUMMARY

Contract: GENV01

Lab Code: CHEM Case No.: Q2364 SAS No.: Q2364 SDG NO.: Q2364

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL02 Date Analyzed: 06/20/2025

Lab Sample No.: AR1660CCC500 Data File : PP073132.D Time Analyzed: 15:23

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.866	4.768	4.968	461.520	500.000	-7.7
Aroclor-1016-2	4.884	4.786	4.986	470.410	500.000	-5.9
Aroclor-1016-3	5.060	4.962	5.162	474.380	500.000	-5.1
Aroclor-1016-4	5.102	5.004	5.204	476.440	500.000	-4.7
Aroclor-1016-5	5.316	5.218	5.418	492.550	500.000	-1.5
Aroclor-1260-1	6.347	6.249	6.449	479.410	500.000	-4.1
Aroclor-1260-2	6.535	6.438	6.638	477.950	500.000	-4.4
Aroclor-1260-3	6.687	6.589	6.789	480.200	500.000	-4.0
Aroclor-1260-4	7.157	7.060	7.260	474.800	500.000	-5.0
Aroclor-1260-5	7.399	7.302	7.502	471.430	500.000	-5.7
Decachlorobiphenyl	8.794	8.697	8.897	50.130	50.000	0.3
Tetrachloro-m-xylene	3.786	3.688	3.888	48.140	50.000	-3.7