

CALIBRATION VERIFICATION SUMMARY

Contract: GENV01

Lab Code: CHEM Case No.: Q2364 SAS No.: Q2364 SDG NO.: Q2364

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL03 Date Analyzed: 06/20/2025

Lab Sample No.: AR1660CCC500 Data File : PP073147.D Time Analyzed: 20:33

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.642	5.546	5.746	454.530	500.000	-9.1
Aroclor-1016-2	5.663	5.567	5.767	484.410	500.000	-3.1
Aroclor-1016-3	5.725	5.629	5.829	479.680	500.000	-4.1
Aroclor-1016-4	5.823	5.727	5.927	498.640	500.000	-0.3
Aroclor-1016-5	6.115	6.019	6.219	469.870	500.000	-6.0
Aroclor-1260-1	7.232	7.136	7.336	501.890	500.000	0.4
Aroclor-1260-2	7.486	7.389	7.589	470.210	500.000	-6.0
Aroclor-1260-3	7.843	7.747	7.947	491.330	500.000	-1.7
Aroclor-1260-4	8.067	7.971	8.171	508.370	500.000	1.7
Aroclor-1260-5	8.384	8.288	8.488	499.520	500.000	-0.1
Decachlorobiphenyl	10.182	10.087	10.287	50.120	50.000	0.2
Tetrachloro-m-xylene	4.491	4.394	4.594	47.120	50.000	-5.8

CALIBRATION VERIFICATION SUMMARY

Contract: GENV01

Lab Code: CHEM Case No.: Q2364 SAS No.: Q2364 SDG NO.: Q2364

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL03 Date Analyzed: 06/20/2025

Lab Sample No.: AR1660CCC500 Data File : PP073147.D Time Analyzed: 20:33

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.866	4.768	4.968	442.290	500.000	-11.5
Aroclor-1016-2	4.884	4.786	4.986	450.070	500.000	-10.0
Aroclor-1016-3	5.061	4.962	5.162	449.590	500.000	-10.1
Aroclor-1016-4	5.102	5.004	5.204	449.020	500.000	-10.2
Aroclor-1016-5	5.316	5.218	5.418	458.800	500.000	-8.2
Aroclor-1260-1	6.347	6.249	6.449	446.790	500.000	-10.6
Aroclor-1260-2	6.536	6.438	6.638	465.130	500.000	-7.0
Aroclor-1260-3	6.688	6.589	6.789	451.030	500.000	-9.8
Aroclor-1260-4	7.158	7.060	7.260	450.160	500.000	-10.0
Aroclor-1260-5	7.400	7.302	7.502	457.340	500.000	-8.5
Decachlorobiphenyl	8.795	8.697	8.897	50.810	50.000	1.6
Tetrachloro-m-xylene	3.787	3.688	3.888	46.400	50.000	-7.2