

## **Report of Analysis**

Client:	G Environn	nental				Date Collected:	06/18/25		
Project:	Ave B					Date Received:	06/18/25		
Client Sample ID:	GAV1B					SDG No.:	Q2364		
Lab Sample ID:	Q2364-02					Matrix:	SOIL		
Analytical Method	-					% Solid:	85.5	Decant	ted.
2		<b>T</b> T :/							iou.
Sample Wt/Vol:	30.02	Units: g				Final Vol:	10000	uL	
Soil Aliquot Vol:		uL				Test:	PCB		
Extraction Type:						Injection Volume :			
GPC Factor :	1.0	PH	:						
Prep Method :	SW3541B								
File ID/Qc Batch: Dilution:			Prep Date			Date Analyzed	Prep Batch ID		
			-			-	PB168558		
PP073130.D	1		06/	20/25 08:30		06/20/25 14:01	PB16	8558	
CAS Number	Parameter		Conc.	Qualifier	MDL		LOQ / CR	QL U	Inits(Dry Weigh
TARGETS									
12674-11-2	Aroclor-1016		4.60	U	4.60		1	9.9	ug/kg
11104-28-2	Aroclor-1221		4.70	U	4.70		1	9.9	ug/kg
11141-16-5	Aroclor-1232		4.30	U	4.30		1	9.9	ug/kg
53469-21-9	Aroclor-1242		4.70	U	4.70		1	9.9	ug/kg
12672-29-6	Aroclor-1248		6.90	U	6.90		1	9.9	ug/kg
11097-69-1	Aroclor-1254		20.7		3.80		1	9.9	ug/kg
37324-23-5	Aroclor-1262		5.90	U	5.90		1	9.9	ug/kg
11100-14-4	Aroclor-1268		4.20	U	4.20		1	9.9	ug/kg
11096-82-5	Aroclor-1260		3.80	U	3.80		1	9.9	ug/kg
SURROGATES									
								0 60 6	CDIZ 20
877-09-8	Tetrachloro-m-xyle	ene	21.1		30 (32)	- 150 (144)	10	06%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates > 25% difference for detected

concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

- J = Estimated Value
- B = Analyte Found in Associated Method Blank
- N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

 $\mathbf{S}=\mathbf{Indicates}$  estimated value where valid five-point calibration

was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit