

Surrogate Summary

SDG No.: **Q2364**

Client: **G Environmental**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PP072990.D	PIBLK-PP072990.D	Tetrachloro-m-xylene	1	20	17.3	86		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	18.3	91		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	18.0	90		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	17.2	86		70 (60)	130 (140)
I.BLK-PP073121.D	PIBLK-PP073121.D	Tetrachloro-m-xylene	1	20	16.1	81		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	17.4	87		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	15.7	79		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	17.3	87		70 (60)	130 (140)
PB168558BL	PB168558BL	Tetrachloro-m-xylene	1	20	20.2	101		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	22.9	114		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	20.5	103		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	22.4	112		30 (32)	150 (175)
PB168558BS	PB168558BS	Tetrachloro-m-xylene	1	20	20.4	102		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	22.2	111		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	19.1	95		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	21.3	107		30 (32)	150 (175)
Q2362-01MS	TP-11MS	Tetrachloro-m-xylene	1	20	21.6	108		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	21.8	109		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	20.0	100		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	21.1	105		30 (32)	150 (175)
Q2362-01MSD	TP-11MSD	Tetrachloro-m-xylene	1	20	19.2	96		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	20.6	103		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	17.5	88		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	20.0	100		30 (32)	150 (175)
Q2364-01	GAV1A	Tetrachloro-m-xylene	1	20	20.2	101		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	19.4	97		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	20.1	100		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	18.6	93		30 (32)	150 (175)
Q2364-02	GAV1B	Tetrachloro-m-xylene	1	20	20.8	104		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	19.0	95		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	21.1	106		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	16.7	84		30 (32)	150 (175)
Q2364-03	GAV2A	Tetrachloro-m-xylene	1	20	20.6	103		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	20.3	102		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	20.6	103		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	19.6	98		30 (32)	150 (175)
I.BLK-PP073136.D	PIBLK-PP073136.D	Tetrachloro-m-xylene	1	20	15.9	79		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	17.6	88		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	15.7	79		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	16.9	84		70 (60)	130 (140)
Q2364-04	GAV2B	Tetrachloro-m-xylene	1	20	20.0	100		30 (32)	150 (144)

() = LABORATORY INHOUSE LIMIT

Surrogate Summary

SDG No.: Q2364

Client: G Environmental

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
Q2364-04	GAV2B	Decachlorobiphenyl	1	20	19.0	95		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	19.7	99		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	18.9	95		30 (32)	150 (175)
I.BLK-PP073151.D	PIBLK-PP073151.D	Tetrachloro-m-xylene	1	20	15.6	78		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	17.2	86		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	15.9	80		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	17.5	87		70 (60)	130 (140)