

Surrogate Summary

SDG No.: **Q2364**

Client: **G Environmental**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2364-02	GAV1B	1,2-Dichloroethane-d4	50	55.9	112	70 (63)	130 (155)
		Dibromofluoromethane	50	55.0	110	70 (70)	130 (134)
		Toluene-d8	50	50.3	100	70 (74)	130 (123)
		4-Bromofluorobenzene	50	42.1	84	70 (17)	130 (146)
Q2364-04	GAV2B	1,2-Dichloroethane-d4	50	57.5	115	70 (63)	130 (155)
		Dibromofluoromethane	50	53.5	107	70 (70)	130 (134)
		Toluene-d8	50	50.5	101	70 (74)	130 (123)
		4-Bromofluorobenzene	50	45.1	90	70 (17)	130 (146)
VY0619SBL01	VY0619SBL01	1,2-Dichloroethane-d4	50	50.4	101	70 (63)	130 (155)
		Dibromofluoromethane	50	50.9	102	70 (70)	130 (134)
		Toluene-d8	50	50.5	101	70 (74)	130 (123)
		4-Bromofluorobenzene	50	40.1	80	70 (17)	130 (146)
VY0619SBS01	VY0619SBS01	1,2-Dichloroethane-d4	50	51.3	102	70 (63)	130 (155)
		Dibromofluoromethane	50	53.0	106	70 (70)	130 (134)
		Toluene-d8	50	51.4	103	70 (74)	130 (123)
		4-Bromofluorobenzene	50	47.9	96	70 (17)	130 (146)
VY0619SBSD01	VY0619SBSD01	1,2-Dichloroethane-d4	50	55.5	111	70 (63)	130 (155)
		Dibromofluoromethane	50	53.8	108	70 (70)	130 (134)
		Toluene-d8	50	53.5	107	70 (74)	130 (123)
		4-Bromofluorobenzene	50	51.5	103	70 (17)	130 (146)

