

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q2364 SAS No.: Q2364 SDG NO.: Q2364
 Lab File ID: VY022742.D Date Analyzed: 06/19/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:10
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	92724	13.347				
UPPER LIMIT	185448	13.847				
LOWER LIMIT	46362	12.847				
EPA SAMPLE NO.						
GAV1B	53874	13.35				
GAV2B	54203	13.35				
VY0619SBL01	52432	13.35				
VY0619SBS01	92139	13.35				
VY0619SBSD01	90185	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.