

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GFEL01  
 Lab Code: CHEM Case No.: Q2368 SAS No.: Q2368 SDG NO.: Q2368  
 Lab File ID: VL042647.D Date Analyzed: 06/19/2025  
 Instrument ID: MSVOA\_L Time Analyzed: 08:56  
 GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: (Y/N) N

|                | IS4<br>AREA # | RT # |  |  |  |  |
|----------------|---------------|------|--|--|--|--|
| 12 HOUR STD    | 0             | 0    |  |  |  |  |
| UPPER LIMIT    | 0             |      |  |  |  |  |
| LOWER LIMIT    | 0             |      |  |  |  |  |
| EPA SAMPLE NO. |               |      |  |  |  |  |
| SV1            | 0             | 0.00 |  |  |  |  |
| SV1DUP         | 0             | 0.00 |  |  |  |  |
| IA1            | 0             | 0.00 |  |  |  |  |
| IA1DL          | 0             | 0.00 |  |  |  |  |
| VL0619ABL01    | 0             | 0.00 |  |  |  |  |
| VL0619ABS01    | 0             | 0.00 |  |  |  |  |

IS4 =

AREA UPPER LIMIT = +40% of internal standard area

AREA LOWER LIMIT = -40% of internal standard area

RT UPPER LIMIT = +0.33 minutes of internal standard RT

RT LOWER LIMIT = -0.33 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.

\* Values outside of QC limits.