

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2369

Client: GFE LLC

Analytical Method: SWTO-15

Datafile : VL042649.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VL0619ABS01	Vinyl Chloride	10	9.30	ppbv	93			70	130	
	Heptane	10	9.50	ppbv	95			70	130	
	1,1-Dichloroethene	10	10.9	ppbv	109			70	130	
	Cyclohexane	10	9.10	ppbv	91			70	130	
	cis-1,2-Dichloroethene	10	9.70	ppbv	97			70	130	
	1,1,1-Trichloroethane	10	10.0	ppbv	100			70	130	
	2,2,4-Trimethylpentane	10	11.1	ppbv	111			70	130	
	Benzene	10	10.8	ppbv	108			70	130	
	Trichloroethene	10	10.2	ppbv	102			70	130	
	Toluene	10	10.5	ppbv	105			70	130	
	Tetrachloroethene	10	9.80	ppbv	98			70	130	
	Ethyl Benzene	10	11.1	ppbv	111			70	130	
	m/p-Xylene	20	23.2	ppbv	116			70	130	
	o-Xylene	10	11.7	ppbv	117			70	130	
	1,3,5-Trimethylbenzene	10	11.5	ppbv	115			70	130	
	1,2,4-Trimethylbenzene	10	11.6	ppbv	116			70	130	
	Naphthalene	10	11.8	ppbv	118			70	130	
	Hexane	10	9.40	ppbv	94			70	130	