

Instrument ID: IC-1

Daily Analysis Runlog For Sequence/QC Batch ID # LB136233

Review By	Iwona	Review On	6/23/2025 4:27:21 PM
Supervise By	Sohil	Supervise On	6/23/2025 4:44:35 PM
SubDirectory	LB136233	Test	Anions
STD. NAME	STD REF.#		
ICAL Standard	WP113186,WP113187,WP113188,WP113189,WP113190,WP113191,WP113192		
ICV Standard	WP113193		
CCV Standard	WP113590		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP113591		
Chk Standard	WP113194,WP113195		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	STD1	STD1	CAL1	05/22/25 11:09	All standards, samples, and	NF/IZ	OK
2	STD2	STD2	CAL2	05/22/25 11:30	QC are filtered through	NF/IZ	OK
3	STD3	STD3	CAL3	05/22/25 11:52	0.45um, filter lot W3160	NF/IZ	OK
4	STD4	STD4	CAL4	05/22/25 12:13		NF/IZ	OK
5	STD5	STD5	CAL5	05/22/25 12:35		NF/IZ	OK
6	STD6	STD6	CAL6	05/22/25 12:56		NF/IZ	OK
7	STD7	STD7	CAL7	05/22/25 13:17		NF/IZ	OK
8	ICV1	ICV1	ICV	05/22/25 13:39		NF/IZ	OK
9	ICB1	ICB1	ICB	05/22/25 14:22		NF/IZ	OK
10	CCV1	CCV1	CCV	06/20/25 10:04		NF/IZ	OK
11	CCB1	CCB1	CCB	06/20/25 10:25		NF/IZ	OK
12	LB136233BLS	LB136233BLS	MB	06/20/25 11:30		NF/IZ	OK
13	Q2371-04	RPXY42025	SAM	06/20/25 11:52		NF/IZ	OK
14	Q2371-05	BBX42025	SAM	06/20/25 12:13		NF/IZ	OK
15	Q2371-05MS	BBX42025MS	MS	06/20/25 12:35	5ml W3092	NF/IZ	OK
16	Q2371-05MSD	BBX42025MSD	MSD	06/20/25 12:56	5ml W3092	NF/IZ	OK
17	LB136233BSS	LB136233BSS	LCS	06/20/25 13:18		NF/IZ	OK
18	CCV2	CCV2	CCV	06/20/25 13:39		NF/IZ	OK

Instrument ID: IC-1

Daily Analysis Runlog For Sequence/QC Batch ID # LB136233

Review By	Iwona	Review On	6/23/2025 4:27:21 PM
Supervise By	Sohil	Supervise On	6/23/2025 4:44:35 PM
SubDirectory	LB136233	Test	Anions
STD. NAME	STD REF.#		
ICAL Standard	WP113186,WP113187,WP113188,WP113189,WP113190,WP113191,WP113192		
ICV Standard	WP113193		
CCV Standard	WP113590		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP113591		
Chk Standard	WP113194,WP113195		

19	CCB2	CCB2	CCB	06/20/25 14:01		NF/IZ	OK
----	------	------	-----	----------------	--	-------	----