

Initial and Continuing Calibration Blank Summary

Client: G Environmental

SDG No.: Q2371

Project: Buff

RunNo.: LB136233

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	05/22/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	05/22/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	05/22/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	05/22/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	05/22/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	05/22/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	05/22/2025
Sample ID: CCB1							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	06/20/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	06/20/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	06/20/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	06/20/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	06/20/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	06/20/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	06/20/2025
Sample ID: CCB2							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	06/20/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	06/20/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	06/20/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	06/20/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	06/20/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	06/20/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	06/20/2025