

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: GENV01

Lab Code: CHEM Case No.: Q2371 SAS No.: Q2371 SDG No.: Q2371

Instrument ID: MSVOA\_X Calibration Date/Time: 06/23/2025 09:03

Lab File ID: VX046804.D Init. Calib. Date(s): 06/17/2025 06/17/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:19 17:18

GC Column: DB-624UI ID: 0.18 (mm)

| COMPOUND                       | RRF   | RRF050 | MIN<br>RRF | %D     | MAX%D |
|--------------------------------|-------|--------|------------|--------|-------|
| Dichlorodifluoromethane        | 0.534 | 0.600  |            | 12.36  | 20    |
| Chloromethane                  | 0.576 | 0.638  | 0.1        | 10.76  | 20    |
| Vinyl Chloride                 | 0.614 | 0.680  |            | 10.75  | 20    |
| Bromomethane                   | 0.346 | 0.403  |            | 16.47  | 20    |
| Chloroethane                   | 0.372 | 0.421  |            | 13.17  | 20    |
| Trichlorofluoromethane         | 0.933 | 1.016  |            | 8.9    | 20    |
| 1,1,2-Trichlorotrifluoroethane | 0.572 | 0.624  |            | 9.09   | 20    |
| 1,1-Dichloroethene             | 0.552 | 0.599  |            | 8.51   | 20    |
| Acetone                        | 0.231 | 0.211  |            | -8.66  | 20    |
| Carbon Disulfide               | 1.668 | 1.680  |            | 0.72   | 20    |
| Methyl tert-butyl Ether        | 1.720 | 1.647  |            | -4.24  | 20    |
| Methyl Acetate                 | 0.586 | 0.514  |            | -12.29 | 20    |
| Methylene Chloride             | 0.641 | 0.646  |            | 0.78   | 20    |
| trans-1,2-Dichloroethene       | 0.591 | 0.616  |            | 4.23   | 20    |
| 1,1-Dichloroethane             | 1.092 | 1.134  | 0.1        | 3.85   | 20    |
| Cyclohexane                    | 1.036 | 1.024  |            | -1.16  | 20    |
| 2-Butanone                     | 0.326 | 0.263  |            | -19.33 | 20    |
| Carbon Tetrachloride           | 0.518 | 0.536  |            | 3.47   | 20    |
| cis-1,2-Dichloroethene         | 0.694 | 0.707  |            | 1.87   | 20    |
| Bromochloromethane             | 0.495 | 0.498  |            | 0.61   | 20    |
| Chloroform                     | 1.092 | 1.125  |            | 3.02   | 20    |
| 1,1,1-Trichloroethane          | 0.958 | 0.974  |            | 1.67   | 20    |
| Methylcyclohexane              | 0.628 | 0.640  |            | 1.91   | 20    |
| Benzene                        | 1.413 | 1.471  |            | 4.11   | 20    |
| 1,2-Dichloroethane             | 0.497 | 0.497  |            | 0      | 20    |
| Trichloroethene                | 0.360 | 0.372  |            | 3.33   | 20    |
| 1,2-Dichloropropane            | 0.350 | 0.367  |            | 4.86   | 20    |
| Bromodichloromethane           | 0.514 | 0.539  |            | 4.86   | 20    |
| 4-Methyl-2-Pentanone           | 0.411 | 0.354  |            | -13.87 | 20    |
| Toluene                        | 0.876 | 0.898  |            | 2.51   | 20    |
| t-1,3-Dichloropropene          | 0.477 | 0.497  |            | 4.19   | 20    |
| cis-1,3-Dichloropropene        | 0.541 | 0.568  |            | 4.99   | 20    |
| 1,1,2-Trichloroethane          | 0.327 | 0.324  |            | -0.92  | 20    |
| 2-Hexanone                     | 0.284 | 0.237  |            | -16.55 | 20    |
| Dibromochloromethane           | 0.385 | 0.396  |            | 2.86   | 20    |
| 1,2-Dibromoethane              | 0.332 | 0.322  |            | -3.01  | 20    |
| Tetrachloroethene              | 0.346 | 0.357  |            | 3.18   | 20    |
| Chlorobenzene                  | 1.104 | 1.136  | 0.3        | 2.9    | 20    |

All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: GENV01

Lab Code: CHEM Case No.: Q2371 SAS No.: Q2371 SDG No.: Q2371

Instrument ID: MSVOA\_X Calibration Date/Time: 06/23/2025 09:03

Lab File ID: VX046804.D Init. Calib. Date(s): 06/17/2025 06/17/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:19 17:18

GC Column: DB-624UI ID: 0.18 (mm)

| COMPOUND                    | RRF   | RRF050 | MIN<br>RRF | %D     | MAX%D |
|-----------------------------|-------|--------|------------|--------|-------|
| Ethyl Benzene               | 1.929 | 1.996  |            | 3.47   | 20    |
| m/p-Xylenes                 | 0.719 | 0.747  |            | 3.89   | 20    |
| o-Xylene                    | 0.673 | 0.721  |            | 7.13   | 20    |
| Styrene                     | 1.179 | 1.237  |            | 4.92   | 20    |
| Bromoform                   | 0.282 | 0.277  | 0.1        | -1.77  | 20    |
| Isopropylbenzene            | 3.682 | 3.893  |            | 5.73   | 20    |
| 1,1,2,2-Tetrachloroethane   | 1.028 | 0.984  | 0.3        | -4.28  | 20    |
| 1,3-Dichlorobenzene         | 1.728 | 1.724  |            | -0.23  | 20    |
| 1,4-Dichlorobenzene         | 1.775 | 1.715  |            | -3.38  | 20    |
| 1,2-Dichlorobenzene         | 1.615 | 1.637  |            | 1.36   | 20    |
| 1,2-Dibromo-3-Chloropropane | 0.203 | 0.170  |            | -16.26 | 20    |
| 1,2,4-Trichlorobenzene      | 1.148 | 1.160  |            | 1.04   | 20    |
| 1,2,3-Trichlorobenzene      | 1.098 | 1.090  |            | -0.73  | 20    |
| 1,2-Dichloroethane-d4       | 0.692 | 0.630  |            | -8.96  | 20    |
| Dibromofluoromethane        | 0.331 | 0.339  |            | 2.42   | 20    |
| Toluene-d8                  | 1.189 | 1.177  |            | -1.01  | 20    |
| 4-Bromofluorobenzene        | 0.443 | 0.431  |            | -2.71  | 20    |

All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.