

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: GENV01

Lab Code: CHEM Case No.: Q2371 SAS No.: Q2371 SDG No.: Q2371

Instrument ID: MSVOA_Y Calibration Date/Time: 06/26/2025 10:07

Lab File ID: VY022838.D Init. Calib. Date(s): 06/23/2025 06/23/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 13:38 15:31

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.428	0.440		2.8	20
Chloromethane	0.816	0.752	0.1	-7.84	20
Vinyl Chloride	1.020	0.980		-3.92	20
Bromomethane	0.802	0.762		-4.99	20
Chloroethane	0.686	0.667		-2.77	20
Trichlorofluoromethane	1.129	1.082		-4.16	20
1,1,2-Trichlorotrifluoroethane	0.516	0.546		5.81	20
1,1-Dichloroethene	0.506	0.545		7.71	20
Acetone	0.105	0.131		24.76	20
Carbon Disulfide	1.635	1.715		4.89	20
Methyl tert-butyl Ether	1.378	1.465		6.31	20
Methyl Acetate	0.349	0.323		-7.45	20
Methylene Chloride	0.666	0.634		-4.8	20
trans-1,2-Dichloroethene	0.578	0.608		5.19	20
1,1-Dichloroethane	1.044	1.094	0.1	4.79	20
Cyclohexane	0.959	1.000		4.28	20
2-Butanone	0.154	0.172		11.69	20
Carbon Tetrachloride	0.486	0.541		11.32	20
cis-1,2-Dichloroethene	0.672	0.699		4.02	20
Bromochloromethane	0.439	0.426		-2.96	20
Chloroform	1.076	1.110		3.26	20
1,1,1-Trichloroethane	0.929	0.993		6.89	20
Methylcyclohexane	0.596	0.679		13.93	20
Benzene	1.417	1.540		8.68	20
1,2-Dichloroethane	0.388	0.410		5.67	20
Trichloroethene	0.356	0.385		8.15	20
1,2-Dichloropropane	0.332	0.357		7.53	20
Bromodichloromethane	0.485	0.522		7.63	20
4-Methyl-2-Pentanone	0.210	0.232		10.48	20
Toluene	0.894	0.997		11.52	20
t-1,3-Dichloropropene	0.437	0.479		9.61	20
cis-1,3-Dichloropropene	0.506	0.554		9.49	20
1,1,2-Trichloroethane	0.244	0.262		7.38	20
2-Hexanone	0.143	0.165		15.39	20
Dibromochloromethane	0.315	0.338		7.3	20
1,2-Dibromoethane	0.228	0.245		7.46	20
Tetrachloroethene	0.472	0.516		9.32	20
Chlorobenzene	1.099	1.206	0.3	9.74	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: GENV01

Lab Code: CHEM Case No.: Q2371 SAS No.: Q2371 SDG No.: Q2371

Instrument ID: MSVOA_Y Calibration Date/Time: 06/26/2025 10:07

Lab File ID: VY022838.D Init. Calib. Date(s): 06/23/2025 06/23/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 13:38 15:31

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.930	2.192		13.57	20
m/p-Xylenes	0.746	0.854		14.48	20
o-Xylene	0.703	0.798		13.51	20
Styrene	1.178	1.349		14.52	20
Bromoform	0.207	0.220	0.1	6.28	20
Isopropylbenzene	3.698	4.255		15.06	20
1,1,2,2-Tetrachloroethane	0.596	0.638	0.3	7.05	20
1,3-Dichlorobenzene	1.683	1.856		10.28	20
1,4-Dichlorobenzene	1.672	1.815		8.55	20
1,2-Dichlorobenzene	1.483	1.605		8.23	20
1,2-Dibromo-3-Chloropropane	0.101	0.101		0	20
1,2,4-Trichlorobenzene	0.838	0.856		2.15	20
1,2,3-Trichlorobenzene	0.724	0.712		-1.66	20
1,2-Dichloroethane-d4	0.558	0.541		-3.05	20
Dibromofluoromethane	0.304	0.302		-0.66	20
Toluene-d8	1.207	1.223		1.33	20
4-Bromofluorobenzene	0.388	0.390		0.51	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.