

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX061725

Review By	Semsettin Yesilyurt	Review On	6/18/2025 5:28:16 PM
Supervise By	Maresh Dadoda	Supervise On	6/18/2025 6:22:48 PM
SubDirectory	VX061725	HP Acquire Method	HP Processing Method 82X061725W.M

STD. NAME	STD REF.#
Tune/Reschk	VP134389
Initial Calibration Stds	VP134390,VP134391,VP134392,VP134393,VP134394,VP134395
CCC	
Internal Standard/PEM	
ICV/I.BLK	VP134396
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX046715.D	17 Jun 2025 08:46		JC/MD	Ok
2	VSTDIC001	VSTDIC001	VX046716.D	17 Jun 2025 10:09	RRF check	JC/MD	Not Ok
3	VSTDIC001	VSTDIC001	VX046717.D	17 Jun 2025 10:58	low response	JC/MD	ReRun
4	VSTDIC005	VSTDIC005	VX046718.D	17 Jun 2025 11:19		JC/MD	Ok,M
5	VSTDIC020	VSTDIC020	VX046719.D	17 Jun 2025 13:59	LR- 10,49	JC/MD	Ok,M
6	VSTDIC050	VSTDIC050	VX046720.D	17 Jun 2025 14:20		JC/MD	Ok,M
7	VSTDIC100	VSTDIC100	VX046721.D	17 Jun 2025 14:41		JC/MD	Ok,M
8	VSTDIC150	VSTDIC150	VX046722.D	17 Jun 2025 15:02		JC/MD	Ok,M
9	IBLK	IBLK	VX046723.D	17 Jun 2025 15:23		JC/MD	Ok
10	VSTDIC001	VSTDIC001	VX046724.D	17 Jun 2025 16:20	spike error	JC/MD	Not Ok
11	VSTDIC001	VSTDIC001	VX046725.D	17 Jun 2025 17:18		JC/MD	Ok,M
12	VSTDICV050	ICVVX061725	VX046726.D	17 Jun 2025 18:00		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX062325

Review By	Mahesh Dadoda	Review On	6/24/2025 8:24:19 AM
Supervise By	Semsettin Yesilyurt	Supervise On	6/24/2025 8:28:55 AM
SubDirectory	VX062325	HP Acquire Method	HP Processing Method 82X061725W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134474
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134475,VP134476

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX046803.D	23 Jun 2025 08:05		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX046804.D	23 Jun 2025 09:03		JC/MD	Ok,M
3	VX0623MBL01	VX0623MBL01	VX046805.D	23 Jun 2025 09:31		JC/MD	Ok
4	VX0623WBL01	VX0623WBL01	VX046806.D	23 Jun 2025 09:52		JC/MD	Ok
5	VX0623MBS01	VX0623MBS01	VX046807.D	23 Jun 2025 10:13		JC/MD	Ok,M
6	VX0623MBSD01	VX0623MBSD01	VX046808.D	23 Jun 2025 10:40		JC/MD	Ok,M
7	Q2355-01	3897	VX046809.D	23 Jun 2025 11:01	cloth material	JC/MD	Ok
8	Q2354-01	3321	VX046810.D	23 Jun 2025 11:22	Need Straight Run	JC/MD	Not Ok
9	IBLK	IBLK	VX046811.D	23 Jun 2025 11:43		JC/MD	Ok
10	VX0623WBS01	VX0623WBS01	VX046812.D	23 Jun 2025 12:04		JC/MD	Ok,M
11	VX0623WBSD01	VX0623WBSD01	VX046813.D	23 Jun 2025 12:42		JC/MD	Ok,M
12	Q2354-01	3321	VX046814.D	23 Jun 2025 13:03		JC/MD	Ok,M
13	IBLK	IBLK	VX046815.D	23 Jun 2025 13:24		JC/MD	Ok
14	IBLK	IBLK	VX046816.D	23 Jun 2025 13:46		JC/MD	Ok
15	Q2371-06	GBUFF1	VX046817.D	23 Jun 2025 14:07	Need Straight Run	JC/MD	Not Ok
16	Q2371-05	BBX42025	VX046818.D	23 Jun 2025 14:28	Need 2500X	JC/MD	Dilution
17	Q2371-04	RPXY42025	VX046819.D	23 Jun 2025 14:50	Need 800X	JC/MD	Dilution
18	IBLK	IBLK	VX046820.D	23 Jun 2025 15:11		JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX062325

Review By	Maresh Dadoda	Review On	6/24/2025 8:24:19 AM
Supervise By	Semsettin Yesilyurt	Supervise On	6/24/2025 8:28:55 AM
SubDirectory	VX062325	HP Acquire Method	HP Processing Method 82X061725W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134474		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134475,VP134476		

19	IBLK	IBLK	VX046821.D	23 Jun 2025 15:33		JC/MD	Ok
20	IBLK	IBLK	VX046822.D	23 Jun 2025 15:55		JC/MD	Ok
21	Q2371-06	GBUFF1	VX046823.D	23 Jun 2025 16:16	need low level soil	JC/MD	Not Ok
22	Q2371-04DL	RPXY42025DL	VX046824.D	23 Jun 2025 16:38		JC/MD	Ok
23	Q2371-05DL	BBX42025DL	VX046825.D	23 Jun 2025 16:59		JC/MD	Ok
24	VSTDCCC050	VSTDCCC050EC	VX046826.D	23 Jun 2025 17:20		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY062325

Review By	Maresh Dadoda	Review On	6/24/2025 8:24:26 AM
Supervise By	Semsettin Yesilyurt	Supervise On	6/24/2025 8:29:32 AM
SubDirectory	VY062325	HP Acquire Method	MSVOA_Y HP Processing Method 82y062325s.m

STD. NAME	STD REF.#
Tune/Reschk	VP134461
Initial Calibration Stds	VP134462,VP134463,VP134464,VP134465,VP134466,VP134467
CCC	
Internal Standard/PEM	VP133934
ICV/I.BLK	VP134468
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022775.D	23 Jun 2025 10:17		SY/MD	Ok
2	VSTDIC005	VSTDIC005	VY022776.D	23 Jun 2025 13:38	LR- 58	SY/MD	Ok,M
3	VSTDIC010	VSTDIC010	VY022777.D	23 Jun 2025 14:00		SY/MD	Ok,M
4	VSTDIC020	VSTDIC020	VY022778.D	23 Jun 2025 14:23		SY/MD	Ok,M
5	VSTDIC050	VSTDIC050	VY022779.D	23 Jun 2025 14:46		SY/MD	Ok,M
6	VSTDIC100	VSTDIC100	VY022780.D	23 Jun 2025 15:08		SY/MD	Ok,M
7	VSTDIC150	VSTDIC150	VY022781.D	23 Jun 2025 15:31		SY/MD	Ok,M
8	VIBLK	VIBLK	VY022782.D	23 Jun 2025 15:54		SY/MD	Ok
9	VSTDICV050	ICVVY062325	VY022783.D	23 Jun 2025 16:17		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY062625

Review By	Maresh Dadoda	Review On	6/27/2025 10:46:03 AM
Supervise By	Semsettin Yesilyurt	Supervise On	6/27/2025 10:51:29 AM
SubDirectory	VY062625	HP Acquire Method	MSVOA_Y HP Processing Method 82y062325s.m
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134542		
CCC	VP134543,VP134544		
Internal Standard/PEM	VP133934		
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022837.D	26 Jun 2025 08:22		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY022838.D	26 Jun 2025 10:07		SY/MD	Ok,M
3	VY0626SBL01	VY0626SBL01	VY022839.D	26 Jun 2025 10:39		SY/MD	Ok
4	VY0626SBS01	VY0626SBS01	VY022840.D	26 Jun 2025 11:09		SY/MD	Ok,M
5	VY0626SBSD01	VY0626SBSD01	VY022841.D	26 Jun 2025 11:32		SY/MD	Ok,M
6	Q2419-01	EO-3-6-25-2025	VY022842.D	26 Jun 2025 12:10	vial-A Internal Standard Fail	SY/MD	ReRun
7	Q2405-03RE	MH-M/N-VOCRE	VY022843.D	26 Jun 2025 12:33	vial-B Internal Standard Fail	SY/MD	Confirms
8	Q2413-01	TP-35	VY022844.D	26 Jun 2025 12:56	vial-B	SY/MD	Ok
9	Q2413-06	TP-72	VY022845.D	26 Jun 2025 13:20	vial-B	SY/MD	Ok
10	Q2371-06	GBUFF1	VY022846.D	26 Jun 2025 13:43	vial-A	SY/MD	Ok
11	Q2419-01RE	EO-3-6-25-2025RE	VY022847.D	26 Jun 2025 14:07	vial-B Internal Standard Fail	SY/MD	Confirms
12	Q2429-03	TP-4-VOC	VY022848.D	26 Jun 2025 14:30	vial-A	SY/MD	Ok
13	Q2430-03	MH-E/F-VOC	VY022849.D	26 Jun 2025 14:54	vial-A	SY/MD	Ok
14	VSTDCCC050	VSTDCCC050EC	VY022850.D	26 Jun 2025 15:19		SY/MD	Ok,M

M : Manual Integration