

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q2371 SAS No.: Q2371 SDG NO.: Q2371
 Lab File ID: VX046804.D Date Analyzed: 06/23/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:03
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	146834	5.55	235057	6.76	202524	10.05
UPPER LIMIT	293668	6.05	470114	7.263	405048	10.549
LOWER LIMIT	73417	5.05	117529	6.263	101262	9.549
EPA SAMPLE NO.						
RPXY42025	167072	5.57	289861	6.77	268205	10.06
RPXY42025DL	116345	5.56	201396	6.77	184795	10.06
BBX42025	111553	5.57	195304	6.77	177101	10.06
BBX42025DL	152903	5.56	268363	6.77	246351	10.06
VX0623MBL01	114433	5.56	198594	6.76	177448	10.05
VX0623MBS01	140099	5.56	228017	6.77	204798	10.06

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.