

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
Lab Code: CHEM Case No.: Q2371 SAS No.: Q2371 SDG NO.: Q2371
Lab File ID: VY022838.D Date Analyzed: 06/26/2025
Instrument ID: MSVOA_Y Time Analyzed: 10:07
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	469939	7.71	769006	8.62	661912	11.41
UPPER LIMIT	939878	8.207	1538010	9.116	1323820	11.914
LOWER LIMIT	234970	7.207	384503	8.116	330956	10.914
EPA SAMPLE NO.						
GBUFF1	299952	7.71	553421	8.62	528624	11.41
VY0626SBL01	328943	7.71	604959	8.62	571725	11.41
VY0626SBS01	434636	7.71	740035	8.62	637859	11.42
VY0626SBSD01	431722	7.71	725692	8.62	622886	11.41

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.