

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q2371 SAS No.: Q2371 SDG NO.: Q2371
 Lab File ID: VX046804.D Date Analyzed: 06/23/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:03
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	98257	12.018				
UPPER LIMIT	196514	12.518				
LOWER LIMIT	49128.5	11.518				
EPA SAMPLE NO.						
RPXY42025	122450	12.02				
RPXY42025DL	91592	12.02				
BBX42025	74038	12.02				
BBX42025DL	124245	12.02				
VX0623MBL01	86635	12.02				
VX0623MBS01	106315	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.