

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q2371 SAS No.: Q2371 SDG NO.: Q2371
 Lab File ID: VY022838.D Date Analyzed: 06/26/2025
 Instrument ID: MSVOA_Y Time Analyzed: 10:07
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	328091	13.347				
UPPER LIMIT	656182	13.847				
LOWER LIMIT	164046	12.847				
EPA SAMPLE NO.						
GBUFF1	237063	13.35				
VY0626SBL01	249132	13.35				
VY0626SBS01	302470	13.35				
VY0626SBSD01	291493	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.