

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB136227

Review By	jaswal	Review On	6/23/2025 11:42:07 PM
Supervise By	Janvi	Supervise On	6/25/2025 8:44:30 AM

STD. NAME	STD REF.#
ICAL Standard	MP86020,MP86021,MP86022,MP86023,MP86024,MP86025
ICV Standard	MP86026
CCV Standard	MP86028
ICSA Standard	
CRI Standard	MP86030
LCS Standard	
Chk Standard	MP86027,MP86029,MP86031,MP86156

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	06/23/25 12:11		MOHAN	OK
2	S0.2	S0.2	CAL2	06/23/25 12:13		MOHAN	OK
3	S2.5	S2.5	CAL3	06/23/25 12:15		MOHAN	OK
4	S5	S5	CAL4	06/23/25 12:18		MOHAN	OK
5	S7.5	S7.5	CAL5	06/23/25 12:23		MOHAN	OK
6	S10	S10	CAL6	06/23/25 12:25		MOHAN	OK
7	ICV01	ICV01	ICV	06/23/25 12:28		MOHAN	OK
8	ICB01	ICB01	ICB	06/23/25 12:30		MOHAN	OK
9	CCV01	CCV01	CCV	06/23/25 12:33		MOHAN	OK
10	CCB01	CCB01	CCB	06/23/25 12:35		MOHAN	OK
11	CRA	CRA	CRDL	06/23/25 12:37		MOHAN	OK
12	HighStd	HighStd	HIGH STD	06/23/25 12:39		MOHAN	OK
13	ChkStd	ChkStd	SAM	06/23/25 12:42		MOHAN	OK
14	PB168567BL	PB168567BL	MB	06/23/25 12:44		MOHAN	OK
15	PB168567BS	PB168567BS	LCS	06/23/25 12:49		MOHAN	OK
16	Q2354-02	3321	SAM	06/23/25 12:51		MOHAN	OK
17	Q2355-02	3897	SAM	06/23/25 12:54		MOHAN	OK
18	Q2362-04	TP-11	SAM	06/23/25 12:56		MOHAN	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB136227

Review By	jaswal	Review On	6/23/2025 11:42:07 PM
Supervise By	Janvi	Supervise On	6/25/2025 8:44:30 AM
STD. NAME	STD REF.#		
ICAL Standard	MP86020,MP86021,MP86022,MP86023,MP86024,MP86025		
ICV Standard	MP86026		
CCV Standard	MP86028		
ICSA Standard			
CRI Standard	MP86030		
LCS Standard			
Chk Standard	MP86027,MP86029,MP86031,MP86156		

19	Q2362-08	EP-6	SAM	06/23/25 12:58		MOHAN	OK
20	Q2367-04	EP-4	SAM	06/23/25 13:00		MOHAN	OK
21	CCV02	CCV02	CCV	06/23/25 13:03		MOHAN	OK
22	CCB02	CCB02	CCB	06/23/25 13:05		MOHAN	OK
23	Q2367-08	EP-5	SAM	06/23/25 13:07		MOHAN	OK
24	Q2367-08DUP	EP-5DUP	DUP	06/23/25 13:09		MOHAN	OK
25	Q2367-08MS	EP-5MS	MS	06/23/25 13:12		MOHAN	OK
26	Q2367-08MSD	EP-5MSD	MSD	06/23/25 13:14		MOHAN	OK
27	PB168568BL	PB168568BL	MB	06/23/25 13:16		MOHAN	OK
28	PB168568BS	PB168568BS	LCS	06/23/25 13:19		MOHAN	OK
29	Q2351-01	FAIRLAWN-SUMP	SAM	06/23/25 13:21		MOHAN	OK
30	Q2352-01	292	SAM	06/23/25 13:23		MOHAN	OK
31	Q2354-04	3310	SAM	06/23/25 13:25		MOHAN	OK
32	Q2354-05	3311	SAM	06/23/25 13:28		MOHAN	OK
33	CCV03	CCV03	CCV	06/23/25 13:30		MOHAN	OK
34	CCB03	CCB03	CCB	06/23/25 13:32		MOHAN	OK
35	Q2355-05	3909	SAM	06/23/25 13:34		MOHAN	OK
36	Q2372-01	GAV1W	SAM	06/23/25 13:37		MOHAN	OK
37	Q2372-02	GAV2W	SAM	06/23/25 13:39		MOHAN	OK
38	Q2372-02DUP	GAV2WDUP	DUP	06/23/25 13:41		MOHAN	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB136227

Review By	jaswal	Review On	6/23/2025 11:42:07 PM
Supervise By	Janvi	Supervise On	6/25/2025 8:44:30 AM
STD. NAME	STD REF.#		
ICAL Standard	MP86020,MP86021,MP86022,MP86023,MP86024,MP86025		
ICV Standard	MP86026		
CCV Standard	MP86028		
ICSA Standard			
CRI Standard	MP86030		
LCS Standard			
Chk Standard	MP86027,MP86029,MP86031,MP86156		

39	Q2372-02MS	GAV2WMS	MS	06/23/25 13:47		MOHAN	OK
40	Q2372-02MSD	GAV2WMSD	MSD	06/23/25 13:49		MOHAN	OK
41	CCV04	CCV04	CCV	06/23/25 13:51		MOHAN	OK
42	CCB04	CCB04	CCB	06/23/25 13:53		MOHAN	OK
43	PB168544TB	PB168544TB	MB	06/23/25 13:56		MOHAN	OK
44	Q2367-08L	EP-5L	SD	06/23/25 13:58		MOHAN	OK
45	Q2367-08A	EP-5A	PS	06/23/25 14:03		MOHAN	OK
46	Q2372-02L	GAV2WL	SD	06/23/25 14:15		MOHAN	OK
47	Q2372-02A	GAV2WA	PS	06/23/25 14:17		MOHAN	OK
48	CCV05	CCV05	CCV	06/23/25 14:20		MOHAN	OK
49	CCB05	CCB05	CCB	06/23/25 14:22		MOHAN	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136255

Review By	Janvi	Review On	6/25/2025 7:44:51 PM
Supervise By	jaswal	Supervise On	6/30/2025 2:47:18 AM
STD. NAME	STD REF.#		
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868		
ICV Standard	MP85934		
CCV Standard	MP85875		
ICSA Standard	MP85873,MP85874		
CRI Standard			
LCS Standard			
Chk Standard	MP85876,MP85877		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	06/24/25 17:55		Jaswal	OK
2	S1	S1	CAL2	06/24/25 18:00		Jaswal	OK
3	S2	S2	CAL3	06/24/25 18:04		Jaswal	OK
4	S3	S3	CAL4	06/24/25 18:29		Jaswal	OK
5	S4	S4	CAL5	06/24/25 18:34		Jaswal	OK
6	S5	S5	CAL6	06/24/25 18:38		Jaswal	OK
7	ICV01	ICV01	ICV	06/24/25 18:43	ICV fail for Mo	Jaswal	OK
8	ICB01	ICB01	ICB	06/24/25 19:08		Jaswal	OK
9	CRI01	CRI01	CRDL	06/24/25 19:12		Jaswal	OK
10	ICSA01	ICSA01	ICSA	06/24/25 19:17		Jaswal	OK
11	ICSAB01	ICSAB01	ICSAB	06/24/25 19:21		Jaswal	OK
12	ICSADL	ICSADL	ICSA	06/24/25 19:25		Jaswal	OK
13	ICSABDL	ICSABDL	ICSAB	06/24/25 19:30		Jaswal	OK
14	CCV01	CCV01	CCV	06/24/25 19:34		Jaswal	OK
15	CCB01	CCB01	CCB	06/24/25 19:39		Jaswal	OK
16	PB168584BL	PB168584BL	MB	06/24/25 19:43		Jaswal	OK
17	PB168584BS	PB168584BS	LCS	06/24/25 19:48		Jaswal	OK
18	PB168583BL	PB168583BL	MB	06/24/25 19:52		Jaswal	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136255

Review By	Janvi	Review On	6/25/2025 7:44:51 PM
Supervise By	jaswal	Supervise On	6/30/2025 2:47:18 AM
STD. NAME	STD REF.#		
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868		
ICV Standard	MP85934		
CCV Standard	MP85875		
ICSA Standard	MP85873,MP85874		
CRI Standard			
LCS Standard			
Chk Standard	MP85876,MP85877		

19	PB168583BS	PB168583BS	LCS	06/24/25 19:56		Jaswal	OK
20	Q2380-01	72-11948	SAM	06/24/25 20:01		Jaswal	OK
21	Q2380-03	VNJ-211	SAM	06/24/25 20:05	Na oversaturated	Jaswal	Dilution
22	Q2380-05	RBR-200059	SAM	06/24/25 20:09	Na oversaturated	Jaswal	Dilution
23	Q2381-01	291431	SAM	06/24/25 20:14	Bad Injection	Jaswal	Not Ok
24	Q2381-01DUP	291431DUP	DUP	06/24/25 20:18	NOT USE	Jaswal	Not Ok
25	Q2381-01L	291431L	SD	06/24/25 20:32	NOT USE	Jaswal	Not Ok
26	CCV02	CCV02	CCV	06/24/25 20:37		Jaswal	OK
27	CCB02	CCB02	CCB	06/24/25 20:41		Jaswal	OK
28	Q2381-01MS	291431MS	MS	06/24/25 20:45	NOT USE	Jaswal	Not Ok
29	Q2381-01MSD	291431MSD	MSD	06/24/25 20:50	NOT USE	Jaswal	Not Ok
30	Q2381-01A	291431A	PS	06/24/25 20:54	NOT USE	Jaswal	Not Ok
31	Q2381-03	VNJ-207	SAM	06/24/25 20:58		Jaswal	OK
32	Q2381-05	72-11929	SAM	06/24/25 21:03		Jaswal	OK
33	Q2383-01RE	RT-5376RE	SAM	06/24/25 21:07	NOT USE	Jaswal	Not Ok
34	Q2384-01RE	OK-01-6202025RE	SAM	06/24/25 21:11	NOT USE	Jaswal	Not Ok
35	Q2386-01RE	SU-1-062025RE	SAM	06/24/25 21:16	NOT USE	Jaswal	Not Ok
36	Q2388-01	TP-12	SAM	06/24/25 21:20		Jaswal	OK
37	Q2389-01	MH-J-I	SAM	06/24/25 21:24		Jaswal	OK
38	CCV03	CCV03	CCV	06/24/25 21:29		Jaswal	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136255

Review By	Janvi	Review On	6/25/2025 7:44:51 PM
Supervise By	jaswal	Supervise On	6/30/2025 2:47:18 AM
STD. NAME	STD REF.#		
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868		
ICV Standard	MP85934		
CCV Standard	MP85875		
ICSA Standard	MP85873,MP85874		
CRI Standard			
LCS Standard			
Chk Standard	MP85876,MP85877		

39	CCB03	CCB03	CCB	06/24/25 21:33		Jaswal	OK
40	Q2372-01	GAV1W	SAM	06/24/25 21:38		Jaswal	OK
41	Q2372-02	GAV2W	SAM	06/24/25 21:42	Zn high	Jaswal	Dilution
42	Q2375-01	RW8-SP100-2025061	SAM	06/24/25 21:47		Jaswal	OK
43	Q2375-02	RW8-SP303-2025061	SAM	06/24/25 21:51		Jaswal	OK
44	Q2382-01	FMI109	SAM	06/24/25 21:56		Jaswal	OK
45	Q2385-01	A5311	SAM	06/24/25 22:00	ICV fail for Mo	Jaswal	Not Ok
46	Q2385-01DUP	A5311DUP	DUP	06/24/25 22:05	ICV fail for Mo	Jaswal	Not Ok
47	Q2385-01L	A5311L	SD	06/24/25 22:09	ICV fail for Mo	Jaswal	Not Ok
48	Q2385-01MS	A5311MS	MS	06/24/25 22:14	ICV fail for Mo	Jaswal	Not Ok
49	Q2385-01MSD	A5311MSD	MSD	06/24/25 22:18	ICV fail for Mo	Jaswal	Not Ok
50	CCV04	CCV04	CCV	06/24/25 22:22		Jaswal	OK
51	LLCCV01	LLCCV01	LLCCV	06/24/25 22:47		Jaswal	OK
52	CCB04	CCB04	CCB	06/24/25 23:01		Jaswal	OK
53	Q2385-01A	A5311A	PS	06/24/25 23:06	ICV fail for Mo	Jaswal	Not Ok
54	Q2383-01	RT-5376	SAM	06/24/25 23:10		Jaswal	OK
55	Q2384-01	OK-01-6202025	SAM	06/24/25 23:15		Jaswal	OK
56	Q2386-01	SU-1-062025	SAM	06/24/25 23:19		Jaswal	OK
57	Q2381-01RE	291431RE	SAM	06/24/25 23:24	NOT USE	Jaswal	Not Ok
58	Q2372-02DL	GAV2WDL	SAM	06/24/25 23:28	NOT USE	Jaswal	Not Ok

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136255

Review By	Janvi	Review On	6/25/2025 7:44:51 PM
Supervise By	jaswal	Supervise On	6/30/2025 2:47:18 AM
STD. NAME	STD REF.#		
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868		
ICV Standard	MP85934		
CCV Standard	MP85875		
ICSA Standard	MP85873,MP85874		
CRI Standard			
LCS Standard			
Chk Standard	MP85876,MP85877		

59	Q2375-01DL	RW8-SP100-2025061	SAM	06/24/25 23:32	NOT USE	Jaswal	Not Ok
60	Q2381-01DL	291431DL	SAM	06/24/25 23:37	NOT USE	Jaswal	Not Ok
61	Q2372-02DL2	GAV2WDL2	SAM	06/24/25 23:41	100x for Zn	Jaswal	Confirms
62	CCV05	CCV05	CCV	06/24/25 23:46		Jaswal	OK
63	CCB05	CCB05	CCB	06/24/25 23:50		Jaswal	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136273

Review By	Janvi	Review On	6/26/2025 11:24:19 AM
Supervise By	jaswal	Supervise On	6/26/2025 3:29:22 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868		
ICV Standard	MP85934		
CCV Standard	MP85875		
ICSA Standard	MP85873,MP85874		
CRI Standard	MP85897		
LCS Standard			
Chk Standard	MP85876,MP85877		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	06/25/25 12:23		Jaswal	OK
2	S1	S1	CAL2	06/25/25 12:27		Jaswal	OK
3	S2	S2	CAL3	06/25/25 12:32		Jaswal	OK
4	S3	S3	CAL4	06/25/25 12:36		Jaswal	OK
5	S4	S4	CAL5	06/25/25 12:40		Jaswal	OK
6	S5	S5	CAL6	06/25/25 12:44		Jaswal	OK
7	ICV01	ICV01	ICV	06/25/25 13:08	ICV fail for Sb,Cu (200.7) (95-105)	Jaswal	OK
8	LLICV01	LLICV01	LLICV	06/25/25 13:20		Jaswal	OK
9	ICB01	ICB01	ICB	06/25/25 14:53		Jaswal	OK
10	CRI01	CRI01	CRDL	06/25/25 15:24		Jaswal	OK
11	ICSA01	ICSA01	ICSA	06/25/25 15:29		Jaswal	OK
12	ICSAB01	ICSAB01	ICSAB	06/25/25 15:51		Jaswal	OK
13	ICSADL	ICSADL	ICSA	06/25/25 15:55		Jaswal	OK
14	ICSABDL	ICSABDL	ICSAB	06/25/25 15:59		Jaswal	OK
15	CCV01	CCV01	CCV	06/25/25 16:04		Jaswal	OK
16	CCB01	CCB01	CCB	06/25/25 16:09		Jaswal	OK
17	Q2385-01	A5311	SAM	06/25/25 16:13		Jaswal	OK
18	Q2385-01DUP	A5311DUP	DUP	06/25/25 16:17		Jaswal	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136273

Review By	Janvi	Review On	6/26/2025 11:24:19 AM
Supervise By	jaswal	Supervise On	6/26/2025 3:29:22 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868		
ICV Standard	MP85934		
CCV Standard	MP85875		
ICSA Standard	MP85873,MP85874		
CRI Standard	MP85897		
LCS Standard			
Chk Standard	MP85876,MP85877		

19	Q2385-01L	A5311L	SD	06/25/25 16:22		Jaswal	OK
20	Q2385-01MS	A5311MS	MS	06/25/25 16:26		Jaswal	OK
21	Q2385-01MSD	A5311MSD	MSD	06/25/25 16:31		Jaswal	OK
22	Q2385-01A	A5311A	PS	06/25/25 16:35		Jaswal	OK
23	Q2354-05DL	3311DL	SAM	06/25/25 16:39	5X for K	Jaswal	Confirms
24	Q2355-02	3897	SAM	06/25/25 16:44		Jaswal	OK
25	Q2362-04	TP-11	SAM	06/25/25 16:48		Jaswal	OK
26	Q2364-04DL	GAV2BDL	SAM	06/25/25 16:53	5X for Zn, Still high	Jaswal	Dilution
27	CCV02	CCV02	CCV	06/25/25 16:57		Jaswal	OK
28	CCB02	CCB02	CCB	06/25/25 17:09		Jaswal	OK
29	Q2364-04DL2	GAV2BDL2	SAM	06/25/25 17:15	10X for Zn	Jaswal	Confirms
30	Q2380-03DL	VNJ-211DL	SAM	06/25/25 17:24	5X for Na	Jaswal	Confirms
31	Q2380-05DL	RBR-200059DL	SAM	06/25/25 17:28	5X for Na	Jaswal	Confirms
32	Q2381-01	291431	SAM	06/25/25 17:33		Jaswal	Not Ok
33	Q2381-01DUP	291431DUP	DUP	06/25/25 17:37	Na oversaturated	Jaswal	Not Ok
34	Q2381-01L	291431L	SD	06/25/25 17:41	Bad Injection	Jaswal	Not Ok
35	Q2381-01MS	291431MS	MS	06/25/25 17:46	Bad Injection	Jaswal	Not Ok
36	Q2381-01MSD	291431MSD	MSD	06/25/25 17:50		Jaswal	Not Ok
37	Q2381-01A	291431A	PS	06/25/25 17:55	Bad Injection	Jaswal	Not Ok
38	CCV03	CCV03	CCV	06/25/25 18:00		Jaswal	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136273

Review By	Janvi	Review On	6/26/2025 11:24:19 AM
Supervise By	jaswal	Supervise On	6/26/2025 3:29:22 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868		
ICV Standard	MP85934		
CCV Standard	MP85875		
ICSA Standard	MP85873,MP85874		
CRI Standard	MP85897		
LCS Standard			
Chk Standard	MP85876,MP85877		

39	CCB03	CCB03	CCB	06/25/25 18:07		Jaswal	OK
40	Q2367-04	EP-4	SAM	06/25/25 18:12		Jaswal	OK
41	Q2386-02	SU-1-062025	SAM	06/25/25 18:16		Jaswal	OK
42	Q2392-01	S-1	SAM	06/25/25 18:21		Jaswal	OK
43	Q2381-01DL	291431DL	SAM	06/25/25 18:29	Not use	Jaswal	Not Ok
44	Q2381-01DUPDL	291431DUPDL	DUP	06/25/25 18:34	Not use	Jaswal	Not Ok
45	Q2381-01LDL	291431LDL	SD	06/25/25 18:38	Not use	Jaswal	Not Ok
46	Q2381-01MSDL	291431MSDL	MS	06/25/25 18:42	Not use	Jaswal	Not Ok
47	Q2381-01MSDDL	291431MSDDL	MSD	06/25/25 18:47	Not use	Jaswal	Not Ok
48	Q2381-01ADL	291431ADL	PS	06/25/25 18:51	Not use	Jaswal	Not Ok
49	CCV04	CCV04	CCV	06/25/25 18:55		Jaswal	OK
50	CCB04	CCB04	CCB	06/25/25 18:59		Jaswal	OK
51	Q2426-01	001-WILLETS-PT-BL	SAM	06/25/25 19:04		Jaswal	OK
52	Q2426-02	002-35th-Ave(May)	SAM	06/25/25 19:09		Jaswal	OK
53	Q2388-04	TP-12	SAM	06/25/25 19:13		Jaswal	OK
54	Q2389-04	MH-J-I	SAM	06/25/25 19:18	Bad Injection	Jaswal	Not Ok
55	Q2391-01	AUD-1623	SAM	06/25/25 19:22		Jaswal	OK
56	Q2394-04	MH-K/L	SAM	06/25/25 19:29	Bad Injection	Jaswal	Not Ok
57	Q2399-04	TP-13	SAM	06/25/25 19:34		Jaswal	OK
58	Q2399-08	EP-7	SAM	06/25/25 19:39		Jaswal	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136273

Review By	Janvi	Review On	6/26/2025 11:24:19 AM
Supervise By	jaswal	Supervise On	6/26/2025 3:29:22 PM

STD. NAME	STD REF.#
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868
ICV Standard	MP85934
CCV Standard	MP85875
ICSA Standard	MP85873,MP85874
CRI Standard	MP85897
LCS Standard	
Chk Standard	MP85876,MP85877

59	Q2405-04	MH-M/N	SAM	06/25/25 19:43		Jaswal	OK
60	PB168571TB	PB168571TB	MB	06/25/25 19:48		Jaswal	OK
61	CCV05	CCV05	CCV	06/25/25 19:52		Jaswal	OK
62	CCB05	CCB05	CCB	06/25/25 19:57		Jaswal	OK
63	PB168613BL	PB168613BL	MB	06/25/25 20:01	CCB fail for Cr	Jaswal	Not Ok
64	PB168613BS	PB168613BS	LCS	06/25/25 20:06	CCB fail for Cr	Jaswal	Not Ok
65	Q2409-02	COP-SOIL-PILE	SAM	06/25/25 20:10	CCB fail for Cr	Jaswal	Not Ok
66	Q2409-02DUP	COP-SOIL-PILEDUP	DUP	06/25/25 20:14	CCB fail for Cr	Jaswal	Not Ok
67	Q2409-02L	COP-SOIL-PILEL	SD	06/25/25 20:19	CCB fail for Cr	Jaswal	Not Ok
68	Q2409-02MS	COP-SOIL-PILEMS	MS	06/25/25 20:23	CCB fail for Cr	Jaswal	Not Ok
69	Q2409-02MSD	COP-SOIL-PILEMSD	MSD	06/25/25 20:28	CCB fail for Cr	Jaswal	Not Ok
70	Q2409-02A	COP-SOIL-PILEA	PS	06/25/25 20:32	CCB fail for Cr	Jaswal	Not Ok
71	Q2381-01DL2	291431DL2	SAM	06/25/25 20:37	CCB fail for Cr	Jaswal	Not Ok
72	Q2381-01DUPDL2	291431DUPDL2	DUP	06/25/25 20:41	CCB fail for Cr	Jaswal	Not Ok
73	CCV06	CCV06	CCV	06/25/25 20:46		Jaswal	OK
74	CCB06	CCB06	CCB	06/25/25 20:50	Fail for Cr	Jaswal	OK
75	Q2381-01LDL2	291431LDL2	SD	06/25/25 20:54	CCB fail for Cr	Jaswal	Not Ok
76	Q2381-01MSDL2	291431MSDL2	MS	06/25/25 20:59	CCB fail for Cr	Jaswal	Not Ok
77	Q2381-01MSDDL2	291431MSDDL2	MSD	06/25/25 21:03	CCB fail for Cr	Jaswal	Not Ok
78	Q2381-01ADL2	291431ADL2	PS	06/25/25 21:07	CCB fail for Cr	Jaswal	Not Ok

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136273

Review By	Janvi	Review On	6/26/2025 11:24:19 AM
Supervise By	jaswal	Supervise On	6/26/2025 3:29:22 PM

STD. NAME	STD REF.#
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868
ICV Standard	MP85934
CCV Standard	MP85875
ICSA Standard	MP85873,MP85874
CRI Standard	MP85897
LCS Standard	
Chk Standard	MP85876,MP85877

79	CCV07	CCV07	CCV	06/25/25 21:12		Jaswal	OK
80	CCB07	CCB07	CCB	06/25/25 21:16		Jaswal	OK